

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: ROYF02
 Lab Code: ACE SDG NO.: Q3447
 Lab File ID: VY023647.D Date Analyzed: 11/03/2025
 Instrument ID: MSVOA_Y Time Analyzed: 09:25
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS4 AREA #	RT #				
12 HOUR STD	317717	13.347				
UPPER LIMIT	635434	13.847				
LOWER LIMIT	158859	12.847				
EPA SAMPLE NO.						
P001-TW2-251023-01	399077	13.35				
VY1103SBL01	440158	13.35				
VY1103SBS01	321134	13.35				
VY1103SBSD01	309437	13.35				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.