



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Initial and Continuing Calibration Verification

Client: Parkway Generation Operating LLC

SDG No.: Q3448

Project: Linden Generating Station - LRSA

RunNo.: LB137634

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: Cyanide	ICV1	mg/L	0.094	0.099	95	85-115	10/24/2025
Sample ID: Cyanide	CCV1	mg/L	0.24	0.25	96	90-110	10/24/2025
Sample ID: Cyanide	CCV2	mg/L	0.24	0.25	96	90-110	10/24/2025

Initial and Continuing Calibration Verification

Client: Parkway Generation Operating LLC

SDG No.: Q3448

Project: Linden Generating Station - LRSA

RunNo.: LB137643

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID:	ICV						
COD		mg/L	50.173	50	100	95-105	09/08/2025
Sample ID:	CCV1						
COD		mg/L	51.184	50	102	95-105	10/24/2025
Sample ID:	CCV2						
COD		mg/L	51.184	50	102	95-105	10/24/2025

Initial and Continuing Calibration Verification

Client: Parkway Generation Operating LLC

SDG No.: Q3448

Project: Linden Generating Station - LRSA

RunNo.: LB137647

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: Phenolics	ICV1	mg/L	1	1	100	90-110	10/24/2025
Sample ID: Phenolics	CCV1	mg/L	0.98	1	98	90-110	10/24/2025
Sample ID: Phenolics	CCV2	mg/L	0.97	1	97	90-110	10/24/2025

Initial and Continuing Calibration Verification

Client: Parkway Generation Operating LLC

SDG No.: Q3448

Project: Linden Generating Station - LRSA

RunNo.: LB137656

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID:	ICV						
Field pH		pH	7.03	7	100	90-110	10/23/2025
Sample ID:	CCV1						
Field pH		pH	7.01	7	100	90-110	10/23/2025
Sample ID:	CCV2						
Field pH		pH	7.02	7	100	90-110	10/23/2025

Initial and Continuing Calibration Verification

Client: Parkway Generation Operating LLC

SDG No.: Q3448

Project: Linden Generating Station - LRSA

RunNo.: LB137656

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
---------	-------	--------	------------	---------------	---------------------------	------------------