

### Laboratory Control Sample Summary

**Client:** Parkway Generation Operating LLC

**SDG No.:** Q3448

**Project:** Linden Generating Station - LRSA

**Run No.:** LB137623

| Analyte   | Units      | True Value | Result | Conc. Qualifier | % Recovery | Dilution Factor | Acceptance Limit %R | Analysis Date |
|-----------|------------|------------|--------|-----------------|------------|-----------------|---------------------|---------------|
| Sample ID | LB137623BS |            |        |                 |            |                 |                     |               |
| BOD5      | mg/L       | 198        | 186    |                 | 94         | 1               | 84.6-115.4          | 10/23/2025    |

### Laboratory Control Sample Summary

**Client:** Parkway Generation Operating LLC

**SDG No.:** Q3448

**Project:** Linden Generating Station - LRSA

**Run No.:** LB137628

| Analyte   | Units      | True Value | Result | Conc. Qualifier | % Recovery | Dilution Factor | Acceptance Limit %R | Analysis Date |
|-----------|------------|------------|--------|-----------------|------------|-----------------|---------------------|---------------|
| Sample ID | LB137628BS |            |        |                 |            |                 |                     |               |
| TDS       | mg/L       | 100        | 95.0   |                 | 95         | 1               | 90-110              | 10/23/2025    |

### Laboratory Control Sample Summary

**Client:** Parkway Generation Operating LLC

**SDG No.:** Q3448

**Project:** Linden Generating Station - LRSA

**Run No.:** LB137643

| Analyte   | Units      | True Value | Result | Conc. Qualifier | % Recovery | Dilution Factor | Acceptance Limit %R | Analysis Date |
|-----------|------------|------------|--------|-----------------|------------|-----------------|---------------------|---------------|
| Sample ID | LB137643BS |            |        |                 |            |                 |                     |               |
| COD       | mg/L       | 50         | 49.2   |                 | 98         | 1               | 90-110              | 10/24/2025    |

### Laboratory Control Sample Summary

**Client:** Parkway Generation Operating LLC

**SDG No.:** Q3448

**Project:** Linden Generating Station - LRSA

**Run No.:** LB137648

| Analyte   | Units      | True Value | Result | Conc. Qualifier | % Recovery | Dilution Factor | Acceptance Limit %R | Analysis Date |
|-----------|------------|------------|--------|-----------------|------------|-----------------|---------------------|---------------|
| Sample ID | LB137648BS |            |        |                 |            |                 |                     |               |
| TSS       | mg/L       | 550        | 531    |                 | 96         | 1               | 90-110              | 10/24/2025    |

### Laboratory Control Sample Summary

**Client:** Parkway Generation Operating LLC

**SDG No.:** Q3448

**Project:** Linden Generating Station - LRSA

**Run No.:** LB137655

| Analyte        | Units      | True Value | Result | Conc. Qualifier | % Recovery | Dilution Factor | Acceptance Limit %R | Analysis Date |
|----------------|------------|------------|--------|-----------------|------------|-----------------|---------------------|---------------|
| Sample ID      | LB137655BS |            |        |                 |            |                 |                     |               |
| Oil and Grease | mg/L       | 20.0       | 16.7   |                 | 84         | 1               | 78-114              | 10/27/2025    |

### Laboratory Control Sample Summary

**Client:** Parkway Generation Operating LLC

**SDG No.:** Q3448

**Project:** Linden Generating Station - LRSA

**Run No.:** LB137647

| Analyte   | Units      | True Value | Result | Conc. Qualifier | % Recovery | Dilution Factor | Acceptance Limit %R | Analysis Date |
|-----------|------------|------------|--------|-----------------|------------|-----------------|---------------------|---------------|
| Sample ID | PB170233BS |            |        |                 |            |                 |                     |               |
| Phenolics | mg/L       | 1          | 1.00   |                 | 100        | 1               | 80-120              | 10/24/2025    |

### Laboratory Control Sample Summary

**Client:** Parkway Generation Operating LLC

**SDG No.:** Q3448

**Project:** Linden Generating Station - LRSA

**Run No.:** LB137634

| Analyte   | Units      | True Value | Result | Conc. Qualifier | % Recovery | Dilution Factor | Acceptance Limit %R | Analysis Date |
|-----------|------------|------------|--------|-----------------|------------|-----------------|---------------------|---------------|
| Sample ID | PB170237BS |            |        |                 |            |                 |                     |               |
| Cyanide   | mg/L       | 0.1        | 0.096  |                 | 96         | 1               | 85-115              | 10/24/2025    |