

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** MATR02
Lab Code: ACE **SDG NO.:** Q3450
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/21/2025 10/21/2025

Client Sample No.: CCAL01 **Date Analyzed:** 10/24/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP076042.D **Time Analyzed:** 18:58

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.792	5.695	5.895	466.180	500.000	-6.8
Aroclor-1016-2	5.814	5.717	5.917	470.850	500.000	-5.8
Aroclor-1016-3	5.876	5.780	5.980	471.910	500.000	-5.6
Aroclor-1016-4	5.974	5.877	6.077	475.240	500.000	-5.0
Aroclor-1016-5	6.266	6.169	6.369	472.190	500.000	-5.6
Aroclor-1260-1	7.383	7.287	7.487	508.760	500.000	1.8
Aroclor-1260-2	7.637	7.540	7.740	477.450	500.000	-4.5
Aroclor-1260-3	7.995	7.899	8.099	498.370	500.000	-0.3
Aroclor-1260-4	8.222	8.127	8.327	531.900	500.000	6.4
Aroclor-1260-5	8.547	8.452	8.652	490.190	500.000	-2.0
Decachlorobiphenyl	10.411	10.319	10.519	49.890	50.000	-0.2
Tetrachloro-m-xylene	4.640	4.542	4.742	45.590	50.000	-8.8

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** MATR02
Lab Code: ACE **SDG NO.:** Q3450
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/21/2025 10/21/2025

Client Sample No.: CCAL01 **Date Analyzed:** 10/24/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP076042.D **Time Analyzed:** 18:58

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.874	4.778	4.978	465.970	500.000	-6.8
Aroclor-1016-2	4.933	4.836	5.036	486.720	500.000	-2.7
Aroclor-1016-3	5.052	4.954	5.154	475.940	500.000	-4.8
Aroclor-1016-4	5.093	4.996	5.196	498.570	500.000	-0.3
Aroclor-1016-5	5.306	5.210	5.410	466.880	500.000	-6.6
Aroclor-1260-1	6.334	6.238	6.438	497.200	500.000	-0.6
Aroclor-1260-2	6.521	6.425	6.625	510.950	500.000	2.2
Aroclor-1260-3	6.674	6.578	6.778	514.410	500.000	2.9
Aroclor-1260-4	7.142	7.047	7.247	563.770	500.000	12.8
Aroclor-1260-5	7.383	7.287	7.487	531.590	500.000	6.3
Decachlorobiphenyl	8.776	8.682	8.882	47.410	50.000	-5.2
Tetrachloro-m-xylene	3.779	3.681	3.881	46.610	50.000	-6.8