

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** MATR02
Lab Code: ACE **SDG NO.:** Q3450
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/21/2025 10/21/2025

Client Sample No.: CCAL02 **Date Analyzed:** 10/25/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP076057.D **Time Analyzed:** 00:22

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.795	5.695	5.895	471.470	500.000	-5.7
Aroclor-1016-2	5.816	5.717	5.917	468.300	500.000	-6.3
Aroclor-1016-3	5.879	5.780	5.980	468.920	500.000	-6.2
Aroclor-1016-4	5.976	5.877	6.077	474.310	500.000	-5.1
Aroclor-1016-5	6.268	6.169	6.369	471.100	500.000	-5.8
Aroclor-1260-1	7.386	7.287	7.487	527.420	500.000	5.5
Aroclor-1260-2	7.639	7.540	7.740	488.330	500.000	-2.3
Aroclor-1260-3	7.997	7.899	8.099	510.250	500.000	2.1
Aroclor-1260-4	8.225	8.127	8.327	539.530	500.000	7.9
Aroclor-1260-5	8.549	8.452	8.652	497.530	500.000	-0.5
Decachlorobiphenyl	10.412	10.319	10.519	50.390	50.000	0.8
Tetrachloro-m-xylene	4.643	4.542	4.742	46.220	50.000	-7.6

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** MATR02
Lab Code: ACE **SDG NO.:** Q3450
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/21/2025 10/21/2025

Client Sample No.: CCAL02 **Date Analyzed:** 10/25/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP076057.D **Time Analyzed:** 00:22

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.878	4.778	4.978	475.300	500.000	-4.9
Aroclor-1016-2	4.937	4.836	5.036	506.010	500.000	1.2
Aroclor-1016-3	5.056	4.954	5.154	486.710	500.000	-2.7
Aroclor-1016-4	5.097	4.996	5.196	501.120	500.000	0.2
Aroclor-1016-5	5.311	5.210	5.410	470.680	500.000	-5.9
Aroclor-1260-1	6.337	6.238	6.438	513.790	500.000	2.8
Aroclor-1260-2	6.524	6.425	6.625	515.860	500.000	3.2
Aroclor-1260-3	6.678	6.578	6.778	518.110	500.000	3.6
Aroclor-1260-4	7.146	7.047	7.247	568.000	500.000	13.6
Aroclor-1260-5	7.386	7.287	7.487	533.710	500.000	6.7
Decachlorobiphenyl	8.779	8.682	8.882	46.800	50.000	-6.4
Tetrachloro-m-xylene	3.782	3.681	3.881	46.920	50.000	-6.2