

DATA REPORTING QUALIFIERS- ORGANIC

For reporting results, the following “ Results Qualifiers” are used:

Value	If the result is a value greater than or equal to the detection limit, report the value
U	Indicates the compound was analyzed for but was not detected. Report the minimum detection limit for the sample with the U, i.e. “10 U”. This is not necessarily the instrument detection limit attainable for this particular sample based on any concentration or dilution that may have been required.
ND	Indicates the analyte was analyzed for, but not detected
J	Indicates an estimated value. This flag is used: (1) When estimating a concentration for a tentatively identified compound (library search hits, where a 1:1 response is assumed.) (2) When the mass spectral data indicated the identification, however the result was less than the specified detection limit greater than zero. If the detection limit was 10ug/L and a concentration of 3 ug/L was calculated report as 3 J. This flag is used when similar situation arise on any organic parameter i.e. Pest, PCB and others.
B	Indicates the analyte was found in the blank as well as the sample report as “12 B”.
E	Indicates the analyte ‘s concentration exceeds the calibrated range of the instrument for that specific analysis.
D	This flag identifies all compounds identified in an analysis at a secondary dilution factor.
P	This flag is used for Pesticide/PCB target analyte when there is >25% difference for detected concentrations between the two GC columns. The lower of the two values is reported on Form 1 and flagged with a “P”.
N	This flag indicates presumptive evidence of a compound. This is only used for tentatively identified compounds (TICs), where the identification is based on a mass spectral library search. It applies to all TIC results. For generic characterization of a TIC, such as chlorinated hydrocarbon, the flag is not used.
A	This flag indicates that a Tentatively Identified Compound is a suspected aldol-condensation product.
Q	Indicates the LCS did not meet the control limits requirements

LAB CHRONICLE

OrderID:	Q3450	OrderDate:	10/23/2025 3:53:00 PM
Client:	Matrix New World Engineering	Project:	PSEG Athenia Substation
Contact:	Christopher Pittarese	Location:	D41

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q3450-01	SS-5R-(2.0-2.5)	SOIL	PCB	8082A	10/23/25	10/24/25	10/27/25	10/23/25
Q3450-02	SS-5R-(3.5-4.0)	SOIL	PCB	8082A	10/23/25	10/24/25	10/27/25	10/23/25
Q3450-08	SS-20-(0.0-0.5)	SOIL	PCB	8082A	10/23/25	10/24/25	10/27/25	10/23/25
Q3450-09	SS-20-(2.0-2.5)	SOIL	PCB	8082A	10/23/25	10/24/25	10/27/25	10/23/25
Q3450-10	SS-20-(3.5-4.0)	SOIL	PCB	8082A	10/23/25	10/24/25	10/24/25	10/23/25
Q3450-11	SS-28-(2.0-2.5)	SOIL	PCB	8082A	10/23/25	10/24/25	10/24/25	10/23/25
Q3450-20	SS-21-(0.0-0.5)	SOIL	PCB	8082A	10/23/25	10/24/25	10/24/25	10/23/25
Q3450-20DL	SS-21-(0.0-0.5)DL	SOIL	PCB	8082A	10/23/25	10/24/25	10/27/25	10/23/25
Q3450-21	SS-21-(2.0-2.5)	SOIL	PCB	8082A	10/23/25	10/24/25	10/24/25	10/23/25
Q3450-22	SS-21-(3.5-4.0)	SOIL	PCB	8082A	10/23/25	10/24/25	10/24/25	10/23/25
Q3450-22DL	SS-21-(3.5-4.0)DL	SOIL	PCB	8082A	10/23/25	10/24/25	10/27/25	10/23/25

Hit Summary Sheet
SW-846

SDG No.: Q3450

Order ID: Q3450

Client: Matrix New World Engineering

Project ID: PSEG Athenia Substation

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Client ID :	SS-5R-(2.0-2.5)							
Q3450-01	SS-5R-(2.0-2.5)	SOIL	Aroclor-1254	45.5		3.50	18.3	ug/kg
			Total Concentration:	45.500				
Client ID :	SS-5R-(3.5-4.0)							
Q3450-02	SS-5R-(3.5-4.0)	SOIL	Aroclor-1254	11.3 J		3.70	19.6	ug/kg
			Total Concentration:	11.300				
Client ID :	SS-20-(0.0-0.5)							
Q3450-08	SS-20-(0.0-0.5)	SOIL	Aroclor-1254	174		3.40	18.2	ug/kg
			Total Concentration:	174.000				
Client ID :	SS-20-(2.0-2.5)							
Q3450-09	SS-20-(2.0-2.5)	SOIL	Aroclor-1254	11.3 J		3.70	19.6	ug/kg
			Total Concentration:	11.300				
Client ID :	SS-20-(3.5-4.0)							
Q3450-10	SS-20-(3.5-4.0)	SOIL	Aroclor-1254	55.2		3.80	20.2	ug/kg
			Total Concentration:	55.200				
Client ID :	SS-21-(0.0-0.5)							
Q3450-20	SS-21-(0.0-0.5)	SOIL	Aroclor-1254	593 E		3.50	18.5	ug/kg
			Total Concentration:	593.000				
Client ID :	SS-21-(0.0-0.5)DL							
Q3450-20DL	SS-21-(0.0-0.5)DL	SOIL	Aroclor-1254	509 D		17.5	92.5	ug/kg
			Total Concentration:	509.000				
Client ID :	SS-21-(3.5-4.0)							
Q3450-22	SS-21-(3.5-4.0)	SOIL	Aroclor-1254	516 E		3.80	20.0	ug/kg
			Total Concentration:	516.000				
Client ID :	SS-21-(3.5-4.0)DL							
Q3450-22DL	SS-21-(3.5-4.0)DL	SOIL	Aroclor-1254	503 D		7.60	40.1	ug/kg
			Total Concentration:	503.000				



QC SUMMARY

Surrogate Summary

SDG No.: **Q3450**

Client: **Matrix New World Engineering**

Analytical Method: **8082A**

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Recovery(%)	Qual	Limits(%)	
								Low	High
I.BLK-PP075887.D	PIBLK-PP075887.D	Tetrachloro-m-xyl	1	20	19.9	100		70 (60)	130 (140)
		Decachlorobiphen	1	20	20.4	102		70 (60)	130 (140)
		Tetrachloro-m-xyl	2	20	18.8	94		70 (60)	130 (140)
		Decachlorobiphen	2	20	19.3	96		70 (60)	130 (140)
I.BLK-PP076046.D	PIBLK-PP076046.D	Tetrachloro-m-xyl	1	20	20.6	103		70 (60)	130 (140)
		Decachlorobiphen	1	20	23.6	118		70 (60)	130 (140)
		Tetrachloro-m-xyl	2	20	17.6	88		70 (60)	130 (140)
		Decachlorobiphen	2	20	20.1	101		70 (60)	130 (140)
Q3450-10	SS-20-(3.5-4.0)	Tetrachloro-m-xyl	1	20	24.1	121		30 (32)	150 (144)
		Decachlorobiphen	1	20	25.8	129		30 (32)	150 (175)
		Tetrachloro-m-xyl	2	20	23.0	115		30 (32)	150 (144)
		Decachlorobiphen	2	20	22.4	112		30 (32)	150 (175)
Q3450-11	SS-28-(2.0-2.5)	Tetrachloro-m-xyl	1	20	23.3	116		30 (32)	150 (144)
		Decachlorobiphen	1	20	25.0	125		30 (32)	150 (175)
		Tetrachloro-m-xyl	2	20	22.7	113		30 (32)	150 (144)
		Decachlorobiphen	2	20	22.0	110		30 (32)	150 (175)
Q3450-20	SS-21-(0.0-0.5)	Tetrachloro-m-xyl	1	20	24.0	120		30 (32)	150 (144)
		Decachlorobiphen	1	20	25.0	125		30 (32)	150 (175)
		Tetrachloro-m-xyl	2	20	22.6	113		30 (32)	150 (144)
		Decachlorobiphen	2	20	21.8	109		30 (32)	150 (175)
Q3450-21	SS-21-(2.0-2.5)	Tetrachloro-m-xyl	1	20	24.5	123		30 (32)	150 (144)
		Decachlorobiphen	1	20	26.4	132		30 (32)	150 (175)
		Tetrachloro-m-xyl	2	20	22.6	113		30 (32)	150 (144)
		Decachlorobiphen	2	20	23.4	117		30 (32)	150 (175)
Q3450-22	SS-21-(3.5-4.0)	Tetrachloro-m-xyl	1	20	20.1	100		30 (32)	150 (144)
		Decachlorobiphen	1	20	17.2	86		30 (32)	150 (175)
		Tetrachloro-m-xyl	2	20	19.0	95		30 (32)	150 (144)
		Decachlorobiphen	2	20	15.1	76		30 (32)	150 (175)
I.BLK-PP076061.D	PIBLK-PP076061.D	Tetrachloro-m-xyl	1	20	20.9	104		70 (60)	130 (140)
		Decachlorobiphen	1	20	23.7	119		70 (60)	130 (140)
		Tetrachloro-m-xyl	2	20	18.2	91		70 (60)	130 (140)
		Decachlorobiphen	2	20	21.0	105		70 (60)	130 (140)
I.BLK-PP076067.D	PIBLK-PP076067.D	Tetrachloro-m-xyl	1	20	22.1	110		70 (60)	130 (140)
		Decachlorobiphen	1	20	22.4	112		70 (60)	130 (140)
		Tetrachloro-m-xyl	2	20	18.9	95		70 (60)	130 (140)
		Decachlorobiphen	2	20	18.2	91		70 (60)	130 (140)
PB170236BL	PB170236BL	Tetrachloro-m-xyl	1	20	27.3	137		30 (32)	150 (144)
		Decachlorobiphen	1	20	29.0	145		30 (32)	150 (175)
		Tetrachloro-m-xyl	2	20	26.1	130		30 (32)	150 (144)
		Decachlorobiphen	2	20	25.2	126		30 (32)	150 (175)
PB170236BS	PB170236BS	Tetrachloro-m-xyl	1	20	26.7	134		30 (32)	150 (144)

() = LABORATORY INHOUSE LIMIT

Surrogate Summary

SDG No.: Q3450

Client: Matrix New World Engineering

Analytical Method: 8082A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Recovery(%)	Qual	Limits(%)	
								Low	High
PB170236BS	PB170236BS	Decachlorobiphen	1	20	28.6	143		30 (32)	150 (175)
		Tetrachloro-m-xyl	2	20	21.3	107		30 (32)	150 (144)
Q3450-20DL	SS-21-(0.0-0.5)DL	Decachlorobiphen	2	20	24.3	121		30 (32)	150 (175)
		Tetrachloro-m-xyl	1	20	21.4	107		30 (32)	150 (144)
		Decachlorobiphen	1	20	24.8	124		30 (32)	150 (175)
		Tetrachloro-m-xyl	2	20	20.5	103		30 (32)	150 (144)
		Decachlorobiphen	2	20	19.4	97		30 (32)	150 (175)
Q3450-22DL	SS-21-(3.5-4.0)DL	Tetrachloro-m-xyl	1	20	21.6	108		30 (32)	150 (144)
		Decachlorobiphen	1	20	17.5	88		30 (32)	150 (175)
		Tetrachloro-m-xyl	2	20	19.7	99		30 (32)	150 (144)
		Decachlorobiphen	2	20	13.8	69		30 (32)	150 (175)
I.BLK-PP076082.D	PIBLK-PP076082.D	Tetrachloro-m-xyl	1	20	18.9	95		70 (60)	130 (140)
		Decachlorobiphen	1	20	20.4	102		70 (60)	130 (140)
		Tetrachloro-m-xyl	2	20	17.1	86		70 (60)	130 (140)
		Decachlorobiphen	2	20	16.7	83		70 (60)	130 (140)
I.BLK-PQ070972.D	PIBLK-PQ070972.D	Tetrachloro-m-xyl	1	20	18.3	91		70 (60)	130 (140)
		Decachlorobiphen	1	20	19.0	95		70 (60)	130 (140)
		Tetrachloro-m-xyl	2	20	19.1	96		70 (60)	130 (140)
		Decachlorobiphen	2	20	19.4	97		70 (60)	130 (140)
I.BLK-PQ071150.D	PIBLK-PQ071150.D	Tetrachloro-m-xyl	1	20	22.5	113		70 (60)	130 (140)
		Decachlorobiphen	1	20	25.7	129		70 (60)	130 (140)
		Tetrachloro-m-xyl	2	20	19.9	100		70 (60)	130 (140)
		Decachlorobiphen	2	20	24.5	122		70 (60)	130 (140)
Q3450-01	SS-5R-(2.0-2.5)	Tetrachloro-m-xyl	1	20	27.2	136		30 (32)	150 (144)
		Decachlorobiphen	1	20	28.0	140		30 (32)	150 (175)
		Tetrachloro-m-xyl	2	20	24.5	122		30 (32)	150 (144)
		Decachlorobiphen	2	20	27.0	135		30 (32)	150 (175)
Q3450-02	SS-5R-(3.5-4.0)	Tetrachloro-m-xyl	1	20	27.0	135		30 (32)	150 (144)
		Decachlorobiphen	1	20	25.4	127		30 (32)	150 (175)
		Tetrachloro-m-xyl	2	20	23.8	119		30 (32)	150 (144)
		Decachlorobiphen	2	20	24.9	124		30 (32)	150 (175)
Q3450-02MS	SS-5R-(3.5-4.0)MS	Tetrachloro-m-xyl	1	20	23.4	117		30 (32)	150 (144)
		Decachlorobiphen	1	20	22.4	112		30 (32)	150 (175)
		Tetrachloro-m-xyl	2	20	18.7	94		30 (32)	150 (144)
		Decachlorobiphen	2	20	20.7	104		30 (32)	150 (175)
Q3450-02MSD	SS-5R-(3.5-4.0)MSD	Tetrachloro-m-xyl	1	20	25.7	128		30 (32)	150 (144)
		Decachlorobiphen	1	20	24.0	120		30 (32)	150 (175)
		Tetrachloro-m-xyl	2	20	22.2	111		30 (32)	150 (144)
		Decachlorobiphen	2	20	22.9	114		30 (32)	150 (175)
Q3450-08	SS-20-(0.0-0.5)	Tetrachloro-m-xyl	1	20	26.8	134		30 (32)	150 (144)
		Decachlorobiphen	1	20	28.1	140		30 (32)	150 (175)

() = LABORATORY INHOUSE LIMIT

Surrogate Summary

SDG No.: Q3450

Client: Matrix New World Engineering

Analytical Method: 8082A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Recovery(%)	Qual	Limits(%)	
								Low	High
Q3450-08	SS-20-(0.0-0.5)	Tetrachloro-m-xyl	2	20	25.1	125		30 (32)	150 (144)
		Decachlorobiphen	2	20	26.6	133		30 (32)	150 (175)
Q3450-09	SS-20-(2.0-2.5)	Tetrachloro-m-xyl	1	20	27.2	136		30 (32)	150 (144)
		Decachlorobiphen	1	20	27.4	137		30 (32)	150 (175)
		Tetrachloro-m-xyl	2	20	24.9	125		30 (32)	150 (144)
		Decachlorobiphen	2	20	26.3	132		30 (32)	150 (175)
I.BLK-PQ071165.D	PIBLK-PQ071165.D	Tetrachloro-m-xyl	1	20	20.3	102		70 (60)	130 (140)
		Decachlorobiphen	1	20	23.5	117		70 (60)	130 (140)
		Tetrachloro-m-xyl	2	20	18.4	92		70 (60)	130 (140)
		Decachlorobiphen	2	20	22.2	111		70 (60)	130 (140)

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.:	<u>Q3450</u>	Analytical Method:	<u>8082A</u>
Client:	<u>Matrix New World Engineering</u>	DataFile :	<u>PQ071153.D</u>

	Parameter	Spike	Sample		Units	Rec	Rec		RPD		Limits		
			Result	Result			Qual	RPD	Qual	Low	High	RPD	
Lab Sample ID:	Q3450-02MS	Client Sample ID:		SS-5R-(3.5-4.0)MS									
	(Column 1)												
	AR1016	192.7	0	121	ug/kg	63				40 (55)	140 (146)		
	AR1260	192.7	0	145	ug/kg	75				40 (54)	140 (119)		
Lab Sample ID:	Q3450-02MS	Client Sample ID:		SS-5R-(3.5-4.0)MS									
	(Column 2)												
	AR1016	192.7	0	118	ug/kg	61				40 (55)	140 (146)		
	AR1260	192.7	0	123	ug/kg	64				40 (54)	140 (119)		

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.:	<u>Q3450</u>	Analytical Method:	<u>8082A</u>
Client:	<u>Matrix New World Engineering</u>	DataFile :	<u>PQ071154.D</u>

		Sample					Rec	RPD		Limits		
	Parameter	Spike	Result	Result	Units	Rec	Qual	RPD	Qual	Low	High	RPD
Lab Sample ID:	Q3450-02MSD (Column 1)	Client Sample ID:		SS-5R-(3.5-4.0)MSD								
	AR1016	192.9	0	162	ug/kg	84		29		40 (55)	140 (146)	30 (15)
	AR1260	192.9	0	166	ug/kg	86		14		40 (54)	140 (119)	30 (15)
Lab Sample ID:	Q3450-02MSD (Column 2)	Client Sample ID:		SS-5R-(3.5-4.0)MSD								
	AR1016	192.9	0	143	ug/kg	74		19		40 (55)	140 (146)	30 (15)
	AR1260	192.9	0	153	ug/kg	79		21		40 (54)	140 (119)	30 (15)

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q3450</u>	Analytical Method:	<u>8082A</u>
Client:	<u>Matrix New World Engineering</u>	Datafile :	<u>PP076069.D</u>

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD	Limits		RPD
								Qual	Low	High	
PB170236BS (Column 1)	AR1016	166.6	145	ug/kg	87				40 (71)	140 (120)	
	AR1260	166.6	149	ug/kg	89				40 (65)	140 (130)	
PB170236BS (Column 2)	AR1016	166.6	127	ug/kg	76				40 (71)	140 (120)	
	AR1260	166.6	143	ug/kg	86				40 (65)	140 (130)	

4C

PESTICIDE METHOD BLANK SUMMARY

Client ID

PB170236BL

Lab Name: Alliance

Contract: MATR02

Lab Code: ACE

SDG NO.: Q3450

Lab Sample ID: PB170236BL

Lab File ID: PP076068.D

Matrix: (soil/water) Solid

Extraction: (Type) SOXH

Sulfur Cleanup: (Y/N) N

Date Extracted: 10/24/2025

Date Analyzed (1): 10/27/2025

Date Analyzed (2): 10/27/2025

Time Analyzed (1): 10:18

Time Analyzed (2): 10:18

Instrument ID (1): ECD_P

Instrument ID (2): ECD_P

GC Column (1): ZB-MR1 ID: 0.32 (mm)

GC Column (2): ZB-MR2 ID: 0.32 (mm)

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED 1	DATE ANALYZED 2
SS-20- (3.5-4.0)	Q3450-10	PP076047.D	10/24/2025	10/24/2025
SS-28- (2.0-2.5)	Q3450-11	PP076048.D	10/24/2025	10/24/2025
SS-21- (0.0-0.5)	Q3450-20	PP076049.D	10/24/2025	10/24/2025
SS-21- (2.0-2.5)	Q3450-21	PP076050.D	10/24/2025	10/24/2025
SS-21- (3.5-4.0)	Q3450-22	PP076051.D	10/24/2025	10/24/2025
PB170236BS	PB170236BS	PP076069.D	10/27/2025	10/27/2025
SS-5R- (2.0-2.5)	Q3450-01	PQ071151.D	10/27/2025	10/27/2025
SS-5R- (3.5-4.0)	Q3450-02	PQ071152.D	10/27/2025	10/27/2025
SS-5R- (3.5-4.0)MS	Q3450-02MS	PQ071153.D	10/27/2025	10/27/2025
SS-5R- (3.5-4.0)MSD	Q3450-02MSD	PQ071154.D	10/27/2025	10/27/2025
SS-20- (0.0-0.5)	Q3450-08	PQ071155.D	10/27/2025	10/27/2025
SS-20- (2.0-2.5)	Q3450-09	PQ071156.D	10/27/2025	10/27/2025

COMMENTS: _____



SAMPLE DATA

Report of Analysis

Client: Matrix New World Engineering	Date Collected: 10/23/25	
Project: PSEG Athenia Substation	Date Received: 10/23/25	
Client Sample ID: SS-5R-(2.0-2.5)	SDG No.: Q3450	
Lab Sample ID: Q3450-01	Matrix: SOIL	
Analytical Method: 8082A	% Solid: 92.9	
Sample Wt/Vol: 30.03 g	Test: PCB	Final Vol: 10000 uL
Prep Method: SW3541	Prep Date: 10/24/25	

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
12674-11-2	Aroclor-1016	4.20	U	1	4.20	18.3	ug/kg	10/27/25 09:26	PB170236
11104-28-2	Aroclor-1221	4.30	U	1	4.30	18.3	ug/kg	10/27/25 09:26	PB170236
11141-16-5	Aroclor-1232	4.00	U	1	4.00	18.3	ug/kg	10/27/25 09:26	PB170236
53469-21-9	Aroclor-1242	4.30	U	1	4.30	18.3	ug/kg	10/27/25 09:26	PB170236
12672-29-6	Aroclor-1248	6.40	U	1	6.40	18.3	ug/kg	10/27/25 09:26	PB170236
11097-69-1	Aroclor-1254	45.5		1	3.50	18.3	ug/kg	10/27/25 09:26	PB170236
37324-23-5	Aroclor-1262	5.40	U	1	5.40	18.3	ug/kg	10/27/25 09:26	PB170236
11100-14-4	Aroclor-1268	3.90	U	1	3.90	18.3	ug/kg	10/27/25 09:26	PB170236
11096-82-5	Aroclor-1260	3.50	U	1	3.50	18.3	ug/kg	10/27/25 09:26	PB170236
SURROGATES									
877-09-8	Tetrachloro-m-xylene	27.2			30 (32) - 150 (144)	136%	SPK: 20		
2051-24-3	Decachlorobiphenyl	28.0			30 (32) - 150 (175)	140%	SPK: 20		

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ102725\
 Data File : PQ071151.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 Oct 2025 09:26
 Operator : YP\AJ
 Sample : Q3450-01
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 SS-5R-(2.0-2.5)

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 27 11:28:52 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Oct 11 01:55:41 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.399	2.776	118.4E6	124.4E6	27.208	24.481
2) SA Decachlor...	8.497	7.531	81424033	136.5E6	28.039	27.033
Target Compounds						
26) L6 AR-1254-1	5.293	4.531	4546338	9896259	23.163	29.347 #
27) L6 AR-1254-2	5.507	4.677	18607410	19639861	64.057	66.093
28) L6 AR-1254-3	5.855	5.057	42076790	59419953	136.965	127.220
29) L6 AR-1254-4	6.136	5.284	34093872	45433959	155.297	153.882
30) L6 AR-1254-5	6.544	5.688	59025852	100.5E6	254.536	239.178



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ102725\
Data File : PQ071151.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Oct 2025 09:26
Operator : YP\AJ
Sample : Q3450-01
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
SS-5R-(2.0-2.5)

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 27 11:28:52 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Oct 11 01:55:41 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.399 min
Delta R.T.: 0.000 min
Response: 118364166
Conc: 27.21 ng/ml

Instrument :
ECD_Q
ClientSampleId :
SS-5R-(2.0-2.5)




#1 Tetrachloro-m-xylene

R.T.: 2.776 min
Delta R.T.: -0.002 min
Response: 124362089
Conc: 24.48 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.497 min
Delta R.T.: -0.003 min
Response: 81424033
Conc: 28.04 ng/ml



#2 Decachlorobiphenyl


R.T.: 7.531 min
Delta R.T.: -0.002 min
Response: 136506492
Conc: 27.03 ng/ml



#26 AR-1254-1


R.T.: 5.293 min
Delta R.T.: 0.000 min
Response: 4546338
Conc: 23.16 ng/ml

Instrument :
ECD_Q
ClientSampleId :
SS-5R-(2.0-2.5)




#26 AR-1254-1

R.T.: 4.531 min
Delta R.T.: -0.002 min
Response: 9896259
Conc: 29.35 ng/ml




#27 AR-1254-2

R.T.: 5.507 min
Delta R.T.: 0.000 min
Response: 18607410
Conc: 64.06 ng/ml



#27 AR-1254-2


R.T.: 4.677 min
Delta R.T.: -0.002 min
Response: 19639861
Conc: 66.09 ng/ml



#28 AR-1254-3


R.T.: 5.855 min
Delta R.T.: 0.000 min
Response: 42076790
Conc: 136.97 ng/ml

Instrument :
ECD_Q
ClientSampleId :
SS-5R-(2.0-2.5)




#28 AR-1254-3

R.T.: 5.057 min
Delta R.T.: -0.002 min
Response: 59419953
Conc: 127.22 ng/ml




#29 AR-1254-4

R.T.: 6.136 min
Delta R.T.: 0.000 min
Response: 34093872
Conc: 155.30 ng/ml



#29 AR-1254-4


R.T.: 5.284 min
Delta R.T.: -0.002 min
Response: 45433959
Conc: 153.88 ng/ml



#30 AR-1254-5

R.T.: 6.544 min
Delta R.T.: 0.000 min
Response: 59025852
Conc: 254.54 ng/ml

Instrument :
ECD_Q
ClientSampleId :
SS-5R-(2.0-2.5)



#30 AR-1254-5

R.T.: 5.688 min
Delta R.T.: -0.002 min
Response: 100538208
Conc: 239.18 ng/ml



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Matrix New World Engineering		Date Collected:	10/23/25
Project:	PSEG Athenia Substation		Date Received:	10/23/25
Client Sample ID:	SS-5R-(3.5-4.0)		SDG No.:	Q3450
Lab Sample ID:	Q3450-02		Matrix:	SOIL
Analytical Method:	8082A		% Solid:	86.3
Sample Wt/Vol:	30.08 g	Final Vol:	10000 uL	Test:
Prep Method:	SW3541	Prep Date	10/24/25	

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
12674-11-2	Aroclor-1016	4.60	U	1	4.60	19.6	ug/kg	10/27/25 09:41	PB170236
11104-28-2	Aroclor-1221	4.70	U	1	4.70	19.6	ug/kg	10/27/25 09:41	PB170236
11141-16-5	Aroclor-1232	4.30	U	1	4.30	19.6	ug/kg	10/27/25 09:41	PB170236
53469-21-9	Aroclor-1242	4.60	U	1	4.60	19.6	ug/kg	10/27/25 09:41	PB170236
12672-29-6	Aroclor-1248	6.80	U	1	6.80	19.6	ug/kg	10/27/25 09:41	PB170236
11097-69-1	Aroclor-1254	11.3	J	1	3.70	19.6	ug/kg	10/27/25 09:41	PB170236
37324-23-5	Aroclor-1262	5.80	U	1	5.80	19.6	ug/kg	10/27/25 09:41	PB170236
11100-14-4	Aroclor-1268	4.20	U	1	4.20	19.6	ug/kg	10/27/25 09:41	PB170236
11096-82-5	Aroclor-1260	3.70	U	1	3.70	19.6	ug/kg	10/27/25 09:41	PB170236
SURROGATES									
877-09-8	Tetrachloro-m-xylene	27.0			30 (32) - 150 (144)	135%	SPK: 20		
2051-24-3	Decachlorobiphenyl	25.4			30 (32) - 150 (175)	127%	SPK: 20		

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ102725\
 Data File : PQ071152.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 Oct 2025 09:41
 Operator : YP\AJ
 Sample : Q3450-02
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 SS-5R-(3.5-4.0)

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 27 11:29:29 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Oct 11 01:55:41 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.399	2.777	117.4E6	120.7E6	26.992	23.763
2)	SA Decachlor...	8.497	7.531	73642448	125.7E6	25.359	24.890
Target Compounds							
26)	L6 AR-1254-1	5.291	4.531	1879086	2547939	9.574	7.556
27)	L6 AR-1254-2	5.507	4.678	5837495	5610499	20.096	18.881
28)	L6 AR-1254-3	5.854	5.058	9901398	13892809	32.230	29.745
29)	L6 AR-1254-4	6.134	5.285	7352340	9923006	33.490	33.609
30)	L6 AR-1254-5	6.543	5.688	12037427	21754901	51.909	51.754



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ102725\
Data File : PQ071152.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Oct 2025 09:41
Operator : YP\AJ
Sample : Q3450-02
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
SS-5R-(3.5-4.0)

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 27 11:29:29 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Oct 11 01:55:41 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.399 min
Delta R.T.: 0.000 min
Response: 117421207
Conc: 26.99 ng/ml

Instrument :
ECD_Q
ClientSampleId :
SS-5R-(3.5-4.0)




#1 Tetrachloro-m-xylene

R.T.: 2.777 min
Delta R.T.: -0.002 min
Response: 120715928
Conc: 23.76 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.497 min
Delta R.T.: -0.003 min
Response: 73642448
Conc: 25.36 ng/ml



#2 Decachlorobiphenyl


R.T.: 7.531 min
Delta R.T.: -0.002 min
Response: 125683039
Conc: 24.89 ng/ml



#26 AR-1254-1


R.T.: 5.291 min
Delta R.T.: -0.002 min
Response: 1879086
Conc: 9.57 ng/ml

Instrument :
ECD_Q
ClientSampleId :
SS-5R-(3.5-4.0)




#26 AR-1254-1

R.T.: 4.531 min
Delta R.T.: -0.001 min
Response: 2547939
Conc: 7.56 ng/ml




#27 AR-1254-2

R.T.: 5.507 min
Delta R.T.: 0.000 min
Response: 5837495
Conc: 20.10 ng/ml



#27 AR-1254-2


R.T.: 4.678 min
Delta R.T.: -0.002 min
Response: 5610499
Conc: 18.88 ng/ml



#28 AR-1254-3


R.T.: 5.854 min
Delta R.T.: 0.000 min
Response: 9901398
Conc: 32.23 ng/ml

Instrument :
ECD_Q
ClientSampleId :
SS-5R-(3.5-4.0)




#28 AR-1254-3

R.T.: 5.058 min
Delta R.T.: -0.001 min
Response: 13892809
Conc: 29.74 ng/ml




#29 AR-1254-4

R.T.: 6.134 min
Delta R.T.: -0.001 min
Response: 7352340
Conc: 33.49 ng/ml



#29 AR-1254-4


R.T.: 5.285 min
Delta R.T.: 0.000 min
Response: 9923006
Conc: 33.61 ng/ml



#30 AR-1254-5

R.T.: 6.543 min
Delta R.T.: 0.000 min
Response: 12037427
Conc: 51.91 ng/ml

Instrument :
ECD_Q
ClientSampleId :
SS-5R-(3.5-4.0)



#30 AR-1254-5

R.T.: 5.688 min
Delta R.T.: -0.002 min
Response: 21754901
Conc: 51.75 ng/ml



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Matrix New World Engineering		Date Collected:	10/23/25
Project:	PSEG Athenia Substation		Date Received:	10/23/25
Client Sample ID:	SS-20-(0.0-0.5)		SDG No.:	Q3450
Lab Sample ID:	Q3450-08		Matrix:	SOIL
Analytical Method:	8082A		% Solid:	93.5
Sample Wt/Vol:	30.01 g	Final Vol:	10000 uL	Test:
Prep Method:	SW3541	Prep Date	10/24/25	

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
12674-11-2	Aroclor-1016	4.20	U	1	4.20	18.2	ug/kg	10/27/25 10:24	PB170236
11104-28-2	Aroclor-1221	4.30	U	1	4.30	18.2	ug/kg	10/27/25 10:24	PB170236
11141-16-5	Aroclor-1232	4.00	U	1	4.00	18.2	ug/kg	10/27/25 10:24	PB170236
53469-21-9	Aroclor-1242	4.30	U	1	4.30	18.2	ug/kg	10/27/25 10:24	PB170236
12672-29-6	Aroclor-1248	6.30	U	1	6.30	18.2	ug/kg	10/27/25 10:24	PB170236
11097-69-1	Aroclor-1254	174		1	3.40	18.2	ug/kg	10/27/25 10:24	PB170236
37324-23-5	Aroclor-1262	5.40	U	1	5.40	18.2	ug/kg	10/27/25 10:24	PB170236
11100-14-4	Aroclor-1268	3.80	U	1	3.80	18.2	ug/kg	10/27/25 10:24	PB170236
11096-82-5	Aroclor-1260	3.50	U	1	3.50	18.2	ug/kg	10/27/25 10:24	PB170236
SURROGATES									
877-09-8	Tetrachloro-m-xylene	26.8			30 (32) - 150 (144)	134%	SPK: 20		
2051-24-3	Decachlorobiphenyl	28.1			30 (32) - 150 (175)	140%	SPK: 20		

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ102725\
 Data File : PQ071155.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 Oct 2025 10:24
 Operator : YP\AJ
 Sample : Q3450-08
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 SS-20-(0.0-0.5)

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 27 11:31:16 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Oct 11 01:55:41 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.399	2.776	116.8E6	127.4E6	26.842	25.074
2) SA Decachlor...	8.495	7.530	81481303	134.3E6	28.058	26.593
Target Compounds						
26) L6 AR-1254-1	5.292	4.531	53593012	87844114	273.046	260.501
27) L6 AR-1254-2	5.506	4.677	103.5E6	115.1E6	356.309	387.376
28) L6 AR-1254-3	5.853	5.058	172.9E6	239.0E6	562.937	511.748
29) L6 AR-1254-4	6.134	5.284	103.5E6	136.4E6	471.483	462.124
30) L6 AR-1254-5	6.542	5.687	180.4E6	297.8E6	778.042	708.513



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ102725\
Data File : PQ071155.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Oct 2025 10:24
Operator : YP\AJ
Sample : Q3450-08
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
SS-20-(0.0-0.5)

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 27 11:31:16 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Oct 11 01:55:41 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.399 min
Delta R.T.: -0.001 min
Response: 116769388
Conc: 26.84 ng/ml

Instrument :
ECD_Q
ClientSampleId :
SS-20-(0.0-0.5)




#1 Tetrachloro-m-xylene

R.T.: 2.776 min
Delta R.T.: -0.002 min
Response: 127376293
Conc: 25.07 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.495 min
Delta R.T.: -0.005 min
Response: 81481303
Conc: 28.06 ng/ml



#2 Decachlorobiphenyl


R.T.: 7.530 min
Delta R.T.: -0.003 min
Response: 134285156
Conc: 26.59 ng/ml



#26 AR-1254-1


R.T.: 5.292 min
Delta R.T.: -0.001 min
Response: 53593012
Conc: 273.05 ng/ml

Instrument :
ECD_Q
ClientSampleId :
SS-20-(0.0-0.5)




#26 AR-1254-1

R.T.: 4.531 min
Delta R.T.: -0.002 min
Response: 87844114
Conc: 260.50 ng/ml




#27 AR-1254-2

R.T.: 5.506 min
Delta R.T.: -0.001 min
Response: 103501873
Conc: 356.31 ng/ml



#27 AR-1254-2


R.T.: 4.677 min
Delta R.T.: -0.002 min
Response: 115110478
Conc: 387.38 ng/ml



#28 AR-1254-3


R.T.: 5.853 min
Delta R.T.: -0.002 min
Response: 172938342
Conc: 562.94 ng/ml

Instrument :
ECD_Q
ClientSampleId :
SS-20-(0.0-0.5)




#28 AR-1254-3

R.T.: 5.058 min
Delta R.T.: -0.002 min
Response: 239019656
Conc: 511.75 ng/ml



#29 AR-1254-4

R.T.: 6.134 min
Delta R.T.: -0.002 min
Response: 103509591
Conc: 471.48 ng/ml




#29 AR-1254-4

R.T.: 5.284 min
Delta R.T.: -0.002 min
Response: 136442788
Conc: 462.12 ng/ml

Response_

Signal: PQ071155.D\ECD1A.ch

#30 AR-1254-5



R.T.: 6.542 min
Delta R.T.: -0.001 min
Response: 180424492
Conc: 778.04 ng/ml


Instrument :
ECD_Q
ClientSampleId :
SS-20-(0.0-0.5)

Time

Response_

Signal: PQ071155.D\ECD2B.ch

#30 AR-1254-5



R.T.: 5.687 min
Delta R.T.: -0.002 min
Response: 297823023
Conc: 708.51 ng/ml

Time



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Matrix New World Engineering		Date Collected:	10/23/25
Project:	PSEG Athenia Substation		Date Received:	10/23/25
Client Sample ID:	SS-20-(2.0-2.5)		SDG No.:	Q3450
Lab Sample ID:	Q3450-09		Matrix:	SOIL
Analytical Method:	8082A		% Solid:	86.3
Sample Wt/Vol:	30.09 g	Final Vol:	10000 uL	Test:
Prep Method:	SW3541	Prep Date	10/24/25	

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
12674-11-2	Aroclor-1016	4.60	U	1	4.60	19.6	ug/kg	10/27/25 10:39	PB170236
11104-28-2	Aroclor-1221	4.70	U	1	4.70	19.6	ug/kg	10/27/25 10:39	PB170236
11141-16-5	Aroclor-1232	4.30	U	1	4.30	19.6	ug/kg	10/27/25 10:39	PB170236
53469-21-9	Aroclor-1242	4.60	U	1	4.60	19.6	ug/kg	10/27/25 10:39	PB170236
12672-29-6	Aroclor-1248	6.80	U	1	6.80	19.6	ug/kg	10/27/25 10:39	PB170236
11097-69-1	Aroclor-1254	11.3	J	1	3.70	19.6	ug/kg	10/27/25 10:39	PB170236
37324-23-5	Aroclor-1262	5.80	U	1	5.80	19.6	ug/kg	10/27/25 10:39	PB170236
11100-14-4	Aroclor-1268	4.20	U	1	4.20	19.6	ug/kg	10/27/25 10:39	PB170236
11096-82-5	Aroclor-1260	3.70	U	1	3.70	19.6	ug/kg	10/27/25 10:39	PB170236
SURROGATES									
877-09-8	Tetrachloro-m-xylene	27.2			30 (32) - 150 (144)	136%	SPK: 20		
2051-24-3	Decachlorobiphenyl	27.4			30 (32) - 150 (175)	137%	SPK: 20		

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ102725\
 Data File : PQ071156.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 Oct 2025 10:39
 Operator : YP\AJ
 Sample : Q3450-09
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 SS-20-(2.0-2.5)

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 10/27/2025
 Supervised By :mohammad ahmed 10/29/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 27 11:32:27 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Oct 11 01:55:41 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.399	2.776	118.4E6	126.6E6	27.206	24.913
2)	SA Decachlor...	8.496	7.531	79502189	132.8E6	27.377	26.304
Target Compounds							
26)	L6 AR-1254-1	5.292	4.530	2481559	3492866	12.643	10.358
27)	L6 AR-1254-2	5.506	4.677	4523855	6232033	15.574	20.972 #
28)	L6 AR-1254-3	5.854	5.057	10115880	12983056	32.928	27.797m
29)	L6 AR-1254-4	6.134	5.283	5851786	8744197	26.655	29.616m
30)	L6 AR-1254-5	6.542	5.687	13252586	24169145	57.149	57.498

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ102725\
Data File : PQ071156.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Oct 2025 10:39
Operator : YP\AJ
Sample : Q3450-09
Misc :
ALS Vial : 16 Sample Multiplier: 1



Instrument :
ECD_Q
ClientSampleId :
SS-20-(2.0-2.5)


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/27/2025
Supervised By :mohammad ahmed 10/29/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 27 11:32:27 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Oct 11 01:55:41 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 3.399 min
Delta R.T.: -0.001 min
Response: 118351534
Conc: 27.21 ng/ml

Instrument :
ECD_Q
ClientSampleId :
SS-20-(2.0-2.5)


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/27/2025
Supervised By :mohammad ahmed 10/29/2025




#1 Tetrachloro-m-xylene

R.T.: 2.776 min
Delta R.T.: -0.002 min
Response: 126556708
Conc: 24.91 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.496 min
Delta R.T.: -0.004 min
Response: 79502189
Conc: 27.38 ng/ml



#2 Decachlorobiphenyl

R.T.: 7.531 min
Delta R.T.: -0.002 min
Response: 132825485
Conc: 26.30 ng/ml




#26 AR-1254-1

R.T.: 5.292 min
Delta R.T.: -0.001 min
Response: 2481559
Conc: 12.64 ng/ml

Instrument :
ECD_Q
ClientSampleId :
SS-20-(2.0-2.5)


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/27/2025
Supervised By :mohammad ahmed 10/29/2025




#26 AR-1254-1

R.T.: 4.530 min
Delta R.T.: -0.002 min
Response: 3492866
Conc: 10.36 ng/ml




#27 AR-1254-2

R.T.: 5.506 min
Delta R.T.: -0.001 min
Response: 4523855
Conc: 15.57 ng/ml



#27 AR-1254-2

R.T.: 4.677 min
Delta R.T.: -0.002 min
Response: 6232033
Conc: 20.97 ng/ml




#28 AR-1254-3

R.T.: 5.854 min
Delta R.T.: -0.001 min
Response: 10115880
Conc: 32.93 ng/ml

Instrument :
ECD_Q
ClientSampleId :
SS-20-(2.0-2.5)


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/27/2025
Supervised By :mohammad ahmed 10/29/2025




#28 AR-1254-3

R.T.: 5.057 min
Delta R.T.: -0.003 min
Response: 12983056
Conc: 27.80 ng/ml m




#29 AR-1254-4

R.T.: 6.134 min
Delta R.T.: -0.001 min
Response: 5851786
Conc: 26.65 ng/ml



#29 AR-1254-4

R.T.: 5.283 min
Delta R.T.: -0.003 min
Response: 8744197
Conc: 29.62 ng/ml m




#30 AR-1254-5

R.T.: 6.542 min
Delta R.T.: -0.001 min
Response: 13252586
Conc: 57.15 ng/ml

Instrument :
ECD_Q
ClientSampleId :
SS-20-(2.0-2.5)

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/27/2025
Supervised By :mohammad ahmed 10/29/2025



#30 AR-1254-5

R.T.: 5.687 min
Delta R.T.: -0.002 min
Response: 24169145
Conc: 57.50 ng/ml

Report of Analysis

Client:	Matrix New World Engineering		Date Collected:	10/23/25	
Project:	PSEG Athenia Substation		Date Received:	10/23/25	
Client Sample ID:	SS-20-(3.5-4.0)		SDG No.:	Q3450	
Lab Sample ID:	Q3450-10		Matrix:	SOIL	
Analytical Method:	8082A		% Solid:	84	
Sample Wt/Vol:	30.07 g	Final Vol:	10000 uL	Test:	PCB
Prep Method:	SW3541	Prep Date	10/24/25		

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
12674-11-2	Aroclor-1016	4.70	U	1	4.70	20.2	ug/kg	10/24/25 20:52	PB170236
11104-28-2	Aroclor-1221	4.80	U	1	4.80	20.2	ug/kg	10/24/25 20:52	PB170236
11141-16-5	Aroclor-1232	4.40	U	1	4.40	20.2	ug/kg	10/24/25 20:52	PB170236
53469-21-9	Aroclor-1242	4.80	U	1	4.80	20.2	ug/kg	10/24/25 20:52	PB170236
12672-29-6	Aroclor-1248	7.00	U	1	7.00	20.2	ug/kg	10/24/25 20:52	PB170236
11097-69-1	Aroclor-1254	55.2		1	3.80	20.2	ug/kg	10/24/25 20:52	PB170236
37324-23-5	Aroclor-1262	6.00	U	1	6.00	20.2	ug/kg	10/24/25 20:52	PB170236
11100-14-4	Aroclor-1268	4.30	U	1	4.30	20.2	ug/kg	10/24/25 20:52	PB170236
11096-82-5	Aroclor-1260	3.80	U	1	3.80	20.2	ug/kg	10/24/25 20:52	PB170236
SURROGATES									
877-09-8	Tetrachloro-m-xylene	24.1			30 (32) - 150 (144)	121%	SPK: 20		
2051-24-3	Decachlorobiphenyl	25.8			30 (32) - 150 (175)	129%	SPK: 20		

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102425\
 Data File : PP076047.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 24 Oct 2025 20:52
 Operator : YP\AJ
 Sample : Q3450-10
 Misc :
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 SS-20-(3.5-4.0)

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 10/27/2025
 Supervised By :mohammad ahmed 10/28/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 27 05:09:22 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Oct 24 15:40:32 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.640	3.779	34497841	152.2E6	24.140	22.996
2)	SA Decachlor...	10.411	8.777	27763719	184.0E6	25.790	22.378
Target Compounds							
26)	L6 AR-1254-1	6.639	5.653	2313097	21935559	28.952m	38.893 #
27)	L6 AR-1254-2	6.856	5.801	6552449	39782760	65.535	75.050
28)	L6 AR-1254-3	7.218	6.204	12960078	99046244	124.670	116.376
29)	L6 AR-1254-4	7.500	6.430	15873106	103.6E6	189.838	164.524
30)	L6 AR-1254-5	7.917	6.847	26406087	190.2E6	288.368	264.263

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102425\
Data File : PP076047.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 24 Oct 2025 20:52
Operator : YP\AJ
Sample : Q3450-10
Misc :
ALS Vial : 31 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
SS-20-(3.5-4.0)


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/27/2025
Supervised By :mohammad ahmed 10/28/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 27 05:09:22 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 24 15:40:32 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





#1 Tetrachloro-m-xylene

R.T.: 4.640 min
Delta R.T.: -0.002 min
Response: 34497841
Conc: 24.14 ng/ml

Instrument :

ECD_P

ClientSampleId :


SS-20-(3.5-4.0)

Manual Integrations

APPROVED


Reviewed By :Yogesh Patel 10/27/2025

Supervised By :mohammad ahmed 10/28/2025




#1 Tetrachloro-m-xylene

R.T.: 3.779 min
Delta R.T.: -0.002 min
Response: 152196349
Conc: 23.00 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.411 min
Delta R.T.: -0.008 min
Response: 27763719
Conc: 25.79 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.777 min
Delta R.T.: -0.005 min
Response: 184016645
Conc: 22.38 ng/ml




#26 AR-1254-1

R.T.: 6.639 min
Delta R.T.: -0.008 min
Response: 2313097
Conc: 28.95 ng/ml

Instrument :
ECD_P
ClientSampleId :
SS-20-(3.5-4.0)


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/27/2025
Supervised By :mohammad ahmed 10/28/2025




#26 AR-1254-1

R.T.: 5.653 min
Delta R.T.: -0.009 min
Response: 21935559
Conc: 38.89 ng/ml




#27 AR-1254-2

R.T.: 6.856 min
Delta R.T.: -0.008 min
Response: 6552449
Conc: 65.53 ng/ml



#27 AR-1254-2

R.T.: 5.801 min
Delta R.T.: -0.007 min
Response: 39782760
Conc: 75.05 ng/ml




#28 AR-1254-3

R.T.: 7.218 min
Delta R.T.: -0.008 min
Response: 12960078
Conc: 124.67 ng/ml

Instrument :
ECD_P
ClientSampleId :
SS-20-(3.5-4.0)


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/27/2025
Supervised By :mohammad ahmed 10/28/2025




#28 AR-1254-3

R.T.: 6.204 min
Delta R.T.: -0.006 min
Response: 99046244
Conc: 116.38 ng/ml




#29 AR-1254-4

R.T.: 7.500 min
Delta R.T.: -0.008 min
Response: 15873106
Conc: 189.84 ng/ml



#29 AR-1254-4

R.T.: 6.430 min
Delta R.T.: -0.008 min
Response: 103614141
Conc: 164.52 ng/ml




#30 AR-1254-5

R.T.: 7.917 min
Delta R.T.: -0.008 min
Response: 26406087
Conc: 288.37 ng/ml

Instrument :
ECD_P
ClientSampleId :
SS-20-(3.5-4.0)

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/27/2025
Supervised By :mohammad ahmed 10/28/2025



#30 AR-1254-5

R.T.: 6.847 min
Delta R.T.: -0.007 min
Response: 190238505
Conc: 264.26 ng/ml



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Matrix New World Engineering		Date Collected:	10/23/25
Project:	PSEG Athenia Substation		Date Received:	10/23/25
Client Sample ID:	SS-28-(2.0-2.5)		SDG No.:	Q3450
Lab Sample ID:	Q3450-11		Matrix:	SOIL
Analytical Method:	8082A		% Solid:	93
Sample Wt/Vol:	30.02 g	Final Vol:	10000 uL	Test:
Prep Method:	SW3541	Prep Date	10/24/25	

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
12674-11-2	Aroclor-1016	4.20	U	1	4.20	18.3	ug/kg	10/24/25 21:08	PB170236
11104-28-2	Aroclor-1221	4.30	U	1	4.30	18.3	ug/kg	10/24/25 21:08	PB170236
11141-16-5	Aroclor-1232	4.00	U	1	4.00	18.3	ug/kg	10/24/25 21:08	PB170236
53469-21-9	Aroclor-1242	4.30	U	1	4.30	18.3	ug/kg	10/24/25 21:08	PB170236
12672-29-6	Aroclor-1248	6.40	U	1	6.40	18.3	ug/kg	10/24/25 21:08	PB170236
11097-69-1	Aroclor-1254	3.40	U	1	3.40	18.3	ug/kg	10/24/25 21:08	PB170236
37324-23-5	Aroclor-1262	5.40	U	1	5.40	18.3	ug/kg	10/24/25 21:08	PB170236
11100-14-4	Aroclor-1268	3.90	U	1	3.90	18.3	ug/kg	10/24/25 21:08	PB170236
11096-82-5	Aroclor-1260	3.50	U	1	3.50	18.3	ug/kg	10/24/25 21:08	PB170236
SURROGATES									
877-09-8	Tetrachloro-m-xylene	23.3			30 (32) - 150 (144)	116%	SPK: 20		
2051-24-3	Decachlorobiphenyl	25.0			30 (32) - 150 (175)	125%	SPK: 20		

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102425\
Data File : PP076048.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 24 Oct 2025 21:08
Operator : YP\AJ
Sample : Q3450-11
Misc :
ALS Vial : 32 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
SS-28-(2.0-2.5)

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 24 23:40:37 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 24 15:40:32 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.643	3.782	33230144	150.1E6	23.253	22.678
2) SA Decachlor...	10.414	8.779	26956497	181.2E6	25.040	22.030

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102425\
Data File : PP076048.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 24 Oct 2025 21:08
Operator : YP\AJ
Sample : Q3450-11
Misc :
ALS Vial : 32 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
SS-28-(2.0-2.5)

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 24 23:40:37 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 24 15:40:32 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.643 min
Delta R.T.: 0.001 min
Response: 33230144
Conc: 23.25 ng/ml

Instrument :
ECD_P
ClientSampleId :
SS-28-(2.0-2.5)




#1 Tetrachloro-m-xylene

R.T.: 3.782 min
Delta R.T.: 0.000 min
Response: 150088081
Conc: 22.68 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.414 min
Delta R.T.: -0.005 min
Response: 26956497
Conc: 25.04 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.779 min
Delta R.T.: -0.003 min
Response: 181156077
Conc: 22.03 ng/ml



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Matrix New World Engineering	Date Collected:	10/23/25
Project:	PSEG Athenia Substation	Date Received:	10/23/25
Client Sample ID:	SS-21-(0.0-0.5)	SDG No.:	Q3450
Lab Sample ID:	Q3450-20	Matrix:	SOIL
Analytical Method:	8082A	% Solid:	91.7
Sample Wt/Vol:	30.05 g	Final Vol:	10000 uL
Prep Method:	SW3541	Test:	PCB
	Prep Date		10/24/25

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
12674-11-2	Aroclor-1016	4.30	U	1	4.30	18.5	ug/kg	10/24/25 21:24	PB170236
11104-28-2	Aroclor-1221	4.40	U	1	4.40	18.5	ug/kg	10/24/25 21:24	PB170236
11141-16-5	Aroclor-1232	4.00	U	1	4.00	18.5	ug/kg	10/24/25 21:24	PB170236
53469-21-9	Aroclor-1242	4.40	U	1	4.40	18.5	ug/kg	10/24/25 21:24	PB170236
12672-29-6	Aroclor-1248	6.40	U	1	6.40	18.5	ug/kg	10/24/25 21:24	PB170236
11097-69-1	Aroclor-1254	593	E	1	3.50	18.5	ug/kg	10/24/25 21:24	PB170236
37324-23-5	Aroclor-1262	5.50	U	1	5.50	18.5	ug/kg	10/24/25 21:24	PB170236
11100-14-4	Aroclor-1268	3.90	U	1	3.90	18.5	ug/kg	10/24/25 21:24	PB170236
11096-82-5	Aroclor-1260	3.50	U	1	3.50	18.5	ug/kg	10/24/25 21:24	PB170236
SURROGATES									
877-09-8	Tetrachloro-m-xylene	24.0			30 (32) - 150 (144)	120%	SPK: 20		
2051-24-3	Decachlorobiphenyl	25.0			30 (32) - 150 (175)	125%	SPK: 20		

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102425\
 Data File : PP076049.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 24 Oct 2025 21:24
 Operator : YP\AJ
 Sample : Q3450-20
 Misc :
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 SS-21-(0.0-0.5)

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 27 05:12:05 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Oct 24 15:40:32 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.640	3.779	34345481	149.5E6	24.033	22.583
2) SA Decachlor...	10.411	8.776	26890948	179.3E6	24.979	21.803
Target Compounds						
26) L6 AR-1254-1	6.641	5.656	44357794	459.7E6	555.212	815.008 #
27) L6 AR-1254-2	6.856	5.804	97377943	593.5E6	973.933	1119.679
28) L6 AR-1254-3	7.219	6.205	174.6E6	1691.9E6	1679.422	1987.948
29) L6 AR-1254-4	7.501	6.431	106.6E6	1034.7E6	1274.780	1642.925 #
30) L6 AR-1254-5	7.917	6.848	155.2E6	1873.1E6	1694.686	2601.893 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102425\
Data File : PP076049.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 24 Oct 2025 21:24
Operator : YP\AJ
Sample : Q3450-20
Misc :
ALS Vial : 33 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
SS-21-(0.0-0.5)

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 27 05:12:05 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 24 15:40:32 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.640 min
Delta R.T.: -0.002 min
Response: 34345481
Conc: 24.03 ng/ml

Instrument :
ECD_P
ClientSampleId :
SS-21-(0.0-0.5)




#1 Tetrachloro-m-xylene

R.T.: 3.779 min
Delta R.T.: -0.002 min
Response: 149460452
Conc: 22.58 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.411 min
Delta R.T.: -0.008 min
Response: 26890948
Conc: 24.98 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.776 min
Delta R.T.: -0.006 min
Response: 179283186
Conc: 21.80 ng/ml



#26 AR-1254-1


R.T.: 6.641 min
Delta R.T.: -0.005 min
Response: 44357794
Conc: 555.21 ng/ml

Instrument :
ECD_P
ClientSampleId :
SS-21-(0.0-0.5)




#26 AR-1254-1

R.T.: 5.656 min
Delta R.T.: -0.005 min
Response: 459664343
Conc: 815.01 ng/ml




#27 AR-1254-2

R.T.: 6.856 min
Delta R.T.: -0.007 min
Response: 97377943
Conc: 973.93 ng/ml



#27 AR-1254-2


R.T.: 5.804 min
Delta R.T.: -0.005 min
Response: 593520103
Conc: 1119.68 ng/ml



#28 AR-1254-3


R.T.: 7.219 min
Delta R.T.: -0.007 min
Response: 174583913
Conc: 1679.42 ng/ml

Instrument :
ECD_P
ClientSampleId :
SS-21-(0.0-0.5)




#28 AR-1254-3

R.T.: 6.205 min
Delta R.T.: -0.006 min
Response: 1691924436
Conc: 1987.95 ng/ml



#29 AR-1254-4

R.T.: 7.501 min
Delta R.T.: -0.008 min
Response: 106589163
Conc: 1274.78 ng/ml




#29 AR-1254-4

R.T.: 6.431 min
Delta R.T.: -0.006 min
Response: 1034683862
Conc: 1642.93 ng/ml

Signal: PP076049.D\ECD1A.ch

#30 AR-1254-5




R.T.: 7.917 min
Delta R.T.: -0.008 min
Response: 155183728
Conc: 1694.69 ng/ml

Instrument :
ECD_P
ClientSampleId :
SS-21-(0.0-0.5)

Signal: PP076049.D\ECD2B.ch

#30 AR-1254-5



R.T.: 6.848 min
Delta R.T.: -0.006 min
Response: 1873057153
Conc: 2601.89 ng/ml

Report of Analysis

Client: Matrix New World Engineering	Date Collected: 10/23/25
Project: PSEG Athenia Substation	Date Received: 10/23/25
Client Sample ID: SS-21-(0.0-0.5)DL	SDG No.: Q3450
Lab Sample ID: Q3450-20DL	Matrix: SOIL
Analytical Method: 8082A	% Solid: 91.7
Sample Wt/Vol: 30.05 g Final Vol: 10000 uL	Test: PCB
Prep Method: SW3541 Prep Date 10/24/25	

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
12674-11-2	Aroclor-1016	21.5	UD	5	21.5	92.5	ug/kg	10/27/25 11:42	PB170236
11104-28-2	Aroclor-1221	21.9	UD	5	21.9	92.5	ug/kg	10/27/25 11:42	PB170236
11141-16-5	Aroclor-1232	20.2	UD	5	20.2	92.5	ug/kg	10/27/25 11:42	PB170236
53469-21-9	Aroclor-1242	21.8	UD	5	21.8	92.5	ug/kg	10/27/25 11:42	PB170236
12672-29-6	Aroclor-1248	32.2	UD	5	32.2	92.5	ug/kg	10/27/25 11:42	PB170236
11097-69-1	Aroclor-1254	509	D	5	17.5	92.5	ug/kg	10/27/25 11:42	PB170236
37324-23-5	Aroclor-1262	27.3	UD	5	27.3	92.5	ug/kg	10/27/25 11:42	PB170236
11100-14-4	Aroclor-1268	19.6	UD	5	19.6	92.5	ug/kg	10/27/25 11:42	PB170236
11096-82-5	Aroclor-1260	17.6	UD	5	17.6	92.5	ug/kg	10/27/25 11:42	PB170236
SURROGATES									
877-09-8	Tetrachloro-m-xylene	21.4			30 (32) - 150 (144)	107%	SPK: 20		
2051-24-3	Decachlorobiphenyl	24.8			30 (32) - 150 (175)	124%	SPK: 20		

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102725\
 Data File : PP076072.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 Oct 2025 11:42
 Operator : YP\AJ
 Sample : Q3450-20DL 5X
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 SS-21-(0.0-0.5)DL

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 10/28/2025
 Supervised By :mohammad ahmed 10/29/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 27 11:55:27 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Oct 24 15:40:32 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.639	3.780	6137390	27135235	4.295m	4.100
2) SA Decachlor...	10.407	8.777	5339287	31864282	4.960	3.875
Target Compounds						
26) L6 AR-1254-1	6.641	5.657	11528188	95819976	144.295	169.894
27) L6 AR-1254-2	6.856	5.804	22107812	114.2E6	221.113	215.422
28) L6 AR-1254-3	7.219	6.205	37628464	275.6E6	361.970	323.852
29) L6 AR-1254-4	7.500	6.431	24138913	172.4E6	288.695	273.687
30) L6 AR-1254-5	7.917	6.848	35280033	234.1E6	385.276	325.201

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102725\
Data File : PP076072.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Oct 2025 11:42
Operator : YP\AJ
Sample : Q3450-20DL 5X
Misc :
ALS Vial : 12 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
SS-21-(0.0-0.5)DL


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/28/2025
Supervised By :mohammad ahmed 10/29/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 27 11:55:27 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 24 15:40:32 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.639 min
Delta R.T.: -0.003 min
Response: 6137390
Conc: 4.29 ng/ml

Instrument :
ECD_P
ClientSampleId :
SS-21-(0.0-0.5)DL


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/28/2025
Supervised By :mohammad ahmed 10/29/2025




#1 Tetrachloro-m-xylene

R.T.: 3.780 min
Delta R.T.: -0.002 min
Response: 27135235
Conc: 4.10 ng/ml



#2 Decachlorobiphenyl


R.T.: 10.407 min
Delta R.T.: -0.012 min
Response: 5339287
Conc: 4.96 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.777 min
Delta R.T.: -0.005 min
Response: 31864282
Conc: 3.88 ng/ml

Response_ Signal: PP076072.D\ECD1A.ch



#26 AR-1254-1


R.T.: 6.641 min
Delta R.T.: -0.006 min
Response: 11528188
Conc: 144.29 ng/ml

Instrument :
ECD_P
ClientSampleId :
SS-21-(0.0-0.5)DL

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/28/2025
Supervised By :mohammad ahmed 10/29/2025


Response_ Signal: PP076072.D\ECD2B.ch



#26 AR-1254-1

R.T.: 5.657 min
Delta R.T.: -0.004 min
Response: 95819976
Conc: 169.89 ng/ml


Response_ Signal: PP076072.D\ECD1A.ch



#27 AR-1254-2


R.T.: 6.856 min
Delta R.T.: -0.008 min
Response: 22107812
Conc: 221.11 ng/ml

Response_ Signal: PP076072.D\ECD2B.ch



#27 AR-1254-2

R.T.: 5.804 min
Delta R.T.: -0.004 min
Response: 114191174
Conc: 215.42 ng/ml




#28 AR-1254-3

R.T.: 7.219 min
Delta R.T.: -0.007 min
Response: 37628464
Conc: 361.97 ng/ml

Instrument :
ECD_P
ClientSampleId :
SS-21-(0.0-0.5)DL


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/28/2025
Supervised By :mohammad ahmed 10/29/2025




#28 AR-1254-3

R.T.: 6.205 min
Delta R.T.: -0.005 min
Response: 275627803
Conc: 323.85 ng/ml




#29 AR-1254-4

R.T.: 7.500 min
Delta R.T.: -0.008 min
Response: 24138913
Conc: 288.70 ng/ml



#29 AR-1254-4

R.T.: 6.431 min
Delta R.T.: -0.006 min
Response: 172363131
Conc: 273.69 ng/ml




#30 AR-1254-5

R.T.: 7.917 min
Delta R.T.: -0.008 min
Response: 35280033
Conc: 385.28 ng/ml

Instrument :
ECD_P
ClientSampleId :
SS-21-(0.0-0.5)DL

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/28/2025
Supervised By :mohammad ahmed 10/29/2025



#30 AR-1254-5

R.T.: 6.848 min
Delta R.T.: -0.005 min
Response: 234106565
Conc: 325.20 ng/ml



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Matrix New World Engineering	Date Collected:	10/23/25
Project:	PSEG Athenia Substation	Date Received:	10/23/25
Client Sample ID:	SS-21-(2.0-2.5)	SDG No.:	Q3450
Lab Sample ID:	Q3450-21	Matrix:	SOIL
Analytical Method:	8082A	% Solid:	93.2
Sample Wt/Vol:	30.08 g	Final Vol:	10000 uL
Prep Method:	SW3541	Test:	PCB
	Prep Date		10/24/25

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
12674-11-2	Aroclor-1016	4.20	U	1	4.20	18.2	ug/kg	10/24/25 21:41	PB170236
11104-28-2	Aroclor-1221	4.30	U	1	4.30	18.2	ug/kg	10/24/25 21:41	PB170236
11141-16-5	Aroclor-1232	4.00	U	1	4.00	18.2	ug/kg	10/24/25 21:41	PB170236
53469-21-9	Aroclor-1242	4.30	U	1	4.30	18.2	ug/kg	10/24/25 21:41	PB170236
12672-29-6	Aroclor-1248	6.30	U	1	6.30	18.2	ug/kg	10/24/25 21:41	PB170236
11097-69-1	Aroclor-1254	3.40	U	1	3.40	18.2	ug/kg	10/24/25 21:41	PB170236
37324-23-5	Aroclor-1262	5.40	U	1	5.40	18.2	ug/kg	10/24/25 21:41	PB170236
11100-14-4	Aroclor-1268	3.90	U	1	3.90	18.2	ug/kg	10/24/25 21:41	PB170236
11096-82-5	Aroclor-1260	3.50	U	1	3.50	18.2	ug/kg	10/24/25 21:41	PB170236
SURROGATES									
877-09-8	Tetrachloro-m-xylene	24.5			30 (32) - 150 (144)	123%	SPK: 20		
2051-24-3	Decachlorobiphenyl	26.4			30 (32) - 150 (175)	132%	SPK: 20		

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102425\
Data File : PP076050.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 24 Oct 2025 21:41
Operator : YP\AJ
Sample : Q3450-21
Misc :
ALS Vial : 34 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
SS-21-(2.0-2.5)

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 24 23:41:10 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 24 15:40:32 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.643	3.783	35015538	149.5E6	24.502	22.590
2)	SA Decachlor...	10.413	8.779	28463996	192.1E6	26.441	23.367

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102425\
Data File : PP076050.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 24 Oct 2025 21:41
Operator : YP\AJ
Sample : Q3450-21
Misc :
ALS Vial : 34 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
SS-21-(2.0-2.5)

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 24 23:41:10 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 24 15:40:32 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.643 min
Delta R.T.: 0.001 min
Response: 35015538
Conc: 24.50 ng/ml

Instrument :
ECD_P
ClientSampleId :
SS-21-(2.0-2.5)




#1 Tetrachloro-m-xylene

R.T.: 3.783 min
Delta R.T.: 0.002 min
Response: 149509931
Conc: 22.59 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.413 min
Delta R.T.: -0.006 min
Response: 28463996
Conc: 26.44 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.779 min
Delta R.T.: -0.003 min
Response: 192147524
Conc: 23.37 ng/ml



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Matrix New World Engineering		Date Collected:	10/23/25
Project:	PSEG Athenia Substation		Date Received:	10/23/25
Client Sample ID:	SS-21-(3.5-4.0)		SDG No.:	Q3450
Lab Sample ID:	Q3450-22		Matrix:	SOIL
Analytical Method:	8082A		% Solid:	84.7
Sample Wt/Vol:	30.04 g	Final Vol:	10000 uL	Test:
Prep Method:	SW3541	Prep Date	10/24/25	

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
12674-11-2	Aroclor-1016	4.70	U	1	4.70	20.0	ug/kg	10/24/25 21:57	PB170236
11104-28-2	Aroclor-1221	4.80	U	1	4.80	20.0	ug/kg	10/24/25 21:57	PB170236
11141-16-5	Aroclor-1232	4.40	U	1	4.40	20.0	ug/kg	10/24/25 21:57	PB170236
53469-21-9	Aroclor-1242	4.70	U	1	4.70	20.0	ug/kg	10/24/25 21:57	PB170236
12672-29-6	Aroclor-1248	7.00	U	1	7.00	20.0	ug/kg	10/24/25 21:57	PB170236
11097-69-1	Aroclor-1254	516	E	1	3.80	20.0	ug/kg	10/24/25 21:57	PB170236
37324-23-5	Aroclor-1262	5.90	U	1	5.90	20.0	ug/kg	10/24/25 21:57	PB170236
11100-14-4	Aroclor-1268	4.20	U	1	4.20	20.0	ug/kg	10/24/25 21:57	PB170236
11096-82-5	Aroclor-1260	3.80	U	1	3.80	20.0	ug/kg	10/24/25 21:57	PB170236
SURROGATES									
877-09-8	Tetrachloro-m-xylene	20.1			30 (32) - 150 (144)	100%	SPK: 20		
2051-24-3	Decachlorobiphenyl	17.2			30 (32) - 150 (175)	86%	SPK: 20		

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102425\
 Data File : PP076051.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 24 Oct 2025 21:57
 Operator : YP\AJ
 Sample : Q3450-22
 Misc :
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 SS-21-(3.5-4.0)

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 27 05:16:26 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Oct 24 15:40:32 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.639	3.779	28698180	125.8E6	20.081	19.010
2) SA Decachlor...	10.409	8.777	18498607	124.5E6	17.184	15.145
Target Compounds						
26) L6 AR-1254-1	6.640	5.657	12072300	127.8E6	151.105	226.514 #
27) L6 AR-1254-2	6.856	5.804	41710840	229.4E6	417.174	432.824
28) L6 AR-1254-3	7.218	6.204	98268988	940.9E6	945.305	1105.574
29) L6 AR-1254-4	7.500	6.431	112.9E6	1078.5E6	1350.585	1712.431 #
30) L6 AR-1254-5	7.916	6.848	206.9E6	2226.6E6	2259.030	3092.969 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102425\
Data File : PP076051.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 24 Oct 2025 21:57
Operator : YP\AJ
Sample : Q3450-22
Misc :
ALS Vial : 35 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
SS-21-(3.5-4.0)

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 27 05:16:26 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 24 15:40:32 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.639 min
Delta R.T.: -0.003 min
Response: 28698180
Conc: 20.08 ng/ml

Instrument :
ECD_P
ClientSampleId :
SS-21-(3.5-4.0)




#1 Tetrachloro-m-xylene

R.T.: 3.779 min
Delta R.T.: -0.002 min
Response: 125816731
Conc: 19.01 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.409 min
Delta R.T.: -0.010 min
Response: 18498607
Conc: 17.18 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.777 min
Delta R.T.: -0.005 min
Response: 124535846
Conc: 15.14 ng/ml



#26 AR-1254-1


R.T.: 6.640 min
Delta R.T.: -0.007 min
Response: 12072300
Conc: 151.10 ng/ml

Instrument :
ECD_P
ClientSampleId :
SS-21-(3.5-4.0)




#26 AR-1254-1

R.T.: 5.657 min
Delta R.T.: -0.005 min
Response: 127753836
Conc: 226.51 ng/ml




#27 AR-1254-2

R.T.: 6.856 min
Delta R.T.: -0.008 min
Response: 41710840
Conc: 417.17 ng/ml



#27 AR-1254-2


R.T.: 5.804 min
Delta R.T.: -0.005 min
Response: 229431687
Conc: 432.82 ng/ml



#28 AR-1254-3


R.T.: 7.218 min
Delta R.T.: -0.008 min
Response: 98268988
Conc: 945.31 ng/ml

Instrument :
ECD_P
ClientSampleId :
SS-21-(3.5-4.0)




#28 AR-1254-3

R.T.: 6.204 min
Delta R.T.: -0.006 min
Response: 940943554
Conc: 1105.57 ng/ml



#29 AR-1254-4

R.T.: 7.500 min
Delta R.T.: -0.009 min
Response: 112927507
Conc: 1350.58 ng/ml




#29 AR-1254-4

R.T.: 6.431 min
Delta R.T.: -0.006 min
Response: 1078457206
Conc: 1712.43 ng/ml

Signal: PP076051.D\ECD1A.ch

#30 AR-1254-5




R.T.: 7.916 min
Delta R.T.: -0.009 min
Response: 206861114
Conc: 2259.03 ng/ml

Instrument :
ECD_P
ClientSampleId :
SS-21-(3.5-4.0)

Signal: PP076051.D\ECD2B.ch

#30 AR-1254-5



R.T.: 6.848 min
Delta R.T.: -0.006 min
Response: 2226573531
Conc: 3092.97 ng/ml



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Matrix New World Engineering		Date Collected:	10/23/25
Project:	PSEG Athenia Substation		Date Received:	10/23/25
Client Sample ID:	SS-21-(3.5-4.0)DL		SDG No.:	Q3450
Lab Sample ID:	Q3450-22DL		Matrix:	SOIL
Analytical Method:	8082A		% Solid:	84.7
Sample Wt/Vol:	30.04 g	Final Vol:	10000 uL	Test:
Prep Method:	SW3541	Prep Date	10/24/25	

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
12674-11-2	Aroclor-1016	9.30	UD	2	9.30	40.1	ug/kg	10/27/25 11:58	PB170236
11104-28-2	Aroclor-1221	9.50	UD	2	9.50	40.1	ug/kg	10/27/25 11:58	PB170236
11141-16-5	Aroclor-1232	8.80	UD	2	8.80	40.1	ug/kg	10/27/25 11:58	PB170236
53469-21-9	Aroclor-1242	9.50	UD	2	9.50	40.1	ug/kg	10/27/25 11:58	PB170236
12672-29-6	Aroclor-1248	14.0	UD	2	14.0	40.1	ug/kg	10/27/25 11:58	PB170236
11097-69-1	Aroclor-1254	503	D	2	7.60	40.1	ug/kg	10/27/25 11:58	PB170236
37324-23-5	Aroclor-1262	11.8	UD	2	11.8	40.1	ug/kg	10/27/25 11:58	PB170236
11100-14-4	Aroclor-1268	8.50	UD	2	8.50	40.1	ug/kg	10/27/25 11:58	PB170236
11096-82-5	Aroclor-1260	7.60	UD	2	7.60	40.1	ug/kg	10/27/25 11:58	PB170236
SURROGATES									
877-09-8	Tetrachloro-m-xylene	21.6			30 (32) - 150 (144)	108%	SPK: 20		
2051-24-3	Decachlorobiphenyl	17.5			30 (32) - 150 (175)	88%	SPK: 20		

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102725\
 Data File : PP076073.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 Oct 2025 11:58
 Operator : YP\AJ
 Sample : Q3450-22DL 2X
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 SS-21-(3.5-4.0)DL

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 10/28/2025
 Supervised By :mohammad ahmed 10/29/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 27 12:11:39 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Oct 24 15:40:32 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.637	3.778	15408285	65330300	10.782	9.871
2)	SA Decachlor...	10.404	8.771	9443447	56936322	8.772	6.924
Target Compounds							
26)	L6 AR-1254-1	6.635	5.667	9652989	102.5E6	120.823m	181.709m#
27)	L6 AR-1254-2	6.851	5.799	23356369	116.0E6	233.600m	218.818m
28)	L6 AR-1254-3	7.214	6.202	62673662	469.4E6	602.894	551.583
29)	L6 AR-1254-4	7.497	6.428	61806136	538.1E6	739.186	854.358
30)	L6 AR-1254-5	7.913	6.844	106.8E6	1004.3E6	1165.820	1395.048

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102725\
Data File : PP076073.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Oct 2025 11:58
Operator : YP\AJ
Sample : Q3450-22DL 2X
Misc :
ALS Vial : 13 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
SS-21-(3.5-4.0)DL


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/28/2025
Supervised By :mohammad ahmed 10/29/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 27 12:11:39 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 24 15:40:32 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.637 min
Delta R.T.: -0.005 min
Response: 15408285
Conc: 10.78 ng/ml

Instrument :
ECD_P
ClientSampleId :
SS-21-(3.5-4.0)DL


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/28/2025
Supervised By :mohammad ahmed 10/29/2025




#1 Tetrachloro-m-xylene

R.T.: 3.778 min
Delta R.T.: -0.003 min
Response: 65330300
Conc: 9.87 ng/ml



#2 Decachlorobiphenyl


R.T.: 10.404 min
Delta R.T.: -0.015 min
Response: 9443447
Conc: 8.77 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.771 min
Delta R.T.: -0.011 min
Response: 56936322
Conc: 6.92 ng/ml

Response_ Signal: PP076073.D\ECD1A.ch



#26 AR-1254-1


R.T.: 6.635 min
Delta R.T.: -0.011 min
Response: 9652989
Conc: 120.82 ng/ml

Instrument :
ECD_P
ClientSampleId :
SS-21-(3.5-4.0)DL

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/28/2025
Supervised By :mohammad ahmed 10/29/2025


Response_ Signal: PP076073.D\ECD2B.ch



#26 AR-1254-1

R.T.: 5.667 min
Delta R.T.: 0.005 min
Response: 102483861
Conc: 181.71 ng/ml m


Response_ Signal: PP076073.D\ECD1A.ch



#27 AR-1254-2


R.T.: 6.851 min
Delta R.T.: -0.013 min
Response: 23356369
Conc: 233.60 ng/ml m

Response_ Signal: PP076073.D\ECD2B.ch



#27 AR-1254-2

R.T.: 5.799 min
Delta R.T.: -0.010 min
Response: 115991207
Conc: 218.82 ng/ml m




#28 AR-1254-3

R.T.: 7.214 min
Delta R.T.: -0.012 min
Response: 62673662
Conc: 602.89 ng/ml

Instrument :
ECD_P
ClientSampleId :
SS-21-(3.5-4.0)DL


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/28/2025
Supervised By :mohammad ahmed 10/29/2025




#28 AR-1254-3

R.T.: 6.202 min
Delta R.T.: -0.009 min
Response: 469446908
Conc: 551.58 ng/ml




#29 AR-1254-4

R.T.: 7.497 min
Delta R.T.: -0.012 min
Response: 61806136
Conc: 739.19 ng/ml



#29 AR-1254-4

R.T.: 6.428 min
Delta R.T.: -0.009 min
Response: 538058919
Conc: 854.36 ng/ml




#30 AR-1254-5

R.T.: 7.913 min
Delta R.T.: -0.012 min
Response: 106755060
Conc: 1165.82 ng/ml

Instrument :
ECD_P
ClientSampleId :
SS-21-(3.5-4.0)DL

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/28/2025
Supervised By :mohammad ahmed 10/29/2025



#30 AR-1254-5

R.T.: 6.844 min
Delta R.T.: -0.010 min
Response: 1004270507
Conc: 1395.05 ng/ml



CALIBRATION SUMMARY

Calibration Times: **09:06** **17:01**

LAB FILE ID: RT 1000 = PP075888.D RT 750 = PP075889.D
RT 500 = PP075890.D RT 250 = PP075891.D RT 050 = PP075892.D

[illegible]

[illegible]

Calibration Times: 09:06 17:01

LAB FILE ID: RT 1000 = PP075888.D RT 750 = PP075889.D
RT 500 = PP075890.D RT 250 = PP075891.D RT 050 = PP075892.D

[illegible]

RETENTION TIMES OF INITIAL CALIBRATION

Decachlorobiphenyl	8.79	8.78	8.79	8.79	8.79	8.79	8.69	8.89
Tetrachloro-m-xylene	3.78	3.78	3.79	3.79	3.79	3.78	3.68	3.88



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION FACTOR OF INITIAL CALIBRATION

Lab Name: Alliance
Lab Code: ACE
Instrument ID: ECD_P

Contract: MATR02
SDG NO.: Q3450

Calibration Date(s): 10/21/2025 10/21/2025
Calibration Times: 09:06 17:01

GC Column: ZB-MR1 ID: 0.32 (mm)

LAB FILE ID:		CF 1000 = <u>PP075888.D</u>		CF 750 = <u>PP075889.D</u>			
CF 500 = <u>PP075890.D</u>		CF 250 = <u>PP075891.D</u>		CF 050 = <u>PP075892.D</u>			
COMPOUND		CF 1000	CF 750	CF 500	CF 250	CF 050	% RSD
Aroclor-1016-1	(1)	53893845	57848439	60228768	63056856	60602640	59126110
Aroclor-1016-2	(2)	81387326	86593457	91038396	94931280	88270120	88444116
Aroclor-1016-3	(3)	50025937	53493744	56764402	59544472	56926600	55351031
Aroclor-1016-4	(4)	41815325	44617539	46982858	48544784	45240520	45440205
Aroclor-1016-5	(5)	39075321	41377448	43150568	44782124	48642160	43405524
Aroclor-1260-1	(1)	64655691	67908073	70553690	74100180	66565440	68756615
Aroclor-1260-2	(2)	80744063	85234851	88594752	93987136	99325680	89577296
Aroclor-1260-3	(3)	64964687	67552583	69003280	72715708	72196540	69286560
Aroclor-1260-4	(4)	65448766	70711301	72542976	74553116	59600740	68571380
Aroclor-1260-5	(5)	139384380	146230631	152452738	155039368	146650940	147951611
Decachlorobiphenyl		1012132950	1071302787	1124785220	1137889120	1036523400	1076526695
Tetrachloro-m-xylene		1370219590	1453116893	1490571040	1483020960	1348537200	1429093137
Aroclor-1242-1	(1)	44019933	46315467	47674416	50255084	51268840	47906748
Aroclor-1242-2	(2)	64942800	67715737	70979556	73767248	76553360	70791740
Aroclor-1242-3	(3)	40140584	42044764	44589646	47135032	54298900	45641785
Aroclor-1242-4	(4)	33543618	35162207	36420600	36967860	39702420	36359341
Aroclor-1242-5	(5)	36405140	38662289	41888622	39296052	48467620	40943945
Decachlorobiphenyl		1136280900	1209279307	1231278960	1289202560	1357807000	1244769745
Tetrachloro-m-xylene		1466601570	1512648573	1545825660	1595207720	1573043800	1538665465
Aroclor-1248-1	(1)	33471805	35785521	36963520	39121904	34010240	35870598
Aroclor-1248-2	(2)	44816839	47584043	49402910	52441868	49476340	48744400
Aroclor-1248-3	(3)	49477903	52446567	61105936	57335280	50958020	54264741
Aroclor-1248-4	(4)	60435029	64167197	67505182	72174088	67539280	66364155
Aroclor-1248-5	(5)	58134534	61713151	65347636	71297512	67810300	64860627
Decachlorobiphenyl		1140637530	1190681093	1243623480	1306886720	1250686400	1226503045
Tetrachloro-m-xylene		1468701040	1529744213	1558324940	1626496320	1450854600	1526824223
Aroclor-1254-1	(1)	66261960	71898345	78270020	82464920	100572060	79893461
Aroclor-1254-2	(2)	87851253	92570580	96669370	102809560	120020480	99984249
Aroclor-1254-3	(3)	92035771	96900691	99787992	105988040	125061300	103954759
Aroclor-1254-4	(4)	73136868	76689867	79390286	83967352	104884540	83613783
Aroclor-1254-5	(5)	85440842	89720104	92252192	95013656	95427040	91570767
Decachlorobiphenyl		1194970230	1245235800	1296058720	1332579960	1514859800	1316740902
Tetrachloro-m-xylene		1499217980	1559170693	1583264260	1614158480	1720822800	1595326843
Aroclor-1268-1	(1)	186733090	195066952	205606148	210022232	235253460	206536376



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION FACTOR OF INITIAL CALIBRATION

Aroclor-1268-2	(2)	172486766	179135464	188184104	190078376	211424800	188261902	8
Aroclor-1268-3	(3)	145245892	150281380	156894280	158090072	165408840	155184093	5
Aroclor-1268-4	(4)	71036137	74805685	76392484	76627212	94373740	78647052	12
Aroclor-1268-5	(5)	446420076	458917541	475352492	479939416	541304600	480386825	8
Decachlorobiphenyl		1988394870	2075434813	2173906840	2211711440	2356033600	2161096313	6
Tetrachloro-m-xylene		1515083040	1549265533	1603737900	1591085800	1659449000	1583724255	3



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION FACTOR OF INITIAL CALIBRATION

Lab Name: Alliance
Lab Code: ACE
Instrument ID: ECD_P

Contract: MATR02
SDG NO.: Q3450

Calibration Date(s): 10/21/2025 10/21/2025
Calibration Times: 09:06 17:01

GC Column: ZB-MR2 ID: 0.32 (mm)

LAB FILE ID:		CF 1000 = <u>PP075888.D</u>		CF 750 = <u>PP075889.D</u>			
CF 500 = <u>PP075890.D</u>		CF 250 = <u>PP075891.D</u>		CF 050 = <u>PP075892.D</u>			
COMPOUND		CF 1000	CF 750	CF 500	CF 250	CF 050	% RSD
Aroclor-1016-1	(1)	708597432	755130533	750168220	768814916	718149060	740172032 3
Aroclor-1016-2	(2)	344688013	362255933	349033076	368003724	285681000	341932349 10
Aroclor-1016-3	(3)	184756021	198805969	201689770	204443556	237017160	205342495 9
Aroclor-1016-4	(4)	185429480	195612065	193130110	203643812	156612400	186885573 10
Aroclor-1016-5	(5)	202087874	223186380	221264342	249936892	243439240	227982946 8
Aroclor-1260-1	(1)	519115007	547143451	551523748	550034516	570658900	547695124 3
Aroclor-1260-2	(2)	816933203	847057133	836099344	841169992	727408040	813733542 6
Aroclor-1260-3	(3)	571186815	591919587	603061842	595667512	451948620	562756875 11
Aroclor-1260-4	(4)	562542796	557220415	556309062	533352684	387232680	519331527 14
Aroclor-1260-5	(5)	1443277342	1495374137	1448146362	1379476524	1011499060	1355554685 15
Decachlorobiphenyl		7970697090	8448130240	8416417320	8375691800	7903993200	8222985930 3
Tetrachloro-m-xylene		6687145430	7018651333	6804923780	6758234840	5822487600	6618288597 7
Aroclor-1242-1	(1)	571147464	588341980	597723718	569760388	546790280	574752766 3
Aroclor-1242-2	(2)	271622570	270496133	289123856	291348184	264965740	277511297 4
Aroclor-1242-3	(3)	148417684	150685240	162169734	157714788	168071120	157411713 5
Aroclor-1242-4	(4)	195275059	191450107	210824584	191270784	136740280	185112163 15
Aroclor-1242-5	(5)	229588702	236959084	244202752	234962408	205474700	230237529 6
Decachlorobiphenyl		8905080650	8993413907	9094156860	9113318400	9309338200	9083061603 2
Tetrachloro-m-xylene		7556133750	7417192413	7366207320	7201170040	6123421000	7132824905 8
Aroclor-1248-1	(1)	362937087	375551145	389160858	390611340	359825360	375617158 4
Aroclor-1248-2	(2)	226906631	241863020	243838414	257801160	259130280	245907901 5
Aroclor-1248-3	(3)	258884180	267447340	286041186	300772772	321318420	286892780 9
Aroclor-1248-4	(4)	264944699	288582168	281543548	284810560	293675780	282711351 4
Aroclor-1248-5	(5)	531974990	530976477	538260328	580030476	417051840	519658822 12
Decachlorobiphenyl		8995204870	9175718960	9269974040	9050625160	7344080200	8767120646 9
Tetrachloro-m-xylene		7343926730	7496283907	7432157100	7125361600	5416154400	6962776747 13
Aroclor-1254-1	(1)	570889099	597066183	605466528	586480320	460097860	563999998 11
Aroclor-1254-2	(2)	484331632	496825463	519240264	512723972	637282480	530080762 12
Aroclor-1254-3	(3)	866771846	893592493	892235862	873540968	729313040	851090842 8
Aroclor-1254-4	(4)	656603428	687178148	673132272	636725912	495267720	629781496 12
Aroclor-1254-5	(5)	714919063	793859644	718111150	679323440	693198580	719882375 6
Decachlorobiphenyl		9294247170	9652211800	9737198260	9430246600	10010120400	9624804846 3
Tetrachloro-m-xylene		7467434200	7759548133	7675903540	7263970760	7005463400	7434464007 4
Aroclor-1268-1	(1)	1831610079	1813456776	1858863308	1732000476	1817331660	1810652460 3



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION FACTOR OF INITIAL CALIBRATION

Aroclor-1268-2	(2)	1800363506	1772458256	1795415328	1662166904	1645167940	1735114387	4
Aroclor-1268-3	(3)	1369797429	1358701644	1402746042	1358715304	1345500120	1367092108	2
Aroclor-1268-4	(4)	542764588	536784571	540251440	501746924	465672060	517443917	6
Aroclor-1268-5	(5)	3855476951	3900359425	4013489614	3707995788	3381331260	3771730608	6
Decachlorobiphenyl		16718332370	16659738347	17037476180	16197346520	15850998600	16492778403	3
Tetrachloro-m-xylene		7499400880	7531972947	7553431460	7073074800	6855756800	7302727377	4



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Lab Name: Alliance Contract: MATR02

Lab Code: ACE SDG NO.: Q3450

Instrument ID: ECD_P Date(s) Analyzed: 10/21/2025 10/21/2025

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
Aroclor-1221	500	1	4.84	4.74	4.94	20176600
		2	4.93	4.83	5.03	15061400
		3	5.01	4.91	5.11	46915600
		4	0.00			0
		5	0.00			0
Aroclor-1232	500	1	5.01	4.91	5.11	39314800
		2	5.53	5.43	5.63	20627600
		3	5.82	5.72	5.92	38639600
		4	5.98	5.88	6.08	20172600
		5	6.07	5.97	6.17	15662900
Aroclor-1262	500	1	8.23	8.13	8.33	84026000
		2	8.56	8.46	8.66	163669000
		3	8.88	8.78	8.98	121787000
		4	8.97	8.87	9.07	92111600
		5	9.64	9.54	9.74	67824600



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Lab Name: Alliance **Contract:** MATR02
Lab Code: ACE **SDG NO.:** Q3450
Instrument ID: ECD_P **Date(s) Analyzed:** 10/21/2025 10/21/2025
GC Column: ZB-MR2 **ID:** 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
Aroclor-1221	500	1	3.99	3.89	4.09	90009000
		2	4.08	3.98	4.18	63177800
		3	4.15	4.05	4.25	212036000
		4	0.00			0
		5	0.00			0
Aroclor-1232	500	1	4.15	4.05	4.25	168044000
		2	4.88	4.78	4.98	297516000
		3	5.06	4.96	5.16	81766800
		4	5.14	5.04	5.24	84943400
		5	5.31	5.21	5.41	100651000
Aroclor-1262	500	1	6.89	6.79	6.99	986952000
		2	7.15	7.05	7.25	703208000
		3	7.67	7.57	7.77	606960000
		4	7.74	7.64	7.84	1101340000
		5	8.23	8.13	8.33	461492000

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
 Data File : PP075888.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Oct 2025 09:06
 Operator : YP\AJ
 Sample : AR1660ICC1000
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660ICC1000

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 21 17:20:08 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Oct 21 17:14:53 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.646	3.784	137.0E6	668.7E6	95.793	99.127
2) SA Decachlor...	10.422	8.784	101.2E6	797.1E6	94.728	97.280
Target Compounds						
3) L1 AR-1016-1	5.799	4.880	53893845	708.6E6	944.490	971.503
4) L1 AR-1016-2	5.820	4.938	81387326	344.7E6	944.028	993.737
5) L1 AR-1016-3	5.883	5.058	50025937	184.8E6	936.900	956.181
6) L1 AR-1016-4	5.980	5.099	41815325	185.4E6	941.806	979.658
7) L1 AR-1016-5	6.272	5.313	39075321	202.1E6	950.438	954.703
31) L7 AR-1260-1	7.391	6.341	64655691	519.1E6	956.379	969.730
32) L7 AR-1260-2	7.643	6.527	80744063	816.9E6	953.639	988.405
33) L7 AR-1260-3	8.001	6.681	64964687	571.2E6	969.854	972.855
34) L7 AR-1260-4	8.229	7.150	65448766	562.5E6	948.590	1005.572
35) L7 AR-1260-5	8.555	7.390	139.4E6	1443.3E6	955.220	998.316



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075888.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2025 09:06
Operator : YP\AJ
Sample : AR1660ICC1000
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660ICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 17:20:08 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 21 17:14:53 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
 Data File : PP075889.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Oct 2025 09:22
 Operator : YP\AJ
 Sample : AR1660ICC750
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660ICC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 21 17:23:54 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Oct 21 17:14:53 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.642	3.781	109.0E6	526.4E6	75.790	76.994
2) SA Decachlor...	10.419	8.782	80347709	633.6E6	75.133	76.538
Target Compounds						
3) L1 AR-1016-1	5.795	4.877	43386329	566.3E6	756.866	767.445
4) L1 AR-1016-2	5.817	4.935	64945093	271.7E6	752.204	771.869
5) L1 AR-1016-3	5.879	5.055	40120308	149.1E6	750.922	764.309
6) L1 AR-1016-4	5.977	5.096	33463154	146.7E6	752.456	766.543
7) L1 AR-1016-5	6.270	5.309	31033086	167.4E6	753.210	776.704
31) L7 AR-1260-1	7.387	6.338	50931055	410.4E6	752.240	760.963
32) L7 AR-1260-2	7.640	6.525	63926138	635.3E6	753.332	762.324
33) L7 AR-1260-3	7.999	6.678	50664437	443.9E6	754.232	754.073
34) L7 AR-1260-4	8.226	7.147	53033476	417.9E6	762.329	748.026
35) L7 AR-1260-5	8.552	7.387	109.7E6	1121.5E6	751.069	766.981



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075889.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2025 09:22
Operator : YP\AJ
Sample : AR1660ICC750
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 17:23:54 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 21 17:14:53 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
 Data File : PP075890.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Oct 2025 09:38
 Operator : YP\AJ
 Sample : AR1660ICC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 21 17:17:19 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Oct 21 17:14:53 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.642	3.781	74528552	340.2E6	50.000	50.000
2) SA Decachlor...	10.419	8.782	56239261	420.8E6	50.000	50.000
Target Compounds						
3) L1 AR-1016-1	5.795	4.878	30114384	375.1E6	500.000	500.000
4) L1 AR-1016-2	5.817	4.936	45519198	174.5E6	500.000	500.000
5) L1 AR-1016-3	5.880	5.054	28382201	100.8E6	500.000	500.000
6) L1 AR-1016-4	5.977	5.096	23491429	96565055	500.000	500.000
7) L1 AR-1016-5	6.269	5.310	21575284	110.6E6	500.000	500.000
31) L7 AR-1260-1	7.387	6.338	35276845	275.8E6	500.000	500.000
32) L7 AR-1260-2	7.640	6.525	44297376	418.0E6	500.000	500.000
33) L7 AR-1260-3	7.999	6.678	34501640	301.5E6	500.000	500.000
34) L7 AR-1260-4	8.227	7.147	36271488	278.2E6	500.000	500.000
35) L7 AR-1260-5	8.552	7.387	76226369	724.1E6	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075890.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2025 09:38
Operator : YP\AJ
Sample : AR1660ICC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 17:17:19 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 21 17:14:53 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
 Data File : PP075891.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Oct 2025 09:54
 Operator : YP\AJ
 Sample : AR1660ICC250
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660ICC250

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 21 17:27:12 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Oct 21 17:14:53 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.645	3.784	37075524	169.0E6	25.583	24.784
2) SA Decachlor...	10.422	8.784	28447228	209.4E6	26.182	25.220
Target Compounds						
3) L1 AR-1016-1	5.798	4.879	15764214	192.2E6	268.295	257.757
4) L1 AR-1016-2	5.819	4.938	23732820	92000931	268.205	258.433
5) L1 AR-1016-3	5.882	5.057	14886118	51110889	270.868	258.889
6) L1 AR-1016-4	5.979	5.098	12136196	50910953	266.787	261.815
7) L1 AR-1016-5	6.272	5.311	11195531	62484223	265.950	278.799
31) L7 AR-1260-1	7.390	6.340	18525045	137.5E6	267.300	253.727
32) L7 AR-1260-2	7.643	6.526	23496784	210.3E6	269.643	251.752
33) L7 AR-1260-3	8.001	6.680	18178927	148.9E6	265.157	252.205
34) L7 AR-1260-4	8.229	7.150	18638279	133.3E6	263.200	241.399
35) L7 AR-1260-5	8.554	7.389	38759842	344.9E6	261.402	239.232



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075891.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2025 09:54
Operator : YP\AJ
Sample : AR1660ICC250
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660ICC250

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 17:27:12 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 21 17:14:53 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
 Data File : PP075892.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Oct 2025 10:11
 Operator : YP\AJ
 Sample : AR1660ICC050
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660ICC050

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 10/22/2025
 Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 21 17:33:37 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Oct 21 17:14:53 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.645	3.781	6742686	29112438	4.718	4.278m
2) SA Decachlor...	10.422	8.783	5182617	39519966	4.814	4.806
Target Compounds						
3) L1 AR-1016-1	5.798	4.880	3030132	35907453	51.249	48.512
4) L1 AR-1016-2	5.819	4.943	4413506	14284050	49.902	41.774
5) L1 AR-1016-3	5.880	5.052	2846330	11850858	50.609m	58.316m
6) L1 AR-1016-4	5.978	5.094	2262026	7830620	48.653m	41.996m
7) L1 AR-1016-5	6.272	5.308	2432108	12171962	56.032	53.433m
31) L7 AR-1260-1	7.390	6.339	3328272	28532945	48.407	52.096
32) L7 AR-1260-2	7.643	6.528	4966284	36370402	55.441	44.696
33) L7 AR-1260-3	8.001	6.681	3609827	22597431	52.100	40.155
34) L7 AR-1260-4	8.229	7.150	2980037	19361634	43.459	37.282
35) L7 AR-1260-5	8.555	7.388	7332547	50574953	49.560	37.309

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075892.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2025 10:11
Operator : YP\AJ
Sample : AR1660ICC050
Misc :
ALS Vial : 7 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
AR1660ICC050

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/22/2025
Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 17:33:37 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 21 17:14:53 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
 Data File : PP075893.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Oct 2025 10:27
 Operator : YP\AJ
 Sample : AR1221ICC500
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1221ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 21 16:37:58 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Oct 21 16:37:41 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.643	3.782	75268417	347.9E6	50.000	50.000
2) SA Decachlor...	10.419	8.781	55765598	413.6E6	50.000	50.000
Target Compounds						
8) L2 AR-1221-1	4.843	3.993	10088314	45004530	500.000	500.000
9) L2 AR-1221-2	4.929	4.076	7530715	31588869	500.000	500.000
10) L2 AR-1221-3	5.005	4.154	23457783	106.0E6	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075893.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2025 10:27
Operator : YP\AJ
Sample : AR1221ICC500
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1221ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 16:37:58 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 21 16:37:41 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
 Data File : PP075894.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Oct 2025 10:43
 Operator : YP\AJ
 Sample : AR1232ICC500
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1232ICC500

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 10/22/2025
 Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 21 16:45:39 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Oct 21 16:45:19 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.644	3.783	76762692	367.2E6	50.000	50.000
2) SA Decachlor...	10.418	8.783	57573853	421.9E6	50.000	50.000
Target Compounds						
11) L3 AR-1232-1	5.006	4.154	19657396	84022049	500.000	500.000
12) L3 AR-1232-2	5.532	4.877	10313847	148.8E6	500.000	500.000
13) L3 AR-1232-3	5.818	5.056	19319843	40883369	500.000	500.000
14) L3 AR-1232-4	5.978	5.136	10086255	42471740	500.000	480.688m
15) L3 AR-1232-5	6.068	5.311	7831441	50325540	500.000	500.000

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075894.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2025 10:43
Operator : YP\AJ
Sample : AR1232ICC500
Misc :
ALS Vial : 9 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
AR1232ICC500

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/22/2025
Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 16:45:39 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 21 16:45:19 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
 Data File : PP075895.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Oct 2025 10:59
 Operator : YP\AJ
 Sample : AR1242ICC1000
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1242ICC1000

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 10/22/2025
 Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 21 16:57:47 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Oct 21 16:53:53 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.646	3.784	146.7E6	755.6E6	97.370	101.273
2)	SA Decachlor...	10.422	8.784	113.6E6	890.5E6	95.988	98.950
Target Compounds							
16)	L4 AR-1242-1	5.799	4.880	44019933	571.1E6	960.145	977.263
17)	L4 AR-1242-2	5.820	4.938	64942800	271.6E6	955.587	968.789
18)	L4 AR-1242-3	5.882	5.058	40140584	148.4E6	947.491	955.722
19)	L4 AR-1242-4	5.980	5.140	33543618	195.3E6	958.879	946.968m
20)	L4 AR-1242-5	6.710	5.660	36405140	229.6E6	929.963	969.155

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075895.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2025 10:59
Operator : YP\AJ
Sample : AR1242ICC1000
Misc :
ALS Vial : 10 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
AR1242ICC1000

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/22/2025
Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 16:57:47 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 21 16:53:53 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
 Data File : PP075896.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Oct 2025 11:16
 Operator : YP\AJ
 Sample : AR1242ICC750
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1242ICC750

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 10/22/2025
 Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 21 17:03:07 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Oct 21 16:53:53 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.644	3.782	113.4E6	556.3E6	75.213	74.705
2)	SA Decachlor...	10.420	8.782	90695948	674.5E6	76.069	74.966
Target Compounds							
16)	L4 AR-1242-1	5.797	4.877	34736600	441.3E6	755.090	753.335
17)	L4 AR-1242-2	5.818	4.936	50786803	202.9E6	748.192	732.177
18)	L4 AR-1242-3	5.881	5.056	31533573	113.0E6	746.210	735.014
19)	L4 AR-1242-4	5.978	5.138	26371655	143.6E6	752.570	715.813m
20)	L4 AR-1242-5	6.708	5.658	28996717	177.7E6	743.785	750.134

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075896.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2025 11:16
Operator : YP\AJ
Sample : AR1242ICC750
Misc :
ALS Vial : 11 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
AR1242ICC750

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/22/2025
Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 17:03:07 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 21 16:53:53 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
 Data File : PP075897.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Oct 2025 11:48
 Operator : YP\AJ
 Sample : AR1242ICC500
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1242ICC500

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 10/22/2025
 Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 21 16:54:12 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Oct 21 16:53:53 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.647	3.784	77291283	368.3E6	50.000	50.000
2) SA Decachlor...	10.422	8.785	61563948	454.7E6	50.000	50.000
Target Compounds						
16) L4 AR-1242-1	5.799	4.880	23837208	298.9E6	498.805	500.000
17) L4 AR-1242-2	5.821	4.938	35489778	144.6E6	500.000	500.000
18) L4 AR-1242-3	5.884	5.059	22294823	81084867	500.000	500.000
19) L4 AR-1242-4	5.981	5.141	18210300	105.4E6	500.000	515.181m
20) L4 AR-1242-5	6.710	5.660	20944311	122.1E6	500.000	500.000

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075897.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2025 11:48
Operator : YP\AJ
Sample : AR1242ICC500
Misc :
ALS Vial : 12 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
AR1242ICC500

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/22/2025
Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 16:54:12 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 21 16:53:53 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
 Data File : PP075898.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Oct 2025 12:04
 Operator : YP\AJ
 Sample : AR1242ICC250
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1242ICC250

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 10/22/2025
 Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 21 17:06:44 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Oct 21 16:53:53 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.645	3.783	39880193	180.0E6	26.064	24.377
2)	SA Decachlor...	10.420	8.782	32230064	227.8E6	26.494	25.240
Target Compounds							
16)	L4 AR-1242-1	5.797	4.880	12563771	142.4E6	266.938	244.850
17)	L4 AR-1242-2	5.819	4.936	18441812	72837046	265.919	259.532
18)	L4 AR-1242-3	5.881	5.057	11783758	39428697	271.031	254.795
19)	L4 AR-1242-4	5.979	5.139	9241965	47817696	260.164	239.194m
20)	L4 AR-1242-5	6.707	5.659	9824013	58740602	248.605m	248.450

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075898.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2025 12:04
Operator : YP\AJ
Sample : AR1242ICC250
Misc :
ALS Vial : 13 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
AR1242ICC250

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/22/2025
Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 17:06:44 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 21 16:53:53 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
 Data File : PP075899.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Oct 2025 12:21
 Operator : YP\AJ
 Sample : AR1242ICC050
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1242ICC050

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 10/22/2025
 Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 21 17:10:30 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Oct 21 16:53:53 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.646	3.783	7865219	30617105	5.112	4.301m
2)	SA Decachlor...	10.422	8.783	6789035	46546691	5.454	5.125
Target Compounds							
16)	L4 AR-1242-1	5.798	4.882	2563442	27339514	53.509	47.567
17)	L4 AR-1242-2	5.820	4.939	3827668	13248287	54.069	47.740
18)	L4 AR-1242-3	5.883	5.054	2714945	8403556	59.484	53.386
19)	L4 AR-1242-4	5.979	5.138	1985121	6837014	54.597	36.934 #
20)	L4 AR-1242-5	6.708	5.660	2423381	10273735	59.725m	44.622 #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075899.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2025 12:21
Operator : YP\AJ
Sample : AR1242ICC050
Misc :
ALS Vial : 14 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
AR1242ICC050

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/22/2025
Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 17:10:30 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 21 16:53:53 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
 Data File : PP075900.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Oct 2025 12:41
 Operator : YP\AJ
 Sample : AR1248ICC1000
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1248ICC1000

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 10/22/2025
 Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 21 14:17:09 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Oct 21 14:13:23 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.647	3.780	146.9E6	734.4E6	97.039	99.403
2) SA Decachlor...	10.426	8.783	114.1E6	899.5E6	95.681	98.496
Target Compounds						
21) L5 AR-1248-1	5.799	4.876	33471805	362.9E6	950.427	965.133
22) L5 AR-1248-2	6.071	5.095	44816839	226.9E6	951.326	964.032
23) L5 AR-1248-3	6.274	5.136	49477903	258.9E6	894.849	947.454m
24) L5 AR-1248-4	6.672	5.309	60435029	264.9E6	944.739	969.626
25) L5 AR-1248-5	6.711	5.698	58134534	532.0E6	941.586	994.127

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075900.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2025 12:41
Operator : YP\AJ
Sample : AR1248ICC1000
Misc :
ALS Vial : 15 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
AR1248ICC1000

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/22/2025
Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 14:17:09 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 21 14:13:23 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
 Data File : PP075901.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Oct 2025 12:57
 Operator : YP\AJ
 Sample : AR1248ICC750
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1248ICC750

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 10/22/2025
 Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 21 14:20:31 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Oct 21 14:13:23 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.644	3.784	114.7E6	562.2E6	75.534	75.729
2) SA Decachlor...	10.422	8.784	89301082	688.2E6	74.939	75.236
Target Compounds						
21) L5 AR-1248-1	5.797	4.879	26839141	281.7E6	758.019	749.338
22) L5 AR-1248-2	6.068	5.099	35688032	181.4E6	755.016	763.662
23) L5 AR-1248-3	6.271	5.139	39334925	200.6E6	723.821	728.506m
24) L5 AR-1248-4	6.670	5.311	48125398	216.4E6	751.539	777.551
25) L5 AR-1248-5	6.709	5.701	46284863	398.2E6	749.774	746.121

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075901.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2025 12:57
Operator : YP\AJ
Sample : AR1248ICC750
Misc :
ALS Vial : 16 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
AR1248ICC750

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/22/2025
Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 14:20:31 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 21 14:13:23 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
 Data File : PP075902.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Oct 2025 13:13
 Operator : YP\AJ
 Sample : AR1248ICC500
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1248ICC500

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 10/22/2025
 Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 21 14:13:37 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Oct 21 14:13:23 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.645	3.784	77916247	371.6E6	50.000	50.000
2)	SA Decachlor...	10.424	8.784	62181174	463.5E6	50.000	50.000
Target Compounds							
21)	L5 AR-1248-1	5.798	4.877	18481760	194.6E6	500.000	500.000
22)	L5 AR-1248-2	6.069	5.098	24701455	121.9E6	500.000	500.000
23)	L5 AR-1248-3	6.273	5.138	30552968	143.0E6	500.000	491.112m
24)	L5 AR-1248-4	6.671	5.312	33752591	140.8E6	500.000	500.000
25)	L5 AR-1248-5	6.710	5.701	32673818	269.1E6	500.000	500.000

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075902.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2025 13:13
Operator : YP\AJ
Sample : AR1248ICC500
Misc :
ALS Vial : 17 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
AR1248ICC500

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/22/2025
Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 14:13:37 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 21 14:13:23 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
 Data File : PP075903.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Oct 2025 13:30
 Operator : YP\AJ
 Sample : AR1248ICC250
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1248ICC250

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 10/22/2025
 Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 21 14:24:34 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Oct 21 14:13:23 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.646	3.783	40662408	178.1E6	26.305	24.238
2) SA Decachlor...	10.423	8.784	32672168	226.3E6	26.770	24.802
Target Compounds						
21) L5 AR-1248-1	5.799	4.878	9780476	97652835	269.170	257.276
22) L5 AR-1248-2	6.070	5.099	13110467	64450290	269.977	265.662
23) L5 AR-1248-3	6.273	5.138	14333820	75193193	260.182	273.023m
24) L5 AR-1248-4	6.671	5.312	18043522	71202640	273.096	254.322
25) L5 AR-1248-5	6.710	5.701	17824378	145.0E6	277.971	265.917

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075903.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2025 13:30
Operator : YP\AJ
Sample : AR1248ICC250
Misc :
ALS Vial : 18 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
AR1248ICC250

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/22/2025
Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 14:24:34 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 21 14:13:23 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
 Data File : PP075904.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Oct 2025 13:46
 Operator : YP\AJ
 Sample : AR1248ICC050
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1248ICC050

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 10/22/2025
 Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 21 14:29:55 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Oct 21 14:13:23 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.645	3.780	7254273	27080772	4.751	3.914m
2) SA Decachlor...	10.422	8.783	6253432	36720401	5.099	4.188
Target Compounds						
21) L5 AR-1248-1	5.798	4.873	1700512	17991268	47.407	47.898
22) L5 AR-1248-2	6.069	5.101	2473817	12956514	50.751	52.688
23) L5 AR-1248-3	6.273	5.138	2547901	16065921	46.953	56.000
24) L5 AR-1248-4	6.669	5.313	3376964	14683789	49.218m	51.704m
25) L5 AR-1248-5	6.708	5.730	3390515	20852592	50.206m	38.783m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075904.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2025 13:46
Operator : YP\AJ
Sample : AR1248ICC050
Misc :
ALS Vial : 19 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
AR1248ICC050

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/22/2025
Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 14:29:55 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 21 14:13:23 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
 Data File : PP075905.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Oct 2025 14:02
 Operator : YP\AJ
 Sample : AR1254ICC1000
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254ICC1000

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 10/22/2025
 Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 21 15:57:57 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Oct 21 15:24:42 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.646	3.783	149.9E6	746.7E6	97.273	98.623
2)	SA Decachlor...	10.424	8.784	119.5E6	929.4E6	95.942	97.673
Target Compounds							
26)	L6 AR-1254-1	6.646	5.662	66261960	570.9E6	872.670m	970.606
27)	L6 AR-1254-2	6.864	5.809	87851253	484.3E6	952.211	965.216
28)	L6 AR-1254-3	7.226	6.211	92035771	866.8E6	959.587	985.524
29)	L6 AR-1254-4	7.509	6.437	73136868	656.6E6	959.001	987.570
30)	L6 AR-1254-5	7.925	6.854	85440842	714.9E6	961.668	997.772

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075905.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2025 14:02
Operator : YP\AJ
Sample : AR1254ICC1000
Misc :
ALS Vial : 20 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
AR1254ICC1000

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/22/2025
Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 15:57:57 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 21 15:24:42 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
 Data File : PP075906.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Oct 2025 14:18
 Operator : YP\AJ
 Sample : AR1254ICC750
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254ICC750

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 10/22/2025
 Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 21 16:03:45 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Oct 21 15:24:42 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.644	3.783	116.9E6	582.0E6	75.579	76.230
2)	SA Decachlor...	10.423	8.784	93392685	723.9E6	74.989	75.714
Target Compounds							
26)	L6 AR-1254-1	6.645	5.661	53923759	447.8E6	731.053m	757.518
27)	L6 AR-1254-2	6.863	5.808	69427935	372.6E6	751.680	745.041
28)	L6 AR-1254-3	7.225	6.210	72675518	670.2E6	755.137	757.967
29)	L6 AR-1254-4	7.507	6.436	57517400	515.4E6	752.790	766.592
30)	L6 AR-1254-5	7.923	6.851	67290078	595.4E6	754.900	785.342m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075906.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2025 14:18
Operator : YP\AJ
Sample : AR1254ICC750
Misc :
ALS Vial : 21 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
AR1254ICC750

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/22/2025
Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 16:03:45 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 21 15:24:42 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
 Data File : PP075907.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Oct 2025 14:35
 Operator : YP\AJ
 Sample : AR1254ICC500
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254ICC500

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 10/22/2025
 Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 21 15:24:59 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Oct 21 15:24:42 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.647	3.785	79163213	383.8E6	50.000	50.000
2) SA Decachlor...	10.424	8.785	64802936	486.9E6	50.000	50.000
Target Compounds						
26) L6 AR-1254-1	6.648	5.663	39135010	302.7E6	480.703m	500.000
27) L6 AR-1254-2	6.866	5.810	48334685	259.6E6	500.000	500.000
28) L6 AR-1254-3	7.228	6.212	49893996	446.1E6	500.000	500.000
29) L6 AR-1254-4	7.510	6.438	39695143	336.6E6	500.000	500.000
30) L6 AR-1254-5	7.926	6.854	46126096	359.1E6	500.000	500.000

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075907.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2025 14:35
Operator : YP\AJ
Sample : AR1254ICC500
Misc :
ALS Vial : 22 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
AR1254ICC500

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/22/2025
Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 15:24:59 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 21 15:24:42 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
 Data File : PP075908.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Oct 2025 14:51
 Operator : YP\AJ
 Sample : AR1254ICC250
 Misc :
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254ICC250

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 10/22/2025
 Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 21 16:07:49 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Oct 21 15:24:42 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.646	3.784	40353962	181.6E6	25.803	24.079
2) SA Decachlor...	10.422	8.784	33314499	235.8E6	26.290	24.742
Target Compounds						
26) L6 AR-1254-1	6.647	5.662	20616230	146.6E6	278.580m	248.519
27) L6 AR-1254-2	6.864	5.809	25702390	128.2E6	270.622	254.691
28) L6 AR-1254-3	7.226	6.210	26497010	218.4E6	268.520	247.733
29) L6 AR-1254-4	7.508	6.436	20991838	159.2E6	268.108	239.944
30) L6 AR-1254-5	7.925	6.853	23753414	169.8E6	262.160	233.749

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075908.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2025 14:51
Operator : YP\AJ
Sample : AR1254ICC250
Misc :
ALS Vial : 23 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
AR1254ICC250

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/22/2025
Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 16:07:49 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 21 15:24:42 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
 Data File : PP075909.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Oct 2025 15:07
 Operator : YP\AJ
 Sample : AR1254ICC050
 Misc :
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254ICC050

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 10/22/2025
 Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 21 16:12:03 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Oct 21 15:24:42 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.642	3.780	8604114	35027317	5.393	4.744m
2) SA Decachlor...	10.419	8.782	7574299	50050602	5.752	5.200
Target Compounds						
26) L6 AR-1254-1	6.644	5.659	5028603	23004893	64.906	40.789 #
27) L6 AR-1254-2	6.860	5.807	6001024	31864124	60.020	60.390
28) L6 AR-1254-3	7.222	6.206	6253065	36465652	60.152	42.846 #
29) L6 AR-1254-4	7.504	6.433	5244227	24763386	62.720	39.321 #
30) L6 AR-1254-5	7.921	6.850	4771352	34659929	52.106	48.147

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075909.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2025 15:07
Operator : YP\AJ
Sample : AR1254ICC050
Misc :
ALS Vial : 24 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
AR1254ICC050

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/22/2025
Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 16:12:03 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 21 15:24:42 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
 Data File : PP075910.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Oct 2025 15:40
 Operator : YP\AJ
 Sample : AR1262ICC500
 Misc :
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1262ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 21 16:21:11 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Oct 21 16:20:51 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.647	3.785	76058633	352.0E6	50.000	50.000
2) SA Decachlor...	10.425	8.786	61011705	448.1E6	50.000	50.000
Target Compounds						
36) L8 AR-1262-1	8.232	6.893	42013007	493.5E6	500.000	500.000
37) L8 AR-1262-2	8.557	7.150	81834582	351.6E6	500.000	500.000
38) L8 AR-1262-3	8.883	7.672	60893486	303.5E6	500.000	500.000
39) L8 AR-1262-4	8.971	7.736	46055772	550.7E6	500.000	500.000
40) L8 AR-1262-5	9.642	8.231	33912301	230.7E6	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075910.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2025 15:40
Operator : YP\AJ
Sample : AR1262ICC500
Misc :
ALS Vial : 25 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1262ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 16:21:11 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 21 16:20:51 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
 Data File : PP075911.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Oct 2025 15:56
 Operator : YP\AJ
 Sample : AR1268ICC1000
 Misc :
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1268ICC1000

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 21 18:02:42 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Oct 21 17:59:29 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.646	3.784	151.5E6	749.9E6	97.157	99.641
2) SA Decachlor...	10.424	8.785	198.8E6	1671.8E6	95.543	99.055
Target Compounds						
41) L9 AR-1268-1	8.878	7.673	186.7E6	1831.6E6	951.896	992.615
42) L9 AR-1268-2	8.973	7.737	172.5E6	1800.4E6	956.477	1001.376
43) L9 AR-1268-3	9.213	7.944	145.2E6	1369.8E6	961.447	988.116
44) L9 AR-1268-4	9.640	8.230	71036137	542.8E6	963.668	1002.321
45) L9 AR-1268-5	10.071	8.526	446.4E6	3855.5E6	968.612	979.920



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075911.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2025 15:56
Operator : YP\AJ
Sample : AR1268ICC1000
Misc :
ALS Vial : 26 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1268ICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 18:02:42 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 21 17:59:29 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
 Data File : PP075912.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Oct 2025 16:12
 Operator : YP\AJ
 Sample : AR1268ICC750
 Misc :
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1268ICC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 21 18:06:09 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Oct 21 17:59:29 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.646	3.783	116.2E6	564.9E6	74.674	75.037
2) SA Decachlor...	10.423	8.784	155.7E6	1249.5E6	74.863	74.351
Target Compounds						
41) L9 AR-1268-1	8.877	7.671	146.3E6	1360.1E6	747.184	741.339
42) L9 AR-1268-2	8.973	7.736	134.4E6	1329.3E6	746.666	742.894
43) L9 AR-1268-3	9.211	7.943	112.7E6	1019.0E6	747.385	739.990
44) L9 AR-1268-4	9.640	8.229	56104264	402.6E6	757.366	745.626
45) L9 AR-1268-5	10.070	8.525	344.2E6	2925.3E6	747.861	745.651



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075912.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2025 16:12
Operator : YP\AJ
Sample : AR1268ICC750
Misc :
ALS Vial : 27 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1268ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 18:06:09 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 21 17:59:29 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
 Data File : PP075913.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Oct 2025 16:29
 Operator : YP\AJ
 Sample : AR1268ICC500
 Misc :
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1268ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 21 17:59:51 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Oct 21 17:59:29 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.647	3.785	80186895	377.7E6	50.000	50.000
2) SA Decachlor...	10.424	8.785	108.7E6	851.9E6	50.000	50.000
Target Compounds						
41) L9 AR-1268-1	8.878	7.673	102.8E6	929.4E6	500.000	500.000
42) L9 AR-1268-2	8.973	7.737	94092052	897.7E6	500.000	500.000
43) L9 AR-1268-3	9.213	7.943	78447140	701.4E6	500.000	500.000
44) L9 AR-1268-4	9.641	8.229	38196242	270.1E6	500.000	500.000
45) L9 AR-1268-5	10.070	8.526	237.7E6	2006.7E6	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075913.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2025 16:29
Operator : YP\AJ
Sample : AR1268ICC500
Misc :
ALS Vial : 28 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1268ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 17:59:51 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 21 17:59:29 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
 Data File : PP075914.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Oct 2025 16:45
 Operator : YP\AJ
 Sample : AR1268ICC250
 Misc :
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1268ICC250

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 21 18:11:19 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Oct 21 17:59:29 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.648	3.785	39777145	176.8E6	25.420	23.849
2) SA Decachlor...	10.427	8.786	55292786	404.9E6	26.176	24.316
Target Compounds						
41) L9 AR-1268-1	8.880	7.673	52505558	433.0E6	263.374	239.361
42) L9 AR-1268-2	8.975	7.737	47519594	415.5E6	260.422	236.426
43) L9 AR-1268-3	9.214	7.943	39522518	339.7E6	258.947	247.491
44) L9 AR-1268-4	9.642	8.231	19156803	125.4E6	256.397	236.500
45) L9 AR-1268-5	10.073	8.527	120.0E6	927.0E6	257.945	239.576



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075914.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2025 16:45
Operator : YP\AJ
Sample : AR1268ICC250
Misc :
ALS Vial : 29 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1268ICC250

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 18:11:19 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 21 17:59:29 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
 Data File : PP075915.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Oct 2025 17:01
 Operator : YP\AJ
 Sample : AR1268ICC050
 Misc :
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1268ICC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 21 18:15:07 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Oct 21 17:59:29 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.646	3.785	8297245	34278784	5.239	4.694
2) SA Decachlor...	10.424	8.786	11780168	79254993	5.451	4.805
Target Compounds						
41) L9 AR-1268-1	8.878	7.671	11762673	90866583	56.952	50.184
42) L9 AR-1268-2	8.973	7.736	10571240	82258397	56.152	47.408
43) L9 AR-1268-3	9.212	7.944	8270442	67275006	53.294	49.210
44) L9 AR-1268-4	9.641	8.228	4718687	23283603	59.998	44.997 #
45) L9 AR-1268-5	10.071	8.525	27065230	169.1E6	56.340	44.825



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075915.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2025 17:01
Operator : YP\AJ
Sample : AR1268ICC050
Misc :
ALS Vial : 30 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1268ICC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 18:15:07 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 21 17:59:29 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
 Data File : PP075916.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Oct 2025 17:34
 Operator : YP\AJ
 Sample : PP102125ICV500
 Misc :
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 ICVPP102125

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 21 18:26:54 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Oct 21 18:24:24 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm


Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.646	3.784	73895870	340.9E6	51.708	51.507
2) SA Decachlor...	10.423	8.784	58947888	432.2E6	54.757	52.559
Target Compounds						
3) L1 AR-1016-1	5.799	4.879	30579076	390.9E6	517.184	528.086
4) L1 AR-1016-2	5.820	4.937	46283521	183.5E6	523.308	536.724
5) L1 AR-1016-3	5.883	5.057	29014100	102.8E6	524.184	500.470
6) L1 AR-1016-4	5.980	5.097	23843232	102.7E6	524.717	549.750
7) L1 AR-1016-5	6.273	5.312	22384300	107.6E6	515.702	472.171
31) L7 AR-1260-1	7.391	6.340	38313477	286.2E6	557.233	522.496
32) L7 AR-1260-2	7.644	6.526	47189702	452.3E6	526.804	555.781
33) L7 AR-1260-3	8.002	6.680	37186118	328.9E6	536.700	584.366
34) L7 AR-1260-4	8.230	7.149	40498377	308.8E6	590.602	594.624
35) L7 AR-1260-5	8.556	7.389	79566551	796.8E6	537.788	587.781

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Instrument :
ECD_P
ClientSampleId :
ICVPP102125

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
 Data File : PP075917.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Oct 2025 18:06
 Operator : YP\AJ
 Sample : AR1242ICV500
 Misc :
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 ICVPP102125AR1242

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 10/24/2025
 Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 22 08:16:42 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Oct 21 18:24:24 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.646	3.784	79486082	365.2E6	55.620	55.173
2) SA Decachlor...	10.423	8.785	64567530	484.9E6	59.978m	58.969
Target Compounds						
16) L4 AR-1242-1	5.799	4.879	24532478	301.7E6	512.088	524.891
17) L4 AR-1242-2	5.820	4.937	37086619	138.8E6	523.883	499.981
18) L4 AR-1242-3	5.883	5.057	22990427	78773054	503.714	500.427
19) L4 AR-1242-4	5.980	5.139	18909181	99357590	520.064	536.743m
20) L4 AR-1242-5	6.710	5.660	21676788	123.9E6	529.426	538.030

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075917.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2025 18:06
Operator : YP\AJ
Sample : AR1242ICV500
Misc :
ALS Vial : 32 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
ICVPP102125AR1242

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/24/2025
Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 22 08:16:42 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 21 18:24:24 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
 Data File : PP075918.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Oct 2025 18:39
 Operator : YP\AJ
 Sample : AR1248ICV500
 Misc :
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 ICVPP102125AR1248

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 10/24/2025
 Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 22 08:17:44 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Oct 21 18:24:24 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.647	3.784	77354608	375.4E6	54.128	56.717
2) SA Decachlor...	10.425	8.785	62090346	474.3E6	57.677	57.676
Target Compounds						
21) L5 AR-1248-1	5.799	4.878	18454423	193.8E6	514.472	515.973
22) L5 AR-1248-2	6.070	5.099	24718037	124.2E6	507.095	505.124
23) L5 AR-1248-3	6.273	5.138	27349933	132.7E6	504.009	462.616m
24) L5 AR-1248-4	6.671	5.312	33802442	144.9E6	509.348	512.423
25) L5 AR-1248-5	6.711	5.701	33065578	279.8E6	509.794	538.456

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075918.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2025 18:39
Operator : YP\AJ
Sample : AR1248ICV500
Misc :
ALS Vial : 33 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
ICVPP102125AR1248

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/24/2025
Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 22 08:17:44 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 21 18:24:24 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
 Data File : PP075919.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Oct 2025 08:49
 Operator : YP\AJ
 Sample : AR1254ICV500
 Misc :
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 ICVPP102125AR1254

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 10/24/2025
 Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 22 09:02:43 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Oct 21 18:24:24 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.654	3.784	73513288	361.3E6	51.441	54.591
2) SA Decachlor...	10.439	8.789	61994754	468.2E6	57.588	56.937
Target Compounds						
26) L6 AR-1254-1	6.657	5.662	33305659	337.3E6	416.876	597.997m#
27) L6 AR-1254-2	6.873	5.811	45460472	248.5E6	454.676	468.721
28) L6 AR-1254-3	7.235	6.214	47667437	420.5E6	458.540	494.093
29) L6 AR-1254-4	7.517	6.439	37597676	337.3E6	449.659	535.605
30) L6 AR-1254-5	7.934	6.855	44185805	348.6E6	482.532	484.302

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075919.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Oct 2025 08:49
Operator : YP\AJ
Sample : AR1254ICV500
Misc :
ALS Vial : 34 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
ICVPP102125AR1254

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/24/2025
Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 22 09:02:43 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 21 18:24:24 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
 Data File : PP075920.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Oct 2025 09:07
 Operator : YP\AJ
 Sample : AR1268ICV500
 Misc :
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 ICVPP102125AR1268

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 24 15:40:49 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Oct 24 15:38:49 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.648	3.782	80645176	382.7E6	50.921	52.402
2)	SA Decachlor...	10.430	8.784	112.8E6	913.8E6	52.199	55.404
Target Compounds							
41)	L9 AR-1268-1	8.881	7.671	106.5E6	989.5E6	515.825	546.466
42)	L9 AR-1268-2	8.976	7.736	98184801	960.8E6	521.533	553.714
43)	L9 AR-1268-3	9.216	7.942	81566200	749.4E6	525.609	548.157
44)	L9 AR-1268-4	9.645	8.228	39672536	280.8E6	504.438	542.576
45)	L9 AR-1268-5	10.076	8.525	246.0E6	2111.9E6	512.030	559.930



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075920.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Oct 2025 09:07
Operator : YP\AJ
Sample : AR1268ICV500
Misc :
ALS Vial : 35 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
ICVPP102125AR1268

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 24 15:40:49 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 24 15:38:49 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Calibration Times: 08:57 16:43

LAB FILE ID:	RT 1000 =	<u>PQ070973.D</u>	RT 750 =	<u>PQ070974.D</u>
	RT 500 =	PO070975.D	RT 250 =	PO070976.D
			RT 050 =	PO070977.D

[illegible]

[illegible]

Calibration Times: 08:57 16:43

LAB FILE ID:	RT 1000 =	<u>PQ070973.D</u>	RT 750 =	<u>PQ070974.D</u>
	RT 500 =	PO070975.D	RT 250 =	PO070976.D
			RT 050 =	PO070977.D

[illegible]

[illegible]

CALIBRATION FACTOR OF INITIAL CALIBRATION

Lab Name: Alliance
Lab Code: ACE
Instrument ID: ECD_Q

Contract: MATR02
SDG NO.: Q3450

Calibration Date(s): 10/10/2025 10/10/2025
Calibration Times: 08:57 16:43

GC Column: ZB-MR1 **ID:** 0.32 (mm)

LAB FILE ID:		CF 1000 = <u>PQ070973.D</u>		CF 750 = <u>PQ070974.D</u>			
CF 500 = <u>PQ070975.D</u>		CF 250 = <u>PQ070976.D</u>		CF 050 = <u>PQ070977.D</u>			
COMPOUND		CF 1000	CF 750	CF 500	CF 250	CF 050	% RSD
Aroclor-1016-1	(1)	152632258	161381460	169566884	175492100	161647220	5
Aroclor-1016-2	(2)	235274296	256838307	264605432	271518004	253235720	5
Aroclor-1016-3	(3)	145675938	156782704	164416996	169596776	159119400	6
Aroclor-1016-4	(4)	120162951	129378400	134848252	138303760	126953020	5
Aroclor-1016-5	(5)	117231883	128545740	133336348	136306768	128593380	6
Aroclor-1260-1	(1)	208928914	234195365	241877518	247859876	239493660	6
Aroclor-1260-2	(2)	243695794	275266251	287710272	296720404	279652200	7
Aroclor-1260-3	(3)	178017013	196424459	204653606	209590628	207855920	6
Aroclor-1260-4	(4)	191556315	214263107	223690164	228814336	216092480	7
Aroclor-1260-5	(5)	400986941	436898331	447507618	451939420	431553500	5
Decachlorobiphenyl		2603997920	2888430733	3017037660	3097201720	2913248400	6
Tetrachloro-m-xylene		4048991050	4420325093	4447782420	4497118040	4337178200	4
Aroclor-1242-1	(1)	117929900	120501580	121917368	124177888	110233240	5
Aroclor-1242-2	(2)	179238730	188762288	194471730	197719720	193371880	4
Aroclor-1242-3	(3)	111298009	117265485	120728326	123754708	121434700	4
Aroclor-1242-4	(4)	91702396	97376287	98944784	101196952	108558560	6
Aroclor-1242-5	(5)	63322787	67097277	82718820	102665812	77804800	20
Decachlorobiphenyl		3082568640	3243406933	3352805760	3359769360	3286790800	3
Tetrachloro-m-xylene		4441893010	4577290613	4623065280	4605855840	4702516200	2
Aroclor-1248-1	(1)	89951061	92802913	99278242	97399904	67891380	14
Aroclor-1248-2	(2)	126849870	131764987	138900528	142257364	142462060	5
Aroclor-1248-3	(3)	152980867	157544657	166146850	167481096	162764480	4
Aroclor-1248-4	(4)	167077290	173126060	182827440	186713196	157721500	7
Aroclor-1248-5	(5)	163667698	168136284	132211082	121389384	159185140	14
Decachlorobiphenyl		3177348440	3248518907	3402766540	3434790520	3353916400	3
Tetrachloro-m-xylene		4546508980	4599154480	4739413180	4685828360	4518221600	2
Aroclor-1254-1	(1)	186317660	198212097	200308250	222421948	174131700	9
Aroclor-1254-2	(2)	272465681	301191412	300558232	318683640	259516480	8
Aroclor-1254-3	(3)	294493285	320137152	318530722	335107472	267769020	9
Aroclor-1254-4	(4)	212636724	232647637	229798130	237628972	184991000	10
Aroclor-1254-5	(5)	229095618	241104957	242576814	249415576	197285560	9
Decachlorobiphenyl		3135993520	3237057933	3300114540	3432442200	2724290800	9
Tetrachloro-m-xylene		4449551480	4584042107	4604497820	4625575480	3701892800	9
Aroclor-1268-1	(1)	555593291	572438855	583365466	598015136	603747960	3



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION FACTOR OF INITIAL CALIBRATION

Aroclor-1268-2	(2)	510325576	524057337	534075228	546100468	539884140	530888550	3
Aroclor-1268-3	(3)	416538407	427532203	436646098	443325192	442459640	433300308	3
Aroclor-1268-4	(4)	179373252	185531188	189624920	193669728	190155160	187670850	3
Aroclor-1268-5	(5)	1283167757	1308010297	1321380362	1326514116	1318957680	1311606042	1
Decachlorobiphenyl		5757514290	5913646120	6037365900	6192444040	6223042400	6024802550	3
Tetrachloro-m-xylene		4734874420	4871269707	4916225920	4964521920	4932484000	4883875193	2

CALIBRATION FACTOR OF INITIAL CALIBRATION

Lab Name: Alliance
Lab Code: ACE
Instrument ID: ECD_Q

Contract: MATR02
SDG NO.: Q3450

Calibration Date(s): 10/10/2025 10/10/2025
Calibration Times: 08:57 16:43

GC Column: ZB-MR2 **ID:** 0.32 (mm)

LAB FILE ID:		CF 1000 = <u>PQ070973.D</u>		CF 750 = <u>PQ070974.D</u>			
CF 500 = <u>PQ070975.D</u>		CF 250 = <u>PQ070976.D</u>		CF 050 = <u>PQ070977.D</u>			
COMPOUND		CF 1000	CF 750	CF 500	CF 250	CF 050	% RSD
Aroclor-1016-1	(1)	198094739	217218485	228189888	237783008	238723760	224001976
Aroclor-1016-2	(2)	306974778	336928535	345611618	359075648	352024660	340123048
Aroclor-1016-3	(3)	164197912	180535471	187035142	195263012	182262860	181858879
Aroclor-1016-4	(4)	139678855	154088441	161162796	170890484	161242760	157412667
Aroclor-1016-5	(5)	171310856	187991799	196963118	205633060	197310680	191841903
Aroclor-1260-1	(1)	322104344	352779140	365742242	382158164	379245720	360405922
Aroclor-1260-2	(2)	392921251	432078265	450339936	466790948	457838380	439993756
Aroclor-1260-3	(3)	374441100	409997701	423595070	436050024	430913160	414999411
Aroclor-1260-4	(4)	324169048	356082777	368301826	380117112	370143240	359762801
Aroclor-1260-5	(5)	773923407	822481391	873089514	880740576	851175200	840282018
Decachlorobiphenyl		4511855410	4959881947	5165429800	5359918520	5250729800	5049563095
Tetrachloro-m-xylene		4790824740	5183285133	5313707940	5302586120	4809696600	5080020107
Aroclor-1242-1	(1)	149120136	160366853	160219200	165386980	179219740	162862582
Aroclor-1242-2	(2)	228406508	237737935	244055020	248772624	247903320	241375081
Aroclor-1242-3	(3)	123267302	131780304	132089372	135429308	136535900	131820437
Aroclor-1242-4	(4)	107929318	113409847	120761922	128344628	102401880	114569519
Aroclor-1242-5	(5)	175313049	196019013	175387368	167275268	156704980	174139936
Decachlorobiphenyl		5194557400	5483323560	5616478880	5698756520	6104877800	5619598832
Tetrachloro-m-xylene		5511103540	5712989200	5725543240	5446452960	5436162600	5566450308
Aroclor-1248-1	(1)	118147108	120641916	127443132	132200888	134031720	126492953
Aroclor-1248-2	(2)	172958749	177705395	189325858	196101732	206725040	188563355
Aroclor-1248-3	(3)	167593672	172934769	178216680	182062792	197467520	179655087
Aroclor-1248-4	(4)	207432918	214191484	230649658	232116576	237264740	224331075
Aroclor-1248-5	(5)	211078475	216720032	243430792	250826296	242157620	232842643
Decachlorobiphenyl		5357284330	5460780160	5730809660	5844375240	6130897000	5704829278
Tetrachloro-m-xylene		5499268270	5571294053	5722936200	5664492400	5408993600	5573396905
Aroclor-1254-1	(1)	326099704	341297420	345505280	377987212	295169840	337211891
Aroclor-1254-2	(2)	283781194	302336180	304849560	326229636	268575420	297154398
Aroclor-1254-3	(3)	455841488	477545104	484627186	501619228	415693040	467065209
Aroclor-1254-4	(4)	291449906	303039177	308578712	317073776	256115840	295251482
Aroclor-1254-5	(5)	417470468	427079459	439495106	452352324	365350180	420349507
Decachlorobiphenyl		5360763820	5577066440	5716894620	5972893640	5256420800	5576807864
Tetrachloro-m-xylene		5311856950	5451428933	5484489360	5522984680	4238156400	5201783265
Aroclor-1268-1	(1)	931455367	961712979	973652708	974423320	1030986200	974446115



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION FACTOR OF INITIAL CALIBRATION

Aroclor-1268-2	(2)	858872879	891158693	900594916	906701084	936258500	898717214	3
Aroclor-1268-3	(3)	746016914	769745548	779615614	791820108	821422560	781724149	4
Aroclor-1268-4	(4)	330162118	339473573	345636662	356393996	350955500	344524370	3
Aroclor-1268-5	(5)	2374681378	2422263161	2432472106	2427996872	2444443520	2420371407	1
Decachlorobiphenyl		10401085910	10736618840	10941521580	11269357960	11896046400	11048926138	5
Tetrachloro-m-xylene		5581659480	5729850200	5733063640	5828463760	5669436400	5708494696	2



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Lab Name: Alliance **Contract:** MATR02
Lab Code: ACE **SDG NO.:** Q3450
Instrument ID: ECD_Q **Date(s) Analyzed:** 10/10/2025 10/10/2025
GC Column: ZB-MR1 **ID:** 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
Aroclor-1221	500	1	3.59	3.49	3.69	59190800
		2	3.67	3.57	3.77	43146800
		3	3.74	3.64	3.84	133548000
		4	0.00			0
		5	0.00			0
Aroclor-1232	500	1	3.74	3.64	3.84	108948000
		2	4.23	4.13	4.33	57502800
		3	4.50	4.40	4.60	107099000
		4	4.65	4.55	4.75	54014200
		5	4.79	4.69	4.89	43713200
Aroclor-1262	500	1	6.84	6.74	6.94	274018000
		2	7.15	7.05	7.25	498476000
		3	7.42	7.32	7.52	328660000
		4	7.49	7.39	7.59	252340000
		5	7.99	7.89	8.09	164350000



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Lab Name: Alliance **Contract:** MATR02
Lab Code: ACE **SDG NO.:** Q3450
Instrument ID: ECD_Q **Date(s) Analyzed:** 10/10/2025 10/10/2025
GC Column: ZB-MR2 **ID:** 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
Aroclor-1221	500	1	2.98	2.88	3.08	73819200
		2	3.05	2.95	3.15	60003200
		3	3.12	3.02	3.22	171184000
		4	0.00			0
		5	0.00			0
Aroclor-1232	500	1	3.12	3.02	3.22	139342000
		2	3.79	3.69	3.89	138900000
		3	3.95	3.85	4.05	73450200
		4	4.04	3.94	4.14	63034400
		5	4.20	4.10	4.30	71607200
Aroclor-1262	500	1	5.73	5.63	5.83	452486000
		2	5.98	5.88	6.08	427702000
		3	6.50	6.40	6.60	332708000
		4	6.56	6.46	6.66	616820000
		5	7.05	6.95	7.15	302986000

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
 Data File : PQ070973.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Oct 2025 08:57
 Operator : YP\AJ
 Sample : AR1660ICC1000
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 AR1660ICC1000

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 10 10:54:32 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Oct 10 10:49:25 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.401	2.778	404.9E6	479.1E6	95.307	94.825
2) SA Decachlor...	8.496	7.534	260.4E6	451.2E6	92.652	93.246
Target Compounds						
3) L1 AR-1016-1	4.480	3.778	152.6E6	198.1E6	947.441	929.401
4) L1 AR-1016-2	4.499	3.793	235.3E6	307.0E6	941.324	940.794
5) L1 AR-1016-3	4.556	3.954	145.7E6	164.2E6	939.563	934.980
6) L1 AR-1016-4	4.646	4.002	120.2E6	139.7E6	942.413	928.587
7) L1 AR-1016-5	4.931	4.197	117.2E6	171.3E6	935.728	930.345
31) L7 AR-1260-1	6.020	5.190	208.9E6	322.1E6	926.912	936.559
32) L7 AR-1260-2	6.275	5.379	243.7E6	392.9E6	917.174	931.909
33) L7 AR-1260-3	6.626	5.520	178.0E6	374.4E6	930.393	938.406
34) L7 AR-1260-4	6.840	5.980	191.6E6	324.2E6	922.615	936.268
35) L7 AR-1260-5	7.146	6.224	401.0E6	773.9E6	945.173	939.790



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ070973.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 08:57
Operator : YP\AJ
Sample : AR1660ICC1000
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1660ICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 10:54:32 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 10 10:49:25 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
 Data File : PQ070974.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Oct 2025 09:11
 Operator : YP\AJ
 Sample : AR1660ICC750
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 AR1660ICC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 10 10:57:39 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Oct 10 10:49:25 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.400	2.778	331.5E6	388.7E6	76.997	76.286
2) SA Decachlor...	8.496	7.534	216.6E6	372.0E6	76.373	76.242
Target Compounds						
3) L1 AR-1016-1	4.480	3.778	121.0E6	162.9E6	750.874	759.502
4) L1 AR-1016-2	4.499	3.794	192.6E6	252.7E6	763.674	766.122
5) L1 AR-1016-3	4.556	3.954	117.6E6	135.4E6	755.578	763.875
6) L1 AR-1016-4	4.646	4.002	97033800	115.6E6	757.308	762.093
7) L1 AR-1016-5	4.930	4.197	96409305	141.0E6	762.905	760.395
31) L7 AR-1260-1	6.020	5.189	175.6E6	264.6E6	769.253	762.765
32) L7 AR-1260-2	6.276	5.378	206.4E6	324.1E6	767.783	762.288
33) L7 AR-1260-3	6.627	5.519	147.3E6	307.5E6	763.182	763.633
34) L7 AR-1260-4	6.840	5.980	160.7E6	267.1E6	765.822	764.087
35) L7 AR-1260-5	7.146	6.224	327.7E6	616.9E6	764.763	749.377



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ070974.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 09:11
Operator : YP\AJ
Sample : AR1660ICC750
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1660ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 10:57:39 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 10 10:49:25 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
 Data File : PQ070975.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Oct 2025 09:26
 Operator : YP\AJ
 Sample : AR1660ICC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 AR1660ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 10 10:49:52 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Oct 10 10:49:25 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.400	2.778	222.4E6	265.7E6	50.000	50.000
2) SA Decachlor...	8.496	7.533	150.9E6	258.3E6	50.000	50.000
Target Compounds						
3) L1 AR-1016-1	4.480	3.778	84783442	114.1E6	500.000	500.000
4) L1 AR-1016-2	4.499	3.793	132.3E6	172.8E6	500.000	500.000
5) L1 AR-1016-3	4.556	3.953	82208498	93517571	500.000	500.000
6) L1 AR-1016-4	4.646	4.002	67424126	80581398	500.000	500.000
7) L1 AR-1016-5	4.931	4.197	66668174	98481559	500.000	500.000
31) L7 AR-1260-1	6.019	5.189	120.9E6	182.9E6	500.000	500.000
32) L7 AR-1260-2	6.275	5.378	143.9E6	225.2E6	500.000	500.000
33) L7 AR-1260-3	6.626	5.519	102.3E6	211.8E6	500.000	500.000
34) L7 AR-1260-4	6.840	5.979	111.8E6	184.2E6	500.000	500.000
35) L7 AR-1260-5	7.145	6.224	223.8E6	436.5E6	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ070975.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 09:26
Operator : YP\AJ
Sample : AR1660ICC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1660ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 10:49:52 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 10 10:49:25 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
 Data File : PQ070976.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Oct 2025 09:40
 Operator : YP\AJ
 Sample : AR1660ICC250
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 AR1660ICC250

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 10 11:01:33 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Oct 10 10:49:25 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.400	2.778	112.4E6	132.6E6	25.824	25.753
2) SA Decachlor...	8.495	7.534	77430043	134.0E6	26.685	26.803
Target Compounds						
3) L1 AR-1016-1	4.480	3.778	43873025	59445752	266.271	269.814
4) L1 AR-1016-2	4.499	3.794	67879501	89768912	264.062	266.260
5) L1 AR-1016-3	4.557	3.954	42399194	48815753	266.464	268.576
6) L1 AR-1016-4	4.646	4.003	34575940	42722621	264.598	273.066
7) L1 AR-1016-5	4.931	4.197	34076692	51408265	264.457	269.895
31) L7 AR-1260-1	6.019	5.190	61964969	95539541	265.698	268.599
32) L7 AR-1260-2	6.275	5.379	74180101	116.7E6	268.916	267.943
33) L7 AR-1260-3	6.626	5.520	52397657	109.0E6	265.747	265.224
34) L7 AR-1260-4	6.840	5.980	57203584	95029278	266.583	266.063
35) L7 AR-1260-5	7.146	6.224	113.0E6	220.2E6	260.134	262.889



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ070976.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 09:40
Operator : YP\AJ
Sample : AR1660ICC250
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1660ICC250

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 11:01:33 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 10 10:49:25 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
 Data File : PQ070977.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Oct 2025 09:55
 Operator : YP\AJ
 Sample : AR1660ICC050
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 AR1660ICC050

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 10/13/2025
 Supervised By :mohammad ahmed 10/14/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 10 11:05:13 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Oct 10 10:49:25 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.400	2.778	21685891	24048483	4.985	4.734
2) SA Decachlor...	8.495	7.533	14566242	26253649	5.016	5.199
Target Compounds						
3) L1 AR-1016-1	4.480	3.778	8082361	11936188	49.239	53.286
4) L1 AR-1016-2	4.498	3.793	12661786	17601233	49.403	51.750
5) L1 AR-1016-3	4.557	3.953	7955970	9113143	50.000	50.111
6) L1 AR-1016-4	4.645	4.002	6347651	8062138	48.855	51.217
7) L1 AR-1016-5	4.930	4.197	6429669	9865534	49.919	51.425
31) L7 AR-1260-1	6.019	5.190	11974683	18962286	51.162m	52.614
32) L7 AR-1260-2	6.274	5.378	13982610	22891919	50.322m	52.028
33) L7 AR-1260-3	6.626	5.519	10392796	21545658	52.144	51.917
34) L7 AR-1260-4	6.840	5.980	10804624	18507162	50.281	51.443
35) L7 AR-1260-5	7.145	6.224	21577675	42558760	49.744	50.648

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ070977.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 09:55
Operator : YP\AJ
Sample : AR1660ICC050
Misc :
ALS Vial : 7 Sample Multiplier: 1



Instrument :
ECD_Q
ClientSampleId :
AR1660ICC050

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/13/2025
Supervised By :mohammad ahmed 10/14/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 11:05:13 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 10 10:49:25 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
 Data File : PQ070978.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Oct 2025 10:09
 Operator : YP\AJ
 Sample : AR1221ICC500
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 AR1221ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 10 11:55:54 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Oct 10 11:55:37 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.400	2.778	220.4E6	255.3E6	50.000	50.000
2) SA Decachlor...	8.496	7.534	149.9E6	262.1E6	50.000	50.000
Target Compounds						
8) L2 AR-1221-1	3.594	2.977	29595379	36909578	500.000	500.000
9) L2 AR-1221-2	3.672	3.053	21573372	30001568	500.000	500.000
10) L2 AR-1221-3	3.739	3.122	66773962	85591891	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ070978.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 10:09
Operator : YP\AJ
Sample : AR1221ICC500
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1221ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 11:55:54 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 10 11:55:37 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
 Data File : PQ070979.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Oct 2025 10:24
 Operator : YP\AJ
 Sample : AR1232ICC500
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 AR1232ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 10 12:00:39 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Oct 10 11:55:37 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.400	2.778	234.5E6	276.3E6	50.000	50.000
2) SA Decachlor...	8.497	7.534	163.6E6	280.6E6	50.000	50.000
Target Compounds						
11) L3 AR-1232-1	3.739	3.121	54474047	69671226	500.000	500.000
12) L3 AR-1232-2	4.226	3.793	28751418	69449831	500.000	500.000
13) L3 AR-1232-3	4.499	3.953	53549480	36725094	500.000	500.000
14) L3 AR-1232-4	4.646	4.039	27007118	31517168	500.000	500.000
15) L3 AR-1232-5	4.785	4.197	21856633	35803624	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ070979.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 10:24
Operator : YP\AJ
Sample : AR1232ICC500
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1232ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 12:00:39 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 10 11:55:37 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
 Data File : PQ070980.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Oct 2025 10:38
 Operator : YP\AJ
 Sample : AR1242ICC1000
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 AR1242ICC1000

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 10 12:08:51 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Oct 10 12:05:14 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.400	2.778	444.2E6	551.1E6	98.001	98.092
2)	SA Decachlor...	8.495	7.534	308.3E6	519.5E6	95.801	96.097
Target Compounds							
16)	L4 AR-1242-1	4.480	3.778	117.9E6	149.1E6	983.375	964.120
17)	L4 AR-1242-2	4.499	3.793	179.2E6	228.4E6	959.238	966.879
18)	L4 AR-1242-3	4.556	3.953	111.3E6	123.3E6	959.357	965.452
19)	L4 AR-1242-4	4.646	4.039	91702396	107.9E6	962.012	943.887
20)	L4 AR-1242-5	5.349	4.534	63322787	175.3E6	867.188	999.788



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ070980.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 10:38
Operator : YP\AJ
Sample : AR1242ICC1000
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1242ICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 12:08:51 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 10 12:05:14 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
 Data File : PQ070981.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Oct 2025 10:53
 Operator : YP\AJ
 Sample : AR1242ICC750
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 AR1242ICC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 10 12:12:26 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Oct 10 12:10:08 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.400	2.778	343.3E6	428.5E6	75.493	75.838
2) SA Decachlor...	8.496	7.534	243.3E6	411.2E6	75.399	75.716
Target Compounds						
16) L4 AR-1242-1	4.480	3.778	90376185	120.3E6	752.406	768.194
17) L4 AR-1242-2	4.498	3.794	141.6E6	178.3E6	755.086	753.183
18) L4 AR-1242-3	4.556	3.954	87949114	98835228	755.378	765.893
19) L4 AR-1242-4	4.646	4.039	73032215	85057385	760.690	745.897
20) L4 AR-1242-5	5.348	4.534	50322958	147.0E6	708.312	806.708



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ070981.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 10:53
Operator : YP\AJ
Sample : AR1242ICC750
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1242ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 12:12:26 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 10 12:10:08 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
 Data File : PQ070982.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Oct 2025 11:08
 Operator : YP\AJ
 Sample : AR1242ICC500
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 AR1242ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 10 12:05:29 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Oct 10 12:05:14 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.400	2.778	231.2E6	286.3E6	50.000	50.000
2) SA Decachlor...	8.496	7.533	167.6E6	280.8E6	50.000	50.000
Target Compounds						
16) L4 AR-1242-1	4.480	3.778	60958684	80109600	500.000	500.000
17) L4 AR-1242-2	4.499	3.793	97235865	122.0E6	500.000	500.000
18) L4 AR-1242-3	4.556	3.954	60364163	66044686	500.000	500.000
19) L4 AR-1242-4	4.646	4.040	49472392	60380961	500.000	500.000
20) L4 AR-1242-5	5.351	4.534	41359410	87693684	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ070982.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 11:08
Operator : YP\AJ
Sample : AR1242ICC500
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1242ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 12:05:29 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 10 12:05:14 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
 Data File : PQ070983.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Oct 2025 11:22
 Operator : YP\AJ
 Sample : AR1242ICC250
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 AR1242ICC250

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 10 12:15:55 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Oct 10 12:10:08 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.401	2.778	115.1E6	136.2E6	25.240	24.319
2) SA Decachlor...	8.497	7.534	83994234	142.5E6	25.768	25.912
Target Compounds						
16) L4 AR-1242-1	4.481	3.778	31044472	41346745	256.287	260.414
17) L4 AR-1242-2	4.499	3.794	49429930	62193156	260.092	259.416
18) L4 AR-1242-3	4.557	3.954	30938677	33857327	261.612	259.162
19) L4 AR-1242-4	4.647	4.040	25299238	32086157	259.999	272.815
20) L4 AR-1242-5	5.352	4.534	25666453	41818817	325.093	234.281 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ070983.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 11:22
Operator : YP\AJ
Sample : AR1242ICC250
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1242ICC250

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 12:15:55 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 10 12:10:08 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
 Data File : PQ070984.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Oct 2025 11:37
 Operator : YP\AJ
 Sample : AR1242ICC050
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 AR1242ICC050

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 10/13/2025
 Supervised By :mohammad ahmed 10/14/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 10 12:20:34 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Oct 10 12:10:08 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.400	2.778	23512581	27180813	5.122	4.883
2) SA Decachlor...	8.496	7.534	16433954	30524389	5.033	5.432
Target Compounds						
16) L4 AR-1242-1	4.481	3.777	5511662	8960987	46.335	57.042m
17) L4 AR-1242-2	4.499	3.793	9668594	12395166	50.697	51.291m
18) L4 AR-1242-3	4.556	3.953	6071735	6826795	51.068	51.803m
19) L4 AR-1242-4	4.646	4.039	5427928	5120094	54.521	44.579m
20) L4 AR-1242-5	5.317	4.534	3890240	7835249	49.488m	45.137m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ070984.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 11:37
Operator : YP\AJ
Sample : AR1242ICC050
Misc :
ALS Vial : 14 Sample Multiplier: 1



Instrument :
ECD_Q
ClientSampleId :
AR1242ICC050

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/13/2025
Supervised By :mohammad ahmed 10/14/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 12:20:34 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 10 12:10:08 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
 Data File : PQ070985.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Oct 2025 11:51
 Operator : YP\AJ
 Sample : AR1248ICC1000
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 AR1248ICC1000

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 10 12:40:50 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Oct 10 12:37:43 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.400	2.778	454.7E6	549.9E6	97.923	98.007
2) SA Decachlor...	8.496	7.534	317.7E6	535.7E6	96.574	96.631
Target Compounds						
21) L5 AR-1248-1	4.480	3.778	89951061	118.1E6	950.710	962.148
22) L5 AR-1248-2	4.743	4.002	126.8E6	173.0E6	954.654	954.823
23) L5 AR-1248-3	4.930	4.039	153.0E6	167.6E6	958.744	969.281
24) L5 AR-1248-4	5.318	4.197	167.1E6	207.4E6	954.987	947.004
25) L5 AR-1248-5	5.353	4.572	163.7E6	211.1E6	1106.316	928.819



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ070985.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 11:51
Operator : YP\AJ
Sample : AR1248ICC1000
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1248ICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 12:40:50 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 10 12:37:43 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
 Data File : PQ070986.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Oct 2025 12:06
 Operator : YP\AJ
 Sample : AR1248ICC750
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 AR1248ICC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 10 12:46:34 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Oct 10 12:37:43 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.400	2.778	344.9E6	417.8E6	74.527	74.644
2) SA Decachlor...	8.495	7.533	243.6E6	409.6E6	74.366	74.245
Target Compounds						
21) L5 AR-1248-1	4.480	3.777	69602185	90481437	740.364	741.181
22) L5 AR-1248-2	4.744	4.002	98823740	133.3E6	745.811	740.453
23) L5 AR-1248-3	4.931	4.039	118.2E6	129.7E6	743.646	750.086
24) L5 AR-1248-4	5.318	4.197	129.8E6	160.6E6	744.762	738.847
25) L5 AR-1248-5	5.352	4.571	126.1E6	162.5E6	815.290	726.458



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ070986.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 12:06
Operator : YP\AJ
Sample : AR1248ICC750
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1248ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 12:46:34 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 10 12:37:43 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
 Data File : PQ070987.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Oct 2025 12:20
 Operator : YP\AJ
 Sample : AR1248ICC500
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 AR1248ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 10 12:37:58 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Oct 10 12:37:43 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.400	2.778	237.0E6	286.1E6	50.000	50.000
2) SA Decachlor...	8.496	7.532	170.1E6	286.5E6	50.000	50.000
Target Compounds						
21) L5 AR-1248-1	4.480	3.778	49639121	63721566	500.000	500.000
22) L5 AR-1248-2	4.744	4.002	69450264	94662929	500.000	500.000
23) L5 AR-1248-3	4.931	4.039	83073425	89108340	500.000	500.000
24) L5 AR-1248-4	5.318	4.197	91413720	115.3E6	500.000	500.000
25) L5 AR-1248-5	5.351	4.571	66105541	121.7E6	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ070987.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 12:20
Operator : YP\AJ
Sample : AR1248ICC500
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1248ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 12:37:58 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 10 12:37:43 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
 Data File : PQ070988.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Oct 2025 12:35
 Operator : YP\AJ
 Sample : AR1248ICC250
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 AR1248ICC250

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 10/13/2025
 Supervised By :mohammad ahmed 10/14/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 10 12:50:45 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Oct 10 12:37:43 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.401	2.778	117.1E6	141.6E6	25.232	25.223
2)	SA Decachlor...	8.497	7.534	85869763	146.1E6	25.897	26.099
Target Compounds							
21)	L5 AR-1248-1	4.481	3.778	24349976	33050222	256.699	265.233
22)	L5 AR-1248-2	4.744	4.002	35564341	49025433	263.550	266.409
23)	L5 AR-1248-3	4.931	4.039	41870274	45515698	260.002	259.790
24)	L5 AR-1248-4	5.318	4.196	46678299	58029144	263.071	262.398m
25)	L5 AR-1248-5	5.351	4.570	30347346	62706574	207.360	273.971m#

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ070988.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 12:35
Operator : YP\AJ
Sample : AR1248ICC250
Misc :
ALS Vial : 18 Sample Multiplier: 1



Instrument :
ECD_Q
ClientSampleId :
AR1248ICC250

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/13/2025
Supervised By :mohammad ahmed 10/14/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 12:50:45 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 10 12:37:43 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
 Data File : PQ070989.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Oct 2025 12:50
 Operator : YP\AJ
 Sample : AR1248ICC050
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 AR1248ICC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 10 14:31:10 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Oct 10 14:30:55 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.400	2.778	22591108	27044968	4.892	4.853
2) SA Decachlor...	8.497	7.534	16769582	30654485	5.046	5.373
Target Compounds						
21) L5 AR-1248-1	4.481	3.778	3394569	6701586	37.943	52.980 #
22) L5 AR-1248-2	4.743	4.002	7123103	10336252	52.204	54.816
23) L5 AR-1248-3	4.931	4.039	8138224	9873376	50.428	54.957
24) L5 AR-1248-4	5.318	4.197	7886075	11863237	45.455	52.883
25) L5 AR-1248-5	5.352	4.571	7959257	12107881	53.447	52.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ070989.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 12:50
Operator : YP\AJ
Sample : AR1248ICC050
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1248ICC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 14:31:10 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 10 14:30:55 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
 Data File : PQ070990.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Oct 2025 13:04
 Operator : YP\AJ
 Sample : AR1254ICC1000
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 AR1254ICC1000

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 10 14:53:43 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Oct 10 14:52:12 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.400	2.778	445.0E6	531.2E6	98.289	98.401
2) SA Decachlor...	8.496	7.533	313.6E6	536.1E6	97.450	96.785
Target Compounds						
26) L6 AR-1254-1	5.292	4.533	186.3E6	326.1E6	963.814	971.106
27) L6 AR-1254-2	5.507	4.680	272.5E6	283.8E6	950.975	964.208
28) L6 AR-1254-3	5.855	5.060	294.5E6	455.8E6	960.789	969.392
29) L6 AR-1254-4	6.135	5.286	212.6E6	291.4E6	961.211	971.453
30) L6 AR-1254-5	6.543	5.690	229.1E6	417.5E6	971.418	974.299



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ070990.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 13:04
Operator : YP\AJ
Sample : AR1254ICC1000
Misc :
ALS Vial : 20 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1254ICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 14:53:43 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 10 14:52:12 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
 Data File : PQ070991.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Oct 2025 13:19
 Operator : YP\AJ
 Sample : AR1254ICC750
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 AR1254ICC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 10 14:59:55 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Oct 10 14:59:41 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.401	2.778	343.8E6	408.9E6	75.627	75.492
2) SA Decachlor...	8.496	7.534	242.8E6	418.3E6	75.295	75.344
Target Compounds						
26) L6 AR-1254-1	5.293	4.533	148.7E6	256.0E6	762.565	758.137
27) L6 AR-1254-2	5.507	4.680	225.9E6	226.8E6	775.187	763.504
28) L6 AR-1254-3	5.854	5.060	240.1E6	358.2E6	771.902	757.733
29) L6 AR-1254-4	6.134	5.286	174.5E6	227.3E6	775.397	755.024
30) L6 AR-1254-5	6.544	5.690	180.8E6	320.3E6	761.088	748.361



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ070991.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 13:19
Operator : YP\AJ
Sample : AR1254ICC750
Misc :
ALS Vial : 21 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1254ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 14:59:55 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 10 14:59:41 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
 Data File : PQ070992.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Oct 2025 13:33
 Operator : YP\AJ
 Sample : AR1254ICC500
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 AR1254ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 10 14:49:59 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Oct 10 14:49:02 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.400	2.778	230.2E6	274.2E6	50.000	50.000
2) SA Decachlor...	8.497	7.534	165.0E6	285.8E6	50.000	50.000
Target Compounds						
26) L6 AR-1254-1	5.293	4.532	100.2E6	172.8E6	500.000	500.000
27) L6 AR-1254-2	5.507	4.680	150.3E6	152.4E6	500.000	500.000
28) L6 AR-1254-3	5.855	5.060	159.3E6	242.3E6	500.000	500.000
29) L6 AR-1254-4	6.136	5.286	114.9E6	154.3E6	500.000	500.000
30) L6 AR-1254-5	6.543	5.689	121.3E6	219.7E6	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ070992.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 13:33
Operator : YP\AJ
Sample : AR1254ICC500
Misc :
ALS Vial : 22 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1254ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 14:49:59 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 10 14:49:02 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
 Data File : PQ070993.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Oct 2025 13:48
 Operator : YP\AJ
 Sample : AR1254ICC250
 Misc :
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 AR1254ICC250

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 10 15:02:42 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Oct 10 15:02:26 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.401	2.778	115.6E6	138.1E6	25.327	25.369
2) SA Decachlor...	8.497	7.534	85811055	149.3E6	26.191	26.396
Target Compounds						
26) L6 AR-1254-1	5.293	4.533	55605487	94496803	275.527	271.759
27) L6 AR-1254-2	5.507	4.679	79670910	81557409	267.151	268.017
28) L6 AR-1254-3	5.855	5.060	83776868	125.4E6	264.224	261.310
29) L6 AR-1254-4	6.135	5.286	59407243	79268444	260.355	259.866
30) L6 AR-1254-5	6.544	5.690	62353894	113.1E6	259.216	260.512



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ070993.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 13:48
Operator : YP\AJ
Sample : AR1254ICC250
Misc :
ALS Vial : 23 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1254ICC250

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 15:02:42 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 10 15:02:26 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
 Data File : PQ070994.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Oct 2025 15:15
 Operator : YP\AJ
 Sample : AR1254ICC050
 Misc :
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 AR1254ICC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 10 15:31:28 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Oct 10 15:31:02 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.400	2.777	18509464	21190782	4.213	4.074
2) SA Decachlor...	8.497	7.533	13621454	26282104	4.302	4.713
Target Compounds						
26) L6 AR-1254-1	5.294	4.533	8706585	14758492	44.358	43.766
27) L6 AR-1254-2	5.507	4.679	12975824	13428771	44.670	45.191
28) L6 AR-1254-3	5.856	5.060	13388451	20784652	43.581	44.501
29) L6 AR-1254-4	6.135	5.286	9249550	12805792	42.131	43.372
30) L6 AR-1254-5	6.544	5.690	9864278	18267509	42.538	43.458



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ070994.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 15:15
Operator : YP\AJ
Sample : AR1254ICC050
Misc :
ALS Vial : 24 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1254ICC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 15:31:28 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 10 15:31:02 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
 Data File : PQ070995.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Oct 2025 15:30
 Operator : YP\AJ
 Sample : AR1262ICC500
 Misc :
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 AR1262ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 10 15:45:51 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Oct 10 15:45:32 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.400	2.778	233.7E6	278.1E6	50.000	50.000
2) SA Decachlor...	8.497	7.533	168.1E6	302.5E6	50.000	50.000
Target Compounds						
36) L8 AR-1262-1	6.841	5.733	137.0E6	226.2E6	500.000	500.000
37) L8 AR-1262-2	7.146	5.979	249.2E6	213.9E6	500.000	500.000
38) L8 AR-1262-3	7.419	6.500	164.3E6	166.4E6	500.000	500.000
39) L8 AR-1262-4	7.492	6.561	126.2E6	308.4E6	500.000	500.000
40) L8 AR-1262-5	7.988	7.051	82174920	151.5E6	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ070995.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 15:30
Operator : YP\AJ
Sample : AR1262ICC500
Misc :
ALS Vial : 25 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1262ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 15:45:51 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 10 15:45:32 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
 Data File : PQ070996.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Oct 2025 15:44
 Operator : YP\AJ
 Sample : AR1268ICC1000
 Misc :
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 AR1268ICC1000

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 10 16:44:44 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Oct 10 16:41:42 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.400	2.778	473.5E6	558.2E6	98.121	98.662
2) SA Decachlor...	8.496	7.533	575.8E6	1040.1E6	97.627	97.468
Target Compounds						
41) L9 AR-1268-1	7.418	6.501	555.6E6	931.5E6	975.616	977.850
42) L9 AR-1268-2	7.494	6.563	510.3E6	858.9E6	977.260	976.287
43) L9 AR-1268-3	7.676	6.763	416.5E6	746.0E6	976.432	977.977
44) L9 AR-1268-4	7.988	7.051	179.4E6	330.2E6	972.218	977.102
45) L9 AR-1268-5	8.271	7.318	1283.2E6	2374.7E6	985.329	987.978



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ070996.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 15:44
Operator : YP\AJ
Sample : AR1268ICC1000
Misc :
ALS Vial : 26 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1268ICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 16:44:20 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 10 16:41:42 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
 Data File : PQ070997.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Oct 2025 15:59
 Operator : YP\AJ
 Sample : AR1268ICC750
 Misc :
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 AR1268ICC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 10 16:47:10 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Oct 10 16:41:42 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.400	2.778	365.3E6	429.7E6	75.472	75.638
2) SA Decachlor...	8.496	7.533	443.5E6	805.2E6	75.137	75.305
Target Compounds						
41) L9 AR-1268-1	7.417	6.501	429.3E6	721.3E6	752.594	754.792
42) L9 AR-1268-2	7.495	6.564	393.0E6	668.4E6	751.776	756.465
43) L9 AR-1268-3	7.677	6.763	320.6E6	577.3E6	751.101	754.528
44) L9 AR-1268-4	7.988	7.051	139.1E6	254.6E6	752.792	752.326
45) L9 AR-1268-5	8.270	7.318	981.0E6	1816.7E6	752.199	753.877



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ070997.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 15:59
Operator : YP\AJ
Sample : AR1268ICC750
Misc :
ALS Vial : 27 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1268ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 16:47:10 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 10 16:41:42 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
 Data File : PQ070998.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Oct 2025 16:13
 Operator : YP\AJ
 Sample : AR1268ICC500
 Misc :
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 AR1268ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 10 16:41:57 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Oct 10 16:41:42 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.401	2.778	245.8E6	286.7E6	50.000	50.000
2) SA Decachlor...	8.497	7.533	301.9E6	547.1E6	50.000	50.000
Target Compounds						
41) L9 AR-1268-1	7.418	6.501	291.7E6	486.8E6	500.000	500.000
42) L9 AR-1268-2	7.496	6.564	267.0E6	450.3E6	500.000	500.000
43) L9 AR-1268-3	7.677	6.763	218.3E6	389.8E6	500.000	500.000
44) L9 AR-1268-4	7.988	7.051	94812460	172.8E6	500.000	500.000
45) L9 AR-1268-5	8.271	7.318	660.7E6	1216.2E6	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ070998.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 16:13
Operator : YP\AJ
Sample : AR1268ICC500
Misc :
ALS Vial : 28 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1268ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 16:41:57 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 10 16:41:42 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
 Data File : PQ070999.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Oct 2025 16:28
 Operator : YP\AJ
 Sample : AR1268ICC250
 Misc :
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 AR1268ICC250

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 10 16:50:34 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Oct 10 16:41:42 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.400	2.778	124.1E6	145.7E6	25.476	25.482
2) SA Decachlor...	8.496	7.533	154.8E6	281.7E6	25.909	25.997
Target Compounds						
41) L9 AR-1268-1	7.418	6.500	149.5E6	243.6E6	258.947	253.674
42) L9 AR-1268-2	7.495	6.563	136.5E6	226.7E6	258.257	254.883
43) L9 AR-1268-3	7.677	6.763	110.8E6	198.0E6	257.143	256.485
44) L9 AR-1268-4	7.988	7.051	48417432	89098499	258.848	259.826
45) L9 AR-1268-5	8.270	7.317	331.6E6	607.0E6	253.196	251.413



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ070999.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 16:28
Operator : YP\AJ
Sample : AR1268ICC250
Misc :
ALS Vial : 29 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1268ICC250

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 16:50:34 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 10 16:41:42 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
 Data File : PQ071000.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Oct 2025 16:43
 Operator : YP\AJ
 Sample : AR1268ICC050
 Misc :
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 AR1268ICC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 10 17:16:32 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Oct 10 17:16:19 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.400	2.778	24662420	28347182	5.050	4.966
2) SA Decachlor...	8.497	7.533	31115212	59480232	5.165	5.383
Target Compounds						
41) L9 AR-1268-1	7.418	6.500	30187398	51549310	51.812	52.901
42) L9 AR-1268-2	7.495	6.563	26994207	46812925	50.847	52.089
43) L9 AR-1268-3	7.676	6.763	22122982	41071128	51.057	52.539
44) L9 AR-1268-4	7.989	7.050	9507758	17547775	50.662	50.933
45) L9 AR-1268-5	8.271	7.317	65947884	122.2E6	50.280	50.497



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ071000.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 16:43
Operator : YP\AJ
Sample : AR1268ICC050
Misc :
ALS Vial : 30 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1268ICC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 17:16:32 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 10 17:16:19 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
 Data File : PQ071001.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Oct 2025 16:57
 Operator : YP\AJ
 Sample : PQ101025ICV
 Misc :
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 ICVPQ101025

Manual Integrations
 APPROVED

Reviewed By :Rahul Chavli 10/31/2025
 Supervised By :Yogesh Patel 10/31/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 10 17:29:38 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Oct 10 17:24:38 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.400	2.778	237.3E6	265.1E6	54.558	52.183
2)	SA Decachlor...	8.497	7.533	168.1E6	302.1E6	57.872	59.825
Target Compounds							
3)	L1 AR-1016-1	4.480	3.777	84536347	121.7E6	515.013	543.285m
4)	L1 AR-1016-2	4.499	3.792	137.9E6	157.0E6	537.936	461.544m
5)	L1 AR-1016-3	4.556	3.953	86370882	89605925	542.809	492.722m
6)	L1 AR-1016-4	4.646	4.001	72647096	77616175	559.128	493.075m
7)	L1 AR-1016-5	4.931	4.196	71678763	95334887	556.500	496.945m
31)	L7 AR-1260-1	6.019	5.188	124.6E6	179.0E6	531.360m	496.698m
32)	L7 AR-1260-2	6.275	5.377	157.7E6	216.4E6	569.990m	491.786m
33)	L7 AR-1260-3	6.627	5.518	115.0E6	195.1E6	576.989	470.083
34)	L7 AR-1260-4	6.840	5.979	126.1E6	186.2E6	586.767	517.623
35)	L7 AR-1260-5	7.146	6.224	259.3E6	433.0E6	597.827	515.349

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ071001.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 16:57
Operator : YP\AJ
Sample : PQ101025ICV
Misc :
ALS Vial : 31 Sample Multiplier: 1



Instrument :
ECD_Q
ClientSampleId :
ICVPQ101025

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 10/31/2025
Supervised By :Yogesh Patel 10/31/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 17:29:38 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 10 17:24:38 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
 Data File : PQ071002.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Oct 2025 17:41
 Operator : YP\AJ
 Sample : AR1242ICV500
 Misc :
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 ICVPQ101025AR1242

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 10/13/2025
 Supervised By :mohammad ahmed 10/14/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 11 01:19:16 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Oct 11 01:16:43 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.400	2.778	255.6E6	301.0E6	55.692	54.076
2) SA Decachlor...	8.496	7.534	182.2E6	328.4E6	55.793	58.440
Target Compounds						
16) L4 AR-1242-1	4.480	3.778	67446984	88802083	567.010	545.258
17) L4 AR-1242-2	4.499	3.793	106.5E6	130.9E6	558.407	542.137
18) L4 AR-1242-3	4.556	3.952	65846450	69528801	553.814	527.451m
19) L4 AR-1242-4	4.646	4.039	55324188	66246241	555.710	578.219
20) L4 AR-1242-5	5.349	4.534	35965359	88786139	456.866	509.855

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ071002.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 17:41
Operator : YP\AJ
Sample : AR1242ICV500
Misc :
ALS Vial : 32 Sample Multiplier: 1



Instrument :
ECD_Q
ClientSampleId :
ICVPQ101025AR1242

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/13/2025
Supervised By :mohammad ahmed 10/14/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 11 01:19:16 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Oct 11 01:16:43 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
 Data File : PQ071003.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Oct 2025 18:25
 Operator : YP\AJ
 Sample : AR1248ICV500
 Misc :
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 ICVPQ101025AR1248

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 10/13/2025
 Supervised By :mohammad ahmed 10/14/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 11 01:26:29 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Oct 11 01:23:25 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.400	2.778	253.0E6	292.6E6	54.792	52.505
2) SA Decachlor...	8.497	7.534	179.8E6	324.3E6	54.114	56.854
Target Compounds						
21) L5 AR-1248-1	4.480	3.777	47223685	65282112	527.847	516.093
22) L5 AR-1248-2	4.743	4.002	73173417	97020820	536.277	514.526
23) L5 AR-1248-3	4.931	4.039	87732864	87534945	543.629	487.239
24) L5 AR-1248-4	5.317	4.196	98470443	117.2E6	567.576m	522.551m
25) L5 AR-1248-5	5.349	4.571	59958766	126.5E6	402.630m	543.081 #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ071003.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 18:25
Operator : YP\AJ
Sample : AR1248ICV500
Misc :
ALS Vial : 33 Sample Multiplier: 1



Instrument :
ECD_Q
ClientSampleId :
ICVPQ101025AR1248

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/13/2025
Supervised By :mohammad ahmed 10/14/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 11 01:26:29 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Oct 11 01:23:25 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
 Data File : PQ071004.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Oct 2025 19:08
 Operator : YP\AJ
 Sample : AR1254ICV500
 Misc :
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 ICVPQ101025AR1254

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 11 01:34:52 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Oct 11 01:30:57 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.400	2.778	247.0E6	287.2E6	56.214	55.217
2) SA Decachlor...	8.498	7.533	175.2E6	316.3E6	55.323	56.717
Target Compounds						
26) L6 AR-1254-1	5.293	4.533	115.3E6	191.1E6	587.461	566.643
27) L6 AR-1254-2	5.507	4.680	167.9E6	162.6E6	578.152	547.137
28) L6 AR-1254-3	5.855	5.060	176.3E6	253.6E6	573.790	542.980
29) L6 AR-1254-4	6.134	5.286	124.7E6	161.4E6	568.016	546.628
30) L6 AR-1254-5	6.544	5.690	130.6E6	228.4E6	563.253	543.396



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ071004.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 19:08
Operator : YP\AJ
Sample : AR1254ICV500
Misc :
ALS Vial : 34 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
ICVPQ101025AR1254

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 11 01:34:52 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Oct 11 01:30:57 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
 Data File : PQ071005.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Oct 2025 19:52
 Operator : YP\AJ
 Sample : AR1268ICV500
 Misc :
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 ICVPQ101025AR1268

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 11 01:39:57 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Oct 11 01:37:12 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.401	2.777	256.6E6	296.1E6	52.549	51.872
2) SA Decachlor...	8.497	7.533	317.0E6	560.4E6	52.621	50.721
Target Compounds						
41) L9 AR-1268-1	7.418	6.500	307.3E6	498.7E6	527.367	511.825
42) L9 AR-1268-2	7.495	6.563	283.4E6	464.0E6	533.785	516.248
43) L9 AR-1268-3	7.677	6.763	233.9E6	395.9E6	539.709	506.489
44) L9 AR-1268-4	7.989	7.051	100.4E6	175.1E6	534.727	508.315
45) L9 AR-1268-5	8.271	7.318	696.8E6	1236.0E6	531.266	510.673



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ071005.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 19:52
Operator : YP\AJ
Sample : AR1268ICV500
Misc :
ALS Vial : 35 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
ICVPQ101025AR1268

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 11 01:39:57 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Oct 11 01:37:12 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: MATR02

Lab Code: ACE

SDG NO.: Q3450

Continuing Calib Date: 10/24/2025

Initial Calibration Date(s): 10/21/2025 10/21/2025

Continuing Calib Time: 18:58

Initial Calibration Time(s): 09:06 17:01

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	5.79	5.80	5.70	5.90	0.01
Aroclor-1016-2	(2)	5.81	5.82	5.72	5.92	0.01
Aroclor-1016-3	(3)	5.88	5.88	5.78	5.98	0.00
Aroclor-1016-4	(4)	5.97	5.98	5.88	6.08	0.01
Aroclor-1016-5	(5)	6.27	6.27	6.17	6.37	0.00
Aroclor-1260-1	(1)	7.38	7.39	7.29	7.49	0.01
Aroclor-1260-2	(2)	7.64	7.64	7.54	7.74	0.00
Aroclor-1260-3	(3)	8.00	8.00	7.90	8.10	0.01
Aroclor-1260-4	(4)	8.22	8.23	8.13	8.33	0.01
Aroclor-1260-5	(5)	8.55	8.55	8.45	8.65	0.00
Tetrachloro-m-xylene		4.64	4.64	4.54	4.74	0.00
Decachlorobiphenyl		10.41	10.42	10.32	10.52	0.01



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: MATR02

Lab Code: ACE

SDG NO.: Q3450

Continuing Calib Date: 10/24/2025

Initial Calibration Date(s): 10/21/2025 10/21/2025

Continuing Calib Time: 18:58

Initial Calibration Time(s): 09:06 17:01

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	4.87	4.88	4.78	4.98	0.01
Aroclor-1016-2	(2)	4.93	4.94	4.84	5.04	0.01
Aroclor-1016-3	(3)	5.05	5.05	4.95	5.15	0.00
Aroclor-1016-4	(4)	5.09	5.10	5.00	5.20	0.01
Aroclor-1016-5	(5)	5.31	5.31	5.21	5.41	0.00
Aroclor-1260-1	(1)	6.33	6.34	6.24	6.44	0.01
Aroclor-1260-2	(2)	6.52	6.53	6.43	6.63	0.01
Aroclor-1260-3	(3)	6.67	6.68	6.58	6.78	0.01
Aroclor-1260-4	(4)	7.14	7.15	7.05	7.25	0.01
Aroclor-1260-5	(5)	7.38	7.39	7.29	7.49	0.01
Tetrachloro-m-xylene		3.78	3.78	3.68	3.88	0.00
Decachlorobiphenyl		8.78	8.78	8.68	8.88	0.00



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** MATR02
Lab Code: ACE **SDG NO.:** Q3450
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/21/2025 10/21/2025

Client Sample No.: CCAL01 **Date Analyzed:** 10/24/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP076042.D **Time Analyzed:** 18:58

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.792	5.695	5.895	466.180	500.000	-6.8
Aroclor-1016-2	5.814	5.717	5.917	470.850	500.000	-5.8
Aroclor-1016-3	5.876	5.780	5.980	471.910	500.000	-5.6
Aroclor-1016-4	5.974	5.877	6.077	475.240	500.000	-5.0
Aroclor-1016-5	6.266	6.169	6.369	472.190	500.000	-5.6
Aroclor-1260-1	7.383	7.287	7.487	508.760	500.000	1.8
Aroclor-1260-2	7.637	7.540	7.740	477.450	500.000	-4.5
Aroclor-1260-3	7.995	7.899	8.099	498.370	500.000	-0.3
Aroclor-1260-4	8.222	8.127	8.327	531.900	500.000	6.4
Aroclor-1260-5	8.547	8.452	8.652	490.190	500.000	-2.0
Decachlorobiphenyl	10.411	10.319	10.519	49.890	50.000	-0.2
Tetrachloro-m-xylene	4.640	4.542	4.742	45.590	50.000	-8.8



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** MATR02
Lab Code: ACE **SDG NO.:** Q3450
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/21/2025 10/21/2025

Client Sample No.: CCAL01 **Date Analyzed:** 10/24/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP076042.D **Time Analyzed:** 18:58

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.874	4.778	4.978	465.970	500.000	-6.8
Aroclor-1016-2	4.933	4.836	5.036	486.720	500.000	-2.7
Aroclor-1016-3	5.052	4.954	5.154	475.940	500.000	-4.8
Aroclor-1016-4	5.093	4.996	5.196	498.570	500.000	-0.3
Aroclor-1016-5	5.306	5.210	5.410	466.880	500.000	-6.6
Aroclor-1260-1	6.334	6.238	6.438	497.200	500.000	-0.6
Aroclor-1260-2	6.521	6.425	6.625	510.950	500.000	2.2
Aroclor-1260-3	6.674	6.578	6.778	514.410	500.000	2.9
Aroclor-1260-4	7.142	7.047	7.247	563.770	500.000	12.8
Aroclor-1260-5	7.383	7.287	7.487	531.590	500.000	6.3
Decachlorobiphenyl	8.776	8.682	8.882	47.410	50.000	-5.2
Tetrachloro-m-xylene	3.779	3.681	3.881	46.610	50.000	-6.8

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102425\
 Data File : PP076042.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 24 Oct 2025 18:58
 Operator : YP\AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 24 23:38:56 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Oct 24 15:40:32 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.640	3.779	65154729	308.5E6	45.592	46.609
2) SA Decachlor...	10.411	8.776	53708533	389.8E6	49.891	47.409
Target Compounds						
3) L1 AR-1016-1	5.792	4.874	27563278	344.9E6	466.178	465.975
4) L1 AR-1016-2	5.814	4.933	41643707	166.4E6	470.848	486.715
5) L1 AR-1016-3	5.876	5.052	26120781	97730819	471.911	475.941
6) L1 AR-1016-4	5.974	5.093	21594789	93175401	475.235	498.569
7) L1 AR-1016-5	6.266	5.306	20495798	106.4E6	472.193	466.880
31) L7 AR-1260-1	7.383	6.334	34980707	272.3E6	508.761	497.196
32) L7 AR-1260-2	7.637	6.521	42768798	415.8E6	477.451	510.946
33) L7 AR-1260-3	7.995	6.674	34530147	289.5E6	498.367	514.409
34) L7 AR-1260-4	8.222	7.142	36473340	292.8E6	531.903	563.770
35) L7 AR-1260-5	8.547	7.383	72524197	720.6E6	490.189	531.589


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102425\
Data File : PP076042.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 24 Oct 2025 18:58
Operator : YP\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 24 23:38:56 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 24 15:40:32 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: MATR02

Lab Code: ACE

SDG NO.: Q3450

Continuing Calib Date: 10/25/2025

Initial Calibration Date(s): 10/21/2025 10/21/2025

Continuing Calib Time: 00:22

Initial Calibration Time(s): 09:06 17:01

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	5.80	5.80	5.70	5.90	0.00
Aroclor-1016-2	(2)	5.82	5.82	5.72	5.92	0.00
Aroclor-1016-3	(3)	5.88	5.88	5.78	5.98	0.00
Aroclor-1016-4	(4)	5.98	5.98	5.88	6.08	0.00
Aroclor-1016-5	(5)	6.27	6.27	6.17	6.37	0.00
Aroclor-1260-1	(1)	7.39	7.39	7.29	7.49	0.00
Aroclor-1260-2	(2)	7.64	7.64	7.54	7.74	0.00
Aroclor-1260-3	(3)	8.00	8.00	7.90	8.10	0.00
Aroclor-1260-4	(4)	8.23	8.23	8.13	8.33	0.00
Aroclor-1260-5	(5)	8.55	8.55	8.45	8.65	0.00
Tetrachloro-m-xylene		4.64	4.64	4.54	4.74	0.00
Decachlorobiphenyl		10.41	10.42	10.32	10.52	0.01



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: MATR02

Lab Code: ACE

SDG NO.: Q3450

Continuing Calib Date: 10/25/2025

Initial Calibration Date(s): 10/21/2025 10/21/2025

Continuing Calib Time: 00:22

Initial Calibration Time(s): 09:06 17:01

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Aroclor-1016-1	(1)	4.88	4.88	4.78	4.98	0.00
Aroclor-1016-2	(2)	4.94	4.94	4.84	5.04	0.00
Aroclor-1016-3	(3)	5.06	5.05	4.95	5.15	-0.01
Aroclor-1016-4	(4)	5.10	5.10	5.00	5.20	0.00
Aroclor-1016-5	(5)	5.31	5.31	5.21	5.41	0.00
Aroclor-1260-1	(1)	6.34	6.34	6.24	6.44	0.00
Aroclor-1260-2	(2)	6.52	6.53	6.43	6.63	0.01
Aroclor-1260-3	(3)	6.68	6.68	6.58	6.78	0.00
Aroclor-1260-4	(4)	7.15	7.15	7.05	7.25	0.00
Aroclor-1260-5	(5)	7.39	7.39	7.29	7.49	0.00
Tetrachloro-m-xylene		3.78	3.78	3.68	3.88	0.00
Decachlorobiphenyl		8.78	8.78	8.68	8.88	0.00



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** MATR02
Lab Code: ACE **SDG NO.:** Q3450
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/21/2025 10/21/2025

Client Sample No.: CCAL02 **Date Analyzed:** 10/25/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP076057.D **Time Analyzed:** 00:22

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.795	5.695	5.895	471.470	500.000	-5.7
Aroclor-1016-2	5.816	5.717	5.917	468.300	500.000	-6.3
Aroclor-1016-3	5.879	5.780	5.980	468.920	500.000	-6.2
Aroclor-1016-4	5.976	5.877	6.077	474.310	500.000	-5.1
Aroclor-1016-5	6.268	6.169	6.369	471.100	500.000	-5.8
Aroclor-1260-1	7.386	7.287	7.487	527.420	500.000	5.5
Aroclor-1260-2	7.639	7.540	7.740	488.330	500.000	-2.3
Aroclor-1260-3	7.997	7.899	8.099	510.250	500.000	2.1
Aroclor-1260-4	8.225	8.127	8.327	539.530	500.000	7.9
Aroclor-1260-5	8.549	8.452	8.652	497.530	500.000	-0.5
Decachlorobiphenyl	10.412	10.319	10.519	50.390	50.000	0.8
Tetrachloro-m-xylene	4.643	4.542	4.742	46.220	50.000	-7.6



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** MATR02
Lab Code: ACE **SDG NO.:** Q3450
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/21/2025 10/21/2025

Client Sample No.: CCAL02 **Date Analyzed:** 10/25/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP076057.D **Time Analyzed:** 00:22

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.878	4.778	4.978	475.300	500.000	-4.9
Aroclor-1016-2	4.937	4.836	5.036	506.010	500.000	1.2
Aroclor-1016-3	5.056	4.954	5.154	486.710	500.000	-2.7
Aroclor-1016-4	5.097	4.996	5.196	501.120	500.000	0.2
Aroclor-1016-5	5.311	5.210	5.410	470.680	500.000	-5.9
Aroclor-1260-1	6.337	6.238	6.438	513.790	500.000	2.8
Aroclor-1260-2	6.524	6.425	6.625	515.860	500.000	3.2
Aroclor-1260-3	6.678	6.578	6.778	518.110	500.000	3.6
Aroclor-1260-4	7.146	7.047	7.247	568.000	500.000	13.6
Aroclor-1260-5	7.386	7.287	7.487	533.710	500.000	6.7
Decachlorobiphenyl	8.779	8.682	8.882	46.800	50.000	-6.4
Tetrachloro-m-xylene	3.782	3.681	3.881	46.920	50.000	-6.2

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102425\
 Data File : PP076057.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 25 Oct 2025 00:22
 Operator : YP\AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 25 05:33:15 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Oct 24 15:40:32 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.643	3.782	66049864	310.6E6	46.218	46.925
2) SA Decachlor...	10.412	8.779	54250357	384.8E6	50.394	46.799
Target Compounds						
3) L1 AR-1016-1	5.795	4.878	27876164	351.8E6	471.470	475.300
4) L1 AR-1016-2	5.816	4.937	41418265	173.0E6	468.299	506.011
5) L1 AR-1016-3	5.879	5.056	25955144	99942279	468.919	486.710
6) L1 AR-1016-4	5.976	5.097	21552666	93652976	474.308	501.125
7) L1 AR-1016-5	6.268	5.311	20448172	107.3E6	471.096	470.680
31) L7 AR-1260-1	7.386	6.337	36263643	281.4E6	527.420	513.786
32) L7 AR-1260-2	7.639	6.524	43743655	419.8E6	488.334	515.864
33) L7 AR-1260-3	7.997	6.678	35353730	291.6E6	510.254	518.109
34) L7 AR-1260-4	8.225	7.146	36996195	295.0E6	539.528	568.004
35) L7 AR-1260-5	8.549	7.386	73609646	723.5E6	497.525	533.711



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102425\
Data File : PP076057.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 25 Oct 2025 00:22
Operator : YP\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 25 05:33:15 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 24 15:40:32 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: MATR02

Lab Code: ACE

SDG NO.: Q3450

Continuing Calib Date: 10/27/2025

Initial Calibration Date(s): 10/21/2025 10/21/2025

Continuing Calib Time: 08:56

Initial Calibration Time(s): 09:06 17:01

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	5.79	5.80	5.70	5.90	0.01
Aroclor-1016-2	(2)	5.81	5.82	5.72	5.92	0.01
Aroclor-1016-3	(3)	5.87	5.88	5.78	5.98	0.01
Aroclor-1016-4	(4)	5.97	5.98	5.88	6.08	0.01
Aroclor-1016-5	(5)	6.26	6.27	6.17	6.37	0.01
Aroclor-1260-1	(1)	7.38	7.39	7.29	7.49	0.01
Aroclor-1260-2	(2)	7.63	7.64	7.54	7.74	0.01
Aroclor-1260-3	(3)	7.99	8.00	7.90	8.10	0.01
Aroclor-1260-4	(4)	8.22	8.23	8.13	8.33	0.01
Aroclor-1260-5	(5)	8.54	8.55	8.45	8.65	0.01
Tetrachloro-m-xylene		4.64	4.64	4.54	4.74	0.00
Decachlorobiphenyl		10.41	10.42	10.32	10.52	0.02



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: MATR02

Lab Code: ACE

SDG NO.: Q3450

Continuing Calib Date: 10/27/2025

Initial Calibration Date(s): 10/21/2025 10/21/2025

Continuing Calib Time: 08:56

Initial Calibration Time(s): 09:06 17:01

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	4.87	4.88	4.78	4.98	0.01
Aroclor-1016-2	(2)	4.93	4.94	4.84	5.04	0.01
Aroclor-1016-3	(3)	5.05	5.05	4.95	5.15	0.00
Aroclor-1016-4	(4)	5.09	5.10	5.00	5.20	0.01
Aroclor-1016-5	(5)	5.31	5.31	5.21	5.41	0.00
Aroclor-1260-1	(1)	6.33	6.34	6.24	6.44	0.01
Aroclor-1260-2	(2)	6.52	6.53	6.43	6.63	0.01
Aroclor-1260-3	(3)	6.67	6.68	6.58	6.78	0.01
Aroclor-1260-4	(4)	7.14	7.15	7.05	7.25	0.01
Aroclor-1260-5	(5)	7.38	7.39	7.29	7.49	0.01
Tetrachloro-m-xylene		3.78	3.78	3.68	3.88	0.00
Decachlorobiphenyl		8.78	8.78	8.68	8.88	0.01



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** MATR02
Lab Code: ACE **SDG NO.:** Q3450
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/21/2025 10/21/2025

Client Sample No.: CCAL03 **Date Analyzed:** 10/27/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP076063.D **Time Analyzed:** 08:56

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.789	5.695	5.895	527.890	500.000	5.6
Aroclor-1016-2	5.811	5.717	5.917	519.780	500.000	4.0
Aroclor-1016-3	5.873	5.780	5.980	525.330	500.000	5.1
Aroclor-1016-4	5.971	5.877	6.077	527.600	500.000	5.5
Aroclor-1016-5	6.263	6.169	6.369	513.510	500.000	2.7
Aroclor-1260-1	7.380	7.287	7.487	555.310	500.000	11.1
Aroclor-1260-2	7.633	7.540	7.740	520.510	500.000	4.1
Aroclor-1260-3	7.991	7.899	8.099	543.790	500.000	8.8
Aroclor-1260-4	8.219	8.127	8.327	568.050	500.000	13.6
Aroclor-1260-5	8.544	8.452	8.652	527.260	500.000	5.5
Decachlorobiphenyl	10.405	10.319	10.519	52.540	50.000	5.1
Tetrachloro-m-xylene	4.637	4.542	4.742	53.280	50.000	6.6



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** MATR02
Lab Code: ACE **SDG NO.:** Q3450
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/21/2025 10/21/2025

Client Sample No.: CCAL03 **Date Analyzed:** 10/27/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP076063.D **Time Analyzed:** 08:56

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.873	4.778	4.978	528.360	500.000	5.7
Aroclor-1016-2	4.930	4.836	5.036	524.210	500.000	4.8
Aroclor-1016-3	5.051	4.954	5.154	458.440	500.000	-8.3
Aroclor-1016-4	5.092	4.996	5.196	553.310	500.000	10.7
Aroclor-1016-5	5.306	5.210	5.410	471.040	500.000	-5.8
Aroclor-1260-1	6.333	6.238	6.438	534.930	500.000	7.0
Aroclor-1260-2	6.520	6.425	6.625	523.920	500.000	4.8
Aroclor-1260-3	6.673	6.578	6.778	544.970	500.000	9.0
Aroclor-1260-4	7.142	7.047	7.247	557.220	500.000	11.4
Aroclor-1260-5	7.381	7.287	7.487	542.600	500.000	8.5
Decachlorobiphenyl	8.775	8.682	8.882	47.580	50.000	-4.8
Tetrachloro-m-xylene	3.779	3.681	3.881	52.390	50.000	4.8

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102725\
 Data File : PP076063.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 Oct 2025 08:56
 Operator : YP\AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660CCC500

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 10/28/2025
 Supervised By :mohammad ahmed 10/29/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 27 11:00:04 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Oct 24 15:40:32 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.637	3.779	76140554	346.7E6	53.279	52.387
2) SA Decachlor...	10.405	8.775	56558628	391.2E6	52.538	47.576
Target Compounds						
3) L1 AR-1016-1	5.789	4.873	31211797	391.1E6	527.885	528.362m
4) L1 AR-1016-2	5.811	4.930	45971798	179.2E6	519.784	524.212m
5) L1 AR-1016-3	5.873	5.051	29077501	94136772	525.329	458.438
6) L1 AR-1016-4	5.971	5.092	23974413	103.4E6	527.604	553.305
7) L1 AR-1016-5	6.263	5.306	22289187	107.4E6	513.510	471.035
31) L7 AR-1260-1	7.380	6.333	38181326	293.0E6	555.311	534.927
32) L7 AR-1260-2	7.633	6.520	46626204	426.3E6	520.514	523.919
33) L7 AR-1260-3	7.991	6.673	37677160	306.7E6	543.787	544.973
34) L7 AR-1260-4	8.219	7.142	38952287	289.4E6	568.055	557.215
35) L7 AR-1260-5	8.544	7.381	78009133	735.5E6	527.261	542.599

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102725\
Data File : PP076063.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Oct 2025 08:56
Operator : YP\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
AR1660CCC500

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/28/2025
Supervised By :mohammad ahmed 10/29/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 27 11:00:04 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 24 15:40:32 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: MATR02

Lab Code: ACE

SDG NO.: Q3450

Continuing Calib Date: 10/27/2025

Initial Calibration Date(s): 10/21/2025

10/21/2025

Continuing Calib Time: 14:08

Initial Calibration Time(s): 09:06

17:01

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	5.79	5.80	5.70	5.90	0.01
Aroclor-1016-2	(2)	5.82	5.82	5.72	5.92	0.00
Aroclor-1016-3	(3)	5.88	5.88	5.78	5.98	0.00
Aroclor-1016-4	(4)	5.98	5.98	5.88	6.08	0.01
Aroclor-1016-5	(5)	6.27	6.27	6.17	6.37	0.00
Aroclor-1260-1	(1)	7.38	7.39	7.29	7.49	0.01
Aroclor-1260-2	(2)	7.64	7.64	7.54	7.74	0.00
Aroclor-1260-3	(3)	8.00	8.00	7.90	8.10	0.01
Aroclor-1260-4	(4)	8.22	8.23	8.13	8.33	0.01
Aroclor-1260-5	(5)	8.55	8.55	8.45	8.65	0.00
Tetrachloro-m-xylene		4.64	4.64	4.54	4.74	0.00
Decachlorobiphenyl		10.41	10.42	10.32	10.52	0.01



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: MATR02

Lab Code: ACE

SDG NO.: Q3450

Continuing Calib Date: 10/27/2025

Initial Calibration Date(s): 10/21/2025

10/21/2025

Continuing Calib Time: 14:08

Initial Calibration Time(s): 09:06

17:01

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	4.88	4.88	4.78	4.98	0.00
Aroclor-1016-2	(2)	4.94	4.94	4.84	5.04	0.01
Aroclor-1016-3	(3)	5.05	5.05	4.95	5.15	0.00
Aroclor-1016-4	(4)	5.10	5.10	5.00	5.20	0.01
Aroclor-1016-5	(5)	5.31	5.31	5.21	5.41	0.00
Aroclor-1260-1	(1)	6.34	6.34	6.24	6.44	0.00
Aroclor-1260-2	(2)	6.52	6.53	6.43	6.63	0.01
Aroclor-1260-3	(3)	6.68	6.68	6.58	6.78	0.00
Aroclor-1260-4	(4)	7.14	7.15	7.05	7.25	0.01
Aroclor-1260-5	(5)	7.38	7.39	7.29	7.49	0.01
Tetrachloro-m-xylene		3.78	3.78	3.68	3.88	0.00
Decachlorobiphenyl		8.78	8.78	8.68	8.88	0.00



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** MATR02
Lab Code: ACE **SDG NO.:** Q3450
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/21/2025 10/21/2025

Client Sample No.: CCAL04 **Date Analyzed:** 10/27/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP076078.D **Time Analyzed:** 14:08

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.793	5.695	5.895	489.480	500.000	-2.1
Aroclor-1016-2	5.815	5.717	5.917	493.840	500.000	-1.2
Aroclor-1016-3	5.877	5.780	5.980	492.480	500.000	-1.5
Aroclor-1016-4	5.975	5.877	6.077	498.560	500.000	-0.3
Aroclor-1016-5	6.267	6.169	6.369	482.900	500.000	-3.4
Aroclor-1260-1	7.384	7.287	7.487	535.750	500.000	7.2
Aroclor-1260-2	7.636	7.540	7.740	493.780	500.000	-1.2
Aroclor-1260-3	7.995	7.899	8.099	512.820	500.000	2.6
Aroclor-1260-4	8.222	8.127	8.327	550.900	500.000	10.2
Aroclor-1260-5	8.547	8.452	8.652	502.490	500.000	0.5
Decachlorobiphenyl	10.409	10.319	10.519	50.530	50.000	1.1
Tetrachloro-m-xylene	4.642	4.542	4.742	49.760	50.000	-0.5



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** MATR02
Lab Code: ACE **SDG NO.:** Q3450
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/21/2025 10/21/2025

Client Sample No.: CCAL04 **Date Analyzed:** 10/27/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP076078.D **Time Analyzed:** 14:08

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.876	4.778	4.978	518.620	500.000	3.7
Aroclor-1016-2	4.935	4.836	5.036	531.450	500.000	6.3
Aroclor-1016-3	5.054	4.954	5.154	501.290	500.000	0.3
Aroclor-1016-4	5.095	4.996	5.196	541.930	500.000	8.4
Aroclor-1016-5	5.308	5.210	5.410	547.650	500.000	9.5
Aroclor-1260-1	6.336	6.238	6.438	486.190	500.000	-2.8
Aroclor-1260-2	6.522	6.425	6.625	481.180	500.000	-3.8
Aroclor-1260-3	6.675	6.578	6.778	504.200	500.000	0.8
Aroclor-1260-4	7.144	7.047	7.247	522.340	500.000	4.5
Aroclor-1260-5	7.384	7.287	7.487	484.740	500.000	-3.1
Decachlorobiphenyl	8.777	8.682	8.882	45.370	50.000	-9.3
Tetrachloro-m-xylene	3.782	3.681	3.881	53.460	50.000	6.9

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102725\
 Data File : PP076078.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 Oct 2025 14:08
 Operator : YP\AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 28 01:02:38 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Oct 24 15:40:32 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.642	3.782	71108591	353.8E6	49.758	53.460
2) SA Decachlor...	10.409	8.777	54396935	373.1E6	50.530	45.368
Target Compounds						
3) L1 AR-1016-1	5.793	4.876	28941333	383.9E6	489.485	518.615
4) L1 AR-1016-2	5.815	4.935	43677020	181.7E6	493.837	531.450
5) L1 AR-1016-3	5.877	5.054	27259037	102.9E6	492.476	501.291
6) L1 AR-1016-4	5.975	5.095	22654846	101.3E6	498.564	541.928
7) L1 AR-1016-5	6.267	5.308	20960588	124.9E6	482.901	547.651
31) L7 AR-1260-1	7.384	6.336	36836407	266.3E6	535.751	486.193
32) L7 AR-1260-2	7.636	6.522	44231121	391.6E6	493.776	481.177
33) L7 AR-1260-3	7.995	6.675	35531662	283.7E6	512.822	504.203
34) L7 AR-1260-4	8.222	7.144	37775708	271.3E6	550.896	522.340
35) L7 AR-1260-5	8.547	7.384	74344445	657.1E6	502.492	484.740



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102725\
Data File : PP076078.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Oct 2025 14:08
Operator : YP\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 28 01:02:38 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 24 15:40:32 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: MATR02

Lab Code: ACE

SDG NO.: Q3450

Continuing Calib Date: 10/27/2025

Initial Calibration Date(s): 10/10/2025 10/10/2025

Continuing Calib Time: 08:57

Initial Calibration Time(s): 08:57 16:43

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	4.48	4.48	4.38	4.58	0.00
Aroclor-1016-2	(2)	4.50	4.50	4.40	4.60	0.00
Aroclor-1016-3	(3)	4.56	4.56	4.46	4.66	0.00
Aroclor-1016-4	(4)	4.65	4.65	4.55	4.75	0.00
Aroclor-1016-5	(5)	4.93	4.93	4.83	5.03	0.00
Aroclor-1260-1	(1)	6.02	6.02	5.92	6.12	0.00
Aroclor-1260-2	(2)	6.28	6.28	6.18	6.38	0.00
Aroclor-1260-3	(3)	6.63	6.63	6.53	6.73	0.00
Aroclor-1260-4	(4)	6.84	6.84	6.74	6.94	0.00
Aroclor-1260-5	(5)	7.15	7.15	7.05	7.25	0.00
Tetrachloro-m-xylene		3.40	3.40	3.30	3.50	0.00
Decachlorobiphenyl		8.50	8.50	8.40	8.60	0.00



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: MATR02

Lab Code: ACE

SDG NO.: Q3450

Continuing Calib Date: 10/27/2025

Initial Calibration Date(s): 10/10/2025 10/10/2025

Continuing Calib Time: 08:57

Initial Calibration Time(s): 08:57 16:43

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	3.78	3.78	3.68	3.88	0.00
Aroclor-1016-2	(2)	3.79	3.79	3.69	3.89	0.00
Aroclor-1016-3	(3)	3.95	3.95	3.85	4.05	0.00
Aroclor-1016-4	(4)	4.00	4.00	3.90	4.10	0.00
Aroclor-1016-5	(5)	4.19	4.20	4.10	4.30	0.01
Aroclor-1260-1	(1)	5.19	5.19	5.09	5.29	0.00
Aroclor-1260-2	(2)	5.38	5.38	5.28	5.48	0.00
Aroclor-1260-3	(3)	5.52	5.52	5.42	5.62	0.00
Aroclor-1260-4	(4)	5.98	5.98	5.88	6.08	0.00
Aroclor-1260-5	(5)	6.22	6.22	6.12	6.32	0.00
Tetrachloro-m-xylene		2.78	2.78	2.68	2.88	0.00
Decachlorobiphenyl		7.53	7.53	7.43	7.63	0.00



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** MATR02
Lab Code: ACE **SDG NO.:** Q3450
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/10/2025 10/10/2025

Client Sample No.: CCAL05 **Date Analyzed:** 10/27/2025
Lab Sample No.: AR1660CCC500 **Data File :** PQ071149.D **Time Analyzed:** 08:57

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.480	4.380	4.580	582.660	500.000	16.5
Aroclor-1016-2	4.499	4.399	4.599	562.340	500.000	12.5
Aroclor-1016-3	4.556	4.456	4.656	584.520	500.000	16.9
Aroclor-1016-4	4.646	4.546	4.746	585.390	500.000	17.1
Aroclor-1016-5	4.931	4.831	5.031	576.280	500.000	15.3
Aroclor-1260-1	6.020	5.919	6.119	558.050	500.000	11.6
Aroclor-1260-2	6.276	6.175	6.375	557.500	500.000	11.5
Aroclor-1260-3	6.626	6.526	6.726	583.750	500.000	16.8
Aroclor-1260-4	6.841	6.740	6.940	592.250	500.000	18.5
Aroclor-1260-5	7.146	7.045	7.245	579.210	500.000	15.8
Decachlorobiphenyl	8.497	8.396	8.596	59.340	50.000	18.7
Tetrachloro-m-xylene	3.400	3.300	3.500	57.140	50.000	14.3



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** MATR02
Lab Code: ACE **SDG NO.:** Q3450
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/10/2025 10/10/2025

Client Sample No.: CCAL05 **Date Analyzed:** 10/27/2025
Lab Sample No.: AR1660CCC500 **Data File :** PQ071149.D **Time Analyzed:** 08:57

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	3.776	3.678	3.878	461.700	500.000	-7.7
Aroclor-1016-2	3.791	3.693	3.893	465.260	500.000	-6.9
Aroclor-1016-3	3.951	3.853	4.053	480.210	500.000	-4.0
Aroclor-1016-4	4.000	3.902	4.102	488.300	500.000	-2.3
Aroclor-1016-5	4.194	4.097	4.297	480.390	500.000	-3.9
Aroclor-1260-1	5.187	5.089	5.289	477.270	500.000	-4.5
Aroclor-1260-2	5.376	5.278	5.478	476.660	500.000	-4.7
Aroclor-1260-3	5.517	5.419	5.619	480.230	500.000	-4.0
Aroclor-1260-4	5.977	5.879	6.079	481.030	500.000	-3.8
Aroclor-1260-5	6.222	6.124	6.324	480.940	500.000	-3.8
Decachlorobiphenyl	7.531	7.433	7.633	56.210	50.000	12.4
Tetrachloro-m-xylene	2.776	2.678	2.878	49.690	50.000	-0.6

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ102725\
 Data File : PQ071149.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 Oct 2025 08:57
 Operator : YP\AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 AR1660CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 27 11:27:45 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Oct 11 01:55:41 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.400	2.776	248.6E6	252.4E6	57.145	49.691
2) SA Decachlor...	8.497	7.531	172.3E6	283.8E6	59.344	56.205
Target Compounds						
3) L1 AR-1016-1	4.480	3.776	95639952	103.4E6	582.659	461.704
4) L1 AR-1016-2	4.499	3.791	144.1E6	158.2E6	562.335	465.258
5) L1 AR-1016-3	4.556	3.951	93008322	87331257	584.523	480.214
6) L1 AR-1016-4	4.646	4.000	76059348	76865356	585.390	488.305
7) L1 AR-1016-5	4.931	4.194	74226869	92158770	576.283	480.389
31) L7 AR-1260-1	6.020	5.187	130.8E6	172.0E6	558.047	477.273
32) L7 AR-1260-2	6.276	5.376	154.2E6	209.7E6	557.496	476.659
33) L7 AR-1260-3	6.626	5.517	116.3E6	199.3E6	583.746	480.230
34) L7 AR-1260-4	6.841	5.977	127.3E6	173.1E6	592.247	481.031
35) L7 AR-1260-5	7.146	6.222	251.2E6	404.1E6	579.210	480.935



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ102725\
Data File : PQ071149.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Oct 2025 08:57
Operator : YP\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1660CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 27 11:27:45 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Oct 11 01:55:41 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: MATR02

Lab Code: ACE

SDG NO.: Q3450

Continuing Calib Date: 10/27/2025

Initial Calibration Date(s): 10/10/2025 10/10/2025

Continuing Calib Time: 12:35

Initial Calibration Time(s): 08:57 16:43

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	4.48	4.48	4.38	4.58	0.00
Aroclor-1016-2	(2)	4.50	4.50	4.40	4.60	0.00
Aroclor-1016-3	(3)	4.56	4.56	4.46	4.66	0.00
Aroclor-1016-4	(4)	4.65	4.65	4.55	4.75	0.01
Aroclor-1016-5	(5)	4.93	4.93	4.83	5.03	0.00
Aroclor-1260-1	(1)	6.02	6.02	5.92	6.12	0.00
Aroclor-1260-2	(2)	6.27	6.28	6.18	6.38	0.01
Aroclor-1260-3	(3)	6.63	6.63	6.53	6.73	0.00
Aroclor-1260-4	(4)	6.84	6.84	6.74	6.94	0.00
Aroclor-1260-5	(5)	7.15	7.15	7.05	7.25	0.01
Tetrachloro-m-xylene		3.40	3.40	3.30	3.50	0.00
Decachlorobiphenyl		8.50	8.50	8.40	8.60	0.01



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: MATR02

Lab Code: ACE

SDG NO.: Q3450

Continuing Calib Date: 10/27/2025

Initial Calibration Date(s): 10/10/2025 10/10/2025

Continuing Calib Time: 12:35

Initial Calibration Time(s): 08:57 16:43

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	3.78	3.78	3.68	3.88	0.00
Aroclor-1016-2	(2)	3.79	3.79	3.69	3.89	0.00
Aroclor-1016-3	(3)	3.95	3.95	3.85	4.05	0.00
Aroclor-1016-4	(4)	4.00	4.00	3.90	4.10	0.00
Aroclor-1016-5	(5)	4.20	4.20	4.10	4.30	0.00
Aroclor-1260-1	(1)	5.19	5.19	5.09	5.29	0.00
Aroclor-1260-2	(2)	5.38	5.38	5.28	5.48	0.00
Aroclor-1260-3	(3)	5.52	5.52	5.42	5.62	0.00
Aroclor-1260-4	(4)	5.98	5.98	5.88	6.08	0.00
Aroclor-1260-5	(5)	6.22	6.22	6.12	6.32	0.00
Tetrachloro-m-xylene		2.78	2.78	2.68	2.88	0.00
Decachlorobiphenyl		7.53	7.53	7.43	7.63	0.00



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** MATR02
Lab Code: ACE **SDG NO.:** Q3450
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/10/2025 10/10/2025

Client Sample No.: CCAL06 **Date Analyzed:** 10/27/2025
Lab Sample No.: AR1660CCC500 **Data File :** PQ071161.D **Time Analyzed:** 12:35

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.479	4.380	4.580	550.580	500.000	10.1
Aroclor-1016-2	4.498	4.399	4.599	559.440	500.000	11.9
Aroclor-1016-3	4.556	4.456	4.656	571.120	500.000	14.2
Aroclor-1016-4	4.645	4.546	4.746	570.660	500.000	14.1
Aroclor-1016-5	4.930	4.831	5.031	565.870	500.000	13.2
Aroclor-1260-1	6.019	5.919	6.119	557.440	500.000	11.5
Aroclor-1260-2	6.274	6.175	6.375	551.940	500.000	10.4
Aroclor-1260-3	6.626	6.526	6.726	579.440	500.000	15.9
Aroclor-1260-4	6.839	6.740	6.940	591.430	500.000	18.3
Aroclor-1260-5	7.145	7.045	7.245	581.000	500.000	16.2
Decachlorobiphenyl	8.495	8.396	8.596	59.520	50.000	19.0
Tetrachloro-m-xylene	3.399	3.300	3.500	56.350	50.000	12.7



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** MATR02
Lab Code: ACE **SDG NO.:** Q3450
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/10/2025 10/10/2025

Client Sample No.: CCAL06 **Date Analyzed:** 10/27/2025
Lab Sample No.: AR1660CCC500 **Data File :** PQ071161.D **Time Analyzed:** 12:35

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	3.776	3.678	3.878	459.340	500.000	-8.1
Aroclor-1016-2	3.791	3.693	3.893	461.520	500.000	-7.7
Aroclor-1016-3	3.951	3.853	4.053	473.930	500.000	-5.2
Aroclor-1016-4	4.000	3.902	4.102	480.760	500.000	-3.8
Aroclor-1016-5	4.195	4.097	4.297	472.690	500.000	-5.5
Aroclor-1260-1	5.186	5.089	5.289	469.950	500.000	-6.0
Aroclor-1260-2	5.376	5.278	5.478	469.430	500.000	-6.1
Aroclor-1260-3	5.517	5.419	5.619	470.970	500.000	-5.8
Aroclor-1260-4	5.977	5.879	6.079	477.010	500.000	-4.6
Aroclor-1260-5	6.221	6.124	6.324	476.200	500.000	-4.8
Decachlorobiphenyl	7.531	7.433	7.633	55.990	50.000	12.0
Tetrachloro-m-xylene	2.776	2.678	2.878	50.180	50.000	0.4

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ102725\
 Data File : PQ071161.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 Oct 2025 12:35
 Operator : YP\AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 AR1660CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 27 12:58:57 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Oct 11 01:55:41 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.399	2.776	245.1E6	254.9E6	56.346	50.181
2) SA Decachlor...	8.495	7.531	172.8E6	282.7E6	59.521	55.992
Target Compounds						
3) L1 AR-1016-1	4.479	3.776	90374544	102.9E6	550.581	459.336
4) L1 AR-1016-2	4.498	3.791	143.4E6	157.0E6	559.440	461.520
5) L1 AR-1016-3	4.556	3.951	90876424	86188973	571.125	473.933
6) L1 AR-1016-4	4.645	4.000	74144880	75678013	570.656	480.762
7) L1 AR-1016-5	4.930	4.195	72885811	90681279	565.871	472.688
31) L7 AR-1260-1	6.019	5.186	130.7E6	169.4E6	557.445	469.948
32) L7 AR-1260-2	6.274	5.376	152.7E6	206.5E6	551.943	469.434
33) L7 AR-1260-3	6.626	5.517	115.5E6	195.5E6	579.442	470.969
34) L7 AR-1260-4	6.839	5.977	127.1E6	171.6E6	591.425	477.012
35) L7 AR-1260-5	7.145	6.221	252.0E6	400.1E6	580.996	476.203



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ102725\
Data File : PQ071161.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Oct 2025 12:35
Operator : YP\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1660CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 27 12:58:57 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Oct 11 01:55:41 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Analytical Sequence

Client: Matrix New World Engineering	SDG No.: Q3450
Project: PSEG Athenia Substation	Instrument ID: ECD_P
GC Column: ZB-MR1	ID: 0.32 (mm) Inst. Calib. Date(s): 10/21/2025 10/21/2025

THE ANALYTICAL SEQUENCE OF PERFORMANCE EVALUATION MIXTURES, BLANKS, SAMPLES,
AND STANDARDS IS GIVEN BELOW:

CLIENT ID	LAB SAMPLE ID	DATE ANALYZED	TIME ANALYZED	DATAFILE	DCB RT #	TCX RT #
IBLK	IBLK	10/21/2025	08:49	PP075887.D	10.42	4.64
AR1660ICC1000	AR1660ICC1000	10/21/2025	09:06	PP075888.D	10.42	4.65
AR1660ICC750	AR1660ICC750	10/21/2025	09:22	PP075889.D	10.42	4.64
AR1660ICC500	AR1660ICC500	10/21/2025	09:38	PP075890.D	10.42	4.64
AR1660ICC250	AR1660ICC250	10/21/2025	09:54	PP075891.D	10.42	4.65
AR1660ICC050	AR1660ICC050	10/21/2025	10:11	PP075892.D	10.42	4.65
AR1221ICC500	AR1221ICC500	10/21/2025	10:27	PP075893.D	10.42	4.64
AR1232ICC500	AR1232ICC500	10/21/2025	10:43	PP075894.D	10.42	4.64
AR1242ICC1000	AR1242ICC1000	10/21/2025	10:59	PP075895.D	10.42	4.65
AR1242ICC750	AR1242ICC750	10/21/2025	11:16	PP075896.D	10.42	4.64
AR1242ICC500	AR1242ICC500	10/21/2025	11:48	PP075897.D	10.42	4.65
AR1242ICC250	AR1242ICC250	10/21/2025	12:04	PP075898.D	10.42	4.65
AR1242ICC050	AR1242ICC050	10/21/2025	12:21	PP075899.D	10.42	4.65
AR1248ICC1000	AR1248ICC1000	10/21/2025	12:41	PP075900.D	10.43	4.65
AR1248ICC750	AR1248ICC750	10/21/2025	12:57	PP075901.D	10.42	4.64
AR1248ICC500	AR1248ICC500	10/21/2025	13:13	PP075902.D	10.42	4.65
AR1248ICC250	AR1248ICC250	10/21/2025	13:30	PP075903.D	10.42	4.65
AR1248ICC050	AR1248ICC050	10/21/2025	13:46	PP075904.D	10.42	4.65
AR1254ICC1000	AR1254ICC1000	10/21/2025	14:02	PP075905.D	10.42	4.65
AR1254ICC750	AR1254ICC750	10/21/2025	14:18	PP075906.D	10.42	4.64
AR1254ICC500	AR1254ICC500	10/21/2025	14:35	PP075907.D	10.42	4.65
AR1254ICC250	AR1254ICC250	10/21/2025	14:51	PP075908.D	10.42	4.65
AR1254ICC050	AR1254ICC050	10/21/2025	15:07	PP075909.D	10.42	4.64
AR1262ICC500	AR1262ICC500	10/21/2025	15:40	PP075910.D	10.43	4.65
AR1268ICC1000	AR1268ICC1000	10/21/2025	15:56	PP075911.D	10.42	4.65
AR1268ICC750	AR1268ICC750	10/21/2025	16:12	PP075912.D	10.42	4.65
AR1268ICC500	AR1268ICC500	10/21/2025	16:29	PP075913.D	10.42	4.65
AR1268ICC250	AR1268ICC250	10/21/2025	16:45	PP075914.D	10.43	4.65
AR1268ICC050	AR1268ICC050	10/21/2025	17:01	PP075915.D	10.42	4.65
AR1660CCC500	AR1660CCC500	10/24/2025	18:58	PP076042.D	10.41	4.64
IBLK	IBLK	10/24/2025	20:36	PP076046.D	10.41	4.64
SS-20-(3.5-4.0)	Q3450-10	10/24/2025	20:52	PP076047.D	10.41	4.64
SS-28-(2.0-2.5)	Q3450-11	10/24/2025	21:08	PP076048.D	10.41	4.64
SS-21-(0.0-0.5)	Q3450-20	10/24/2025	21:24	PP076049.D	10.41	4.64
SS-21-(2.0-2.5)	Q3450-21	10/24/2025	21:41	PP076050.D	10.41	4.64
SS-21-(3.5-4.0)	Q3450-22	10/24/2025	21:57	PP076051.D	10.41	4.64
AR1660CCC500	AR1660CCC500	10/25/2025	00:22	PP076057.D	10.41	4.64
IBLK	IBLK	10/25/2025	02:16	PP076061.D	10.41	4.64
AR1660CCC500	AR1660CCC500	10/27/2025	08:56	PP076063.D	10.41	4.64
IBLK	IBLK	10/27/2025	10:02	PP076067.D	10.40	4.64
PB170236BL	PB170236BL	10/27/2025	10:18	PP076068.D	10.40	4.64
PB170236BS	PB170236BS	10/27/2025	10:34	PP076069.D	10.41	4.64

Analytical Sequence

SS-21-(0.0-0.5)DL	Q3450-20DL	10/27/2025	11:42	PP076072.D	10.41	4.64
SS-21-(3.5-4.0)DL	Q3450-22DL	10/27/2025	11:58	PP076073.D	10.40	4.64
AR1660CCC500	AR1660CCC500	10/27/2025	14:08	PP076078.D	10.41	4.64
I.BLK	I.BLK	10/27/2025	15:13	PP076082.D	10.41	4.64
I.BLK	I.BLK	10/10/2025	08:42	PQ070972.D	8.50	3.40
AR1660ICC1000	AR1660ICC1000	10/10/2025	08:57	PQ070973.D	8.50	3.40
AR1660ICC750	AR1660ICC750	10/10/2025	09:11	PQ070974.D	8.50	3.40
AR1660ICC500	AR1660ICC500	10/10/2025	09:26	PQ070975.D	8.50	3.40
AR1660ICC250	AR1660ICC250	10/10/2025	09:40	PQ070976.D	8.50	3.40
AR1660ICC050	AR1660ICC050	10/10/2025	09:55	PQ070977.D	8.50	3.40
AR1221ICC500	AR1221ICC500	10/10/2025	10:09	PQ070978.D	8.50	3.40
AR1232ICC500	AR1232ICC500	10/10/2025	10:24	PQ070979.D	8.50	3.40
AR1242ICC1000	AR1242ICC1000	10/10/2025	10:38	PQ070980.D	8.50	3.40
AR1242ICC750	AR1242ICC750	10/10/2025	10:53	PQ070981.D	8.50	3.40
AR1242ICC500	AR1242ICC500	10/10/2025	11:08	PQ070982.D	8.50	3.40
AR1242ICC250	AR1242ICC250	10/10/2025	11:22	PQ070983.D	8.50	3.40
AR1242ICC050	AR1242ICC050	10/10/2025	11:37	PQ070984.D	8.50	3.40
AR1248ICC1000	AR1248ICC1000	10/10/2025	11:51	PQ070985.D	8.50	3.40
AR1248ICC750	AR1248ICC750	10/10/2025	12:06	PQ070986.D	8.50	3.40
AR1248ICC500	AR1248ICC500	10/10/2025	12:20	PQ070987.D	8.50	3.40
AR1248ICC250	AR1248ICC250	10/10/2025	12:35	PQ070988.D	8.50	3.40
AR1248ICC050	AR1248ICC050	10/10/2025	12:50	PQ070989.D	8.50	3.40
AR1254ICC1000	AR1254ICC1000	10/10/2025	13:04	PQ070990.D	8.50	3.40
AR1254ICC750	AR1254ICC750	10/10/2025	13:19	PQ070991.D	8.50	3.40
AR1254ICC500	AR1254ICC500	10/10/2025	13:33	PQ070992.D	8.50	3.40
AR1254ICC250	AR1254ICC250	10/10/2025	13:48	PQ070993.D	8.50	3.40
AR1254ICC050	AR1254ICC050	10/10/2025	15:15	PQ070994.D	8.50	3.40
AR1262ICC500	AR1262ICC500	10/10/2025	15:30	PQ070995.D	8.50	3.40
AR1268ICC1000	AR1268ICC1000	10/10/2025	15:44	PQ070996.D	8.50	3.40
AR1268ICC750	AR1268ICC750	10/10/2025	15:59	PQ070997.D	8.50	3.40
AR1268ICC500	AR1268ICC500	10/10/2025	16:13	PQ070998.D	8.50	3.40
AR1268ICC250	AR1268ICC250	10/10/2025	16:28	PQ070999.D	8.50	3.40
AR1268ICC050	AR1268ICC050	10/10/2025	16:43	PQ071000.D	8.50	3.40
AR1660CCC500	AR1660CCC500	10/27/2025	08:57	PQ071149.D	8.50	3.40
I.BLK	I.BLK	10/27/2025	09:11	PQ071150.D	8.50	3.40
SS-5R-(2.0-2.5)	Q3450-01	10/27/2025	09:26	PQ071151.D	8.50	3.40
SS-5R-(3.5-4.0)	Q3450-02	10/27/2025	09:41	PQ071152.D	8.50	3.40
SS-5R-(3.5-4.0)MS	Q3450-02MS	10/27/2025	09:55	PQ071153.D	8.50	3.40
SS-5R-(3.5-4.0)MSD	Q3450-02MSD	10/27/2025	10:10	PQ071154.D	8.50	3.40
SS-20-(0.0-0.5)	Q3450-08	10/27/2025	10:24	PQ071155.D	8.50	3.40
SS-20-(2.0-2.5)	Q3450-09	10/27/2025	10:39	PQ071156.D	8.50	3.40
AR1660CCC500	AR1660CCC500	10/27/2025	12:35	PQ071161.D	8.50	3.40
I.BLK	I.BLK	10/27/2025	14:18	PQ071165.D	8.50	3.40

Analytical Sequence

Client: Matrix New World Engineering	SDG No.: Q3450
Project: PSEG Athenia Substation	Instrument ID: ECD_P
GC Column: ZB-MR2	ID: 0.32 (mm) Inst. Calib. Date(s): 10/21/2025 10/21/2025

THE ANALYTICAL SEQUENCE OF PERFORMANCE EVALUATION MIXTURES, BLANKS, SAMPLES,
AND STANDARDS IS GIVEN BELOW:

CLIENT ID	LAB SAMPLE ID	DATE ANALYZED	TIME ANALYZED	DATAFILE	DCB RT #	TCX RT #
IBLK	IBLK	10/21/2025	08:49	PP075887.D	8.78	3.78
AR1660ICC1000	AR1660ICC1000	10/21/2025	09:06	PP075888.D	8.78	3.78
AR1660ICC750	AR1660ICC750	10/21/2025	09:22	PP075889.D	8.78	3.78
AR1660ICC500	AR1660ICC500	10/21/2025	09:38	PP075890.D	8.78	3.78
AR1660ICC250	AR1660ICC250	10/21/2025	09:54	PP075891.D	8.78	3.78
AR1660ICC050	AR1660ICC050	10/21/2025	10:11	PP075892.D	8.78	3.78
AR1221ICC500	AR1221ICC500	10/21/2025	10:27	PP075893.D	8.78	3.78
AR1232ICC500	AR1232ICC500	10/21/2025	10:43	PP075894.D	8.78	3.78
AR1242ICC1000	AR1242ICC1000	10/21/2025	10:59	PP075895.D	8.78	3.78
AR1242ICC750	AR1242ICC750	10/21/2025	11:16	PP075896.D	8.78	3.78
AR1242ICC500	AR1242ICC500	10/21/2025	11:48	PP075897.D	8.79	3.78
AR1242ICC250	AR1242ICC250	10/21/2025	12:04	PP075898.D	8.78	3.78
AR1242ICC050	AR1242ICC050	10/21/2025	12:21	PP075899.D	8.78	3.78
AR1248ICC1000	AR1248ICC1000	10/21/2025	12:41	PP075900.D	8.78	3.78
AR1248ICC750	AR1248ICC750	10/21/2025	12:57	PP075901.D	8.78	3.78
AR1248ICC500	AR1248ICC500	10/21/2025	13:13	PP075902.D	8.78	3.78
AR1248ICC250	AR1248ICC250	10/21/2025	13:30	PP075903.D	8.78	3.78
AR1248ICC050	AR1248ICC050	10/21/2025	13:46	PP075904.D	8.78	3.78
AR1254ICC1000	AR1254ICC1000	10/21/2025	14:02	PP075905.D	8.78	3.78
AR1254ICC750	AR1254ICC750	10/21/2025	14:18	PP075906.D	8.78	3.78
AR1254ICC500	AR1254ICC500	10/21/2025	14:35	PP075907.D	8.79	3.79
AR1254ICC250	AR1254ICC250	10/21/2025	14:51	PP075908.D	8.78	3.78
AR1254ICC050	AR1254ICC050	10/21/2025	15:07	PP075909.D	8.78	3.78
AR1262ICC500	AR1262ICC500	10/21/2025	15:40	PP075910.D	8.79	3.79
AR1268ICC1000	AR1268ICC1000	10/21/2025	15:56	PP075911.D	8.79	3.78
AR1268ICC750	AR1268ICC750	10/21/2025	16:12	PP075912.D	8.78	3.78
AR1268ICC500	AR1268ICC500	10/21/2025	16:29	PP075913.D	8.79	3.79
AR1268ICC250	AR1268ICC250	10/21/2025	16:45	PP075914.D	8.79	3.79
AR1268ICC050	AR1268ICC050	10/21/2025	17:01	PP075915.D	8.79	3.79
AR1660CCC500	AR1660CCC500	10/24/2025	18:58	PP076042.D	8.78	3.78
IBLK	IBLK	10/24/2025	20:36	PP076046.D	8.78	3.78
SS-20-(3.5-4.0)	Q3450-10	10/24/2025	20:52	PP076047.D	8.78	3.78
SS-28-(2.0-2.5)	Q3450-11	10/24/2025	21:08	PP076048.D	8.78	3.78
SS-21-(0.0-0.5)	Q3450-20	10/24/2025	21:24	PP076049.D	8.78	3.78
SS-21-(2.0-2.5)	Q3450-21	10/24/2025	21:41	PP076050.D	8.78	3.78
SS-21-(3.5-4.0)	Q3450-22	10/24/2025	21:57	PP076051.D	8.78	3.78
AR1660CCC500	AR1660CCC500	10/25/2025	00:22	PP076057.D	8.78	3.78
IBLK	IBLK	10/25/2025	02:16	PP076061.D	8.78	3.78
AR1660CCC500	AR1660CCC500	10/27/2025	08:56	PP076063.D	8.78	3.78
IBLK	IBLK	10/27/2025	10:02	PP076067.D	8.77	3.78
PB170236BL	PB170236BL	10/27/2025	10:18	PP076068.D	8.77	3.78
PB170236BS	PB170236BS	10/27/2025	10:34	PP076069.D	8.78	3.78

Analytical Sequence

SS-21-(0.0-0.5)DL	Q3450-20DL	10/27/2025	11:42	PP076072.D	8.78	3.78
SS-21-(3.5-4.0)DL	Q3450-22DL	10/27/2025	11:58	PP076073.D	8.77	3.78
AR1660CCC500	AR1660CCC500	10/27/2025	14:08	PP076078.D	8.78	3.78
I.BLK	I.BLK	10/27/2025	15:13	PP076082.D	8.77	3.78
I.BLK	I.BLK	10/10/2025	08:42	PQ070972.D	7.53	2.78
AR1660ICC1000	AR1660ICC1000	10/10/2025	08:57	PQ070973.D	7.53	2.78
AR1660ICC750	AR1660ICC750	10/10/2025	09:11	PQ070974.D	7.53	2.78
AR1660ICC500	AR1660ICC500	10/10/2025	09:26	PQ070975.D	7.53	2.78
AR1660ICC250	AR1660ICC250	10/10/2025	09:40	PQ070976.D	7.53	2.78
AR1660ICC050	AR1660ICC050	10/10/2025	09:55	PQ070977.D	7.53	2.78
AR1221ICC500	AR1221ICC500	10/10/2025	10:09	PQ070978.D	7.53	2.78
AR1232ICC500	AR1232ICC500	10/10/2025	10:24	PQ070979.D	7.53	2.78
AR1242ICC1000	AR1242ICC1000	10/10/2025	10:38	PQ070980.D	7.53	2.78
AR1242ICC750	AR1242ICC750	10/10/2025	10:53	PQ070981.D	7.53	2.78
AR1242ICC500	AR1242ICC500	10/10/2025	11:08	PQ070982.D	7.53	2.78
AR1242ICC250	AR1242ICC250	10/10/2025	11:22	PQ070983.D	7.53	2.78
AR1242ICC050	AR1242ICC050	10/10/2025	11:37	PQ070984.D	7.53	2.78
AR1248ICC1000	AR1248ICC1000	10/10/2025	11:51	PQ070985.D	7.53	2.78
AR1248ICC750	AR1248ICC750	10/10/2025	12:06	PQ070986.D	7.53	2.78
AR1248ICC500	AR1248ICC500	10/10/2025	12:20	PQ070987.D	7.53	2.78
AR1248ICC250	AR1248ICC250	10/10/2025	12:35	PQ070988.D	7.53	2.78
AR1248ICC050	AR1248ICC050	10/10/2025	12:50	PQ070989.D	7.53	2.78
AR1254ICC1000	AR1254ICC1000	10/10/2025	13:04	PQ070990.D	7.53	2.78
AR1254ICC750	AR1254ICC750	10/10/2025	13:19	PQ070991.D	7.53	2.78
AR1254ICC500	AR1254ICC500	10/10/2025	13:33	PQ070992.D	7.53	2.78
AR1254ICC250	AR1254ICC250	10/10/2025	13:48	PQ070993.D	7.53	2.78
AR1254ICC050	AR1254ICC050	10/10/2025	15:15	PQ070994.D	7.53	2.78
AR1262ICC500	AR1262ICC500	10/10/2025	15:30	PQ070995.D	7.53	2.78
AR1268ICC1000	AR1268ICC1000	10/10/2025	15:44	PQ070996.D	7.53	2.78
AR1268ICC750	AR1268ICC750	10/10/2025	15:59	PQ070997.D	7.53	2.78
AR1268ICC500	AR1268ICC500	10/10/2025	16:13	PQ070998.D	7.53	2.78
AR1268ICC250	AR1268ICC250	10/10/2025	16:28	PQ070999.D	7.53	2.78
AR1268ICC050	AR1268ICC050	10/10/2025	16:43	PQ071000.D	7.53	2.78
AR1660CCC500	AR1660CCC500	10/27/2025	08:57	PQ071149.D	7.53	2.78
I.BLK	I.BLK	10/27/2025	09:11	PQ071150.D	7.53	2.78
SS-5R-(2.0-2.5)	Q3450-01	10/27/2025	09:26	PQ071151.D	7.53	2.78
SS-5R-(3.5-4.0)	Q3450-02	10/27/2025	09:41	PQ071152.D	7.53	2.78
SS-5R-(3.5-4.0)MS	Q3450-02MS	10/27/2025	09:55	PQ071153.D	7.53	2.78
SS-5R-(3.5-4.0)MSD	Q3450-02MSD	10/27/2025	10:10	PQ071154.D	7.53	2.78
SS-20-(0.0-0.5)	Q3450-08	10/27/2025	10:24	PQ071155.D	7.53	2.78
SS-20-(2.0-2.5)	Q3450-09	10/27/2025	10:39	PQ071156.D	7.53	2.78
AR1660CCC500	AR1660CCC500	10/27/2025	12:35	PQ071161.D	7.53	2.78
I.BLK	I.BLK	10/27/2025	14:18	PQ071165.D	7.53	2.78



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

IDENTIFICATION SUMMARY
FOR MULTICOMPONENT ANALYTES

SAMPLE NO.

SS-20-(3.5-4.0)

Lab Name: Alliance

Contract: MATR02

Lab Code: ACE

SDG NO.: Q3450

Lab Sample ID: Q3450-10

Date(s) Analyzed: 10/24/2025 10/24/2025

Instrument ID (1): ECD_P

Instrument ID (2): ECD_P

GC Column: (1): ZB-MR1 ID: 0.32 (mm)

GC Column: (2): ZB-MR2 ID: 0.32 (mm)

Data file PP076047.D

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	MEAN CONCENTRATION	%RPD
			FROM	TO			
Aroclor-1254	1	6.639	6.589	6.689	11.5	55.2	5.59
COLUMN 1	2	6.856	6.806	6.906	25.9		
	3	7.218	7.168	7.268	49.4		
	4	7.5	7.45	7.55	75.2		
	5	7.917	7.867	7.967	114		
COLUMN 2	1	5.653	5.603	5.703	15.4	52.2	
	2	5.801	5.751	5.851	29.7		
	3	6.204	6.154	6.254	46.1		
	4	6.43	6.38	6.48	65.1		
	5	6.847	6.797	6.897	105		

IDENTIFICATION SUMMARY
FOR MULTICOMPONENT ANALYTES

SAMPLE NO.

SS-21-(0.0-0.5)

Lab Name: Alliance

Contract: MATR02

Lab Code: ACE

SDG NO.: Q3450

Lab Sample ID: Q3450-20

Date(s) Analyzed: 10/24/2025 10/24/2025

Instrument ID (1): ECD_P

Instrument ID (2): ECD_P

GC Column: (1): ZB-MR1 ID: 0.32 (mm)

GC Column: (2): ZB-MR2 ID: 0.32 (mm)

Data file PP076049.D

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	MEAN CONCENTRATION	%RPD
			FROM	TO			
Aroclor-1254	1	6.641	6.591	6.691	201	448	27.86
COLUMN 1	2	6.856	6.806	6.906	353		
	3	7.219	7.169	7.269	609		
	4	7.501	7.451	7.551	463		
	5	7.917	7.867	7.967	615		
COLUMN 2	1	5.656	5.606	5.706	296	593	
	2	5.804	5.754	5.854	406		
	3	6.205	6.155	6.255	721		
	4	6.431	6.381	6.481	596		
	5	6.848	6.798	6.898	944		

IDENTIFICATION SUMMARY
FOR MULTICOMPONENT ANALYTES

SAMPLE NO.

SS-21-(3.5-4.0)

Lab Name: Alliance

Contract: MATR02

Lab Code: ACE

SDG NO.: Q3450

Lab Sample ID: Q3450-22

Date(s) Analyzed: 10/24/2025 10/24/2025

Instrument ID (1): ECD_P

Instrument ID (2): ECD_P

GC Column: (1): ZB-MR1 ID: 0.32 (mm)

GC Column: (2): ZB-MR2 ID: 0.32 (mm)

Data file PP076051.D

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	MEAN CONCENTRATION	%RPD
			FROM	TO			
Aroclor-1254	1	6.64	6.59	6.69	59.4	403	24.59
COLUMN 1	2	6.856	6.806	6.906	164		
	3	7.218	7.168	7.268	372		
	4	7.5	7.45	7.55	531		
	5	7.916	7.866	7.966	888		
	1	5.657	5.607	5.707	89.0		
COLUMN 2	2	5.804	5.754	5.854	170	516	
	3	6.204	6.154	6.254	435		
	4	6.431	6.381	6.481	673		
	5	6.848	6.798	6.898	1220		

IDENTIFICATION SUMMARY
FOR MULTICOMPONENT ANALYTES

SAMPLE NO.

PB170236BS

Lab Name: Alliance

Contract: MATR02

Lab Code: ACE

SDG NO.: Q3450

Lab Sample ID: PB170236BS

Date(s) Analyzed: 10/27/2025 10/27/2025

Instrument ID (1): ECD_P

Instrument ID (2): ECD_P

GC Column: (1): ZB-MR1 ID: 0.32 (mm)

GC Column: (2): ZB-MR2 ID: 0.32 (mm)

Data file PP076069.D

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	MEAN CONCENTRATION	%RPD
			FROM	TO			
Aroclor-1016	1	5.793	5.743	5.843	146	145	13.24
COLUMN 1	2	5.814	5.764	5.864	147		
	3	5.877	5.827	5.927	145		
	4	5.974	5.924	6.024	146		
	5	6.266	6.216	6.316	140		
	1	4.877	4.827	4.927	132		
COLUMN 2	2	4.935	4.885	4.985	131	127	
	3	5.054	5.004	5.104	122		
	4	5.095	5.045	5.145	129		
	5	5.309	5.259	5.359	120		
Aroclor-1260	1	7.383	7.333	7.433	167	149	4.11
COLUMN 1	2	7.636	7.586	7.686	153		
	3	7.994	7.944	8.044	134		
	4	8.22	8.17	8.27	157		
	5	8.546	8.496	8.596	135		
	1	6.335	6.285	6.385	147	143	
COLUMN 2	2	6.522	6.472	6.572	147		
	3	6.675	6.625	6.725	152		
	4	7.144	7.094	7.194	142		
	5	7.384	7.334	7.434	129		

IDENTIFICATION SUMMARY
FOR MULTICOMPONENT ANALYTES

SAMPLE NO.

SS-21-(0.0-0.5)DL

Lab Name: Alliance

Contract: MATR02

Lab Code: ACE

SDG NO.: Q3450

Lab Sample ID: Q3450-20DL

Date(s) Analyzed: 10/27/2025 10/27/2025

Instrument ID (1): ECD_P

Instrument ID (2): ECD_P

GC Column: (1): ZB-MR1 ID: 0.32 (mm)

GC Column: (2): ZB-MR2 ID: 0.32 (mm)

Data file PP076072.D

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	MEAN CONCENTRATION	%RPD
			FROM	TO			
Aroclor-1254	1	6.641	6.591	6.691	262	509	6.91
COLUMN 1	2	6.856	6.806	6.906	401		
	3	7.219	7.169	7.269	657		
	4	7.5	7.45	7.55	524		
	5	7.917	7.867	7.967	699		
	1	5.657	5.607	5.707	308		
COLUMN 2	2	5.804	5.754	5.854	391	475	
	3	6.205	6.155	6.255	588		
	4	6.431	6.381	6.481	497		
	5	6.848	6.798	6.898	590		

IDENTIFICATION SUMMARY
FOR MULTICOMPONENT ANALYTES

SAMPLE NO.

SS-21-(3.5-4.0)DL

Lab Name: Alliance

Contract: MATR02

Lab Code: ACE

SDG NO.: Q3450

Lab Sample ID: Q3450-22DL

Date(s) Analyzed: 10/27/2025 10/27/2025

Instrument ID (1): ECD_P

Instrument ID (2): ECD_P

GC Column: (1): ZB-MR1 ID: 0.32 (mm)

GC Column: (2): ZB-MR2 ID: 0.32 (mm)

Data file PP076073.D

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	MEAN CONCENTRATION	%RPD
			FROM	TO			
Aroclor-1254	1	6.635	6.585	6.685	95.0	450	11.12
COLUMN 1	2	6.851	6.801	6.901	184		
	3	7.214	7.164	7.264	474		
	4	7.497	7.447	7.547	581		
	5	7.913	7.863	7.963	916		
	1	5.667	5.617	5.717	143		
COLUMN 2	2	5.799	5.749	5.849	172	503	
	3	6.202	6.152	6.252	434		
	4	6.428	6.378	6.478	672		
	5	6.844	6.794	6.894	1100		

IDENTIFICATION SUMMARY
FOR MULTICOMPONENT ANALYTES

SAMPLE NO.

SS-5R-(2.0-2.5)

Lab Name: Alliance

Contract: MATR02

Lab Code: ACE

SDG NO.: Q3450

Lab Sample ID: Q3450-01

Date(s) Analyzed: 10/27/2025 10/27/2025

Instrument ID (1): ECD_Q

Instrument ID (2): ECD_Q

GC Column: (1): ZB-MR1 ID: 0.32 (mm)

GC Column: (2): ZB-MR2 ID: 0.32 (mm)

Data file PQ071151.D

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	MEAN CONCENTRATION	%RPD
			FROM	TO			
Aroclor-1254	1	5.293	5.243	5.343	8.30	45.5	3.12
COLUMN 1	2	5.507	5.457	5.557	23.0		
	3	5.855	5.805	5.905	49.1		
	4	6.136	6.086	6.186	55.7		
	5	6.544	6.494	6.594	91.2		
	1	4.531	4.481	4.581	10.5		
COLUMN 2	2	4.677	4.627	4.727	23.7	44.1	
	3	5.057	5.007	5.107	45.6		
	4	5.284	5.234	5.334	55.2		
	5	5.688	5.638	5.738	85.7		

IDENTIFICATION SUMMARY
FOR MULTICOMPONENT ANALYTES

SAMPLE NO.

SS-5R-(3.5-4.0)

Lab Name: Alliance

Contract: MATR02

Lab Code: ACE

SDG NO.: Q3450

Lab Sample ID: Q3450-02

Date(s) Analyzed: 10/27/2025 10/27/2025

Instrument ID (1): ECD_Q

Instrument ID (2): ECD_Q

GC Column: (1): ZB-MR1 ID: 0.32 (mm)

GC Column: (2): ZB-MR2 ID: 0.32 (mm)

Data file PQ071152.D

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	MEAN CONCENTRATION	%RPD
			FROM	TO			
Aroclor-1254	1	5.291	5.241	5.341	3.69	11.3	3.6
COLUMN 1	2	5.507	5.457	5.557	7.74		
	3	5.854	5.804	5.904	12.4		
	4	6.134	6.084	6.184	12.9		
	5	6.543	6.493	6.593	20.0		
	1	4.531	4.481	4.581	2.91		
COLUMN 2	2	4.678	4.628	4.728	7.27	10.9	
	3	5.058	5.008	5.108	11.5		
	4	5.285	5.235	5.335	12.9		
	5	5.688	5.638	5.738	19.9		

IDENTIFICATION SUMMARY
FOR MULTICOMPONENT ANALYTES

SAMPLE NO.

SS-5R-(3.5-4.0)MS

Lab Name: Alliance

Contract: MATR02

Lab Code: ACE

SDG NO.: Q3450

Lab Sample ID: Q3450-02MS

Date(s) Analyzed: 10/27/2025 10/27/2025

Instrument ID (1): ECD_Q

Instrument ID (2): ECD_Q

GC Column: (1): ZB-MR1 ID: 0.32 (mm)

GC Column: (2): ZB-MR2 ID: 0.32 (mm)

Data file PQ071153.D

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	MEAN CONCENTRATION	%RPD
			FROM	TO			
Aroclor-1016	1	4.481	4.431	4.531	59.9	121	2.51
COLUMN 1	2	4.498	4.448	4.548	123		
	3	4.555	4.505	4.605	140		
	4	4.645	4.595	4.695	148		
	5	4.93	4.88	4.98	135		
COLUMN 2	1	3.775	3.725	3.825	131	118	
	2	3.791	3.741	3.841	110		
	3	3.951	3.901	4.001	119		
	4	3.999	3.949	4.049	115		
	5	4.195	4.145	4.245	114		
Aroclor-1254	1	5.291	5.241	5.341	73.6	105	13.52
COLUMN 1	2	5.506	5.456	5.556	65.1		
	3	5.854	5.804	5.904	52.0		
	4	6.132	6.082	6.182	65.2		
	5	6.542	6.492	6.592	271		
COLUMN 2	1	4.531	4.481	4.581	57.2	91.7	
	2	4.677	4.627	4.727	82.7		
	3	5.058	5.008	5.108	47.0		
	4	5.284	5.234	5.334	62.9		
	5	5.687	5.637	5.737	209		

IDENTIFICATION SUMMARY
FOR MULTICOMPONENT ANALYTES

Aroclor-1260	1	6.018	5.968	6.068	162	145	16.42
	2	6.275	6.225	6.325	162		
	3	6.626	6.576	6.676	125		
	4	6.84	6.79	6.89	137		
	5	7.146	7.096	7.196	139		
COLUMN 1							
	1	5.186	5.136	5.236	135		
	2	5.377	5.327	5.427	123		
	3	5.516	5.466	5.566	141		
	4	5.978	5.928	6.028	109		
COLUMN 2	5	6.222	6.172	6.272	107	123	

IDENTIFICATION SUMMARY
FOR MULTICOMPONENT ANALYTES

SAMPLE NO.

SS-5R-(3.5-4.0)MSD

Lab Name: Alliance

Contract: MATR02

Lab Code: ACE

SDG NO.: Q3450

Lab Sample ID: Q3450-02MSD

Date(s) Analyzed: 10/27/2025 10/27/2025

Instrument ID (1): ECD_Q

Instrument ID (2): ECD_Q

GC Column: (1): ZB-MR1 ID: 0.32 (mm)

GC Column: (2): ZB-MR2 ID: 0.32 (mm)

Data file PQ071154.D

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	MEAN CONCENTRATION	%RPD
			FROM	TO			
Aroclor-1016	1	4.478	4.428	4.528	155	162	12.46
COLUMN 1	2	4.497	4.447	4.547	160		
	3	4.555	4.505	4.605	166		
	4	4.645	4.595	4.695	166		
	5	4.93	4.88	4.98	161		
COLUMN 2	1	3.775	3.725	3.825	150	143	
	2	3.791	3.741	3.841	128		
	3	3.952	3.902	4.002	153		
	4	4	3.95	4.05	150		
	5	4.195	4.145	4.245	136		
Aroclor-1254	1	5.292	5.242	5.342	91.0	148	13.72
COLUMN 1	2	5.507	5.457	5.557	89.9		
	3	5.854	5.804	5.904	99.6		
	4	6.134	6.084	6.184	114		
	5	6.543	6.493	6.593	347		
COLUMN 2	1	4.531	4.481	4.581	74.2	129	
	2	4.678	4.628	4.728	97.9		
	3	5.058	5.008	5.108	90.8		
	4	5.284	5.234	5.334	102		
	5	5.687	5.637	5.737	280		

IDENTIFICATION SUMMARY
FOR MULTICOMPONENT ANALYTES

Aroclor-1260	1	6.019	5.969	6.069	196	166	8.15
	2	6.275	6.225	6.325	197		
	3	6.625	6.575	6.675	137		
	4	6.84	6.79	6.89	153		
	5	7.145	7.095	7.195	147		
COLUMN 1							
	1	5.187	5.137	5.237	183		
	2	5.376	5.326	5.426	165		
	3	5.516	5.466	5.566	176		
	4	5.977	5.927	6.027	121		
COLUMN 2	5	6.221	6.171	6.271	120	153	

IDENTIFICATION SUMMARY
FOR MULTICOMPONENT ANALYTES

SAMPLE NO.

SS-20-(0.0-0.5)

Lab Name: Alliance

Contract: MATR02

Lab Code: ACE

SDG NO.: Q3450

Lab Sample ID: Q3450-08

Date(s) Analyzed: 10/27/2025 10/27/2025

Instrument ID (1): ECD_Q

Instrument ID (2): ECD_Q

GC Column: (1): ZB-MR1 ID: 0.32 (mm)

GC Column: (2): ZB-MR2 ID: 0.32 (mm)

Data file PQ071155.D

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	MEAN CONCENTRATION	%RPD
			FROM	TO			
Aroclor-1254	1	5.292	5.242	5.342	97.3	174	4.71
COLUMN 1	2	5.506	5.456	5.556	127		
	3	5.853	5.803	5.903	201		
	4	6.134	6.084	6.184	168		
	5	6.542	6.492	6.592	277		
	1	4.531	4.481	4.581	92.8		
COLUMN 2	2	4.677	4.627	4.727	138	166	
	3	5.058	5.008	5.108	182		
	4	5.284	5.234	5.334	165		
	5	5.687	5.637	5.737	253		

IDENTIFICATION SUMMARY
FOR MULTICOMPONENT ANALYTES

SAMPLE NO.

SS-20-(2.0-2.5)

Lab Name: Alliance

Contract: MATR02

Lab Code: ACE

SDG NO.: Q3450

Lab Sample ID: Q3450-09

Date(s) Analyzed: 10/27/2025 10/27/2025

Instrument ID (1): ECD_Q

Instrument ID (2): ECD_Q

GC Column: (1): ZB-MR1 ID: 0.32 (mm)

GC Column: (2): ZB-MR2 ID: 0.32 (mm)

Data file PQ071156.D

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	MEAN CONCENTRATION	%RPD
			FROM	TO			
Aroclor-1254	1	5.292	5.242	5.342	4.87	11.2	0.89
COLUMN 1	2	5.506	5.456	5.556	6.00		
	3	5.854	5.804	5.904	12.7		
	4	6.134	6.084	6.184	10.3		
	5	6.542	6.492	6.592	22.0		
	1	4.53	4.48	4.58	3.99		
COLUMN 2	2	4.677	4.627	4.727	8.08	11.3	
	3	5.057	5.007	5.107	10.7		
	4	5.283	5.233	5.333	11.4		
	5	5.687	5.637	5.737	22.1		



QC SAMPLE DATA



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Matrix New World Engineering		Date Collected:	
Project:	PSEG Athenia Substation		Date Received:	
Client Sample ID:	PB170236BL		SDG No.:	Q3450
Lab Sample ID:	PB170236BL		Matrix:	SOIL
Analytical Method:	8082A		% Solid:	100
Sample Wt/Vol:	30.02 g	Final Vol: 10000 uL	Test:	PCB
Prep Method:	SW3541	Prep Date 10/24/25		

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
12674-11-2	Aroclor-1016	3.90	U	1	3.90	17.0	ug/kg	10/27/25 10:18	PB170236
11104-28-2	Aroclor-1221	4.00	U	1	4.00	17.0	ug/kg	10/27/25 10:18	PB170236
11141-16-5	Aroclor-1232	3.70	U	1	3.70	17.0	ug/kg	10/27/25 10:18	PB170236
53469-21-9	Aroclor-1242	4.00	U	1	4.00	17.0	ug/kg	10/27/25 10:18	PB170236
12672-29-6	Aroclor-1248	5.90	U	1	5.90	17.0	ug/kg	10/27/25 10:18	PB170236
11097-69-1	Aroclor-1254	3.20	U	1	3.20	17.0	ug/kg	10/27/25 10:18	PB170236
37324-23-5	Aroclor-1262	5.00	U	1	5.00	17.0	ug/kg	10/27/25 10:18	PB170236
11100-14-4	Aroclor-1268	3.60	U	1	3.60	17.0	ug/kg	10/27/25 10:18	PB170236
11096-82-5	Aroclor-1260	3.20	U	1	3.20	17.0	ug/kg	10/27/25 10:18	PB170236
SURROGATES									
877-09-8	Tetrachloro-m-xylene	27.3			30 (32) - 150 (144)	137%	SPK: 20		
2051-24-3	Decachlorobiphenyl	29.0			30 (32) - 150 (175)	145%	SPK: 20		

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102725\
 Data File : PP076068.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 Oct 2025 10:18
 Operator : YP\AJ
 Sample : PB170236BL
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 PB170236BL

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/28/2025
 Supervised By :mohammad ahmed 10/29/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 27 11:02:59 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Oct 24 15:40:32 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.637	3.780	39070596	172.6E6	27.339m	26.077
2) SA Decachlor...	10.403	8.773	31178050	207.4E6	28.962m	25.220

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102725\
Data File : PP076068.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Oct 2025 10:18
Operator : YP\AJ
Sample : PB170236BL
Misc :
ALS Vial : 7 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
PB170236BL

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/28/2025
Supervised By :mohammad ahmed 10/29/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 27 11:02:59 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 24 15:40:32 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Response_ Signal: PP076068.D\ECD1A.ch



#1 Tetrachloro-m-xylene


R.T.: 4.637 min
Delta R.T.: -0.005 min
Response: 39070596
Conc: 27.34 ng/ml

Instrument :
ECD_P
ClientSampleId :
PB170236BL

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/28/2025
Supervised By :mohammad ahmed 10/29/2025


Response_ Signal: PP076068.D\ECD2B.ch



#1 Tetrachloro-m-xylene

R.T.: 3.780 min
Delta R.T.: -0.002 min
Response: 172582422
Conc: 26.08 ng/ml


Response_ Signal: PP076068.D\ECD1A.ch



#2 Decachlorobiphenyl

R.T.: 10.403 min
Delta R.T.: -0.015 min
Response: 31178050
Conc: 28.96 ng/ml m

Response_ Signal: PP076068.D\ECD2B.ch



#2 Decachlorobiphenyl

R.T.: 8.773 min
Delta R.T.: -0.009 min
Response: 207383403
Conc: 25.22 ng/ml



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Matrix New World Engineering		Date Collected:	10/21/25	
Project:	PSEG Athenia Substation		Date Received:	10/21/25	
Client Sample ID:	PIBLK-PP075887.D		SDG No.:	Q3450	
Lab Sample ID:	I.BLK-PP075887.D		Matrix:	WATER	
Analytical Method:	8082A		% Solid:	0	
Sample Wt/Vol:	1000 mL	Final Vol:	10000 uL	Test:	PCB
Prep Method:	5030	Prep Date:			

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
12674-11-2	Aroclor-1016	0.097	U	1	0.097	0.50	ug/L	10/21/25	PP102125
11104-28-2	Aroclor-1221	0.13	U	1	0.13	0.50	ug/L	10/21/25	PP102125
11141-16-5	Aroclor-1232	0.096	U	1	0.096	0.50	ug/L	10/21/25	PP102125
53469-21-9	Aroclor-1242	0.12	U	1	0.12	0.50	ug/L	10/21/25	PP102125
12672-29-6	Aroclor-1248	0.071	U	1	0.071	0.50	ug/L	10/21/25	PP102125
11097-69-1	Aroclor-1254	0.094	U	1	0.094	0.50	ug/L	10/21/25	PP102125
11096-82-5	Aroclor-1260	0.081	U	1	0.081	0.50	ug/L	10/21/25	PP102125
37324-23-5	Aroclor-1262	0.14	U	1	0.14	0.50	ug/L	10/21/25	PP102125
11100-14-4	Aroclor-1268	0.11	U	1	0.11	0.50	ug/L	10/21/25	PP102125
SURROGATES									
877-09-8	Tetrachloro-m-xylene	18.8			70 (60) - 130 (140)	94%	SPK: 20		
2051-24-3	Decachlorobiphenyl	19.3			70 (60) - 130 (140)	96%	SPK: 20		

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
 Data File : PP075887.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Oct 2025 08:49
 Operator : YP\AJ
 Sample : I.BLK
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 I.BLK

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 10/22/2025
 Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 21 18:26:13 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Oct 21 18:24:24 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.644	3.783	28503905	124.7E6	19.945	18.843
2) SA Decachlor...	10.419	8.784	21974278	158.6E6	20.412m	19.281

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075887.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2025 08:49
Operator : YP\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
I.BLK

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/22/2025
Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 18:26:13 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 21 18:24:24 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Response_ Signal: PP075887.D\ECD1A.ch



#1 Tetrachloro-m-xylene


R.T.: 4.644 min
Delta R.T.: 0.002 min
Response: 28503905
Conc: 19.95 ng/ml

Instrument :
ECD_P
ClientSampleId :
I.BLK

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/22/2025
Supervised By :mohammad ahmed 10/25/2025


Response_ Signal: PP075887.D\ECD2B.ch



#1 Tetrachloro-m-xylene

R.T.: 3.783 min
Delta R.T.: 0.002 min
Response: 124709992
Conc: 18.84 ng/ml


Response_ Signal: PP075887.D\ECD1A.ch



#2 Decachlorobiphenyl

R.T.: 10.419 min
Delta R.T.: 0.000 min
Response: 21974278
Conc: 20.41 ng/ml m

Response_ Signal: PP075887.D\ECD2B.ch



#2 Decachlorobiphenyl

R.T.: 8.784 min
Delta R.T.: 0.002 min
Response: 158550080
Conc: 19.28 ng/ml



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Matrix New World Engineering		Date Collected:	10/24/25
Project:	PSEG Athenia Substation		Date Received:	10/24/25
Client Sample ID:	PIBLK-PP076046.D		SDG No.:	Q3450
Lab Sample ID:	I.BLK-PP076046.D		Matrix:	WATER
Analytical Method:	8082A		% Solid:	0
Sample Wt/Vol:	1000 mL	Final Vol:	10000 uL	Test:
Prep Method:	5030	Prep Date:		PCB

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
12674-11-2	Aroclor-1016	0.097	U	1	0.097	0.50	ug/L	10/24/25	pp102425
11104-28-2	Aroclor-1221	0.13	U	1	0.13	0.50	ug/L	10/24/25	pp102425
11141-16-5	Aroclor-1232	0.096	U	1	0.096	0.50	ug/L	10/24/25	pp102425
53469-21-9	Aroclor-1242	0.12	U	1	0.12	0.50	ug/L	10/24/25	pp102425
12672-29-6	Aroclor-1248	0.071	U	1	0.071	0.50	ug/L	10/24/25	pp102425
11097-69-1	Aroclor-1254	0.094	U	1	0.094	0.50	ug/L	10/24/25	pp102425
11096-82-5	Aroclor-1260	0.081	U	1	0.081	0.50	ug/L	10/24/25	pp102425
37324-23-5	Aroclor-1262	0.14	U	1	0.14	0.50	ug/L	10/24/25	pp102425
11100-14-4	Aroclor-1268	0.11	U	1	0.11	0.50	ug/L	10/24/25	pp102425
SURROGATES									
877-09-8	Tetrachloro-m-xylene	17.6			70 (60) - 130 (140)	88%	SPK: 20		
2051-24-3	Decachlorobiphenyl	20.1			70 (60) - 130 (140)	101%	SPK: 20		

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102425\
Data File : PP076046.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 24 Oct 2025 20:36
Operator : YP\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 24 23:40:03 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 24 15:40:32 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.642	3.782	29442757	116.3E6	20.602	17.572
2)	SA Decachlor...	10.413	8.778	25427395	165.5E6	23.620	20.126

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102425\
Data File : PP076046.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 24 Oct 2025 20:36
Operator : YP\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1


Instrument :
ECD_P
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 24 23:40:03 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 24 15:40:32 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm



Response_ Signal: PP076046.D\ECD1A.ch




#1 Tetrachloro-m-xylene

R.T.: 4.642 min
Delta R.T.: 0.000 min
Response: 29442757
Conc: 20.60 ng/ml

Instrument :
ECD_P
ClientSampleId :
I.BLK


Time Response_ Signal: PP076046.D\ECD2B.ch



#1 Tetrachloro-m-xylene

R.T.: 3.782 min
Delta R.T.: 0.000 min
Response: 116298499
Conc: 17.57 ng/ml


Time Response_ Signal: PP076046.D\ECD1A.ch



#2 Decachlorobiphenyl

R.T.: 10.413 min
Delta R.T.: -0.006 min
Response: 25427395
Conc: 23.62 ng/ml

Time Response_ Signal: PP076046.D\ECD2B.ch



#2 Decachlorobiphenyl

R.T.: 8.778 min
Delta R.T.: -0.004 min
Response: 165496707
Conc: 20.13 ng/ml



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Matrix New World Engineering		Date Collected:	10/25/25	
Project:	PSEG Athenia Substation		Date Received:	10/25/25	
Client Sample ID:	PIBLK-PP076061.D		SDG No.:	Q3450	
Lab Sample ID:	I.BLK-PP076061.D		Matrix:	WATER	
Analytical Method:	8082A		% Solid:	0	
Sample Wt/Vol:	1000 mL	Final Vol:	10000 uL	Test:	PCB
Prep Method:	5030	Prep Date:			

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
12674-11-2	Aroclor-1016	0.097	U	1	0.097	0.50	ug/L	10/25/25	pp102425
11104-28-2	Aroclor-1221	0.13	U	1	0.13	0.50	ug/L	10/25/25	pp102425
11141-16-5	Aroclor-1232	0.096	U	1	0.096	0.50	ug/L	10/25/25	pp102425
53469-21-9	Aroclor-1242	0.12	U	1	0.12	0.50	ug/L	10/25/25	pp102425
12672-29-6	Aroclor-1248	0.071	U	1	0.071	0.50	ug/L	10/25/25	pp102425
11097-69-1	Aroclor-1254	0.094	U	1	0.094	0.50	ug/L	10/25/25	pp102425
11096-82-5	Aroclor-1260	0.081	U	1	0.081	0.50	ug/L	10/25/25	pp102425
37324-23-5	Aroclor-1262	0.14	U	1	0.14	0.50	ug/L	10/25/25	pp102425
11100-14-4	Aroclor-1268	0.11	U	1	0.11	0.50	ug/L	10/25/25	pp102425
SURROGATES									
877-09-8	Tetrachloro-m-xylene	18.2			70 (60) - 130 (140)	91%	SPK: 20		
2051-24-3	Decachlorobiphenyl	21.0			70 (60) - 130 (140)	105%	SPK: 20		

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102425\
 Data File : PP076061.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 25 Oct 2025 02:16
 Operator : YP\AJ
 Sample : I.BLK
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 I.BLK

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 25 05:34:19 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Oct 24 15:40:32 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.643	3.783	29833113	120.5E6	20.876	18.205
2) SA Decachlor...	10.411	8.778	25539926	172.5E6	23.724	20.975

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102425\
Data File : PP076061.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 25 Oct 2025 02:16
Operator : YP\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1


Instrument :
ECD_P
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 25 05:34:19 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 24 15:40:32 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Response_ Signal: PP076061.D\ECD1A.ch




#1 Tetrachloro-m-xylene

R.T.: 4.643 min
Delta R.T.: 0.000 min
Response: 29833113
Conc: 20.88 ng/ml

Instrument :
ECD_P
ClientSampleId :
I.BLK


Response_ Signal: PP076061.D\ECD2B.ch



#1 Tetrachloro-m-xylene

R.T.: 3.783 min
Delta R.T.: 0.001 min
Response: 120487723
Conc: 18.21 ng/ml


Response_ Signal: PP076061.D\ECD1A.ch



#2 Decachlorobiphenyl

R.T.: 10.411 min
Delta R.T.: -0.008 min
Response: 25539926
Conc: 23.72 ng/ml

Response_ Signal: PP076061.D\ECD2B.ch



#2 Decachlorobiphenyl

R.T.: 8.778 min
Delta R.T.: -0.004 min
Response: 172478139
Conc: 20.98 ng/ml



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Matrix New World Engineering		Date Collected:	10/27/25
Project:	PSEG Athenia Substation		Date Received:	10/27/25
Client Sample ID:	PIBLK-PP076067.D		SDG No.:	Q3450
Lab Sample ID:	I.BLK-PP076067.D		Matrix:	WATER
Analytical Method:	8082A		% Solid:	0
Sample Wt/Vol:	1000 mL	Final Vol:	10000 uL	Test:
Prep Method:	5030	Prep Date:		PCB

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
12674-11-2	Aroclor-1016	0.097	U	1	0.097	0.50	ug/L	10/27/25	PP102725
11104-28-2	Aroclor-1221	0.13	U	1	0.13	0.50	ug/L	10/27/25	PP102725
11141-16-5	Aroclor-1232	0.096	U	1	0.096	0.50	ug/L	10/27/25	PP102725
53469-21-9	Aroclor-1242	0.12	U	1	0.12	0.50	ug/L	10/27/25	PP102725
12672-29-6	Aroclor-1248	0.071	U	1	0.071	0.50	ug/L	10/27/25	PP102725
11097-69-1	Aroclor-1254	0.094	U	1	0.094	0.50	ug/L	10/27/25	PP102725
11096-82-5	Aroclor-1260	0.081	U	1	0.081	0.50	ug/L	10/27/25	PP102725
37324-23-5	Aroclor-1262	0.14	U	1	0.14	0.50	ug/L	10/27/25	PP102725
11100-14-4	Aroclor-1268	0.11	U	1	0.11	0.50	ug/L	10/27/25	PP102725
SURROGATES									
877-09-8	Tetrachloro-m-xylene	18.9			70 (60) - 130 (140)	95%	SPK: 20		
2051-24-3	Decachlorobiphenyl	18.2			70 (60) - 130 (140)	91%	SPK: 20		

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102725\
Data File : PP076067.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Oct 2025 10:02
Operator : YP\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 27 11:02:24 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 24 15:40:32 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.638	3.778	31506659	125.1E6	22.047	18.907
2)	SA Decachlor...	10.404	8.773	24084542	150.0E6	22.372	18.238

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102725\
Data File : PP076067.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Oct 2025 10:02
Operator : YP\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1


Instrument :
ECD_P
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 27 11:02:24 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 24 15:40:32 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Response_ Signal: PP076067.D\ECD1A.ch




#1 Tetrachloro-m-xylene

R.T.: 4.638 min
Delta R.T.: -0.004 min
Response: 31506659
Conc: 22.05 ng/ml

Instrument :
ECD_P
ClientSampleId :
I.BLK


Time Response_ Signal: PP076067.D\ECD2B.ch



#1 Tetrachloro-m-xylene

R.T.: 3.778 min
Delta R.T.: -0.003 min
Response: 125130938
Conc: 18.91 ng/ml


Time Response_ Signal: PP076067.D\ECD1A.ch



#2 Decachlorobiphenyl

R.T.: 10.404 min
Delta R.T.: -0.015 min
Response: 24084542
Conc: 22.37 ng/ml

Time Response_ Signal: PP076067.D\ECD2B.ch



#2 Decachlorobiphenyl

R.T.: 8.773 min
Delta R.T.: -0.009 min
Response: 149972909
Conc: 18.24 ng/ml



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Matrix New World Engineering		Date Collected:	10/27/25	
Project:	PSEG Athenia Substation		Date Received:	10/27/25	
Client Sample ID:	PIBLK-PP076082.D		SDG No.:	Q3450	
Lab Sample ID:	I.BLK-PP076082.D		Matrix:	WATER	
Analytical Method:	8082A		% Solid:	0	
Sample Wt/Vol:	1000 mL	Final Vol:	10000 uL	Test:	PCB
Prep Method:	5030	Prep Date:			

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
12674-11-2	Aroclor-1016	0.097	U	1	0.097	0.50	ug/L	10/27/25	pp102725
11104-28-2	Aroclor-1221	0.13	U	1	0.13	0.50	ug/L	10/27/25	pp102725
11141-16-5	Aroclor-1232	0.096	U	1	0.096	0.50	ug/L	10/27/25	pp102725
53469-21-9	Aroclor-1242	0.12	U	1	0.12	0.50	ug/L	10/27/25	pp102725
12672-29-6	Aroclor-1248	0.071	U	1	0.071	0.50	ug/L	10/27/25	pp102725
11097-69-1	Aroclor-1254	0.094	U	1	0.094	0.50	ug/L	10/27/25	pp102725
11096-82-5	Aroclor-1260	0.081	U	1	0.081	0.50	ug/L	10/27/25	pp102725
37324-23-5	Aroclor-1262	0.14	U	1	0.14	0.50	ug/L	10/27/25	pp102725
11100-14-4	Aroclor-1268	0.11	U	1	0.11	0.50	ug/L	10/27/25	pp102725
SURROGATES									
877-09-8	Tetrachloro-m-xylene	17.1			70 (60) - 130 (140)	86%	SPK: 20		
2051-24-3	Decachlorobiphenyl	16.7			70 (60) - 130 (140)	83%	SPK: 20		

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102725\
Data File : PP076082.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Oct 2025 15:13
Operator : YP\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 28 01:03:28 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 24 15:40:32 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.639	3.778	27067383	113.2E6	18.940	17.111
2) SA Decachlor...	10.406	8.774	21916774	137.1E6	20.359	16.668

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102725\
Data File : PP076082.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Oct 2025 15:13
Operator : YP\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1


Instrument :
ECD_P
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 28 01:03:28 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 24 15:40:32 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm



Response_ Signal: PP076082.D\ECD1A.ch




#1 Tetrachloro-m-xylene

R.T.: 4.639 min
Delta R.T.: -0.003 min
Response: 27067383
Conc: 18.94 ng/ml

Instrument :
ECD_P
ClientSampleId :
I.BLK


Time Response_ Signal: PP076082.D\ECD2B.ch



#1 Tetrachloro-m-xylene

R.T.: 3.778 min
Delta R.T.: -0.003 min
Response: 113248293
Conc: 17.11 ng/ml


Time Response_ Signal: PP076082.D\ECD1A.ch



#2 Decachlorobiphenyl

R.T.: 10.406 min
Delta R.T.: -0.013 min
Response: 21916774
Conc: 20.36 ng/ml

Time Response_ Signal: PP076082.D\ECD2B.ch



#2 Decachlorobiphenyl

R.T.: 8.774 min
Delta R.T.: -0.008 min
Response: 137057682
Conc: 16.67 ng/ml

Report of Analysis

Client: Matrix New World Engineering	Date Collected: 10/10/25	
Project: PSEG Athenia Substation	Date Received: 10/10/25	
Client Sample ID: PIBLK-PQ070972.D	SDG No.: Q3450	
Lab Sample ID: I.BLK-PQ070972.D	Matrix: WATER	
Analytical Method: 8082A	% Solid: 0	
Sample Wt/Vol: 1000 mL	Final Vol: 10000 uL	Test: PCB
Prep Method: 5030	Prep Date	

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
12674-11-2	Aroclor-1016	0.097	U	1	0.097	0.50	ug/L	10/10/25	pq101025
11104-28-2	Aroclor-1221	0.13	U	1	0.13	0.50	ug/L	10/10/25	pq101025
11141-16-5	Aroclor-1232	0.096	U	1	0.096	0.50	ug/L	10/10/25	pq101025
53469-21-9	Aroclor-1242	0.12	U	1	0.12	0.50	ug/L	10/10/25	pq101025
12672-29-6	Aroclor-1248	0.071	U	1	0.071	0.50	ug/L	10/10/25	pq101025
11097-69-1	Aroclor-1254	0.094	U	1	0.094	0.50	ug/L	10/10/25	pq101025
11096-82-5	Aroclor-1260	0.081	U	1	0.081	0.50	ug/L	10/10/25	pq101025
37324-23-5	Aroclor-1262	0.14	U	1	0.14	0.50	ug/L	10/10/25	pq101025
11100-14-4	Aroclor-1268	0.11	U	1	0.11	0.50	ug/L	10/10/25	pq101025
SURROGATES									
877-09-8	Tetrachloro-m-xylene	18.3			70 (60) - 130 (140)	91%	SPK: 20		
2051-24-3	Decachlorobiphenyl	19.0			70 (60) - 130 (140)	95%	SPK: 20		

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ070972.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 08:42
Operator : YP\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 17:32:33 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 10 17:24:38 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.400	2.778	79509611	97022264	18.277	19.099
2)	SA Decachlor...	8.496	7.533	55246960	97962147	19.025	19.400

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ070972.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 08:42
Operator : YP\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 17:32:33 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 10 17:24:38 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.400 min
Delta R.T.: 0.000 min
Response: 79509611
Conc: 18.28 ng/ml

Instrument :
ECD_Q
ClientSampleId :
I.BLK




#1 Tetrachloro-m-xylene

R.T.: 2.778 min
Delta R.T.: 0.000 min
Response: 97022264
Conc: 19.10 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.496 min
Delta R.T.: 0.000 min
Response: 55246960
Conc: 19.02 ng/ml



#2 Decachlorobiphenyl

R.T.: 7.533 min
Delta R.T.: 0.000 min
Response: 97962147
Conc: 19.40 ng/ml

Report of Analysis

Client: Matrix New World Engineering	Date Collected: 10/27/25
Project: PSEG Athenia Substation	Date Received: 10/27/25
Client Sample ID: PIBLK-PQ071150.D	SDG No.: Q3450
Lab Sample ID: I.BLK-PQ071150.D	Matrix: WATER
Analytical Method: 8082A	% Solid: 0
Sample Wt/Vol: 1000 mL Final Vol: 10000 uL	Test: PCB
Prep Method: 5030 Prep Date	

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
12674-11-2	Aroclor-1016	0.097	U	1	0.097	0.50	ug/L	10/27/25	PQ102725
11104-28-2	Aroclor-1221	0.13	U	1	0.13	0.50	ug/L	10/27/25	PQ102725
11141-16-5	Aroclor-1232	0.096	U	1	0.096	0.50	ug/L	10/27/25	PQ102725
53469-21-9	Aroclor-1242	0.12	U	1	0.12	0.50	ug/L	10/27/25	PQ102725
12672-29-6	Aroclor-1248	0.071	U	1	0.071	0.50	ug/L	10/27/25	PQ102725
11097-69-1	Aroclor-1254	0.094	U	1	0.094	0.50	ug/L	10/27/25	PQ102725
11096-82-5	Aroclor-1260	0.081	U	1	0.081	0.50	ug/L	10/27/25	PQ102725
37324-23-5	Aroclor-1262	0.14	U	1	0.14	0.50	ug/L	10/27/25	PQ102725
11100-14-4	Aroclor-1268	0.11	U	1	0.11	0.50	ug/L	10/27/25	PQ102725
SURROGATES									
877-09-8	Tetrachloro-m-xylene	19.9			70 (60) - 130 (140)	100%	SPK: 20		
2051-24-3	Decachlorobiphenyl	24.5			70 (60) - 130 (140)	122%	SPK: 20		

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ102725\
 Data File : PQ071150.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 Oct 2025 09:11
 Operator : YP\AJ
 Sample : I.BLK
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 I.BLK

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 27 11:28:19 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Oct 11 01:55:41 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.401	2.776	97916618	101.1E6	22.508	19.909
2) SA Decachlor...	8.497	7.532	74642604	123.6E6	25.704	24.472

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ102725\
Data File : PQ071150.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Oct 2025 09:11
Operator : YP\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 27 11:28:19 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Oct 11 01:55:41 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.401 min
Delta R.T.: 0.000 min
Response: 97916618
Conc: 22.51 ng/ml

Instrument :
ECD_Q
ClientSampleId :
I.BLK




#1 Tetrachloro-m-xylene

R.T.: 2.776 min
Delta R.T.: -0.002 min
Response: 101136067
Conc: 19.91 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.497 min
Delta R.T.: -0.003 min
Response: 74642604
Conc: 25.70 ng/ml



#2 Decachlorobiphenyl

R.T.: 7.532 min
Delta R.T.: -0.001 min
Response: 123572351
Conc: 24.47 ng/ml

Report of Analysis

Client: Matrix New World Engineering	Date Collected: 10/27/25	
Project: PSEG Athenia Substation	Date Received: 10/27/25	
Client Sample ID: PIBLK-PQ071165.D	SDG No.: Q3450	
Lab Sample ID: I.BLK-PQ071165.D	Matrix: WATER	
Analytical Method: 8082A	% Solid: 0	
Sample Wt/Vol: 1000 mL	Final Vol: 10000 uL	Test: PCB
Prep Method: 5030	Prep Date	

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
12674-11-2	Aroclor-1016	0.097	U	1	0.097	0.50	ug/L	10/27/25	PQ102725
11104-28-2	Aroclor-1221	0.13	U	1	0.13	0.50	ug/L	10/27/25	PQ102725
11141-16-5	Aroclor-1232	0.096	U	1	0.096	0.50	ug/L	10/27/25	PQ102725
53469-21-9	Aroclor-1242	0.12	U	1	0.12	0.50	ug/L	10/27/25	PQ102725
12672-29-6	Aroclor-1248	0.071	U	1	0.071	0.50	ug/L	10/27/25	PQ102725
11097-69-1	Aroclor-1254	0.094	U	1	0.094	0.50	ug/L	10/27/25	PQ102725
11096-82-5	Aroclor-1260	0.081	U	1	0.081	0.50	ug/L	10/27/25	PQ102725
37324-23-5	Aroclor-1262	0.14	U	1	0.14	0.50	ug/L	10/27/25	PQ102725
11100-14-4	Aroclor-1268	0.11	U	1	0.11	0.50	ug/L	10/27/25	PQ102725
SURROGATES									
877-09-8	Tetrachloro-m-xylene	18.4			70 (60) - 130 (140)	92%	SPK: 20		
2051-24-3	Decachlorobiphenyl	22.2			70 (60) - 130 (140)	111%	SPK: 20		

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ102725\
Data File : PQ071165.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Oct 2025 14:18
Operator : YP\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 28 00:23:56 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Oct 11 01:55:41 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.399	2.776	88339084	93561979	20.307	18.418
2) SA Decachlor...	8.497	7.530	68135285	112.1E6	23.463	22.205

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ102725\
Data File : PQ071165.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Oct 2025 14:18
Operator : YP\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 28 00:23:56 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Oct 11 01:55:41 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.399 min
Delta R.T.: -0.001 min
Response: 88339084
Conc: 20.31 ng/ml

Instrument :
ECD_Q
ClientSampleId :
I.BLK




#1 Tetrachloro-m-xylene

R.T.: 2.776 min
Delta R.T.: -0.002 min
Response: 93561979
Conc: 18.42 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.497 min
Delta R.T.: -0.003 min
Response: 68135285
Conc: 23.46 ng/ml



#2 Decachlorobiphenyl

R.T.: 7.530 min
Delta R.T.: -0.003 min
Response: 112123551
Conc: 22.20 ng/ml



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Matrix New World Engineering		Date Collected:	
Project:	PSEG Athenia Substation		Date Received:	
Client Sample ID:	PB170236BS		SDG No.:	Q3450
Lab Sample ID:	PB170236BS		Matrix:	SOIL
Analytical Method:	8082A		% Solid:	100
Sample Wt/Vol:	30.01 g	Final Vol: 10000 uL	Test:	PCB
Prep Method:	SW3541	Prep Date 10/24/25		

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
12674-11-2	Aroclor-1016	145		1	3.90	17.0	ug/kg	10/27/25 10:34	PB170236
11104-28-2	Aroclor-1221	4.00	U	1	4.00	17.0	ug/kg	10/27/25 10:34	PB170236
11141-16-5	Aroclor-1232	3.70	U	1	3.70	17.0	ug/kg	10/27/25 10:34	PB170236
53469-21-9	Aroclor-1242	4.00	U	1	4.00	17.0	ug/kg	10/27/25 10:34	PB170236
12672-29-6	Aroclor-1248	5.90	U	1	5.90	17.0	ug/kg	10/27/25 10:34	PB170236
11097-69-1	Aroclor-1254	3.20	U	1	3.20	17.0	ug/kg	10/27/25 10:34	PB170236
37324-23-5	Aroclor-1262	5.00	U	1	5.00	17.0	ug/kg	10/27/25 10:34	PB170236
11100-14-4	Aroclor-1268	3.60	U	1	3.60	17.0	ug/kg	10/27/25 10:34	PB170236
11096-82-5	Aroclor-1260	149		1	3.20	17.0	ug/kg	10/27/25 10:34	PB170236
SURROGATES									
877-09-8	Tetrachloro-m-xylene	26.7			30 (32) - 150 (144)	134%	SPK: 20		
2051-24-3	Decachlorobiphenyl	28.6			30 (32) - 150 (175)	143%	SPK: 20		

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102725\
 Data File : PP076069.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 Oct 2025 10:34
 Operator : YP\AJ
 Sample : PB170236BS
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 PB170236BS

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 10/28/2025
 Supervised By :mohammad ahmed 10/29/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 27 11:18:04 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Oct 24 15:40:32 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.640	3.780	38149794	141.0E6	26.695m	21.306m
2) SA Decachlor...	10.408	8.775	30747417	199.5E6	28.562	24.264
Target Compounds						
3) L1 AR-1016-1	5.793	4.877	25825913	292.1E6	436.794	394.687
4) L1 AR-1016-2	5.814	4.935	39096018	134.3E6	442.042	392.888
5) L1 AR-1016-3	5.877	5.054	24165535	75316147	436.587	366.783
6) L1 AR-1016-4	5.974	5.095	19941002	72452629	438.840	387.684
7) L1 AR-1016-5	6.266	5.309	18220008	81872028	419.762	359.115
31) L7 AR-1260-1	7.383	6.335	34475825	242.2E6	501.418	442.275
32) L7 AR-1260-2	7.636	6.522	41196225	358.0E6	459.896	440.000
33) L7 AR-1260-3	7.994	6.675	27871115	256.4E6	402.259	455.573
34) L7 AR-1260-4	8.220	7.144	32220662	221.2E6	469.885m	425.862
35) L7 AR-1260-5	8.546	7.384	60142007	524.0E6	406.498	386.522

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102725\
Data File : PP076069.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Oct 2025 10:34
Operator : YP\AJ
Sample : PB170236BS
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :

ECD_P

ClientSampleId :

PB170236BS

Manual Integrations



APPROVED

Reviewed By :Yogesh Patel 10/28/2025

Supervised By :mohammad ahmed 10/29/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 27 11:18:04 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 24 15:40:32 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Matrix New World Engineering	Date Collected:	10/23/25
Project:	PSEG Athenia Substation	Date Received:	10/23/25
Client Sample ID:	SS-5R-(3.5-4.0)MS	SDG No.:	Q3450
Lab Sample ID:	Q3450-02MS	Matrix:	SOIL
Analytical Method:	8082A	% Solid:	86.3
Sample Wt/Vol:	30.06 g	Final Vol:	10000 uL
Prep Method:	SW3541	Test:	PCB
	Prep Date		10/24/25

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
12674-11-2	Aroclor-1016	121		1	4.60	19.7	ug/kg	10/27/25 09:55	PB170236
11104-28-2	Aroclor-1221	4.70	U	1	4.70	19.7	ug/kg	10/27/25 09:55	PB170236
11141-16-5	Aroclor-1232	4.30	U	1	4.30	19.7	ug/kg	10/27/25 09:55	PB170236
53469-21-9	Aroclor-1242	4.60	U	1	4.60	19.7	ug/kg	10/27/25 09:55	PB170236
12672-29-6	Aroclor-1248	6.80	U	1	6.80	19.7	ug/kg	10/27/25 09:55	PB170236
11097-69-1	Aroclor-1254	105		1	3.70	19.7	ug/kg	10/27/25 09:55	PB170236
37324-23-5	Aroclor-1262	5.80	U	1	5.80	19.7	ug/kg	10/27/25 09:55	PB170236
11100-14-4	Aroclor-1268	4.20	U	1	4.20	19.7	ug/kg	10/27/25 09:55	PB170236
11096-82-5	Aroclor-1260	145		1	3.70	19.7	ug/kg	10/27/25 09:55	PB170236
SURROGATES									
877-09-8	Tetrachloro-m-xylene	23.4			30 (32) - 150 (144)	117%	SPK: 20		
2051-24-3	Decachlorobiphenyl	22.4			30 (32) - 150 (175)	112%	SPK: 20		

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ102725\
 Data File : PQ071153.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 Oct 2025 09:55
 Operator : YP\AJ
 Sample : Q3450-02MS
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 SS-5R-(3.5-4.0)MS

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 10/27/2025
 Supervised By :mohammad ahmed 10/29/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 27 11:30:04 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Oct 11 01:55:41 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.399	2.776	101.9E6	95077060	23.434	18.716m
2) SA Decachlor...	8.495	7.531	64988561	104.5E6	22.379	20.702
Target Compounds						
3) L1 AR-1016-1	4.481	3.775	25519004	75904857	155.467m	338.858m#
4) L1 AR-1016-2	4.498	3.791	81532605	96925792	318.121m	284.973m
5) L1 AR-1016-3	4.555	3.951	57737086	56328796	362.856m	309.739m
6) L1 AR-1016-4	4.645	3.999	49897243	46929761	384.034m	298.132m
7) L1 AR-1016-5	4.930	4.195	44988750	56961892	349.284	296.921
26) L6 AR-1254-1	5.291	4.531	37478695	50045658	190.947m	148.410
27) L6 AR-1254-2	5.506	4.677	49023240	63782741	168.765m	214.645 #
28) L6 AR-1254-3	5.854	5.058	41414456	56889515	134.809m	121.802m
29) L6 AR-1254-4	6.132	5.284	37113579	48175061	169.051m	163.166
30) L6 AR-1254-5	6.542	5.687	163.1E6	227.7E6	703.418	541.766
31) L7 AR-1260-1	6.018	5.186	98581616	125.9E6	420.443m	349.310m
32) L7 AR-1260-2	6.275	5.377	116.0E6	140.7E6	419.237m	319.751
33) L7 AR-1260-3	6.626	5.516	64602779	151.4E6	324.135	364.829
34) L7 AR-1260-4	6.840	5.978	76642151	101.8E6	356.669	282.828
35) L7 AR-1260-5	7.146	6.222	156.2E6	234.0E6	360.118	278.474

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ102725\
Data File : PQ071153.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Oct 2025 09:55
Operator : YP\AJ
Sample : Q3450-02MS
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :

ECD_Q

ClientSampleId :

SS-5R-(3.5-4.0)MS

Manual Integrations



APPROVED

Reviewed By :Yogesh Patel 10/27/2025

Supervised By :mohammad ahmed 10/29/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 27 11:30:04 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Oct 11 01:55:41 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Report of Analysis

Client: Matrix New World Engineering	Date Collected: 10/23/25
Project: PSEG Athenia Substation	Date Received: 10/23/25
Client Sample ID: SS-5R-(3.5-4.0)MSD	SDG No.: Q3450
Lab Sample ID: Q3450-02MSD	Matrix: SOIL
Analytical Method: 8082A	% Solid: 86.3
Sample Wt/Vol: 30.04 g Final Vol: 10000 uL	Test: PCB
Prep Method: SW3541 Prep Date 10/24/25	

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
12674-11-2	Aroclor-1016	162		1	4.60	19.7	ug/kg	10/27/25 10:10	PB170236
11104-28-2	Aroclor-1221	4.70	U	1	4.70	19.7	ug/kg	10/27/25 10:10	PB170236
11141-16-5	Aroclor-1232	4.30	U	1	4.30	19.7	ug/kg	10/27/25 10:10	PB170236
53469-21-9	Aroclor-1242	4.60	U	1	4.60	19.7	ug/kg	10/27/25 10:10	PB170236
12672-29-6	Aroclor-1248	6.90	U	1	6.90	19.7	ug/kg	10/27/25 10:10	PB170236
11097-69-1	Aroclor-1254	148		1	3.70	19.7	ug/kg	10/27/25 10:10	PB170236
37324-23-5	Aroclor-1262	5.80	U	1	5.80	19.7	ug/kg	10/27/25 10:10	PB170236
11100-14-4	Aroclor-1268	4.20	U	1	4.20	19.7	ug/kg	10/27/25 10:10	PB170236
11096-82-5	Aroclor-1260	166		1	3.70	19.7	ug/kg	10/27/25 10:10	PB170236
SURROGATES									
877-09-8	Tetrachloro-m-xylene	25.7			30 (32) - 150 (144)	128%	SPK: 20		
2051-24-3	Decachlorobiphenyl	24.0			30 (32) - 150 (175)	120%	SPK: 20		

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ102725\
 Data File : PQ071154.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 Oct 2025 10:10
 Operator : YP\AJ
 Sample : Q3450-02MSD
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 SS-5R-(3.5-4.0)MSD

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 10/27/2025
 Supervised By :mohammad ahmed 10/29/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 27 11:30:42 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Oct 11 01:55:41 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.399	2.776	111.8E6	112.9E6	25.689	22.228
2) SA Decachlor...	8.496	7.531	69730856	115.6E6	24.012	22.895
Target Compounds						
3) L1 AR-1016-1	4.478	3.775	66133318	86839134	402.898m	387.671m
4) L1 AR-1016-2	4.497	3.791	106.6E6	113.2E6	415.843m	332.751m
5) L1 AR-1016-3	4.555	3.952	68384934	72254323	429.774m	397.310
6) L1 AR-1016-4	4.645	4.000	55794798	61092237	429.424m	388.102
7) L1 AR-1016-5	4.930	4.195	53701684	67804394	416.929	353.439
26) L6 AR-1254-1	5.292	4.531	46316796	64856412	235.975	192.331
27) L6 AR-1254-2	5.507	4.678	67669054	75430352	232.954	253.842
28) L6 AR-1254-3	5.854	5.058	79307585	109.9E6	258.156	235.279
29) L6 AR-1254-4	6.134	5.284	64884968	78163021	295.549	264.734
30) L6 AR-1254-5	6.543	5.687	208.6E6	305.6E6	899.407	726.931
31) L7 AR-1260-1	6.019	5.187	119.0E6	170.6E6	507.455	473.416
32) L7 AR-1260-2	6.275	5.376	141.2E6	188.2E6	510.578	427.657
33) L7 AR-1260-3	6.625	5.516	71042700	189.8E6	356.446	457.353 #
34) L7 AR-1260-4	6.840	5.977	85195497	113.2E6	396.473	314.727
35) L7 AR-1260-5	7.145	6.221	165.8E6	260.8E6	382.125	310.420

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ102725\
Data File : PQ071154.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Oct 2025 10:10
Operator : YP\AJ
Sample : Q3450-02MSD
Misc :
ALS Vial : 14 Sample Multiplier: 1



Instrument :
ECD_Q
ClientSampleId :
SS-5R-(3.5-4.0)MSD

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/27/2025
Supervised By :mohammad ahmed 10/29/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 27 11:30:42 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Oct 11 01:55:41 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Manual Integration Report

Sequence:	PP102125	Instrument	ECD_p

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
I.BLK	PP075887.D	Decachlorobiphenyl	yogesh	10/22/2025 12:24:41 PM	mohammad	10/25/2025 7:28:37	Peak Integrated by Software
AR1660ICC050	PP075892.D	AR-1016-3	yogesh	10/22/2025 12:25:25 PM	mohammad	10/25/2025 7:28:37	Peak Integrated by Software
AR1660ICC050	PP075892.D	AR-1016-3 #2	yogesh	10/22/2025 12:25:25 PM	mohammad	10/25/2025 7:28:37	Peak Integrated by Software
AR1660ICC050	PP075892.D	AR-1016-4	yogesh	10/22/2025 12:25:25 PM	mohammad	10/25/2025 7:28:37	Peak Integrated by Software
AR1660ICC050	PP075892.D	AR-1016-4 #2	yogesh	10/22/2025 12:25:25 PM	mohammad	10/25/2025 7:28:37	Peak Integrated by Software
AR1660ICC050	PP075892.D	AR-1016-5 #2	yogesh	10/22/2025 12:25:25 PM	mohammad	10/25/2025 7:28:37	Peak Integrated by Software
AR1660ICC050	PP075892.D	Tetrachloro-m-xylene #2	yogesh	10/22/2025 12:25:25 PM	mohammad	10/25/2025 7:28:37	Peak Integrated by Software
AR1232ICC500	PP075894.D	AR-1232-4 #2	yogesh	10/22/2025 12:24:44 PM	mohammad	10/25/2025 7:28:37	Peak Integrated by Software
AR1242ICC100 0	PP075895.D	AR-1242-4 #2	yogesh	10/22/2025 12:24:45 PM	mohammad	10/25/2025 7:28:37	Peak Integrated by Software
AR1242ICC750	PP075896.D	AR-1242-4 #2	yogesh	10/22/2025 12:24:46 PM	mohammad	10/25/2025 7:28:37	Peak Integrated by Software
AR1242ICC500	PP075897.D	AR-1242-4 #2	yogesh	10/22/2025 12:24:48 PM	mohammad	10/25/2025 7:28:37	Peak Integrated by Software
AR1242ICC250	PP075898.D	AR-1242-4 #2	yogesh	10/22/2025 12:24:50 PM	mohammad	10/25/2025 7:28:37	Peak Integrated by Software
AR1242ICC250	PP075898.D	AR-1242-5	yogesh	10/22/2025 12:24:50 PM	mohammad	10/25/2025 7:28:37	Peak Integrated by Software

Manual Integration Report

Sequence:	PP102125	Instrument	ECD_p

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
AR1242ICC050	PP075899.D	AR-1242-5	yogesh	10/22/2025 12:24:52 PM	mohammad	10/25/2025 7:28:37	Peak Integrated by Software
AR1242ICC050	PP075899.D	Tetrachloro-m-xylene #2	yogesh	10/22/2025 12:24:52 PM	mohammad	10/25/2025 7:28:37	Peak Integrated by Software
AR1248ICC100 0	PP075900.D	AR-1248-3 #2	yogesh	10/22/2025 12:24:53 PM	mohammad	10/25/2025 7:28:37	Peak Integrated by Software
AR1248ICC750	PP075901.D	AR-1248-3 #2	yogesh	10/22/2025 12:24:55 PM	mohammad	10/25/2025 7:28:37	Peak Integrated by Software
AR1248ICC500	PP075902.D	AR-1248-3 #2	yogesh	10/22/2025 12:24:57 PM	mohammad	10/25/2025 7:28:37	Peak Integrated by Software
AR1248ICC250	PP075903.D	AR-1248-3 #2	yogesh	10/22/2025 12:24:58 PM	mohammad	10/25/2025 7:28:37	Peak Integrated by Software
AR1248ICC050	PP075904.D	AR-1248-4	yogesh	10/22/2025 12:25:00 PM	mohammad	10/25/2025 7:28:37	Peak Integrated by Software
AR1248ICC050	PP075904.D	AR-1248-4 #2	yogesh	10/22/2025 12:25:00 PM	mohammad	10/25/2025 7:28:37	Peak Integrated by Software
AR1248ICC050	PP075904.D	AR-1248-5	yogesh	10/22/2025 12:25:00 PM	mohammad	10/25/2025 7:28:37	Peak Integrated by Software
AR1248ICC050	PP075904.D	AR-1248-5 #2	yogesh	10/22/2025 12:25:00 PM	mohammad	10/25/2025 7:28:37	Peak Integrated by Software
AR1248ICC050	PP075904.D	Tetrachloro-m-xylene #2	yogesh	10/22/2025 12:25:00 PM	mohammad	10/25/2025 7:28:37	Peak Integrated by Software
AR1254ICC100 0	PP075905.D	AR-1254-1	yogesh	10/22/2025 12:25:02 PM	mohammad	10/25/2025 7:28:37	Peak Integrated by Software
AR1254ICC750	PP075906.D	AR-1254-1	yogesh	10/22/2025 12:25:03 PM	mohammad	10/25/2025 7:28:37	Peak Integrated by Software

Manual Integration Report

Sequence:

PP102125

Instrument

ECD_p

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
AR1254ICC750	PP075906.D	AR-1254-5 #2	yogesh	10/22/2025 12:25:03 PM	mohammad	10/25/2025 7:28:37	Peak Integrated by Software
AR1254ICC500	PP075907.D	AR-1254-1	yogesh	10/22/2025 12:25:05 PM	mohammad	10/25/2025 7:28:37	Peak Integrated by Software
AR1254ICC250	PP075908.D	AR-1254-1	yogesh	10/22/2025 12:25:07 PM	mohammad	10/25/2025 7:28:37	Peak Integrated by Software
AR1254ICC050	PP075909.D	Tetrachloro-m-xylene #2	yogesh	10/22/2025 12:25:09 PM	mohammad	10/25/2025 7:28:37	Peak Integrated by Software
AR1242ICV500	PP075917.D	AR-1242-4 #2	yogesh	10/24/2025 4:12:30 PM	mohammad	10/25/2025 7:28:37	Peak Integrated by Software
AR1242ICV500	PP075917.D	Decachlorobiphenyl	yogesh	10/24/2025 4:12:30 PM	mohammad	10/25/2025 7:28:37	Peak Integrated by Software
AR1248ICV500	PP075918.D	AR-1248-3 #2	yogesh	10/24/2025 4:12:35 PM	mohammad	10/25/2025 7:28:37	Peak Integrated by Software
AR1254ICV500	PP075919.D	AR-1254-1 #2	yogesh	10/24/2025 4:12:33 PM	mohammad	10/25/2025 7:28:37	Peak Integrated by Software

Manual Integration Report

Sequence:	PP102425	Instrument	ECD_p

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
AR1242CCC500	PP076013.D	AR-1242-1 #2	yogesh	10/24/2025 4:09:50 PM	mohammad	10/28/2025 6:10:31	Peak Integrated by Software
AR1242CCC500	PP076013.D	AR-1242-2 #2	yogesh	10/24/2025 4:09:50 PM	mohammad	10/28/2025 6:10:31	Peak Integrated by Software
AR1242CCC500	PP076013.D	AR-1242-3 #2	yogesh	10/24/2025 4:09:50 PM	mohammad	10/28/2025 6:10:31	Peak Integrated by Software
AR1242CCC500	PP076013.D	AR-1242-4 #2	yogesh	10/24/2025 4:09:50 PM	mohammad	10/28/2025 6:10:31	Peak Integrated by Software
AR1242CCC500	PP076013.D	AR-1242-5 #2	yogesh	10/24/2025 4:09:50 PM	mohammad	10/28/2025 6:10:31	Peak Integrated by Software
AR1248CCC500	PP076014.D	AR-1248-3 #2	yogesh	10/24/2025 4:09:51 PM	mohammad	10/28/2025 6:10:31	Peak Integrated by Software
AR1248CCC500	PP076014.D	AR-1248-5 #2	yogesh	10/24/2025 4:09:51 PM	mohammad	10/28/2025 6:10:31	Peak Integrated by Software
AR1254CCC500	PP076015.D	AR-1254-1	yogesh	10/24/2025 4:09:57 PM	mohammad	10/28/2025 6:10:31	Peak Integrated by Software
AR1242CCC500	PP076028.D	AR-1242-4 #2	yogesh	10/24/2025 4:10:02 PM	mohammad	10/28/2025 6:10:31	Peak Integrated by Software
AR1248CCC500	PP076029.D	AR-1248-3 #2	yogesh	10/24/2025 4:10:04 PM	mohammad	10/28/2025 6:10:31	Peak Integrated by Software
AR1254CCC500	PP076030.D	AR-1254-1	yogesh	10/27/2025 8:23:51 AM	mohammad	10/28/2025 6:10:31	Peak Integrated by Software
AR1242CCC500	PP076043.D	AR-1242-4 #2	yogesh	10/27/2025 8:24:14 AM	mohammad	10/28/2025 6:10:31	Peak Integrated by Software
AR1248CCC500	PP076044.D	AR-1248-3 #2	yogesh	10/27/2025 8:24:15 AM	mohammad	10/28/2025 6:10:31	Peak Integrated by Software

Manual Integration Report

Sequence:

PP102425

Instrument

ECD_p

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
AR1254CCC500	PP076045.D	AR-1254-1	yogesh	10/27/2025 8:24:17 AM	mohammad	10/28/2025 6:10:31	Peak Integrated by Software
AR1254CCC500	PP076045.D	AR-1254-1 #2	yogesh	10/27/2025 8:24:17 AM	mohammad	10/28/2025 6:10:31	Peak Integrated by Software
AR1254CCC500	PP076045.D	Tetrachloro-m-xylene #2	yogesh	10/27/2025 8:24:17 AM	mohammad	10/28/2025 6:10:31	Peak Integrated by Software
Q3450-10	PP076047.D	AR-1254-1	yogesh	10/27/2025 8:24:19 AM	mohammad	10/28/2025 6:10:31	Peak Integrated by Software
AR1242CCC500	PP076058.D	AR-1242-4 #2	yogesh	10/27/2025 8:24:32 AM	mohammad	10/28/2025 6:10:31	Peak Integrated by Software
AR1248CCC500	PP076059.D	AR-1248-3 #2	yogesh	10/27/2025 8:24:34 AM	mohammad	10/28/2025 6:10:31	Peak Integrated by Software
AR1254CCC500	PP076060.D	AR-1254-1	yogesh	10/27/2025 8:24:36 AM	mohammad	10/28/2025 6:10:31	Peak Integrated by Software
AR1254CCC500	PP076060.D	AR-1254-1 #2	yogesh	10/27/2025 8:24:36 AM	mohammad	10/28/2025 6:10:31	Peak Integrated by Software
AR1254CCC500	PP076060.D	AR-1254-2 #2	yogesh	10/27/2025 8:24:36 AM	mohammad	10/28/2025 6:10:31	Peak Integrated by Software
AR1254CCC500	PP076060.D	AR-1254-3 #2	yogesh	10/27/2025 8:24:36 AM	mohammad	10/28/2025 6:10:31	Peak Integrated by Software
AR1254CCC500	PP076060.D	AR-1254-4 #2	yogesh	10/27/2025 8:24:36 AM	mohammad	10/28/2025 6:10:31	Peak Integrated by Software
AR1254CCC500	PP076060.D	AR-1254-5 #2	yogesh	10/27/2025 8:24:36 AM	mohammad	10/28/2025 6:10:31	Peak Integrated by Software



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

Manual Integration Report

Sequence:

PP102425

Instrument

ECD_p

Sample ID

File ID

Parameter

Review By

Review On

Supervised
By

Supervised On

Reason

Manual Integration Report

Sequence:	PP102725	Instrument	ECD_p

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
AR1660CCC500	PP076063.D	AR-1016-1 #2	yogesh	10/28/2025 7:45:04 AM	mohammad	10/29/2025 5:27:46	Peak Integrated by Software
AR1660CCC500	PP076063.D	AR-1016-2 #2	yogesh	10/28/2025 7:45:04 AM	mohammad	10/29/2025 5:27:46	Peak Integrated by Software
AR1242CCC500	PP076064.D	AR-1242-4 #2	yogesh	10/28/2025 7:45:06 AM	mohammad	10/29/2025 5:27:46	Peak Integrated by Software
AR1248CCC500	PP076065.D	AR-1248-3 #2	yogesh	10/28/2025 7:45:08 AM	mohammad	10/29/2025 5:27:46	Peak Integrated by Software
AR1254CCC500	PP076066.D	AR-1254-1	yogesh	10/28/2025 7:45:20 AM	mohammad	10/29/2025 5:27:46	Peak Integrated by Software
PB170236BL	PP076068.D	Decachlorobiphenyl	yogesh	10/28/2025 7:45:22 AM	mohammad	10/29/2025 5:27:46	Peak Integrated by Software
PB170236BL	PP076068.D	Tetrachloro-m-xylene	yogesh	10/28/2025 7:45:22 AM	mohammad	10/29/2025 5:27:46	Peak Integrated by Software
PB170236BS	PP076069.D	AR-1260-4	yogesh	10/28/2025 7:45:24 AM	mohammad	10/29/2025 5:27:46	Peak Integrated by Software
PB170236BS	PP076069.D	Tetrachloro-m-xylene	yogesh	10/28/2025 7:45:24 AM	mohammad	10/29/2025 5:27:46	Peak Integrated by Software
PB170236BS	PP076069.D	Tetrachloro-m-xylene #2	yogesh	10/28/2025 7:45:24 AM	mohammad	10/29/2025 5:27:46	Peak Integrated by Software
Q3450-20DL	PP076072.D	Tetrachloro-m-xylene	yogesh	10/28/2025 7:45:26 AM	mohammad	10/29/2025 5:27:46	Peak Integrated by Software
Q3450-22DL	PP076073.D	AR-1254-1	yogesh	10/28/2025 7:45:27 AM	mohammad	10/29/2025 5:27:46	Peak Integrated by Software
Q3450-22DL	PP076073.D	AR-1254-1 #2	yogesh	10/28/2025 7:45:27 AM	mohammad	10/29/2025 5:27:46	Peak Integrated by Software

Manual Integration Report

Sequence:	PP102725	Instrument	ECD_p

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
Q3450-22DL	PP076073.D	AR-1254-2	yogesh	10/28/2025 7:45:27 AM	mohammad	10/29/2025 5:27:46	Peak Integrated by Software
Q3450-22DL	PP076073.D	AR-1254-2 #2	yogesh	10/28/2025 7:45:27 AM	mohammad	10/29/2025 5:27:46	Peak Integrated by Software
AR1242CCC50 0	PP076079.D	AR-1242-4 #2	yogesh	10/28/2025 7:45:37 AM	mohammad	10/29/2025 5:27:46	Peak Integrated by Software
AR1248CCC50 0	PP076080.D	AR-1248-3 #2	yogesh	10/28/2025 7:45:40 AM	mohammad	10/29/2025 5:27:46	Peak Integrated by Software
AR1248CCC50 0	PP076080.D	AR-1248-5 #2	yogesh	10/28/2025 7:45:40 AM	mohammad	10/29/2025 5:27:46	Peak Integrated by Software
AR1254CCC50 0	PP076081.D	AR-1254-1	yogesh	10/28/2025 7:45:42 AM	mohammad	10/29/2025 5:27:46	Peak Integrated by Software
AR1242CCC50 0	PP076094.D	AR-1242-4 #2	yogesh	10/28/2025 7:46:25 AM	mohammad	10/29/2025 5:27:46	Peak Integrated by Software
AR1248CCC50 0	PP076095.D	AR-1248-3 #2	yogesh	10/28/2025 7:46:27 AM	mohammad	10/29/2025 5:27:46	Peak Integrated by Software
AR1254CCC50 0	PP076096.D	AR-1254-1	yogesh	10/28/2025 7:46:29 AM	mohammad	10/29/2025 5:27:46	Peak Integrated by Software
AR1254CCC50 0	PP076096.D	AR-1254-1 #2	yogesh	10/28/2025 7:46:29 AM	mohammad	10/29/2025 5:27:46	Peak Integrated by Software
AR1242CCC50 0	PP076105.D	AR-1242-4 #2	yogesh	10/28/2025 7:46:39 AM	mohammad	10/29/2025 5:27:46	Peak Integrated by Software
AR1242CCC50 0	PP076105.D	AR-1242-5 #2	yogesh	10/28/2025 7:46:39 AM	mohammad	10/29/2025 5:27:46	Peak Integrated by Software
AR1248CCC50 0	PP076106.D	AR-1248-3 #2	yogesh	10/28/2025 7:46:41 AM	mohammad	10/29/2025 5:27:46	Peak Integrated by Software



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

Manual Integration Report

Sequence:

PP102725

Instrument

ECD_p

Sample ID

File ID

Parameter

Review By

Review On

Supervised
By

Supervised On

Reason

Manual Integration Report

Sequence:	pq101025	Instrument	ECD_q

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
AR1660ICC050	PQ070977.D	AR-1260-1	yogesh	10/13/2025 8:00:35 AM	mohammad	10/14/2025 2:48:24	Peak Integrated by Software
AR1660ICC050	PQ070977.D	AR-1260-2	yogesh	10/13/2025 8:00:35 AM	mohammad	10/14/2025 2:48:24	Peak Integrated by Software
AR1242ICC050	PQ070984.D	AR-1242-1 #2	yogesh	10/13/2025 8:00:37 AM	mohammad	10/14/2025 2:48:24	Peak Integrated by Software
AR1242ICC050	PQ070984.D	AR-1242-2 #2	yogesh	10/13/2025 8:00:37 AM	mohammad	10/14/2025 2:48:24	Peak Integrated by Software
AR1242ICC050	PQ070984.D	AR-1242-3 #2	yogesh	10/13/2025 8:00:37 AM	mohammad	10/14/2025 2:48:24	Peak Integrated by Software
AR1242ICC050	PQ070984.D	AR-1242-4 #2	yogesh	10/13/2025 8:00:37 AM	mohammad	10/14/2025 2:48:24	Peak Integrated by Software
AR1242ICC050	PQ070984.D	AR-1242-5	yogesh	10/13/2025 8:00:37 AM	mohammad	10/14/2025 2:48:24	Peak Integrated by Software
AR1242ICC050	PQ070984.D	AR-1242-5 #2	yogesh	10/13/2025 8:00:37 AM	mohammad	10/14/2025 2:48:24	Peak Integrated by Software
AR1248ICC250	PQ070988.D	AR-1248-4 #2	yogesh	10/13/2025 8:00:39 AM	mohammad	10/14/2025 2:48:24	Peak Integrated by Software
AR1248ICC250	PQ070988.D	AR-1248-5 #2	yogesh	10/13/2025 8:00:39 AM	mohammad	10/14/2025 2:48:24	Peak Integrated by Software
PQ101025ICV	PQ071001.D	AR-1016-1 #2	yogesh	10/13/2025 8:00:41 AM	mohammad	10/14/2025 2:48:24	Peak Integrated by Software
PQ101025ICV	PQ071001.D	AR-1016-2 #2	yogesh	10/13/2025 8:00:41 AM	mohammad	10/14/2025 2:48:24	Peak Integrated by Software
PQ101025ICV	PQ071001.D	AR-1016-3 #2	yogesh	10/13/2025 8:00:41 AM	mohammad	10/14/2025 2:48:24	Peak Integrated by Software

Manual Integration Report

Sequence:	pq101025	Instrument	ECD_q

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
PQ101025ICV	PQ071001.D	AR-1016-4 #2	yogesh	10/13/2025 8:00:41 AM	mohammad	10/14/2025 2:48:24	Peak Integrated by Software
PQ101025ICV	PQ071001.D	AR-1016-5 #2	yogesh	10/13/2025 8:00:41 AM	mohammad	10/14/2025 2:48:24	Peak Integrated by Software
PQ101025ICV	PQ071001.D	AR-1260-1	yogesh	10/13/2025 8:00:41 AM	mohammad	10/14/2025 2:48:24	Peak Integrated by Software
PQ101025ICV	PQ071001.D	AR-1260-1 #2	yogesh	10/13/2025 8:00:41 AM	mohammad	10/14/2025 2:48:24	Peak Integrated by Software
PQ101025ICV	PQ071001.D	AR-1260-2	yogesh	10/13/2025 8:00:41 AM	mohammad	10/14/2025 2:48:24	Peak Integrated by Software
PQ101025ICV	PQ071001.D	AR-1260-2 #2	yogesh	10/13/2025 8:00:41 AM	mohammad	10/14/2025 2:48:24	Peak Integrated by Software
AR1242ICV500	PQ071002.D	AR-1242-3 #2	yogesh	10/13/2025 8:00:43 AM	mohammad	10/14/2025 2:48:24	Peak Integrated by Software
AR1248ICV500	PQ071003.D	AR-1248-4	yogesh	10/13/2025 8:00:45 AM	mohammad	10/14/2025 2:48:24	Peak Integrated by Software
AR1248ICV500	PQ071003.D	AR-1248-4 #2	yogesh	10/13/2025 8:00:45 AM	mohammad	10/14/2025 2:48:24	Peak Integrated by Software
AR1248ICV500	PQ071003.D	AR-1248-5	yogesh	10/13/2025 8:00:45 AM	mohammad	10/14/2025 2:48:24	Peak Integrated by Software

Manual Integration Report

Sequence:	PQ102725	Instrument	ECD_q

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
Q3450-02MS	PQ071153.D	AR-1016-1	yogesh	10/27/2025 1:46:49 PM	mohammad	10/29/2025 5:27:51	Peak Integrated by Software
Q3450-02MS	PQ071153.D	AR-1016-1 #2	yogesh	10/27/2025 1:46:49 PM	mohammad	10/29/2025 5:27:51	Peak Integrated by Software
Q3450-02MS	PQ071153.D	AR-1016-2	yogesh	10/27/2025 1:46:49 PM	mohammad	10/29/2025 5:27:51	Peak Integrated by Software
Q3450-02MS	PQ071153.D	AR-1016-2 #2	yogesh	10/27/2025 1:46:49 PM	mohammad	10/29/2025 5:27:51	Peak Integrated by Software
Q3450-02MS	PQ071153.D	AR-1016-3	yogesh	10/27/2025 1:46:49 PM	mohammad	10/29/2025 5:27:51	Peak Integrated by Software
Q3450-02MS	PQ071153.D	AR-1016-3 #2	yogesh	10/27/2025 1:46:49 PM	mohammad	10/29/2025 5:27:51	Peak Integrated by Software
Q3450-02MS	PQ071153.D	AR-1016-4	yogesh	10/27/2025 1:46:49 PM	mohammad	10/29/2025 5:27:51	Peak Integrated by Software
Q3450-02MS	PQ071153.D	AR-1016-4 #2	yogesh	10/27/2025 1:46:49 PM	mohammad	10/29/2025 5:27:51	Peak Integrated by Software
Q3450-02MS	PQ071153.D	AR-1254-1	yogesh	10/27/2025 1:46:49 PM	mohammad	10/29/2025 5:27:51	Peak Integrated by Software
Q3450-02MS	PQ071153.D	AR-1254-2	yogesh	10/27/2025 1:46:49 PM	mohammad	10/29/2025 5:27:51	Peak Integrated by Software
Q3450-02MS	PQ071153.D	AR-1254-3	yogesh	10/27/2025 1:46:49 PM	mohammad	10/29/2025 5:27:51	Peak Integrated by Software
Q3450-02MS	PQ071153.D	AR-1254-3 #2	yogesh	10/27/2025 1:46:49 PM	mohammad	10/29/2025 5:27:51	Peak Integrated by Software
Q3450-02MS	PQ071153.D	AR-1254-4	yogesh	10/27/2025 1:46:49 PM	mohammad	10/29/2025 5:27:51	Peak Integrated by Software

Manual Integration Report

Sequence:	PQ102725	Instrument	ECD_q

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
Q3450-02MS	PQ071153.D	AR-1260-1	yogesh	10/27/2025 1:46:49 PM	mohammad	10/29/2025 5:27:51	Peak Integrated by Software
Q3450-02MS	PQ071153.D	AR-1260-1 #2	yogesh	10/27/2025 1:46:49 PM	mohammad	10/29/2025 5:27:51	Peak Integrated by Software
Q3450-02MS	PQ071153.D	AR-1260-2	yogesh	10/27/2025 1:46:49 PM	mohammad	10/29/2025 5:27:51	Peak Integrated by Software
Q3450-02MS	PQ071153.D	Tetrachloro-m-xylene #2	yogesh	10/27/2025 1:46:49 PM	mohammad	10/29/2025 5:27:51	Peak Integrated by Software
Q3450-02MSD	PQ071154.D	AR-1016-1	yogesh	10/27/2025 1:46:51 PM	mohammad	10/29/2025 5:27:51	Peak Integrated by Software
Q3450-02MSD	PQ071154.D	AR-1016-1 #2	yogesh	10/27/2025 1:46:51 PM	mohammad	10/29/2025 5:27:51	Peak Integrated by Software
Q3450-02MSD	PQ071154.D	AR-1016-2	yogesh	10/27/2025 1:46:51 PM	mohammad	10/29/2025 5:27:51	Peak Integrated by Software
Q3450-02MSD	PQ071154.D	AR-1016-2 #2	yogesh	10/27/2025 1:46:51 PM	mohammad	10/29/2025 5:27:51	Peak Integrated by Software
Q3450-02MSD	PQ071154.D	AR-1016-3	yogesh	10/27/2025 1:46:51 PM	mohammad	10/29/2025 5:27:51	Peak Integrated by Software
Q3450-02MSD	PQ071154.D	AR-1016-4	yogesh	10/27/2025 1:46:51 PM	mohammad	10/29/2025 5:27:51	Peak Integrated by Software
Q3450-09	PQ071156.D	AR-1254-3 #2	yogesh	10/27/2025 1:46:53 PM	mohammad	10/29/2025 5:27:51	Peak Integrated by Software
Q3450-09	PQ071156.D	AR-1254-4 #2	yogesh	10/27/2025 1:46:53 PM	mohammad	10/29/2025 5:27:51	Peak Integrated by Software
AR1660CCC50 0	PQ071176.D	AR-1016-3	yogesh	10/28/2025 7:20:00 AM	mohammad	10/29/2025 5:27:51	Peak Integrated by Software

Manual Integration Report

Sequence:	PQ102725	Instrument	ECD_q

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
AR1660CCC500	PQ071176.D	AR-1016-4	yogesh	10/28/2025 7:20:00 AM	mohammad	10/29/2025 5:27:51	Peak Integrated by Software
AR1660CCC500	PQ071176.D	AR-1260-1	yogesh	10/28/2025 7:20:00 AM	mohammad	10/29/2025 5:27:51	Peak Integrated by Software
AR1660CCC500	PQ071176.D	AR-1260-3	yogesh	10/28/2025 7:20:00 AM	mohammad	10/29/2025 5:27:51	Peak Integrated by Software
AR1660CCC500	PQ071176.D	AR-1260-4	yogesh	10/28/2025 7:20:00 AM	mohammad	10/29/2025 5:27:51	Peak Integrated by Software
AR1660CCC500	PQ071176.D	AR-1260-5	yogesh	10/28/2025 7:20:00 AM	mohammad	10/29/2025 5:27:51	Peak Integrated by Software
AR1254CCC500	PQ071179.D	AR-1254-1	yogesh	10/28/2025 7:20:03 AM	mohammad	10/29/2025 5:27:51	Peak Integrated by Software
AR1254CCC500	PQ071179.D	AR-1254-2	yogesh	10/28/2025 7:20:03 AM	mohammad	10/29/2025 5:27:51	Peak Integrated by Software
AR1254CCC500	PQ071179.D	AR-1254-3	yogesh	10/28/2025 7:20:03 AM	mohammad	10/29/2025 5:27:51	Peak Integrated by Software
AR1254CCC500	PQ071179.D	AR-1254-4	yogesh	10/28/2025 7:20:03 AM	mohammad	10/29/2025 5:27:51	Peak Integrated by Software
AR1254CCC500	PQ071179.D	AR-1254-5	yogesh	10/28/2025 7:20:03 AM	mohammad	10/29/2025 5:27:51	Peak Integrated by Software
AR1254CCC500	PQ071179.D	Decachlorobiphenyl	yogesh	10/28/2025 7:20:03 AM	mohammad	10/29/2025 5:27:51	Peak Integrated by Software

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP102125

Review By	yogesh	Review On	10/21/2025 12:01:15 PM
Supervise By	mohammad	Supervise On	10/25/2025 7:28:37 AM
SubDirectory	PP102125	HP Acquire Method	HP Processing Method PP102125
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854 PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862		

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	HEXANE	PP075886.D	21 Oct 2025 08:33	YP\AJ	Ok
2	I.BLK	PP075887.D	21 Oct 2025 08:49	YP\AJ	Ok,M
3	AR1660ICC1000	PP075888.D	21 Oct 2025 09:06	YP\AJ	Ok
4	AR1660ICC750	PP075889.D	21 Oct 2025 09:22	YP\AJ	Ok
5	AR1660ICC500	PP075890.D	21 Oct 2025 09:38	YP\AJ	Ok
6	AR1660ICC250	PP075891.D	21 Oct 2025 09:54	YP\AJ	Ok
7	AR1660ICC050	PP075892.D	21 Oct 2025 10:11	YP\AJ	Ok,M
8	AR1221ICC500	PP075893.D	21 Oct 2025 10:27	YP\AJ	Ok
9	AR1232ICC500	PP075894.D	21 Oct 2025 10:43	YP\AJ	Ok,M
10	AR1242ICC1000	PP075895.D	21 Oct 2025 10:59	YP\AJ	Ok,M
11	AR1242ICC750	PP075896.D	21 Oct 2025 11:16	YP\AJ	Ok,M
12	AR1242ICC500	PP075897.D	21 Oct 2025 11:48	YP\AJ	Ok,M
13	AR1242ICC250	PP075898.D	21 Oct 2025 12:04	YP\AJ	Ok,M
14	AR1242ICC050	PP075899.D	21 Oct 2025 12:21	YP\AJ	Ok,M
15	AR1248ICC1000	PP075900.D	21 Oct 2025 12:41	YP\AJ	Ok,M
16	AR1248ICC750	PP075901.D	21 Oct 2025 12:57	YP\AJ	Ok,M
17	AR1248ICC500	PP075902.D	21 Oct 2025 13:13	YP\AJ	Ok,M
18	AR1248ICC250	PP075903.D	21 Oct 2025 13:30	YP\AJ	Ok,M
19	AR1248ICC050	PP075904.D	21 Oct 2025 13:46	YP\AJ	Ok,M
20	AR1254ICC1000	PP075905.D	21 Oct 2025 14:02	YP\AJ	Ok,M
21	AR1254ICC750	PP075906.D	21 Oct 2025 14:18	YP\AJ	Ok,M

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP102125

Review By	yogesh	Review On	10/21/2025 12:01:15 PM
Supervise By	mohammad	Supervise On	10/25/2025 7:28:37 AM
SubDirectory	PP102125	HP Acquire Method	HP Processing Method PP102125
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854		
CCC			
Internal Standard/PEM			
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

22	AR1254ICC500	PP075907.D	21 Oct 2025 14:35	YP\AJ	Ok,M
23	AR1254ICC250	PP075908.D	21 Oct 2025 14:51	YP\AJ	Ok,M
24	AR1254ICC050	PP075909.D	21 Oct 2025 15:07	YP\AJ	Ok,M
25	AR1262ICC500	PP075910.D	21 Oct 2025 15:40	YP\AJ	Ok
26	AR1268ICC1000	PP075911.D	21 Oct 2025 15:56	YP\AJ	Ok
27	AR1268ICC750	PP075912.D	21 Oct 2025 16:12	YP\AJ	Ok
28	AR1268ICC500	PP075913.D	21 Oct 2025 16:29	YP\AJ	Ok
29	AR1268ICC250	PP075914.D	21 Oct 2025 16:45	YP\AJ	Ok
30	AR1268ICC050	PP075915.D	21 Oct 2025 17:01	YP\AJ	Ok
31	PP102125ICV500	PP075916.D	21 Oct 2025 17:34	YP\AJ	Ok
32	AR1242ICV500	PP075917.D	21 Oct 2025 18:06	YP\AJ	Ok,M
33	AR1248ICV500	PP075918.D	21 Oct 2025 18:39	YP\AJ	Ok,M
34	AR1254ICV500	PP075919.D	22 Oct 2025 08:49	YP\AJ	Ok,M
35	AR1268ICV500	PP075920.D	22 Oct 2025 09:07	YP\AJ	Ok
36	DDT ANALOG	PP075921.D	22 Oct 2025 09:25	YP\AJ	Ok

M : Manual Integration

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP102425

Review By	yogesh	Review On	10/24/2025 11:09:51 AM
Supervise By	mohammad	Supervise On	10/28/2025 6:10:31 AM
SubDirectory	PP102425	HP Acquire Method	HP Processing Method PP102125
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854		
CCC	PP24824,PP24828,PP24832,PP24836,PP24840,PP24844,PP24852		
Internal Standard/PEM			
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	HEXANE	PP076011.D	24 Oct 2025 08:41	YP\AJ	Ok
2	AR1660CCC500	PP076012.D	24 Oct 2025 08:57	YP\AJ	Ok
3	AR1242CCC500	PP076013.D	24 Oct 2025 09:29	YP\AJ	Ok,M
4	AR1248CCC500	PP076014.D	24 Oct 2025 09:45	YP\AJ	Ok,M
5	AR1254CCC500	PP076015.D	24 Oct 2025 10:02	YP\AJ	Ok,M
6	I.BLK	PP076016.D	24 Oct 2025 10:18	YP\AJ	Ok
7	PB170234BL	PP076017.D	24 Oct 2025 10:34	YP\AJ	Ok
8	PB170234BS	PP076018.D	24 Oct 2025 10:51	YP\AJ	Ok
9	Q3437-03	PP076019.D	24 Oct 2025 11:07	YP\AJ	Ok
10	Q3437-05	PP076020.D	24 Oct 2025 11:23	YP\AJ	Ok,M
11	Q3437-11	PP076021.D	24 Oct 2025 11:39	YP\AJ	Ok,M
12	Q3445-01	PP076022.D	24 Oct 2025 11:56	YP\AJ	Ok
13	Q3446-01	PP076023.D	24 Oct 2025 12:12	YP\AJ	Dilution
14	Q3446-02	PP076024.D	24 Oct 2025 12:28	YP\AJ	Dilution
15	Q3414-12DL2	PP076025.D	24 Oct 2025 12:44	YP\AJ	Ok
16	Q3414-13DL2	PP076026.D	24 Oct 2025 13:01	YP\AJ	Ok
17	AR1660CCC500	PP076027.D	24 Oct 2025 14:06	YP\AJ	Ok
18	AR1242CCC500	PP076028.D	24 Oct 2025 14:22	YP\AJ	Ok,M
19	AR1248CCC500	PP076029.D	24 Oct 2025 14:38	YP\AJ	Ok,M
20	AR1254CCC500	PP076030.D	24 Oct 2025 14:54	YP\AJ	Ok,M
21	I.BLK	PP076031.D	24 Oct 2025 15:11	YP\AJ	Ok

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP102425

Review By	yogesh	Review On	10/24/2025 11:09:51 AM
Supervise By	mohammad	Supervise On	10/28/2025 6:10:31 AM
SubDirectory	PP102425	HP Acquire Method	HP Processing Method PP102125
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854		
CCC	PP24824,PP24828,PP24832,PP24836,PP24840,PP24844,PP24852		
Internal Standard/PEM			
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

22	Q3414-07DL2	PP076032.D	24 Oct 2025 15:27	YP\AJ	Ok
23	Q3446-06	PP076033.D	24 Oct 2025 15:43	YP\AJ	Dilution
24	Q3446-07	PP076034.D	24 Oct 2025 16:00	YP\AJ	Dilution
25	Q3446-08	PP076035.D	24 Oct 2025 16:16	YP\AJ	Dilution
26	Q3446-03	PP076036.D	24 Oct 2025 16:32	YP\AJ	Dilution
27	Q3446-04	PP076037.D	24 Oct 2025 16:48	YP\AJ	Dilution
28	Q3446-01DL	PP076038.D	24 Oct 2025 17:05	YP\AJ	Ok,M
29	Q3446-02DL	PP076039.D	24 Oct 2025 17:21	YP\AJ	Ok,M
30	Q3446-05DL	PP076040.D	24 Oct 2025 17:37	YP\AJ	Ok,M
31	Q3446-06DL	PP076041.D	24 Oct 2025 17:53	YP\AJ	Ok,M
32	AR1660CCC500	PP076042.D	24 Oct 2025 18:58	YP\AJ	Ok
33	AR1242CCC500	PP076043.D	24 Oct 2025 19:31	YP\AJ	Ok,M
34	AR1248CCC500	PP076044.D	24 Oct 2025 20:03	YP\AJ	Ok,M
35	AR1254CCC500	PP076045.D	24 Oct 2025 20:19	YP\AJ	Ok,M
36	I.BLK	PP076046.D	24 Oct 2025 20:36	YP\AJ	Ok
37	Q3450-10	PP076047.D	24 Oct 2025 20:52	YP\AJ	Ok,M
38	Q3450-11	PP076048.D	24 Oct 2025 21:08	YP\AJ	Ok
39	Q3450-20	PP076049.D	24 Oct 2025 21:24	YP\AJ	Dilution
40	Q3450-21	PP076050.D	24 Oct 2025 21:41	YP\AJ	Ok
41	Q3450-22	PP076051.D	24 Oct 2025 21:57	YP\AJ	Dilution
42	Q3427-01	PP076052.D	24 Oct 2025 22:13	YP\AJ	Ok,M
43	Q3446-07DL	PP076053.D	24 Oct 2025 22:29	YP\AJ	Ok,M
44	Q3446-08DL	PP076054.D	24 Oct 2025 22:45	YP\AJ	Ok,M

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP102425

Review By	yogesh	Review On	10/24/2025 11:09:51 AM
Supervise By	mohammad	Supervise On	10/28/2025 6:10:31 AM
SubDirectory	PP102425	HP Acquire Method	HP Processing Method PP102125
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854		
CCC	PP24824,PP24828,PP24832,PP24836,PP24840,PP24844,PP24852		
Internal Standard/PEM			
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

45	Q3446-03DL	PP076055.D	24 Oct 2025 23:01	YP\AJ	Ok,M
46	Q3446-04DL	PP076056.D	24 Oct 2025 23:17	YP\AJ	Ok,M
47	AR1660CCC500	PP076057.D	25 Oct 2025 00:22	YP\AJ	Ok
48	AR1242CCC500	PP076058.D	25 Oct 2025 01:11	YP\AJ	Ok,M
49	AR1248CCC500	PP076059.D	25 Oct 2025 01:43	YP\AJ	Ok,M
50	AR1254CCC500	PP076060.D	25 Oct 2025 01:59	YP\AJ	Ok,M
51	I.BLK	PP076061.D	25 Oct 2025 02:16	YP\AJ	Ok

M : Manual Integration

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP102725

Review By	yogesh	Review On	10/27/2025 10:53:48 AM
Supervise By	mohammad	Supervise On	10/29/2025 5:27:46 AM
SubDirectory	PP102725	HP Acquire Method	HP Processing Method PP102125
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854		
CCC	PP24824,PP24828,PP24832,PP24836,PP24840,PP24844,PP24852		
Internal Standard/PEM			
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	HEXANE	PP076062.D	27 Oct 2025 08:40	YP\AJ	Ok
2	AR1660CCC500	PP076063.D	27 Oct 2025 08:56	YP\AJ	Ok,M
3	AR1242CCC500	PP076064.D	27 Oct 2025 09:13	YP\AJ	Ok,M
4	AR1248CCC500	PP076065.D	27 Oct 2025 09:29	YP\AJ	Ok,M
5	AR1254CCC500	PP076066.D	27 Oct 2025 09:45	YP\AJ	Ok,M
6	I.BLK	PP076067.D	27 Oct 2025 10:02	YP\AJ	Ok
7	PB170236BL	PP076068.D	27 Oct 2025 10:18	YP\AJ	Ok,M
8	PB170236BS	PP076069.D	27 Oct 2025 10:34	YP\AJ	Ok,M
9	Q3439-01	PP076070.D	27 Oct 2025 11:06	YP\AJ	ReRun
10	Q3449-01	PP076071.D	27 Oct 2025 11:26	YP\AJ	Ok
11	Q3450-20DL	PP076072.D	27 Oct 2025 11:42	YP\AJ	Ok,M
12	Q3450-22DL	PP076073.D	27 Oct 2025 11:58	YP\AJ	Ok,M
13	Q3446-05	PP076074.D	27 Oct 2025 12:14	YP\AJ	Not Ok
14	Q3447-03	PP076075.D	27 Oct 2025 12:30	YP\AJ	ReRun
15	Q3447-07	PP076076.D	27 Oct 2025 12:47	YP\AJ	Ok,M
16	Q3446-05	PP076077.D	27 Oct 2025 13:03	YP\AJ	Dilution
17	AR1660CCC500	PP076078.D	27 Oct 2025 14:08	YP\AJ	Ok
18	AR1242CCC500	PP076079.D	27 Oct 2025 14:24	YP\AJ	Ok,M
19	AR1248CCC500	PP076080.D	27 Oct 2025 14:40	YP\AJ	Ok,M
20	AR1254CCC500	PP076081.D	27 Oct 2025 14:57	YP\AJ	Ok,M
21	I.BLK	PP076082.D	27 Oct 2025 15:13	YP\AJ	Ok

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP102725

Review By	yogesh	Review On	10/27/2025 10:53:48 AM
Supervise By	mohammad	Supervise On	10/29/2025 5:27:46 AM
SubDirectory	PP102725	HP Acquire Method	HP Processing Method PP102125
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854		
CCC	PP24824,PP24828,PP24832,PP24836,PP24840,PP24844,PP24852		
Internal Standard/PEM			
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

22	Q3456-01	PP076083.D	27 Oct 2025 15:29	YP\AJ	Ok,M
23	Q3457-01	PP076084.D	27 Oct 2025 15:45	YP\AJ	Ok,M
24	Q3460-01	PP076085.D	27 Oct 2025 16:02	YP\AJ	Ok
25	Q3460-17	PP076086.D	27 Oct 2025 16:18	YP\AJ	Ok,M
26	Q3447-05	PP076087.D	27 Oct 2025 16:34	YP\AJ	Ok,M
27	Q3447-01	PP076088.D	27 Oct 2025 16:50	YP\AJ	ReRun
28	Q3459-01	PP076089.D	27 Oct 2025 17:07	YP\AJ	Ok,M
29	Q3459-02	PP076090.D	27 Oct 2025 17:23	YP\AJ	Ok,M
30	Q3459-03	PP076091.D	27 Oct 2025 17:39	YP\AJ	Ok,M
31	Q3459-04	PP076092.D	27 Oct 2025 17:55	YP\AJ	Ok
32	AR1660CCC500	PP076093.D	27 Oct 2025 19:00	YP\AJ	Ok
33	AR1242CCC500	PP076094.D	27 Oct 2025 19:49	YP\AJ	Ok,M
34	AR1248CCC500	PP076095.D	27 Oct 2025 20:05	YP\AJ	Ok,M
35	AR1254CCC500	PP076096.D	27 Oct 2025 20:21	YP\AJ	Ok,M
36	I.BLK	PP076097.D	27 Oct 2025 20:38	YP\AJ	Ok
37	PB170272BL	PP076098.D	27 Oct 2025 20:54	YP\AJ	Ok
38	PB170272BS	PP076099.D	27 Oct 2025 21:10	YP\AJ	Ok,M
39	Q3459-05	PP076100.D	27 Oct 2025 21:26	YP\AJ	Ok,M
40	Q3459-06	PP076101.D	27 Oct 2025 21:42	YP\AJ	Ok,M
41	Q3459-07	PP076102.D	27 Oct 2025 21:59	YP\AJ	Ok,M
42	Q3459-08	PP076103.D	27 Oct 2025 22:15	YP\AJ	Ok,M
43	AR1660CCC500	PP076104.D	27 Oct 2025 23:20	YP\AJ	Ok
44	AR1242CCC500	PP076105.D	28 Oct 2025 00:08	YP\AJ	Ok,M

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP102725

Review By	yogesh	Review On	10/27/2025 10:53:48 AM
Supervise By	mohammad	Supervise On	10/29/2025 5:27:46 AM
SubDirectory	PP102725	HP Acquire Method	HP Processing Method PP102125
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854		
CCC	PP24824,PP24828,PP24832,PP24836,PP24840,PP24844,PP24852		
Internal Standard/PEM			
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

45	AR1248CCC500	PP076106.D	28 Oct 2025 00:25	YP\AJ	Ok,M
46	AR1254CCC500	PP076107.D	28 Oct 2025 00:41	YP\AJ	Ok
47	I.BLK	PP076108.D	28 Oct 2025 00:57	YP\AJ	Ok

M : Manual Integration

Instrument ID: ECD_Q

Daily Analysis Runlog For Sequence/QC Batch ID # PQ101025

Review By	yogesh	Review On	10/10/2025 10:11:25 AM
Supervise By	mohammad	Supervise On	10/14/2025 2:48:24 AM
SubDirectory	PQ101025	HP Acquire Method	HP Processing Method PQ101025
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854 PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862		

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	HEXANE	PQ070971.D	10 Oct 2025 08:27	YP\AJ	Ok
2	I.BLK	PQ070972.D	10 Oct 2025 08:42	YP\AJ	Ok
3	AR1660ICC1000	PQ070973.D	10 Oct 2025 08:57	YP\AJ	Ok
4	AR1660ICC750	PQ070974.D	10 Oct 2025 09:11	YP\AJ	Ok
5	AR1660ICC500	PQ070975.D	10 Oct 2025 09:26	YP\AJ	Ok
6	AR1660ICC250	PQ070976.D	10 Oct 2025 09:40	YP\AJ	Ok
7	AR1660ICC050	PQ070977.D	10 Oct 2025 09:55	YP\AJ	Ok,M
8	AR1221ICC500	PQ070978.D	10 Oct 2025 10:09	YP\AJ	Ok
9	AR1232ICC500	PQ070979.D	10 Oct 2025 10:24	YP\AJ	Ok
10	AR1242ICC1000	PQ070980.D	10 Oct 2025 10:38	YP\AJ	Ok
11	AR1242ICC750	PQ070981.D	10 Oct 2025 10:53	YP\AJ	Ok
12	AR1242ICC500	PQ070982.D	10 Oct 2025 11:08	YP\AJ	Ok
13	AR1242ICC250	PQ070983.D	10 Oct 2025 11:22	YP\AJ	Ok
14	AR1242ICC050	PQ070984.D	10 Oct 2025 11:37	YP\AJ	Ok,M
15	AR1248ICC1000	PQ070985.D	10 Oct 2025 11:51	YP\AJ	Ok
16	AR1248ICC750	PQ070986.D	10 Oct 2025 12:06	YP\AJ	Ok
17	AR1248ICC500	PQ070987.D	10 Oct 2025 12:20	YP\AJ	Ok
18	AR1248ICC250	PQ070988.D	10 Oct 2025 12:35	YP\AJ	Ok,M
19	AR1248ICC050	PQ070989.D	10 Oct 2025 12:50	YP\AJ	Ok
20	AR1254ICC1000	PQ070990.D	10 Oct 2025 13:04	YP\AJ	Ok
21	AR1254ICC750	PQ070991.D	10 Oct 2025 13:19	YP\AJ	Ok

Instrument ID: ECD_Q

Daily Analysis Runlog For Sequence/QC Batch ID # PQ101025

Review By	yogesh	Review On	10/10/2025 10:11:25 AM
Supervise By	mohammad	Supervise On	10/14/2025 2:48:24 AM
SubDirectory	PQ101025	HP Acquire Method	HP Processing Method PQ101025
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854		
CCC			
Internal Standard/PEM			
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

22	AR1254ICC500	PQ070992.D	10 Oct 2025 13:33	YPIAJ	Ok
23	AR1254ICC250	PQ070993.D	10 Oct 2025 13:48	YPIAJ	Ok
24	AR1254ICC050	PQ070994.D	10 Oct 2025 15:15	YPIAJ	Ok
25	AR1262ICC500	PQ070995.D	10 Oct 2025 15:30	YPIAJ	Ok
26	AR1268ICC1000	PQ070996.D	10 Oct 2025 15:44	YPIAJ	Ok
27	AR1268ICC750	PQ070997.D	10 Oct 2025 15:59	YPIAJ	Ok
28	AR1268ICC500	PQ070998.D	10 Oct 2025 16:13	YPIAJ	Ok
29	AR1268ICC250	PQ070999.D	10 Oct 2025 16:28	YPIAJ	Ok
30	AR1268ICC050	PQ071000.D	10 Oct 2025 16:43	YPIAJ	Ok
31	PQ101025ICV	PQ071001.D	10 Oct 2025 16:57	YPIAJ	Ok,M
32	AR1242ICV500	PQ071002.D	10 Oct 2025 17:41	YPIAJ	Ok,M
33	AR1248ICV500	PQ071003.D	10 Oct 2025 18:25	YPIAJ	Ok,M
34	AR1254ICV500	PQ071004.D	10 Oct 2025 19:08	YPIAJ	Ok
35	AR1268ICV500	PQ071005.D	10 Oct 2025 19:52	YPIAJ	Ok
36	DDT ANALOG	PQ071006.D	10 Oct 2025 20:36	YPIAJ	Ok

M : Manual Integration

Instrument ID: ECD_Q

Daily Analysis Runlog For Sequence/QC Batch ID # PQ102725

Review By	yogesh	Review On	10/27/2025 10:54:57 AM
Supervise By	mohammad	Supervise On	10/29/2025 5:27:51 AM
SubDirectory	PQ102725	HP Acquire Method	HP Processing Method PQ101025
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854		
CCC	PP24824,PP24828,PP24832,PP24836,PP24840,PP24844,PP24852		
Internal Standard/PEM			
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	HEXANE	PQ071148.D	27 Oct 2025 08:42	YP\AJ	Ok
2	AR1660CCC500	PQ071149.D	27 Oct 2025 08:57	YP\AJ	Ok
3	I.BLK	PQ071150.D	27 Oct 2025 09:11	YP\AJ	Ok
4	Q3450-01	PQ071151.D	27 Oct 2025 09:26	YP\AJ	Ok
5	Q3450-02	PQ071152.D	27 Oct 2025 09:41	YP\AJ	Ok
6	Q3450-02MS	PQ071153.D	27 Oct 2025 09:55	YP\AJ	Ok,M
7	Q3450-02MSD	PQ071154.D	27 Oct 2025 10:10	YP\AJ	Ok,M
8	Q3450-08	PQ071155.D	27 Oct 2025 10:24	YP\AJ	Ok
9	Q3450-09	PQ071156.D	27 Oct 2025 10:39	YP\AJ	Ok,M
10	Q3437-16	PQ071157.D	27 Oct 2025 10:54	YP\AJ	Ok
11	Q3419-01	PQ071158.D	27 Oct 2025 11:08	YP\AJ	Ok,M
12	Q3419-02MS	PQ071159.D	27 Oct 2025 11:23	YP\AJ	Ok
13	Q3419-03MSD	PQ071160.D	27 Oct 2025 11:37	YP\AJ	Ok
14	AR1660CCC500	PQ071161.D	27 Oct 2025 12:35	YP\AJ	Ok
15	AR1242CCC500	PQ071162.D	27 Oct 2025 13:34	YP\AJ	Ok
16	AR1248CCC500	PQ071163.D	27 Oct 2025 13:49	YP\AJ	Ok
17	AR1254CCC500	PQ071164.D	27 Oct 2025 14:03	YP\AJ	Ok
18	I.BLK	PQ071165.D	27 Oct 2025 14:18	YP\AJ	Ok
19	PB170255BL	PQ071166.D	27 Oct 2025 14:32	YP\AJ	Not Ok
20	PB170255BS	PQ071167.D	27 Oct 2025 14:47	YP\AJ	Not Ok
21	Q3451-01	PQ071168.D	27 Oct 2025 15:01	YP\AJ	Ok

Instrument ID: ECD_Q

Daily Analysis Runlog For Sequence/QC Batch ID # PQ102725

Review By	yogesh	Review On	10/27/2025 10:54:57 AM
Supervise By	mohammad	Supervise On	10/29/2025 5:27:51 AM
SubDirectory	PQ102725	HP Acquire Method	HP Processing Method PQ101025
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854		
CCC	PP24824,PP24828,PP24832,PP24836,PP24840,PP24844,PP24852		
Internal Standard/PEM			
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

22	Q3451-05	PQ071169.D	27 Oct 2025 15:16	YP\AJ	Ok
23	Q3451-05MS	PQ071170.D	27 Oct 2025 15:31	YP\AJ	Ok
24	Q3451-05MSD	PQ071171.D	27 Oct 2025 15:45	YP\AJ	Ok
25	Q3452-01	PQ071172.D	27 Oct 2025 16:00	YP\AJ	Ok
26	Q3460-05	PQ071173.D	27 Oct 2025 16:14	YP\AJ	ReRun
27	Q3460-09	PQ071174.D	27 Oct 2025 16:29	YP\AJ	ReRun
28	Q3460-13	PQ071175.D	27 Oct 2025 16:43	YP\AJ	Ok,M
29	AR1660CCC500	PQ071176.D	27 Oct 2025 17:42	YP\AJ	Ok,M
30	AR1242CCC500	PQ071177.D	27 Oct 2025 18:40	YP\AJ	Ok
31	AR1248CCC500	PQ071178.D	27 Oct 2025 18:55	YP\AJ	Ok
32	AR1254CCC500	PQ071179.D	27 Oct 2025 19:09	YP\AJ	Ok,M
33	I.BLK	PQ071180.D	27 Oct 2025 19:24	YP\AJ	Ok

M : Manual Integration

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP102125

Review By	yogesh	Review On	10/21/2025 12:01:15 PM
Supervise By	mohammad	Supervise On	10/25/2025 7:28:37 AM
SubDirectory	PP102125	HP Acquire Method	HP Processing Method PP102125

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854 PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PP075886.D	21 Oct 2025 08:33		YP\AJ	Ok
2	I.BLK	I.BLK	PP075887.D	21 Oct 2025 08:49		YP\AJ	Ok,M
3	AR1660ICC1000	AR1660ICC1000	PP075888.D	21 Oct 2025 09:06		YP\AJ	Ok
4	AR1660ICC750	AR1660ICC750	PP075889.D	21 Oct 2025 09:22		YP\AJ	Ok
5	AR1660ICC500	AR1660ICC500	PP075890.D	21 Oct 2025 09:38		YP\AJ	Ok
6	AR1660ICC250	AR1660ICC250	PP075891.D	21 Oct 2025 09:54		YP\AJ	Ok
7	AR1660ICC050	AR1660ICC050	PP075892.D	21 Oct 2025 10:11		YP\AJ	Ok,M
8	AR1221ICC500	AR1221ICC500	PP075893.D	21 Oct 2025 10:27		YP\AJ	Ok
9	AR1232ICC500	AR1232ICC500	PP075894.D	21 Oct 2025 10:43		YP\AJ	Ok,M
10	AR1242ICC1000	AR1242ICC1000	PP075895.D	21 Oct 2025 10:59		YP\AJ	Ok,M
11	AR1242ICC750	AR1242ICC750	PP075896.D	21 Oct 2025 11:16		YP\AJ	Ok,M
12	AR1242ICC500	AR1242ICC500	PP075897.D	21 Oct 2025 11:48		YP\AJ	Ok,M
13	AR1242ICC250	AR1242ICC250	PP075898.D	21 Oct 2025 12:04		YP\AJ	Ok,M
14	AR1242ICC050	AR1242ICC050	PP075899.D	21 Oct 2025 12:21		YP\AJ	Ok,M
15	AR1248ICC1000	AR1248ICC1000	PP075900.D	21 Oct 2025 12:41		YP\AJ	Ok,M
16	AR1248ICC750	AR1248ICC750	PP075901.D	21 Oct 2025 12:57		YP\AJ	Ok,M
17	AR1248ICC500	AR1248ICC500	PP075902.D	21 Oct 2025 13:13		YP\AJ	Ok,M
18	AR1248ICC250	AR1248ICC250	PP075903.D	21 Oct 2025 13:30		YP\AJ	Ok,M

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP102125

Review By	yogesh	Review On	10/21/2025 12:01:15 PM
Supervise By	mohammad	Supervise On	10/25/2025 7:28:37 AM
SubDirectory	PP102125	HP Acquire Method	HP Processing Method PP102125
STD. NAME	STD REF.#		
Tune/Reschk	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,P PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP 24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854		
Initial Calibration Stds			
CCC			
Internal Standard/PEM			
ICV/I.BLK			
Surrogate Standard			
MS/MSD Standard	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862		
LCS Standard			

19	AR1248ICC050	AR1248ICC050	PP075904.D	21 Oct 2025 13:46		YPIAJ	Ok,M
20	AR1254ICC1000	AR1254ICC1000	PP075905.D	21 Oct 2025 14:02		YPIAJ	Ok,M
21	AR1254ICC750	AR1254ICC750	PP075906.D	21 Oct 2025 14:18		YPIAJ	Ok,M
22	AR1254ICC500	AR1254ICC500	PP075907.D	21 Oct 2025 14:35		YPIAJ	Ok,M
23	AR1254ICC250	AR1254ICC250	PP075908.D	21 Oct 2025 14:51		YPIAJ	Ok,M
24	AR1254ICC050	AR1254ICC050	PP075909.D	21 Oct 2025 15:07		YPIAJ	Ok,M
25	AR1262ICC500	AR1262ICC500	PP075910.D	21 Oct 2025 15:40		YPIAJ	Ok
26	AR1268ICC1000	AR1268ICC1000	PP075911.D	21 Oct 2025 15:56		YPIAJ	Ok
27	AR1268ICC750	AR1268ICC750	PP075912.D	21 Oct 2025 16:12		YPIAJ	Ok
28	AR1268ICC500	AR1268ICC500	PP075913.D	21 Oct 2025 16:29		YPIAJ	Ok
29	AR1268ICC250	AR1268ICC250	PP075914.D	21 Oct 2025 16:45		YPIAJ	Ok
30	AR1268ICC050	AR1268ICC050	PP075915.D	21 Oct 2025 17:01		YPIAJ	Ok
31	PP102125ICV500	ICVPP102125	PP075916.D	21 Oct 2025 17:34		YPIAJ	Ok
32	AR1242ICV500	ICVPP102125AR1242	PP075917.D	21 Oct 2025 18:06		YPIAJ	Ok,M
33	AR1248ICV500	ICVPP102125AR1248	PP075918.D	21 Oct 2025 18:39		YPIAJ	Ok,M
34	AR1254ICV500	ICVPP102125AR1254	PP075919.D	22 Oct 2025 08:49		YPIAJ	Ok,M
35	AR1268ICV500	ICVPP102125AR1268	PP075920.D	22 Oct 2025 09:07		YPIAJ	Ok
36	DDT ANALOG	DDT ANALOG	PP075921.D	22 Oct 2025 09:25		YPIAJ	Ok

M : Manual Integration

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP102425

Review By	yogesh	Review On	10/24/2025 11:09:51 AM
Supervise By	mohammad	Supervise On	10/28/2025 6:10:31 AM
SubDirectory	PP102425	HP Acquire Method	HP Processing Method PP102125

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854
CCC	PP24824,PP24828,PP24832,PP24836,PP24840,PP24844,PP24852
Internal Standard/PEM	
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PP076011.D	24 Oct 2025 08:41		YPIAJ	Ok
2	AR1660CCC500	AR1660CCC500	PP076012.D	24 Oct 2025 08:57		YPIAJ	Ok
3	AR1242CCC500	AR1242CCC500	PP076013.D	24 Oct 2025 09:29		YPIAJ	Ok,M
4	AR1248CCC500	AR1248CCC500	PP076014.D	24 Oct 2025 09:45		YPIAJ	Ok,M
5	AR1254CCC500	AR1254CCC500	PP076015.D	24 Oct 2025 10:02		YPIAJ	Ok,M
6	I.BLK	I.BLK	PP076016.D	24 Oct 2025 10:18		YPIAJ	Ok
7	PB170234BL	PB170234BL	PP076017.D	24 Oct 2025 10:34		YPIAJ	Ok
8	PB170234BS	PB170234BS	PP076018.D	24 Oct 2025 10:51		YPIAJ	Ok
9	Q3437-03	SS-15R(2.0-2.5)	PP076019.D	24 Oct 2025 11:07	AR1254 hits	YPIAJ	Ok
10	Q3437-05	SS-26(0-0.5)	PP076020.D	24 Oct 2025 11:23	AR1254 hits	YPIAJ	Ok,M
11	Q3437-11	SS-27(0-0.5)	PP076021.D	24 Oct 2025 11:39	AR1254 hits	YPIAJ	Ok,M
12	Q3445-01	EO-01-102325	PP076022.D	24 Oct 2025 11:56		YPIAJ	Ok
13	Q3446-01	SB-14	PP076023.D	24 Oct 2025 12:12	F Flag in TCMX in 1st column, All surrogate fail, AR1242 hits, Need 20X	YPIAJ	Dilution
14	Q3446-02	SB-15	PP076024.D	24 Oct 2025 12:28	AR1242 hits, Need 20X	YPIAJ	Dilution
15	Q3414-12DL2	SB-12DL2	PP076025.D	24 Oct 2025 12:44	AR1242 hit	YPIAJ	Ok
16	Q3414-13DL2	SB-13DL2	PP076026.D	24 Oct 2025 13:01	AR1242 hit	YPIAJ	Ok
17	AR1660CCC500	AR1660CCC500	PP076027.D	24 Oct 2025 14:06		YPIAJ	Ok

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP102425

Review By	yogesh	Review On	10/24/2025 11:09:51 AM
Supervise By	mohammad	Supervise On	10/28/2025 6:10:31 AM
SubDirectory	PP102425	HP Acquire Method	HP Processing Method PP102125
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854		
CCC	PP24824,PP24828,PP24832,PP24836,PP24840,PP24844,PP24852		
Internal Standard/PEM			
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

18	AR1242CCC500	AR1242CCC500	PP076028.D	24 Oct 2025 14:22		YPIAJ	Ok,M
19	AR1248CCC500	AR1248CCC500	PP076029.D	24 Oct 2025 14:38		YPIAJ	Ok,M
20	AR1254CCC500	AR1254CCC500	PP076030.D	24 Oct 2025 14:54		YPIAJ	Ok,M
21	I.BLK	I.BLK	PP076031.D	24 Oct 2025 15:11		YPIAJ	Ok
22	Q3414-07DL2	SB-7DL2	PP076032.D	24 Oct 2025 15:27	AR1242 hit	YPIAJ	Ok
23	Q3446-06	SB-19	PP076033.D	24 Oct 2025 15:43	F Flag in TCMX in 1st column , AR1242+1260 Hit , TCMX high in both column , DCB low in 1st column , Need dilution	YPIAJ	Dilution
24	Q3446-07	SB-20	PP076034.D	24 Oct 2025 16:00	AR1260 Hit , Need dilution , TCMX high in 2nd column	YPIAJ	Dilution
25	Q3446-08	SB-21	PP076035.D	24 Oct 2025 16:16	AR1242+1260 Hit , DCB low in both column , TCMX high in both column , need dilution	YPIAJ	Dilution
26	Q3446-03	SB-16	PP076036.D	24 Oct 2025 16:32	AR1260 Hit , TCMX high in 2nd column , need dilution	YPIAJ	Dilution
27	Q3446-04	SB-17	PP076037.D	24 Oct 2025 16:48	AR1260 Hit , TCMX high in 2nd column , need dilution	YPIAJ	Dilution
28	Q3446-01DL	SB-14DL	PP076038.D	24 Oct 2025 17:05	AR1242 hit	YPIAJ	Ok,M
29	Q3446-02DL	SB-15DL	PP076039.D	24 Oct 2025 17:21	AR1242 hit	YPIAJ	Ok,M
30	Q3446-05DL	SB-18DL	PP076040.D	24 Oct 2025 17:37	AR1242 Hit	YPIAJ	Ok,M
31	Q3446-06DL	SB-19DL	PP076041.D	24 Oct 2025 17:53	AR1242+1260 Hit	YPIAJ	Ok,M
32	AR1660CCC500	AR1660CCC500	PP076042.D	24 Oct 2025 18:58		YPIAJ	Ok
33	AR1242CCC500	AR1242CCC500	PP076043.D	24 Oct 2025 19:31		YPIAJ	Ok,M
34	AR1248CCC500	AR1248CCC500	PP076044.D	24 Oct 2025 20:03		YPIAJ	Ok,M

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP102425

Review By	yogesh	Review On	10/24/2025 11:09:51 AM
Supervise By	mohammad	Supervise On	10/28/2025 6:10:31 AM
SubDirectory	PP102425	HP Acquire Method	HP Processing Method PP102125
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854		
CCC	PP24824,PP24828,PP24832,PP24836,PP24840,PP24844,PP24852		
Internal Standard/PEM			
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

35	AR1254CCC500	AR1254CCC500	PP076045.D	24 Oct 2025 20:19		YPIAJ	Ok,M
36	I.BLK	I.BLK	PP076046.D	24 Oct 2025 20:36		YPIAJ	Ok
37	Q3450-10	SS-20-(3.5-4.0)	PP076047.D	24 Oct 2025 20:52	AR1254 Hit	YPIAJ	Ok,M
38	Q3450-11	SS-28-(2.0-2.5)	PP076048.D	24 Oct 2025 21:08		YPIAJ	Ok
39	Q3450-20	SS-21-(0.0-0.5)	PP076049.D	24 Oct 2025 21:24	AR1254 Hit , Need 5x dilution	YPIAJ	Dilution
40	Q3450-21	SS-21-(2.0-2.5)	PP076050.D	24 Oct 2025 21:41		YPIAJ	Ok
41	Q3450-22	SS-21-(3.5-4.0)	PP076051.D	24 Oct 2025 21:57	AR1254 Hit , Need 2x dilution	YPIAJ	Dilution
42	Q3427-01	TP-8	PP076052.D	24 Oct 2025 22:13		YPIAJ	Ok,M
43	Q3446-07DL	SB-20DL	PP076053.D	24 Oct 2025 22:29	AR1242 Hit	YPIAJ	Ok,M
44	Q3446-08DL	SB-21DL	PP076054.D	24 Oct 2025 22:45	AR1242+1260 Hit	YPIAJ	Ok,M
45	Q3446-03DL	SB-16DL	PP076055.D	24 Oct 2025 23:01	AR1260 Hit	YPIAJ	Ok,M
46	Q3446-04DL	SB-17DL	PP076056.D	24 Oct 2025 23:17	AR1260 Hit	YPIAJ	Ok,M
47	AR1660CCC500	AR1660CCC500	PP076057.D	25 Oct 2025 00:22		YPIAJ	Ok
48	AR1242CCC500	AR1242CCC500	PP076058.D	25 Oct 2025 01:11		YPIAJ	Ok,M
49	AR1248CCC500	AR1248CCC500	PP076059.D	25 Oct 2025 01:43		YPIAJ	Ok,M
50	AR1254CCC500	AR1254CCC500	PP076060.D	25 Oct 2025 01:59		YPIAJ	Ok,M
51	I.BLK	I.BLK	PP076061.D	25 Oct 2025 02:16		YPIAJ	Ok

M : Manual Integration

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP102725

Review By	yogesh	Review On	10/27/2025 10:53:48 AM
Supervise By	mohammad	Supervise On	10/29/2025 5:27:46 AM
SubDirectory	PP102725	HP Acquire Method	HP Processing Method PP102125

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854
CCC	PP24824,PP24828,PP24832,PP24836,PP24840,PP24844,PP24852
Internal Standard/PEM	
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PP076062.D	27 Oct 2025 08:40		YPIAJ	Ok
2	AR1660CCC500	AR1660CCC500	PP076063.D	27 Oct 2025 08:56		YPIAJ	Ok,M
3	AR1242CCC500	AR1242CCC500	PP076064.D	27 Oct 2025 09:13		YPIAJ	Ok,M
4	AR1248CCC500	AR1248CCC500	PP076065.D	27 Oct 2025 09:29		YPIAJ	Ok,M
5	AR1254CCC500	AR1254CCC500	PP076066.D	27 Oct 2025 09:45		YPIAJ	Ok,M
6	I.BLK	I.BLK	PP076067.D	27 Oct 2025 10:02		YPIAJ	Ok
7	PB170236BL	PB170236BL	PP076068.D	27 Oct 2025 10:18		YPIAJ	Ok,M
8	PB170236BS	PB170236BS	PP076069.D	27 Oct 2025 10:34		YPIAJ	Ok,M
9	Q3439-01	TP-9	PP076070.D	27 Oct 2025 11:06	DCB,TCX high 1st column	YPIAJ	ReRun
10	Q3449-01	CF-620-COMP-41	PP076071.D	27 Oct 2025 11:26		YPIAJ	Ok
11	Q3450-20DL	SS-21-(0.0-0.5)DL	PP076072.D	27 Oct 2025 11:42	AR1254 Hit	YPIAJ	Ok,M
12	Q3450-22DL	SS-21-(3.5-4.0)DL	PP076073.D	27 Oct 2025 11:58	AR1254 Hit	YPIAJ	Ok,M
13	Q3446-05	SB-18	PP076074.D	27 Oct 2025 12:14	AR1242 Hit, Need Dilution, f flag	YPIAJ	Not Ok
14	Q3447-03	P001-TW2-251023-01	PP076075.D	27 Oct 2025 12:30	F Flag in TCMX in 2nd column , TCMX high in both column , AR1254+1260 Hit	YPIAJ	ReRun
15	Q3447-07	P001-TW4-251023-01	PP076076.D	27 Oct 2025 12:47	TCMX high in 2nd column	YPIAJ	Ok,M
16	Q3446-05	SB-18	PP076077.D	27 Oct 2025 13:03	TCMX high in 2nd column , AR1242 hit , need 100x dilution	YPIAJ	Dilution
17	AR1660CCC500	AR1660CCC500	PP076078.D	27 Oct 2025 14:08		YPIAJ	Ok

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP102725

Review By	yogesh	Review On	10/27/2025 10:53:48 AM
Supervise By	mohammad	Supervise On	10/29/2025 5:27:46 AM
SubDirectory	PP102725	HP Acquire Method	HP Processing Method PP102125
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854		
CCC	PP24824,PP24828,PP24832,PP24836,PP24840,PP24844,PP24852		
Internal Standard/PEM			
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

18	AR1242CCC500	AR1242CCC500	PP076079.D	27 Oct 2025 14:24		YPIAJ	Ok,M
19	AR1248CCC500	AR1248CCC500	PP076080.D	27 Oct 2025 14:40		YPIAJ	Ok,M
20	AR1254CCC500	AR1254CCC500	PP076081.D	27 Oct 2025 14:57		YPIAJ	Ok,M
21	I.BLK	I.BLK	PP076082.D	27 Oct 2025 15:13		YPIAJ	Ok
22	Q3456-01	OILY-DEBRIS	PP076083.D	27 Oct 2025 15:29		YPIAJ	Ok,M
23	Q3457-01	TR-05-102425	PP076084.D	27 Oct 2025 15:45		YPIAJ	Ok,M
24	Q3460-01	WC-1	PP076085.D	27 Oct 2025 16:02		YPIAJ	Ok
25	Q3460-17	WC-5	PP076086.D	27 Oct 2025 16:18		YPIAJ	Ok,M
26	Q3447-05	P001-TW3-251023-01	PP076087.D	27 Oct 2025 16:34	TCMX high in 2nd column	YPIAJ	Ok,M
27	Q3447-01	P001-TW1-251023-01	PP076088.D	27 Oct 2025 16:50	F Flag in TCMX in 2nd column , TCMX high in both column	YPIAJ	ReRun
28	Q3459-01	Y2404-0071-END-1-1	PP076089.D	27 Oct 2025 17:07	AR1254 Hit	YPIAJ	Ok,M
29	Q3459-02	Y2404-0071-END-1-2	PP076090.D	27 Oct 2025 17:23		YPIAJ	Ok,M
30	Q3459-03	Y2404-0071-END-2-1	PP076091.D	27 Oct 2025 17:39	AR1254 Hit	YPIAJ	Ok,M
31	Q3459-04	Y2404-0071-END-2-2	PP076092.D	27 Oct 2025 17:55		YPIAJ	Ok
32	AR1660CCC500	AR1660CCC500	PP076093.D	27 Oct 2025 19:00		YPIAJ	Ok
33	AR1242CCC500	AR1242CCC500	PP076094.D	27 Oct 2025 19:49		YPIAJ	Ok,M
34	AR1248CCC500	AR1248CCC500	PP076095.D	27 Oct 2025 20:05		YPIAJ	Ok,M
35	AR1254CCC500	AR1254CCC500	PP076096.D	27 Oct 2025 20:21		YPIAJ	Ok,M
36	I.BLK	I.BLK	PP076097.D	27 Oct 2025 20:38		YPIAJ	Ok

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP102725

Review By	yogesh	Review On	10/27/2025 10:53:48 AM
Supervise By	mohammad	Supervise On	10/29/2025 5:27:46 AM
SubDirectory	PP102725	HP Acquire Method	HP Processing Method PP102125
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,P P24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP 24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854		
CCC	PP24824,PP24828,PP24832,PP24836,PP24840,PP24844,PP24852		
Internal Standard/PEM			
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

37	PB170272BL	PB170272BL	PP076098.D	27 Oct 2025 20:54		YPIAJ	Ok
38	PB170272BS	PB170272BS	PP076099.D	27 Oct 2025 21:10		YPIAJ	Ok,M
39	Q3459-05	BC259240LOK-END-1-	PP076100.D	27 Oct 2025 21:26		YPIAJ	Ok,M
40	Q3459-06	BC259240LOK-END-1-	PP076101.D	27 Oct 2025 21:42		YPIAJ	Ok,M
41	Q3459-07	BC271327-END-1-1	PP076102.D	27 Oct 2025 21:59		YPIAJ	Ok,M
42	Q3459-08	BC271327-END-1-2	PP076103.D	27 Oct 2025 22:15		YPIAJ	Ok,M
43	AR1660CCC500	AR1660CCC500	PP076104.D	27 Oct 2025 23:20		YPIAJ	Ok
44	AR1242CCC500	AR1242CCC500	PP076105.D	28 Oct 2025 00:08		YPIAJ	Ok,M
45	AR1248CCC500	AR1248CCC500	PP076106.D	28 Oct 2025 00:25		YPIAJ	Ok,M
46	AR1254CCC500	AR1254CCC500	PP076107.D	28 Oct 2025 00:41		YPIAJ	Ok
47	I.BLK	I.BLK	PP076108.D	28 Oct 2025 00:57		YPIAJ	Ok

M : Manual Integration

Instrument ID: ECD_Q

Daily Analysis Runlog For Sequence/QC Batch ID # PQ101025

Review By	yogesh	Review On	10/10/2025 10:11:25 AM
Supervise By	mohammad	Supervise On	10/14/2025 2:48:24 AM
SubDirectory	PQ101025	HP Acquire Method	HP Processing Method PQ101025

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854 PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PQ070971.D	10 Oct 2025 08:27		YPIAJ	Ok
2	I.BLK	I.BLK	PQ070972.D	10 Oct 2025 08:42		YPIAJ	Ok
3	AR1660ICC1000	AR1660ICC1000	PQ070973.D	10 Oct 2025 08:57		YPIAJ	Ok
4	AR1660ICC750	AR1660ICC750	PQ070974.D	10 Oct 2025 09:11		YPIAJ	Ok
5	AR1660ICC500	AR1660ICC500	PQ070975.D	10 Oct 2025 09:26		YPIAJ	Ok
6	AR1660ICC250	AR1660ICC250	PQ070976.D	10 Oct 2025 09:40		YPIAJ	Ok
7	AR1660ICC050	AR1660ICC050	PQ070977.D	10 Oct 2025 09:55		YPIAJ	Ok,M
8	AR1221ICC500	AR1221ICC500	PQ070978.D	10 Oct 2025 10:09		YPIAJ	Ok
9	AR1232ICC500	AR1232ICC500	PQ070979.D	10 Oct 2025 10:24		YPIAJ	Ok
10	AR1242ICC1000	AR1242ICC1000	PQ070980.D	10 Oct 2025 10:38		YPIAJ	Ok
11	AR1242ICC750	AR1242ICC750	PQ070981.D	10 Oct 2025 10:53		YPIAJ	Ok
12	AR1242ICC500	AR1242ICC500	PQ070982.D	10 Oct 2025 11:08		YPIAJ	Ok
13	AR1242ICC250	AR1242ICC250	PQ070983.D	10 Oct 2025 11:22		YPIAJ	Ok
14	AR1242ICC050	AR1242ICC050	PQ070984.D	10 Oct 2025 11:37		YPIAJ	Ok,M
15	AR1248ICC1000	AR1248ICC1000	PQ070985.D	10 Oct 2025 11:51		YPIAJ	Ok
16	AR1248ICC750	AR1248ICC750	PQ070986.D	10 Oct 2025 12:06		YPIAJ	Ok
17	AR1248ICC500	AR1248ICC500	PQ070987.D	10 Oct 2025 12:20		YPIAJ	Ok
18	AR1248ICC250	AR1248ICC250	PQ070988.D	10 Oct 2025 12:35		YPIAJ	Ok,M

Instrument ID: ECD_Q

Daily Analysis Runlog For Sequence/QC Batch ID # PQ101025

Review By	yogesh	Review On	10/10/2025 10:11:25 AM
Supervise By	mohammad	Supervise On	10/14/2025 2:48:24 AM
SubDirectory	PQ101025	HP Acquire Method	HP Processing Method PQ101025
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,P P24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP 24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854 PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862		

19	AR1248ICC050	AR1248ICC050	PQ070989.D	10 Oct 2025 12:50		YPIAJ	Ok
20	AR1254ICC1000	AR1254ICC1000	PQ070990.D	10 Oct 2025 13:04		YPIAJ	Ok
21	AR1254ICC750	AR1254ICC750	PQ070991.D	10 Oct 2025 13:19		YPIAJ	Ok
22	AR1254ICC500	AR1254ICC500	PQ070992.D	10 Oct 2025 13:33		YPIAJ	Ok
23	AR1254ICC250	AR1254ICC250	PQ070993.D	10 Oct 2025 13:48		YPIAJ	Ok
24	AR1254ICC050	AR1254ICC050	PQ070994.D	10 Oct 2025 15:15		YPIAJ	Ok
25	AR1262ICC500	AR1262ICC500	PQ070995.D	10 Oct 2025 15:30		YPIAJ	Ok
26	AR1268ICC1000	AR1268ICC1000	PQ070996.D	10 Oct 2025 15:44		YPIAJ	Ok
27	AR1268ICC750	AR1268ICC750	PQ070997.D	10 Oct 2025 15:59		YPIAJ	Ok
28	AR1268ICC500	AR1268ICC500	PQ070998.D	10 Oct 2025 16:13		YPIAJ	Ok
29	AR1268ICC250	AR1268ICC250	PQ070999.D	10 Oct 2025 16:28		YPIAJ	Ok
30	AR1268ICC050	AR1268ICC050	PQ071000.D	10 Oct 2025 16:43		YPIAJ	Ok
31	PQ101025ICV	ICVPQ101025	PQ071001.D	10 Oct 2025 16:57		YPIAJ	Ok,M
32	AR1242ICV500	ICVPQ101025AR1242	PQ071002.D	10 Oct 2025 17:41		YPIAJ	Ok,M
33	AR1248ICV500	ICVPQ101025AR1248	PQ071003.D	10 Oct 2025 18:25		YPIAJ	Ok,M
34	AR1254ICV500	ICVPQ101025AR1254	PQ071004.D	10 Oct 2025 19:08		YPIAJ	Ok
35	AR1268ICV500	ICVPQ101025AR1268	PQ071005.D	10 Oct 2025 19:52		YPIAJ	Ok
36	DDT ANALOG	DDT ANALOG	PQ071006.D	10 Oct 2025 20:36		YPIAJ	Ok

M : Manual Integration

Instrument ID: ECD_Q

Daily Analysis Runlog For Sequence/QC Batch ID # PQ102725

Review By	yogesh	Review On	10/27/2025 10:54:57 AM
Supervise By	mohammad	Supervise On	10/29/2025 5:27:51 AM
SubDirectory	PQ102725	HP Acquire Method	HP Processing Method PQ101025

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854
CCC	PP24824,PP24828,PP24832,PP24836,PP24840,PP24844,PP24852
Internal Standard/PEM	
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PQ071148.D	27 Oct 2025 08:42		YP\AJ	Ok
2	AR1660CCC500	AR1660CCC500	PQ071149.D	27 Oct 2025 08:57		YP\AJ	Ok
3	I.BLK	I.BLK	PQ071150.D	27 Oct 2025 09:11		YP\AJ	Ok
4	Q3450-01	SS-5R-(2.0-2.5)	PQ071151.D	27 Oct 2025 09:26	AR1254 Hit	YP\AJ	Ok
5	Q3450-02	SS-5R-(3.5-4.0)	PQ071152.D	27 Oct 2025 09:41	AR1254 Hit	YP\AJ	Ok
6	Q3450-02MS	SS-5R-(3.5-4.0)MS	PQ071153.D	27 Oct 2025 09:55		YP\AJ	Ok,M
7	Q3450-02MSD	SS-5R-(3.5-4.0)MSD	PQ071154.D	27 Oct 2025 10:10		YP\AJ	Ok,M
8	Q3450-08	SS-20-(0.0-0.5)	PQ071155.D	27 Oct 2025 10:24	AR1254 Hit	YP\AJ	Ok
9	Q3450-09	SS-20-(2.0-2.5)	PQ071156.D	27 Oct 2025 10:39	AR1254 Hit	YP\AJ	Ok,M
10	Q3437-16	SS-23(0-0.5)	PQ071157.D	27 Oct 2025 10:54	AR1254 Hit, TCX high 1st column	YP\AJ	Ok
11	Q3419-01	PR-132-S16-015094-20	PQ071158.D	27 Oct 2025 11:08	AR1242 hits	YP\AJ	Ok,M
12	Q3419-02MS	PR-132-S16-015094-20	PQ071159.D	27 Oct 2025 11:23		YP\AJ	Ok
13	Q3419-03MSD	PR-132-S16-015094-20	PQ071160.D	27 Oct 2025 11:37		YP\AJ	Ok
14	AR1660CCC500	AR1660CCC500	PQ071161.D	27 Oct 2025 12:35		YP\AJ	Ok
15	AR1242CCC500	AR1242CCC500	PQ071162.D	27 Oct 2025 13:34		YP\AJ	Ok
16	AR1248CCC500	AR1248CCC500	PQ071163.D	27 Oct 2025 13:49		YP\AJ	Ok
17	AR1254CCC500	AR1254CCC500	PQ071164.D	27 Oct 2025 14:03		YP\AJ	Ok

Instrument ID: ECD_Q

Daily Analysis Runlog For Sequence/QC Batch ID # PQ102725

Review By	yogesh	Review On	10/27/2025 10:54:57 AM
Supervise By	mohammad	Supervise On	10/29/2025 5:27:51 AM
SubDirectory	PQ102725	HP Acquire Method	HP Processing Method PQ101025
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854		
CCC	PP24824,PP24828,PP24832,PP24836,PP24840,PP24844,PP24852		
Internal Standard/PEM			
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

18	I.BLK	I.BLK	PQ071165.D	27 Oct 2025 14:18		YPIAJ	Ok
19	PB170255BL	PB170255BL	PQ071166.D	27 Oct 2025 14:32	DCB high 1st column for PSEG	YPIAJ	Not Ok
20	PB170255BS	PB170255BS	PQ071167.D	27 Oct 2025 14:47	DCB high 1st column for PSEG	YPIAJ	Not Ok
21	Q3451-01	TP-11	PQ071168.D	27 Oct 2025 15:01	Tcmx high in 1st column	YPIAJ	Ok
22	Q3451-05	TP-12	PQ071169.D	27 Oct 2025 15:16		YPIAJ	Ok
23	Q3451-05MS	TP-12MS	PQ071170.D	27 Oct 2025 15:31	Tcmx high in 1st column	YPIAJ	Ok
24	Q3451-05MSD	TP-12MSD	PQ071171.D	27 Oct 2025 15:45	Tcmx high in 1st column	YPIAJ	Ok
25	Q3452-01	JB-1	PQ071172.D	27 Oct 2025 16:00		YPIAJ	Ok
26	Q3460-05	WC-2	PQ071173.D	27 Oct 2025 16:14	TCMX & DCB high in 1st column	YPIAJ	ReRun
27	Q3460-09	WC-3	PQ071174.D	27 Oct 2025 16:29	TCMX & DCB high in 1st column	YPIAJ	ReRun
28	Q3460-13	WC-4	PQ071175.D	27 Oct 2025 16:43	DCB high in 1st column, AR1260 hits	YPIAJ	Ok,M
29	AR1660CCC500	AR1660CCC500	PQ071176.D	27 Oct 2025 17:42		YPIAJ	Ok,M
30	AR1242CCC500	AR1242CCC500	PQ071177.D	27 Oct 2025 18:40		YPIAJ	Ok
31	AR1248CCC500	AR1248CCC500	PQ071178.D	27 Oct 2025 18:55		YPIAJ	Ok
32	AR1254CCC500	AR1254CCC500	PQ071179.D	27 Oct 2025 19:09		YPIAJ	Ok,M
33	I.BLK	I.BLK	PQ071180.D	27 Oct 2025 19:24		YPIAJ	Ok

M : Manual Integration



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 10/27/2025

OVENTEMP IN Celsius(°C): 107
Time IN: 17:25
In Date: 10/24/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 104
Time OUT: 08:27
Out Date: 10/25/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID-OVEN

QC:LB137635

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
Q3449-01	CF-620-COMP-41	1	1.14	10.13	11.27	10.24	89.8	
Q3449-02	CF-620-VOC-41	2	1.15	10.74	11.89	11.28	94.3	
Q3449-03	CF-620-121	3	1.16	10.81	11.97	11.05	91.5	
Q3449-04	CF-620-122	4	1.14	10.40	11.54	10.35	88.6	
Q3449-05	CF-620-123	5	1.19	10.37	11.56	9.88	83.8	
Q3450-01	SS-5R-(2.0-2.5)	6	1.19	10.00	11.19	10.48	92.9	
Q3450-02	SS-5R-(3.5-4.0)	7	1.15	10.95	12.1	10.6	86.3	
Q3450-03	SS-10R-(2.0-2.5)	8	1.18	10.53	11.71	11.00	93.3	
Q3450-04	SS-10R-(3.5-4.0)	9	1.12	10.86	11.98	10.52	86.6	
Q3450-05	SS-13R-(3.5-4.0)	10	1.15	10.58	11.73	10.88	92.0	
Q3450-06	SS-13R-(5.0-5.5)	11	1.17	10.15	11.32	9.71	84.1	
Q3450-07	SS-13R-(5.0-5.5)	12	1.13	10.68	11.81	10.24	85.3	
Q3450-08	SS-20-(0.0-0.5)	13	1.15	9.72	10.87	10.24	93.5	
Q3450-09	SS-20-(2.0-2.5)	14	1.14	10.79	11.93	10.45	86.3	
Q3450-10	SS-20-(3.5-4.0)	15	1.17	10.15	11.32	9.7	84.0	
Q3450-11	SS-28-(2.0-2.5)	16	1.18	10.44	11.62	10.89	93.0	
Q3450-12	SS-29-(20-2.5)	17	1.13	11.27	12.4	10.24	80.8	
Q3450-13	SS-30-(0.0-0.5)	18	1.18	10.14	11.32	10.42	91.1	
Q3450-14	SS-30-(2.0-2.5)	19	1.18	10.12	11.3	10.04	87.5	
Q3450-15	SS-30-(3.5-4.0)	20	1.15	10.14	11.29	9.74	84.7	
Q3450-16	SS-31-(0.0-0.5)	21	1.17	11.01	12.18	10.93	88.6	
Q3450-17	SS-34-(0.0-0.5)	22	1.11	10.41	11.52	10.14	86.7	
Q3450-18	SS-34-(2.0-2.5)	23	1.18	10.67	11.85	11.00	92.0	
Q3450-19	SS-34-(3.5-4.0)	24	1.14	10.04	11.18	10.79	96.1	
Q3450-20	SS-21-(0.0-0.5)	25	1.14	10.28	11.42	10.57	91.7	
Q3450-21	SS-21-(2.0-2.5)	26	1.14	9.99	11.13	10.45	93.2	
Q3450-22	SS-21-(3.5-4.0)	27	1.18	10.56	11.74	10.12	84.7	
Q3450-23	SS-22-(0.0-0.5)	28	1.18	10.29	11.47	10.59	91.4	

PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 10/27/2025

OVENTEMP IN Celsius(°C): 107
Time IN: 17:25
In Date: 10/24/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 104
Time OUT: 08:27
Out Date: 10/25/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID-OVEN

QC:LB137635

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
Q3450-24	SS-22-(2.0-2.5)	29	1.14	10.09	11.23	10.32	91.0	
Q3450-25	SS-22-(3.5-4.0)	30	1.13	10.66	11.79	9.95	82.7	
Q3451-01	TP-11	31	1.18	10.88	12.06	10.95	89.8	
Q3451-02	TP-11-EPH	32	1.13	10.40	11.53	10.07	86.0	
Q3451-03	TP-11-VOC	33	1.16	10.54	11.7	10.99	93.3	
Q3451-05	TP-12	34	1.19	10.28	11.47	10.35	89.1	
Q3451-06	TP-12-EPH	35	1.15	11.08	12.23	10.62	85.5	
Q3451-07	TP-12-VOC	36	1.17	10.05	11.22	10.24	90.2	
Q3452-01	JB-1	37	1.12	10.26	11.38	8.46	71.5	
Q3452-02	JB-1-EPH	38	1.15	10.80	11.95	9.15	74.1	
Q3452-03	JB-1-VOC	39	1.16	10.37	11.53	10.32	88.3	
Q3453-01	102225-A	40	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
Q3453-02	102225-B	41	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
Q3453-03	102325-A	42	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
Q3453-04	102325-B	43	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
Q3456-01	OILY-DEBRIS	50	1.00	1.00	2.00	2.00	100.0	oily-debris
Q3457-01	TR-05-102425	51	1.14	10.12	11.26	10.47	92.2	
Q3457-02	TR-05-102425-E2	52	1.16	10.37	11.53	10.88	93.7	
Q3458-01	TRE-25-0116	53	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
Q3460-01	WC-1	54	1.18	10.42	11.6	10.9	93.3	
Q3460-02	WC-1-EPH	55	1.19	10.51	11.7	11.00	93.3	
Q3460-03	WC-1-VOC	56	1.19	10.04	11.23	10.52	92.9	
Q3460-05	WC-2	57	1.14	10.24	11.38	11.04	96.7	
Q3460-06	WC-2-EPH	58	1.18	10.48	11.66	11.3	96.6	
Q3460-07	WC-2-VOC	59	1.13	9.83	10.96	10.6	96.3	
Q3460-09	WC-3	60	1.19	10.78	11.97	11.36	94.3	
Q3460-10	WC-3-EPH	61	1.11	10.30	11.41	10.81	94.2	
Q3460-11	WC-3-VOC	62	1.16	10.12	11.28	10.74	94.7	



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 10/27/2025

OVENTEMP IN Celsius(°C): 107
Time IN: 17:25
In Date: 10/24/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 104
Time OUT: 08:27
Out Date: 10/25/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID-OVEN

QC:LB137635

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
Q3460-13	WC-4	63	1.13	10.80	11.93	11.41	95.2	
Q3460-14	WC-4-EPH	64	1.18	10.43	11.61	11.02	94.3	
Q3460-15	WC-4-VOC	65	1.18	9.96	11.14	10.6	94.6	
Q3460-17	WC-5	66	1.17	10.38	11.55	10.99	94.6	
Q3460-18	WC-5-EPH	67	1.19	10.50	11.69	11.14	94.8	
Q3460-19	WC-5-VOC	68	1.16	10.44	11.6	11.09	95.1	
Q3463-05	SVOC-GPC-BLANK	44	1.00	1.00	2.00	2.00	100.0	
Q3463-06	PEST-GPC-BLANK	45	1.00	1.00	2.00	2.00	100.0	
Q3463-07	PEST-GPC-BLANK-SPIKE	46	1.00	1.00	2.00	2.00	100.0	
Q3463-08	SVOC-GPC2-BLANK	47	1.00	1.00	2.00	2.00	100.0	
Q3463-09	PEST-GPC2-BLANK	48	1.00	1.00	2.00	2.00	100.0	
Q3463-10	PEST -GPC2-BLANK-SPIKE	49	1.00	1.00	2.00	2.00	100.0	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$

WORKLIST(Hardcopy Internal Chain)

137635

WorkList Name : %1-102425

WorkList ID : 192653

Department : Wet-Chemistry

Date : 10-24-2025 08:07:25

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q3449-01	CF-620-COMP-41	Solid	Percent Solids	Cool 4 deg C	PSEG03	D21	10/23/2025	Chemtech -SO
Q3449-02	CF-620-VOC-41	Solid	Percent Solids	Cool 4 deg C	PSEG03	D21	10/23/2025	Chemtech -SO
Q3449-03	CF-620-121	Solid	Percent Solids	Cool 4 deg C	PSEG03	D21	10/23/2025	Chemtech -SO
Q3449-04	CF-620-122	Solid	Percent Solids	Cool 4 deg C	PSEG03	D21	10/23/2025	Chemtech -SO
Q3449-05	CF-620-123	Solid	Percent Solids	Cool 4 deg C	PSEG03	D21	10/23/2025	Chemtech -SO
Q3450-01	SS-5R-(2.0-2.5)	Solid	Percent Solids	Cool 4 deg C	MATR02	D41	10/23/2025	Chemtech -SO
Q3450-02	SS-5R-(3.5-4.0)	Solid	Percent Solids	Cool 4 deg C	MATR02	D41	10/23/2025	Chemtech -SO
Q3450-03	SS-10R-(2.0-2.5)	Solid	Percent Solids	Cool 4 deg C	MATR02	D41	10/23/2025	Chemtech -SO
Q3450-04	SS-10R-(3.5-4.0)	Solid	Percent Solids	Cool 4 deg C	MATR02	D41	10/23/2025	Chemtech -SO
Q3450-05	SS-13R-(3.5-4.0)	Solid	Percent Solids	Cool 4 deg C	MATR02	D41	10/23/2025	Chemtech -SO
Q3450-06	SS-13R-(5.0-5.5)	Solid	Percent Solids	Cool 4 deg C	MATR02	D41	10/23/2025	Chemtech -SO
Q3450-07	SS-13R-(5.0-5.5)	Solid	Percent Solids	Cool 4 deg C	MATR02	D41	10/23/2025	Chemtech -SO
Q3450-08	SS-20-(0.0-0.5)	Solid	Percent Solids	Cool 4 deg C	MATR02	D41	10/23/2025	Chemtech -SO
Q3450-09	SS-20-(2.0-2.5)	Solid	Percent Solids	Cool 4 deg C	MATR02	D41	10/23/2025	Chemtech -SO
Q3450-10	SS-20-(3.5-4.0)	Solid	Percent Solids	Cool 4 deg C	MATR02	D41	10/23/2025	Chemtech -SO
Q3450-11	SS-28-(2.0-2.5)	Solid	Percent Solids	Cool 4 deg C	MATR02	D41	10/23/2025	Chemtech -SO
Q3450-12	SS-29-(20-2.5)	Solid	Percent Solids	Cool 4 deg C	MATR02	D41	10/23/2025	Chemtech -SO
Q3450-13	SS-30-(0.0-0.5)	Solid	Percent Solids	Cool 4 deg C	MATR02	D41	10/23/2025	Chemtech -SO
Q3450-14	SS-30-(2.0-2.5)	Solid	Percent Solids	Cool 4 deg C	MATR02	D41	10/23/2025	Chemtech -SO
Q3450-15	SS-30-(3.5-4.0)	Solid	Percent Solids	Cool 4 deg C	MATR02	D41	10/23/2025	Chemtech -SO
Q3450-16	SS-31-(0.0-0.5)	Solid	Percent Solids	Cool 4 deg C	MATR02	D41	10/23/2025	Chemtech -SO

Date/Time 10/24/25 15:00

Raw Sample Received by: [Signature]

Raw Sample Relinquished by: [Signature]

Date/Time 10/24/25

Raw Sample Received by: [Signature]

Raw Sample Relinquished by: [Signature]

WORKLIST(Hardcopy Internal Chain)

137135

WorkList Name : %1-102425 WorkList ID : 192653 Department : Wet-Chemistry Date : 10-24-2025 08:07:25

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q3450-17	SS-34-(0.0-0.5)	Solid	Percent Solids	Cool 4 deg C	MATR02	D41	10/23/2025	Chemtech -SO
Q3450-18	SS-34-(2.0-2.5)	Solid	Percent Solids	Cool 4 deg C	MATR02	D41	10/23/2025	Chemtech -SO
Q3450-19	SS-34-(3.5-4.0)	Solid	Percent Solids	Cool 4 deg C	MATR02	D41	10/23/2025	Chemtech -SO
Q3450-20	SS-21-(0.0-0.5)	Solid	Percent Solids	Cool 4 deg C	MATR02	D41	10/23/2025	Chemtech -SO
Q3450-21	SS-21-(2.0-2.5)	Solid	Percent Solids	Cool 4 deg C	MATR02	D41	10/23/2025	Chemtech -SO
Q3450-22	SS-21-(3.5-4.0)	Solid	Percent Solids	Cool 4 deg C	MATR02	D41	10/23/2025	Chemtech -SO
Q3450-23	SS-22-(0.0-0.5)	Solid	Percent Solids	Cool 4 deg C	MATR02	D41	10/23/2025	Chemtech -SO
Q3450-24	SS-22-(2.0-2.5)	Solid	Percent Solids	Cool 4 deg C	MATR02	D41	10/23/2025	Chemtech -SO
Q3450-25	SS-22-(3.5-4.0)	Solid	Percent Solids	Cool 4 deg C	MATR02	D41	10/23/2025	Chemtech -SO
Q3451-01	TP-11	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	10/23/2025	Chemtech -SO
Q3451-02	TP-11-EPH	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	10/23/2025	Chemtech -SO
Q3451-03	TP-11-VOC	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	10/23/2025	Chemtech -SO
Q3451-05	TP-12	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	10/23/2025	Chemtech -SO
Q3451-06	TP-12-EPH	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	10/23/2025	Chemtech -SO
Q3451-07	TP-12-VOC	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	10/23/2025	Chemtech -SO
Q3452-01	JB-1	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	10/23/2025	Chemtech -SO
Q3452-02	JB-1-EPH	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	10/24/2025	Chemtech -SO
Q3452-03	JB-1-VOC	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	10/24/2025	Chemtech -SO
Q3453-01	102225-A	Solid	Percent Solids	Cool 4 deg C	PSEG03	D41	10/24/2025	Chemtech -SO
Q3453-02	102225-B	Solid	Percent Solids	Cool 4 deg C	PSEG03	D41	10/24/2025	Chemtech -SO
Q3453-03	102325-A	Solid	Percent Solids	Cool 4 deg C	PSEG03	D41	10/24/2025	Chemtech -SO

Date/Time: 10/24/25 15:00
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]
 Date/Time: 10/24/25 17:35
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]

WORKLIST(Hardcopy Internal Chain)

8137435

WorkList Name : %1-102425 WorkList ID : 192653 Department : Wet-Chemistry Date : 10-24-2025 08:07:25

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q3453-04	102325-B	Solid	Percent Solids	Cool 4 deg C	PSEG03	D41	10/24/2025	Chemtech -SO
Q3456-01	OILY-DEBRIS	Solid	Percent Solids	Cool 4 deg C	PSEG03	D41	10/24/2025	Chemtech -SO
Q3457-01	TR-05-102425	Solid	Percent Solids	Cool 4 deg C	PSEG05	D31	10/24/2025	Chemtech -SO
Q3457-02	TR-05-102425-E2	Solid	Percent Solids	Cool 4 deg C	PSEG05	D31	10/24/2025	Chemtech -SO
Q3458-01	TRE-25-0116	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	10/24/2025	Chemtech -SO
Q3460-01	WC-1	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	10/24/2025	Chemtech -SO
Q3460-02	WC-1-EPH	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	10/24/2025	Chemtech -SO
Q3460-03	WC-1-VOC	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	10/24/2025	Chemtech -SO
Q3460-05	WC-2	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	10/24/2025	Chemtech -SO
Q3460-06	WC-2-EPH	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	10/24/2025	Chemtech -SO
Q3460-07	WC-2-VOC	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	10/24/2025	Chemtech -SO
Q3460-09	WC-3	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	10/24/2025	Chemtech -SO
Q3460-10	WC-3-EPH	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	10/24/2025	Chemtech -SO
Q3460-11	WC-3-VOC	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	10/24/2025	Chemtech -SO
Q3460-13	WC-4	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	10/24/2025	Chemtech -SO
Q3460-14	WC-4-EPH	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	10/24/2025	Chemtech -SO
Q3460-15	WC-4-VOC	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	10/24/2025	Chemtech -SO
Q3460-17	WC-5	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	10/24/2025	Chemtech -SO
Q3460-18	WC-5-EPH	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	10/24/2025	Chemtech -SO
Q3460-19	WC-5-VOC	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	10/24/2025	Chemtech -SO
Q3463-05	SVOC-GPC-BLANK	Solid	Percent Solids	Cool 4 deg C	CHEM02	J32	10/17/2025	Chemtech -SO

Date/Time 10/24/25 15:00
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]
 Date/Time 10/24/25 17:35
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]

WORKLIST(Hardcopy Internal Chain)

137435

WorkList Name : %1-102425 WorkList ID : 192653 Department : Wet-Chemistry Date : 10-24-2025 08:07:25

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q3463-06	PEST-GPC-BLANK	Solid	Percent Solids	Cool 4 deg C	CHEM02	J32	10/17/2025	Chemtech -SO
Q3463-07	PEST-GPC-BLANK-SPIKE	Solid	Percent Solids	Cool 4 deg C	CHEM02	J32	10/17/2025	Chemtech -SO
Q3463-08	SVOC-GPC2-BLANK	Solid	Percent Solids	Cool 4 deg C	CHEM02	J32	10/17/2025	Chemtech -SO
Q3463-09	PEST-GPC2-BLANK	Solid	Percent Solids	Cool 4 deg C	CHEM02	J32	10/17/2025	Chemtech -SO
Q3463-10	PEST -GPC2-BLANK-SPIKE	Solid	Percent Solids	Cool 4 deg C	CHEM02	J32	10/17/2025	Chemtech -SO

Date/Time 10/24/25 13:00
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]

Date/Time 10/24/25 14:35
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]

SOP ID: M3541-ASE Extraction-15

Clean Up SOP #: Acid Cleanup

Extraction Start Date : 10/24/2025

Matrix : Solid

Extraction Start Time : 12:05

Weigh By: EH

Extraction By: RJ

Extraction End Date : 10/24/2025

Balance check: EH

Filter By: RJ

Extraction End Time : 15:25

Balance ID: EX-SC-2

pH Meter ID: N/A

Concentration By: EH

pH Strip Lot#: N/A

Hood ID: 3,7

Supervisor By : RUPESH

Extraction Method: ☐ Separatory Funnel ☐ Continuous Liquid/Liquid ☐ Sonication ☐ Waste Dilution ☒ Soxhlet

Standard Name	MLS USED	Concentration ug/mL	STD REF. # FROM LOG
Spike Sol 1	1.0ML	5000 PPB	PP24930
Surrogate	1.0ML	200 PPB	PP24663
N/A	N/A	N/A	N/A
N/A	N/A	N/A	N/A
N/A	N/A	N/A	N/A

Chemical Used	ML/SAMPLE USED	Lot Number
Hexane/Acetone/1:1	N/A	EP2643
Baked Na2SO4	N/A	EP2652
Sand	N/A	E3951
Hexane	N/A	E3978
H2SO4 1:1	N/A	EP2610
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

Extraction Conformance/Non-Conformance Comments:

40ML Vial Lot # 03-40BTS721. Q3447-07 used Limited volume as sample is tar matrix.

KD Bath ID: N/A

Envap ID: N-EVAP-02

KD Bath Temperature: N/A

Envap Temperature: 40 °C

Date / Time	Prepped Sample Relinquished By/Location	Received By/Location
10/24/25	RS (Ext Lab)	R. Pesh...
15:30	Preparation Group	Analysis Group

Analytical Method: M3541-ASE Extraction-15

Concentration Date: 10/24/2025

Sample ID	Client Sample ID	Test	g/mL	PH	Surr/Spike By:		Final Vol. (mL)	JarID	Comments	Prep Pos
					AddedBy	VerifiedBy				
PB170236BL	ABLK236	PCB	30.02	N/A	ritesh	Evelyn	10			U1-1
PB170236BS	ALCS236	PCB	30.01	N/A	ritesh	Evelyn	10			2
Q3427-01	TP-8	PCB	30.06	N/A	ritesh	Evelyn	10	E	Oily Mat.	3
Q3447-01	P001-TW1-251023-01	PCB	30.07	N/A	ritesh	Evelyn	10	E	Oily Mat.	4
Q3447-03	P001-TW2-251023-01	PCB	30.04	N/A	ritesh	Evelyn	10	E	Tar Mat.	5
Q3447-05	P001-TW3-251023-01	PCB	30.09	N/A	ritesh	Evelyn	10	E	Tar Mat.	6
Q3447-07	P001-TW4-251023-01	PCB	10.05	N/A	ritesh	Evelyn	10	E		U2-1
Q3449-01	CF-620-COMP-41	PCB	30.02	N/A	ritesh	Evelyn	10	E		2
Q3450-01	SS-5R-(2.0-2.5)	PCB	30.03	N/A	ritesh	Evelyn	10			3
Q3450-02	SS-5R-(3.5-4.0)	PCB	30.08	N/A	ritesh	Evelyn	10			4
Q3450-02MS	SS-5R-(3.5-4.0)MS	PCB	30.06	N/A	ritesh	Evelyn	10			5
Q3450-02MS D	SS-5R-(3.5-4.0)MSD	PCB	30.04	N/A	ritesh	Evelyn	10			6
Q3450-08	SS-20-(0.0-0.5)	PCB	30.01	N/A	ritesh	Evelyn	10			U3-1
Q3450-09	SS-20-(2.0-2.5)	PCB	30.09	N/A	ritesh	Evelyn	10			2
Q3450-10	SS-20-(3.5-4.0)	PCB	30.07	N/A	ritesh	Evelyn	10			3
Q3450-11	SS-28-(2.0-2.5)	PCB	30.02	N/A	ritesh	Evelyn	10			4
Q3450-20	SS-21-(0.0-0.5)	PCB	30.05	N/A	ritesh	Evelyn	10			5
Q3450-21	SS-21-(2.0-2.5)	PCB	30.08	N/A	ritesh	Evelyn	10			6
Q3450-22	SS-21-(3.5-4.0)	PCB	30.04	N/A	ritesh	Evelyn	10			U6-1

* Extracts relinquished on the same date as received.

RS
10/24

10/23/25 10:10:15

WORKLIST(Hardcopy Internal Chain)

WorkList Name : Q3450 WorkList ID : 192676 Department : Extraction Date : 10-24-2025 10:40:05

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q3427-01	TP-8	Solid	PCB	Cool 4 deg C	PSEG03	D41	10/21/2025	8082A
Q3447-01	P001-TW1-251023-01	Solid	PCB	Cool 4 deg C	ROYF02	D41	10/23/2025	8082A
Q3447-03	P001-TW2-251023-01	Solid	PCB	Cool 4 deg C	ROYF02	D41	10/23/2025	8082A
Q3447-05	P001-TW3-251023-01	Solid	PCB	Cool 4 deg C	ROYF02	D41	10/23/2025	8082A
Q3447-07	P001-TW4-251023-01	Solid	PCB	Cool 4 deg C	ROYF02	D41	10/23/2025	8082A
Q3449-01	CF-620-COMP-41	Solid	PCB	Cool 4 deg C	PSEG03	D21	10/23/2025	8082A
Q3450-01	SS-5R-(2.0-2.5)	Solid	PCB	Cool 4 deg C	MATR02	D41	10/23/2025	8082A
Q3450-02	SS-5R-(3.5-4.0)	Solid	PCB	Cool 4 deg C	MATR02	D41	10/23/2025	8082A
Q3450-08	SS-20-(0.0-0.5)	Solid	PCB	Cool 4 deg C	MATR02	D41	10/23/2025	8082A
Q3450-09	SS-20-(2.0-2.5)	Solid	PCB	Cool 4 deg C	MATR02	D41	10/23/2025	8082A
Q3450-10	SS-20-(3.5-4.0)	Solid	PCB	Cool 4 deg C	MATR02	D41	10/23/2025	8082A
Q3450-11	SS-28-(2.0-2.5)	Solid	PCB	Cool 4 deg C	MATR02	D41	10/23/2025	8082A
Q3450-20	SS-21-(0.0-0.5)	Solid	PCB	Cool 4 deg C	MATR02	D41	10/23/2025	8082A
Q3450-21	SS-21-(2.0-2.5)	Solid	PCB	Cool 4 deg C	MATR02	D41	10/23/2025	8082A
Q3450-22	SS-21-(3.5-4.0)	Solid	PCB	Cool 4 deg C	MATR02	D41	10/23/2025	8082A

Date/Time 10/24/25 10:10
Raw Sample Received by: RJ (Ext Lab)
Raw Sample Relinquished by: [Signature]

Date/Time 10/24/25 11:10
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: RJ (Ext Lab)

SOP ID:	M3541-ASE Extraction-15		
Clean Up SOP #:	Acid Cleanup	Extraction Start Date :	10/31/2025
Matrix :	Solid	Extraction Start Time :	12:10
Weight By:	EH	Extraction By:	RJ
Balance check:	EH	Filter By:	RJ
Balance ID:	EX-SC-2	pH Meter ID:	N/A
pH Strip Lot#:	N/A	Hood ID:	3,7
Extraction Method:	☐ Separatory Funnel	☐ Continuous Liquid/Liquid	☐ Sonication
		☐ Waste Dilution	☒ Soxhlet

Standard Name	MLS USED	Concentration ug/mL	STD REF. # FROM LOG
Spike Sol 1	1.0ML	5000 PPB	PP24930
Surrogate	1.0ML	200 PPB	PP24663
N/A	N/A	N/A	N/A
N/A	N/A	N/A	N/A
N/A	N/A	N/A	N/A

Chemical Used	ML/SAMPLE USED	Lot Number
Hexane/Acetone/1:1	N/A	EP2643
Baked Na2SO4	N/A	EP2655
Hexane	N/A	E3981
H2SO4 1:1	N/A	EP2610
Sand	N/A	E3951
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

Extraction Conformance/Non-Conformance Comments:

40ML Vial Lot # 03-40BTS721. Q3502-04 used Limited volume as sample is oil.


KD Bath ID:	N/A	Envap ID:	N-EVAP-02
KD Bath Temperature:	N/A	Envap Temperature:	40 °C

Date / Time	Prepped Sample Relinquished By/Location	Received By/Location
10/31/25	RS (Ext Lab)	Y-P. Pest/PCB.
15:15	Preparation Group	Analysis Group

Analytical Method: M3541-ASE Extraction-15

Concentration Date: 10/31/2025

Sample ID	Client Sample ID	Test	g/ mL	PH	Surr/ Spike By:		Final Vol. (mL)	JarID	Comments	Prep Pos
					AddedBy	VerifiedBy				
PB170358BL	ABLK358	PCB	30.01	N/A	ritesh	Evelyn	10			U2-1
PB170358BS	ALCS358	PCB	30.03	N/A	ritesh	Evelyn	10			2
Q3437-15	SS-32(0-0.5)	PCB	30.06	N/A	ritesh	Evelyn	10			3
Q3450-04	SS-10R-(3.5-4.0)	PCB	30.09	N/A	ritesh	Evelyn	10			4
Q3450-04MS	SS-10R-(3.5-4.0)MS	PCB	30.02	N/A	ritesh	Evelyn	10			5
Q3450-04MS D	SS-10R-(3.5-4.0)MSD	PCB	30.04	N/A	ritesh	Evelyn	10			6
Q3450-05	SS-13R-(3.5-4.0)	PCB	30.07	N/A	ritesh	Evelyn	10			U3-1
Q3450-06	SS-13R-(5.0-5.5)	PCB	30.03	N/A	ritesh	Evelyn	10			2
Q3450-23	SS-22-(0.0-0.5)	PCB	30.05	N/A	ritesh	Evelyn	10			3
Q3450-25	SS-22-(3.5-4.0)	PCB	30.04	N/A	ritesh	Evelyn	10			4
Q3502-04	PARTS-CLEANER	PCB	1.09	N/A	ritesh	Evelyn	10		Oil	



170358
12:10

WORKLIST(Hardcopy Internal Chain)

WorkList Name : Q3437P WorkList ID : 192830 Department : Extraction Date : 10-31-2025 12:04:12

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q3437-15	SS-32(0-0.5)	Solid	PCB	Cool 4 deg C	MATR02	D31	10/22/2025	8082A
Q3450-04	SS-10R-(3.5-4.0)	Solid	PCB	Cool 4 deg C	MATR02	D41	10/23/2025	8082A
Q3450-05	SS-13R-(3.5-4.0)	Solid	PCB	Cool 4 deg C	MATR02	D41	10/23/2025	8082A
Q3450-06	SS-13R-(5.0-5.5)	Solid	PCB	Cool 4 deg C	MATR02	D41	10/23/2025	8082A
Q3450-23	SS-22-(0.0-0.5)	Solid	PCB	Cool 4 deg C	MATR02	D41	10/23/2025	8082A
Q3450-25	SS-22-(3.5-4.0)	Solid	PCB	Cool 4 deg C	MATR02	D41	10/23/2025	8082A
Q3502-04	PARTS-CLEANER	Solid	PCB	Cool 4 deg C	PSEG03	D41	10/30/2025	8082A

Date/Time 10/31/25 12:05
Raw Sample Received by: RJC(Ext-66b)
Raw Sample Relinquished by: [Signature]

Date/Time 10/31/25 12:20
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: RJC(Ext-66b)

Prep Standard - Chemical Standard Summary

Order ID : Q3450

Test : PCB

Prepbatch ID : PB170236,

Sequence ID/Qc Batch ID: pp102425,PP102725,PQ102725,

Standard ID :

EP2610,EP2643,EP2652,PP24329,PP24330,PP24381,PP24663,PP24805,PP24806,PP24807,PP24808,PP24809,PP24810,PP24811,PP24812,PP24814,PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862,PP24930,

Chemical ID :

E3804,E3875,E3944,E3951,E3962,E3963,E3971,E3972,E3978,M6157,P11522,P11592,P12700,P12703,P12952,P12953,P12957,P13356,P13373,P13375,P13590,P13592,P13695,P13697,P13706,P13786,P13789,P13833,P13879,P13881,P13885,W3112,W3177,W3234,



<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
314	1.1 H2SO4 SOLN	EP2610	05/07/2025	11/07/2025	RUPESHKUMAR SHAH	Extraction_SC ALE_2 (EX-SC-2)	None	Riteshkumar Patel 05/07/2025
<u>FROM</u>	1000.00000ml of M6157 + 1000.00000ml of W3112 = Final Quantity: 2000.000 ml							

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
230	1:1ACETONE/HEXANE	EP2643	09/16/2025	03/16/2026	Evelyn Huang	None	None	Riteshkumar Patel 09/16/2025
<u>FROM</u> 4000.00000ml of E3971 + 4000.00000ml of E3972 = Final Quantity: 8000.000 ml								



<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3923	Baked Sodium Sulfate	EP2652	10/10/2025	01/28/2026	Evelyn Huang	Extraction_SC ALE_2 (EX-SC-2)	None	Riteshkumar Patel 10/10/2025
<u>FROM</u> 4000.00000gram of E3875 = Final Quantity: 4000.000 gram								

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
84	Pest/PCB Surrogate Stock 20 PPM	PP24329	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza 04/03/2025
<u>FROM</u> 1.00000ml of P13356 + 9.00000ml of W3177 = Final Quantity: 10.000 ml								

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
202	AR1660 1000/100 ppb working solution 1st source	PP24330	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.10000ml of P13697 + 99.40000ml of W3177 + 0.50000ml of PP24329 = Final Quantity: 100.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1893	AR1254 1000 PPB Working Sol. 2nd Source	PP24381	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 1.00000ml of P12957 + 98.50000ml of W3177 + 0.50000ml of PP24329 = Final Quantity: 100.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
465	200 PPB Pest/PCB Surrogate Spike	PP24663	06/24/2025	12/24/2025	Abdul Mirza	None	None	Yogesh Patel
								07/21/2025

FROM 1.00000ml of P13786 + 999.00000ml of E3944 = Final Quantity: 1000.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
84	Pest/PCB Surrogate Stock 20 PPM	PP24805	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza
								08/19/2025

FROM 1.00000ml of P13789 + 9.00000ml of E3962 = Final Quantity: 10.000 ml



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
215	AR1242 1000 PPB WORKING STD	PP24808	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza
								08/19/2025

FROM 0.10000ml of P13706 + 99.40000ml of E3962 + 0.50000ml of PP24805 = Final Quantity: 100.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
216	AR1248 1000 PPB WORKING STD	PP24809	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza
								08/19/2025

FROM 0.10000ml of P13695 + 99.40000ml of E3962 + 0.50000ml of PP24805 = Final Quantity: 100.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
217	AR1254 1000 PPB WORKING STD	PP24810	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza
								08/19/2025

FROM 0.10000ml of P11592 + 99.40000ml of E3962 + 0.50000ml of PP24805 = Final Quantity: 100.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1529	AR1262 1000 PPB Working Solution	PP24811	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza
								08/19/2025

FROM 0.10000ml of P13881 + 99.40000ml of E3962 + 0.50000ml of PP24805 = Final Quantity: 100.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1532	AR1268 1000 PPB Working Solution	PP24812	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza
								08/19/2025

FROM 0.10000ml of P13885 + 99.40000ml of E3962 + 0.50000ml of PP24805 = Final Quantity: 100.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
404	AR1660 100 PPM Stock Solution 2nd Source	PP24814	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza
								08/19/2025

FROM 1.00000ml of P12952 + 9.00000ml of E3804 = Final Quantity: 10.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
405	AR1660 1000/100 PPB ICV STD	PP24815	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza
								08/19/2025

FROM 98.50000ml of E3962 + 0.50000ml of PP24805 + 1.00000ml of PP24814 = Final Quantity: 100.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3789	AR1221 1000 PPB WORKING SOL.2ND SOURCE(AGILENT)	PP24816	08/18/2025	02/15/2026	Yogesh Patel	None	None	Abdul Mirza
								08/19/2025

FROM 1.00000ml of P13373 + 98.50000ml of E3962 + 0.50000ml of PP24805 = Final Quantity: 100.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1887	AR1232 1000 PPB Working Sol. 2nd Source	PP24817	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza
								08/19/2025

FROM 1.00000ml of P12700 + 98.50000ml of E3962 + 0.50000ml of PP24805 = Final Quantity: 100.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1889	AR1242 1000 PPB Working Sol. 2nd Source	PP24818	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza
								08/19/2025

FROM 1.00000ml of P13590 + 98.50000ml of E3962 + 0.50000ml of PP24805 = Final Quantity: 100.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1890	AR1248 1000 PPB Working Sol. 2nd Source	PP24819	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza 08/19/2025

FROM 1.00000ml of P13592 + 98.50000ml of E3962 + 0.50000ml of PP24805 = Final Quantity: 100.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
2193	AR1254 100 PPM STOCK SOLUTION	PP24820	08/18/2025	02/15/2026	Yogesh Patel	None	None	Abdul Mirza 08/19/2025

FROM 1.00000ml of P13833 + 9.00000ml of E3963 = Final Quantity: 10.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3757	AR1262 1000 PPB Working Solution second source	PP24821	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza
								08/19/2025

FROM 1.00000ml of P12703 + 98.50000ml of E3962 + 0.50000ml of PP24805 = Final Quantity: 100.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3817	AR1268 1000 ppb Working Soln. 2nd source	PP24822	08/18/2025	02/15/2026	Yogesh Patel	None	None	Abdul Mirza
								08/19/2025

FROM 1.00000ml of P11522 + 98.50000ml of E3962 + 0.50000ml of PP24805 = Final Quantity: 100.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
203	AR1660 750 PPB STD	PP24823	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza
								08/19/2025

FROM 0.25000ml of W3234 + 0.75000ml of PP24330 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
204	AR1660 500 PPB STD	PP24824	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza
								08/19/2025

FROM 0.50000ml of W3234 + 0.50000ml of PP24330 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
205	AR1660 250 PPB STD	PP24825	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza
								08/19/2025

FROM 0.75000ml of W3234 + 0.25000ml of PP24330 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
206	AR1660 50 PPB STD	PP24826	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza
								08/19/2025

FROM 0.90000ml of W3234 + 0.10000ml of PP24824 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1079	AR1221 750 PPB STD	PP24827	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza
								08/19/2025

FROM 0.25000ml of W3234 + 0.75000ml of PP24806 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
222	AR1221 500 PPB STD	PP24828	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza
								08/19/2025

FROM 0.50000ml of W3234 + 0.50000ml of PP24806 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1080	AR1221 250 PPB STD	PP24829	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza
								08/19/2025

FROM 0.75000ml of W3234 + 0.25000ml of PP24806 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1081	AR1221 50 PPB STD	PP24830	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza
								08/19/2025

FROM 0.90000ml of W3234 + 0.10000ml of PP24828 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1063	AR1232 750 PPB STD	PP24831	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza
								08/19/2025

FROM 0.25000ml of W3234 + 0.75000ml of PP24807 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
223	AR1232 500 PPB STD	PP24832	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza
								08/19/2025

FROM 0.50000ml of W3234 + 0.50000ml of PP24807 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1064	AR1232 250 PPB STD	PP24833	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza
								08/19/2025

FROM 0.75000ml of W3234 + 0.25000ml of PP24807 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1065	AR1232 50 PPB STD	PP24834	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza
								08/19/2025

FROM 0.90000ml of W3234 + 0.10000ml of PP24832 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1067	AR1242 750 PPB STD	PP24835	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza
								08/19/2025

FROM 0.25000ml of W3234 + 0.75000ml of PP24808 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
224	AR1242 500 PPB STD	PP24836	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza
								08/19/2025

FROM 0.50000ml of W3234 + 0.50000ml of PP24808 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1068	AR1242 250 PPB STD	PP24837	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza
								08/19/2025

FROM 0.75000ml of W3234 + 0.25000ml of PP24808 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1069	AR1242 50 PPB STD	PP24838	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza
								08/19/2025

FROM 0.90000ml of W3234 + 0.10000ml of PP24836 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1075	AR1248 750 PPB STD	PP24839	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza
								08/19/2025

FROM 0.25000ml of W3234 + 0.75000ml of PP24809 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
225	AR1248 500 PPB STD	PP24840	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza
								08/19/2025

FROM 0.50000ml of W3234 + 0.50000ml of PP24809 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1076	AR1248 250 PPB STD	PP24841	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza
								08/19/2025

FROM 0.75000ml of W3234 + 0.25000ml of PP24809 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1077	AR1248 50 PPB STD	PP24842	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza
								08/19/2025

FROM 0.90000ml of W3234 + 0.10000ml of PP24840 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1071	AR1254 750 PPB STD	PP24843	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza
								08/19/2025

FROM 0.25000ml of W3234 + 0.75000ml of PP24810 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
226	AR1254 500 PPB STD	PP24844	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza
								08/19/2025

FROM 0.50000ml of W3234 + 0.50000ml of PP24810 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1072	AR1254 250 PPB STD	PP24845	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza
								08/19/2025

FROM 0.75000ml of W3234 + 0.25000ml of PP24810 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1073	AR1254 50 PPB STD	PP24846	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza
								08/19/2025

FROM 0.90000ml of W3234 + 0.10000ml of PP24844 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3753	AR1262 750 PPB STD	PP24847	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza
								08/19/2025

FROM 0.25000ml of W3234 + 0.75000ml of PP24811 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1530	AR1262 500 PPB STD	PP24848	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza
								08/19/2025

FROM 0.50000ml of W3234 + 0.50000ml of PP24811 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3754	AR1262 250 PPB STD	PP24849	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza
								08/19/2025

FROM 0.75000ml of W3234 + 0.25000ml of PP24811 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3755	AR1262 50 PPB STD	PP24850	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza
								08/19/2025

FROM 0.90000ml of W3234 + 0.10000ml of PP24848 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3820	AR1268 750 PPB STD	PP24851	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza
								08/19/2025

FROM 0.25000ml of W3234 + 0.75000ml of PP24812 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1533	AR1268 500 PPB STD	PP24852	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza
								08/19/2025

FROM 0.50000ml of W3234 + 0.50000ml of PP24812 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3821	AR1268 250 PPB STD	PP24853	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza
								08/19/2025

FROM 0.75000ml of W3234 + 0.25000ml of PP24812 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3822	AR1268 50 PPB STD	PP24854	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza
								08/19/2025

FROM 0.90000ml of W3234 + 0.10000ml of PP24852 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
406	AR1660 500 PPB ICV	PP24855	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza
								08/19/2025

FROM 0.50000ml of W3234 + 0.50000ml of PP24815 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3790	AR1221 500 PPB ICV(AGILENT)	PP24856	08/18/2025	02/15/2026	Yogesh Patel	None	None	Abdul Mirza
								08/19/2025

FROM 0.50000ml of W3234 + 0.50000ml of PP24816 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1888	AR1232 500 PPB ICV	PP24857	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza
								08/19/2025

FROM 0.50000ml of W3234 + 0.50000ml of PP24817 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1891	AR1242 500 PPB ICV	PP24858	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza
								08/19/2025

FROM 0.50000ml of W3234 + 0.50000ml of PP24818 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1892	AR1248 500 PPB ICV	PP24859	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza
								08/19/2025

FROM 0.50000ml of W3234 + 0.50000ml of PP24819 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1894	AR1254 500 PPB ICV	PP24860	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza
								08/19/2025

FROM 0.50000ml of W3234 + 0.50000ml of PP24381 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3758	AR1262 500 PPB STD ICV	PP24861	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza
								08/19/2025

FROM 0.50000ml of W3234 + 0.50000ml of PP24821 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3823	AR1268 500 PPB STD ICV	PP24862	08/18/2025	02/15/2026	Yogesh Patel	None	None	Abdul Mirza
								08/19/2025

FROM 0.50000ml of W3234 + 0.50000ml of PP24822 = Final Quantity: 1.000 ml



<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3857	5000 PPB PCB SPIKE SOLUTION 2ND SOURCE	PP24930	09/18/2025	03/18/2026	Abdul Mirza	None	None	Yogesh Patel 09/19/2025
<u>FROM</u> 0.50000ml of P12953 + 99.50000ml of E3972 = Final Quantity: 100.000 ml								

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	9005-05 / Acetone Ultra (cs/4x4L)	24E0761004	02/18/2026	10/01/2024 / Rajesh	09/25/2024 / Rajesh	E3804

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	PC19631-100 / SODIUM SULFATE, ANHYDROUS, PEST GRADE, 1	417203	01/28/2026	07/28/2025 / RUPESH	01/29/2025 / Rajesh	E3875

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9254-03 / Acetone, Ultra Resi (cs/4x4L)	24H1462005	05/24/2027	06/20/2025 / RUPESH	05/14/2025 / RUPESH	E3944

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-3382-05 / Sand, Purified (cs/4x2.5kg)	25A2756718	12/31/2028	07/09/2025 / RUPESH	04/28/2020 / RUPESH	E3951

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9262-03 / Hexane, Ultra-Resi (cs/4x4L)	25C0362005	04/30/2026	08/05/2025 / RUPESH	07/30/2025 / RUPESH	E3962

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9254-03 / Acetone, Ultra Resi (cs/4x4L)	24H2762008	02/15/2026	08/15/2025 / RUPESH	08/07/2025 / RUPESH	E3963

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9262-03 / Hexane, Ultra-Resi (cs/4x4L)	25C0362005	04/30/2026	09/15/2025 / Evelyn	09/04/2025 / Riteshkumar	E3971

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9254-03 / Acetone, Ultra Resi (cs/4x4L)	24H1462005	05/24/2027	09/16/2025 / Evelyn	09/04/2025 / Riteshkumar	E3972

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9262-03 / Hexane, Ultra-Resi (cs/4x4L)	25C0362006	04/13/2026	10/13/2025 / RUPESH	10/08/2025 / Riteshkumar	E3978

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9673-33 / Sulfuric Acid, Instra-Analyzed (cs/6c2.5L)	24i1262013	11/07/2025	05/07/2025 / RUPESH	02/18/2025 / Mohan	M6157

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Agilent Technologies	PP-382-1 / Aroclor 1268	0006587800	09/18/2025	03/18/2025 / yogesh	02/21/2022 / Ankita	P11522

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32011 / PCB Mix, Aroclor 1254, 1000ug/mL, Hexane, 1mL/ampul	A0175403	02/18/2026	08/18/2025 / yogesh	03/18/2022 / Abdul	P11592

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc	91867 / Aroclor 1232 100 ug/mL	020823	02/18/2026	08/18/2025 / yogesh	08/07/2023 / Ankita	P12700

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc	x9166 / Aroclor 1262 100 ug/mL	060523	02/18/2026	08/18/2025 / yogesh	08/07/2023 / Ankita	P12703

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	20064 / Aroclor 1016/1260	022023	02/18/2026	08/18/2025 / yogesh	12/20/2023 / Yogesh	P12952

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	20064 / Aroclor 1016/1260	022023	03/18/2026	09/18/2025 / Abdul	12/20/2023 / Yogesh	P12953

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	/ Arochlor 1254	121823	04/03/2025	10/03/2024 / Ankita	12/20/2023 / Yogesh	P12957

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32000 / Pesticide Mix, CLP method, Pesticide Surrogate Mix, 200ug/mL, Acetone, 1mL	A0206810	09/18/2025	03/18/2025 / yogesh	04/22/2024 / Abdul	P13356

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Agilent Technologies	PP-292-1 / Aroclor 1221	0006783205	09/18/2025	03/18/2025 / yogesh	05/02/2024 / Ankita	P13373

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32007 / PCB Mix, Aroclor 1221, 1000ug/mL, Hexane, 1mL/ampul	A0207511	02/18/2026	08/18/2025 / yogesh	05/03/2024 / Abdul	P13375

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Agilent Technologies	PP-312-1 / Aroclor 1242	0006665550	02/18/2026	08/18/2025 / yogesh	10/14/2024 / Ankita	P13590

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Agilent Technologies	PP-342-1 / Aroclor 1248	0006726317	02/18/2026	08/18/2025 / yogesh	10/14/2024 / Ankita	P13592

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32010 / PCB Mix, Aroclor 1248, 1000ug/mL, Hexane, 1mL/ampul	A0214733	02/18/2026	08/18/2025 / yogesh	10/17/2024 / yogesh	P13695

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32039 / PCB Mix, Aroclor 1016/1260, 1000ug/mL, hexane, 1mL/ampul	A0210629	09/18/2025	03/18/2025 / yogesh	10/17/2024 / yogesh	P13697

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32009 / PCB Mix, Aroclor 1242, 1000ug/mL, Hexane, 1mL/ampul	A0210232	02/18/2026	08/18/2025 / yogesh	10/17/2024 / yogesh	P13706

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32000 / Pesticide Mix, CLP method, Pesticide Surrogate Mix, 200ug/mL, Acetone, 1mL	A0214495	12/24/2025	06/24/2025 / Abdul	11/19/2024 / Ankita	P13786

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32000 / Pesticide Mix, CLP method, Pesticide Surrogate Mix, 200ug/mL, Acetone, 1mL	A0214495	02/18/2026	08/18/2025 / yogesh	11/19/2024 / Ankita	P13789

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32011 / PCB Mix, Aroclor 1254, 1000ug/mL, Hexane, 1mL/ampul	a0217391	02/18/2026	08/18/2025 / yogesh	12/09/2024 / Ankita	P13833

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32008 / PCB Mix, Aroclor 1232, 1000ug/mL, Hexane, 1mL/ampul	A0219655	02/18/2026	08/18/2025 / yogesh	01/23/2025 / Ankita	P13879

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32409 / PCB Stock Solution, Aroclor 1262 Std, 1mL, Hexane	A0217041	02/18/2026	08/18/2025 / yogesh	01/23/2025 / Ankita	P13881

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32410 / PCB Stock Solution, Aroclor 1268 Std, 1mL, Hexane	A0217264	02/18/2026	08/18/2025 / yogesh	01/23/2025 / Ankita	P13885

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	DIW / DI Water	Daily Lab-Certified	07/03/2029	07/03/2024 / Iwona	07/03/2024 / Iwona	W3112

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9262-03 / Hexane, Ultra-Resi (cs/4x4L)	24G1962003	08/22/2025	02/03/2025 / jignesh	01/31/2025 / jignesh	W3177

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9262-03 / Hexane, Ultra-Resi (cs/4x4L)	25C0362005	04/30/2026	07/28/2025 / jignesh	07/25/2025 / jignesh	W3234

Material No.: 9005-05
Batch No.: 24E0761004
Manufactured Date: 2024-05-02
Retest Date: 2029-05-01
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
Assay ((CH ₃) ₂ CO) (by GC, corrected for water)	≥ 99.5 %	99.8 %
Color (APHA)	≤ 10	< 5
Residue after Evaporation	≤ 5 ppm	< 1 ppm
Titration Acid (μeq/g)	≤ 0.3	0.1
Titration Base (μeq/g)	≤ 0.5	0.1
Water (H ₂ O)	≤ 0.5 %	0.1 %
Solubility in H ₂ O	Passes Test	Passes Test
Chloride (Cl)	≤ 0.2 ppm	< 0.2 ppm
Phosphate (PO ₄)	≤ 0.05 ppm	< 0.05 ppm
Trace Impurities – Aluminum (Al)	≤ 50.0 ppb	< 5.0 ppb
Arsenic and Antimony (as As)	≤ 5.0 ppb	< 5.0 ppb
Trace Impurities – Barium (Ba)	≤ 20.0 ppb	< 1.0 ppb
Trace Impurities – Beryllium (Be)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities – Bismuth (Bi)	≤ 20.0 ppb	< 10.0 ppb
Trace Impurities – Boron (B)	≤ 10.0 ppb	< 5.0 ppb
Trace Impurities – Cadmium (Cd)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities – Calcium (Ca)	≤ 25.0 ppb	3.6 ppb
Trace Impurities – Chromium (Cr)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities – Cobalt (Co)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities – Copper (Cu)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities – Gallium (Ga)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities – Germanium (Ge)	≤ 10.0 ppb	< 10.0 ppb
Trace Impurities – Gold (Au)	≤ 20 ppb	< 5 ppb
Trace Impurities – Iron (Fe)	≤ 20.0 ppb	< 1.0 ppb
Trace Impurities – Lead (Pb)	≤ 10.0 ppb	< 10.0 ppb
Trace Impurities – Lithium (Li)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities – Magnesium (Mg)	≤ 20 ppb	< 1 ppb
Trace Impurities – Manganese (Mn)	≤ 10.0 ppb	< 1.0 ppb

Recd by RP on 9/25/24

E 3804

>>> Continued on page 2 >>>

Acetone
CMOS

 **avantor™**



Material No.: 9005-05
Batch No.: 24E0761004

Test	Specification	Result
Trace Impurities – Molybdenum (Mo)	≤ 10.0 ppb	< 5.0 ppb
Trace Impurities – Nickel (Ni)	≤ 10.0 ppb	< 5.0 ppb
Trace Impurities – Niobium (Nb)	≤ 50.0 ppb	< 1.0 ppb
Trace Impurities – Potassium (K)	≤ 10.0 ppb	< 10.0 ppb
Trace Impurities – Silicon (Si)	≤ 50 ppb	< 10 ppb
Trace Impurities – Silver (Ag)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities – Sodium (Na)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities – Strontium (Sr)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities – Tantalum (Ta)	≤ 50.0 ppb	< 5.0 ppb
Trace Impurities – Thallium (Tl)	≤ 10.0 ppb	< 5.0 ppb
Trace Impurities – Tin (Sn)	≤ 20.0 ppb	< 10.0 ppb
Trace Impurities – Titanium (Ti)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities – Vanadium (V)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities – Zinc (Zn)	≤ 20.0 ppb	7.9 ppb
Trace Impurities – Zirconium (Zr)	≤ 10.0 ppb	< 1.0 ppb
Particle Count – 0.5 µm and greater (Rion KS42AF)	≤ 100 par/ml	8 par/ml
Particle Count – 1.0 µm and greater (Rion KS42AF)	≤ 8 par/ml	2 par/ml

>>> Continued on page 3 >>>

Acetone
CMOS

 **avantor™**



Material No.: 9005-05
Batch No.: 24E0761004

Test	Specification	Result

For Microelectronic Use

Country of Origin: USA
Packaging Site: Paris Mfg Ctr & DC



Michelle Bales
Sr. Manager, Quality Assurance



**PRODUCTOS
QUÍMICOS
MONTERREY, S.A. DE C.V.**

MIRADOR 201, COL. MIRADOR
MONTERREY, N.L. MÉXICO
CP 64070
TEL +52 81 13 52 67 67
www.pqm.com.mx

CERTIFICATE OF ANALYSIS

PRODUCT :	SODIUM SULFATE CRYSTALS ANHYDROUS		
QUALITY :	ACS (CODE RMB3375)	FORMULA :	Na ₂ SO ₄
SPECIFICATION NUMBER:	6399	RELEASE DATE:	MAY/23/2024
LOT NUMBER :	417203		

TEST	SPECIFICATIONS	LOT VALUES
Assay (Na ₂ SO ₄)	Min. 99.0%	99.8 %
pH of a 5% solution at 25°C	5.2 - 9.2	6.2
Insoluble matter	Max. 0.01%	0.001 %
Loss on ignition	Max. 0.5%	0.1 %
Chloride (Cl)	Max. 0.001%	<0.001 %
Nitrogen compounds (as N)	Max. 5 ppm	<5 ppm
Phosphate (PO ₄)	Max. 0.001%	<0.001 %
Heavy metals (as Pb)	Max. 5 ppm	<5 ppm
Iron (Fe)	Max. 0.001%	<0.001 %
Calcium (Ca)	Max. 0.01%	0.001 %
Magnesium (Mg)	Max. 0.005%	0.001 %
Potassium (K)	Max. 0.008%	0.001 %
Extraction-concentration suitability	Passes test	Passes test
Appearance	Passes test	Passes test
Identification	Passes test	Passes test
Solubility and foreign matter	Passes test	Passes test
Retained on US Standard No. 10 sieve	Max. 1%	0.2 %
Retained on US Standard No. 60 sieve	Min. 94%	96.2 %
Through US Standard No. 60 sieve	Max. 5%	3.5 %
Through US Standard No. 100 sieve	Max. 10%	0.1 %

COMMENTS

QC: PhC Irma Belmares

If you need further details, please call our factory or contact our local distributor.

RE-02-01, Ed. 3

E 3875

Acetone
BAKER RESI-ANALYZED® Reagent
For Organic Residue Analysis

Avantor™



Material No.: 9254-03

Batch No.: 24H1462005

Manufactured Date: 2024-05-24

Expiration Date: 2027-05-24

Revision No.: 0

Certificate of Analysis

Test	Specification	Result
Assay ((CH ₃) ₂ CO) (by GC, corrected for water)	>= 99.4 %	99.8 %
Color (APHA)	<= 10	5
Residue after Evaporation	<= 1.0 ppm	0.2 ppm
Substances Reducing Permanganate	Passes Test	Passes Test
Titration Acid (μeq/g)	<= 0.3	0.2
Titration Base (μeq/g)	<= 0.6	<0.1
Water (H ₂ O)	<= 0.5 %	0.2 %
FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL)	<= 5	<1
ECD Sensitive Impurities (as Heptachlor Epoxide) Single Peak (pg/mL)	<= 10	1

For Laboratory, Research, or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: United States
Packaging Site: Phillipsburg Mfg Ctr & DC

E3944

Jamie Croak
Director Quality Operations, Bioscience Production

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

Avantor Performance Materials LLC



Material	BDH9274-2.5KG
Material Description	BDH SAND STDD OTTAWA W+I 2.5KG
Grade	NOT APPLICABLE
Batch	25A2756718
Reassay Date	12/31/2028
CAS Number	14808-60-7
Molecular Formula	SiO ₂
Molecular Mass	60.09
Date of Manufacture	12/05/2024
Storage	Room Temperature

Characteristics	Specifications	Measured Values
Appearance	Beige granules.	Beige granules.
Moisture	<= 0.1 %	0.1 %
Particle Size 30-40 mesh	>= 80 %	99 %
CUSTOMER PART # BDH9274-2.5KG		

Received on 1/12/25.

E3951

Internal ID #: 793

Signature	Additional Information
We certify that this batch conforms to the specifications listed above. This document has been electronically produced and is valid without a signature. Leona Edwardson, Quality Control Sr. Manager - Solon VWR Chemicals, LLC. 28600 Fountain Parkway, Solon OH 44139 USA	Analysis may have been rounded to significant digits in specification limits Product meets analytical specifications of the grades listed.

n-Hexane 95%
ULTRA RESI-ANALYZED
For Organic Residue Analysis



Material No.: 9262-03
Batch No.: 25C0362005
Manufactured Date: 2025-01-29
Expiration Date: 2026-04-30
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL)	≤ 5	1
ECD Sensitive Impurities (as HeptachlorEpoxide) Single Peak (pg/mL)	≤ 10	6
ECD-Sensitive Impurities (as EthyleneDibromide) - Single Impurity Peak (ng/mL)	≤ 5	5
Assay (Total Saturated C ₆ Isomers) (by GC, corrected for water)	$\geq 99.5 \%$	100.0 %
Assay (as n-Hexane) (by GC, corrected for water)	$\geq 95 \%$	100 %
Color (APHA)	≤ 10	10
Residue after Evaporation	≤ 1.0 ppm	0.1 ppm
Substances Darkened by H ₂ SO ₄	Passes Test	Passes Test
Water (by KF, coulometric)	$\leq 0.05 \%$	$< 0.01 \%$

For Laboratory, Research, or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: United States
Packaging Site: Phillipsburg Mfg Ctr & DC

Received on 7/30/25

E3962

Jamie Croak
Director Quality Operations, Bioscience Production

Acetone.
BAKER RESI-ANALYZED® Reagent
For Organic Residue Analysis

Avantor™



Material No.: 9254-03
Batch No.: 24H2762008
Manufactured Date: 2024-04-18
Expiration Date: 2027-04-18
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
Assay ((CH ₃) ₂ CO) (by GC, corrected for water)	>= 99.4 %	100.0 %
Color (APHA)	<= 10	5
Residue after Evaporation	<= 1.0 ppm	0.0 ppm
Substances Reducing Permanganate	Passes Test	Passes Test
Titration Acid (μeq/g)	<= 0.3	0.2
Titration Base (μeq/g)	<= 0.6	<0.1
Water (H ₂ O)	<= 0.5 %	<0.1 %
FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL)	<= 5	1
ECD Sensitive Impurities (as Heptachlor Epoxide) Single Peak (pg/mL)	<= 10	1

For Laboratory, Research, or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: United States
Packaging Site: Phillipsburg Mfg Ctr & DC

Received on 8/6/25

E 3963

J. Croak

Jamie Croak
Director Quality Operations, Bioscience Production

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

Avantor Performance Materials, LLC

100 Matsonford Rd, Suite 200, Radnor, PA, 19087, U.S.A. Phone 610.386.1700

n-Hexane 95%
ULTRA RESI-ANALYZED
For Organic Residue Analysis



Material No.: 9262-03
Batch No.: 25C0362005
Manufactured Date: 2025-01-29
Expiration Date: 2026-04-30
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL)	≤ 5	1
ECD Sensitive Impurities (as HeptachlorEpoxide) Single Peak (pg/mL)	≤ 10	6
ECD-Sensitive Impurities (as EthyleneDibromide) - Single Impurity Peak (ng/mL)	≤ 5	5
Assay (Total Saturated C ₆ Isomers) (by GC, corrected for water)	$\geq 99.5 \%$	100.0 %
Assay (as n-Hexane) (by GC, corrected for water)	$\geq 95 \%$	100 %
Color (APHA)	≤ 10	10
Residue after Evaporation	≤ 1.0 ppm	0.1 ppm
Substances Darkened by H ₂ SO ₄	Passes Test	Passes Test
Water (by KF, coulometric)	$\leq 0.05 \%$	$< 0.01 \%$

For Laboratory, Research, or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: United States
Packaging Site: Phillipsburg Mfg Ctr & DC

E3971

Jamie Croak
Director Quality Operations, Bioscience Production

Acetone
BAKER RESI-ANALYZED® Reagent
For Organic Residue Analysis

avantor™



Material No.: 9254-03

Batch No.: 24H1462005

Manufactured Date: 2024-05-24

Expiration Date: 2027-05-24

Revision No.: 0

Certificate of Analysis

Test	Specification	Result
Assay ((CH ₃) ₂ CO) (by GC, corrected for water)	>= 99.4 %	99.8 %
Color (APHA)	<= 10	5
Residue after Evaporation	<= 1.0 ppm	0.2 ppm
Substances Reducing Permanganate	Passes Test	Passes Test
Titration Acid (µeq/g)	<= 0.3	0.2
Titration Base (µeq/g)	<= 0.6	<0.1
Water (H ₂ O)	<= 0.5 %	0.2 %
FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL)	<= 5	<1
ECD Sensitive Impurities (as Heptachlor Epoxide) Single Peak (pg/mL)	<= 10	1

For Laboratory, Research, or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: United States
Packaging Site: Phillipsburg Mfg Ctr & DC

E3972

Jamie Croak
Director Quality Operations, Bioscience Production

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

Avantor Performance Materials LLC

n-Hexane 95%
ULTRA RESI-ANALYZED
For Organic Residue Analysis

 **avantorsTM**



Material No.: 9262-03

Batch No.: 25C0362006

Manufactured Date: 2025-01-29

Expiration Date: 2026-04-30

Revision No.: 0

Certificate of Analysis

Test	Specification	Result
FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL)	≤ 5	1
ECD Sensitive Impurities (as Heptachlor Epoxide) Single Peak (pg/mL)	≤ 10	6
ECD-Sensitive Impurities (as Ethylene Dibromide) - Single Impurity Peak (ng/mL)	≤ 5	4
Assay (Total Saturated C ₆ Isomers) (by GC, corrected for water)	$\geq 99.5 \%$	100.0 %
Assay (as n-Hexane) (by GC, corrected for water)	$\geq 95 \%$	100 %
Color (APHA)	≤ 10	10
Residue after Evaporation	≤ 1.0 ppm	0.2 ppm
Substances Darkened by H ₂ SO ₄	Passes Test	Passes Test
Water (by KF, coulometric)	$\leq 0.05 \%$	$< 0.01 \%$

For Laboratory, Research, or Manufacturing Use

MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: United States

Packaging Site: Phillipsburg Mfg Ctr & DC

E 3978



Jamie Croak
Director Quality Operations, Bioscience Production

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

Avantor Performance Materials LLC

Sulfuric Acid
BAKER INSTRA-ANALYZED® Reagent
For Trace Metal Analysis
Low Selenium

avantor™



M6157
MS

Material No.: 9673-33

Batch No.: 24I1262013

Manufactured Date: 2024-08-07

Retest Date: 2029-08-06

Revision No.: 0

Certificate of Analysis

Test	Specification	Result
ACS – Assay (H ₂ SO ₄)	95.0 – 98.0 %	96.2 %
Appearance	Passes Test	Passes Test
ACS – Color (APHA)	<= 10	5
ACS – Residue after Ignition	<= 3 ppm	<1 ppm
ACS – Substances Reducing Permanganate(as SO ₂)	<= 2 ppm	<2 ppm
Ammonium (NH ₄)	<= 1 ppm	<1 ppm
Chloride (Cl)	<= 0.1 ppm	<0.1 ppm
Nitrate (NO ₃)	<= 0.2 ppm	0.1 ppm
Phosphate (PO ₄)	<= 0.5 ppm	<0.1 ppm
Trace Impurities – Aluminum (Al)	<= 30.0 ppb	<5.0 ppb
Arsenic & Antimony (as As)	<= 4.0 ppb	<2.0 ppb
Trace Impurities – Boron (B)	<= 10.0 ppb	<5.0 ppb
Trace Impurities – Cadmium (Cd)	<= 2.0 ppb	<1.0 ppb
Trace Impurities – Chromium (Cr)	<= 6.0 ppb	<1.0 ppb
Trace Impurities – Cobalt (Co)	<= 0.5 ppb	<0.3 ppb
Trace Impurities – Copper (Cu)	<= 1.0 ppb	<1.0 ppb
Trace Impurities – Gold (Au)	<= 10.0 ppb	<5.0 ppb
Heavy Metals (as Pb)	<= 500.0 ppb	<100.0 ppb
Trace Impurities – Iron (Fe)	<= 50.0 ppb	<1.0 ppb
Trace Impurities – Lead (Pb)	<= 0.5 ppb	<0.5 ppb
Trace Impurities – Magnesium (Mg)	<= 7.0 ppb	<1.0 ppb
Trace Impurities – Manganese (Mn)	<= 1.0 ppb	<1.0 ppb
Trace Impurities – Mercury (Hg)	<= 0.5 ppb	<0.1 ppb
Trace Impurities – Nickel (Ni)	<= 2.0 ppb	<0.3 ppb
Trace Impurities – Potassium (K)	<= 500.0 ppb	<10.0 ppb
Trace Impurities – Selenium (Se)	<= 50.0 ppb	7.2 ppb
Trace Impurities – Silicon (Si)	<= 100.0 ppb	12.8 ppb
Trace Impurities – Silver (Ag)	<= 1.0 ppb	<1.0 ppb

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

Avantor Performance Materials LLC

Sulfuric Acid

BAKER INSTRA-ANALYZED® Reagent

For Trace Metal Analysis

Low Selenium

 **avantor**™



Material No.: 9673-33

Batch No.: 2411262013

Test	Specification	Result
Trace Impurities – Sodium (Na)	≤ 500.0 ppb	< 5.0 ppb
Trace Impurities – Strontium (Sr)	≤ 5.0 ppb	< 1.0 ppb
Trace Impurities – Tin (Sn)	≤ 5.0 ppb	1.1 ppb
Trace Impurities – Zinc (Zn)	≤ 5.0 ppb	< 1.0 ppb

For Laboratory, Research, or Manufacturing Use

Country of Origin: United States

Packaging Site: Phillipsburg Mfg Ctr & DC



Jamie Croak

Director Quality Operations, Bioscience Production

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

Avantor Performance Materials LLC



Certificate of Analysis

P11518
↓
P11522
AJ
02/21/22

Product Name: Aroclor 1268 Standard

Product Number: PP-382-1

Lot Issue Date: 09-Feb-2021

Lot Number: 0006587800

Expiration Date: 31-Mar-2029

Description:

This analytical reference material (RM) was manufactured and verified in accordance with an ISO 9001 registered quality system and analyte concentrations were verified by an ISO 17025 accredited laboratory. The concentration and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed below.

Analyte	CAS#	Analyte Lot	Concentration ± Uncertainty
Aroclor 1268	011100-14-4	RM00937	100.0 ± 0.5 µg/mL

Matrix: isooctane (2,2,4-trimethylpentane)

Storage Conditions: Store at Room Temperature (15° to 30°C).

Traceability:

The balances used for these measurements are calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z540.3, ISO 9001, ISO 17025, and ISO 17034. Calibrated Class A glassware is used for volumetric measurements. Thermometers are calibrated against a NIST traceable thermometer in accordance with NIST Special Publication 1088.

Homogeneity:

This RM was unitized according to an in-house procedure and is guaranteed to be homogeneous. There is no minimum sub-sample size required.

Intended Use:

This RM is intended for the preparation of working reference samples for use in routine laboratory analyses, calibration of instruments, validation of analytical methods, assessments of measurement methods, and continuing calibration verification.

Instructions for Use:

Sample aliquots for analysis should be withdrawn at 20°C to 25°C immediately after opening the container and should be processed without delay for the certified values to be valid within the stated uncertainties.

Hazards:

Refer to the Safety Data Sheet on www.agilent.com for information regarding this RM.


Expiration of Certification:

The certification of this RM is valid until the expiration date specified above, provided the RM is handled and stored in accordance with the instructions given in this certificate. This certification is nullified if the RM is damaged, contaminated, or otherwise modified.

Maintenance of Certification:

If substantive changes are noted that affect the certification before the expiration of this certificate, Agilent will notify the purchaser.

Sample lot approver:


Monica Bourgeois
QMS Representative



ISO 17034 Cert
No. AR-1936

RM was produced in accordance with TUV USA Inc registered ISO 9001 Quality Management System. Cert # 56 100 18560026
Page: 1 of 1

www.agilent.com/quality/
CSD-QA-015.1



ISO 17025 Cert
No. AT-1937



CERTIFIED REFERENCE MATERIAL

110 Benner Circle
Bellefonte, PA 16823-8812
Tel: (800)356-1688
Fax: (814)353-1309

www.restek.com

Certificate of Analysis



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32011 Lot No.: A0175403
Description : Aroclor® 1254 Standard
Aroclor® 1254 Standard 1,000 µg/mL, Hexane, 1mL/ampul
Container Size : 2 mL Pkg Amt: > 1 mL
Expiration Date : November 30, 2027 Storage: 25°C nominal
Handling: This product contains PCBs. Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	Grav. Conc. (weight/volume)	Expanded Uncertainty (95% C.L.; K=2)
1	Aroclor 1254 CAS # 11097-69-1 (Lot 124-191-B) Purity ----%	1,000.7 µg/mL	+/- 5.9437 µg/mL Gravimetric +/- 31.7284 µg/mL Unstressed +/- 41.4406 µg/mL Stressed

Solvent: Hexane
CAS # 110-54-3
Purity 99%

P11588
↓
P11592 / (S)

AR
04/30/2022

Column:

30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:

helium-constant pressure 20 psi.

Temp. Program:

200°C to 300°C
@ 25°C/min. (hold 10 min.)

Inj. Temp:


250°C

Det. Temp:

300°C

Det. Type:

ECD



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Cathleen Soltis

Cathleen Soltis - Mix Technician

Date Mixed: 15-Aug-2021

Balance: 1128360905

Alexis Shelow

Alexis Shelow - Operations Tech I

Date Passed: 17-Aug-2021

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

P11588
↓
P11592 / (5)

AR
04/30/22



CERTIFIED WEIGHT REPORT

Part Number: **91867**
 Lot Number: **020823**
 Description: **WP 037 - Aroclor 1232**
PCB Technical Mixture
 Expiration Date: **020833**
 Recommended Storage: **Ambient (20 °C)**
 Nominal Concentration ($\mu\text{g/mL}$): **100**
 NIST Test ID#: **6UTB**

Solvent(
Aceton

Weight(s) shown below were combined and diluted to (mL): **100.0**

5E-05 Balance Uncertainty

0.057 Flask Uncertainty

Compound	RM#	Lot Number	Nominal Conc ($\mu\text{g/mL}$)	Purity (%)	Uncertainty Purity	Target Weight (g)
1. Aroclor 1232	17	45-6A	100	100	0.5	0.01000

- The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
- Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
- Standards are certified (+/-) 0.5% of the stated value, unless otherwise stated.
- All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
- Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measure Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).

Comments

GC3-M1 Analysis by Melissa Stonier

Column ID SPB-608 30 meter X 0.53mm X 5 μm film thickness

Flow rates: Helium (carrier) = 5mL/min, Helium (make-up) = 25mL/min


Hydrogen (make-up) = 30mL/min, Air (make-up) = 350mL/min

Oven Profile: Temp 1 = 150°C (Time 1 = 4 min), Temp 2 = 290°C (Time 2 = 13.5 min)

Rate = 8°C/min, Total run time = 35 min

Injector temp. = 200°C, FID Temp. = 300°C. FID Signal = Edaq Channel 1

Standard injection = 1.5 μL , Range=3





CERTIFIED WEIGHT REPORT

Part Number:
Lot Number:
Description:

Expiration Date:
Recommended Storage:
Nominal Concentration (µg/mL):
NIST Test ID#:

X9166
060523
Aroclor 1262
Appendix IX Compound
060533
Ambient (20 °C)
100
6UTB

Solvent(s):
Methanol


5E-05
0.002

Balance Uncertainty
Flask Uncertainty

Volume(s) shown below were combined and diluted to (mL): 20.0

Compound	Part Number	Lot Number	Dil. Factor	Initial Vol. (mL)	Uncertainty Pipette (mL)	Co
1. Aroclor 1262	70444	080322	0.10	2.00	0.017	

Method GC7MSD-7.M: Column:(30m X 0.25mm ID X 0.25µm film thickn min.), Rate = 8°C/min., Injector B= 200°C, Detector B = 290°C.



- The certified value is the concentration calculated from gravimetric and volumetric me
- Standards are prepared gravimetrically using balances that are calibrated with weight
- Standards are certified (+/-) 0.5% of the stated value, unless otherwise stated.
- All Standards, after opening ampule, should be stored with caps tight and under appro
- Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).



Certified Reference Material CRM

CERTIFIED WEIGHT REPORT

Part Number:
20064

Lot Number:
022023

Description:
CLP PCB'S - Aroclor Mix
Aroclors 1016 & 1260

Expiration Date:
022033

Recommended Storage:
Ambient (20 °C)

Nominal Concentration (µg/mL):
1000

NIST Test ID#:
6UTB

Solvent(s):
Hexane

Lot#
273615

Formulated By:
Benson Chan

Reviewed By:
Pedro L. Rentas

DATE
022023

DATE
022023

5E-05 Balance Uncertainty
0.010 Flask Uncertainty

Nominal Conc (µg/mL):
200.0

Weight(s) shown below were combined and diluted to (mL):

Part Number:
20064

Lot Number:
022023

Description:
CLP PCB'S - Aroclor Mix
Aroclors 1016 & 1260

Expiration Date:
022033

Recommended Storage:
Ambient (20 °C)

Nominal Concentration (µg/mL):
1000

NIST Test ID#:
6UTB

Solvent(s):
Hexane

Lot#
273615

Formulated By:
Benson Chan

Reviewed By:
Pedro L. Rentas

DATE
022023

DATE
022023

5E-05 Balance Uncertainty
0.010 Flask Uncertainty


Nominal Conc (µg/mL):
200.0

Weight(s) shown below were combined and diluted to (mL):

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Purity Uncertainty	Target Weight(g)	Actual Weight(g)	Actual Conc (µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	CAS#	OSHA PEL (TWA)	LD50
1. Aroclor 1016	15	020491JC	1000	100	0.2	0.20004	0.20060	1002.8	4.0	12674-11-2	N/A	N/A
2. Aroclor 1260	21	020491JC	1000	100	0.2	0.20004	0.20081	1003.9	4.0	11086-82-5	0.5mg/m3	orl-rat 1315mg/kg

* The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
* Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
* All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
* Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).

Comments
GC3-M1 Analysis by Melissa Stortier
Column ID SPB-608 30 meter X 0.53mm X5µm film thickness
Flow rates: Helium (carrier) = 5mL/min, Helium (make-up) = 25mL/min
Hydrogen (make-up) = 30mL/min, Air (make-up) = 350mL/min
Oven Profile: Temp 1 = 150°C (Time 1 = 4 min), Temp 2 = 280°C (Time 2 = 13.5 min)
Rate = 8°C/min, Total run time = 35 min
Injector temp. = 200°C, FID Temp. = 300°C, FID Signal = Etdaq Channel 1
Standard Injection = 1.5µL, Range=3





Certified Reference Material CRM

CERTIFIED WEIGHT REPORT

Part Number:
20064

Lot Number:
022023

Description:
CLP PCB'S - Aroclor Mix
Aroclors 1016 & 1260

Expiration Date:
022033

Recommended Storage:
Ambient (20 °C)

Nominal Concentration (µg/mL):
1000

NIST Test ID#:
6UTB

Solvent(s):
Hexane

Lot#
273615

Formulated By:
Benson Chan

Reviewed By:
Pedro L. Rentas

DATE
022023

DATE
022023

5E-05 Balance Uncertainty
0.010 Flask Uncertainty

Nominal Conc (µg/mL):
200.0

Weight(s) shown below were combined and diluted to (mL):

Part Number:
20064

Lot Number:
022023

Description:
CLP PCB'S - Aroclor Mix
Aroclors 1016 & 1260

Expiration Date:
022033

Recommended Storage:
Ambient (20 °C)

Nominal Concentration (µg/mL):
1000

NIST Test ID#:
6UTB

Solvent(s):
Hexane

Lot#
273615

Formulated By:
Benson Chan

Reviewed By:
Pedro L. Rentas

DATE
022023

DATE
022023

5E-05 Balance Uncertainty
0.010 Flask Uncertainty


Nominal Conc (µg/mL):
200.0

Weight(s) shown below were combined and diluted to (mL):

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Purity Uncertainty	Target Weight(g)	Actual Weight(g)	Actual Conc (µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	CAS#	OSHA PEL (TWA)	LD50
1. Aroclor 1016	15	020491JC	1000	100	0.2	0.20004	0.20060	1002.8	4.0	12674-11-2	N/A	N/A
2. Aroclor 1260	21	020491JC	1000	100	0.2	0.20004	0.20081	1003.9	4.0	11086-82-5	0.5mg/m3	orl-rat 1315mg/kg

* The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
* Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
* All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
* Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).

Comments
GC3-M1 Analysis by Melissa Stortier
Column ID SPB-608 30 meter X 0.53mm X5µm film thickness
Flow rates: Helium (carrier) = 5mL/min, Helium (make-up) = 25mL/min
Hydrogen (make-up) = 30mL/min, Air (make-up) = 350mL/min
Oven Profile: Temp 1 = 150°C (Time 1 = 4 min), Temp 2 = 280°C (Time 2 = 13.5 min)
Rate = 8°C/min, Total run time = 35 min
Injector temp. = 200°C, FID Temp. = 300°C, FID Signal = Etdaq Channel 1
Standard Injection = 1.5µL, Range=3





CERTIFIED WEIGHT REPORT

Part Number: 99139
Lot Number: 121823
Description: Aroclor 1254

Expiration Date: 121833
Recommended Storage: Ambient (20 °C)
Nominal Concentration (µg/mL): 100
NIST Test ID#: 6UTB

Volume(s) shown below were combined and diluted to (mL): 20.0
Note: Aroclor 1254 is a mix of isomers.

Solvent(s): Iso-octane
Lot#: 82227

Formulated By: Anthony Mahoney	121823	DATE
Reviewed By: Pedro L. Rentas	121823	DATE


P12956 / Y.P.
12/18/23
P12957

Compound	Part Number	Lot Number	Dilution Factor	Initial Vol. (mL)	Initial Uncertainty	Pipette (mL)	Conc. (µg/mL)	Initial Conc. (µg/mL)	Final Conc. (µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	SDS Information		
											(Solvent Safety Info. On Attached pg.)	CAS#	OSHA PEL (TWA) LD50
1. Aroclor 1254	79100	121823	0.10	2.00	0.017	0.017	1003.3	100.1	11097-69-1	1.8	0.5mg/m3 (skin)		or-rat 1295mg/kg

* The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
* Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
* Standards are certified (+/-) 0.5% of the stated value, unless otherwise stated.
* All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
* Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).

Comments
GC3-M1 Analysis by Melissa Stonier
Column ID SPB-608 30 meter X 0.53mm X 5µm film thickness
Flow rates: Helium (carrier) = 5mL/min, Helium (make-up) = 25mL/min
Hydrogen (make-up) = 30mL/min, Air (make-up) = 350mL/min
Oven Profile: Temp 1 = 150 °C (Time 1 = 4 min), Temp 2 = 280 °C (Time 2 = 13.5 min)
Rate = 8 °C/min, Total run time = 35 min
Injector temp. = 200 °C, FID Temp. = 300 °C, FID Signal = EDAQ Channel 1
Standard injection = 1.5µL, Range=3

Peak No.	Name	FID RT (min.)
1	Total Aroclor 1254	18.12





110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32000 **Lot No.:** A0206810

Description : Pesticide Surrogate Mix
Pesticide Surrogate Mix 200 µg/mL, Acetone, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : April 30, 2030 **Storage:** 10°C or colder

Handling: Contains PCBs - sonicate prior to use. **Ship:** Ambient

P13348
↓
P13357
10
BAUF
04/25/2024

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	2,4,5,6-Tetrachloro-m-xylene	877-09-8	RP220407	99%	200.3 µg/mL	+/- 11.1143
2	Decachlorobiphenyl (BZ# 209)	2051-24-3	30638	99%	200.6 µg/mL	+/- 11.1298

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Acetone
CAS # 67-64-1
Purity 99%

Tech Tips:

Decachlorobiphenyl has poor solubility in most organic solvents. The maximum concentration that can be prepared in acetone, hexane, or isooctane is 200µg/mL. Temperature will affect the solubility as well. Storing solutions at reduced temperatures will cause decachlorobiphenyl to precipitate.

Products containing decachlorobiphenyl must be sonicated for a minimum of 10 minutes prior to opening the ampul. Because each ultrasonic bath operates at a different energy level, 10 minutes is a guideline only. Longer sonication time will not affect product quality.

These precautions apply to working solutions prepared in your laboratory as well. The amount of compound that precipitates depends on concentration AND temperature. If you store your standards at a temperature lower than 4°C (even dilute solutions), allow extra sonication time.

Quality Confirmation Test

Column:

30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:

helium-constant pressure 20 psi.

Temp. Program:

200°C to 300°C
@ 25°C/min. (hold 10 min.)

Inj. Temp:

250°C

Det. Temp:

300°C

Det. Type:


ECD

Split Vent:


10 ml/min.

Inj. Vol

1µl




This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Laith Clemente - Operations Technician I

Date Mixed: 22-Jan-2024

Balance Serial # 1128360905


Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 24-Jan-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

P 13348
↓
P 13357
SAUF
04/25/2025



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32007 **Lot No.:** A0207511

Description : Aroclor® 1221 Standard

Aroclor® 1221 Standard 1,000 µg/mL, Hexane, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : May 31, 2030 **Storage:** 25°C nominal

Handling: This product contains PCBs. **Ship:** Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Aroclor 1221	11104-28-2	13192500	----%	1,002.7 µg/mL	+/- 55.6424

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Hexane
CAS # 110-54-3
Purity 99%

P 13374
↓
P 13375
/

[Signature]
05/6/2024

Quality Confirmation Test

Column:

30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:

helium-constant pressure 20 psi.

Temp. Program:

200°C to 300°C
@ 25°C/min. (hold 10 min.)

Inj. Temp:

250°C

Det. Temp:

300°C

Det. Type:


ECD

Split Vent:


10 ml/min.

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Laith Clemente - Operations Technician I

Date Mixed: 06-Feb-2024

Balance Serial # 1128360905


Jennifer Pollino - Operations Tech III - ARM QC


Date Passed: 08-Feb-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

P13374 } ②
↓
P13375 }


05/6/2024

ISO 17034



Agilent

Trusted Answers

Reference Material Certificate

Product Name: Aroclor 1242 Standard

Lot Number: 0006665550

Product Number: PP-312-1

Lot Issue Date: 08-Feb-2022

Storage Conditions: Store at Room Temperature (15° to 30°C).

Expiration Date: 31-Jan-2027

Component Name	CERTIFIED VALUES		CAS#	Analyte Lot
	Concentration	Expanded Uncertainty		
Aroclor 1242	100.4	± 0.5 µg/mL	053469-21-9	NT01020

Matrix: isooctane (2,2,4-trimethylpentane)

Description:

This document is prepared in accordance with ISO 17034 and Guide 31. This analytical reference material standard was manufactured and verified in accordance with an ISO 9001 registered quality system and analyte concentrations were verified by an ISO 17025 accredited laboratory. The concentration and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed above.

Traceability:

The balances used for these measurements are calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z540.3, ISO 9001, ISO 17025, and ISO 17034. Calibrated Class A glassware is used for volumetric measurements. Thermometers are calibrated against a NIST traceable thermometer in accordance with NIST Special Publication 1088.

Homogeneity:

This analytical reference standard was unitized according to an in-house procedure and is guaranteed to be homogeneous. There is no minimum sub-sample size required.

Instructions for Use:

Sample aliquots for analysis should be withdrawn at 20°C to 25°C immediately after opening the container and should be processed without delay for the certified values to be valid within the stated uncertainties.

Safety:

Refer to the Safety Data Sheet on www.agilent.com for information regarding this analytical reference material.

Intended Use:

This analytical reference standard is intended for the preparation of working reference samples for use in routine laboratory analyses, calibration of instruments, validation of analytical methods, assessments of measurement methods, and continuing calibration verification.

Expiration of Certification:

The certification of this analytical reference standard is valid until the expiration date specified above, provided the material is handled and stored in accordance with the instructions given in this certificate. This certification is nullified if the material is damaged, contaminated, or otherwise modified.

P13589
↓
P13590AJ
10/14/24

ISO 17034



Maintenance of Certification:

If substantive changes are noted that affect the certification before the expiration of this certificate, Agilent will notify the purchaser.

Sample lot approver:

Monica Bourgeois
QMS Representative



ISO 17034 Cert
No. AR-1936

RM was produced in accordance with the TUV/SUD registered ISO 9001:2015
Quality Management System. Cert# 95T215321

Page: 2 of 2

www.agilent.com/quality/

CSD-QA-015.1



ISO 17025
Cert No. AT-

Reference Material Certificate
Product Information Sheet

Product Name: Aroclor 1248 Standard

Lot Number: 0006726317

Product Number: PP-342-1

Lot Issue Date: 27-Jan-2023

Storage Conditions: Store at Room Temperature (15° to 30°C).

Expiration Date: 28-Feb-2031

Component Name	Concentration	Uncertainty	CAS#	Analyte Lot
Aroclor 1248	100.3 ±	0.5 µg/mL	012672-29-6	NT01582

Matrix: isooctane (2,2,4-trimethylpentane)

Description:

This document is prepared in accordance with ISO 17034 and Guide 31. This analytical reference material (RM) standard was manufactured and verified in accordance with an ISO 9001 registered quality system and analyte concentrations were verified by an ISO 17025 accredited laboratory. The concentration and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed above. Purity values are taken from approved vendor raw material certificates.

Traceability:

The balances used for these measurements are calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z540.3, ISO 9001, ISO 17025, and ISO 17034. Calibrated Class A glassware is used for volumetric measurements. Thermometers are calibrated against a NIST traceable thermometer in accordance with NIST Special Publication 1088.

Homogeneity:

This analytical reference (RM) standard was unitized according to an in-house procedure and is guaranteed to be homogeneous. There is no minimum sub-sample size required.

Instructions for Use:

Sample aliquots for analysis should be withdrawn at 20°C to 25°C immediately after opening the container and should be processed without delay for the certified values to be valid within the stated uncertainties.

Safety:

Refer to the Safety Data Sheet on www.agilent.com for information regarding this analytical reference material.

Intended Use:

This analytical reference (RM) standard is intended for the preparation of working reference samples for use in routine laboratory analyses, calibration of instruments, validation of analytical methods, assessments of measurement methods, and continuing calibration verification.

Expiration of Certification:

The certification of this analytical reference standard (RM) is valid until the expiration date specified above, provided the material is handled and stored in accordance with the instructions given in this certificate. This certification is nullified if the material is damaged, contaminated, or otherwise modified.

P13591
↓
P13592

AS
10/14/2024



Maintenance of Certification:

If substantive changes are noted that affect the certification before the expiration of this certificate, Agilent will notify the purchaser.

Sample lot approver:

Monica Bourgeois
QMS Representative



ISO 17034 Cert
No. AR-1936

RM was produced in accordance with the TUV/SUD registered ISO
9001:2015 Quality Management System. Cert# 951215321

Page: 2 of 2

www.agilent.com/quality/

CSD-QA-015.1

ISO 17025



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32010 **Lot No.:** A0214733
Description : Aroclor® 1248 Standard
Aroclor® 1248 Standard 1,000µg/mL, Hexane, 1mL/ampul
Container Size : 2 mL **Pkg Amt:** > 1 mL
Expiration Date : November 30, 2030 **Storage:** 25°C nominal
Handling: This product contains PCBs. **Ship:** Ambient

P13695 } Y.P.
P13696 } 10/17/24

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Aroclor 1248	12672-29-6	15242800	----%	1,005.0 µg/mL	+/- 55.7700

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Hexane
CAS # 110-54-3
Purity 99%

Quality Confirmation Test

Column:

30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:

helium-constant pressure 20 psi.

Temp. Program:

200°C to 300°C
@ 25°C/min. (hold 10 min.)

Inj. Temp:

250°C

Det. Temp:

300°C

Det. Type:


ECD

Split Vent:

10 ml/min.

Inj. Vol

0.2µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Rebecca Gingerich - Operations Tech II

Date Mixed: 02-Aug-2024

Balance Serial # C322230531

Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 13-Aug-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32039 **Lot No.:** A0210629

Description : Aroclor® 1016/1260 Mix
Aroclor® 1016/1260 Mix 1,000 µg/mL, Hexane, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : July 31, 2030 **Storage:** 25°C nominal

Handling: This product contains PCBs. **Ship:** Ambient

P13697
↓
P13701 } Y.P.
10/19/24

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Aroclor 1016	12674-11-2	07	----%	1,005.3 µg/mL	+/- 55.7809
2	Aroclor 1260	11096-82-5	1320657	----%	1,000.0 µg/mL	+/- 55.4850

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Hexane
CAS # 110-54-3
Purity 99%

Quality Confirmation Test

Column:

30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:

helium-constant pressure 20 psi.

Temp. Program:

200°C to 300°C
@ 25°C/min. (hold 10 min.)

Inj. Temp:

250°C

Det. Temp:

300°C

Det. Type:


ECD

Split Vent:


10 ml/min.

Inj. Vol

0.2µl




This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Laith Clemente - Operations Technician I

Date Mixed: 22-Apr-2024

Balance Serial # B442140311


Dillan Murphy - Operations Technician I

Date Passed: 24-Apr-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32009 **Lot No.:** A0210232

Description : Aroclor® 1242 Standard

Aroclor® 1242 Standard 1,000 µg/mL, Hexane, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : July 31, 2030 **Storage:** 25°C nominal

Handling: This product contains PCBs. **Ship:** Ambient

P13706 } y.p.
P13707 } 10/17/24

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Aroclor 1242	53469-21-9	01141	----%	1,000.7 µg/mL	+/- 55.5295

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Hexane
CAS # 110-54-3
Purity 99%

Quality Confirmation Test

Column:

30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:

helium-constant pressure 20 psi.

Temp. Program:

200°C to 300°C
@ 25°C/min. (hold 10 min.)

Inj. Temp:

250°C

Det. Temp:

300°C

Det. Type:


ECD

Split Vent:

10 ml/min.

Inj. Vol

0.2µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Amanda Miller - Operations Tech III - ARM QC

Date Mixed: 11-Apr-2024

Balance Serial # B442140311


Christie Mills - Operations Lead Tech - ARM QC

Date Passed: 17-Apr-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.




110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32000 **Lot No.:** A0214495

Description : Pesticide Surrogate Mix

Pesticide Surrogate Mix 200 µg/mL, Acetone, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : October 31, 2030 **Storage:** 10°C or colder

Handling: Contains PCBs - sonicate prior to use. **Ship:** Ambient

P19785
AJ
11/19/24
P19789

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	2,4,5,6-Tetrachloro-m-xylene	877-09-8	RP220407	99%	200.2 µg/mL	+/- 11.1087
2	Decachlorobiphenyl (BZ# 209)	2051-24-3	30679	99%	201.4 µg/mL	+/- 11.1753

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Acetone
CAS # 67-64-1
Purity 99%

Tech Tips:

Decachlorobiphenyl has poor solubility in most organic solvents. The maximum concentration that can be prepared in acetone, hexane, or isooctane is 200µg/mL. Temperature will affect the solubility as well. Storing solutions at reduced temperatures will cause decachlorobiphenyl to precipitate.

Products containing decachlorobiphenyl must be sonicated for a minimum of 10 minutes prior to opening the ampul. Because each ultrasonic bath operates at a different energy level, 10 minutes is a guideline only. Longer sonication time will not affect product quality.

These precautions apply to working solutions prepared in your laboratory as well. The amount of compound that precipitates depends on concentration AND temperature. If you store your standards at a temperature lower than 4°C (even dilute solutions), allow extra sonication time.

Quality Confirmation Test

Column:
30m x 25mm x 2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:
helium-constant pressure 20 psi.

Temp. Program:
200°C to 300°C
@ 25°C/min. (hold 10 min.)


Inj. Temp:
250°C

Det. Temp:
300°C

Det. Type:
ECD

Split Vent:
10 ml/min.

Inj. Vol
1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

A. O. P.
Aaron Enyart - Operations Tech I

Date Mixed: 29-Jul-2024 **Balance Serial #** B345965662

Jennifer Pollino
Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 01-Aug-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397




110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32000 **Lot No.:** A0214495

Description : Pesticide Surrogate Mix

Pesticide Surrogate Mix 200 µg/mL, Acetone, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : October 31, 2030 **Storage:** 10°C or colder

Handling: Contains PCBs - sonicate prior to use. **Ship:** Ambient

P19785
AJ
11/19/24
P19789

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	2,4,5,6-Tetrachloro-m-xylene	877-09-8	RP220407	99%	200.2 µg/mL	+/- 11.1087
2	Decachlorobiphenyl (BZ# 209)	2051-24-3	30679	99%	201.4 µg/mL	+/- 11.1753

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Acetone
CAS # 67-64-1
Purity 99%

Tech Tips:

Decachlorobiphenyl has poor solubility in most organic solvents. The maximum concentration that can be prepared in acetone, hexane, or isooctane is 200µg/mL. Temperature will affect the solubility as well. Storing solutions at reduced temperatures will cause decachlorobiphenyl to precipitate.

Products containing decachlorobiphenyl must be sonicated for a minimum of 10 minutes prior to opening the ampul. Because each ultrasonic bath operates at a different energy level, 10 minutes is a guideline only. Longer sonication time will not affect product quality.

These precautions apply to working solutions prepared in your laboratory as well. The amount of compound that precipitates depends on concentration AND temperature. If you store your standards at a temperature lower than 4°C (even dilute solutions), allow extra sonication time.

Quality Confirmation Test

Column:
30m x 25mm x 2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:
helium-constant pressure 20 psi.

Temp. Program:
200°C to 300°C
@ 25°C/min. (hold 10 min.)


Inj. Temp:
250°C

Det. Temp:
300°C

Det. Type:
ECD

Split Vent:
10 ml/min.

Inj. Vol
1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

A. O. P.
Aaron Enyart - Operations Tech I

Date Mixed: 29-Jul-2024 **Balance Serial #** B345965662

Jennifer Pollino
Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 01-Aug-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32011 **Lot No.:** A0217391

Description : Aroclor® 1254 Standard

Aroclor® 1254 Standard 1,000 µg/mL, Hexane, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : January 31, 2031 **Storage:** 25°C nominal

Handling: This product contains PCBs. **Ship:** Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Aroclor 1254	11097-69-1	124-191-B	----%	1,004.7 µg/mL	+/- 55.7515

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Hexane
CAS # 110-54-3
Purity 99%

P13830
↓
P13832
↓
P13834
AJ
12/09/24

Quality Confirmation Test

Column:

30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:

helium-constant pressure 20 psi.

Temp. Program:

200°C to 300°C
@ 25°C/min. (hold 10 min.)

Inj. Temp:

250°C

Det. Temp:

300°C

Det. Type:


ECD

Split Vent:

300 ml/min.

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Michael Maye
Michael Maye - Operations Tech I

Date Mixed: 02-Oct-2024

Balance Serial # C322230531

Jennifer Pollino
Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 07-Oct-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32008

Lot No.: A0219655

Description : Aroclor® 1232 Standard

Aroclor® 1232 Standard 1,000 µg/mL, Hexane, 1mL/ampul

Container Size : 2 mL

Pkg Amt: > 1 mL

Expiration Date : March 31, 2031

Storage: 25°C nominal

Handling:

This product contains PCBs.

Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Aroclor 1232	11141-16-5	15665-01	---%	1,007.0 µg/mL	+/- 55.8810

Solvent: Hexane

CAS # 110-54-3

Purity 99%

* Expanded Uncertainty displayed in same units as Grav. Conc.

P13878



AJ
01/28/25

P13880

Quality Confirmation Test

Column:

30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:

helium-constant pressure 20 psi.

Temp. Program:

200°C to 300°C
@ 25°C/min. (hold 10 min.)

Inj. Temp:

250°C

Det. Temp:

300°C

Det. Type:


ECD

Split Vent:

10 ml/min.

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Michael Maye
Michael Maye - Operations Tech I

Date Mixed: 02-Dec-2024

Balance Serial #

C322230531

Brittany Federhik
Brittany Federhiko - Operations Tech I

Date Passed: 05-Dec-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397



CERTIFIED REFERENCE MATERIAL

110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com



Certificate of Analysis

chromatographic plus

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32409 **Lot No.:** A0217041

Description : Aroclor® 1262 Standard

Aroclor® 1262 Standard 1,000 µg/mL, 1mL/ampul, Hexane

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : December 31, 2030 **Storage:** 25°C nominal

Handling: This product contains PCBs. **Ship:** Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Aroclor 1262	37324-23-5	10849100	-----%	1,000.0 µg/mL	+/- 55.4925

Solvent: Hexane
CAS # 110-54-3
Purity 99%

* Expanded Uncertainty displayed in same units as Grav. Conc.

P13881
AJ
01128125

Quality Confirmation Test

Column:
30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:
helium-constant pressure 20 psi.

Temp. Program:
200°C to 300°C
@ 25°C/min. (hold 10 min.)


Inj. Temp:
250°C

Det. Temp:
300°C

Det. Type:
ECD

Split Vent:
300 ml/min.

Inj. Vol
0.2µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Rebecca Gingerich - Operations Tech II


Jennifer Pollino - Operations Tech III - ARM QC

Date Mixed: 25-Sep-2024 Balance Serial # C322230531

Date Passed: 30-Sep-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397




110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32410

Lot No.: A0217264

Description : Aroclor® 1268 Standard

Aroclor® 1268 Standard 1,000 µg/mL, 1mL/ampul, Hexane

Container Size : 2 mL

Pkg Amt: > 1 mL

Expiration Date : January 31, 2031

Storage: 25°C nominal

Handling:

This product contains PCBs.

Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Aroclor 1268	11100-14-4	12353400	----%	1,009.8 µg/mL	+/- 56.0364

Solvent: Hexane

CAS # 110-54-3

Purity 99%

* Expanded Uncertainty displayed in same units as Grav. Conc.

P13884

+

AJ
01/28/25

P13886

Quality Confirmation Test

Column:

30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:

helium-constant pressure 20 psi.

Temp. Program:

200°C to 300°C
@ 25°C/min. (hold 10 min.)

Inj. Temp:

250°C

Det. Temp:

300°C

Det. Type:


ECD

Split Vent:

Split ratio 500:1

Inj. Vol

0.2µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Stacey Wanner - Operations Technician I

Balance Serial # 1128360905

Date Mixed: 01-Oct-2024

Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 02-Oct-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

n-Hexane 95%
ULTRA RESI-ANALYZED
For Organic Residue Analysis

avantor™



Material No.: 9262-03
Batch No.: 24G1962003
Manufactured Date: 2024-05-23
Expiration Date: 2025-08-22
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL)	≤ 5	3
ECD Sensitive Impurities (as Heptachlor Epoxide) Single Peak (pg/mL)	≤ 10	1
ECD-Sensitive Impurities (as Ethylene Dibromide) - Single Impurity Peak (ng/mL)	≤ 5	1
Assay (Total Saturated C ₆ Isomers) (by GC, corrected for water)	≥ 99.5 %	99.7 %
Assay (as n-Hexane) (by GC, corrected for water)	≥ 95 %	98 %
Color (APHA)	≤ 10	5
Residue after Evaporation	≤ 1.0 ppm	0.1 ppm
Substances Darkened by H ₂ SO ₄	Passes Test	Passes Test
Water (by KF, coulometric)	≤ 0.05 %	< 0.01 %

For Laboratory, Research, or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: USA
Packaging Site: Phillipsburg Mfg Ctr & DC

Jamie Croak
Director Quality Operations, Bioscience Production

n-Hexane 95%
ULTRA RESI-ANALYZED
For Organic Residue Analysis

 **avantor**™



Material No.: 9262-03

Batch No.: 25C0362005

Manufactured Date: 2025-01-29

Expiration Date: 2026-04-30

Revision No.: 0

W3234 JB
OP4421. 07/28/2025

Certificate of Analysis

Test	Specification	Result
FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL)	<= 5	1
ECD Sensitive Impurities (as HeptachlorEpoxide) Single Peak (pg/mL)	<= 10	6
ECD-Sensitive Impurities (as EthyleneDibromide) - Single Impurity Peak (ng/mL)	<= 5	5
Assay (Total Saturated C ₆ Isomers) (by GC, corrected for water)	>= 99.5 %	100.0 %
Assay (as n-Hexane) (by GC, corrected for water)	>= 95 %	100 %
Color (APHA)	<= 10	10
Residue after Evaporation	<= 1.0 ppm	0.1 ppm
Substances Darkened by H ₂ SO ₄	Passes Test	Passes Test
Water (by KF, coulometric)	<= 0.05 %	<0.01 %

For Laboratory, Research, or Manufacturing Use

MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: United States

Packaging Site: Phillipsburg Mfg Ctr & DC



Jamie Croak
Director Quality Operations, Bioscience Production

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

Avantor Performance Materials LLC



SHIPPING DOCUMENTS



284 Sheffield Street, Mountainside, NJ 07092
(908) 789-8900 • Fax (908) 789-8922
www.chemtech.net

ALLIANCE PROJECT NO.

QUOTE NO.

COC Number

Q3450
2046803

CLIENT INFORMATION

REPORT TO BE SENT TO:

COMPANY: Matrix New World Eng

ADDRESS: 26 Columbia Tpke

CITY: Florham Park STATE: NJ ZIP: 07932

ATTENTION:

PHONE: 973-240-1800 FAX:

CLIENT PROJECT INFORMATION

PROJECT NAME: PSEG Athenia Substation

PROJECT NO.: 24-0935 LOCATION: Clifton

PROJECT MANAGER: Chris Pittarese

e-mail: cpittarese@empnwe.com

PHONE: 973-240-1800 FAX:

CLIENT BILLING INFORMATION

BILL TO: SAME AS CLIENT PO#: INFO

ADDRESS:

CITY

STATE:

ZIP:

ATTENTION:

PHONE:

ANALYSIS

DATA TURNAROUND INFORMATION

FAX (RUSH) DAYS*

HARDCOPY (DATA PACKAGE): DAYS*

EDD: 24HR (1 DAY) DAYS*

*TO BE APPROVED BY CHEMTECH

STANDARD HARDCOPY TURNAROUND TIME IS 10 BUSINESS

DATA DELIVERABLE INFORMATION

☐ Level 1 (Results Only) ☐ Level 4 (QC + Full Raw Data)

☐ Level 2 (Results + QC) ☒ NJ Reduced ☐ US EPA CLP

☐ Level 3 (Results + QC) ☐ NYS ASP A ☐ NYS ASP B

+ Raw Data ☐ Other

☐ EDD FORMAT

1	2	3	4	5	6	7	8	9
PCBS								

ALLIANCE SAMPLE ID	PROJECT SAMPLE IDENTIFICATION	SAMPLE MATRIX	SAMPLE TYPE		SAMPLE COLLECTION		# OF BOTTLES	PRESERVATIVES									COMMENTS	
			COMP	GRAB	DATE	TIME		1	2	3	4	5	6	7	8	9	← Specify Preservatives A-HCl D-NaOH B-HNO3 E-ICE C-H2SO4 F-OTHER	
1.	SS-5R(2.0-2.5)	S		X	10/23	1010	1	X										
2.	SS-5R(3.5-4.0)					1005	1	X										
3.	SS-10R(2.0-2.5)					1135	1	H										
4.	SS-10R(3.5-4.0)					1132	1	H										
5.	SS-13R(3.5-4.0)					0955	1	H										
6.	SS-13R(4.5-5.0)					0945	1	H										
7.	SS-13R(5.0-5.5)					0950	1	H										
8.	SS-20(0.0-0.5)					1125	1	X										
9.	SS-20(2.0-2.5)					1120	1	X										
10.	SS-20(3.5-4.0)					1117	1	X										

SAMPLE CUSTODY MUST BE DOCUMENTED BELOW EACH TIME SAMPLES CHANGE POSSESSION INCLUDING COURIER DELIVERY

RELINQUISHED BY SAMPLER: 1. Madison D. Goff 10/23/13 544	DATE/TIME: 1. 544	RECEIVED BY: 1. [Signature]	Conditions of bottles or coolers at receipt: ☐ COMPLIANT ☐ NON COMPLIANT ☐ COOLER TEMP 2.3 °C
RELINQUISHED BY SAMPLER: 2.	DATE/TIME:	RECEIVED BY: 2.	Comments: X=run H=hold 246011
RELINQUISHED BY SAMPLER: 3.	DATE/TIME:	RECEIVED BY: 3.	Page ____ of CLIENT: ☐ Hand Delivered ☐ Other Shipment Complete ☐ YES ☐ NO

CLIENT INFORMATION

CLIENT PROJECT INFORMATION

CLIENT BILLING INFORMATION

REPORT TO BE SENT TO:

COMPANY: Matrix New World Eng.
ADDRESS: 26 Columbia Tpk
CITY: Florham Park STATE: NJ ZIP: 07932
ATTENTION:
PHONE: 973-240-1800 FAX:

PROJECT NAME: SEG Athenia Substation
PROJECT NO.: 24-0935 LOCATION: Clifton
PROJECT MANAGER: Cpn's Pitta resc
e-mail: cpittaresc@mnme.com
PHONE: 973-240-1800 FAX:

BILL TO: SAME AS CLIENT PO#: INFD
ADDRESS:
CITY: STATE: ZIP:
ATTENTION: PHONE:

ANALYSIS

DATA TURNAROUND INFORMATION

DATA DELIVERABLE INFORMATION

FAX (RUSH) _____ DAYS*
HARDCOPY (DATA PACKAGE): _____ DAYS*
EDD: 24HR (1 DAY) DAYS*
*TO BE APPROVED BY CHEMTECH
STANDARD HARDCOPY TURNAROUND TIME IS 10 BUSINESS

☐ Level 1 (Results Only) ☐ Level 4 (QC + Full Raw Data)
☐ Level 2 (Results + QC) ☒ NJ Reduced ☐ US EPA CLP
☐ Level 3 (Results + QC) ☐ NYS ASP A ☐ NYS ASP B
+ Raw Data ☐ Other _____
☐ EDD FORMAT _____

1	2	3	4	5	6	7	8	9

PRESERVATIVES

COMMENTS

ALLIANCE SAMPLE ID	PROJECT SAMPLE IDENTIFICATION	SAMPLE MATRIX	SAMPLE TYPE		SAMPLE COLLECTION		# OF BOTTLES	PRESERVATIVES									COMMENTS
			COMP	GRAB	DATE	TIME		1	2	3	4	5	6	7	8	9	
1.	SS-28 (2.0-2.5)	S		2	10/23	1130	1	X									
2.	SS-29 (2.0-2.5)					1140	1	H									
3.	SS-30 (0.0-0.5)					1155	1	H									
4.	SS-30 (2.0-2.5)					1150	1	H									
5.	SS-30 (3.5-4.0)					1145	1	H									
6.	SS-31 (0.0-0.5)					1115	1	H									
7.	SS-34 (0.0-0.5)					0925	1	H									
8.	SS-34 (2.0-2.5)					0930	1	H									
9.	SS-34 (3.5-4.0)					0935	1	H									
10.	SS-21 (0.0-0.5)					1335	1	X									

SAMPLE CUSTODY MUST BE DOCUMENTED BELOW EACH TIME SAMPLES CHANGE POSSESSION INCLUDING COURIER DELIVERY

RELINQUISHED BY SAMPLER:	DATE/TIME:	RECEIVED BY:	Conditions of bottles or coolers at receipt: ☐ COMPLIANT ☐ NON COMPLIANT ☐ COOLER TEMP
1. <u>Madison Tigan</u>	<u>10/23 154</u>	1. <u>[Signature]</u>	Comments: <u>X=run H=hold</u> <u>2.3°C</u>
RELINQUISHED BY SAMPLER:	DATE/TIME:	RECEIVED BY:	
2.		2.	
RELINQUISHED BY SAMPLER:	DATE/TIME:	RECEIVED BY:	
3.		3.	

Page ____ of CLIENT: ☐ Hand Delivered ☐ Other Shipment Complete ☐ YES ☐ NO

CLIENT INFORMATION

CLIENT PROJECT INFORMATION

CLIENT BILLING INFORMATION

REPORT TO BE SENT TO:

COMPANY: Matrix New World Eng

ADDRESS: 26 Columbia Tpke

CITY: Florsham Park STATE: NJ ZIP: 07922

ATTENTION:

PHONE: 973-240-1800 FAX:

PROJECT NAME: BEG Athenia Substation

PROJECT NO: 24-0935 LOCATION: Clifton

PROJECT MANAGER: Chris Pittarese

e-mail: cpittarese@mnne.com

PHONE: 973-240-1800 FAX:

BILL TO: SAME AS CLIENT PO#: INV#

ADDRESS:

CITY: STATE: ZIP:

ATTENTION: PHONE:

ANALYSIS

DATA TURNAROUND INFORMATION

FAX (RUSH) DAYS*

HARDCOPY (DATA PACKAGE): DAYS*

EDD: 24HR (1 DAY) DAYS*

*TO BE APPROVED BY CHEMTECH

STANDARD HARDCOPY TURNAROUND TIME IS 10 BUSINESS

DATA DELIVERABLE INFORMATION

☐ Level 1 (Results Only) ☐ Level 4 (QC + Full Raw Data)

☐ Level 2 (Results + QC) ☒ NJ Reduced ☐ US EPA CLP

☐ Level 3 (Results + QC) ☐ NYS ASP A ☐ NYS ASP B
+ Raw Data ☐ Other

☐ EDD FORMAT

PCBS
1 2 3 4 5 6 7 8 9

PRESERVATIVES

COMMENTS

ALLIANCE
SAMPLE
ID

PROJECT
SAMPLE IDENTIFICATION

SAMPLE
MATRIX

SAMPLE
TYPE

SAMPLE
COLLECTION

OF BOTTLES

COMP

GRAB

DATE

TIME

1

2

3

4

5

6

7

8

9

← Specify Preservatives
A-HCl D-NaOH
B-HNO3 E-ICE
C-H2SO4 F-OTHER

1.

SS-21 (20-25)

S

X

10/23

1333

1

X

2.

SS-21 (35-40)

1330

1

X

3.

SS-22 (0.0-05)

1340

1

H

4.

SS-22 (20-25)

1342

1

H

5.

SS-22 (35-40)

1338

1

H

6.

7.

8.

9.

10.

SAMPLE CUSTODY MUST BE DOCUMENTED BELOW EACH TIME SAMPLES CHANGE POSSESSION INCLUDING COURIER DELIVERY

RELINQUISHED BY SAMPLER:

DATE/TIME:

RECEIVED BY:

1. [Signature]

10/23 1544

1. [Signature]

Conditions of bottles or coolers at receipt: ☐ COMPLIANT ☐ NON COMPLIANT ☐ COOLER TEMP

Comments: X=run H=hold

2-3 °C

RELINQUISHED BY SAMPLER:

DATE/TIME:

RECEIVED BY:

2.

2.

RELINQUISHED BY SAMPLER:

DATE/TIME:

RECEIVED BY:

3.

3.

Page ____ of ____

CLIENT: ☐ Hand Delivered ☐ Other

Shipment Complete

☐ YES ☐ NO

[Signature]

Laboratory Certification

Certified By	License No.
Connecticut	PH-0830
DOD ELAP (ANAB)	L2219
Maine	2024021
Maryland	296
New Hampshire	255425
New Jersey	20012
New York	11376
Pennsylvania	68-00548
Soil Permit	525-24-234-08441
Texas	TX-C25-00189
Virginia	460312