

### CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: MATR02

Lab Code: ACE

SDG NO.: Q3450

Continuing Calib Date: 10/31/2025

Initial Calibration Date(s): 10/21/2025

10/21/2025

Continuing Calib Time: 16:04

Initial Calibration Time(s): 09:06

17:01

GC Column: ZB-MR1 ID: 0.32 (mm)

| COMPOUND             |     | CCAL<br>RT | AVG<br>RT | RT WINDOW |       | DIFF<br>RT |
|----------------------|-----|------------|-----------|-----------|-------|------------|
|                      |     |            |           | FROM      | TO    |            |
| Aroclor-1016-1       | (1) | 5.80       | 5.80      | 5.70      | 5.90  | 0.00       |
| Aroclor-1016-2       | (2) | 5.82       | 5.82      | 5.72      | 5.92  | 0.00       |
| Aroclor-1016-3       | (3) | 5.88       | 5.88      | 5.78      | 5.98  | 0.00       |
| Aroclor-1016-4       | (4) | 5.98       | 5.98      | 5.88      | 6.08  | 0.00       |
| Aroclor-1016-5       | (5) | 6.27       | 6.27      | 6.17      | 6.37  | 0.00       |
| Aroclor-1260-1       | (1) | 7.39       | 7.39      | 7.29      | 7.49  | 0.00       |
| Aroclor-1260-2       | (2) | 7.64       | 7.64      | 7.54      | 7.74  | 0.00       |
| Aroclor-1260-3       | (3) | 8.00       | 8.00      | 7.90      | 8.10  | 0.00       |
| Aroclor-1260-4       | (4) | 8.23       | 8.23      | 8.13      | 8.33  | 0.00       |
| Aroclor-1260-5       | (5) | 8.56       | 8.55      | 8.45      | 8.65  | -0.01      |
| Tetrachloro-m-xylene |     | 4.65       | 4.64      | 4.54      | 4.74  | -0.01      |
| Decachlorobiphenyl   |     | 10.43      | 10.42     | 10.32     | 10.52 | -0.01      |

### CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: MATR02

Lab Code: ACE

SDG NO.: Q3450

Continuing Calib Date: 10/31/2025

Initial Calibration Date(s): 10/21/2025

10/21/2025

Continuing Calib Time: 16:04

Initial Calibration Time(s): 09:06

17:01

GC Column: ZB-MR2 ID: 0.32 (mm)

| COMPOUND             |     | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|----------------------|-----|------------|-----------|-----------|------|------------|
|                      |     |            |           | FROM      | TO   |            |
| Aroclor-1016-1       | (1) | 4.88       | 4.88      | 4.78      | 4.98 | 0.00       |
| Aroclor-1016-2       | (2) | 4.93       | 4.94      | 4.84      | 5.04 | 0.01       |
| Aroclor-1016-3       | (3) | 5.05       | 5.05      | 4.95      | 5.15 | 0.00       |
| Aroclor-1016-4       | (4) | 5.10       | 5.10      | 5.00      | 5.20 | 0.00       |
| Aroclor-1016-5       | (5) | 5.31       | 5.31      | 5.21      | 5.41 | 0.00       |
| Aroclor-1260-1       | (1) | 6.34       | 6.34      | 6.24      | 6.44 | 0.00       |
| Aroclor-1260-2       | (2) | 6.52       | 6.53      | 6.43      | 6.63 | 0.01       |
| Aroclor-1260-3       | (3) | 6.68       | 6.68      | 6.58      | 6.78 | 0.00       |
| Aroclor-1260-4       | (4) | 7.14       | 7.15      | 7.05      | 7.25 | 0.01       |
| Aroclor-1260-5       | (5) | 7.39       | 7.39      | 7.29      | 7.49 | 0.00       |
| Tetrachloro-m-xylene |     | 3.78       | 3.78      | 3.68      | 3.88 | 0.00       |
| Decachlorobiphenyl   |     | 8.78       | 8.78      | 8.68      | 8.88 | 0.00       |