

### CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: MATR02

Lab Code: ACE

SDG NO.: Q3450

Continuing Calib Date: 11/03/2025

Initial Calibration Date(s): 10/21/2025 10/21/2025

Continuing Calib Time: 09:21

Initial Calibration Time(s): 09:06 17:01

GC Column: ZB-MR1 ID: 0.32 (mm)

| COMPOUND             | CCAL<br>RT | AVG<br>RT | RT WINDOW<br>FROM   TO |       | DIFF<br>RT |
|----------------------|------------|-----------|------------------------|-------|------------|
| Aroclor-1248-1 (1)   | 5.79       | 5.80      | 5.70                   | 5.90  | 0.01       |
| Aroclor-1248-2 (2)   | 6.06       | 6.07      | 5.97                   | 6.17  | 0.01       |
| Aroclor-1248-3 (3)   | 6.27       | 6.27      | 6.17                   | 6.37  | 0.00       |
| Aroclor-1248-4 (4)   | 6.67       | 6.67      | 6.57                   | 6.77  | 0.01       |
| Aroclor-1248-5 (5)   | 6.70       | 6.71      | 6.61                   | 6.81  | 0.01       |
| Tetrachloro-m-xylene | 4.64       | 4.65      | 4.55                   | 4.75  | 0.01       |
| Decachlorobiphenyl   | 10.41      | 10.42     | 10.32                  | 10.52 | 0.01       |

### CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: MATR02

Lab Code: ACE

SDG NO.: Q3450

Continuing Calib Date: 11/03/2025

Initial Calibration Date(s): 10/21/2025

10/21/2025

Continuing Calib Time: 09:21

Initial Calibration Time(s): 09:06

17:01

GC Column: ZB-MR2 ID: 0.32 (mm)

| COMPOUND             | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|----------------------|------------|-----------|-----------|------|------------|
|                      |            |           | FROM      | TO   |            |
| Aroclor-1248-1 (1)   | 4.87       | 4.88      | 4.78      | 4.98 | 0.01       |
| Aroclor-1248-2 (2)   | 5.09       | 5.10      | 5.00      | 5.20 | 0.01       |
| Aroclor-1248-3 (3)   | 5.13       | 5.14      | 5.04      | 5.24 | 0.01       |
| Aroclor-1248-4 (4)   | 5.31       | 5.31      | 5.21      | 5.41 | 0.00       |
| Aroclor-1248-5 (5)   | 5.70       | 5.70      | 5.60      | 5.80 | 0.00       |
| Tetrachloro-m-xylene | 3.78       | 3.78      | 3.68      | 3.88 | 0.00       |
| Decachlorobiphenyl   | 8.78       | 8.78      | 8.68      | 8.88 | 0.01       |