

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** MATR02
Lab Code: ACE **SDG NO.:** Q3450
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/04/2025 11/04/2025

Client Sample No.: CCAL02 **Date Analyzed:** 11/12/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO115046.D **Time Analyzed:** 15:55

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.487	5.387	5.587	486.740	500.000	-2.7
Aroclor-1016-2	5.509	5.408	5.608	467.970	500.000	-6.4
Aroclor-1016-3	5.570	5.470	5.670	485.340	500.000	-2.9
Aroclor-1016-4	5.666	5.566	5.766	460.290	500.000	-7.9
Aroclor-1016-5	5.958	5.858	6.058	456.330	500.000	-8.7
Aroclor-1260-1	7.071	6.971	7.171	467.020	500.000	-6.6
Aroclor-1260-2	7.325	7.225	7.425	495.090	500.000	-1.0
Aroclor-1260-3	7.682	7.582	7.782	497.620	500.000	-0.5
Aroclor-1260-4	7.904	7.803	8.003	484.630	500.000	-3.1
Aroclor-1260-5	8.212	8.110	8.310	487.270	500.000	-2.5
Decachlorobiphenyl	9.932	9.830	10.030	48.700	50.000	-2.6
Tetrachloro-m-xylene	4.347	4.248	4.448	49.330	50.000	-1.3

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** MATR02
Lab Code: ACE **SDG NO.:** Q3450
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/04/2025 11/04/2025

Client Sample No.: CCAL02 **Date Analyzed:** 11/12/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO115046.D **Time Analyzed:** 15:55

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.726	4.628	4.828	463.320	500.000	-7.3
Aroclor-1016-2	4.746	4.647	4.847	469.030	500.000	-6.2
Aroclor-1016-3	4.920	4.821	5.021	468.660	500.000	-6.3
Aroclor-1016-4	4.962	4.864	5.064	475.910	500.000	-4.8
Aroclor-1016-5	5.175	5.076	5.276	483.800	500.000	-3.2
Aroclor-1260-1	6.204	6.106	6.306	453.970	500.000	-9.2
Aroclor-1260-2	6.392	6.294	6.494	457.030	500.000	-8.6
Aroclor-1260-3	6.545	6.446	6.646	463.220	500.000	-7.4
Aroclor-1260-4	7.016	6.917	7.117	464.740	500.000	-7.1
Aroclor-1260-5	7.257	7.158	7.358	467.270	500.000	-6.5
Decachlorobiphenyl	8.640	8.541	8.741	50.800	50.000	1.6
Tetrachloro-m-xylene	3.648	3.548	3.748	47.940	50.000	-4.1