

**CALIBRATION VERIFICATION SUMMARY**

**Lab Name:** Alliance **Contract:** MATR02  
**Lab Code:** ACE **SDG NO.:** Q3450  
**GC Column:** ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/21/2025 10/21/2025

**Client Sample No.:** CCAL09 **Date Analyzed:** 11/03/2025  
**Lab Sample No.:** AR1660CCC500 **Data File :** PP076292.D **Time Analyzed:** 08:44

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.793	5.695	5.895	549.830	500.000	10.0
Aroclor-1016-2	5.815	5.717	5.917	540.140	500.000	8.0
Aroclor-1016-3	5.877	5.780	5.980	546.540	500.000	9.3
Aroclor-1016-4	5.974	5.877	6.077	550.230	500.000	10.0
Aroclor-1016-5	6.267	6.169	6.369	532.730	500.000	6.5
Aroclor-1260-1	7.384	7.287	7.487	580.700	500.000	16.1
Aroclor-1260-2	7.637	7.540	7.740	549.670	500.000	9.9
Aroclor-1260-3	7.995	7.899	8.099	576.350	500.000	15.3
Aroclor-1260-4	8.223	8.127	8.327	597.300	500.000	19.5
Aroclor-1260-5	8.547	8.452	8.652	550.880	500.000	10.2
Decachlorobiphenyl	10.411	10.319	10.519	52.840	50.000	5.7
Tetrachloro-m-xylene	4.641	4.542	4.742	54.140	50.000	8.3

**CALIBRATION VERIFICATION SUMMARY**

**Lab Name:** Alliance **Contract:** MATR02  
**Lab Code:** ACE **SDG NO.:** Q3450  
**GC Column:** ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/21/2025 10/21/2025

**Client Sample No.:** CCAL09 **Date Analyzed:** 11/03/2025  
**Lab Sample No.:** AR1660CCC500 **Data File :** PP076292.D **Time Analyzed:** 08:44

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.875	4.778	4.978	535.300	500.000	7.1
Aroclor-1016-2	4.932	4.836	5.036	535.190	500.000	7.0
Aroclor-1016-3	5.052	4.954	5.154	481.690	500.000	-3.7
Aroclor-1016-4	5.092	4.996	5.196	565.650	500.000	13.1
Aroclor-1016-5	5.307	5.210	5.410	469.810	500.000	-6.0
Aroclor-1260-1	6.334	6.238	6.438	553.960	500.000	10.8
Aroclor-1260-2	6.520	6.425	6.625	558.660	500.000	11.7
Aroclor-1260-3	6.673	6.578	6.778	580.770	500.000	16.2
Aroclor-1260-4	7.141	7.047	7.247	631.580	500.000	26.3
Aroclor-1260-5	7.382	7.287	7.487	589.330	500.000	17.9
Decachlorobiphenyl	8.774	8.682	8.882	51.320	50.000	2.6
Tetrachloro-m-xylene	3.780	3.681	3.881	57.770	50.000	15.5