

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO110425

Review By	yogesh	Review On	11/4/2025 12:15:43 PM
Supervise By	rahul	Supervise On	11/5/2025 8:40:26 AM
SubDirectory	PO110425	HP Acquire Method	HP Processing Method PO110425
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,P P24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP 24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854		
CCC			
Internal Standard/PEM			
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	Sampleld	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PO114869.D	04 Nov 2025 09:03		YP/AJ	Ok
2	I.BLK	I.BLK	PO114870.D	04 Nov 2025 09:21		YP/AJ	Ok
3	AR1660ICC1000	AR1660ICC1000	PO114871.D	04 Nov 2025 09:39		YP/AJ	Ok
4	AR1660ICC750	AR1660ICC750	PO114872.D	04 Nov 2025 09:58		YP/AJ	Ok
5	AR1660ICC500	AR1660ICC500	PO114873.D	04 Nov 2025 10:16		YP/AJ	Ok
6	AR1660ICC250	AR1660ICC250	PO114874.D	04 Nov 2025 10:34		YP/AJ	Ok,M
7	AR1660ICC050	AR1660ICC050	PO114875.D	04 Nov 2025 10:53		YP/AJ	Ok,M
8	AR1221ICC500	AR1221ICC500	PO114876.D	04 Nov 2025 11:11		YP/AJ	Ok
9	AR1232ICC500	AR1232ICC500	PO114877.D	04 Nov 2025 11:30		YP/AJ	Ok
10	AR1242ICC1000	AR1242ICC1000	PO114878.D	04 Nov 2025 11:48		YP/AJ	Ok
11	AR1242ICC750	AR1242ICC750	PO114879.D	04 Nov 2025 12:06		YP/AJ	Ok
12	AR1242ICC500	AR1242ICC500	PO114880.D	04 Nov 2025 12:25		YP/AJ	Ok
13	AR1242ICC250	AR1242ICC250	PO114881.D	04 Nov 2025 12:42		YP/AJ	Ok
14	AR1242ICC050	AR1242ICC050	PO114882.D	04 Nov 2025 13:00		YP/AJ	Ok,M
15	AR1248ICC1000	AR1248ICC1000	PO114883.D	04 Nov 2025 13:19		YP/AJ	Ok
16	AR1248ICC750	AR1248ICC750	PO114884.D	04 Nov 2025 13:37		YP/AJ	Ok
17	AR1248ICC500	AR1248ICC500	PO114885.D	04 Nov 2025 13:56		YP/AJ	Ok
18	AR1248ICC250	AR1248ICC250	PO114886.D	04 Nov 2025 14:14		YP/AJ	Ok

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO110425

Review By	yogesh	Review On	11/4/2025 12:15:43 PM
Supervise By	rahul	Supervise On	11/5/2025 8:40:26 AM
SubDirectory	PO110425	HP Acquire Method	HP Processing Method PO110425
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,P P24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP 24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862		

19	AR1248ICC050	AR1248ICC050	PO114887.D	04 Nov 2025 14:33		YP/AJ	Ok
20	AR1254ICC1000	AR1254ICC1000	PO114888.D	04 Nov 2025 14:51		YP/AJ	Ok
21	AR1254ICC750	AR1254ICC750	PO114889.D	04 Nov 2025 15:09		YP/AJ	Ok
22	AR1254ICC500	AR1254ICC500	PO114890.D	04 Nov 2025 15:28		YP/AJ	Ok
23	AR1254ICC250	AR1254ICC250	PO114891.D	04 Nov 2025 16:04		YP/AJ	Ok
24	AR1254ICC050	AR1254ICC050	PO114892.D	04 Nov 2025 16:23		YP/AJ	Ok,M
25	AR1262ICC500	AR1262ICC500	PO114893.D	04 Nov 2025 17:00		YP/AJ	Ok
26	AR1268ICC1000	AR1268ICC1000	PO114894.D	04 Nov 2025 17:18		YP/AJ	Ok
27	AR1268ICC750	AR1268ICC750	PO114895.D	04 Nov 2025 17:36		YP/AJ	Ok
28	AR1268ICC500	AR1268ICC500	PO114896.D	04 Nov 2025 17:55		YP/AJ	Ok
29	AR1268ICC250	AR1268ICC250	PO114897.D	04 Nov 2025 18:13		YP/AJ	Ok,M
30	AR1268ICC050	AR1268ICC050	PO114898.D	04 Nov 2025 18:31		YP/AJ	Ok
31	PO110425ICV500	ICVPO110425	PO114899.D	04 Nov 2025 19:08		YP/AJ	Ok
32	AR1242ICV500	ICVPO110425AR1242	PO114900.D	04 Nov 2025 20:21		YP/AJ	Ok,M
33	AR1248ICV500	ICVPO110425AR1248	PO114901.D	04 Nov 2025 20:58		YP/AJ	Ok
34	AR1254ICV500	ICVPO110425AR1254	PO114902.D	04 Nov 2025 21:35		YP/AJ	Ok
35	AR1268ICV500	ICVPO110425AR1268	PO114903.D	04 Nov 2025 22:12		YP/AJ	Ok
36	DDT ANALOG	DDT ANALOG	PO114904.D	04 Nov 2025 22:48		YP/AJ	Ok

M : Manual Integration

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO111225

Review By	rahul	Review On	11/13/2025 1:15:59 PM
Supervise By	yogesh	Supervise On	11/13/2025 1:22:24 PM
SubDirectory	PO111225	HP Acquire Method	HP Processing Method PP102125

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854
CCC	PP24824,PP24828,PP24832,PP24836,PP24840,PP24844,PP24852
Internal Standard/PEM	
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	Sampleld	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PO115030.D	12 Nov 2025 09:31		YP/AJ	Ok
2	AR1660CCC500	AR1660CCC500	PO115031.D	12 Nov 2025 09:49		YP/AJ	Ok
3	AR1242CCC500	AR1242CCC500	PO115032.D	12 Nov 2025 10:08		YP/AJ	Ok
4	AR1248CCC500	AR1248CCC500	PO115033.D	12 Nov 2025 10:27		YP/AJ	Ok
5	AR1254CCC500	AR1254CCC500	PO115034.D	12 Nov 2025 10:45		YP/AJ	Ok
6	I.BLK	I.BLK	PO115035.D	12 Nov 2025 11:03		YP/AJ	Ok
7	Q3561-01DL	1027-A-DDL	PO115036.D	12 Nov 2025 11:22	TCMX high in 2nd column , AR1248 hit	YP/AJ	Ok,M
8	PB170497BL	PB170497BL	PO115037.D	12 Nov 2025 11:58		YP/AJ	Ok
9	PB170497BS	PB170497BS	PO115038.D	12 Nov 2025 12:16		YP/AJ	Ok
10	Q3597-01	TP-1	PO115039.D	12 Nov 2025 12:34		YP/AJ	Ok
11	Q3597-01MS	TP-1MS	PO115040.D	12 Nov 2025 12:53		YP/AJ	Ok
12	Q3597-01MSD	TP-1MSD	PO115041.D	12 Nov 2025 13:11		YP/AJ	Ok
13	Q3597-04	TP-2	PO115042.D	12 Nov 2025 13:29		YP/AJ	Ok
14	Q3597-07	TP-3	PO115043.D	12 Nov 2025 13:48		YP/AJ	Ok
15	Q3597-10	TP-4	PO115044.D	12 Nov 2025 14:06		YP/AJ	Ok
16	Q3450-07	SS-13R-(4.5-5.0)	PO115045.D	12 Nov 2025 15:36	AR1254 Hit	YP/AJ	Ok
17	AR1660CCC500	AR1660CCC500	PO115046.D	12 Nov 2025 15:55		YP/AJ	Ok,NR

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO111225

Review By	rahul	Review On	11/13/2025 1:15:59 PM
Supervise By	yogesh	Supervise On	11/13/2025 1:22:24 PM
SubDirectory	PO111225	HP Acquire Method	HP Processing Method PP102125
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,P24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854		
CCC	PP24824,PP24828,PP24832,PP24836,PP24840,PP24844,PP24852		
Internal Standard/PEM			
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

18	AR1242CCC500	AR1242CCC500	PO115047.D	12 Nov 2025 16:13		YP/AJ	Ok
19	AR1248CCC500	AR1248CCC500	PO115048.D	12 Nov 2025 16:31		YP/AJ	Ok
20	AR1254CCC500	AR1254CCC500	PO115049.D	12 Nov 2025 16:50		YP/AJ	Ok
21	I.BLK	I.BLK	PO115050.D	12 Nov 2025 17:07		YP/AJ	Ok
22	PB170499BL	PB170499BL	PO115051.D	12 Nov 2025 17:26		YP/AJ	Ok
23	PB170499BS	PB170499BS	PO115052.D	12 Nov 2025 17:44		YP/AJ	Ok
24	Q3591-01	KMH4680-END-1-1	PO115053.D	12 Nov 2025 18:02		YP/AJ	Ok,M
25	Q3591-02	KMH4680-END-1-2	PO115054.D	12 Nov 2025 18:21		YP/AJ	Ok
26	Q3591-03	M4-END-A	PO115055.D	12 Nov 2025 18:39	AR1254 Hit	YP/AJ	Ok,M
27	Q3591-04	M4-END-B	PO115056.D	12 Nov 2025 18:58	AR1254 Hit	YP/AJ	Ok,M
28	Q3591-05	M5-END-A	PO115057.D	12 Nov 2025 19:16	AR1254 Hit	YP/AJ	Ok,M
29	Q3591-06	M5-END-B	PO115058.D	12 Nov 2025 19:34	AR1254 Hit	YP/AJ	Ok,M
30	Q3596-01	11-7-25-A	PO115059.D	12 Nov 2025 19:53	AR1242+1254 Hit	YP/AJ	Ok,M
31	Q3596-02	11-7-25-B	PO115060.D	12 Nov 2025 20:11	AR1242 Hit	YP/AJ	Ok,M
32	AR1660CCC500	AR1660CCC500	PO115061.D	12 Nov 2025 21:25		YP/AJ	Ok
33	AR1242CCC500	AR1242CCC500	PO115062.D	12 Nov 2025 22:19		YP/AJ	Ok
34	AR1248CCC500	AR1248CCC500	PO115063.D	12 Nov 2025 22:38		YP/AJ	Ok
35	AR1254CCC500	AR1254CCC500	PO115064.D	12 Nov 2025 22:56		YP/AJ	Ok
36	I.BLK	I.BLK	PO115065.D	12 Nov 2025 23:13		YP/AJ	Ok

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO111225

Review By	rahul	Review On	11/13/2025 1:15:59 PM		
Supervise By	yogesh	Supervise On	11/13/2025 1:22:24 PM		
SubDirectory	PO111225	HP Acquire Method	HP Processing Method	PP102125	
STD. NAME	STD REF.#				
Tune/Reschk					
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,P P24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP 24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854				
CCC	PP24824,PP24828,PP24832,PP24836,PP24840,PP24844,PP24852				
Internal Standard/PEM					
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862				
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

M : Manual Integration

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP102125

Review By	yogesh	Review On	10/21/2025 12:01:15 PM
Supervise By	mohammad	Supervise On	10/25/2025 7:28:37 AM
SubDirectory	PP102125	HP Acquire Method	HP Processing Method PP102125

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,P P24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP 24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854
CCC	
Internal Standard/PEM	
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	Sampleld	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PP075886.D	21 Oct 2025 08:33		YPIAJ	Ok
2	I.BLK	I.BLK	PP075887.D	21 Oct 2025 08:49		YPIAJ	Ok,M
3	AR1660ICC1000	AR1660ICC1000	PP075888.D	21 Oct 2025 09:06		YPIAJ	Ok
4	AR1660ICC750	AR1660ICC750	PP075889.D	21 Oct 2025 09:22		YPIAJ	Ok
5	AR1660ICC500	AR1660ICC500	PP075890.D	21 Oct 2025 09:38		YPIAJ	Ok
6	AR1660ICC250	AR1660ICC250	PP075891.D	21 Oct 2025 09:54		YPIAJ	Ok
7	AR1660ICC050	AR1660ICC050	PP075892.D	21 Oct 2025 10:11		YPIAJ	Ok,M
8	AR1221ICC500	AR1221ICC500	PP075893.D	21 Oct 2025 10:27		YPIAJ	Ok
9	AR1232ICC500	AR1232ICC500	PP075894.D	21 Oct 2025 10:43		YPIAJ	Ok,M
10	AR1242ICC1000	AR1242ICC1000	PP075895.D	21 Oct 2025 10:59		YPIAJ	Ok,M
11	AR1242ICC750	AR1242ICC750	PP075896.D	21 Oct 2025 11:16		YPIAJ	Ok,M
12	AR1242ICC500	AR1242ICC500	PP075897.D	21 Oct 2025 11:48		YPIAJ	Ok,M
13	AR1242ICC250	AR1242ICC250	PP075898.D	21 Oct 2025 12:04		YPIAJ	Ok,M
14	AR1242ICC050	AR1242ICC050	PP075899.D	21 Oct 2025 12:21		YPIAJ	Ok,M
15	AR1248ICC1000	AR1248ICC1000	PP075900.D	21 Oct 2025 12:41		YPIAJ	Ok,M
16	AR1248ICC750	AR1248ICC750	PP075901.D	21 Oct 2025 12:57		YPIAJ	Ok,M
17	AR1248ICC500	AR1248ICC500	PP075902.D	21 Oct 2025 13:13		YPIAJ	Ok,M
18	AR1248ICC250	AR1248ICC250	PP075903.D	21 Oct 2025 13:30		YPIAJ	Ok,M

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP102125

Review By	yogesh	Review On	10/21/2025 12:01:15 PM
Supervise By	mohammad	Supervise On	10/25/2025 7:28:37 AM
SubDirectory	PP102125	HP Acquire Method	HP Processing Method PP102125
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854 PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862		

19	AR1248ICC050	AR1248ICC050	PP075904.D	21 Oct 2025 13:46		YPIAJ	Ok,M
20	AR1254ICC1000	AR1254ICC1000	PP075905.D	21 Oct 2025 14:02		YPIAJ	Ok,M
21	AR1254ICC750	AR1254ICC750	PP075906.D	21 Oct 2025 14:18		YPIAJ	Ok,M
22	AR1254ICC500	AR1254ICC500	PP075907.D	21 Oct 2025 14:35		YPIAJ	Ok,M
23	AR1254ICC250	AR1254ICC250	PP075908.D	21 Oct 2025 14:51		YPIAJ	Ok,M
24	AR1254ICC050	AR1254ICC050	PP075909.D	21 Oct 2025 15:07		YPIAJ	Ok,M
25	AR1262ICC500	AR1262ICC500	PP075910.D	21 Oct 2025 15:40		YPIAJ	Ok
26	AR1268ICC1000	AR1268ICC1000	PP075911.D	21 Oct 2025 15:56		YPIAJ	Ok
27	AR1268ICC750	AR1268ICC750	PP075912.D	21 Oct 2025 16:12		YPIAJ	Ok
28	AR1268ICC500	AR1268ICC500	PP075913.D	21 Oct 2025 16:29		YPIAJ	Ok
29	AR1268ICC250	AR1268ICC250	PP075914.D	21 Oct 2025 16:45		YPIAJ	Ok
30	AR1268ICC050	AR1268ICC050	PP075915.D	21 Oct 2025 17:01		YPIAJ	Ok
31	PP102125ICV500	ICVPP102125	PP075916.D	21 Oct 2025 17:34		YPIAJ	Ok
32	AR1242ICV500	ICVPP102125AR1242	PP075917.D	21 Oct 2025 18:06		YPIAJ	Ok,M
33	AR1248ICV500	ICVPP102125AR1248	PP075918.D	21 Oct 2025 18:39		YPIAJ	Ok,M
34	AR1254ICV500	ICVPP102125AR1254	PP075919.D	22 Oct 2025 08:49		YPIAJ	Ok,M
35	AR1268ICV500	ICVPP102125AR1268	PP075920.D	22 Oct 2025 09:07		YPIAJ	Ok
36	DDT ANALOG	DDT ANALOG	PP075921.D	22 Oct 2025 09:25		YPIAJ	Ok

M : Manual Integration

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP102425

Review By	yogesh	Review On	10/24/2025 11:09:51 AM
Supervise By	mohammad	Supervise On	10/28/2025 6:10:31 AM
SubDirectory	PP102425	HP Acquire Method	HP Processing Method PP102125

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,P P24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP 24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854
CCC	PP24824,PP24828,PP24832,PP24836,PP24840,PP24844,PP24852
Internal Standard/PEM	
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PP076011.D	24 Oct 2025 08:41		YPIAJ	Ok
2	AR1660CCC500	AR1660CCC500	PP076012.D	24 Oct 2025 08:57		YPIAJ	Ok
3	AR1242CCC500	AR1242CCC500	PP076013.D	24 Oct 2025 09:29		YPIAJ	Ok,M
4	AR1248CCC500	AR1248CCC500	PP076014.D	24 Oct 2025 09:45		YPIAJ	Ok,M
5	AR1254CCC500	AR1254CCC500	PP076015.D	24 Oct 2025 10:02		YPIAJ	Ok,M
6	I.BLK	I.BLK	PP076016.D	24 Oct 2025 10:18		YPIAJ	Ok
7	PB170234BL	PB170234BL	PP076017.D	24 Oct 2025 10:34		YPIAJ	Ok
8	PB170234BS	PB170234BS	PP076018.D	24 Oct 2025 10:51		YPIAJ	Ok
9	Q3437-03	SS-15R(2.0-2.5)	PP076019.D	24 Oct 2025 11:07	AR1254 hits	YPIAJ	Ok
10	Q3437-05	SS-26(0-0.5)	PP076020.D	24 Oct 2025 11:23	AR1254 hits	YPIAJ	Ok,M
11	Q3437-11	SS-27(0-0.5)	PP076021.D	24 Oct 2025 11:39	AR1254 hits	YPIAJ	Ok,M
12	Q3445-01	EO-01-102325	PP076022.D	24 Oct 2025 11:56		YPIAJ	Ok
13	Q3446-01	SB-14	PP076023.D	24 Oct 2025 12:12	F Flag in TCMX in 1st column, All surrogate fail, AR1242 hits, Need 20X	YPIAJ	Dilution
14	Q3446-02	SB-15	PP076024.D	24 Oct 2025 12:28	AR1242 hits, Need 20X	YPIAJ	Dilution
15	Q3414-12DL2	SB-12DL2	PP076025.D	24 Oct 2025 12:44	AR1242 hit	YPIAJ	Ok
16	Q3414-13DL2	SB-13DL2	PP076026.D	24 Oct 2025 13:01	AR1242 hit	YPIAJ	Ok
17	AR1660CCC500	AR1660CCC500	PP076027.D	24 Oct 2025 14:06		YPIAJ	Ok

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP102425

Review By	yogesh	Review On	10/24/2025 11:09:51 AM
Supervise By	mohammad	Supervise On	10/28/2025 6:10:31 AM
SubDirectory	PP102425	HP Acquire Method	HP Processing Method PP102125
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854		
CCC	PP24824,PP24828,PP24832,PP24836,PP24840,PP24844,PP24852		
Internal Standard/PEM			
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

18	AR1242CCC500	AR1242CCC500	PP076028.D	24 Oct 2025 14:22		YPIAJ	Ok,M
19	AR1248CCC500	AR1248CCC500	PP076029.D	24 Oct 2025 14:38		YPIAJ	Ok,M
20	AR1254CCC500	AR1254CCC500	PP076030.D	24 Oct 2025 14:54		YPIAJ	Ok,M
21	I.BLK	I.BLK	PP076031.D	24 Oct 2025 15:11		YPIAJ	Ok
22	Q3414-07DL2	SB-7DL2	PP076032.D	24 Oct 2025 15:27	AR1242 hit	YPIAJ	Ok
23	Q3446-06	SB-19	PP076033.D	24 Oct 2025 15:43	F Flag in TCMX in 1st column , AR1242+1260 Hit , TCMX high in both column , DCB low in 1st column , Need dilution	YPIAJ	Dilution
24	Q3446-07	SB-20	PP076034.D	24 Oct 2025 16:00	AR1260 Hit , Need dilution , TCMX high in 2nd column	YPIAJ	Dilution
25	Q3446-08	SB-21	PP076035.D	24 Oct 2025 16:16	AR1242+1260 Hit , DCB low in both column , TCMX high in both column , need dilution	YPIAJ	Dilution
26	Q3446-03	SB-16	PP076036.D	24 Oct 2025 16:32	AR1260 Hit , TCMX high in 2nd column , need dilution	YPIAJ	Dilution
27	Q3446-04	SB-17	PP076037.D	24 Oct 2025 16:48	AR1260 Hit , TCMX high in 2nd column , need dilution	YPIAJ	Dilution
28	Q3446-01DL	SB-14DL	PP076038.D	24 Oct 2025 17:05	AR1242 hit	YPIAJ	Ok,M
29	Q3446-02DL	SB-15DL	PP076039.D	24 Oct 2025 17:21	AR1242 hit	YPIAJ	Ok,M
30	Q3446-05DL	SB-18DL	PP076040.D	24 Oct 2025 17:37	AR1242 Hit	YPIAJ	Ok,M
31	Q3446-06DL	SB-19DL	PP076041.D	24 Oct 2025 17:53	AR1242+1260 Hit	YPIAJ	Ok,M
32	AR1660CCC500	AR1660CCC500	PP076042.D	24 Oct 2025 18:58		YPIAJ	Ok
33	AR1242CCC500	AR1242CCC500	PP076043.D	24 Oct 2025 19:31		YPIAJ	Ok,M
34	AR1248CCC500	AR1248CCC500	PP076044.D	24 Oct 2025 20:03		YPIAJ	Ok,M

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP102425

Review By	yogesh	Review On	10/24/2025 11:09:51 AM
Supervise By	mohammad	Supervise On	10/28/2025 6:10:31 AM
SubDirectory	PP102425	HP Acquire Method	HP Processing Method PP102125
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,P P24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP 24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854		
CCC	PP24824,PP24828,PP24832,PP24836,PP24840,PP24844,PP24852		
Internal Standard/PEM			
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

35	AR1254CCC500	AR1254CCC500	PP076045.D	24 Oct 2025 20:19		YPIAJ	Ok,M
36	I.BLK	I.BLK	PP076046.D	24 Oct 2025 20:36		YPIAJ	Ok
37	Q3450-10	SS-20-(3.5-4.0)	PP076047.D	24 Oct 2025 20:52	AR1254 Hit	YPIAJ	Ok,M
38	Q3450-11	SS-28-(2.0-2.5)	PP076048.D	24 Oct 2025 21:08		YPIAJ	Ok
39	Q3450-20	SS-21-(0.0-0.5)	PP076049.D	24 Oct 2025 21:24	AR1254 Hit , Need 5x dilution	YPIAJ	Dilution
40	Q3450-21	SS-21-(2.0-2.5)	PP076050.D	24 Oct 2025 21:41		YPIAJ	Ok
41	Q3450-22	SS-21-(3.5-4.0)	PP076051.D	24 Oct 2025 21:57	AR1254 Hit , Need 2x dilution	YPIAJ	Dilution
42	Q3427-01	TP-8	PP076052.D	24 Oct 2025 22:13		YPIAJ	Ok,M
43	Q3446-07DL	SB-20DL	PP076053.D	24 Oct 2025 22:29	AR1242 Hit	YPIAJ	Ok,M
44	Q3446-08DL	SB-21DL	PP076054.D	24 Oct 2025 22:45	AR1242+1260 Hit	YPIAJ	Ok,M
45	Q3446-03DL	SB-16DL	PP076055.D	24 Oct 2025 23:01	AR1260 Hit	YPIAJ	Ok,M
46	Q3446-04DL	SB-17DL	PP076056.D	24 Oct 2025 23:17	AR1260 Hit	YPIAJ	Ok,M
47	AR1660CCC500	AR1660CCC500	PP076057.D	25 Oct 2025 00:22		YPIAJ	Ok
48	AR1242CCC500	AR1242CCC500	PP076058.D	25 Oct 2025 01:11		YPIAJ	Ok,M
49	AR1248CCC500	AR1248CCC500	PP076059.D	25 Oct 2025 01:43		YPIAJ	Ok,M
50	AR1254CCC500	AR1254CCC500	PP076060.D	25 Oct 2025 01:59		YPIAJ	Ok,M
51	I.BLK	I.BLK	PP076061.D	25 Oct 2025 02:16		YPIAJ	Ok

M : Manual Integration

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP102725

Review By	yogesh	Review On	10/27/2025 10:53:48 AM
Supervise By	mohammad	Supervise On	10/29/2025 5:27:46 AM
SubDirectory	PP102725	HP Acquire Method	HP Processing Method PP102125

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,P P24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP 24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854
CCC	PP24824,PP24828,PP24832,PP24836,PP24840,PP24844,PP24852
Internal Standard/PEM	
ICV/ILBLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PP076062.D	27 Oct 2025 08:40		YPIAJ	Ok
2	AR1660CCC500	AR1660CCC500	PP076063.D	27 Oct 2025 08:56		YPIAJ	Ok,M
3	AR1242CCC500	AR1242CCC500	PP076064.D	27 Oct 2025 09:13		YPIAJ	Ok,M
4	AR1248CCC500	AR1248CCC500	PP076065.D	27 Oct 2025 09:29		YPIAJ	Ok,M
5	AR1254CCC500	AR1254CCC500	PP076066.D	27 Oct 2025 09:45		YPIAJ	Ok,M
6	I.BLK	I.BLK	PP076067.D	27 Oct 2025 10:02		YPIAJ	Ok
7	PB170236BL	PB170236BL	PP076068.D	27 Oct 2025 10:18		YPIAJ	Ok,M
8	PB170236BS	PB170236BS	PP076069.D	27 Oct 2025 10:34		YPIAJ	Ok,M
9	Q3439-01	TP-9	PP076070.D	27 Oct 2025 11:06	DCB,TCX high 1st column	YPIAJ	ReRun
10	Q3449-01	CF-620-COMP-41	PP076071.D	27 Oct 2025 11:26		YPIAJ	Ok
11	Q3450-20DL	SS-21-(0.0-0.5)DL	PP076072.D	27 Oct 2025 11:42	AR1254 Hit	YPIAJ	Ok,M
12	Q3450-22DL	SS-21-(3.5-4.0)DL	PP076073.D	27 Oct 2025 11:58	AR1254 Hit	YPIAJ	Ok,M
13	Q3446-05	SB-18	PP076074.D	27 Oct 2025 12:14	AR1242 Hit, Need Dilution, f flag	YPIAJ	Not Ok
14	Q3447-03	P001-TW2-251023-01	PP076075.D	27 Oct 2025 12:30	F Flag in TCMX in 2nd column , TCMX high in both column , AR1254+1260 Hit	YPIAJ	ReRun
15	Q3447-07	P001-TW4-251023-01	PP076076.D	27 Oct 2025 12:47	TCMX high in 2nd column	YPIAJ	Ok,M
16	Q3446-05	SB-18	PP076077.D	27 Oct 2025 13:03	TCMX high in 2nd column , AR1242 hit , need 100x dilution	YPIAJ	Dilution
17	AR1660CCC500	AR1660CCC500	PP076078.D	27 Oct 2025 14:08		YPIAJ	Ok

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP102725

Review By	yogesh	Review On	10/27/2025 10:53:48 AM		
Supervise By	mohammad	Supervise On	10/29/2025 5:27:46 AM		
SubDirectory	PP102725	HP Acquire Method	HP Processing Method	PP102125	
STD. NAME	STD REF.#				
Tune/Reschk Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854				
CCC Internal Standard/PEM	PP24824,PP24828,PP24832,PP24836,PP24840,PP24844,PP24852				
ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862				

18	AR1242CCC500	AR1242CCC500	PP076079.D	27 Oct 2025 14:24		YPIAJ	Ok,M
19	AR1248CCC500	AR1248CCC500	PP076080.D	27 Oct 2025 14:40		YPIAJ	Ok,M
20	AR1254CCC500	AR1254CCC500	PP076081.D	27 Oct 2025 14:57		YPIAJ	Ok,M
21	I.BLK	I.BLK	PP076082.D	27 Oct 2025 15:13		YPIAJ	Ok
22	Q3456-01	OILY-DEBRIS	PP076083.D	27 Oct 2025 15:29		YPIAJ	Ok,M
23	Q3457-01	TR-05-102425	PP076084.D	27 Oct 2025 15:45		YPIAJ	Ok,M
24	Q3460-01	WC-1	PP076085.D	27 Oct 2025 16:02		YPIAJ	Ok
25	Q3460-17	WC-5	PP076086.D	27 Oct 2025 16:18		YPIAJ	Ok,M
26	Q3447-05	P001-TW3-251023-01	PP076087.D	27 Oct 2025 16:34	TCMX high in 2nd column	YPIAJ	Ok,M
27	Q3447-01	P001-TW1-251023-01	PP076088.D	27 Oct 2025 16:50	F Flag in TCMX in 2nd column , TCMX high in both column	YPIAJ	ReRun
28	Q3459-01	Y2404-0071-END-1-1	PP076089.D	27 Oct 2025 17:07	AR1254 Hit	YPIAJ	Ok,M
29	Q3459-02	Y2404-0071-END-1-2	PP076090.D	27 Oct 2025 17:23		YPIAJ	Ok,M
30	Q3459-03	Y2404-0071-END-2-1	PP076091.D	27 Oct 2025 17:39	AR1254 Hit	YPIAJ	Ok,M
31	Q3459-04	Y2404-0071-END-2-2	PP076092.D	27 Oct 2025 17:55		YPIAJ	Ok
32	AR1660CCC500	AR1660CCC500	PP076093.D	27 Oct 2025 19:00		YPIAJ	Ok
33	AR1242CCC500	AR1242CCC500	PP076094.D	27 Oct 2025 19:49		YPIAJ	Ok,M
34	AR1248CCC500	AR1248CCC500	PP076095.D	27 Oct 2025 20:05		YPIAJ	Ok,M
35	AR1254CCC500	AR1254CCC500	PP076096.D	27 Oct 2025 20:21		YPIAJ	Ok,M
36	I.BLK	I.BLK	PP076097.D	27 Oct 2025 20:38		YPIAJ	Ok

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP102725

Review By	yogesh	Review On	10/27/2025 10:53:48 AM
Supervise By	mohammad	Supervise On	10/29/2025 5:27:46 AM
SubDirectory	PP102725	HP Acquire Method	HP Processing Method PP102125
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,P P24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP 24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854		
CCC	PP24824,PP24828,PP24832,PP24836,PP24840,PP24844,PP24852		
Internal Standard/PEM			
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

37	PB170272BL	PB170272BL	PP076098.D	27 Oct 2025 20:54		YPIAJ	Ok
38	PB170272BS	PB170272BS	PP076099.D	27 Oct 2025 21:10		YPIAJ	Ok,M
39	Q3459-05	BC259240LOK-END-1	PP076100.D	27 Oct 2025 21:26		YPIAJ	Ok,M
40	Q3459-06	BC259240LOK-END-1	PP076101.D	27 Oct 2025 21:42		YPIAJ	Ok,M
41	Q3459-07	BC271327-END-1-1	PP076102.D	27 Oct 2025 21:59		YPIAJ	Ok,M
42	Q3459-08	BC271327-END-1-2	PP076103.D	27 Oct 2025 22:15		YPIAJ	Ok,M
43	AR1660CCC500	AR1660CCC500	PP076104.D	27 Oct 2025 23:20		YPIAJ	Ok
44	AR1242CCC500	AR1242CCC500	PP076105.D	28 Oct 2025 00:08		YPIAJ	Ok,M
45	AR1248CCC500	AR1248CCC500	PP076106.D	28 Oct 2025 00:25		YPIAJ	Ok,M
46	AR1254CCC500	AR1254CCC500	PP076107.D	28 Oct 2025 00:41		YPIAJ	Ok
47	I.BLK	I.BLK	PP076108.D	28 Oct 2025 00:57		YPIAJ	Ok

M : Manual Integration

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP103125

Review By	rahul	Review On	11/3/2025 8:18:24 AM
Supervise By	yogesh	Supervise On	11/3/2025 8:49:19 AM
SubDirectory	PP103125	HP Acquire Method	HP Processing Method PP102125

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854
CCC	PP24824,PP24828,PP24832,PP24836,PP24840,PP24844,PP24852
Internal Standard/PEM	
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PP076241.D	31 Oct 2025 09:26		YPIAJ	Ok
2	AR1660CCC500	AR1660CCC500	PP076242.D	31 Oct 2025 09:43	AR1260-04 high in second column	YPIAJ	Ok
3	AR1242CCC500	AR1242CCC500	PP076243.D	31 Oct 2025 09:59		YPIAJ	Ok,M
4	AR1248CCC500	AR1248CCC500	PP076244.D	31 Oct 2025 10:15		YPIAJ	Ok,M
5	AR1254CCC500	AR1254CCC500	PP076245.D	31 Oct 2025 10:32		YPIAJ	Ok,M
6	I.BLK	I.BLK	PP076246.D	31 Oct 2025 10:48		YPIAJ	Ok
7	Q3482-04	WC-4	PP076247.D	31 Oct 2025 11:09		YPIAJ	Ok,M
8	Q3500-01	459	PP076248.D	31 Oct 2025 13:21	TCMX high in 2nd column ,	YPIAJ	Ok,M
9	Q3501-01	461	PP076249.D	31 Oct 2025 13:38	DCB low in 2nd and high in 1st column , TCMX low in 2nd column	YPIAJ	Not Ok
10	Q3502-01	CATCH-CAN-USED-OI	PP076250.D	31 Oct 2025 13:54		YPIAJ	Ok,M
11	Q3503-01	322	PP076251.D	31 Oct 2025 14:10		YPIAJ	Ok,M
12	Q3503-03	323	PP076252.D	31 Oct 2025 14:26	DCB Low in 2nd column , AR1260 Hit	YPIAJ	Ok,M
13	Q3503-05	324	PP076253.D	31 Oct 2025 14:43		YPIAJ	Ok,M
14	Q3505-02	10-10-25-PLASTIC	PP076254.D	31 Oct 2025 14:59	DCB Low in 2nd column	YPIAJ	Ok,M
15	Q3505-03	10-17-25-RAGS	PP076255.D	31 Oct 2025 15:15		YPIAJ	Ok,M
16	AR1660CCC500	AR1660CCC500	PP076256.D	31 Oct 2025 16:04		YPIAJ	Ok,M
17	AR1242CCC500	AR1242CCC500	PP076257.D	31 Oct 2025 16:20		YPIAJ	Ok,M

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP103125

Review By	rahul	Review On	11/3/2025 8:18:24 AM
Supervise By	yogesh	Supervise On	11/3/2025 8:49:19 AM
SubDirectory	PP103125	HP Acquire Method	HP Processing Method PP102125
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854		
CCC	PP24824,PP24828,PP24832,PP24836,PP24840,PP24844,PP24852		
Internal Standard/PEM			
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

18	AR1248CCC500	AR1248CCC500	PP076258.D	31 Oct 2025 16:37		YPIAJ	Ok,M
19	AR1254CCC500	AR1254CCC500	PP076259.D	31 Oct 2025 16:53		YPIAJ	Ok,M
20	I.BLK	I.BLK	PP076260.D	31 Oct 2025 17:09		YPIAJ	Ok
21	PB170347BL	PB170347BL	PP076261.D	31 Oct 2025 17:26		YPIAJ	Ok
22	PB170347BS	PB170347BS	PP076262.D	31 Oct 2025 17:42		YPIAJ	Ok,M
23	PB170347BSD	PB170347BSD	PP076263.D	31 Oct 2025 17:58		YPIAJ	Ok,M
24	Q3484-01	TRE-25-0118	PP076264.D	31 Oct 2025 18:14	AR1254 hit	YPIAJ	Ok,M
25	Q3492-01	252611	PP076265.D	31 Oct 2025 18:31		YPIAJ	Ok,M
26	Q3495-03	TW-1025-1	PP076266.D	31 Oct 2025 18:47		YPIAJ	Ok,M
27	Q3500-02	462	PP076267.D	31 Oct 2025 19:03	AR1248+1260 Hit	YPIAJ	Ok,M
28	Q3500-03	463	PP076268.D	31 Oct 2025 19:19	AR1248+1260 Hit	YPIAJ	Ok,M
29	Q3500-04	465	PP076269.D	31 Oct 2025 19:36	DCB Low in 2nd column	YPIAJ	Ok,M
30	Q3503-07	320	PP076270.D	31 Oct 2025 19:52	DCB low and TCMX high in 2nd column	YPIAJ	ReRun
31	AR1660CCC500	AR1660CCC500	PP076271.D	31 Oct 2025 20:57		YPIAJ	Ok
32	AR1242CCC500	AR1242CCC500	PP076272.D	31 Oct 2025 21:46	TCMX high in 2nd column	YPIAJ	Ok,M
33	AR1248CCC500	AR1248CCC500	PP076273.D	31 Oct 2025 22:02		YPIAJ	Ok,M
34	AR1254CCC500	AR1254CCC500	PP076274.D	31 Oct 2025 22:18		YPIAJ	Ok,M
35	I.BLK	I.BLK	PP076275.D	31 Oct 2025 22:34		YPIAJ	Ok
36	PB170358BL	PB170358BL	PP076276.D	31 Oct 2025 22:51		YPIAJ	Ok

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP103125

Review By	rahul	Review On	11/3/2025 8:18:24 AM
Supervise By	yogesh	Supervise On	11/3/2025 8:49:19 AM
SubDirectory	PP103125	HP Acquire Method	HP Processing Method PP102125
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,P P24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP 24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854		
CCC	PP24824,PP24828,PP24832,PP24836,PP24840,PP24844,PP24852		
Internal Standard/PEM			
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

37	PB170358BS	PB170358BS	PP076277.D	31 Oct 2025 23:07		YPIAJ	Ok,M
38	Q3437-15	SS-32(0-0.5)	PP076278.D	31 Oct 2025 23:23	AR1254 Hit	YPIAJ	Ok,M
39	Q3450-04	SS-10R-(3.5-4.0)	PP076279.D	31 Oct 2025 23:40		YPIAJ	Ok,M
40	Q3450-04MS	SS-10R-(3.5-4.0)MS	PP076280.D	31 Oct 2025 23:56		YPIAJ	Ok,M
41	Q3450-04MSD	SS-10R-(3.5-4.0)MSD	PP076281.D	01 Nov 2025 00:12		YPIAJ	Ok,M
42	Q3450-05	SS-13R-(3.5-4.0)	PP076282.D	01 Nov 2025 00:28		YPIAJ	Ok
43	Q3450-06	SS-13R-(5.0-5.5)	PP076283.D	01 Nov 2025 00:45		YPIAJ	Ok,M
44	Q3450-23	SS-22-(0.0-0.5)	PP076284.D	01 Nov 2025 01:01	AR1254 Hit , Need 5x dilution	YPIAJ	Dilution
45	Q3450-25	SS-22-(3.5-4.0)	PP076285.D	01 Nov 2025 01:17	AR1254 Hit	YPIAJ	Ok,M
46	AR1660CCC500	AR1660CCC500	PP076286.D	01 Nov 2025 02:22		YPIAJ	Ok,M
47	AR1242CCC500	AR1242CCC500	PP076287.D	01 Nov 2025 03:11	TCMX high in 2nd column	YPIAJ	Ok,M
48	AR1248CCC500	AR1248CCC500	PP076288.D	01 Nov 2025 03:28	TCMX high in 2nd column	YPIAJ	Ok,M
49	AR1254CCC500	AR1254CCC500	PP076289.D	01 Nov 2025 03:44	TCMX high in 2nd column	YPIAJ	Ok,M
50	I.BLK	I.BLK	PP076290.D	01 Nov 2025 04:00		YPIAJ	Ok

M : Manual Integration

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP110325

Review By	rahul	Review On	11/4/2025 8:41:32 AM
Supervise By	yogesh	Supervise On	11/4/2025 10:10:47 AM
SubDirectory	PP110325	HP Acquire Method	HP Processing Method PP102125

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854
CCC	PP24824,PP24828,PP24832,PP24836,PP24840,PP24844,PP24852
Internal Standard/PEM	
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	Sampleld	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PP076291.D	03 Nov 2025 08:28		YPIAJ	Ok
2	AR1660CCC500	AR1660CCC500	PP076292.D	03 Nov 2025 08:44	AR1260-04 high in second column	YPIAJ	Ok
3	AR1242CCC500	AR1242CCC500	PP076293.D	03 Nov 2025 09:04		YPIAJ	Ok,M
4	AR1248CCC500	AR1248CCC500	PP076294.D	03 Nov 2025 09:21		YPIAJ	Ok,M
5	AR1254CCC500	AR1254CCC500	PP076295.D	03 Nov 2025 09:37		YPIAJ	Ok,M
6	I.BLK	I.BLK	PP076296.D	03 Nov 2025 09:53		YPIAJ	Ok
7	Q3450-23DL	SS-22-(0.0-0.5)DL	PP076297.D	03 Nov 2025 10:09	AR1254 Hit	YPIAJ	Ok,M
8	PB170370BL	PB170370BL	PP076298.D	03 Nov 2025 11:53		YPIAJ	Ok,M
9	PB170370BS	PB170370BS	PP076299.D	03 Nov 2025 12:09		YPIAJ	Ok,M
10	Q3515-01	SU-04-103125	PP076300.D	03 Nov 2025 12:25	TCMX high in 2nd column	YPIAJ	Ok,M
11	Q3515-01MS	SU-04-103125MS	PP076301.D	03 Nov 2025 12:41	TCMX high in 2nd column	YPIAJ	Ok,M
12	Q3515-01MSD	SU-04-103125MSD	PP076302.D	03 Nov 2025 12:58	TCMX high in 2nd column	YPIAJ	Ok,M
13	Q3516-01	HD-01-103125	PP076303.D	03 Nov 2025 13:14		YPIAJ	Ok,M
14	AR1660CCC500	AR1660CCC500	PP076304.D	03 Nov 2025 14:19		YPIAJ	Ok
15	AR1242CCC500	AR1242CCC500	PP076305.D	03 Nov 2025 14:35		YPIAJ	Ok,M
16	AR1248CCC500	AR1248CCC500	PP076306.D	03 Nov 2025 14:51		YPIAJ	Ok,M
17	AR1254CCC500	AR1254CCC500	PP076307.D	03 Nov 2025 15:08		YPIAJ	Ok,M

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP110325

Review By	rahul	Review On	11/4/2025 8:41:32 AM		
Supervise By	yogesh	Supervise On	11/4/2025 10:10:47 AM		
SubDirectory	PP110325	HP Acquire Method	HP Processing Method	PP102125	

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,P P24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP 24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854
CCC	PP24824,PP24828,PP24832,PP24836,PP24840,PP24844,PP24852
Internal Standard/PEM	
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

18	I.BLK	I.BLK	PP076308.D	03 Nov 2025 15:24		YPIAJ	Ok
----	-------	-------	------------	-------------------	--	-------	----

M : Manual Integration

Instrument ID: ECD_Q

Daily Analysis Runlog For Sequence/QC Batch ID # PQ101025

Review By	yogesh	Review On	10/10/2025 10:11:25 AM
Supervise By	mohammad	Supervise On	10/14/2025 2:48:24 AM
SubDirectory	PQ101025	HP Acquire Method	HP Processing Method PQ101025

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,P P24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP 24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854
CCC	
Internal Standard/PEM	
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	Sampleld	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PQ070971.D	10 Oct 2025 08:27		YPIAJ	Ok
2	I.BLK	I.BLK	PQ070972.D	10 Oct 2025 08:42		YPIAJ	Ok
3	AR1660ICC1000	AR1660ICC1000	PQ070973.D	10 Oct 2025 08:57		YPIAJ	Ok
4	AR1660ICC750	AR1660ICC750	PQ070974.D	10 Oct 2025 09:11		YPIAJ	Ok
5	AR1660ICC500	AR1660ICC500	PQ070975.D	10 Oct 2025 09:26		YPIAJ	Ok
6	AR1660ICC250	AR1660ICC250	PQ070976.D	10 Oct 2025 09:40		YPIAJ	Ok
7	AR1660ICC050	AR1660ICC050	PQ070977.D	10 Oct 2025 09:55		YPIAJ	Ok,M
8	AR1221ICC500	AR1221ICC500	PQ070978.D	10 Oct 2025 10:09		YPIAJ	Ok
9	AR1232ICC500	AR1232ICC500	PQ070979.D	10 Oct 2025 10:24		YPIAJ	Ok
10	AR1242ICC1000	AR1242ICC1000	PQ070980.D	10 Oct 2025 10:38		YPIAJ	Ok
11	AR1242ICC750	AR1242ICC750	PQ070981.D	10 Oct 2025 10:53		YPIAJ	Ok
12	AR1242ICC500	AR1242ICC500	PQ070982.D	10 Oct 2025 11:08		YPIAJ	Ok
13	AR1242ICC250	AR1242ICC250	PQ070983.D	10 Oct 2025 11:22		YPIAJ	Ok
14	AR1242ICC050	AR1242ICC050	PQ070984.D	10 Oct 2025 11:37		YPIAJ	Ok,M
15	AR1248ICC1000	AR1248ICC1000	PQ070985.D	10 Oct 2025 11:51		YPIAJ	Ok
16	AR1248ICC750	AR1248ICC750	PQ070986.D	10 Oct 2025 12:06		YPIAJ	Ok
17	AR1248ICC500	AR1248ICC500	PQ070987.D	10 Oct 2025 12:20		YPIAJ	Ok
18	AR1248ICC250	AR1248ICC250	PQ070988.D	10 Oct 2025 12:35		YPIAJ	Ok,M

Instrument ID: ECD_Q

Daily Analysis Runlog For Sequence/QC Batch ID # PQ101025

Review By	yogesh	Review On	10/10/2025 10:11:25 AM
Supervise By	mohammad	Supervise On	10/14/2025 2:48:24 AM
SubDirectory	PQ101025	HP Acquire Method	HP Processing Method PQ101025
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,P P24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP 24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862		

19	AR1248ICC050	AR1248ICC050	PQ070989.D	10 Oct 2025 12:50		YPIAJ	Ok
20	AR1254ICC1000	AR1254ICC1000	PQ070990.D	10 Oct 2025 13:04		YPIAJ	Ok
21	AR1254ICC750	AR1254ICC750	PQ070991.D	10 Oct 2025 13:19		YPIAJ	Ok
22	AR1254ICC500	AR1254ICC500	PQ070992.D	10 Oct 2025 13:33		YPIAJ	Ok
23	AR1254ICC250	AR1254ICC250	PQ070993.D	10 Oct 2025 13:48		YPIAJ	Ok
24	AR1254ICC050	AR1254ICC050	PQ070994.D	10 Oct 2025 15:15		YPIAJ	Ok
25	AR1262ICC500	AR1262ICC500	PQ070995.D	10 Oct 2025 15:30		YPIAJ	Ok
26	AR1268ICC1000	AR1268ICC1000	PQ070996.D	10 Oct 2025 15:44		YPIAJ	Ok
27	AR1268ICC750	AR1268ICC750	PQ070997.D	10 Oct 2025 15:59		YPIAJ	Ok
28	AR1268ICC500	AR1268ICC500	PQ070998.D	10 Oct 2025 16:13		YPIAJ	Ok
29	AR1268ICC250	AR1268ICC250	PQ070999.D	10 Oct 2025 16:28		YPIAJ	Ok
30	AR1268ICC050	AR1268ICC050	PQ071000.D	10 Oct 2025 16:43		YPIAJ	Ok
31	PQ101025ICV500	ICVPQ101025	PQ071001.D	10 Oct 2025 16:57		YPIAJ	Ok,M
32	AR1242ICV500	ICVPQ101025AR1242	PQ071002.D	10 Oct 2025 17:41		YPIAJ	Ok,M
33	AR1248ICV500	ICVPQ101025AR1248	PQ071003.D	10 Oct 2025 18:25		YPIAJ	Ok,M
34	AR1254ICV500	ICVPQ101025AR1254	PQ071004.D	10 Oct 2025 19:08		YPIAJ	Ok
35	AR1268ICV500	ICVPQ101025AR1268	PQ071005.D	10 Oct 2025 19:52		YPIAJ	Ok
36	DDT ANALOG	DDT ANALOG	PQ071006.D	10 Oct 2025 20:36		YPIAJ	Ok

M : Manual Integration

Instrument ID: ECD_Q

Daily Analysis Runlog For Sequence/QC Batch ID # PQ102725

Review By	yogesh	Review On	10/27/2025 10:54:57 AM
Supervise By	mohammad	Supervise On	10/29/2025 5:27:51 AM
SubDirectory	PQ102725	HP Acquire Method	HP Processing Method PQ101025

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,P P24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP 24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854
CCC	PP24824,PP24828,PP24832,PP24836,PP24840,PP24844,PP24852
Internal Standard/PEM	
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	Sampleld	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PQ071148.D	27 Oct 2025 08:42		YPIAJ	Ok
2	AR1660CCC500	AR1660CCC500	PQ071149.D	27 Oct 2025 08:57		YPIAJ	Ok
3	I.BLK	I.BLK	PQ071150.D	27 Oct 2025 09:11		YPIAJ	Ok
4	Q3450-01	SS-5R-(2.0-2.5)	PQ071151.D	27 Oct 2025 09:26	AR1254 Hit	YPIAJ	Ok
5	Q3450-02	SS-5R-(3.5-4.0)	PQ071152.D	27 Oct 2025 09:41	AR1254 Hit	YPIAJ	Ok
6	Q3450-02MS	SS-5R-(3.5-4.0)MS	PQ071153.D	27 Oct 2025 09:55		YPIAJ	Ok,M
7	Q3450-02MSD	SS-5R-(3.5-4.0)MSD	PQ071154.D	27 Oct 2025 10:10		YPIAJ	Ok,M
8	Q3450-08	SS-20-(0.0-0.5)	PQ071155.D	27 Oct 2025 10:24	AR1254 Hit	YPIAJ	Ok
9	Q3450-09	SS-20-(2.0-2.5)	PQ071156.D	27 Oct 2025 10:39	AR1254 Hit	YPIAJ	Ok,M
10	Q3437-16	SS-23(0-0.5)	PQ071157.D	27 Oct 2025 10:54	AR1254 Hit, TCX high 1st column	YPIAJ	Ok
11	Q3419-01	PR-132-S16-015094-20	PQ071158.D	27 Oct 2025 11:08	AR1242 hits	YPIAJ	Ok,M
12	Q3419-02MS	PR-132-S16-015094-20	PQ071159.D	27 Oct 2025 11:23		YPIAJ	Ok
13	Q3419-03MSD	PR-132-S16-015094-20	PQ071160.D	27 Oct 2025 11:37		YPIAJ	Ok
14	AR1660CCC500	AR1660CCC500	PQ071161.D	27 Oct 2025 12:35		YPIAJ	Ok
15	AR1242CCC500	AR1242CCC500	PQ071162.D	27 Oct 2025 13:34		YPIAJ	Ok
16	AR1248CCC500	AR1248CCC500	PQ071163.D	27 Oct 2025 13:49		YPIAJ	Ok
17	AR1254CCC500	AR1254CCC500	PQ071164.D	27 Oct 2025 14:03		YPIAJ	Ok

Instrument ID: ECD_Q

Daily Analysis Runlog For Sequence/QC Batch ID # PQ102725

Review By	yogesh	Review On	10/27/2025 10:54:57 AM		
Supervise By	mohammad	Supervise On	10/29/2025 5:27:51 AM		
SubDirectory	PQ102725	HP Acquire Method	HP Processing Method	PQ101025	
STD. NAME	STD REF.#				
Tune/Reschk					
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,P P24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP 24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854				
CCC	PP24824,PP24828,PP24832,PP24836,PP24840,PP24844,PP24852				
Internal Standard/PEM					
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862				
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

18	I.BLK	I.BLK	PQ071165.D	27 Oct 2025 14:18		YPIAJ	Ok
19	PB170255BL	PB170255BL	PQ071166.D	27 Oct 2025 14:32	DCB high 1st column for PSEG	YPIAJ	Not Ok
20	PB170255BS	PB170255BS	PQ071167.D	27 Oct 2025 14:47	DCB high 1st column for PSEG	YPIAJ	Not Ok
21	Q3451-01	TP-11	PQ071168.D	27 Oct 2025 15:01	Tcmx high in 1st column	YPIAJ	Ok
22	Q3451-05	TP-12	PQ071169.D	27 Oct 2025 15:16		YPIAJ	Ok
23	Q3451-05MS	TP-12MS	PQ071170.D	27 Oct 2025 15:31	Tcmx high in 1st column	YPIAJ	Ok
24	Q3451-05MSD	TP-12MSD	PQ071171.D	27 Oct 2025 15:45	Tcmx high in 1st column	YPIAJ	Ok
25	Q3452-01	JB-1	PQ071172.D	27 Oct 2025 16:00		YPIAJ	Ok
26	Q3460-05	WC-2	PQ071173.D	27 Oct 2025 16:14	TCMX & DCB high in 1st column	YPIAJ	ReRun
27	Q3460-09	WC-3	PQ071174.D	27 Oct 2025 16:29	TCMX & DCB high in 1st column	YPIAJ	ReRun
28	Q3460-13	WC-4	PQ071175.D	27 Oct 2025 16:43	DCB high in 1st column, AR1260 hits	YPIAJ	Ok,M
29	AR1660CCC500	AR1660CCC500	PQ071176.D	27 Oct 2025 17:42		YPIAJ	Ok,M
30	AR1242CCC500	AR1242CCC500	PQ071177.D	27 Oct 2025 18:40		YPIAJ	Ok
31	AR1248CCC500	AR1248CCC500	PQ071178.D	27 Oct 2025 18:55		YPIAJ	Ok
32	AR1254CCC500	AR1254CCC500	PQ071179.D	27 Oct 2025 19:09		YPIAJ	Ok,M
33	I.BLK	I.BLK	PQ071180.D	27 Oct 2025 19:24		YPIAJ	Ok

M : Manual Integration