284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Cover Page

Order ID : Q3450

Project ID : PSEG Athenia Substation

Client : Matrix New World Engineering

Lab Sample Number	Client Sample Number
Q3450-01	SS-5R-(2.0-2.5)
Q3450-02	SS-5R-(3.5-4.0)
Q3450-04	SS-10R-(3.5-4.0)
Q3450-05	SS-13R-(3.5-4.0)
Q3450-06	SS-13R-(5.0-5.5)
Q3450-08	SS-20-(0.0-0.5)
Q3450-09	SS-20-(2.0-2.5)
Q3450-10	SS-20-(3.5-4.0)
Q3450-11	SS-28-(2.0-2.5)
Q3450-20	SS-21-(0.0-0.5)
Q3450-21	SS-21-(2.0-2.5)
Q3450-22	SS-21-(3.5-4.0)
Q3450-23	SS-22-(0.0-0.5)
Q3450-25	SS-22-(3.5-4.0)

I certify that the data package is in compliance with the terms and conditions of the contract,both technically and for completeness, for other than the conditions detailed above. Release of the data contained in this hard copy data package has been authorized by the laboratory manager or his designee, as verified by the following signature.

Signature : _____

Date: 11/12/2025

NYDOH CERTIFICATION NO - 11376

NJDEP CERTIFICATION NO - 20012

284 Sheffield Street, Mountainside, NJ 07092
Phone: 908 789 8900 Fax: 908 789 8922

CASE NARRATIVE

Matrix New World Engineering
Project Name: PSEG Athenia Substation
Project # N/A
Order ID # Q3450
Test Name: PCB

A. Number of Samples and Date of Receipt:
14 Solid samples were received on 10/23/2025.

B. Parameters
According to the Chain of Custody document, the following analyses were requested: PCB. This data package contains results for PCB.

C. Analytical Techniques:
The analyses were performed on instrument GCECD_Q.The front column is ZB-MR1 which is 30 meters, 0.32 mm ID, 0.5 um df, Catalogue # 7HM-G016-17. The rear column is ZB-MR2 which is 30 meters, 0.32 mm ID, 0.25 µm; Catalogue # 7HM-G017-11.The analyses were performed on instrument GCECD_P. The front column is ZB-MR1 which is 30 meters, 0.32 mm ID, 0.5 um df, Catalogue # 7HM-G016-17. The rear column is ZB-MR2 which is 30 meters, 0.32 mm ID, 0.25 µm; Catalogue # 7HM-G017-11.The analysis of PCBs was based on method 8082A and extraction was done based on method 3541.

D. QA/ QC Samples:
The Holding Times were met for all analysis.
The Surrogate recoveries were met for all analysis.
The Retention Times were met for all analysis.
The MS recoveries met the requirements for all compounds.
The MSD recoveries met the requirements for all compounds.
The RPD were met for all analysis.
The Blank Spike met requirements for all compounds.
The Blank analysis did not indicate the presence of lab contamination.
The Initial Calibration met the requirements.
The Continuous Calibration File ID PP076292.D met the requirements except for Aroclor-1260(Peak-04) is failing in 2nd column, however it is passed in 1st column therefore no corrective action was taken.

Samples SS-21-(0.0-0.5), SS-21-(3.5-4.0) and SS-22-(0.0-0.5) were diluted due to high concentrations.

E. Additional Comments:
The soil samples results are based on a dry weight basis.

284 Sheffield Street, Mountainside, NJ 07092
Phone: 908 789 8900 Fax: 908 789 8922

F. Manual Integration Comments:

Please refer to the Manual integration Report included with the Run Logs for information on the manual integrations performed.

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature_____

284 Sheffield Street, Mountainside NJ 07092 Tel. 908-789-8900 Fax: 908-789-8922

DATA REPORTING QUALIFIERS- ORGANIC

For reporting results, the following " Results Qualifiers" are used:

Value	If the result is a value greater than or equal to the detection limit, report the value
U	Indicates the compound was analyzed for but was not detected. Report the minimum detection limit for the sample with the U, i.e. "10 U". This is not necessarily the instrument detection limit attainable for this particular sample based on any concentration or dilution that may have been required.
ND	Indicates the analyte was analyzed for, but not detected
J	Indicates an estimated value**.** This flag is used: (1) When estimating a concentration for a tentatively identified compound (library search hits, where a 1:1 response is assumed.) (2) When the mass spectral data indicated the identification, however the result was less than the specified detection limit greater than zero. If the detection limit was 10ug/L and a concentration of 3 ug/L was calculated report as 3 J. This is flag is used when similar situation arise on any organic parameter i.e. Pest, PCB and others.
B	Indicates the analyte was found in the blank as well as the sample report as "12 B".
E	Indicates the analyte 's concentration exceeds the calibrated range of the instrument for that specific analysis.
D	This flag identifies all compounds identified in an analysis at a secondary dilution factor.
P	This flag is used for Pesticide/PCB target analyte when there is >25% difference for detected concentrations between the two GC columns. The lower of the two values is reported on Form 1 and flagged with a "P".
N	This flag indicates presumptive evidence of a compound. This is only used for tentatively identified compounds (TICs), where the identification is based on a mass spectral library search. It applies to all TIC results. For generic characterization of a TIC, such as chlorinated hydrocarbon, the flag is not used.
A	This flag indicates that a Tentatively Identified Compound is a suspected aldol-condensation product.
Q	Indicates the LCS did not meet the control limits requirements

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

APPENDIX A

QA REVIEW GENERAL DOCUMENTATION

Project #: **Q3450**

	Completed
For thorough review, the report must have the following:	
GENERAL:	
Are all original paperwork present (chain of custody, record of communication,airbill, sample management lab chronicle, login page)	✓
Check chain-of-custody for proper relinquish/return of samples	✓
Is the chain of custody signed and complete	✓
Check internal chain-of-custody for proper relinquish/return of samples /sample extracts	✓
Collect information for each project id from server. Were all requirements followed	✓
COVER PAGE:	
Do numbers of samples correspond to the number of samples in the Chain of Custody on login page	✓
Do lab numbers and client Ids on cover page agree with the Chain of Custody	✓
CHAIN OF CUSTODY:	
Do requested analyses on Chain of Custody agree with form I results	✓
Do requested analyses on Chain of Custody agree with the log-in page	✓
Were the correct method log-in for analysis according to the Analytical Request and Chain of Castody	✓
Were the samples received within hold time	✓
Were any problems found with the samples at arrival recorded in the Sample Management Laboratory Chronicle	✓
ANALYTICAL:	
Was method requirement followed?	✓
Was client requirement followed?	✓
Does the case narrative summarize all QC failure?	✓
All runlogs and manual integration are reviewed for requirements	✓
All manual calculations and /or hand notations verified	✓

QA Review Signature: **SOHIL JODHANI** **Date:** **11/12/2025**

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

LAB CHRONICLE

OrderID:	Q3450	**OrderDate:**	10/23/2025 3:53:00 PM
Client:	Matrix New World Engineering	**Project:**	PSEG Athenia Substation
Contact:	Christopher Pittarese	**Location:**	D41

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q3450-01	**SS-5R-(2.0-2.5)**	**SOIL**			**10/23/25**			**10/23/25**
			PCB	8082A		10/24/25	10/27/25	
Q3450-02	**SS-5R-(3.5-4.0)**	**SOIL**			**10/23/25**			**10/23/25**
			PCB	8082A		10/24/25	10/27/25	
Q3450-04	**SS-10R-(3.5-4.0)**	**SOIL**			**10/23/25**			**10/23/25**
			PCB	8082A		10/31/25	10/31/25	
Q3450-05	**SS-13R-(3.5-4.0)**	**SOIL**			**10/23/25**			**10/23/25**
			PCB	8082A		10/31/25	11/01/25	
Q3450-06	**SS-13R-(5.0-5.5)**	**SOIL**			**10/23/25**			**10/23/25**
			PCB	8082A		10/31/25	11/01/25	
Q3450-07	**SS-13R-(4.5-5.0)**	**SOIL**			**10/23/25**			**10/23/25**
			PCB	8082A		11/12/25	11/12/25	
Q3450-08	**SS-20-(0.0-0.5)**	**SOIL**			**10/23/25**			**10/23/25**
			PCB	8082A		10/24/25	10/27/25	
Q3450-09	**SS-20-(2.0-2.5)**	**SOIL**			**10/23/25**			**10/23/25**
			PCB	8082A		10/24/25	10/27/25	
Q3450-10	**SS-20-(3.5-4.0)**	**SOIL**			**10/23/25**			**10/23/25**
			PCB	8082A		10/24/25	10/24/25	
Q3450-11	**SS-28-(2.0-2.5)**	**SOIL**			**10/23/25**			**10/23/25**
			PCB	8082A		10/24/25	10/24/25	
Q3450-20	**SS-21-(0.0-0.5)**	**SOIL**			**10/23/25**			**10/23/25**
			PCB	8082A		10/24/25	10/24/25	
Q3450-20DL	**SS-21-(0.0-0.5)DL**	**SOIL**			**10/23/25**			**10/23/25**

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

LAB CHRONICLE

			PCB	8082A		10/24/25	10/27/25	
Q3450-21	**SS-21-(2.0-2.5)**	**SOIL**			**10/23/25**			**10/23/25**
			PCB	8082A		10/24/25	10/24/25	
Q3450-22	**SS-21-(3.5-4.0)**	**SOIL**			**10/23/25**			**10/23/25**
			PCB	8082A		10/24/25	10/24/25	
Q3450-22DL	**SS-21-(3.5-4.0)DL**	**SOIL**			**10/23/25**			**10/23/25**
			PCB	8082A		10/24/25	10/27/25	
Q3450-23	**SS-22-(0.0-0.5)**	**SOIL**			**10/23/25**			**10/23/25**
			PCB	8082A		10/31/25	11/01/25	
Q3450-23DL	**SS-22-(0.0-0.5)DL**	**SOIL**			**10/23/25**			**10/23/25**
			PCB	8082A		10/31/25	11/03/25	
Q3450-25	**SS-22-(3.5-4.0)**	**SOIL**			**10/23/25**			**10/23/25**
			PCB	8082A		10/31/25	11/01/25	

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Hit Summary Sheet
SW-846

SDG No.: Q3450 **Order ID:** Q3450

Client: Matrix New World Engineering **Project ID:** PSEG Athenia Substation

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Client ID :	**SS-5R-(2.0-2.5)**							
Q3450-01	SS-5R-(2.0-2.5)	SOIL	Aroclor-1254	45.5		3.50	18.3	ug/kg
			Total Concentration:	**45.500**				
Client ID :	**SS-5R-(3.5-4.0)**							
Q3450-02	SS-5R-(3.5-4.0)	SOIL	Aroclor-1254	11.3	J	3.70	19.6	ug/kg
			Total Concentration:	**11.300**				
Client ID :	**SS-13R-(4.5-5.0)**							
Q3450-07	SS-13R-(4.5-5.0)	SOIL	Aroclor-1254	13.8	J	3.80	19.9	ug/kg
			Total Concentration:	**13.800**				
Client ID :	**SS-20-(0.0-0.5)**							
Q3450-08	SS-20-(0.0-0.5)	SOIL	Aroclor-1254	174		3.40	18.2	ug/kg
			Total Concentration:	**174.000**				
Client ID :	**SS-20-(2.0-2.5)**							
Q3450-09	SS-20-(2.0-2.5)	SOIL	Aroclor-1254	11.3	J	3.70	19.6	ug/kg
			Total Concentration:	**11.300**				
Client ID :	**SS-20-(3.5-4.0)**							
Q3450-10	SS-20-(3.5-4.0)	SOIL	Aroclor-1254	55.2		3.80	20.2	ug/kg
			Total Concentration:	**55.200**				
Client ID :	**SS-21-(0.0-0.5)**							
Q3450-20	SS-21-(0.0-0.5)	SOIL	Aroclor-1254	593	E	3.50	18.5	ug/kg
			Total Concentration:	**593.000**				
Client ID :	**SS-21-(0.0-0.5)DL**							
Q3450-20DL	SS-21-(0.0-0.5)DL	SOIL	Aroclor-1254	509	D	17.5	92.5	ug/kg
			Total Concentration:	**509.000**				
Client ID :	**SS-21-(3.5-4.0)**							
Q3450-22	SS-21-(3.5-4.0)	SOIL	Aroclor-1254	516	E	3.80	20.0	ug/kg
			Total Concentration:	**516.000**				
Client ID :	**SS-21-(3.5-4.0)DL**							
Q3450-22DL	SS-21-(3.5-4.0)DL	SOIL	Aroclor-1254	503	D	7.60	40.1	ug/kg

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Hit Summary Sheet
SW-846

SDG No.: Q3450 **Order ID:** Q3450

Client: Matrix New World Engineering **Project ID:** PSEG Athenia Substation

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
			Total Concentration:	**503.000**				
Client ID :	**SS-22-(0.0-0.5)**							
Q3450-23	SS-22-(0.0-0.5)	SOIL	Aroclor-1254	1200	E	3.50	18.6	ug/kg
			Total Concentration:	**1,200.000**				
Client ID :	**SS-22-(0.0-0.5)DL**							
Q3450-23DL	SS-22-(0.0-0.5)DL	SOIL	Aroclor-1254	881	D	17.5	92.8	ug/kg
			Total Concentration:	**881.000**				
Client ID :	**SS-22-(3.5-4.0)**							
Q3450-25	SS-22-(3.5-4.0)	SOIL	Aroclor-1254	381		3.90	20.5	ug/kg
			Total Concentration:	**381.000**				



QC SUMMARY

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Surrogate Summary

SDG No.: Q3450

Client: Matrix New World Engineering

Analytical Method: 8082A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Recovery(%)	Qual	Limits(%) Low	Limits(%) High
I.BLK-PO114870.D	PIBLK-PO114870.D	Tetrachloro-m-xyl	1	20	15.5	77		70 (60)	130 (140)
		Decachlorobiphen	1	20	16.8	84		70 (60)	130 (140)
		Tetrachloro-m-xyl	2	20	16.2	81		70 (60)	130 (140)
		Decachlorobiphen	2	20	17.3	86		70 (60)	130 (140)
I.BLK-PO115035.D	PIBLK-PO115035.D	Tetrachloro-m-xyl	1	20	19.6	98		70 (60)	130 (140)
		Decachlorobiphen	1	20	19.5	98		70 (60)	130 (140)
		Tetrachloro-m-xyl	2	20	19.8	99		70 (60)	130 (140)
		Decachlorobiphen	2	20	21.2	106		70 (60)	130 (140)
PB170497BL	PB170497BL	Tetrachloro-m-xyl	1	20	22.0	110		30 (21)	150 (165)
		Decachlorobiphen	1	20	21.6	108		30 (10)	150 (170)
		Tetrachloro-m-xyl	2	20	21.4	107		30 (21)	150 (165)
		Decachlorobiphen	2	20	23.0	115		30 (10)	150 (170)
PB170497BS	PB170497BS	Tetrachloro-m-xyl	1	20	21.8	109		30 (21)	150 (165)
		Decachlorobiphen	1	20	21.3	107		30 (10)	150 (170)
		Tetrachloro-m-xyl	2	20	20.2	101		30 (21)	150 (165)
		Decachlorobiphen	2	20	22.6	113		30 (10)	150 (170)
Q3597-01MS	TP-1MS	Tetrachloro-m-xyl	1	20	25.9	130		30 (30)	150 (150)
		Decachlorobiphen	1	20	24.3	122		30 (30)	150 (150)
		Tetrachloro-m-xyl	2	20	25.0	125		30 (30)	150 (150)
		Decachlorobiphen	2	20	25.9	130		30 (30)	150 (150)
Q3597-01MSD	TP-1MSD	Tetrachloro-m-xyl	1	20	26.0	130		30 (30)	150 (150)
		Decachlorobiphen	1	20	24.2	121		30 (30)	150 (150)
		Tetrachloro-m-xyl	2	20	25.1	125		30 (30)	150 (150)
		Decachlorobiphen	2	20	26.1	130		30 (30)	150 (150)
Q3450-07	SS-13R-(4.5-5.0)	Tetrachloro-m-xyl	1	20	23.1	116		30 (21)	150 (165)
		Decachlorobiphen	1	20	21.1	106		30 (10)	150 (170)
		Tetrachloro-m-xyl	2	20	23.4	117		30 (21)	150 (165)
		Decachlorobiphen	2	20	22.9	115		30 (10)	150 (170)
I.BLK-PO115050.D	PIBLK-PO115050.D	Tetrachloro-m-xyl	1	20	19.6	98		70 (60)	130 (140)
		Decachlorobiphen	1	20	19.6	98		70 (60)	130 (140)
		Tetrachloro-m-xyl	2	20	19.6	98		70 (60)	130 (140)
		Decachlorobiphen	2	20	20.9	105		70 (60)	130 (140)
I.BLK-PP075887.D	PIBLK-PP075887.D	Tetrachloro-m-xyl	1	20	19.9	100		70 (60)	130 (140)
		Decachlorobiphen	1	20	20.4	102		70 (60)	130 (140)
		Tetrachloro-m-xyl	2	20	18.8	94		70 (60)	130 (140)
		Decachlorobiphen	2	20	19.3	96		70 (60)	130 (140)
I.BLK-PP076046.D	PIBLK-PP076046.D	Tetrachloro-m-xyl	1	20	20.6	103		70 (60)	130 (140)
		Decachlorobiphen	1	20	23.6	118		70 (60)	130 (140)
		Tetrachloro-m-xyl	2	20	17.6	88		70 (60)	130 (140)
		Decachlorobiphen	2	20	20.1	101		70 (60)	130 (140)
Q3450-10	SS-20-(3.5-4.0)	Tetrachloro-m-xyl	1	20	24.1	121		30 (21)	150 (165)

( ) = LABORATORY INHOUSE LIMIT

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Surrogate Summary

SDG No.: Q3450

Client: Matrix New World Engineering

Analytical Method: 8082A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Recovery(%)	Qual	Limits(%) Low	Limits(%) High
Q3450-10	SS-20-(3.5-4.0)	Decachlorobiphen	1	20	25.8	129		30 (10)	150 (170)
		Tetrachloro-m-xyl	2	20	23.0	115		30 (21)	150 (165)
		Decachlorobiphen	2	20	22.4	112		30 (10)	150 (170)
Q3450-11	SS-28-(2.0-2.5)	Tetrachloro-m-xyl	1	20	23.3	116		30 (21)	150 (165)
		Decachlorobiphen	1	20	25.0	125		30 (10)	150 (170)
		Tetrachloro-m-xyl	2	20	22.7	113		30 (21)	150 (165)
		Decachlorobiphen	2	20	22.0	110		30 (10)	150 (170)
Q3450-20	SS-21-(0.0-0.5)	Tetrachloro-m-xyl	1	20	24.0	120		30 (21)	150 (165)
		Decachlorobiphen	1	20	25.0	125		30 (10)	150 (170)
		Tetrachloro-m-xyl	2	20	22.6	113		30 (21)	150 (165)
		Decachlorobiphen	2	20	21.8	109		30 (10)	150 (170)
Q3450-21	SS-21-(2.0-2.5)	Tetrachloro-m-xyl	1	20	24.5	123		30 (21)	150 (165)
		Decachlorobiphen	1	20	26.4	132		30 (10)	150 (170)
		Tetrachloro-m-xyl	2	20	22.6	113		30 (21)	150 (165)
		Decachlorobiphen	2	20	23.4	117		30 (10)	150 (170)
Q3450-22	SS-21-(3.5-4.0)	Tetrachloro-m-xyl	1	20	20.1	100		30 (21)	150 (165)
		Decachlorobiphen	1	20	17.2	86		30 (10)	150 (170)
		Tetrachloro-m-xyl	2	20	19.0	95		30 (21)	150 (165)
		Decachlorobiphen	2	20	15.1	76		30 (10)	150 (170)
I.BLK-PP076061.D	PIBLK-PP076061.D	Tetrachloro-m-xyl	1	20	20.9	104		70 (60)	130 (140)
		Decachlorobiphen	1	20	23.7	119		70 (60)	130 (140)
		Tetrachloro-m-xyl	2	20	18.2	91		70 (60)	130 (140)
		Decachlorobiphen	2	20	21.0	105		70 (60)	130 (140)
I.BLK-PP076067.D	PIBLK-PP076067.D	Tetrachloro-m-xyl	1	20	22.1	110		70 (60)	130 (140)
		Decachlorobiphen	1	20	22.4	112		70 (60)	130 (140)
		Tetrachloro-m-xyl	2	20	18.9	95		70 (60)	130 (140)
		Decachlorobiphen	2	20	18.2	91		70 (60)	130 (140)
PB170236BL	PB170236BL	Tetrachloro-m-xyl	1	20	27.3	137		30 (21)	150 (165)
		Decachlorobiphen	1	20	29.0	145		30 (10)	150 (170)
		Tetrachloro-m-xyl	2	20	26.1	130		30 (21)	150 (165)
		Decachlorobiphen	2	20	25.2	126		30 (10)	150 (170)
PB170236BS	PB170236BS	Tetrachloro-m-xyl	1	20	26.7	134		30 (21)	150 (165)
		Decachlorobiphen	1	20	28.6	143		30 (10)	150 (170)
		Tetrachloro-m-xyl	2	20	21.3	107		30 (21)	150 (165)
		Decachlorobiphen	2	20	24.3	121		30 (10)	150 (170)
Q3450-20DL	SS-21-(0.0-0.5)DL	Tetrachloro-m-xyl	1	20	21.4	107		30 (21)	150 (165)
		Decachlorobiphen	1	20	24.8	124		30 (10)	150 (170)
		Tetrachloro-m-xyl	2	20	20.5	103		30 (21)	150 (165)
		Decachlorobiphen	2	20	19.4	97		30 (10)	150 (170)
Q3450-22DL	SS-21-(3.5-4.0)DL	Tetrachloro-m-xyl	1	20	21.6	108		30 (21)	150 (165)
		Decachlorobiphen	1	20	17.5	88		30 (10)	150 (170)

( ) = LABORATORY INHOUSE LIMIT

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Surrogate Summary

SDG No.: Q3450

Client: Matrix New World Engineering

Analytical Method: 8082A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Recovery(%)	Qual	Limits(%) Low	Limits(%) High
Q3450-22DL	SS-21-(3.5-4.0)DL	Tetrachloro-m-xyl	2	20	19.7	99		30 (21)	150 (165)
		Decachlorobiphen	2	20	13.8	69		30 (10)	150 (170)
I.BLK-PP076082.D	PIBLK-PP076082.D	Tetrachloro-m-xyl	1	20	18.9	95		70 (60)	130 (140)
		Decachlorobiphen	1	20	20.4	102		70 (60)	130 (140)
		Tetrachloro-m-xyl	2	20	17.1	86		70 (60)	130 (140)
		Decachlorobiphen	2	20	16.7	83		70 (60)	130 (140)
I.BLK-PP076275.D	PIBLK-PP076275.D	Tetrachloro-m-xyl	1	20	21.6	108		70 (60)	130 (140)
		Decachlorobiphen	1	20	21.0	105		70 (60)	130 (140)
		Tetrachloro-m-xyl	2	20	19.3	97		70 (60)	130 (140)
		Decachlorobiphen	2	20	17.5	87		70 (60)	130 (140)
PB170358BL	PB170358BL	Tetrachloro-m-xyl	1	20	28.4	142		30 (21)	150 (165)
		Decachlorobiphen	1	20	27.5	138		30 (10)	150 (170)
		Tetrachloro-m-xyl	2	20	25.8	129		30 (21)	150 (165)
		Decachlorobiphen	2	20	23.1	115		30 (10)	150 (170)
PB170358BS	PB170358BS	Tetrachloro-m-xyl	1	20	27.9	140		30 (21)	150 (165)
		Decachlorobiphen	1	20	25.9	130		30 (10)	150 (170)
		Tetrachloro-m-xyl	2	20	24.7	123		30 (21)	150 (165)
		Decachlorobiphen	2	20	21.8	109		30 (10)	150 (170)
Q3450-04	SS-10R-(3.5-4.0)	Tetrachloro-m-xyl	1	20	17.1	86		30 (21)	150 (165)
		Decachlorobiphen	1	20	13.4	67		30 (10)	150 (170)
		Tetrachloro-m-xyl	2	20	15.2	76		30 (21)	150 (165)
		Decachlorobiphen	2	20	10.9	55		30 (10)	150 (170)
Q3450-04MS	SS-10R-(3.5-4.0)MS	Tetrachloro-m-xyl	1	20	25.4	127		30 (21)	150 (165)
		Decachlorobiphen	1	20	22.3	112		30 (10)	150 (170)
		Tetrachloro-m-xyl	2	20	22.8	114		30 (21)	150 (165)
		Decachlorobiphen	2	20	19.4	97		30 (10)	150 (170)
Q3450-04MSD	SS-10R-(3.5-4.0)MSD	Tetrachloro-m-xyl	1	20	24.5	123		30 (21)	150 (165)
		Decachlorobiphen	1	20	21.6	108		30 (10)	150 (170)
		Tetrachloro-m-xyl	2	20	21.9	110		30 (21)	150 (165)
		Decachlorobiphen	2	20	18.2	91		30 (10)	150 (170)
Q3450-05	SS-13R-(3.5-4.0)	Tetrachloro-m-xyl	1	20	22.7	114		30 (21)	150 (165)
		Decachlorobiphen	1	20	21.2	106		30 (10)	150 (170)
		Tetrachloro-m-xyl	2	20	21.4	107		30 (21)	150 (165)
		Decachlorobiphen	2	20	17.6	88		30 (10)	150 (170)
Q3450-06	SS-13R-(5.0-5.5)	Tetrachloro-m-xyl	1	20	22.4	112		30 (21)	150 (165)
		Decachlorobiphen	1	20	20.1	101		30 (10)	150 (170)
		Tetrachloro-m-xyl	2	20	21.6	108		30 (21)	150 (165)
		Decachlorobiphen	2	20	16.8	84		30 (10)	150 (170)
Q3450-23	SS-22-(0.0-0.5)	Tetrachloro-m-xyl	1	20	22.0	110		30 (21)	150 (165)
		Decachlorobiphen	1	20	19.4	97		30 (10)	150 (170)
		Tetrachloro-m-xyl	2	20	20.4	102		30 (21)	150 (165)

( ) = LABORATORY INHOUSE LIMIT

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Surrogate Summary

SDG No.: Q3450

Client: Matrix New World Engineering

Analytical Method: 8082A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Recovery(%)	Qual	Limits(%) Low	Limits(%) High
Q3450-23	SS-22-(0.0-0.5)	Decachlorobiphen	2	20	15.9	80		30 (10)	150 (170)
Q3450-25	SS-22-(3.5-4.0)	Tetrachloro-m-xyl	1	20	19.7	98		30 (21)	150 (165)
		Decachlorobiphen	1	20	13.6	68		30 (10)	150 (170)
		Tetrachloro-m-xyl	2	20	18.4	92		30 (21)	150 (165)
		Decachlorobiphen	2	20	11.0	55		30 (10)	150 (170)
I.BLK-PP076290.D	PIBLK-PP076290.D	Tetrachloro-m-xyl	1	20	22.0	110		70 (60)	130 (140)
		Decachlorobiphen	1	20	21.1	106		70 (60)	130 (140)
		Tetrachloro-m-xyl	2	20	19.2	96		70 (60)	130 (140)
		Decachlorobiphen	2	20	17.5	88		70 (60)	130 (140)
I.BLK-PP076296.D	PIBLK-PP076296.D	Tetrachloro-m-xyl	1	20	20.1	101		70 (60)	130 (140)
		Decachlorobiphen	1	20	21.1	105		70 (60)	130 (140)
		Tetrachloro-m-xyl	2	20	17.7	88		70 (60)	130 (140)
		Decachlorobiphen	2	20	18.8	94		70 (60)	130 (140)
Q3450-23DL	SS-22-(0.0-0.5)DL	Tetrachloro-m-xyl	1	20	15.8	79		30 (21)	150 (165)
		Decachlorobiphen	1	20	16.4	82		30 (10)	150 (170)
		Tetrachloro-m-xyl	2	20	11.3	57		30 (21)	150 (165)
		Decachlorobiphen	2	20	11.9	59		30 (10)	150 (170)
I.BLK-PP076308.D	PIBLK-PP076308.D	Tetrachloro-m-xyl	1	20	20.4	102		70 (60)	130 (140)
		Decachlorobiphen	1	20	19.9	99		70 (60)	130 (140)
		Tetrachloro-m-xyl	2	20	19.0	95		70 (60)	130 (140)
		Decachlorobiphen	2	20	15.1	76		70 (60)	130 (140)
I.BLK-PQ070972.D	PIBLK-PQ070972.D	Tetrachloro-m-xyl	1	20	18.3	91		70 (60)	130 (140)
		Decachlorobiphen	1	20	19.0	95		70 (60)	130 (140)
		Tetrachloro-m-xyl	2	20	19.1	96		70 (60)	130 (140)
		Decachlorobiphen	2	20	19.4	97		70 (60)	130 (140)
I.BLK-PQ071150.D	PIBLK-PQ071150.D	Tetrachloro-m-xyl	1	20	22.5	113		70 (60)	130 (140)
		Decachlorobiphen	1	20	25.7	129		70 (60)	130 (140)
		Tetrachloro-m-xyl	2	20	19.9	100		70 (60)	130 (140)
		Decachlorobiphen	2	20	24.5	122		70 (60)	130 (140)
Q3450-01	SS-5R-(2.0-2.5)	Tetrachloro-m-xyl	1	20	27.2	136		30 (21)	150 (165)
		Decachlorobiphen	1	20	28.0	140		30 (10)	150 (170)
		Tetrachloro-m-xyl	2	20	24.5	122		30 (21)	150 (165)
		Decachlorobiphen	2	20	27.0	135		30 (10)	150 (170)
Q3450-02	SS-5R-(3.5-4.0)	Tetrachloro-m-xyl	1	20	27.0	135		30 (21)	150 (165)
		Decachlorobiphen	1	20	25.4	127		30 (10)	150 (170)
		Tetrachloro-m-xyl	2	20	23.8	119		30 (21)	150 (165)
		Decachlorobiphen	2	20	24.9	124		30 (10)	150 (170)
Q3450-02MS	SS-5R-(3.5-4.0)MS	Tetrachloro-m-xyl	1	20	23.4	117		30 (21)	150 (165)
		Decachlorobiphen	1	20	22.4	112		30 (10)	150 (170)
		Tetrachloro-m-xyl	2	20	18.7	94		30 (21)	150 (165)
		Decachlorobiphen	2	20	20.7	104		30 (10)	150 (170)

( ) = LABORATORY INHOUSE LIMIT

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Surrogate Summary

SDG No.: Q3450

Client: Matrix New World Engineering

Analytical Method: 8082A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Recovery(%)	Qual	Limits(%) Low	Limits(%) High
Q3450-02MSD	SS-5R-(3.5-4.0)MSD	Tetrachloro-m-xyl	1	20	25.7	128		30 (21)	150 (165)
		Decachlorobiphen	1	20	24.0	120		30 (10)	150 (170)
		Tetrachloro-m-xyl	2	20	22.2	111		30 (21)	150 (165)
		Decachlorobiphen	2	20	22.9	114		30 (10)	150 (170)
Q3450-08	SS-20-(0.0-0.5)	Tetrachloro-m-xyl	1	20	26.8	134		30 (21)	150 (165)
		Decachlorobiphen	1	20	28.1	140		30 (10)	150 (170)
		Tetrachloro-m-xyl	2	20	25.1	125		30 (21)	150 (165)
		Decachlorobiphen	2	20	26.6	133		30 (10)	150 (170)
Q3450-09	SS-20-(2.0-2.5)	Tetrachloro-m-xyl	1	20	27.2	136		30 (21)	150 (165)
		Decachlorobiphen	1	20	27.4	137		30 (10)	150 (170)
		Tetrachloro-m-xyl	2	20	24.9	125		30 (21)	150 (165)
		Decachlorobiphen	2	20	26.3	132		30 (10)	150 (170)
I.BLK-PQ071165.D	PIBLK-PQ071165.D	Tetrachloro-m-xyl	1	20	20.3	102		70 (60)	130 (140)
		Decachlorobiphen	1	20	23.5	117		70 (60)	130 (140)
		Tetrachloro-m-xyl	2	20	18.4	92		70 (60)	130 (140)
		Decachlorobiphen	2	20	22.2	111		70 (60)	130 (140)

( ) = LABORATORY INHOUSE LIMIT

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q3450 **Analytical Method:** 8082A

Client: Matrix New World Engineering **DataFile :** PO115040.D

Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Limits Low	Limits High	RPD
Lab Sample ID: **Q3597-01MS** **Client Sample ID:** **TP-1MS** **(Column 1)**											
AR1016	170.9	0	160	ug/kg	94				40 (40)	140 (140)	
AR1260	170.9	0	146	ug/kg	85				40 (40)	140 (140)	
Lab Sample ID: **Q3597-01MS** **Client Sample ID:** **TP-1MS** **(Column 2)**											
AR1016	170.9	0	157	ug/kg	92				40 (40)	140 (140)	
AR1260	170.9	0	145	ug/kg	85				40 (40)	140 (140)	

( ) = LABORATORY INHOUSE LIMIT

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q3450 **Analytical Method:** 8082A

Client: Matrix New World Engineering **DataFile :** PO115041.D

Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Limits Low	Limits High	Limits RPD
Lab Sample ID: Q3597-01MSD (Column 1)	**Client Sample ID: TP-1MSD**										
AR1016	170.7	0	158	ug/kg	93		1		40 (40)	140 (140)	30 (30)
AR1260	170.7	0	147	ug/kg	86		1		40 (40)	140 (140)	30 (30)
Lab Sample ID: Q3597-01MSD (Column 2)	**Client Sample ID: TP-1MSD**										
AR1016	170.7	0	158	ug/kg	93		1		40 (40)	140 (140)	30 (30)
AR1260	170.7	0	145	ug/kg	85		0		40 (40)	140 (140)	30 (30)

( ) = LABORATORY INHOUSE LIMIT

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q3450 **Analytical Method:** 8082A

Client: Matrix New World Engineering **DataFile :** PP076280.D

	Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Limits Low	Limits High	Limits RPD
Lab Sample ID:	**Q3450-04MS (Column 1)**	**Client Sample ID:**		**SS-10R-(3.5-4.0)MS**								
	AR1016	192.3	0	193	ug/kg	100				40 (39)	140 (159)	
	AR1260	192.3	0	221	ug/kg	115				40 (19)	140 (160)	
Lab Sample ID:	**Q3450-04MS (Column 2)**	**Client Sample ID:**		**SS-10R-(3.5-4.0)MS**								
	AR1016	192.3	0	175	ug/kg	91				40 (39)	140 (159)	
	AR1260	192.3	0	209	ug/kg	109				40 (19)	140 (160)	

( ) = LABORATORY INHOUSE LIMIT

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q3450 **Analytical Method:** 8082A

Client: Matrix New World Engineering **DataFile :** PP076281.D

Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Limits Low	Limits High	Limits RPD
Lab Sample ID: Q3450-04MSD (Column 1)	**Client Sample ID: SS-10R-(3.5-4.0)MSD**										
AR1016	192.2	0	189	ug/kg	98		2		40 (39)	140 (159)	30 (21)
AR1260	192.2	0	215	ug/kg	112		3		40 (19)	140 (160)	30 (15)
Lab Sample ID: Q3450-04MSD (Column 2)	**Client Sample ID: SS-10R-(3.5-4.0)MSD**										
AR1016	192.2	0	173	ug/kg	90		1		40 (39)	140 (159)	30 (21)
AR1260	192.2	0	206	ug/kg	107		2		40 (19)	140 (160)	30 (15)

( ) = LABORATORY INHOUSE LIMIT

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q3450 **Analytical Method:** 8082A

Client: Matrix New World Engineering **DataFile :** PQ071153.D

Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Limits Low	Limits High	RPD
Lab Sample ID: **Q3450-02MS (Column 1)**	**Client Sample ID:**		**SS-5R-(3.5-4.0)MS**								
AR1016	192.7	0	121	ug/kg	63				40 (39)	140 (159)	
AR1260	192.7	0	145	ug/kg	75				40 (19)	140 (160)	
Lab Sample ID: **Q3450-02MS (Column 2)**	**Client Sample ID:**		**SS-5R-(3.5-4.0)MS**								
AR1016	192.7	0	118	ug/kg	61				40 (39)	140 (159)	
AR1260	192.7	0	123	ug/kg	64				40 (19)	140 (160)	

( ) = LABORATORY INHOUSE LIMIT

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q3450 **Analytical Method:** 8082A

Client: Matrix New World Engineering **DataFile :** PQ071154.D

Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Limits Low	Limits High	Limits RPD
Lab Sample ID: Q3450-02MSD (Column 1)	**Client Sample ID: SS-5R-(3.5-4.0)MSD**										
AR1016	192.9	0	162	ug/kg	84		29		40 (39)	140 (159)	30 (21)
AR1260	192.9	0	166	ug/kg	86		14		40 (19)	140 (160)	30 (15)
Lab Sample ID: Q3450-02MSD (Column 2)	**Client Sample ID: SS-5R-(3.5-4.0)MSD**										
AR1016	192.9	0	143	ug/kg	74		19		40 (39)	140 (159)	30 (21)
AR1260	192.9	0	153	ug/kg	79		21		40 (19)	140 (160)	30 (15)

( ) = LABORATORY INHOUSE LIMIT

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q3450 **Analytical Method:** 8082A

Client: Matrix New World Engineering **Datafile :** PO115038.D

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD Qual	Limits Low	Limits High	RPD
PB170497BS (Column 1)	AR1016	166.5	145	ug/kg	87				40 (71)	140 (120)	
	AR1260	166.5	137	ug/kg	82				40 (65)	140 (130)	
PB170497BS (Column 2)	AR1016	166.5	142	ug/kg	85				40 (71)	140 (120)	
	AR1260	166.5	133	ug/kg	80				40 (65)	140 (130)	

( ) = LABORATORY INHOUSE LIMIT

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q3450 **Analytical Method:** 8082A

Client: Matrix New World Engineering **Datafile :** PP076069.D

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD Qual	Limits Low	Limits High	RPD
PB170236BS (Column 1)	AR1016	166.6	145	ug/kg	87				40 (71)	140 (120)	
	AR1260	166.6	149	ug/kg	89				40 (65)	140 (130)	
PB170236BS (Column 2)	AR1016	166.6	127	ug/kg	76				40 (71)	140 (120)	
	AR1260	166.6	143	ug/kg	86				40 (65)	140 (130)	

( ) = LABORATORY INHOUSE LIMIT

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q3450 **Analytical Method:** 8082A

Client: Matrix New World Engineering **Datafile :** PP076277.D

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD Qual	Limits Low	Limits High	RPD
PB170358BS (Column 1)	AR1016	166.5	170	ug/kg	102				40 (71)	140 (120)	
	AR1260	166.5	166	ug/kg	100				40 (65)	140 (130)	
PB170358BS (Column 2)	AR1016	166.5	151	ug/kg	91				40 (71)	140 (120)	
	AR1260	166.5	168	ug/kg	101				40 (65)	140 (130)	

( ) = LABORATORY INHOUSE LIMIT

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

4C

PESTICIDE METHOD BLANK SUMMARY

Client ID: PB170236BL

Lab Name: Alliance | Contract: MATR02

Lab Code: ACE | SDG NO.: Q3450

Lab Sample ID: PB170236BL | Lab File ID: PP076068.D

Matrix: (soil/water) Solid | Extraction: (Type) SOXH

Sulfur Cleanup: (Y/N) N | Date Extracted: 10/24/2025

Date Analyzed (1): 10/27/2025 | Date Analyzed (2): 10/27/2025

Time Analyzed (1): 10:18 | Time Analyzed (2): 10:18

Instrument ID (1): ECD_P | Instrument ID (2): ECD_P

GC Column (1): ZB-MR1 ID: 0.32 (mm) | GC Column (2): ZB-MR2 ID: 0.32 (mm)

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED 1	DATE ANALYZED 2
SS-20-(3.5-4.0)	Q3450-10	PP076047.D	10/24/2025	10/24/2025
SS-28-(2.0-2.5)	Q3450-11	PP076048.D	10/24/2025	10/24/2025
SS-21-(0.0-0.5)	Q3450-20	PP076049.D	10/24/2025	10/24/2025
SS-21-(2.0-2.5)	Q3450-21	PP076050.D	10/24/2025	10/24/2025
SS-21-(3.5-4.0)	Q3450-22	PP076051.D	10/24/2025	10/24/2025
PB170236BS	PB170236BS	PP076069.D	10/27/2025	10/27/2025
SS-5R-(2.0-2.5)	Q3450-01	PQ071151.D	10/27/2025	10/27/2025
SS-5R-(3.5-4.0)	Q3450-02	PQ071152.D	10/27/2025	10/27/2025
SS-5R-(3.5-4.0)MS	Q3450-02MS	PQ071153.D	10/27/2025	10/27/2025
SS-5R-(3.5-4.0)MSD	Q3450-02MSD	PQ071154.D	10/27/2025	10/27/2025
SS-20-(0.0-0.5)	Q3450-08	PQ071155.D	10/27/2025	10/27/2025
SS-20-(2.0-2.5)	Q3450-09	PQ071156.D	10/27/2025	10/27/2025

COMMENTS:

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

4C

PESTICIDE METHOD BLANK SUMMARY

Client ID: PB170358BL

Lab Name: Alliance | Contract: MATR02

Lab Code: ACE | SDG NO.: Q3450

Lab Sample ID: PB170358BL | Lab File ID: PP076276.D

Matrix: (soil/water) Solid | Extraction: (Type) SOXH

Sulfur Cleanup: (Y/N) N | Date Extracted: 10/31/2025

Date Analyzed (1): 10/31/2025 | Date Analyzed (2): 10/31/2025

Time Analyzed (1): 22:51 | Time Analyzed (2): 22:51

Instrument ID (1): ECD_P | Instrument ID (2): ECD_P

GC Column (1): ZB-MR1 ID: 0.32 (mm) | GC Column (2): ZB-MR2 ID: 0.32 (mm)

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED 1	DATE ANALYZED 2
PB170358BS	PB170358BS	PP076277.D	10/31/2025	10/31/2025
SS-10R-(3.5-4.0)	Q3450-04	PP076279.D	10/31/2025	10/31/2025
SS-10R-(3.5-4.0)MS	Q3450-04MS	PP076280.D	10/31/2025	10/31/2025
SS-10R-(3.5-4.0)MSD	Q3450-04MSD	PP076281.D	11/01/2025	11/01/2025
SS-13R-(3.5-4.0)	Q3450-05	PP076282.D	11/01/2025	11/01/2025
SS-13R-(5.0-5.5)	Q3450-06	PP076283.D	11/01/2025	11/01/2025
SS-22-(0.0-0.5)	Q3450-23	PP076284.D	11/01/2025	11/01/2025
SS-22-(3.5-4.0)	Q3450-25	PP076285.D	11/01/2025	11/01/2025

COMMENTS:

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

4C

PESTICIDE METHOD BLANK SUMMARY

Client ID: PB170497BL

Lab Name: Alliance | Contract: MATR02

Lab Code: ACE | SDG NO.: Q3450

Lab Sample ID: PB170497BL | Lab File ID: PO115037.D

Matrix: (soil/water) Solid | Extraction: (Type) SOXH

Sulfur Cleanup: (Y/N) N | Date Extracted: 11/12/2025

Date Analyzed (1): 11/12/2025 | Date Analyzed (2): 11/12/2025

Time Analyzed (1): 11:58 | Time Analyzed (2): 11:58

Instrument ID (1): ECD_O | Instrument ID (2): ECD_O

GC Column (1): ZB-MR1 ID: 0.32 (mm) | GC Column (2): ZB-MR2 ID: 0.32 (mm)

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED 1	DATE ANALYZED 2
PB170497BS	PB170497BS	PO115038.D	11/12/2025	11/12/2025
TP-1MS	Q3597-01MS	PO115040.D	11/12/2025	11/12/2025
TP-1MSD	Q3597-01MSD	PO115041.D	11/12/2025	11/12/2025
SS-13R-(4.5-5.0)	Q3450-07	PO115045.D	11/12/2025	11/12/2025

COMMENTS:



SAMPLE DATA

Raw Data: **PQ071151.D**

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Matrix New World Engineering	Date Collected:	10/23/25
Project:	PSEG Athenia Substation	Date Received:	10/23/25
Client Sample ID:	SS-5R-(2.0-2.5)	SDG No.:	Q3450
Lab Sample ID:	Q3450-01	Matrix:	SOIL
Analytical Method:	8082A	% Solid:	92.9
Sample Wt/Vol:	30.03 g	Final Vol: 10000 uL	Test: PCB
Prep Method:	SW3541B	Prep Date: 10/24/25	

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
12674-11-2	Aroclor-1016	4.20	U	1	4.20	18.3	ug/kg	10/27/25 09:26	PB170236
11104-28-2	Aroclor-1221	4.30	U	1	4.30	18.3	ug/kg	10/27/25 09:26	PB170236
11141-16-5	Aroclor-1232	4.00	U	1	4.00	18.3	ug/kg	10/27/25 09:26	PB170236
53469-21-9	Aroclor-1242	4.30	U	1	4.30	18.3	ug/kg	10/27/25 09:26	PB170236
12672-29-6	Aroclor-1248	6.40	U	1	6.40	18.3	ug/kg	10/27/25 09:26	PB170236
11097-69-1	Aroclor-1254	45.5		1	3.50	18.3	ug/kg	10/27/25 09:26	PB170236
37324-23-5	Aroclor-1262	5.40	U	1	5.40	18.3	ug/kg	10/27/25 09:26	PB170236
11100-14-4	Aroclor-1268	3.90	U	1	3.90	18.3	ug/kg	10/27/25 09:26	PB170236
11096-82-5	Aroclor-1260	3.50	U	1	3.50	18.3	ug/kg	10/27/25 09:26	PB170236
SURROGATES									
877-09-8	Tetrachloro-m-xylene	27.2			30 (21) - 150 (165)	136%	SPK: 20		
2051-24-3	Decachlorobiphenyl	28.0			30 (10) - 150 (170)	140%	SPK: 20		

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
E = Value Exceeds Calibration Range
P = Indicates >25% difference for detected concentrations between the two GC columns
Q = indicates LCS control criteria did not meet requirements
M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
B = Analyte Found in Associated Method Blank
N = Presumptive Evidence of a Compound
* = Values outside of QC limits
D = Dilution
S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.
() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ102725\
Data File : PQ071151.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Oct 2025 09:26
Operator : YP\AJ
Sample : Q3450-01
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
SS-5R-(2.0-2.5)

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 27 11:28:52 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Oct 11 01:55:41 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	3.399	2.776	118.4E6	124.4E6	27.208	24.481
2)	SA Decachlor...	8.497	7.531	81424033	136.5E6	28.039	27.033
	Target Compounds						
26)	L6 AR-1254-1	5.293	4.531	4546338	9896259	23.163	29.347 #
27)	L6 AR-1254-2	5.507	4.677	18607410	19639861	64.057	66.093
28)	L6 AR-1254-3	5.855	5.057	42076790	59419953	136.965	127.220
29)	L6 AR-1254-4	6.136	5.284	34093872	45433959	155.297	153.882
30)	L6 AR-1254-5	6.544	5.688	59025852	100.5E6	254.536	239.178

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Sample Results: **PQ071151.D**


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ102725\
Data File : PQ071151.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Oct 2025 09:26
Operator : YP\AJ
Sample : Q3450-01
Misc :
ALS Vial : 11 Sample Multiplier: 1


Instrument :
ECD_Q
ClientSampleId :
SS-5R-(2.0-2.5)

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 27 11:28:52 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Oct 11 01:55:41 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm








#1 Tetrachloro-m-xylene


R.T.: 3.399 min
Delta R.T.: 0.000 min
Response: 118364166
Conc: 27.21 ng/ml

Instrument :
ECD_Q
ClientSampleId :
SS-5R-(2.0-2.5)




#1 Tetrachloro-m-xylene

R.T.: 2.776 min
Delta R.T.: -0.002 min
Response: 124362089
Conc: 24.48 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.497 min
Delta R.T.: -0.003 min
Response: 81424033
Conc: 28.04 ng/ml



#2 Decachlorobiphenyl


R.T.: 7.531 min
Delta R.T.: -0.002 min
Response: 136506492
Conc: 27.03 ng/ml



#26 AR-1254-1


R.T.: 5.293 min
Delta R.T.: 0.000 min
Response: 4546338
Conc: 23.16 ng/ml

Instrument :
ECD_Q
ClientSampleId :
SS-5R-(2.0-2.5)




#26 AR-1254-1

R.T.: 4.531 min
Delta R.T.: -0.002 min
Response: 9896259
Conc: 29.35 ng/ml




#27 AR-1254-2

R.T.: 5.507 min
Delta R.T.: 0.000 min
Response: 18607410
Conc: 64.06 ng/ml



#27 AR-1254-2


R.T.: 4.677 min
Delta R.T.: -0.002 min
Response: 19639861
Conc: 66.09 ng/ml



#28 AR-1254-3


R.T.: 5.855 min
Delta R.T.: 0.000 min
Response: 42076790
Conc: 136.97 ng/ml

Instrument :
ECD_Q
ClientSampleId :
SS-5R-(2.0-2.5)




#28 AR-1254-3

R.T.: 5.057 min
Delta R.T.: -0.002 min
Response: 59419953
Conc: 127.22 ng/ml




#29 AR-1254-4

R.T.: 6.136 min
Delta R.T.: 0.000 min
Response: 34093872
Conc: 155.30 ng/ml



#29 AR-1254-4


R.T.: 5.284 min
Delta R.T.: -0.002 min
Response: 45433959
Conc: 153.88 ng/ml



#30 AR-1254-5

R.T.: 6.544 min
Delta R.T.: 0.000 min
Response: 59025852
Conc: 254.54 ng/ml

Instrument :
ECD_Q
ClientSampleId :
SS-5R-(2.0-2.5)



#30 AR-1254-5

R.T.: 5.688 min
Delta R.T.: -0.002 min
Response: 100538208
Conc: 239.18 ng/ml

Raw Data: **PQ071152.D**

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Matrix New World Engineering	Date Collected:	10/23/25
Project:	PSEG Athenia Substation	Date Received:	10/23/25
Client Sample ID:	SS-5R-(3.5-4.0)	SDG No.:	Q3450
Lab Sample ID:	Q3450-02	Matrix:	SOIL
Analytical Method:	8082A	% Solid:	86.3
Sample Wt/Vol:	30.08 g	Final Vol: 10000 uL	Test: PCB
Prep Method:	SW3541B	Prep Date: 10/24/25	

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
12674-11-2	Aroclor-1016	4.60	U	1	4.60	19.6	ug/kg	10/27/25 09:41	PB170236
11104-28-2	Aroclor-1221	4.70	U	1	4.70	19.6	ug/kg	10/27/25 09:41	PB170236
11141-16-5	Aroclor-1232	4.30	U	1	4.30	19.6	ug/kg	10/27/25 09:41	PB170236
53469-21-9	Aroclor-1242	4.60	U	1	4.60	19.6	ug/kg	10/27/25 09:41	PB170236
12672-29-6	Aroclor-1248	6.80	U	1	6.80	19.6	ug/kg	10/27/25 09:41	PB170236
11097-69-1	Aroclor-1254	11.3	J	1	3.70	19.6	ug/kg	10/27/25 09:41	PB170236
37324-23-5	Aroclor-1262	5.80	U	1	5.80	19.6	ug/kg	10/27/25 09:41	PB170236
11100-14-4	Aroclor-1268	4.20	U	1	4.20	19.6	ug/kg	10/27/25 09:41	PB170236
11096-82-5	Aroclor-1260	3.70	U	1	3.70	19.6	ug/kg	10/27/25 09:41	PB170236
SURROGATES									
877-09-8	Tetrachloro-m-xylene	27.0			30 (21) - 150 (165)	135%	SPK: 20		
2051-24-3	Decachlorobiphenyl	25.4			30 (10) - 150 (170)	127%	SPK: 20		

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
E = Value Exceeds Calibration Range
P = Indicates >25% difference for detected concentrations between the two GC columns
Q = indicates LCS control criteria did not meet requirements
M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
B = Analyte Found in Associated Method Blank
N = Presumptive Evidence of a Compound
* = Values outside of QC limits
D = Dilution
S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.
() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ102725\
Data File : PQ071152.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Oct 2025 09:41
Operator : YP\AJ
Sample : Q3450-02
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
SS-5R-(3.5-4.0)

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 27 11:29:29 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Oct 11 01:55:41 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	3.399	2.777	117.4E6	120.7E6	26.992	23.763
2)	SA Decachlor...	8.497	7.531	73642448	125.7E6	25.359	24.890
	Target Compounds						
26)	L6 AR-1254-1	5.291	4.531	1879086	2547939	9.574	7.556
27)	L6 AR-1254-2	5.507	4.678	5837495	5610499	20.096	18.881
28)	L6 AR-1254-3	5.854	5.058	9901398	13892809	32.230	29.745
29)	L6 AR-1254-4	6.134	5.285	7352340	9923006	33.490	33.609
30)	L6 AR-1254-5	6.543	5.688	12037427	21754901	51.909	51.754


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ102725\
Data File : PQ071152.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Oct 2025 09:41
Operator : YP\AJ
Sample : Q3450-02
Misc :
ALS Vial : 12 Sample Multiplier: 1


Instrument :
ECD_Q
ClientSampleId :
SS-5R-(3.5-4.0)

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 27 11:29:29 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Oct 11 01:55:41 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm








#1 Tetrachloro-m-xylene


R.T.: 3.399 min
Delta R.T.: 0.000 min
Response: 117421207
Conc: 26.99 ng/ml

Instrument :
ECD_Q
ClientSampleId :
SS-5R-(3.5-4.0)




#1 Tetrachloro-m-xylene

R.T.: 2.777 min
Delta R.T.: -0.002 min
Response: 120715928
Conc: 23.76 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.497 min
Delta R.T.: -0.003 min
Response: 73642448
Conc: 25.36 ng/ml



#2 Decachlorobiphenyl


R.T.: 7.531 min
Delta R.T.: -0.002 min
Response: 125683039
Conc: 24.89 ng/ml



#26 AR-1254-1


R.T.: 5.291 min
Delta R.T.: -0.002 min
Response: 1879086
Conc: 9.57 ng/ml

Instrument :
ECD_Q
ClientSampleId :
SS-5R-(3.5-4.0)




#26 AR-1254-1

R.T.: 4.531 min
Delta R.T.: -0.001 min
Response: 2547939
Conc: 7.56 ng/ml




#27 AR-1254-2

R.T.: 5.507 min
Delta R.T.: 0.000 min
Response: 5837495
Conc: 20.10 ng/ml



#27 AR-1254-2


R.T.: 4.678 min
Delta R.T.: -0.002 min
Response: 5610499
Conc: 18.88 ng/ml



#28 AR-1254-3


R.T.: 5.854 min
Delta R.T.: 0.000 min
Response: 9901398
Conc: 32.23 ng/ml

Instrument :
ECD_Q
ClientSampleId :
SS-5R-(3.5-4.0)




#28 AR-1254-3

R.T.: 5.058 min
Delta R.T.: -0.001 min
Response: 13892809
Conc: 29.74 ng/ml




#29 AR-1254-4

R.T.: 6.134 min
Delta R.T.: -0.001 min
Response: 7352340
Conc: 33.49 ng/ml



#29 AR-1254-4


R.T.: 5.285 min
Delta R.T.: 0.000 min
Response: 9923006
Conc: 33.61 ng/ml



#30 AR-1254-5

R.T.: 6.543 min
Delta R.T.: 0.000 min
Response: 12037427
Conc: 51.91 ng/ml

Instrument :
ECD_Q
ClientSampleId :
SS-5R-(3.5-4.0)



#30 AR-1254-5

R.T.: 5.688 min
Delta R.T.: -0.002 min
Response: 21754901
Conc: 51.75 ng/ml

Raw Data: **PP076279.D**

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Matrix New World Engineering	Date Collected:	10/23/25
Project:	PSEG Athenia Substation	Date Received:	10/23/25
Client Sample ID:	SS-10R-(3.5-4.0)	SDG No.:	Q3450
Lab Sample ID:	Q3450-04	Matrix:	SOIL
Analytical Method:	8082A	% Solid:	86.6
Sample Wt/Vol:	30.09 g	Final Vol: 10000 uL	Test: PCB
Prep Method:	SW3541B	Prep Date: 10/31/25	

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
12674-11-2	Aroclor-1016	4.50	U	1	4.50	19.6	ug/kg	10/31/25 23:40	PB170358
11104-28-2	Aroclor-1221	4.60	U	1	4.60	19.6	ug/kg	10/31/25 23:40	PB170358
11141-16-5	Aroclor-1232	4.30	U	1	4.30	19.6	ug/kg	10/31/25 23:40	PB170358
53469-21-9	Aroclor-1242	4.60	U	1	4.60	19.6	ug/kg	10/31/25 23:40	PB170358
12672-29-6	Aroclor-1248	6.80	U	1	6.80	19.6	ug/kg	10/31/25 23:40	PB170358
11097-69-1	Aroclor-1254	3.70	U	1	3.70	19.6	ug/kg	10/31/25 23:40	PB170358
37324-23-5	Aroclor-1262	5.80	U	1	5.80	19.6	ug/kg	10/31/25 23:40	PB170358
11100-14-4	Aroclor-1268	4.10	U	1	4.10	19.6	ug/kg	10/31/25 23:40	PB170358
11096-82-5	Aroclor-1260	3.70	U	1	3.70	19.6	ug/kg	10/31/25 23:40	PB170358
SURROGATES									
877-09-8	Tetrachloro-m-xylene	17.1			30 (21) - 150 (165)	86%	SPK: 20		
2051-24-3	Decachlorobiphenyl	13.4			30 (10) - 150 (170)	67%	SPK: 20		

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
E = Value Exceeds Calibration Range
P = Indicates >25% difference for detected concentrations between the two GC columns
Q = indicates LCS control criteria did not meet requirements
M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
B = Analyte Found in Associated Method Blank
N = Presumptive Evidence of a Compound
* = Values outside of QC limits
D = Dilution
S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.
() = Laboratory InHouse Limit

Quantitation Report (QT Reviewed)

Sample Results: **PP076279.D**

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP103125\
Data File : PP076279.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 31 Oct 2025 23:40
Operator : YP\AJ
Sample : Q3450-04
Misc :
ALS Vial : 29 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
SS-10R-(3.5-4.0)

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 11/03/2025
Supervised By :Yogesh Patel 11/03/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 01 02:47:44 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 24 15:40:32 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.644	3.782	24507961	100.9E6	17.149	15.241
2)	SA Decachlor...	10.421	8.780	14440943	89924937	13.414	10.936
	Target Compounds						

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Sample Results: **PP076279.D**

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP103125\
Data File : PP076279.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 31 Oct 2025 23:40
Operator : YP\AJ
Sample : Q3450-04
Misc :
ALS Vial : 29 Sample Multiplier: 1


Instrument :
ECD_P
ClientSampleId :
SS-10R-(3.5-4.0)


Manual Integrations APPROVED


Reviewed By :Rahul Chavli 11/03/2025
Supervised By :Yogesh Patel 11/03/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 01 02:47:44 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 24 15:40:32 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm








#1 Tetrachloro-m-xylene

R.T.: 4.644 min
Delta R.T.: 0.002 min
Response: 24507961
Conc: 17.15 ng/ml

Instrument :
ECD_P
ClientSampleId :
SS-10R-(3.5-4.0)


Manual Integrations APPROVED

Reviewed By :Rahul Chavli 11/03/2025
Supervised By :Yogesh Patel 11/03/2025




#1 Tetrachloro-m-xylene

R.T.: 3.782 min
Delta R.T.: 0.000 min
Response: 100866397
Conc: 15.24 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.421 min
Delta R.T.: 0.003 min
Response: 14440943
Conc: 13.41 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.780 min
Delta R.T.: -0.002 min
Response: 89924937
Conc: 10.94 ng/ml

Raw Data: **PP076282.D**

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Matrix New World Engineering	Date Collected:	10/23/25
Project:	PSEG Athenia Substation	Date Received:	10/23/25
Client Sample ID:	SS-13R-(3.5-4.0)	SDG No.:	Q3450
Lab Sample ID:	Q3450-05	Matrix:	SOIL
Analytical Method:	8082A	% Solid:	92
Sample Wt/Vol:	30.07 g	Final Vol: 10000 uL	Test: PCB
Prep Method:	SW3541B	Prep Date: 10/31/25	

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
12674-11-2	Aroclor-1016	4.30	U	1	4.30	18.4	ug/kg	11/01/25 00:28	PB170358
11104-28-2	Aroclor-1221	4.40	U	1	4.40	18.4	ug/kg	11/01/25 00:28	PB170358
11141-16-5	Aroclor-1232	4.00	U	1	4.00	18.4	ug/kg	11/01/25 00:28	PB170358
53469-21-9	Aroclor-1242	4.30	U	1	4.30	18.4	ug/kg	11/01/25 00:28	PB170358
12672-29-6	Aroclor-1248	6.40	U	1	6.40	18.4	ug/kg	11/01/25 00:28	PB170358
11097-69-1	Aroclor-1254	3.50	U	1	3.50	18.4	ug/kg	11/01/25 00:28	PB170358
37324-23-5	Aroclor-1262	5.40	U	1	5.40	18.4	ug/kg	11/01/25 00:28	PB170358
11100-14-4	Aroclor-1268	3.90	U	1	3.90	18.4	ug/kg	11/01/25 00:28	PB170358
11096-82-5	Aroclor-1260	3.50	U	1	3.50	18.4	ug/kg	11/01/25 00:28	PB170358
SURROGATES									
877-09-8	Tetrachloro-m-xylene	22.7			30 (21) - 150 (165)	114%	SPK: 20		
2051-24-3	Decachlorobiphenyl	21.2			30 (10) - 150 (170)	106%	SPK: 20		

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
E = Value Exceeds Calibration Range
P = Indicates >25% difference for detected concentrations between the two GC columns
Q = indicates LCS control criteria did not meet requirements
M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
B = Analyte Found in Associated Method Blank
N = Presumptive Evidence of a Compound
* = Values outside of QC limits
D = Dilution
S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.
() = Laboratory InHouse Limit

Sample Results: **PP076282.D**

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP103125\
Data File : PP076282.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 01 Nov 2025 00:28
Operator : YP\AJ
Sample : Q3450-05
Misc :
ALS Vial : 32 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
SS-13R-(3.5-4.0)

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 01 02:48:46 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 24 15:40:32 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.645	3.781	32485556	141.6E6	22.732	21.397
2)	SA Decachlor...	10.423	8.778	22797108	145.0E6	21.177	17.628
	Target Compounds						

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Sample Results: **PP076282.D**


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP103125\
Data File : PP076282.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 01 Nov 2025 00:28
Operator : YP\AJ
Sample : Q3450-05
Misc :
ALS Vial : 32 Sample Multiplier: 1


Instrument :
ECD_P
ClientSampleId :
SS-13R-(3.5-4.0)

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 01 02:48:46 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 24 15:40:32 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm








#1 Tetrachloro-m-xylene


R.T.: 4.645 min
Delta R.T.: 0.003 min
Response: 32485556
Conc: 22.73 ng/ml

Instrument :
ECD_P
ClientSampleId :
SS-13R-(3.5-4.0)




#1 Tetrachloro-m-xylene

R.T.: 3.781 min
Delta R.T.: 0.000 min
Response: 141611198
Conc: 21.40 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.423 min
Delta R.T.: 0.004 min
Response: 22797108
Conc: 21.18 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.778 min
Delta R.T.: -0.004 min
Response: 144955939
Conc: 17.63 ng/ml

Raw Data: **PP076283.D**

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Matrix New World Engineering	Date Collected:	10/23/25
Project:	PSEG Athenia Substation	Date Received:	10/23/25
Client Sample ID:	SS-13R-(5.0-5.5)	SDG No.:	Q3450
Lab Sample ID:	Q3450-06	Matrix:	SOIL
Analytical Method:	8082A	% Solid:	84.1
Sample Wt/Vol:	30.03 g	Final Vol: 10000 uL	Test: PCB
Prep Method:	SW3541B	Prep Date: 10/31/25	

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
12674-11-2	Aroclor-1016	4.70	U	1	4.70	20.2	ug/kg	11/01/25 00:45	PB170358
11104-28-2	Aroclor-1221	4.80	U	1	4.80	20.2	ug/kg	11/01/25 00:45	PB170358
11141-16-5	Aroclor-1232	4.40	U	1	4.40	20.2	ug/kg	11/01/25 00:45	PB170358
53469-21-9	Aroclor-1242	4.80	U	1	4.80	20.2	ug/kg	11/01/25 00:45	PB170358
12672-29-6	Aroclor-1248	7.00	U	1	7.00	20.2	ug/kg	11/01/25 00:45	PB170358
11097-69-1	Aroclor-1254	3.80	U	1	3.80	20.2	ug/kg	11/01/25 00:45	PB170358
37324-23-5	Aroclor-1262	6.00	U	1	6.00	20.2	ug/kg	11/01/25 00:45	PB170358
11100-14-4	Aroclor-1268	4.30	U	1	4.30	20.2	ug/kg	11/01/25 00:45	PB170358
11096-82-5	Aroclor-1260	3.80	U	1	3.80	20.2	ug/kg	11/01/25 00:45	PB170358
SURROGATES									
877-09-8	Tetrachloro-m-xylene	22.4			30 (21) - 150 (165)	112%	SPK: 20		
2051-24-3	Decachlorobiphenyl	20.1			30 (10) - 150 (170)	101%	SPK: 20		

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
E = Value Exceeds Calibration Range
P = Indicates >25% difference for detected concentrations between the two GC columns
Q = indicates LCS control criteria did not meet requirements
M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
B = Analyte Found in Associated Method Blank
N = Presumptive Evidence of a Compound
* = Values outside of QC limits
D = Dilution
S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.
() = Laboratory InHouse Limit

Quantitation Report (QT Reviewed)

Sample Results: **PP076283.D**

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP103125\
Data File : PP076283.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 01 Nov 2025 00:45
Operator : YP\AJ
Sample : Q3450-06
Misc :
ALS Vial : 33 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
SS-13R-(5.0-5.5)

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 11/03/2025
Supervised By :Yogesh Patel 11/03/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 01 02:49:05 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 24 15:40:32 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.644	3.781	31993783	142.8E6	22.387m	21.578
2)	SA Decachlor...	10.422	8.779	21671724	138.3E6	20.131	16.815
	Target Compounds						

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Sample Results: **PP076283.D**

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP103125\
Data File : PP076283.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 01 Nov 2025 00:45
Operator : YP\AJ
Sample : Q3450-06
Misc :
ALS Vial : 33 Sample Multiplier: 1


Instrument :
ECD_P
ClientSampleId :
SS-13R-(5.0-5.5)


Manual Integrations APPROVED


Reviewed By :Rahul Chavli 11/03/2025
Supervised By :Yogesh Patel 11/03/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 01 02:49:05 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 24 15:40:32 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm








#1 Tetrachloro-m-xylene

R.T.: 4.644 min
Delta R.T.: 0.002 min
Response: 31993783
Conc: 22.39 ng/ml

Instrument :
ECD_P
ClientSampleId :
SS-13R-(5.0-5.5)


Manual Integrations APPROVED

Reviewed By :Rahul Chavli 11/03/2025
Supervised By :Yogesh Patel 11/03/2025




#1 Tetrachloro-m-xylene

R.T.: 3.781 min
Delta R.T.: 0.000 min
Response: 142808703
Conc: 21.58 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.422 min
Delta R.T.: 0.003 min
Response: 21671724
Conc: 20.13 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.779 min
Delta R.T.: -0.003 min
Response: 138267932
Conc: 16.81 ng/ml

Raw Data: **PO115045.D**

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Matrix New World Engineering	Date Collected:	10/23/25
Project:	PSEG Athenia Substation	Date Received:	10/23/25
Client Sample ID:	SS-13R-(4.5-5.0)	SDG No.:	Q3450
Lab Sample ID:	Q3450-07	Matrix:	SOIL
Analytical Method:	8082A	% Solid:	85.3
Sample Wt/Vol:	30.08 g	Final Vol: 10000 uL	Test: PCB
Prep Method:	SW3541B	Prep Date: 11/12/25	

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
12674-11-2	Aroclor-1016	4.60	U	1	4.60	19.9	ug/kg	11/12/25 15:36	PB170497
11104-28-2	Aroclor-1221	4.70	U	1	4.70	19.9	ug/kg	11/12/25 15:36	PB170497
11141-16-5	Aroclor-1232	4.30	U	1	4.30	19.9	ug/kg	11/12/25 15:36	PB170497
53469-21-9	Aroclor-1242	4.70	U	1	4.70	19.9	ug/kg	11/12/25 15:36	PB170497
12672-29-6	Aroclor-1248	6.90	U	1	6.90	19.9	ug/kg	11/12/25 15:36	PB170497
11097-69-1	Aroclor-1254	13.8	J	1	3.80	19.9	ug/kg	11/12/25 15:36	PB170497
37324-23-5	Aroclor-1262	5.90	U	1	5.90	19.9	ug/kg	11/12/25 15:36	PB170497
11100-14-4	Aroclor-1268	4.20	U	1	4.20	19.9	ug/kg	11/12/25 15:36	PB170497
11096-82-5	Aroclor-1260	3.80	U	1	3.80	19.9	ug/kg	11/12/25 15:36	PB170497
SURROGATES									
877-09-8	Tetrachloro-m-xylene	23.4			30 (21) - 150 (165)	117%	SPK: 20		
2051-24-3	Decachlorobiphenyl	22.9			30 (10) - 150 (170)	115%	SPK: 20		

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
E = Value Exceeds Calibration Range
P = Indicates >25% difference for detected concentrations between the two GC columns
Q = indicates LCS control criteria did not meet requirements
M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
B = Analyte Found in Associated Method Blank
N = Presumptive Evidence of a Compound
* = Values outside of QC limits
D = Dilution
S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.
() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO111225\
Data File : PO115045.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Nov 2025 15:36
Operator : YP/AJ
Sample : Q3450-07
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
SS-13R-(4.5-5.0)

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 13 10:14:52 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Nov 05 06:00:07 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.347	3.648	162.6E6	104.6E6	23.108	23.426
2)	SA Decachlor...	9.933	8.642	93256374	89547537	21.098	22.938
	Target Compounds						
26)	L6 AR-1254-1	6.326	5.528	3648455	3249674	14.396	10.826
27)	L6 AR-1254-2	6.547	5.672	5839323	3809651	15.763	14.130
28)	L6 AR-1254-3	6.905	6.075	13185812	11532465	34.261	27.312
29)	L6 AR-1254-4	7.189	6.303	10730347	8865135	32.643	29.809
30)	L6 AR-1254-5	7.604	6.719	24036098	26864413	79.628	67.579


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO111225\
Data File : PO115045.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Nov 2025 15:36
Operator : YP/AJ
Sample : Q3450-07
Misc :
ALS Vial : 16 Sample Multiplier: 1


Instrument :
ECD_O
ClientSampleId :
SS-13R-(4.5-5.0)

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 13 10:14:52 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Nov 05 06:00:07 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm








#1 Tetrachloro-m-xylene


R.T.: 4.347 min
Delta R.T.: 0.000 min
Response: 162637860
Conc: 23.11 ng/ml

Instrument :
ECD_O
ClientSampleId :
SS-13R-(4.5-5.0)




#1 Tetrachloro-m-xylene

R.T.: 3.648 min
Delta R.T.: 0.000 min
Response: 104639705
Conc: 23.43 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.933 min
Delta R.T.: 0.003 min
Response: 93256374
Conc: 21.10 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.642 min
Delta R.T.: 0.001 min
Response: 89547537
Conc: 22.94 ng/ml



#26 AR-1254-1


R.T.: 6.326 min
Delta R.T.: -0.005 min
Response: 3648455
Conc: 14.40 ng/ml

Instrument :
ECD_O
ClientSampleId :
SS-13R-(4.5-5.0)




#26 AR-1254-1

R.T.: 5.528 min
Delta R.T.: 0.002 min
Response: 3249674
Conc: 10.83 ng/ml




#27 AR-1254-2

R.T.: 6.547 min
Delta R.T.: 0.000 min
Response: 5839323
Conc: 15.76 ng/ml



#27 AR-1254-2


R.T.: 5.672 min
Delta R.T.: -0.002 min
Response: 3809651
Conc: 14.13 ng/ml



#28 AR-1254-3


R.T.: 6.905 min
Delta R.T.: -0.002 min
Response: 13185812
Conc: 34.26 ng/ml

Instrument :
ECD_O
ClientSampleId :
SS-13R-(4.5-5.0)




#28 AR-1254-3

R.T.: 6.075 min
Delta R.T.: -0.001 min
Response: 11532465
Conc: 27.31 ng/ml




#29 AR-1254-4

R.T.: 7.189 min
Delta R.T.: 0.000 min
Response: 10730347
Conc: 32.64 ng/ml



#29 AR-1254-4


R.T.: 6.303 min
Delta R.T.: 0.000 min
Response: 8865135
Conc: 29.81 ng/ml



#30 AR-1254-5

R.T.: 7.604 min
Delta R.T.: 0.001 min
Response: 24036098
Conc: 79.63 ng/ml

Instrument :
ECD_O
ClientSampleId :
SS-13R-(4.5-5.0)



#30 AR-1254-5

R.T.: 6.719 min
Delta R.T.: 0.000 min
Response: 26864413
Conc: 67.58 ng/ml

Raw Data: **PQ071155.D**

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Matrix New World Engineering	Date Collected:	10/23/25
Project:	PSEG Athenia Substation	Date Received:	10/23/25
Client Sample ID:	SS-20-(0.0-0.5)	SDG No.:	Q3450
Lab Sample ID:	Q3450-08	Matrix:	SOIL
Analytical Method:	8082A	% Solid:	93.5
Sample Wt/Vol:	30.01 g	Final Vol: 10000 uL	Test: PCB
Prep Method:	SW3541B	Prep Date: 10/24/25	

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
12674-11-2	Aroclor-1016	4.20	U	1	4.20	18.2	ug/kg	10/27/25 10:24	PB170236
11104-28-2	Aroclor-1221	4.30	U	1	4.30	18.2	ug/kg	10/27/25 10:24	PB170236
11141-16-5	Aroclor-1232	4.00	U	1	4.00	18.2	ug/kg	10/27/25 10:24	PB170236
53469-21-9	Aroclor-1242	4.30	U	1	4.30	18.2	ug/kg	10/27/25 10:24	PB170236
12672-29-6	Aroclor-1248	6.30	U	1	6.30	18.2	ug/kg	10/27/25 10:24	PB170236
11097-69-1	Aroclor-1254	174		1	3.40	18.2	ug/kg	10/27/25 10:24	PB170236
37324-23-5	Aroclor-1262	5.40	U	1	5.40	18.2	ug/kg	10/27/25 10:24	PB170236
11100-14-4	Aroclor-1268	3.80	U	1	3.80	18.2	ug/kg	10/27/25 10:24	PB170236
11096-82-5	Aroclor-1260	3.50	U	1	3.50	18.2	ug/kg	10/27/25 10:24	PB170236
SURROGATES									
877-09-8	Tetrachloro-m-xylene	26.8			30 (21) - 150 (165)	134%	SPK: 20		
2051-24-3	Decachlorobiphenyl	28.1			30 (10) - 150 (170)	140%	SPK: 20		

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
E = Value Exceeds Calibration Range
P = Indicates >25% difference for detected concentrations between the two GC columns
Q = indicates LCS control criteria did not meet requirements
M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
B = Analyte Found in Associated Method Blank
N = Presumptive Evidence of a Compound
* = Values outside of QC limits
D = Dilution
S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.
() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ102725\
Data File : PQ071155.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Oct 2025 10:24
Operator : YP\AJ
Sample : Q3450-08
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
SS-20-(0.0-0.5)

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 27 11:31:16 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Oct 11 01:55:41 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	3.399	2.776	116.8E6	127.4E6	26.842	25.074
2)	SA Decachlor...	8.495	7.530	81481303	134.3E6	28.058	26.593
	Target Compounds						
26)	L6 AR-1254-1	5.292	4.531	53593012	87844114	273.046	260.501
27)	L6 AR-1254-2	5.506	4.677	103.5E6	115.1E6	356.309	387.376
28)	L6 AR-1254-3	5.853	5.058	172.9E6	239.0E6	562.937	511.748
29)	L6 AR-1254-4	6.134	5.284	103.5E6	136.4E6	471.483	462.124
30)	L6 AR-1254-5	6.542	5.687	180.4E6	297.8E6	778.042	708.513


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ102725\
Data File : PQ071155.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Oct 2025 10:24
Operator : YP\AJ
Sample : Q3450-08
Misc :
ALS Vial : 15 Sample Multiplier: 1


Instrument :
ECD_Q
ClientSampleId :
SS-20-(0.0-0.5)

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 27 11:31:16 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Oct 11 01:55:41 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm








#1 Tetrachloro-m-xylene


R.T.: 3.399 min
Delta R.T.: -0.001 min
Response: 116769388
Conc: 26.84 ng/ml

Instrument :
ECD_Q
ClientSampleId :
SS-20-(0.0-0.5)




#1 Tetrachloro-m-xylene

R.T.: 2.776 min
Delta R.T.: -0.002 min
Response: 127376293
Conc: 25.07 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.495 min
Delta R.T.: -0.005 min
Response: 81481303
Conc: 28.06 ng/ml



#2 Decachlorobiphenyl


R.T.: 7.530 min
Delta R.T.: -0.003 min
Response: 134285156
Conc: 26.59 ng/ml



#26 AR-1254-1


R.T.: 5.292 min
Delta R.T.: -0.001 min
Response: 53593012
Conc: 273.05 ng/ml

Instrument :
ECD_Q
ClientSampleId :
SS-20-(0.0-0.5)




#26 AR-1254-1

R.T.: 4.531 min
Delta R.T.: -0.002 min
Response: 87844114
Conc: 260.50 ng/ml




#27 AR-1254-2

R.T.: 5.506 min
Delta R.T.: -0.001 min
Response: 103501873
Conc: 356.31 ng/ml



#27 AR-1254-2


R.T.: 4.677 min
Delta R.T.: -0.002 min
Response: 115110478
Conc: 387.38 ng/ml



#28 AR-1254-3


R.T.: 5.853 min
Delta R.T.: -0.002 min
Response: 172938342
Conc: 562.94 ng/ml

Instrument :
ECD_Q
ClientSampleId :
SS-20-(0.0-0.5)




#28 AR-1254-3

R.T.: 5.058 min
Delta R.T.: -0.002 min
Response: 239019656
Conc: 511.75 ng/ml




#29 AR-1254-4

R.T.: 6.134 min
Delta R.T.: -0.002 min
Response: 103509591
Conc: 471.48 ng/ml



#29 AR-1254-4


R.T.: 5.284 min
Delta R.T.: -0.002 min
Response: 136442788
Conc: 462.12 ng/ml



#30 AR-1254-5

R.T.: 6.542 min
Delta R.T.: -0.001 min
Response: 180424492
Conc: 778.04 ng/ml

Instrument :
ECD_Q
ClientSampleId :
SS-20-(0.0-0.5)



#30 AR-1254-5

R.T.: 5.687 min
Delta R.T.: -0.002 min
Response: 297823023
Conc: 708.51 ng/ml

Raw Data: **PQ071156.D**

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Matrix New World Engineering	Date Collected:	10/23/25
Project:	PSEG Athenia Substation	Date Received:	10/23/25
Client Sample ID:	SS-20-(2.0-2.5)	SDG No.:	Q3450
Lab Sample ID:	Q3450-09	Matrix:	SOIL
Analytical Method:	8082A	% Solid:	86.3
Sample Wt/Vol:	30.09 g	Final Vol: 10000 uL	Test: PCB
Prep Method:	SW3541B	Prep Date: 10/24/25	

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
12674-11-2	Aroclor-1016	4.60	U	1	4.60	19.6	ug/kg	10/27/25 10:39	PB170236
11104-28-2	Aroclor-1221	4.70	U	1	4.70	19.6	ug/kg	10/27/25 10:39	PB170236
11141-16-5	Aroclor-1232	4.30	U	1	4.30	19.6	ug/kg	10/27/25 10:39	PB170236
53469-21-9	Aroclor-1242	4.60	U	1	4.60	19.6	ug/kg	10/27/25 10:39	PB170236
12672-29-6	Aroclor-1248	6.80	U	1	6.80	19.6	ug/kg	10/27/25 10:39	PB170236
11097-69-1	Aroclor-1254	11.3	J	1	3.70	19.6	ug/kg	10/27/25 10:39	PB170236
37324-23-5	Aroclor-1262	5.80	U	1	5.80	19.6	ug/kg	10/27/25 10:39	PB170236
11100-14-4	Aroclor-1268	4.20	U	1	4.20	19.6	ug/kg	10/27/25 10:39	PB170236
11096-82-5	Aroclor-1260	3.70	U	1	3.70	19.6	ug/kg	10/27/25 10:39	PB170236
SURROGATES									
877-09-8	Tetrachloro-m-xylene	27.2			30 (21) - 150 (165)	136%	SPK: 20		
2051-24-3	Decachlorobiphenyl	27.4			30 (10) - 150 (170)	137%	SPK: 20		

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
E = Value Exceeds Calibration Range
P = Indicates >25% difference for detected concentrations between the two GC columns
Q = indicates LCS control criteria did not meet requirements
M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
B = Analyte Found in Associated Method Blank
N = Presumptive Evidence of a Compound
* = Values outside of QC limits
D = Dilution
S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.
() = Laboratory InHouse Limit

Sample Results: **PQ071156.D**

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ102725\
Data File : PQ071156.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Oct 2025 10:39
Operator : YP\AJ
Sample : Q3450-09
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
SS-20-(2.0-2.5)

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/27/2025
Supervised By :mohammad ahmed 10/29/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 27 11:32:27 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Oct 11 01:55:41 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	3.399	2.776	118.4E6	126.6E6	27.206	24.913
2)	SA Decachlor...	8.496	7.531	79502189	132.8E6	27.377	26.304
	Target Compounds						
26)	L6 AR-1254-1	5.292	4.530	2481559	3492866	12.643	10.358
27)	L6 AR-1254-2	5.506	4.677	4523855	6232033	15.574	20.972 #
28)	L6 AR-1254-3	5.854	5.057	10115880	12983056	32.928	27.797m
29)	L6 AR-1254-4	6.134	5.283	5851786	8744197	26.655	29.616m
30)	L6 AR-1254-5	6.542	5.687	13252586	24169145	57.149	57.498

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ102725\
Data File : PQ071156.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Oct 2025 10:39
Operator : YP\AJ
Sample : Q3450-09
Misc :
ALS Vial : 16 Sample Multiplier: 1


Instrument :
ECD_Q
ClientSampleId :
SS-20-(2.0-2.5)


Manual Integrations APPROVED


Reviewed By :Yogesh Patel 10/27/2025
Supervised By :mohammad ahmed 10/29/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 27 11:32:27 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Oct 11 01:55:41 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm








#1 Tetrachloro-m-xylene

R.T.: 3.399 min
Delta R.T.: -0.001 min
Response: 118351534
Conc: 27.21 ng/ml

Instrument :
ECD_Q
ClientSampleId :
SS-20-(2.0-2.5)


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/27/2025
Supervised By :mohammad ahmed 10/29/2025




#1 Tetrachloro-m-xylene

R.T.: 2.776 min
Delta R.T.: -0.002 min
Response: 126556708
Conc: 24.91 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.496 min
Delta R.T.: -0.004 min
Response: 79502189
Conc: 27.38 ng/ml



#2 Decachlorobiphenyl

R.T.: 7.531 min
Delta R.T.: -0.002 min
Response: 132825485
Conc: 26.30 ng/ml




#26 AR-1254-1

R.T.: 5.292 min
Delta R.T.: -0.001 min
Response: 2481559
Conc: 12.64 ng/ml

Instrument :
ECD_Q
ClientSampleId :
SS-20-(2.0-2.5)


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/27/2025
Supervised By :mohammad ahmed 10/29/2025




#26 AR-1254-1

R.T.: 4.530 min
Delta R.T.: -0.002 min
Response: 3492866
Conc: 10.36 ng/ml




#27 AR-1254-2

R.T.: 5.506 min
Delta R.T.: -0.001 min
Response: 4523855
Conc: 15.57 ng/ml



#27 AR-1254-2

R.T.: 4.677 min
Delta R.T.: -0.002 min
Response: 6232033
Conc: 20.97 ng/ml




#28 AR-1254-3

R.T.: 5.854 min
Delta R.T.: -0.001 min
Response: 10115880
Conc: 32.93 ng/ml

Instrument :
ECD_Q
ClientSampleId :
SS-20-(2.0-2.5)


Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/27/2025
Supervised By :mohammad ahmed 10/29/2025




#28 AR-1254-3

R.T.: 5.057 min
Delta R.T.: -0.003 min
Response: 12983056
Conc: 27.80 ng/ml m




#29 AR-1254-4

R.T.: 6.134 min
Delta R.T.: -0.001 min
Response: 5851786
Conc: 26.65 ng/ml



#29 AR-1254-4

R.T.: 5.283 min
Delta R.T.: -0.003 min
Response: 8744197
Conc: 29.62 ng/ml m




#30 AR-1254-5

R.T.: 6.542 min
Delta R.T.: -0.001 min
Response: 13252586
Conc: 57.15 ng/ml

Instrument :
ECD_Q
ClientSampleId :
SS-20-(2.0-2.5)

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/27/2025
Supervised By :mohammad ahmed 10/29/2025



#30 AR-1254-5

R.T.: 5.687 min
Delta R.T.: -0.002 min
Response: 24169145
Conc: 57.50 ng/ml

Raw Data: **PP076047.D**

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Matrix New World Engineering	Date Collected:	10/23/25
Project:	PSEG Athenia Substation	Date Received:	10/23/25
Client Sample ID:	SS-20-(3.5-4.0)	SDG No.:	Q3450
Lab Sample ID:	Q3450-10	Matrix:	SOIL
Analytical Method:	8082A	% Solid:	84
Sample Wt/Vol:	30.07 g Final Vol: 10000 uL	Test:	PCB
Prep Method:	SW3541B Prep Date: 10/24/25		

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
12674-11-2	Aroclor-1016	4.70	U	1	4.70	20.2	ug/kg	10/24/25 20:52	PB170236
11104-28-2	Aroclor-1221	4.80	U	1	4.80	20.2	ug/kg	10/24/25 20:52	PB170236
11141-16-5	Aroclor-1232	4.40	U	1	4.40	20.2	ug/kg	10/24/25 20:52	PB170236
53469-21-9	Aroclor-1242	4.80	U	1	4.80	20.2	ug/kg	10/24/25 20:52	PB170236
12672-29-6	Aroclor-1248	7.00	U	1	7.00	20.2	ug/kg	10/24/25 20:52	PB170236
11097-69-1	Aroclor-1254	55.2		1	3.80	20.2	ug/kg	10/24/25 20:52	PB170236
37324-23-5	Aroclor-1262	6.00	U	1	6.00	20.2	ug/kg	10/24/25 20:52	PB170236
11100-14-4	Aroclor-1268	4.30	U	1	4.30	20.2	ug/kg	10/24/25 20:52	PB170236
11096-82-5	Aroclor-1260	3.80	U	1	3.80	20.2	ug/kg	10/24/25 20:52	PB170236
SURROGATES									
877-09-8	Tetrachloro-m-xylene	24.1			30 (21) - 150 (165)	121%	SPK: 20		
2051-24-3	Decachlorobiphenyl	25.8			30 (10) - 150 (170)	129%	SPK: 20		

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
E = Value Exceeds Calibration Range
P = Indicates >25% difference for detected concentrations between the two GC columns
Q = indicates LCS control criteria did not meet requirements
M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
B = Analyte Found in Associated Method Blank
N = Presumptive Evidence of a Compound
* = Values outside of QC limits
D = Dilution
S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.
() = Laboratory InHouse Limit

Sample Results: **PP076047.D**

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102425\
Data File : PP076047.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 24 Oct 2025 20:52
Operator : YP\AJ
Sample : Q3450-10
Misc :
ALS Vial : 31 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
SS-20-(3.5-4.0)

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/27/2025
Supervised By :mohammad ahmed 10/28/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 27 05:09:22 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 24 15:40:32 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.640	3.779	34497841	152.2E6	24.140	22.996
2)	SA Decachlor...	10.411	8.777	27763719	184.0E6	25.790	22.378
	Target Compounds						
26)	L6 AR-1254-1	6.639	5.653	2313097	21935559	28.952m	38.893 #
27)	L6 AR-1254-2	6.856	5.801	6552449	39782760	65.535	75.050
28)	L6 AR-1254-3	7.218	6.204	12960078	99046244	124.670	116.376
29)	L6 AR-1254-4	7.500	6.430	15873106	103.6E6	189.838	164.524
30)	L6 AR-1254-5	7.917	6.847	26406087	190.2E6	288.368	264.263

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102425\
Data File : PP076047.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 24 Oct 2025 20:52
Operator : YP\AJ
Sample : Q3450-10
Misc :
ALS Vial : 31 Sample Multiplier: 1


Instrument :
ECD_P
ClientSampleId :
SS-20-(3.5-4.0)


Manual Integrations APPROVED


Reviewed By :Yogesh Patel 10/27/2025
Supervised By :mohammad ahmed 10/28/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 27 05:09:22 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 24 15:40:32 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm








#1 Tetrachloro-m-xylene

R.T.: 4.640 min
Delta R.T.: -0.002 min
Response: 34497841
Conc: 24.14 ng/ml

Instrument :
ECD_P
ClientSampleId :
SS-20-(3.5-4.0)


Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/27/2025
Supervised By :mohammad ahmed 10/28/2025




#1 Tetrachloro-m-xylene

R.T.: 3.779 min
Delta R.T.: -0.002 min
Response: 152196349
Conc: 23.00 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.411 min
Delta R.T.: -0.008 min
Response: 27763719
Conc: 25.79 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.777 min
Delta R.T.: -0.005 min
Response: 184016645
Conc: 22.38 ng/ml




#26 AR-1254-1

R.T.: 6.639 min
Delta R.T.: -0.008 min
Response: 2313097
Conc: 28.95 ng/ml

Instrument :
ECD_P
ClientSampleId :
SS-20-(3.5-4.0)


Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/27/2025
Supervised By :mohammad ahmed 10/28/2025




#26 AR-1254-1

R.T.: 5.653 min
Delta R.T.: -0.009 min
Response: 21935559
Conc: 38.89 ng/ml




#27 AR-1254-2

R.T.: 6.856 min
Delta R.T.: -0.008 min
Response: 6552449
Conc: 65.53 ng/ml



#27 AR-1254-2

R.T.: 5.801 min
Delta R.T.: -0.007 min
Response: 39782760
Conc: 75.05 ng/ml




#28 AR-1254-3

R.T.: 7.218 min
Delta R.T.: -0.008 min
Response: 12960078
Conc: 124.67 ng/ml

Instrument :
ECD_P
ClientSampleId :
SS-20-(3.5-4.0)


Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/27/2025
Supervised By :mohammad ahmed 10/28/2025




#28 AR-1254-3

R.T.: 6.204 min
Delta R.T.: -0.006 min
Response: 99046244
Conc: 116.38 ng/ml




#29 AR-1254-4

R.T.: 7.500 min
Delta R.T.: -0.008 min
Response: 15873106
Conc: 189.84 ng/ml



#29 AR-1254-4

R.T.: 6.430 min
Delta R.T.: -0.008 min
Response: 103614141
Conc: 164.52 ng/ml




#30 AR-1254-5

R.T.: 7.917 min
Delta R.T.: -0.008 min
Response: 26406087
Conc: 288.37 ng/ml

Instrument :
ECD_P
ClientSampleId :
SS-20-(3.5-4.0)

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/27/2025
Supervised By :mohammad ahmed 10/28/2025



#30 AR-1254-5

R.T.: 6.847 min
Delta R.T.: -0.007 min
Response: 190238505
Conc: 264.26 ng/ml

Raw Data: **PP076048.D**

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Matrix New World Engineering	Date Collected:	10/23/25
Project:	PSEG Athenia Substation	Date Received:	10/23/25
Client Sample ID:	SS-28-(2.0-2.5)	SDG No.:	Q3450
Lab Sample ID:	Q3450-11	Matrix:	SOIL
Analytical Method:	8082A	% Solid:	93
Sample Wt/Vol:	30.02 g	Final Vol: 10000 uL	Test: PCB
Prep Method:	SW3541B	Prep Date: 10/24/25	

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
12674-11-2	Aroclor-1016	4.20	U	1	4.20	18.3	ug/kg	10/24/25 21:08	PB170236
11104-28-2	Aroclor-1221	4.30	U	1	4.30	18.3	ug/kg	10/24/25 21:08	PB170236
11141-16-5	Aroclor-1232	4.00	U	1	4.00	18.3	ug/kg	10/24/25 21:08	PB170236
53469-21-9	Aroclor-1242	4.30	U	1	4.30	18.3	ug/kg	10/24/25 21:08	PB170236
12672-29-6	Aroclor-1248	6.40	U	1	6.40	18.3	ug/kg	10/24/25 21:08	PB170236
11097-69-1	Aroclor-1254	3.40	U	1	3.40	18.3	ug/kg	10/24/25 21:08	PB170236
37324-23-5	Aroclor-1262	5.40	U	1	5.40	18.3	ug/kg	10/24/25 21:08	PB170236
11100-14-4	Aroclor-1268	3.90	U	1	3.90	18.3	ug/kg	10/24/25 21:08	PB170236
11096-82-5	Aroclor-1260	3.50	U	1	3.50	18.3	ug/kg	10/24/25 21:08	PB170236
SURROGATES									
877-09-8	Tetrachloro-m-xylene	23.3			30 (21) - 150 (165)	116%	SPK: 20		
2051-24-3	Decachlorobiphenyl	25.0			30 (10) - 150 (170)	125%	SPK: 20		

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
E = Value Exceeds Calibration Range
P = Indicates >25% difference for detected concentrations between the two GC columns
Q = indicates LCS control criteria did not meet requirements
M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
B = Analyte Found in Associated Method Blank
N = Presumptive Evidence of a Compound
* = Values outside of QC limits
D = Dilution
S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.
() = Laboratory InHouse Limit

Sample Results: **PP076048.D**

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102425\
Data File : PP076048.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 24 Oct 2025 21:08
Operator : YP\AJ
Sample : Q3450-11
Misc :
ALS Vial : 32 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
SS-28-(2.0-2.5)

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 24 23:40:37 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 24 15:40:32 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.643	3.782	33230144	150.1E6	23.253	22.678
2)	SA Decachlor...	10.414	8.779	26956497	181.2E6	25.040	22.030
	Target Compounds						


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102425\
Data File : PP076048.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 24 Oct 2025 21:08
Operator : YP\AJ
Sample : Q3450-11
Misc :
ALS Vial : 32 Sample Multiplier: 1


Instrument :
ECD_P
ClientSampleId :
SS-28-(2.0-2.5)

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 24 23:40:37 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 24 15:40:32 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm








#1 Tetrachloro-m-xylene


R.T.: 4.643 min
Delta R.T.: 0.001 min
Response: 33230144
Conc: 23.25 ng/ml

Instrument :
ECD_P
ClientSampleId :
SS-28-(2.0-2.5)




#1 Tetrachloro-m-xylene

R.T.: 3.782 min
Delta R.T.: 0.000 min
Response: 150088081
Conc: 22.68 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.414 min
Delta R.T.: -0.005 min
Response: 26956497
Conc: 25.04 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.779 min
Delta R.T.: -0.003 min
Response: 181156077
Conc: 22.03 ng/ml

Raw Data: **PP076049.D**

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Matrix New World Engineering	Date Collected:	10/23/25
Project:	PSEG Athenia Substation	Date Received:	10/23/25
Client Sample ID:	SS-21-(0.0-0.5)	SDG No.:	Q3450
Lab Sample ID:	Q3450-20	Matrix:	SOIL
Analytical Method:	8082A	% Solid:	91.7
Sample Wt/Vol:	30.05 g Final Vol: 10000 uL	Test:	PCB
Prep Method:	SW3541B Prep Date: 10/24/25		

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
12674-11-2	Aroclor-1016	4.30	U	1	4.30	18.5	ug/kg	10/24/25 21:24	PB170236
11104-28-2	Aroclor-1221	4.40	U	1	4.40	18.5	ug/kg	10/24/25 21:24	PB170236
11141-16-5	Aroclor-1232	4.00	U	1	4.00	18.5	ug/kg	10/24/25 21:24	PB170236
53469-21-9	Aroclor-1242	4.40	U	1	4.40	18.5	ug/kg	10/24/25 21:24	PB170236
12672-29-6	Aroclor-1248	6.40	U	1	6.40	18.5	ug/kg	10/24/25 21:24	PB170236
11097-69-1	Aroclor-1254	593	E	1	3.50	18.5	ug/kg	10/24/25 21:24	PB170236
37324-23-5	Aroclor-1262	5.50	U	1	5.50	18.5	ug/kg	10/24/25 21:24	PB170236
11100-14-4	Aroclor-1268	3.90	U	1	3.90	18.5	ug/kg	10/24/25 21:24	PB170236
11096-82-5	Aroclor-1260	3.50	U	1	3.50	18.5	ug/kg	10/24/25 21:24	PB170236
SURROGATES									
877-09-8	Tetrachloro-m-xylene	24.0			30 (21) - 150 (165)	120%	SPK: 20		
2051-24-3	Decachlorobiphenyl	25.0			30 (10) - 150 (170)	125%	SPK: 20		

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
E = Value Exceeds Calibration Range
P = Indicates >25% difference for detected concentrations between the two GC columns
Q = indicates LCS control criteria did not meet requirements
M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
B = Analyte Found in Associated Method Blank
N = Presumptive Evidence of a Compound
* = Values outside of QC limits
D = Dilution
S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.
() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102425\
Data File : PP076049.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 24 Oct 2025 21:24
Operator : YP\AJ
Sample : Q3450-20
Misc :
ALS Vial : 33 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
SS-21-(0.0-0.5)

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 27 05:12:05 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 24 15:40:32 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.640	3.779	34345481	149.5E6	24.033	22.583
2)	SA Decachlor...	10.411	8.776	26890948	179.3E6	24.979	21.803
	Target Compounds						
26)	L6 AR-1254-1	6.641	5.656	44357794	459.7E6	555.212	815.008 #
27)	L6 AR-1254-2	6.856	5.804	97377943	593.5E6	973.933	1119.679
28)	L6 AR-1254-3	7.219	6.205	174.6E6	1691.9E6	1679.422	1987.948
29)	L6 AR-1254-4	7.501	6.431	106.6E6	1034.7E6	1274.780	1642.925 #
30)	L6 AR-1254-5	7.917	6.848	155.2E6	1873.1E6	1694.686	2601.893 #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Sample Results: **PP076049.D**


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102425\
Data File : PP076049.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 24 Oct 2025 21:24
Operator : YP\AJ
Sample : Q3450-20
Misc :
ALS Vial : 33 Sample Multiplier: 1


Instrument :
ECD_P
ClientSampleId :
SS-21-(0.0-0.5)

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 27 05:12:05 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 24 15:40:32 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm








#1 Tetrachloro-m-xylene


R.T.: 4.640 min
Delta R.T.: -0.002 min
Response: 34345481
Conc: 24.03 ng/ml

Instrument :
ECD_P
ClientSampleId :
SS-21-(0.0-0.5)




#1 Tetrachloro-m-xylene

R.T.: 3.779 min
Delta R.T.: -0.002 min
Response: 149460452
Conc: 22.58 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.411 min
Delta R.T.: -0.008 min
Response: 26890948
Conc: 24.98 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.776 min
Delta R.T.: -0.006 min
Response: 179283186
Conc: 21.80 ng/ml



#26 AR-1254-1


R.T.: 6.641 min
Delta R.T.: -0.005 min
Response: 44357794
Conc: 555.21 ng/ml

Instrument :
ECD_P
ClientSampleId :
SS-21-(0.0-0.5)




#26 AR-1254-1

R.T.: 5.656 min
Delta R.T.: -0.005 min
Response: 459664343
Conc: 815.01 ng/ml




#27 AR-1254-2

R.T.: 6.856 min
Delta R.T.: -0.007 min
Response: 97377943
Conc: 973.93 ng/ml



#27 AR-1254-2


R.T.: 5.804 min
Delta R.T.: -0.005 min
Response: 593520103
Conc: 1119.68 ng/ml



#28 AR-1254-3


R.T.: 7.219 min
Delta R.T.: -0.007 min
Response: 174583913
Conc: 1679.42 ng/m

Instrument :
ECD_P
ClientSampleId :
SS-21-(0.0-0.5)




#28 AR-1254-3

R.T.: 6.205 min
Delta R.T.: -0.006 min
Response: 1691924436
Conc: 1987.95 ng/ml




#29 AR-1254-4

R.T.: 7.501 min
Delta R.T.: -0.008 min
Response: 106589163
Conc: 1274.78 ng/ml



#29 AR-1254-4


R.T.: 6.431 min
Delta R.T.: -0.006 min
Response: 1034683862
Conc: 1642.93 ng/ml



#30 AR-1254-5

R.T.: 7.917 min
Delta R.T.: -0.008 min
Response: 155183728
Conc: 1694.69 ng/m

Instrument :
ECD_P
ClientSampleId :
SS-21-(0.0-0.5)



#30 AR-1254-5

R.T.: 6.848 min
Delta R.T.: -0.006 min
Response: 1873057153
Conc: 2601.89 ng/ml

Raw Data: **PP076072.D**

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Matrix New World Engineering	Date Collected:	10/23/25
Project:	PSEG Athenia Substation	Date Received:	10/23/25
Client Sample ID:	SS-21-(0.0-0.5)DL	SDG No.:	Q3450
Lab Sample ID:	Q3450-20DL	Matrix:	SOIL
Analytical Method:	8082A	% Solid:	91.7
Sample Wt/Vol:	30.05 g Final Vol: 10000 uL	Test:	PCB
Prep Method:	SW3541B Prep Date: 10/24/25		

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
12674-11-2	Aroclor-1016	21.5	UD	5	21.5	92.5	ug/kg	10/27/25 11:42	PB170236
11104-28-2	Aroclor-1221	21.9	UD	5	21.9	92.5	ug/kg	10/27/25 11:42	PB170236
11141-16-5	Aroclor-1232	20.2	UD	5	20.2	92.5	ug/kg	10/27/25 11:42	PB170236
53469-21-9	Aroclor-1242	21.8	UD	5	21.8	92.5	ug/kg	10/27/25 11:42	PB170236
12672-29-6	Aroclor-1248	32.2	UD	5	32.2	92.5	ug/kg	10/27/25 11:42	PB170236
11097-69-1	Aroclor-1254	509	D	5	17.5	92.5	ug/kg	10/27/25 11:42	PB170236
37324-23-5	Aroclor-1262	27.3	UD	5	27.3	92.5	ug/kg	10/27/25 11:42	PB170236
11100-14-4	Aroclor-1268	19.6	UD	5	19.6	92.5	ug/kg	10/27/25 11:42	PB170236
11096-82-5	Aroclor-1260	17.6	UD	5	17.6	92.5	ug/kg	10/27/25 11:42	PB170236
SURROGATES									
877-09-8	Tetrachloro-m-xylene	21.4			30 (21) - 150 (165)	107%	SPK: 20		
2051-24-3	Decachlorobiphenyl	24.8			30 (10) - 150 (170)	124%	SPK: 20		

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
E = Value Exceeds Calibration Range
P = Indicates >25% difference for detected concentrations between the two GC columns
Q = indicates LCS control criteria did not meet requirements
M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
B = Analyte Found in Associated Method Blank
N = Presumptive Evidence of a Compound
* = Values outside of QC limits
D = Dilution
S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.
() = Laboratory InHouse Limit

Sample Results: **PP076072.D**

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102725\
Data File : PP076072.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Oct 2025 11:42
Operator : YP\AJ
Sample : Q3450-20DL 5X
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
SS-21-(0.0-0.5)DL

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/28/2025
Supervised By :mohammad ahmed 10/29/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 27 11:55:27 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 24 15:40:32 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.639	3.780	6137390	27135235	4.295m	4.100
2)	SA Decachlor...	10.407	8.777	5339287	31864282	4.960	3.875
	Target Compounds						
26)	L6 AR-1254-1	6.641	5.657	11528188	95819976	144.295	169.894
27)	L6 AR-1254-2	6.856	5.804	22107812	114.2E6	221.113	215.422
28)	L6 AR-1254-3	7.219	6.205	37628464	275.6E6	361.970	323.852
29)	L6 AR-1254-4	7.500	6.431	24138913	172.4E6	288.695	273.687
30)	L6 AR-1254-5	7.917	6.848	35280033	234.1E6	385.276	325.201

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Sample Results: **PP076072.D**

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102725\
Data File : PP076072.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Oct 2025 11:42
Operator : YP\AJ
Sample : Q3450-20DL 5X
Misc :
ALS Vial : 12 Sample Multiplier: 1


Instrument :
ECD_P
ClientSampleId :
SS-21-(0.0-0.5)DL


Manual Integrations APPROVED


Reviewed By :Yogesh Patel 10/28/2025
Supervised By :mohammad ahmed 10/29/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 27 11:55:27 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 24 15:40:32 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm








#1 Tetrachloro-m-xylene

R.T.: 4.639 min
Delta R.T.: -0.003 min
Response: 6137390
Conc: 4.29 ng/ml

Instrument :
ECD_P
ClientSampleId :
SS-21-(0.0-0.5)DL


Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/28/2025
Supervised By :mohammad ahmed 10/29/2025




#1 Tetrachloro-m-xylene

R.T.: 3.780 min
Delta R.T.: -0.002 min
Response: 27135235
Conc: 4.10 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.407 min
Delta R.T.: -0.012 min
Response: 5339287
Conc: 4.96 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.777 min
Delta R.T.: -0.005 min
Response: 31864282
Conc: 3.88 ng/ml




#26 AR-1254-1

R.T.: 6.641 min
Delta R.T.: -0.006 min
Response: 11528188
Conc: 144.29 ng/ml

Instrument :
ECD_P
ClientSampleId :
SS-21-(0.0-0.5)DL


Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/28/2025
Supervised By :mohammad ahmed 10/29/2025




#26 AR-1254-1

R.T.: 5.657 min
Delta R.T.: -0.004 min
Response: 95819976
Conc: 169.89 ng/ml




#27 AR-1254-2

R.T.: 6.856 min
Delta R.T.: -0.008 min
Response: 22107812
Conc: 221.11 ng/ml



#27 AR-1254-2

R.T.: 5.804 min
Delta R.T.: -0.004 min
Response: 114191174
Conc: 215.42 ng/ml




#28 AR-1254-3

R.T.: 7.219 min
Delta R.T.: -0.007 min
Response: 37628464
Conc: 361.97 ng/ml

Instrument :
ECD_P
ClientSampleId :
SS-21-(0.0-0.5)DL


Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/28/2025
Supervised By :mohammad ahmed 10/29/2025




#28 AR-1254-3

R.T.: 6.205 min
Delta R.T.: -0.005 min
Response: 275627803
Conc: 323.85 ng/ml




#29 AR-1254-4

R.T.: 7.500 min
Delta R.T.: -0.008 min
Response: 24138913
Conc: 288.70 ng/ml



#29 AR-1254-4

R.T.: 6.431 min
Delta R.T.: -0.006 min
Response: 172363131
Conc: 273.69 ng/ml




#30 AR-1254-5

R.T.: 7.917 min
Delta R.T.: -0.008 min
Response: 35280033
Conc: 385.28 ng/ml

Instrument :
ECD_P
ClientSampleId :
SS-21-(0.0-0.5)DL

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/28/2025
Supervised By :mohammad ahmed 10/29/2025



#30 AR-1254-5

R.T.: 6.848 min
Delta R.T.: -0.005 min
Response: 234106565
Conc: 325.20 ng/ml

Raw Data: **PP076050.D**

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Matrix New World Engineering	Date Collected:	10/23/25
Project:	PSEG Athenia Substation	Date Received:	10/23/25
Client Sample ID:	SS-21-(2.0-2.5)	SDG No.:	Q3450
Lab Sample ID:	Q3450-21	Matrix:	SOIL
Analytical Method:	8082A	% Solid:	93.2
Sample Wt/Vol:	30.08 g Final Vol: 10000 uL	Test:	PCB
Prep Method:	SW3541B Prep Date: 10/24/25		

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
12674-11-2	Aroclor-1016	4.20	U	1	4.20	18.2	ug/kg	10/24/25 21:41	PB170236
11104-28-2	Aroclor-1221	4.30	U	1	4.30	18.2	ug/kg	10/24/25 21:41	PB170236
11141-16-5	Aroclor-1232	4.00	U	1	4.00	18.2	ug/kg	10/24/25 21:41	PB170236
53469-21-9	Aroclor-1242	4.30	U	1	4.30	18.2	ug/kg	10/24/25 21:41	PB170236
12672-29-6	Aroclor-1248	6.30	U	1	6.30	18.2	ug/kg	10/24/25 21:41	PB170236
11097-69-1	Aroclor-1254	3.40	U	1	3.40	18.2	ug/kg	10/24/25 21:41	PB170236
37324-23-5	Aroclor-1262	5.40	U	1	5.40	18.2	ug/kg	10/24/25 21:41	PB170236
11100-14-4	Aroclor-1268	3.90	U	1	3.90	18.2	ug/kg	10/24/25 21:41	PB170236
11096-82-5	Aroclor-1260	3.50	U	1	3.50	18.2	ug/kg	10/24/25 21:41	PB170236
SURROGATES									
877-09-8	Tetrachloro-m-xylene	24.5			30 (21) - 150 (165)	123%	SPK: 20		
2051-24-3	Decachlorobiphenyl	26.4			30 (10) - 150 (170)	132%	SPK: 20		

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
E = Value Exceeds Calibration Range
P = Indicates >25% difference for detected concentrations between the two GC columns
Q = indicates LCS control criteria did not meet requirements
M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
B = Analyte Found in Associated Method Blank
N = Presumptive Evidence of a Compound
* = Values outside of QC limits
D = Dilution
S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.
() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102425\
Data File : PP076050.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 24 Oct 2025 21:41
Operator : YP\AJ
Sample : Q3450-21
Misc :
ALS Vial : 34 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
SS-21-(2.0-2.5)

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 24 23:41:10 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 24 15:40:32 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.643	3.783	35015538	149.5E6	24.502	22.590
2)	SA Decachlor...	10.413	8.779	28463996	192.1E6	26.441	23.367
	Target Compounds						

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Sample Results: **PP076050.D**


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102425\
Data File : PP076050.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 24 Oct 2025 21:41
Operator : YP\AJ
Sample : Q3450-21
Misc :
ALS Vial : 34 Sample Multiplier: 1


Instrument :
ECD_P
ClientSampleId :
SS-21-(2.0-2.5)

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 24 23:41:10 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 24 15:40:32 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm








#1 Tetrachloro-m-xylene


R.T.: 4.643 min
Delta R.T.: 0.001 min
Response: 35015538
Conc: 24.50 ng/ml

Instrument :
ECD_P
ClientSampleId :
SS-21-(2.0-2.5)




#1 Tetrachloro-m-xylene

R.T.: 3.783 min
Delta R.T.: 0.002 min
Response: 149509931
Conc: 22.59 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.413 min
Delta R.T.: -0.006 min
Response: 28463996
Conc: 26.44 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.779 min
Delta R.T.: -0.003 min
Response: 192147524
Conc: 23.37 ng/ml

Raw Data: **PP076051.D**

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Matrix New World Engineering	Date Collected:	10/23/25
Project:	PSEG Athenia Substation	Date Received:	10/23/25
Client Sample ID:	SS-21-(3.5-4.0)	SDG No.:	Q3450
Lab Sample ID:	Q3450-22	Matrix:	SOIL
Analytical Method:	8082A	% Solid:	84.7
Sample Wt/Vol:	30.04 g Final Vol: 10000 uL	Test:	PCB
Prep Method:	SW3541B Prep Date: 10/24/25		

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
12674-11-2	Aroclor-1016	4.70	U	1	4.70	20.0	ug/kg	10/24/25 21:57	PB170236
11104-28-2	Aroclor-1221	4.80	U	1	4.80	20.0	ug/kg	10/24/25 21:57	PB170236
11141-16-5	Aroclor-1232	4.40	U	1	4.40	20.0	ug/kg	10/24/25 21:57	PB170236
53469-21-9	Aroclor-1242	4.70	U	1	4.70	20.0	ug/kg	10/24/25 21:57	PB170236
12672-29-6	Aroclor-1248	7.00	U	1	7.00	20.0	ug/kg	10/24/25 21:57	PB170236
11097-69-1	Aroclor-1254	516	E	1	3.80	20.0	ug/kg	10/24/25 21:57	PB170236
37324-23-5	Aroclor-1262	5.90	U	1	5.90	20.0	ug/kg	10/24/25 21:57	PB170236
11100-14-4	Aroclor-1268	4.20	U	1	4.20	20.0	ug/kg	10/24/25 21:57	PB170236
11096-82-5	Aroclor-1260	3.80	U	1	3.80	20.0	ug/kg	10/24/25 21:57	PB170236
SURROGATES									
877-09-8	Tetrachloro-m-xylene	20.1			30 (21) - 150 (165)	100%	SPK: 20		
2051-24-3	Decachlorobiphenyl	17.2			30 (10) - 150 (170)	86%	SPK: 20		

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
E = Value Exceeds Calibration Range
P = Indicates >25% difference for detected concentrations between the two GC columns
Q = indicates LCS control criteria did not meet requirements
M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
B = Analyte Found in Associated Method Blank
N = Presumptive Evidence of a Compound
* = Values outside of QC limits
D = Dilution
S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.
() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102425\
Data File : PP076051.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 24 Oct 2025 21:57
Operator : YP\AJ
Sample : Q3450-22
Misc :
ALS Vial : 35 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
SS-21-(3.5-4.0)

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 27 05:16:26 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 24 15:40:32 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.639	3.779	28698180	125.8E6	20.081	19.010
2)	SA Decachlor...	10.409	8.777	18498607	124.5E6	17.184	15.145
	Target Compounds						
26)	L6 AR-1254-1	6.640	5.657	12072300	127.8E6	151.105	226.514 #
27)	L6 AR-1254-2	6.856	5.804	41710840	229.4E6	417.174	432.824
28)	L6 AR-1254-3	7.218	6.204	98268988	940.9E6	945.305	1105.574
29)	L6 AR-1254-4	7.500	6.431	112.9E6	1078.5E6	1350.585	1712.431 #
30)	L6 AR-1254-5	7.916	6.848	206.9E6	2226.6E6	2259.030	3092.969 #


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102425\
Data File : PP076051.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 24 Oct 2025 21:57
Operator : YP\AJ
Sample : Q3450-22
Misc :
ALS Vial : 35 Sample Multiplier: 1


Instrument :
ECD_P
ClientSampleId :
SS-21-(3.5-4.0)

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 27 05:16:26 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 24 15:40:32 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm








#1 Tetrachloro-m-xylene


R.T.: 4.639 min
Delta R.T.: -0.003 min
Response: 28698180
Conc: 20.08 ng/ml

Instrument :
ECD_P
ClientSampleId :
SS-21-(3.5-4.0)




#1 Tetrachloro-m-xylene

R.T.: 3.779 min
Delta R.T.: -0.002 min
Response: 125816731
Conc: 19.01 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.409 min
Delta R.T.: -0.010 min
Response: 18498607
Conc: 17.18 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.777 min
Delta R.T.: -0.005 min
Response: 124535846
Conc: 15.14 ng/ml



#26 AR-1254-1


R.T.: 6.640 min
Delta R.T.: -0.007 min
Response: 12072300
Conc: 151.10 ng/ml

Instrument :
ECD_P
ClientSampleId :
SS-21-(3.5-4.0)




#26 AR-1254-1

R.T.: 5.657 min
Delta R.T.: -0.005 min
Response: 127753836
Conc: 226.51 ng/ml




#27 AR-1254-2

R.T.: 6.856 min
Delta R.T.: -0.008 min
Response: 41710840
Conc: 417.17 ng/ml



#27 AR-1254-2


R.T.: 5.804 min
Delta R.T.: -0.005 min
Response: 229431687
Conc: 432.82 ng/ml



#28 AR-1254-3


R.T.: 7.218 min
Delta R.T.: -0.008 min
Response: 98268988
Conc: 945.31 ng/ml

Instrument :
ECD_P
ClientSampleId :
SS-21-(3.5-4.0)




#28 AR-1254-3

R.T.: 6.204 min
Delta R.T.: -0.006 min
Response: 940943554
Conc: 1105.57 ng/ml




#29 AR-1254-4

R.T.: 7.500 min
Delta R.T.: -0.009 min
Response: 112927507
Conc: 1350.58 ng/ml



#29 AR-1254-4


R.T.: 6.431 min
Delta R.T.: -0.006 min
Response: 1078457206
Conc: 1712.43 ng/ml



#30 AR-1254-5

R.T.: 7.916 min
Delta R.T.: -0.009 min
Response: 206861114
Conc: 2259.03 ng/m

Instrument :
ECD_P
ClientSampleId :
SS-21-(3.5-4.0)



#30 AR-1254-5

R.T.: 6.848 min
Delta R.T.: -0.006 min
Response: 2226573531
Conc: 3092.97 ng/ml

Raw Data: **PP076073.D**

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Matrix New World Engineering	Date Collected:	10/23/25
Project:	PSEG Athenia Substation	Date Received:	10/23/25
Client Sample ID:	SS-21-(3.5-4.0)DL	SDG No.:	Q3450
Lab Sample ID:	Q3450-22DL	Matrix:	SOIL
Analytical Method:	8082A	% Solid:	84.7
Sample Wt/Vol:	30.04 g Final Vol: 10000 uL	Test:	PCB
Prep Method:	SW3541B Prep Date: 10/24/25		

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
12674-11-2	Aroclor-1016	9.30	UD	2	9.30	40.1	ug/kg	10/27/25 11:58	PB170236
11104-28-2	Aroclor-1221	9.50	UD	2	9.50	40.1	ug/kg	10/27/25 11:58	PB170236
11141-16-5	Aroclor-1232	8.80	UD	2	8.80	40.1	ug/kg	10/27/25 11:58	PB170236
53469-21-9	Aroclor-1242	9.50	UD	2	9.50	40.1	ug/kg	10/27/25 11:58	PB170236
12672-29-6	Aroclor-1248	14.0	UD	2	14.0	40.1	ug/kg	10/27/25 11:58	PB170236
11097-69-1	Aroclor-1254	503	D	2	7.60	40.1	ug/kg	10/27/25 11:58	PB170236
37324-23-5	Aroclor-1262	11.8	UD	2	11.8	40.1	ug/kg	10/27/25 11:58	PB170236
11100-14-4	Aroclor-1268	8.50	UD	2	8.50	40.1	ug/kg	10/27/25 11:58	PB170236
11096-82-5	Aroclor-1260	7.60	UD	2	7.60	40.1	ug/kg	10/27/25 11:58	PB170236
SURROGATES									
877-09-8	Tetrachloro-m-xylene	21.6			30 (21) - 150 (165)	108%	SPK: 20		
2051-24-3	Decachlorobiphenyl	17.5			30 (10) - 150 (170)	88%	SPK: 20		

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
E = Value Exceeds Calibration Range
P = Indicates >25% difference for detected concentrations between the two GC columns
Q = indicates LCS control criteria did not meet requirements
M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
B = Analyte Found in Associated Method Blank
N = Presumptive Evidence of a Compound
* = Values outside of QC limits
D = Dilution
S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.
() = Laboratory InHouse Limit

Sample Results: **PP076073.D**

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102725\
Data File : PP076073.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Oct 2025 11:58
Operator : YP\AJ
Sample : Q3450-22DL 2X
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
SS-21-(3.5-4.0)DL

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/28/2025
Supervised By :mohammad ahmed 10/29/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 27 12:11:39 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 24 15:40:32 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.637	3.778	15408285	65330300	10.782	9.871
2)	SA Decachlor...	10.404	8.771	9443447	56936322	8.772	6.924
	Target Compounds						
26)	L6 AR-1254-1	6.635	5.667	9652989	102.5E6	120.823m	181.709m#
27)	L6 AR-1254-2	6.851	5.799	23356369	116.0E6	233.600m	218.818m
28)	L6 AR-1254-3	7.214	6.202	62673662	469.4E6	602.894	551.583
29)	L6 AR-1254-4	7.497	6.428	61806136	538.1E6	739.186	854.358
30)	L6 AR-1254-5	7.913	6.844	106.8E6	1004.3E6	1165.820	1395.048

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Sample Results: **PP076073.D**

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102725\
Data File : PP076073.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Oct 2025 11:58
Operator : YP\AJ
Sample : Q3450-22DL 2X
Misc :
ALS Vial : 13 Sample Multiplier: 1


Instrument :
ECD_P
ClientSampleId :
SS-21-(3.5-4.0)DL


Manual Integrations APPROVED


Reviewed By :Yogesh Patel 10/28/2025
Supervised By :mohammad ahmed 10/29/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 27 12:11:39 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 24 15:40:32 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm








#1 Tetrachloro-m-xylene

R.T.: 4.637 min
Delta R.T.: -0.005 min
Response: 15408285
Conc: 10.78 ng/ml

Instrument :
ECD_P
ClientSampleId :
SS-21-(3.5-4.0)DL


Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/28/2025
Supervised By :mohammad ahmed 10/29/2025




#1 Tetrachloro-m-xylene

R.T.: 3.778 min
Delta R.T.: -0.003 min
Response: 65330300
Conc: 9.87 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.404 min
Delta R.T.: -0.015 min
Response: 9443447
Conc: 8.77 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.771 min
Delta R.T.: -0.011 min
Response: 56936322
Conc: 6.92 ng/ml




#26 AR-1254-1

R.T.: 6.635 min
Delta R.T.: -0.011 min
Response: 9652989
Conc: 120.82 ng/ml

Instrument :
ECD_P
ClientSampleId :
SS-21-(3.5-4.0)DL


Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/28/2025
Supervised By :mohammad ahmed 10/29/2025




#26 AR-1254-1

R.T.: 5.667 min
Delta R.T.: 0.005 min
Response: 102483861
Conc: 181.71 ng/ml m




#27 AR-1254-2

R.T.: 6.851 min
Delta R.T.: -0.013 min
Response: 23356369
Conc: 233.60 ng/ml m



#27 AR-1254-2

R.T.: 5.799 min
Delta R.T.: -0.010 min
Response: 115991207
Conc: 218.82 ng/ml m




#28 AR-1254-3

R.T.: 7.214 min
Delta R.T.: -0.012 min
Response: 62673662
Conc: 602.89 ng/ml

Instrument :
ECD_P
ClientSampleId :
SS-21-(3.5-4.0)DL


Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/28/2025
Supervised By :mohammad ahmed 10/29/2025




#28 AR-1254-3

R.T.: 6.202 min
Delta R.T.: -0.009 min
Response: 469446908
Conc: 551.58 ng/ml




#29 AR-1254-4

R.T.: 7.497 min
Delta R.T.: -0.012 min
Response: 61806136
Conc: 739.19 ng/ml



#29 AR-1254-4

R.T.: 6.428 min
Delta R.T.: -0.009 min
Response: 538058919
Conc: 854.36 ng/ml




#30 AR-1254-5

R.T.: 7.913 min
Delta R.T.: -0.012 min
Response: 106755060
Conc: 1165.82 ng/m

Instrument :
ECD_P
ClientSampleId :
SS-21-(3.5-4.0)DL

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/28/2025
Supervised By :mohammad ahmed 10/29/2025



#30 AR-1254-5

R.T.: 6.844 min
Delta R.T.: -0.010 min
Response: 1004270507
Conc: 1395.05 ng/ml

Raw Data: **PP076284.D**

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Matrix New World Engineering	Date Collected:	10/23/25
Project:	PSEG Athenia Substation	Date Received:	10/23/25
Client Sample ID:	SS-22-(0.0-0.5)	SDG No.:	Q3450
Lab Sample ID:	Q3450-23	Matrix:	SOIL
Analytical Method:	8082A	% Solid:	91.4
Sample Wt/Vol:	30.05 g	Final Vol: 10000 uL	Test: PCB
Prep Method:	SW3541B	Prep Date: 10/31/25	

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
12674-11-2	Aroclor-1016	4.30	U	1	4.30	18.6	ug/kg	11/01/25 01:01	PB170358
11104-28-2	Aroclor-1221	4.40	U	1	4.40	18.6	ug/kg	11/01/25 01:01	PB170358
11141-16-5	Aroclor-1232	4.10	U	1	4.10	18.6	ug/kg	11/01/25 01:01	PB170358
53469-21-9	Aroclor-1242	4.40	U	1	4.40	18.6	ug/kg	11/01/25 01:01	PB170358
12672-29-6	Aroclor-1248	6.50	U	1	6.50	18.6	ug/kg	11/01/25 01:01	PB170358
11097-69-1	Aroclor-1254	1200	E	1	3.50	18.6	ug/kg	11/01/25 01:01	PB170358
37324-23-5	Aroclor-1262	5.50	U	1	5.50	18.6	ug/kg	11/01/25 01:01	PB170358
11100-14-4	Aroclor-1268	3.90	U	1	3.90	18.6	ug/kg	11/01/25 01:01	PB170358
11096-82-5	Aroclor-1260	3.50	U	1	3.50	18.6	ug/kg	11/01/25 01:01	PB170358
SURROGATES									
877-09-8	Tetrachloro-m-xylene	22.0			30 (21) - 150 (165)	110%	SPK: 20		
2051-24-3	Decachlorobiphenyl	19.4			30 (10) - 150 (170)	97%	SPK: 20		

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
E = Value Exceeds Calibration Range
P = Indicates >25% difference for detected concentrations between the two GC columns
Q = indicates LCS control criteria did not meet requirements
M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
B = Analyte Found in Associated Method Blank
N = Presumptive Evidence of a Compound
* = Values outside of QC limits
D = Dilution
S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.
() = Laboratory InHouse Limit

Quantitation Report (QT Reviewed)

Sample Results: **PP076284.D**

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP103125\
Data File : PP076284.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 01 Nov 2025 01:01
Operator : YP\AJ
Sample : Q3450-23
Misc :
ALS Vial : 34 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
SS-22-(0.0-0.5)

Manual Integrations APPROVED

Reviewed By :Rahul Chavli

11/03/2025

Supervised By :Yogesh Patel

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 01 02:49:25 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 24 15:40:32 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

11/03/2025

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml	
	System Monitoring Compounds							
1)	SA Tetrachlo...	4.646	3.782	31377746	135.0E6	21.956	20.401	
2)	SA Decachlor...	10.422	8.779	20863650	131.2E6	19.381	15.955	
	Target Compounds							
26)	L6 AR-1254-1	6.646	5.659	123.8E6	1151.2E6	1550.060m	2041.206	#
27)	L6 AR-1254-2	6.863	5.805	203.0E6	1292.8E6	2030.551	2438.853	
28)	L6 AR-1254-3	7.226	6.208	283.8E6	2901.5E6	2730.332	3409.135	
29)	L6 AR-1254-4	7.508	6.434	235.8E6	2336.1E6	2820.438	3709.313	#
30)	L6 AR-1254-5	7.924	6.851	366.0E6	3915.1E6	3996.544	5438.502	#

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Sample Results: **PP076284.D**

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP103125\
Data File : PP076284.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 01 Nov 2025 01:01
Operator : YP\AJ
Sample : Q3450-23
Misc :
ALS Vial : 34 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
SS-22-(0.0-0.5)


Manual Integrations APPROVED


Reviewed By :Rahul Chavli


11/03/2025
Supervised By :Yogesh Patel

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 01 02:49:25 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 24 15:40:32 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm







#1 Tetrachloro-m-xylene

R.T.: 4.646 min
Delta R.T.: 0.004 min
Response: 31377746
Conc: 21.96 ng/ml


Instrument :
ECD_P
ClientSampleId :
SS-22-(0.0-0.5)

Manual Integrations APPROVED

Reviewed By :Rahul Chavli


11/03/2025
Supervised By :Yogesh Patel

11/03/2025




#1 Tetrachloro-m-xylene

R.T.: 3.782 min
Delta R.T.: 0.000 min
Response: 135017485
Conc: 20.40 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.422 min
Delta R.T.: 0.003 min
Response: 20863650
Conc: 19.38 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.779 min
Delta R.T.: -0.003 min
Response: 131195874
Conc: 15.95 ng/ml




#26 AR-1254-1

R.T.: 6.646 min
Delta R.T.: 0.000 min
Response: 123839648
Conc: 1550.06 ng/m

Instrument :
ECD_P
ClientSampleId :
SS-22-(0.0-0.5)

Manual Integrations APPROVED

Reviewed By :Rahul Chavli




#26 AR-1254-1

R.T.: 5.659 min
Delta R.T.: -0.003 min
Response: 1151240111
Conc: 2041.21 ng/ml


11/03/2025
Supervised By :Yogesh Patel

11/03/2025




#27 AR-1254-2

R.T.: 6.863 min
Delta R.T.: 0.000 min
Response: 203023094
Conc: 2030.55 ng/ml



#27 AR-1254-2

R.T.: 5.805 min
Delta R.T.: -0.003 min
Response: 1292789299
Conc: 2438.85 ng/ml



#28 AR-1254-3


R.T.: 7.226 min
Delta R.T.: 0.000 min
Response: 283831015
Conc: 2730.33 ng/m

Instrument :
ECD_P
ClientSampleId :
SS-22-(0.0-0.5)

Manual Integrations APPROVED

Reviewed By :Rahul Chavli

11/03/2025




#28 AR-1254-3

R.T.: 6.208 min
Delta R.T.: -0.003 min
Response: 2901483957
Conc: 3409.14 ng/ml


Supervised By :Yogesh Patel

11/03/2025




#29 AR-1254-4

R.T.: 7.508 min
Delta R.T.: 0.000 min
Response: 235827460
Conc: 2820.44 ng/ml



#29 AR-1254-4

R.T.: 6.434 min
Delta R.T.: -0.003 min
Response: 2336056816
Conc: 3709.31 ng/ml



#30 AR-1254-5

R.T.: 7.924 min
Delta R.T.: 0.000 min
Response: 365966573
Conc: 3996.54 ng/m

Instrument :
ECD_P
ClientSampleId :
SS-22-(0.0-0.5)


Manual Integrations APPROVED

Reviewed By :Rahul Chavli

11/03/2025

Supervised By :Yogesh Patel

11/03/2025



#30 AR-1254-5

R.T.: 6.851 min
Delta R.T.: -0.002 min
Response: 3915081768
Conc: 5438.50 ng/ml

Raw Data: **PP076297.D**

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Matrix New World Engineering	Date Collected:	10/23/25
Project:	PSEG Athenia Substation	Date Received:	10/23/25
Client Sample ID:	SS-22-(0.0-0.5)DL	SDG No.:	Q3450
Lab Sample ID:	Q3450-23DL	Matrix:	SOIL
Analytical Method:	8082A	% Solid:	91.4
Sample Wt/Vol:	30.05 g Final Vol: 10000 uL	Test:	PCB
Prep Method:	SW3541B Prep Date: 10/31/25		

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
12674-11-2	Aroclor-1016	21.6	UD	5	21.6	92.8	ug/kg	11/03/25 10:09	PB170358
11104-28-2	Aroclor-1221	22.0	UD	5	22.0	92.8	ug/kg	11/03/25 10:09	PB170358
11141-16-5	Aroclor-1232	20.3	UD	5	20.3	92.8	ug/kg	11/03/25 10:09	PB170358
53469-21-9	Aroclor-1242	21.9	UD	5	21.9	92.8	ug/kg	11/03/25 10:09	PB170358
12672-29-6	Aroclor-1248	32.3	UD	5	32.3	92.8	ug/kg	11/03/25 10:09	PB170358
11097-69-1	Aroclor-1254	881	D	5	17.5	92.8	ug/kg	11/03/25 10:09	PB170358
37324-23-5	Aroclor-1262	27.4	UD	5	27.4	92.8	ug/kg	11/03/25 10:09	PB170358
11100-14-4	Aroclor-1268	19.7	UD	5	19.7	92.8	ug/kg	11/03/25 10:09	PB170358
11096-82-5	Aroclor-1260	17.6	UD	5	17.6	92.8	ug/kg	11/03/25 10:09	PB170358
SURROGATES									
877-09-8	Tetrachloro-m-xylene	15.8			30 (21) - 150 (165)	79%	SPK: 20		
2051-24-3	Decachlorobiphenyl	16.4			30 (10) - 150 (170)	82%	SPK: 20		

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
E = Value Exceeds Calibration Range
P = Indicates >25% difference for detected concentrations between the two GC columns
Q = indicates LCS control criteria did not meet requirements
M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
B = Analyte Found in Associated Method Blank
N = Presumptive Evidence of a Compound
* = Values outside of QC limits
D = Dilution
S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.
() = Laboratory InHouse Limit

Sample Results: **PP076297.D**

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP110325\
Data File : PP076297.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 03 Nov 2025 10:09
Operator : YP\AJ
Sample : Q3450-23DL 5X
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
SS-22-(0.0-0.5)DL

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 11/04/2025
Supervised By :Yogesh Patel 11/04/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 03 11:23:32 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 24 15:40:32 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.637	3.773	4514069	15013005	3.159	2.268m#
2)	SA Decachlor...	10.405	8.773	3531998	19550676	3.281	2.378 #
	Target Compounds						
26)	L6 AR-1254-1	6.637	5.652	21000162	163.2E6	262.852m	289.383
27)	L6 AR-1254-2	6.853	5.798	34761287	191.5E6	347.668	361.243
28)	L6 AR-1254-3	7.216	6.200	47209842	405.9E6	454.138	476.956
29)	L6 AR-1254-4	7.498	6.426	39785447	337.7E6	475.824	536.186
30)	L6 AR-1254-5	7.914	6.842	60658808	544.9E6	662.425	756.904

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP110325\
Data File : PP076297.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 03 Nov 2025 10:09
Operator : YP\AJ
Sample : Q3450-23DL 5X
Misc :
ALS Vial : 7 Sample Multiplier: 1


Instrument :
ECD_P
ClientSampleId :
SS-22-(0.0-0.5)DL


Manual Integrations APPROVED


Reviewed By :Rahul Chavli 11/04/2025
Supervised By :Yogesh Patel 11/04/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 03 11:23:32 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 24 15:40:32 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm








#1 Tetrachloro-m-xylene

R.T.: 4.637 min
Delta R.T.: -0.005 min
Response: 4514069
Conc: 3.16 ng/ml

Instrument :
ECD_P
ClientSampleId :
SS-22-(0.0-0.5)DL


Manual Integrations APPROVED

Reviewed By :Rahul Chavli 11/04/2025
Supervised By :Yogesh Patel 11/04/2025




#1 Tetrachloro-m-xylene

R.T.: 3.773 min
Delta R.T.: -0.008 min
Response: 15013005
Conc: 2.27 ng/ml m




#2 Decachlorobiphenyl

R.T.: 10.405 min
Delta R.T.: -0.014 min
Response: 3531998
Conc: 3.28 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.773 min
Delta R.T.: -0.009 min
Response: 19550676
Conc: 2.38 ng/ml




#26 AR-1254-1

R.T.: 6.637 min
Delta R.T.: -0.010 min
Response: 21000162
Conc: 262.85 ng/ml

Instrument :
ECD_P
ClientSampleId :
SS-22-(0.0-0.5)DL


Manual Integrations APPROVED

Reviewed By :Rahul Chavli 11/04/2025
Supervised By :Yogesh Patel 11/04/2025




#26 AR-1254-1

R.T.: 5.652 min
Delta R.T.: -0.010 min
Response: 163212125
Conc: 289.38 ng/ml




#27 AR-1254-2

R.T.: 6.853 min
Delta R.T.: -0.011 min
Response: 34761287
Conc: 347.67 ng/ml



#27 AR-1254-2

R.T.: 5.798 min
Delta R.T.: -0.010 min
Response: 191487786
Conc: 361.24 ng/ml




#28 AR-1254-3

R.T.: 7.216 min
Delta R.T.: -0.010 min
Response: 47209842
Conc: 454.14 ng/ml

Instrument :
ECD_P
ClientSampleId :
SS-22-(0.0-0.5)DL


Manual Integrations APPROVED

Reviewed By :Rahul Chavli 11/04/2025
Supervised By :Yogesh Patel 11/04/2025




#28 AR-1254-3

R.T.: 6.200 min
Delta R.T.: -0.011 min
Response: 405932491
Conc: 476.96 ng/ml




#29 AR-1254-4

R.T.: 7.498 min
Delta R.T.: -0.011 min
Response: 39785447
Conc: 475.82 ng/ml



#29 AR-1254-4

R.T.: 6.426 min
Delta R.T.: -0.011 min
Response: 337679889
Conc: 536.19 ng/ml




#30 AR-1254-5

R.T.: 7.914 min
Delta R.T.: -0.011 min
Response: 60658808
Conc: 662.43 ng/ml

Instrument :
ECD_P
ClientSampleId :
SS-22-(0.0-0.5)DL

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 11/04/2025
Supervised By :Yogesh Patel 11/04/2025



#30 AR-1254-5

R.T.: 6.842 min
Delta R.T.: -0.012 min
Response: 544882178
Conc: 756.90 ng/ml

Raw Data: **PP076285.D**

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Matrix New World Engineering	Date Collected:	10/23/25
Project:	PSEG Athenia Substation	Date Received:	10/23/25
Client Sample ID:	SS-22-(3.5-4.0)	SDG No.:	Q3450
Lab Sample ID:	Q3450-25	Matrix:	SOIL
Analytical Method:	8082A	% Solid:	82.7
Sample Wt/Vol:	30.04 g Final Vol: 10000 uL	Test:	PCB
Prep Method:	SW3541B Prep Date: 10/31/25		

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
12674-11-2	Aroclor-1016	4.80	U	1	4.80	20.5	ug/kg	11/01/25 01:17	PB170358
11104-28-2	Aroclor-1221	4.90	U	1	4.90	20.5	ug/kg	11/01/25 01:17	PB170358
11141-16-5	Aroclor-1232	4.50	U	1	4.50	20.5	ug/kg	11/01/25 01:17	PB170358
53469-21-9	Aroclor-1242	4.80	U	1	4.80	20.5	ug/kg	11/01/25 01:17	PB170358
12672-29-6	Aroclor-1248	7.10	U	1	7.10	20.5	ug/kg	11/01/25 01:17	PB170358
11097-69-1	Aroclor-1254	381		1	3.90	20.5	ug/kg	11/01/25 01:17	PB170358
37324-23-5	Aroclor-1262	6.10	U	1	6.10	20.5	ug/kg	11/01/25 01:17	PB170358
11100-14-4	Aroclor-1268	4.30	U	1	4.30	20.5	ug/kg	11/01/25 01:17	PB170358
11096-82-5	Aroclor-1260	3.90	U	1	3.90	20.5	ug/kg	11/01/25 01:17	PB170358
SURROGATES									
877-09-8	Tetrachloro-m-xylene	19.7			30 (21) - 150 (165)	98%	SPK: 20		
2051-24-3	Decachlorobiphenyl	13.6			30 (10) - 150 (170)	68%	SPK: 20		

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
E = Value Exceeds Calibration Range
P = Indicates >25% difference for detected concentrations between the two GC columns
Q = indicates LCS control criteria did not meet requirements
M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
B = Analyte Found in Associated Method Blank
N = Presumptive Evidence of a Compound
* = Values outside of QC limits
D = Dilution
S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.
() = Laboratory InHouse Limit

Sample Results: **PP076285.D**

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP103125\
Data File : PP076285.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 01 Nov 2025 01:17
Operator : YP\AJ
Sample : Q3450-25
Misc :
ALS Vial : 35 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
SS-22-(3.5-4.0)

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 11/03/2025
Supervised By :Yogesh Patel 11/03/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 01 02:49:43 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 24 15:40:32 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.641	3.778	28092571	122.0E6	19.658	18.435
2)	SA Decachlor...	10.417	8.777	14654797	90232389	13.613	10.973
	Target Compounds						
26)	L6 AR-1254-1	6.642	5.654	23441617	213.5E6	293.411	378.628 #
27)	L6 AR-1254-2	6.859	5.801	48601406	273.5E6	486.091	515.987
28)	L6 AR-1254-3	7.221	6.203	78274296	723.2E6	752.965	849.745
29)	L6 AR-1254-4	7.503	6.430	84446227	761.7E6	1009.956	1209.540
30)	L6 AR-1254-5	7.920	6.847	132.2E6	1277.9E6	1443.399	1775.189

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Sample Results: **PP076285.D**

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP103125\
Data File : PP076285.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 01 Nov 2025 01:17
Operator : YP\AJ
Sample : Q3450-25
Misc :
ALS Vial : 35 Sample Multiplier: 1


Instrument :
ECD_P
ClientSampleId :
SS-22-(3.5-4.0)


Manual Integrations APPROVED


Reviewed By :Rahul Chavli 11/03/2025
Supervised By :Yogesh Patel 11/03/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 01 02:49:43 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 24 15:40:32 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm








#1 Tetrachloro-m-xylene

R.T.: 4.641 min
Delta R.T.: 0.000 min
Response: 28092571
Conc: 19.66 ng/ml

Instrument :
ECD_P
ClientSampleId :
SS-22-(3.5-4.0)


Manual Integrations APPROVED

Reviewed By :Rahul Chavli 11/03/2025
Supervised By :Yogesh Patel 11/03/2025




#1 Tetrachloro-m-xylene

R.T.: 3.778 min
Delta R.T.: -0.003 min
Response: 122005455
Conc: 18.43 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.417 min
Delta R.T.: -0.002 min
Response: 14654797
Conc: 13.61 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.777 min
Delta R.T.: -0.005 min
Response: 90232389
Conc: 10.97 ng/ml




#26 AR-1254-1

R.T.: 6.642 min
Delta R.T.: -0.004 min
Response: 23441617
Conc: 293.41 ng/ml

Instrument :
ECD_P
ClientSampleId :
SS-22-(3.5-4.0)


Manual Integrations APPROVED

Reviewed By :Rahul Chavli 11/03/2025
Supervised By :Yogesh Patel 11/03/2025




#26 AR-1254-1

R.T.: 5.654 min
Delta R.T.: -0.008 min
Response: 213546316
Conc: 378.63 ng/ml




#27 AR-1254-2

R.T.: 6.859 min
Delta R.T.: -0.005 min
Response: 48601406
Conc: 486.09 ng/ml



#27 AR-1254-2

R.T.: 5.801 min
Delta R.T.: -0.008 min
Response: 273514973
Conc: 515.99 ng/ml




#28 AR-1254-3

R.T.: 7.221 min
Delta R.T.: -0.005 min
Response: 78274296
Conc: 752.97 ng/ml

Instrument :
ECD_P
ClientSampleId :
SS-22-(3.5-4.0)


Manual Integrations APPROVED

Reviewed By :Rahul Chavli 11/03/2025
Supervised By :Yogesh Patel 11/03/2025




#28 AR-1254-3

R.T.: 6.203 min
Delta R.T.: -0.008 min
Response: 723210154
Conc: 849.74 ng/ml




#29 AR-1254-4

R.T.: 7.503 min
Delta R.T.: -0.005 min
Response: 84446227
Conc: 1009.96 ng/ml



#29 AR-1254-4

R.T.: 6.430 min
Delta R.T.: -0.007 min
Response: 761745745
Conc: 1209.54 ng/ml




#30 AR-1254-5

R.T.: 7.920 min
Delta R.T.: -0.005 min
Response: 132173115
Conc: 1443.40 ng/m

Instrument :
ECD_P
ClientSampleId :
SS-22-(3.5-4.0)

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 11/03/2025
Supervised By :Yogesh Patel 11/03/2025



#30 AR-1254-5

R.T.: 6.847 min
Delta R.T.: -0.007 min
Response: 1277927286
Conc: 1775.19 ng/ml

CALIBRATION SUMMARY

Raw Data

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

RETENTION TIMES OF INITIAL CALIBRATION

Lab Name: Alliance **Contract:** MATR02

Lab Code: ACE **SDG NO.:** Q3450

Instrument ID: ECD_O **Calibration Date(s):** 11/04/2025 11/04/2025

Calibration Times: 09:39 18:31

GC Column: ZB-MR1 ID: 0.32 (mm)

LAB FILE ID: RT 1000 = PO114871.D RT 750 = PO114872.D

RT 500 = PO114873.D RT 250 = PO114874.D RT 050 = PO114875.D

COMPOUND		RT 1000	RT 750	RT 500	RT 250	RT 050	MEAN RT	RT WINDOW FROM	RT WINDOW TO
Aroclor-1016-1	(1)	5.49	5.49	5.49	5.49	5.49	5.49	5.39	5.59
Aroclor-1016-2	(2)	5.51	5.51	5.51	5.51	5.51	5.51	5.41	5.61
Aroclor-1016-3	(3)	5.57	5.57	5.57	5.57	5.57	5.57	5.47	5.67
Aroclor-1016-4	(4)	5.67	5.67	5.67	5.67	5.67	5.67	5.57	5.77
Aroclor-1016-5	(5)	5.96	5.96	5.96	5.96	5.96	5.96	5.86	6.06
Aroclor-1260-1	(1)	7.07	7.07	7.07	7.07	7.07	7.07	6.97	7.17
Aroclor-1260-2	(2)	7.33	7.33	7.33	7.33	7.33	7.33	7.23	7.43
Aroclor-1260-3	(3)	7.68	7.68	7.68	7.68	7.68	7.68	7.58	7.78
Aroclor-1260-4	(4)	7.90	7.90	7.90	7.90	7.90	7.90	7.80	8.00
Aroclor-1260-5	(5)	8.21	8.21	8.21	8.21	8.21	8.21	8.11	8.31
Decachlorobiphenyl		9.93	9.93	9.93	9.93	9.93	9.93	9.83	10.03
Tetrachloro-m-xylene		4.35	4.35	4.35	4.35	4.35	4.35	4.25	4.45
Aroclor-1242-1	(1)	5.49	5.49	5.49	5.49	5.49	5.49	5.39	5.59
Aroclor-1242-2	(2)	5.51	5.51	5.51	5.51	5.51	5.51	5.41	5.61
Aroclor-1242-3	(3)	5.57	5.57	5.57	5.57	5.57	5.57	5.47	5.67
Aroclor-1242-4	(4)	5.67	5.67	5.67	5.67	5.67	5.67	5.57	5.77
Aroclor-1242-5	(5)	6.39	6.39	6.39	6.39	6.39	6.39	6.29	6.49
Decachlorobiphenyl		9.93	9.93	9.93	9.93	9.93	9.93	9.83	10.03
Tetrachloro-m-xylene		4.35	4.35	4.35	4.35	4.35	4.35	4.25	4.45
Aroclor-1248-1	(1)	5.49	5.49	5.49	5.49	5.49	5.49	5.39	5.59
Aroclor-1248-2	(2)	5.76	5.76	5.76	5.76	5.76	5.76	5.66	5.86
Aroclor-1248-3	(3)	5.96	5.96	5.96	5.96	5.96	5.96	5.86	6.06
Aroclor-1248-4	(4)	6.35	6.36	6.35	6.35	6.36	6.35	6.25	6.45
Aroclor-1248-5	(5)	6.39	6.39	6.39	6.39	6.39	6.39	6.29	6.49
Decachlorobiphenyl		9.93	9.93	9.93	9.93	9.93	9.93	9.83	10.03
Tetrachloro-m-xylene		4.35	4.35	4.35	4.35	4.35	4.35	4.25	4.45
Aroclor-1254-1	(1)	6.33	6.33	6.33	6.33	6.33	6.33	6.23	6.43
Aroclor-1254-2	(2)	6.55	6.55	6.55	6.55	6.55	6.55	6.45	6.65
Aroclor-1254-3	(3)	6.91	6.91	6.91	6.91	6.91	6.91	6.81	7.01
Aroclor-1254-4	(4)	7.19	7.19	7.19	7.19	7.19	7.19	7.09	7.29
Aroclor-1254-5	(5)	7.60	7.60	7.60	7.60	7.60	7.60	7.50	7.70
Decachlorobiphenyl		9.93	9.93	9.93	9.93	9.93	9.93	9.83	10.03
Tetrachloro-m-xylene		4.35	4.35	4.35	4.35	4.35	4.35	4.25	4.45
Aroclor-1268-1	(1)	8.51	8.51	8.51	8.51	8.51	8.51	8.41	8.61
Aroclor-1268-2	(2)	8.60	8.60	8.60	8.60	8.60	8.60	8.50	8.70
Aroclor-1268-3	(3)	8.82	8.82	8.82	8.82	8.82	8.82	8.72	8.92
Aroclor-1268-4	(4)	9.22	9.22	9.22	9.22	9.21	9.22	9.12	9.32
Aroclor-1268-5	(5)	9.61	9.61	9.61	9.61	9.61	9.61	9.51	9.71

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

RETENTION TIMES OF INITIAL CALIBRATION

Decachlorobiphenyl	9.93	9.93	9.93	9.93	9.93	9.93	9.83	10.03
Tetrachloro-m-xylene	4.35	4.35	4.35	4.35	4.35	4.35	4.25	4.45

Raw Data

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

RETENTION TIMES OF INITIAL CALIBRATION

Lab Name: Alliance **Contract:** MATR02

Lab Code: ACE **SDG NO.:** Q3450

Instrument ID: ECD_O **Calibration Date(s):** 11/04/2025 11/04/2025

Calibration Times: 09:39 18:31

GC Column: ZB-MR2 ID: 0.32 (mm)

LAB FILE ID: RT 1000 = PO114871.D RT 750 = PO114872.D

RT 500 = PO114873.D RT 250 = PO114874.D RT 050 = PO114875.D

COMPOUND		RT 1000	RT 750	RT 500	RT 250	RT 050	MEAN RT	RT WINDOW FROM	RT WINDOW TO
Aroclor-1016-1	(1)	4.73	4.73	4.73	4.73	4.73	4.73	4.63	4.83
Aroclor-1016-2	(2)	4.75	4.75	4.75	4.75	4.75	4.75	4.65	4.85
Aroclor-1016-3	(3)	4.92	4.92	4.92	4.92	4.92	4.92	4.82	5.02
Aroclor-1016-4	(4)	4.96	4.96	4.96	4.96	4.96	4.96	4.86	5.06
Aroclor-1016-5	(5)	5.18	5.18	5.18	5.18	5.18	5.18	5.08	5.28
Aroclor-1260-1	(1)	6.21	6.21	6.21	6.21	6.21	6.21	6.11	6.31
Aroclor-1260-2	(2)	6.39	6.39	6.39	6.39	6.39	6.39	6.29	6.49
Aroclor-1260-3	(3)	6.55	6.55	6.55	6.55	6.55	6.55	6.45	6.65
Aroclor-1260-4	(4)	7.02	7.02	7.02	7.02	7.02	7.02	6.92	7.12
Aroclor-1260-5	(5)	7.26	7.26	7.26	7.26	7.26	7.26	7.16	7.36
Decachlorobiphenyl		8.64	8.64	8.64	8.64	8.64	8.64	8.54	8.74
Tetrachloro-m-xylene		3.65	3.65	3.65	3.65	3.65	3.65	3.55	3.75
Aroclor-1242-1	(1)	4.73	4.73	4.73	4.73	4.73	4.73	4.63	4.83
Aroclor-1242-2	(2)	4.75	4.75	4.75	4.75	4.75	4.75	4.65	4.85
Aroclor-1242-3	(3)	4.92	4.92	4.92	4.92	4.92	4.92	4.82	5.02
Aroclor-1242-4	(4)	5.01	5.01	5.01	5.01	5.01	5.01	4.91	5.11
Aroclor-1242-5	(5)	5.53	5.53	5.53	5.53	5.53	5.53	5.43	5.63
Decachlorobiphenyl		8.64	8.64	8.64	8.64	8.64	8.64	8.54	8.74
Tetrachloro-m-xylene		3.65	3.65	3.65	3.65	3.65	3.65	3.55	3.75
Aroclor-1248-1	(1)	4.73	4.73	4.73	4.73	4.73	4.73	4.63	4.83
Aroclor-1248-2	(2)	4.96	4.97	4.96	4.96	4.96	4.96	4.86	5.06
Aroclor-1248-3	(3)	5.01	5.01	5.01	5.01	5.01	5.01	4.91	5.11
Aroclor-1248-4	(4)	5.18	5.18	5.18	5.18	5.18	5.18	5.08	5.28
Aroclor-1248-5	(5)	5.57	5.57	5.57	5.57	5.57	5.57	5.47	5.67
Decachlorobiphenyl		8.64	8.64	8.64	8.64	8.64	8.64	8.54	8.74
Tetrachloro-m-xylene		3.65	3.65	3.65	3.65	3.65	3.65	3.55	3.75
Aroclor-1254-1	(1)	5.53	5.53	5.53	5.53	5.53	5.53	5.43	5.63
Aroclor-1254-2	(2)	5.68	5.68	5.67	5.67	5.67	5.67	5.57	5.77
Aroclor-1254-3	(3)	6.08	6.08	6.08	6.08	6.08	6.08	5.98	6.18
Aroclor-1254-4	(4)	6.30	6.30	6.30	6.30	6.30	6.30	6.20	6.40
Aroclor-1254-5	(5)	6.72	6.72	6.72	6.72	6.72	6.72	6.62	6.82
Decachlorobiphenyl		8.64	8.64	8.64	8.64	8.64	8.64	8.54	8.74
Tetrachloro-m-xylene		3.65	3.65	3.65	3.65	3.65	3.65	3.55	3.75
Aroclor-1268-1	(1)	7.54	7.54	7.54	7.54	7.54	7.54	7.44	7.64
Aroclor-1268-2	(2)	7.61	7.61	7.61	7.61	7.61	7.61	7.51	7.71
Aroclor-1268-3	(3)	7.81	7.81	7.81	7.81	7.81	7.81	7.71	7.91
Aroclor-1268-4	(4)	8.10	8.10	8.10	8.10	8.10	8.10	8.00	8.20
Aroclor-1268-5	(5)	8.39	8.39	8.39	8.39	8.39	8.39	8.29	8.49

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

RETENTION TIMES OF INITIAL CALIBRATION

Decachlorobiphenyl	8.64	8.64	8.64	8.64	8.64	8.64	8.54	8.74
Tetrachloro-m-xylene	3.65	3.65	3.65	3.65	3.65	3.65	3.55	3.75

Raw Data

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION FACTOR OF INITIAL CALIBRATION

Lab Name: Alliance **Contract:** MATR02

Lab Code: ACE **SDG NO.:** Q3450

Instrument ID: ECD_O **Calibration Date(s):** 11/04/2025 11/04/2025

Calibration Times: 09:39 18:31

GC Column: ZB-MR1 **ID:** 0.32 **(mm)**

LAB FILE ID: **CF 1000 =** PO114871.D **CF 750 =** PO114872.D

CF 500 = PO114873.D **CF 250 =** PO114874.D **CF 050 =** PO114875.D

COMPOUND		CF 1000	CF 750	CF 500	CF 250	CF 050	CF	% RSD
Aroclor-1016-1	(1)	235183609	242405119	252300612	252330724	247880040	246020021	3
Aroclor-1016-2	(2)	380671765	391174169	403435360	408228592	401844700	397070917	3
Aroclor-1016-3	(3)	214892716	220811199	230510330	231508016	217282560	223000964	3
Aroclor-1016-4	(4)	180735079	184273587	191842228	192742164	194782360	188875084	3
Aroclor-1016-5	(5)	166479276	173610481	179587174	180772736	190536000	178197133	5
Aroclor-1260-1	(1)	291080181	301659089	311914366	329562032	293951540	305633442	5
Aroclor-1260-2	(2)	393658847	402595176	414352186	418536840	350242960	395877202	7
Aroclor-1260-3	(3)	331096941	340266576	347857958	348147780	291720640	331817979	7
Aroclor-1260-4	(4)	303823322	310642864	321126472	321507392	268000240	305020058	7
Aroclor-1260-5	(5)	714393776	724690244	737226162	719415648	670367440	713218654	3
Decachlorobiphenyl		4451734610	4516548787	4612561720	4524869360	3995527600	4420248415	6
Tetrachloro-m-xylene		7020403280	7162783533	7307727940	7156075200	6543618600	7038121711	4
Aroclor-1242-1	(1)	180395803	184480117	189796832	192180524	184350480	186240751	2
Aroclor-1242-2	(2)	290607708	295647931	305830174	302003732	299791360	298776181	2
Aroclor-1242-3	(3)	164449629	167684461	173929682	174410736	176369820	171368866	3
Aroclor-1242-4	(4)	136844790	140633719	146402846	139660276	152376540	143183634	4
Aroclor-1242-5	(5)	138362947	141839337	145002188	138385972	141196180	140957325	2
Decachlorobiphenyl		4729850460	4804993427	4858375920	4746069120	4160138600	4659885505	6
Tetrachloro-m-xylene		7348004600	7445204373	7532777260	7262019280	6694579200	7256516943	5
Aroclor-1248-1	(1)	130126160	135244476	137909000	141742196	154206600	139845686	6
Aroclor-1248-2	(2)	182893435	187680161	194064692	196392780	211353600	194476934	5
Aroclor-1248-3	(3)	203368877	206306241	213983954	212927108	217587500	210834736	3
Aroclor-1248-4	(4)	240181487	246210361	255848644	255957264	264773260	252594203	4
Aroclor-1248-5	(5)	225312398	230738423	239469078	240115800	243213880	235769916	3
Decachlorobiphenyl		4590720290	4647354867	4744288880	4661202760	4211361200	4570985599	5
Tetrachloro-m-xylene		7316541360	7442934853	7528054520	7402608560	7218400400	7381707939	2
Aroclor-1254-1	(1)	240591852	247416493	258765478	257075152	263299240	253429643	3
Aroclor-1254-2	(2)	346021204	353529297	371589652	368897016	412184820	370444398	7
Aroclor-1254-3	(3)	372893964	380446979	397009938	384293176	389678940	384864599	2
Aroclor-1254-4	(4)	315539209	321424513	336840998	326904952	342871860	328716306	3
Aroclor-1254-5	(5)	295373361	300690525	312155076	300474784	300573260	301853401	2
Decachlorobiphenyl		4791769470	4896862187	5061968640	4831244600	4661831200	4848735219	3
Tetrachloro-m-xylene		7410585440	7526809333	7802184660	7481754040	7408041400	7525874975	2
Aroclor-1268-1	(1)	876279188	885717404	893041022	887529864	813508780	871215252	4

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION FACTOR OF INITIAL CALIBRATION

Aroclor-1268-2	(2)	748334854	756337347	765525536	759915712	702804360	746583562	3
Aroclor-1268-3	(3)	663299706	672085827	679866912	671661004	617886840	660960058	4
Aroclor-1268-4	(4)	274155738	277531193	282310020	282161716	286808980	280593529	2
Aroclor-1268-5	(5)	1966581606	1979440064	1985729368	1956798376	1807416340	1939193151	4
Decachlorobiphenyl		8288385350	8396434587	8504946880	8419357920	7722111200	8266247187	4
Tetrachloro-m-xylene		7475792880	7636683133	7655291620	7597134240	7121440400	7497268455	3

Raw Data

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION FACTOR OF INITIAL CALIBRATION

Lab Name: Alliance **Contract:** MATR02

Lab Code: ACE **SDG NO.:** Q3450

Instrument ID: ECD_O **Calibration Date(s):** 11/04/2025 11/04/2025

Calibration Times: 09:39 18:31

GC Column: ZB-MR2 **ID:** 0.32 (mm)

LAB FILE ID: **CF 1000 =** PO114871.D **CF 750 =** PO114872.D

CF 500 = PO114873.D **CF 250 =** PO114874.D **CF 050 =** PO114875.D

COMPOUND		CF 1000	CF 750	CF 500	CF 250	CF 050	CF	% RSD
Aroclor-1016-1	(1)	178795127	185535456	193106802	200932132	192907960	190255495	4
Aroclor-1016-2	(2)	260787154	267435697	277593168	286495324	262010880	270864445	4
Aroclor-1016-3	(3)	140310175	144533207	150201526	154962192	149858780	147973176	4
Aroclor-1016-4	(4)	108607533	112021955	117721682	122827936	125909940	117417809	6
Aroclor-1016-5	(5)	142235561	147493437	153611756	157990124	162236540	152713484	5
Aroclor-1260-1	(1)	272865489	283696256	302549666	306994580	314431380	296107474	5
Aroclor-1260-2	(2)	410849028	420744668	440173148	441650472	436738740	430031211	3
Aroclor-1260-3	(3)	328919579	344471297	362206406	366201912	353957800	351151399	4
Aroclor-1260-4	(4)	295043629	302500340	313130718	316509244	287959420	303028670	4
Aroclor-1260-5	(5)	819932115	830411719	851082708	846532168	760181660	821628074	4
Decachlorobiphenyl		3768213910	3862452533	4051635320	3997632280	3839185200	3903823849	3
Tetrachloro-m-xylene		4444199150	4532189360	4610968540	4573862320	4172678600	4466779594	4
Aroclor-1242-1	(1)	138347633	144438088	149108388	148877560	150336280	146221590	3
Aroclor-1242-2	(2)	199908386	204961789	213926082	211880332	205062860	207147890	3
Aroclor-1242-3	(3)	108216818	112229521	116418372	113485296	109537600	111977521	3
Aroclor-1242-4	(4)	101919263	105301932	110004430	109236748	109193360	107131147	3
Aroclor-1242-5	(5)	135980897	140103160	145653410	150301700	142549620	142917757	4
Decachlorobiphenyl		4093300380	4301738307	4403735460	4338775160	4380898600	4303689581	3
Tetrachloro-m-xylene		4800115840	4683347560	4961464680	4527345840	4437059000	4681866584	4
Aroclor-1248-1	(1)	101027399	105625799	111262386	113093188	115539800	109309714	5
Aroclor-1248-2	(2)	136540550	142594189	148530518	152692968	164144180	148900481	7
Aroclor-1248-3	(3)	143887712	149887199	155474250	158349236	169440400	155407759	6
Aroclor-1248-4	(4)	171266913	177666692	184171754	187513908	195642380	183252329	5
Aroclor-1248-5	(5)	174649655	180220384	187038042	189642400	190092660	184328628	3
Decachlorobiphenyl		4076042330	4166088587	4290124020	4403270520	4144788800	4216062851	3
Tetrachloro-m-xylene		4758509080	4829545453	4892334280	4811209560	4727034000	4803726475	1
Aroclor-1254-1	(1)	275439068	284238351	300243386	299923620	341055960	300180077	8
Aroclor-1254-2	(2)	245333366	253279232	267984098	268318040	313122940	269607535	9
Aroclor-1254-3	(3)	398238109	410614839	430545692	426530308	445294740	422244738	4
Aroclor-1254-4	(4)	284785979	294895399	305027804	301955520	300306220	297394184	3
Aroclor-1254-5	(5)	385708423	392326712	409498118	398089984	402006780	397526003	2
Decachlorobiphenyl		4273373360	4367226933	4604811080	4533103920	4272134600	4410129979	3
Tetrachloro-m-xylene		4791085320	4887004773	5044010700	4876428000	4843036000	4888312959	2
Aroclor-1268-1	(1)	948830555	965398547	981732904	993582148	991082540	976125339	2

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION FACTOR OF INITIAL CALIBRATION

Aroclor-1268-2	(2)	784312077	796694211	811227528	817006596	768929160	795633914	2
Aroclor-1268-3	(3)	672635507	681509284	695723330	696225136	632763280	675771307	4
Aroclor-1268-4	(4)	270862616	275402580	280086838	282372280	267609400	275266743	2
Aroclor-1268-5	(5)	1768544208	1786014605	1818457210	1831952004	1717683400	1784530285	2
Decachlorobiphenyl		7349980390	7437822467	7706201340	7861559120	8321979800	7735508623	5
Tetrachloro-m-xylene		4791521570	4902547960	4915920660	4883273320	4724156000	4843483902	2

Raw Data

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Lab Name: Alliance **Contract:** MATR02

Lab Code: ACE **SDG NO.:** Q3450

Instrument ID: ECD_O **Date(s) Analyzed:** 11/04/2025 11/04/2025

GC Column: ZB-MR1 **ID:** 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW FROM	RT WINDOW TO	CALIBRATION FACTOR
Aroclor-1221	500	1	4.55	4.45	4.65	84124000
		2	4.63	4.53	4.73	60183600
		3	4.71	4.61	4.81	193335000
		4	0.00			0
		5	0.00			0
Aroclor-1232	500	1	4.71	4.61	4.81	150015000
		2	5.22	5.12	5.32	72567400
		3	5.51	5.41	5.61	158675000
		4	5.67	5.57	5.77	73795600
		5	5.76	5.66	5.86	51580600
Aroclor-1262	500	1	7.70	7.60	7.80	453404000
		2	8.23	8.13	8.33	707198000
		3	8.53	8.43	8.63	451420000
		4	8.62	8.52	8.72	336186000
		5	9.24	9.14	9.34	241816000

Raw Data

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Lab Name: Alliance **Contract:** MATR02

Lab Code: ACE **SDG NO.:** Q3450

Instrument ID: ECD_O **Date(s) Analyzed:** 11/04/2025 11/04/2025

GC Column: ZB-MR2 **ID:** 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW FROM	RT WINDOW TO	CALIBRATION FACTOR
Aroclor-1221	500	1	3.86	3.76	3.96	61799200
		2	3.95	3.85	4.05	45637000
		3	4.02	3.92	4.12	140257000
		4	0.00			0
		5	0.00			0
Aroclor-1232	500	1	4.02	3.92	4.12	111493000
		2	4.75	4.65	4.85	112095000
		3	4.92	4.82	5.02	59654200
		4	5.01	4.91	5.11	50889000
		5	5.18	5.08	5.28	56587600
Aroclor-1262	500	1	6.76	6.66	6.86	494206000
		2	7.26	7.16	7.36	795090000
		3	7.54	7.44	7.64	329028000
		4	7.61	7.51	7.71	538192000
		5	8.10	8.00	8.20	238020000

Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
Data File : PO114871.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Nov 2025 09:39
Operator : YP/AJ
Sample : AR1660ICC1000
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1660ICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 04 10:39:30 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Nov 04 10:34:49 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.348	3.649	702.0E6	444.4E6	97.995	98.158
2)	SA Decachlor...	9.931	8.642	445.2E6	376.8E6	98.226	96.376
	Target Compounds						
3)	L1 AR-1016-1	5.487	4.728	235.2E6	178.8E6	964.887	961.518
4)	L1 AR-1016-2	5.509	4.747	380.7E6	260.8E6	970.969	968.784
5)	L1 AR-1016-3	5.571	4.922	214.9E6	140.3E6	964.936	965.952
6)	L1 AR-1016-4	5.666	4.964	180.7E6	108.6E6	970.188	959.731
7)	L1 AR-1016-5	5.958	5.176	166.5E6	142.2E6	962.123	961.547
31)	L7 AR-1260-1	7.072	6.207	291.1E6	272.9E6	965.449	948.413
32)	L7 AR-1260-2	7.325	6.394	393.7E6	410.8E6	974.390	965.542
33)	L7 AR-1260-3	7.683	6.547	331.1E6	328.9E6	975.314	951.837
34)	L7 AR-1260-4	7.904	7.018	303.8E6	295.0E6	972.313	970.260
35)	L7 AR-1260-5	8.212	7.259	714.4E6	819.9E6	984.271	981.358

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


ICAL Form


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
Data File : PO114871.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Nov 2025 09:39
Operator : YP/AJ
Sample : AR1660ICC1000
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1660ICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 04 10:39:30 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Nov 04 10:34:49 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
Data File : PO114872.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Nov 2025 09:58
Operator : YP/AJ
Sample : AR1660ICC750
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1660ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 04 10:42:15 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Nov 04 10:34:49 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.348	3.649	537.2E6	339.9E6	74.991	75.051
2)	SA Decachlor...	9.931	8.642	338.7E6	289.7E6	74.828	74.390
	Target Compounds						
3)	L1 AR-1016-1	5.486	4.728	181.8E6	139.2E6	747.252	748.882
4)	L1 AR-1016-2	5.508	4.747	293.4E6	200.6E6	748.878	746.734
5)	L1 AR-1016-3	5.571	4.921	165.6E6	108.4E6	745.744	747.508
6)	L1 AR-1016-4	5.666	4.964	138.2E6	84016466	744.572	744.934
7)	L1 AR-1016-5	5.958	5.176	130.2E6	110.6E6	751.666	748.544
31)	L7 AR-1260-1	7.071	6.206	226.2E6	212.8E6	750.268	742.996
32)	L7 AR-1260-2	7.326	6.394	301.9E6	315.6E6	748.253	744.378
33)	L7 AR-1260-3	7.683	6.546	255.2E6	258.4E6	751.161	748.419
34)	L7 AR-1260-4	7.904	7.017	233.0E6	226.9E6	747.063	747.386
35)	L7 AR-1260-5	8.211	7.258	543.5E6	622.8E6	749.228	746.944


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
Data File : PO114872.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Nov 2025 09:58
Operator : YP/AJ
Sample : AR1660ICC750
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1660ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 04 10:42:15 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Nov 04 10:34:49 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
Data File : PO114873.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Nov 2025 10:16
Operator : YP/AJ
Sample : AR1660ICC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1660ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 04 10:36:23 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Nov 04 10:34:49 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.348	3.648	365.4E6	230.5E6	50.000	50.000
2)	SA Decachlor...	9.930	8.641	230.6E6	202.6E6	50.000	50.000
	Target Compounds						
3)	L1 AR-1016-1	5.487	4.728	126.2E6	96553401	500.000	500.000
4)	L1 AR-1016-2	5.508	4.747	201.7E6	138.8E6	500.000	500.000
5)	L1 AR-1016-3	5.570	4.921	115.3E6	75100763	500.000	500.000
6)	L1 AR-1016-4	5.666	4.964	95921114	58860841	500.000	500.000
7)	L1 AR-1016-5	5.958	5.176	89793587	76805878	500.000	500.000
31)	L7 AR-1260-1	7.071	6.206	156.0E6	151.3E6	500.000	500.000
32)	L7 AR-1260-2	7.325	6.394	207.2E6	220.1E6	500.000	500.000
33)	L7 AR-1260-3	7.682	6.546	173.9E6	181.1E6	500.000	500.000
34)	L7 AR-1260-4	7.903	7.017	160.6E6	156.6E6	500.000	500.000
35)	L7 AR-1260-5	8.210	7.258	368.6E6	425.5E6	500.000	500.000

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


ICAL Form


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
Data File : PO114873.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Nov 2025 10:16
Operator : YP/AJ
Sample : AR1660ICC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1660ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 04 10:36:23 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Nov 04 10:34:49 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
Data File : PO114874.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Nov 2025 10:34
Operator : YP/AJ
Sample : AR1660ICC250
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1660ICC250

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 11/05/2025
Supervised By :Yogesh Patel 11/05/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 04 10:51:23 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Nov 04 10:51:10 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.348	3.649	178.9E6	114.3E6	24.980	25.185
2)	SA Decachlor...	9.931	8.641	113.1E6	99940807	24.991	25.495
	Target Compounds						
3)	L1 AR-1016-1	5.487	4.728	63082681	50233033	256.898	264.953
4)	L1 AR-1016-2	5.508	4.747	102.1E6	71623831	257.800	262.284
5)	L1 AR-1016-3	5.571	4.922	57877004	38740548	257.884	262.645
6)	L1 AR-1016-4	5.666	4.964	48185541	30706984	257.129	266.335
7)	L1 AR-1016-5	5.958	5.176	45193184	39497531	258.081	262.734
31)	L7 AR-1260-1	7.071	6.206	82390508	76748645	267.021	260.958m
32)	L7 AR-1260-2	7.325	6.393	104.6E6	110.4E6	256.906	256.862m
33)	L7 AR-1260-3	7.682	6.546	87036945	91550478	254.611	261.237
34)	L7 AR-1260-4	7.904	7.017	80376848	79127311	255.753	257.915
35)	L7 AR-1260-5	8.211	7.258	179.9E6	211.6E6	248.441	252.850

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
Data File : PO114874.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Nov 2025 10:34
Operator : YP/AJ
Sample : AR1660ICC250
Misc :
ALS Vial : 6 Sample Multiplier: 1


Instrument :
ECD_O
ClientSampleId :
AR1660ICC250


Manual Integrations APPROVED

Reviewed By :Rahul Chavli 11/05/2025
Supervised By :Yogesh Patel 11/05/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 04 10:51:23 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Nov 04 10:51:10 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
Data File : PO114875.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Nov 2025 10:53
Operator : YP/AJ
Sample : AR1660ICC050
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1660ICC050

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 11/05/2025
Supervised By :Yogesh Patel 11/05/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 04 11:09:54 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Nov 04 11:09:42 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.348	3.649	32718093	20863393	4.649	4.671
2)	SA Decachlor...	9.933	8.641	19977638	19195926	4.520	4.917
	Target Compounds						
3)	L1 AR-1016-1	5.487	4.727	12394002	9645398	50.378	50.697
4)	L1 AR-1016-2	5.509	4.747	20092235	13100544	50.601	48.366
5)	L1 AR-1016-3	5.572	4.921	10864128	7492939	48.718	51.057m
6)	L1 AR-1016-4	5.666	4.964	9739118	6295497	51.564	53.616
7)	L1 AR-1016-5	5.959	5.175	9526800	8111827	53.462	53.507m
31)	L7 AR-1260-1	7.071	6.206	14697577	15721569	48.664m	53.302m
32)	L7 AR-1260-2	7.325	6.392	17512148	21836937	45.034m	51.096m
33)	L7 AR-1260-3	7.682	6.546	14586032	17697890	44.584m	50.400
34)	L7 AR-1260-4	7.904	7.017	13400012	14397971	44.028m	47.514
35)	L7 AR-1260-5	8.211	7.258	33518372	38009083	46.996	46.261

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
Data File : PO114875.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Nov 2025 10:53
Operator : YP/AJ
Sample : AR1660ICC050
Misc :
ALS Vial : 7 Sample Multiplier: 1


Instrument :
ECD_O
ClientSampleId :
AR1660ICC050


Manual Integrations APPROVED

Reviewed By :Rahul Chavli 11/05/2025
Supervised By :Yogesh Patel 11/05/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 04 11:09:54 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Nov 04 11:09:42 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
Data File : PO114876.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Nov 2025 11:11
Operator : YP/AJ
Sample : AR1221ICC500
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1221ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 04 11:57:28 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Nov 04 11:57:11 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.348	3.649	367.5E6	224.3E6	50.000	50.000
2)	SA Decachlor...	9.932	8.641	223.4E6	199.5E6	50.000	50.000
	Target Compounds						
8)	L2 AR-1221-1	4.549	3.860	42062015	30899570	500.000	500.000
9)	L2 AR-1221-2	4.632	3.945	30091751	22818522	500.000	500.000
10)	L2 AR-1221-3	4.706	4.021	96667420	70128648	500.000	500.000


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
Data File : PO114876.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Nov 2025 11:11
Operator : YP/AJ
Sample : AR1221ICC500
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1221ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 04 11:57:28 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Nov 04 11:57:11 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
Data File : PO114877.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Nov 2025 11:30
Operator : YP/AJ
Sample : AR1232ICC500
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1232ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 04 12:15:50 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Nov 04 11:57:11 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.347	3.649	373.4E6	232.5E6	50.000	50.000
2)	SA Decachlor...	9.931	8.641	232.4E6	203.6E6	50.000	50.000
	Target Compounds						
11)	L3 AR-1232-1	4.706	4.020	75007315	55746703	500.000	500.000
12)	L3 AR-1232-2	5.224	4.747	36283683	56047490	500.000	500.000
13)	L3 AR-1232-3	5.508	4.921	79337557	29827092	500.000	500.000
14)	L3 AR-1232-4	5.666	5.005	36897763	25444495	500.000	500.000
15)	L3 AR-1232-5	5.757	5.176	25790349	28293807	500.000	500.000

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


ICAL Form


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
Data File : PO114877.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Nov 2025 11:30
Operator : YP/AJ
Sample : AR1232ICC500
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1232ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 04 12:15:50 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Nov 04 11:57:11 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
Data File : PO114878.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Nov 2025 11:48
Operator : YP/AJ
Sample : AR1242ICC1000
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1242ICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 04 12:52:27 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Nov 04 12:48:08 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.348	3.649	734.8E6	480.0E6	98.758	98.347
2)	SA Decachlor...	9.932	8.642	473.0E6	409.3E6	98.660	96.347
	Target Compounds						
16)	L4 AR-1242-1	5.487	4.728	180.4E6	138.3E6	974.605	962.566
17)	L4 AR-1242-2	5.509	4.747	290.6E6	199.9E6	974.478	966.127
18)	L4 AR-1242-3	5.570	4.921	164.4E6	108.2E6	971.984	963.489
19)	L4 AR-1242-4	5.667	5.005	136.8E6	101.9E6	966.255	961.849
20)	L4 AR-1242-5	6.392	5.526	138.4E6	136.0E6	976.570	965.656


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
Data File : PO114878.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Nov 2025 11:48
Operator : YP/AJ
Sample : AR1242ICC1000
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1242ICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 04 12:52:27 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Nov 04 12:48:08 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
Data File : PO114879.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Nov 2025 12:06
Operator : YP/AJ
Sample : AR1242ICC750
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1242ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 04 12:55:14 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Nov 04 12:48:08 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.348	3.649	558.4E6	351.3E6	75.032	72.950
2)	SA Decachlor...	9.930	8.641	360.4E6	322.6E6	75.113	75.624
	Target Compounds						
16)	L4 AR-1242-1	5.487	4.728	138.4E6	108.3E6	748.334	752.466
17)	L4 AR-1242-2	5.508	4.747	221.7E6	153.7E6	745.677	745.260
18)	L4 AR-1242-3	5.570	4.921	125.8E6	84172141	745.539	749.608
19)	L4 AR-1242-4	5.666	5.006	105.5E6	78976449	746.496	746.880
20)	L4 AR-1242-5	6.391	5.525	106.4E6	105.1E6	750.553	747.461

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


ICAL Form


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
Data File : PO114879.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Nov 2025 12:06
Operator : YP/AJ
Sample : AR1242ICC750
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1242ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 04 12:55:14 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Nov 04 12:48:08 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
Data File : PO114880.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Nov 2025 12:25
Operator : YP/AJ
Sample : AR1242ICC500
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1242ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 04 12:49:50 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Nov 04 12:48:08 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.348	3.649	376.6E6	248.1E6	50.000	50.000
2)	SA Decachlor...	9.932	8.642	242.9E6	220.2E6	50.000	50.000
	Target Compounds						
16)	L4 AR-1242-1	5.488	4.728	94898416	74554194	500.000	500.000
17)	L4 AR-1242-2	5.509	4.747	152.9E6	107.0E6	500.000	500.000
18)	L4 AR-1242-3	5.571	4.922	86964841	58209186	500.000	500.000
19)	L4 AR-1242-4	5.667	5.006	73201423	55002215	500.000	500.000
20)	L4 AR-1242-5	6.393	5.526	72501094	72826705	500.000	500.000


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
Data File : PO114880.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Nov 2025 12:25
Operator : YP/AJ
Sample : AR1242ICC500
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1242ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 04 12:49:50 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Nov 04 12:48:08 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
Data File : PO114881.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Nov 2025 12:42
Operator : YP/AJ
Sample : AR1242ICC250
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1242ICC250

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 04 12:59:45 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Nov 04 12:48:08 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.347	3.649	181.6E6	113.2E6	24.544	23.863
2)	SA Decachlor...	9.930	8.642	118.7E6	108.5E6	24.798	25.317
	Target Compounds						
16)	L4 AR-1242-1	5.485	4.728	48045131	37219390	257.320	256.344
17)	L4 AR-1242-2	5.507	4.747	75500933	52970083	252.915	255.070
18)	L4 AR-1242-3	5.569	4.921	43602684	28371324	256.308	251.994
19)	L4 AR-1242-4	5.665	5.006	34915069	27309187	247.826	256.146
20)	L4 AR-1242-5	6.391	5.526	34596493	37575425	245.544	262.747

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


ICAL Form


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
Data File : PO114881.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Nov 2025 12:42
Operator : YP/AJ
Sample : AR1242ICC250
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1242ICC250

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 04 12:59:45 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Nov 04 12:48:08 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
Data File : PO114882.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Nov 2025 13:00
Operator : YP/AJ
Sample : AR1242ICC050
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1242ICC050

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 11/05/2025
Supervised By :Yogesh Patel 11/05/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 04 13:15:43 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Nov 04 13:15:32 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.349	3.649	33472896	22185295	4.613	4.739
2)	SA Decachlor...	9.932	8.641	20800693	21904493	4.464	5.090
	Target Compounds						
16)	L4 AR-1242-1	5.487	4.729	9217524	7516814	49.493	51.407
17)	L4 AR-1242-2	5.509	4.747	14989568	10253143	50.170	49.497
18)	L4 AR-1242-3	5.571	4.922	8818491	5476880	51.459	48.911
19)	L4 AR-1242-4	5.667	5.007	7618827	5459668	53.210	50.962
20)	L4 AR-1242-5	6.392	5.526	7059809	7127481	50.085	49.823m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
Data File : PO114882.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Nov 2025 13:00
Operator : YP/AJ
Sample : AR1242ICC050
Misc :
ALS Vial : 14 Sample Multiplier: 1


Instrument :
ECD_O
ClientSampleId :
AR1242ICC050


Manual Integrations APPROVED

Reviewed By :Rahul Chavli 11/05/2025
Supervised By :Yogesh Patel 11/05/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 04 13:15:43 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Nov 04 13:15:32 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
Data File : PO114883.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Nov 2025 13:19
Operator : YP/AJ
Sample : AR1248ICC1000
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1248ICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 04 14:15:48 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Nov 04 14:11:27 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.348	3.648	731.7E6	475.9E6	98.575	98.613
2)	SA Decachlor...	9.931	8.641	459.1E6	407.6E6	98.355	97.441
	Target Compounds						
21)	L5 AR-1248-1	5.486	4.727	130.1E6	101.0E6	970.963	951.788
22)	L5 AR-1248-2	5.757	4.963	182.9E6	136.5E6	970.365	957.940
23)	L5 AR-1248-3	5.957	5.005	203.4E6	143.9E6	974.566	961.296
24)	L5 AR-1248-4	6.354	5.175	240.2E6	171.3E6	968.415	963.693
25)	L5 AR-1248-5	6.391	5.567	225.3E6	174.6E6	969.541	965.748

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


ICAL Form


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
Data File : PO114883.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Nov 2025 13:19
Operator : YP/AJ
Sample : AR1248ICC1000
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1248ICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 04 14:15:48 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Nov 04 14:11:27 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
Data File : PO114884.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Nov 2025 13:37
Operator : YP/AJ
Sample : AR1248ICC750
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1248ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 04 14:18:24 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Nov 04 14:11:27 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.349	3.650	558.2E6	362.2E6	75.139	75.043
2)	SA Decachlor...	9.934	8.642	348.6E6	312.5E6	74.784	74.797
	Target Compounds						
21)	L5 AR-1248-1	5.487	4.728	101.4E6	79219349	754.563	747.551
22)	L5 AR-1248-2	5.759	4.965	140.8E6	106.9E6	747.878	750.206
23)	L5 AR-1248-3	5.958	5.007	154.7E6	112.4E6	744.299	750.689
24)	L5 AR-1248-4	6.355	5.177	184.7E6	133.3E6	746.353	749.852
25)	L5 AR-1248-5	6.393	5.568	173.1E6	135.2E6	746.437	748.274

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


ICAL Form


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
Data File : PO114884.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Nov 2025 13:37
Operator : YP/AJ
Sample : AR1248ICC750
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1248ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 04 14:18:24 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Nov 04 14:11:27 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
Data File : PO114885.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Nov 2025 13:56
Operator : YP/AJ
Sample : AR1248ICC500
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1248ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 04 14:12:50 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Nov 04 14:11:27 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.349	3.650	376.4E6	244.6E6	50.000	50.000
2)	SA Decachlor...	9.932	8.642	237.2E6	214.5E6	50.000	50.000
	Target Compounds						
21)	L5 AR-1248-1	5.487	4.728	68954500	55631193	500.000	500.000
22)	L5 AR-1248-2	5.759	4.964	97032346	74265259	500.000	500.000
23)	L5 AR-1248-3	5.959	5.006	107.0E6	77737125	500.000	500.000
24)	L5 AR-1248-4	6.354	5.177	127.9E6	92085877	500.000	500.000
25)	L5 AR-1248-5	6.393	5.568	119.7E6	93519021	500.000	500.000

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


ICAL Form


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
Data File : PO114885.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Nov 2025 13:56
Operator : YP/AJ
Sample : AR1248ICC500
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1248ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 04 14:12:50 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Nov 04 14:11:27 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
Data File : PO114886.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Nov 2025 14:14
Operator : YP/AJ
Sample : AR1248ICC250
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1248ICC250

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 04 14:29:11 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Nov 04 14:11:27 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.349	3.649	185.1E6	120.3E6	24.933	24.939
2)	SA Decachlor...	9.931	8.642	116.5E6	110.1E6	25.002	26.000
	Target Compounds						
21)	L5 AR-1248-1	5.487	4.728	35435549	28273297	260.067	262.392
22)	L5 AR-1248-2	5.758	4.964	49098195	38173242	258.061	263.101
23)	L5 AR-1248-3	5.958	5.006	53231777	39587309	254.519	260.615
24)	L5 AR-1248-4	6.354	5.176	63989316	46878477	256.419	260.212
25)	L5 AR-1248-5	6.393	5.567	60028950	47410600	256.634	259.234


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
Data File : PO114886.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Nov 2025 14:14
Operator : YP/AJ
Sample : AR1248ICC250
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1248ICC250

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 04 14:29:11 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Nov 04 14:11:27 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
Data File : PO114887.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Nov 2025 14:33
Operator : YP/AJ
Sample : AR1248ICC050
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1248ICC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 04 14:49:08 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Nov 04 14:48:57 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.348	3.648	36092002	23635170	4.889	4.920
2)	SA Decachlor...	9.931	8.642	21056806	20723944	4.607	4.915
	Target Compounds						
21)	L5 AR-1248-1	5.487	4.727	7710330	5776990	55.135	52.850
22)	L5 AR-1248-2	5.758	4.963	10567680	8207209	54.339	55.119
23)	L5 AR-1248-3	5.958	5.005	10879375	8472020	51.601	54.515
24)	L5 AR-1248-4	6.355	5.176	13238663	9782119	52.411	53.381
25)	L5 AR-1248-5	6.393	5.567	12160694	9504633	51.579	51.564

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


ICAL Form


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
Data File : PO114887.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Nov 2025 14:33
Operator : YP/AJ
Sample : AR1248ICC050
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1248ICC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 04 14:49:08 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Nov 04 14:48:57 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
Data File : PO114888.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Nov 2025 14:51
Operator : YP/AJ
Sample : AR1254ICC1000
Misc :
ALS Vial : 20 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1254ICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 05 05:36:33 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Nov 05 05:33:43 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.349	3.650	741.1E6	479.1E6	94.981	94.986
2)	SA Decachlor...	9.932	8.642	479.2E6	427.3E6	94.662	92.802
	Target Compounds						
26)	L6 AR-1254-1	6.331	5.527	240.6E6	275.4E6	929.768	917.386
27)	L6 AR-1254-2	6.547	5.675	346.0E6	245.3E6	931.192	915.477
28)	L6 AR-1254-3	6.907	6.076	372.9E6	398.2E6	939.256	924.961
29)	L6 AR-1254-4	7.188	6.304	315.5E6	284.8E6	936.760	933.639
30)	L6 AR-1254-5	7.604	6.721	295.4E6	385.7E6	946.239	941.905

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


ICAL Form


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
Data File : PO114888.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Nov 2025 14:51
Operator : YP/AJ
Sample : AR1254ICC1000
Misc :
ALS Vial : 20 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1254ICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 05 05:36:33 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Nov 05 05:33:43 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
Data File : PO114889.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Nov 2025 15:09
Operator : YP/AJ
Sample : AR1254ICC750
Misc :
ALS Vial : 21 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1254ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 05 05:36:43 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Nov 05 05:33:43 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.349	3.650	564.5E6	366.5E6	72.353	72.665
2)	SA Decachlor...	9.933	8.642	367.3E6	327.5E6	72.554	71.130
	Target Compounds						
26)	L6 AR-1254-1	6.331	5.527	185.6E6	213.2E6	717.106	710.020
27)	L6 AR-1254-2	6.547	5.675	265.1E6	190.0E6	713.548	708.846
28)	L6 AR-1254-3	6.907	6.077	285.3E6	308.0E6	718.711	715.281
29)	L6 AR-1254-4	7.189	6.304	241.1E6	221.2E6	715.674	725.087
30)	L6 AR-1254-5	7.604	6.721	225.5E6	294.2E6	722.455	718.550

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


ICAL Form


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
Data File : PO114889.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Nov 2025 15:09
Operator : YP/AJ
Sample : AR1254ICC750
Misc :
ALS Vial : 21 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1254ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 05 05:36:43 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Nov 05 05:33:43 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
Data File : PO114890.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Nov 2025 15:28
Operator : YP/AJ
Sample : AR1254ICC500
Misc :
ALS Vial : 22 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1254ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 05 05:36:51 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Nov 05 05:33:43 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.349	3.649	390.1E6	252.2E6	50.000	50.000
2)	SA Decachlor...	9.932	8.642	253.1E6	230.2E6	50.000	50.000
	Target Compounds						
26)	L6 AR-1254-1	6.331	5.527	129.4E6	150.1E6	500.000	500.000
27)	L6 AR-1254-2	6.547	5.674	185.8E6	134.0E6	500.000	500.000
28)	L6 AR-1254-3	6.907	6.076	198.5E6	215.3E6	500.000	500.000
29)	L6 AR-1254-4	7.189	6.304	168.4E6	152.5E6	500.000	500.000
30)	L6 AR-1254-5	7.603	6.720	156.1E6	204.7E6	500.000	500.000

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


ICAL Form


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
Data File : PO114890.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Nov 2025 15:28
Operator : YP/AJ
Sample : AR1254ICC500
Misc :
ALS Vial : 22 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1254ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 05 05:36:51 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Nov 05 05:33:43 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
Data File : PO114891.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Nov 2025 16:04
Operator : YP/AJ
Sample : AR1254ICC250
Misc :
ALS Vial : 23 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1254ICC250

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 05 05:36:59 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Nov 05 05:33:43 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.349	3.649	187.0E6	121.9E6	23.973	24.169
2)	SA Decachlor...	9.932	8.642	120.8E6	113.3E6	23.861	24.611
	Target Compounds						
26)	L6 AR-1254-1	6.331	5.527	64268788	74980905	248.367	249.734
27)	L6 AR-1254-2	6.548	5.674	92224254	67079510	248.188	250.312
28)	L6 AR-1254-3	6.907	6.076	96073294	106.6E6	241.992	247.668
29)	L6 AR-1254-4	7.189	6.304	81726238	75488880	242.626	247.482
30)	L6 AR-1254-5	7.604	6.720	75118696	99522496	240.645	243.035

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


ICAL Form


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
Data File : PO114891.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Nov 2025 16:04
Operator : YP/AJ
Sample : AR1254ICC250
Misc :
ALS Vial : 23 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1254ICC250

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 05 05:36:59 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Nov 05 05:33:43 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
Data File : PO114892.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Nov 2025 16:23
Operator : YP/AJ
Sample : AR1254ICC050
Misc :
ALS Vial : 24 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1254ICC050

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 11/05/2025
Supervised By :Yogesh Patel 11/05/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 05 05:37:07 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Nov 05 05:33:43 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.348	3.648	37040207	24215180	4.747	4.801
2)	SA Decachlor...	9.931	8.642	23309156	21360673	4.605m	4.639
	Target Compounds						
26)	L6 AR-1254-1	6.330	5.526	13164962	17052798	50.876m	56.797
27)	L6 AR-1254-2	6.547	5.674	20609241	15656147	55.462m	58.422
28)	L6 AR-1254-3	6.907	6.076	19483947	22264737	49.077	51.713
29)	L6 AR-1254-4	7.188	6.303	17143593	15015311	50.895m	49.226
30)	L6 AR-1254-5	7.604	6.720	15028663	20100339	48.145	49.085

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
Data File : PO114892.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Nov 2025 16:23
Operator : YP/AJ
Sample : AR1254ICC050
Misc :
ALS Vial : 24 Sample Multiplier: 1


Instrument :
ECD_O
ClientSampleId :
AR1254ICC050


Manual Integrations APPROVED

Reviewed By :Rahul Chavli 11/05/2025
Supervised By :Yogesh Patel 11/05/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 05 05:37:07 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Nov 05 05:33:43 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
Data File : PO114893.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Nov 2025 17:00
Operator : YP/AJ
Sample : AR1262ICC500
Misc :
ALS Vial : 25 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1262ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 05 01:08:29 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Nov 05 01:07:49 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.368	3.649	362.5E6	234.2E6	50.000	50.000
2)	SA Decachlor...	9.952	8.641	223.0E6	207.7E6	50.000	50.000
	Target Compounds						
36)	L8 AR-1262-1	7.703	6.759	226.7E6	247.1E6	500.000	500.000
37)	L8 AR-1262-2	8.231	7.258	353.6E6	397.5E6	500.000	500.000
38)	L8 AR-1262-3	8.534	7.542	225.7E6	164.5E6	500.000	500.000
39)	L8 AR-1262-4	8.616	7.605	168.1E6	269.1E6	500.000	500.000
40)	L8 AR-1262-5	9.236	8.097	120.9E6	119.0E6	500.000	500.000


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
Data File : PO114893.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Nov 2025 17:00
Operator : YP/AJ
Sample : AR1262ICC500
Misc :
ALS Vial : 25 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1262ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 05 01:08:29 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Nov 05 01:07:49 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
Data File : PO114894.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Nov 2025 17:18
Operator : YP/AJ
Sample : AR1268ICC1000
Misc :
ALS Vial : 26 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1268ICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 05 01:23:13 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Nov 05 01:21:38 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.349	3.649	747.6E6	479.2E6	97.655	97.469
2)	SA Decachlor...	9.931	8.642	828.8E6	735.0E6	97.454	95.377
	Target Compounds						
41)	L9 AR-1268-1	8.511	7.542	876.3E6	948.8E6	981.231	966.485
42)	L9 AR-1268-2	8.599	7.606	748.3E6	784.3E6	977.544	966.821
43)	L9 AR-1268-3	8.820	7.814	663.3E6	672.6E6	975.632	966.815
44)	L9 AR-1268-4	9.216	8.097	274.2E6	270.9E6	971.116	967.067
45)	L9 AR-1268-5	9.609	8.387	1966.6E6	1768.5E6	990.357	972.552

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


ICAL Form


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
Data File : PO114894.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Nov 2025 17:18
Operator : YP/AJ
Sample : AR1268ICC1000
Misc :
ALS Vial : 26 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1268ICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 05 01:23:13 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Nov 05 01:21:38 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
Data File : PO114895.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Nov 2025 17:36
Operator : YP/AJ
Sample : AR1268ICC750
Misc :
ALS Vial : 27 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1268ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 05 01:23:21 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Nov 05 01:21:38 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.348	3.649	572.8E6	367.7E6	74.818	74.796
2)	SA Decachlor...	9.932	8.641	629.7E6	557.8E6	74.043	72.388
	Target Compounds						
41)	L9 AR-1268-1	8.510	7.542	664.3E6	724.0E6	743.849	737.521
42)	L9 AR-1268-2	8.599	7.606	567.3E6	597.5E6	740.998	736.564
43)	L9 AR-1268-3	8.819	7.814	504.1E6	511.1E6	741.416	734.677
44)	L9 AR-1268-4	9.215	8.096	208.1E6	206.6E6	737.304	737.457
45)	L9 AR-1268-5	9.609	8.387	1484.6E6	1339.5E6	747.625	736.619


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
Data File : PO114895.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Nov 2025 17:36
Operator : YP/AJ
Sample : AR1268ICC750
Misc :
ALS Vial : 27 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1268ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 05 01:23:21 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Nov 05 01:21:38 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
Data File : PO114896.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Nov 2025 17:55
Operator : YP/AJ
Sample : AR1268ICC500
Misc :
ALS Vial : 28 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1268ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 05 01:23:29 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Nov 05 01:21:38 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.348	3.649	382.8E6	245.8E6	50.000	50.000
2)	SA Decachlor...	9.931	8.641	425.2E6	385.3E6	50.000	50.000
	Target Compounds						
41)	L9 AR-1268-1	8.510	7.542	446.5E6	490.9E6	500.000	500.000
42)	L9 AR-1268-2	8.598	7.606	382.8E6	405.6E6	500.000	500.000
43)	L9 AR-1268-3	8.820	7.814	339.9E6	347.9E6	500.000	500.000
44)	L9 AR-1268-4	9.215	8.096	141.2E6	140.0E6	500.000	500.000
45)	L9 AR-1268-5	9.609	8.387	992.9E6	909.2E6	500.000	500.000

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


ICAL Form


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
Data File : PO114896.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Nov 2025 17:55
Operator : YP/AJ
Sample : AR1268ICC500
Misc :
ALS Vial : 28 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1268ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 05 01:23:29 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Nov 05 01:21:38 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
Data File : PO114897.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Nov 2025 18:13
Operator : YP/AJ
Sample : AR1268ICC250
Misc :
ALS Vial : 29 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1268ICC250

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 11/05/2025
Supervised By :Yogesh Patel 11/05/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 05 01:23:37 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Nov 05 01:21:38 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.348	3.649	189.9E6	122.1E6	24.810	24.834
2)	SA Decachlor...	9.932	8.641	210.5E6	196.5E6	24.748	25.504
	Target Compounds						
41)	L9 AR-1268-1	8.511	7.541	221.9E6	248.4E6	248.457	253.017m
42)	L9 AR-1268-2	8.599	7.605	190.0E6	204.3E6	248.168	251.781
43)	L9 AR-1268-3	8.820	7.813	167.9E6	174.1E6	246.983	250.180
44)	L9 AR-1268-4	9.216	8.096	70540429	70593070	249.869	252.040
45)	L9 AR-1268-5	9.610	8.387	489.2E6	458.0E6	246.358	251.855

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
Data File : PO114897.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Nov 2025 18:13
Operator : YP/AJ
Sample : AR1268ICC250
Misc :
ALS Vial : 29 Sample Multiplier: 1


Instrument :
ECD_O
ClientSampleId :
AR1268ICC250


Manual Integrations APPROVED

Reviewed By :Rahul Chavli 11/05/2025
Supervised By :Yogesh Patel 11/05/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 05 01:23:37 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Nov 05 01:21:38 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
Data File : PO114898.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Nov 2025 18:31
Operator : YP/AJ
Sample : AR1268ICC050
Misc :
ALS Vial : 30 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1268ICC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 05 01:23:46 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Nov 05 01:21:38 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.348	3.648	35607202	23620780	4.651	4.805
2)	SA Decachlor...	9.931	8.640	38610556	41609899	4.540	5.400
	Target Compounds						
41)	L9 AR-1268-1	8.511	7.542	40675439	49554127	45.547	50.476
42)	L9 AR-1268-2	8.599	7.606	35140218	38446458	45.903	47.393
43)	L9 AR-1268-3	8.820	7.814	30894342	31638164	45.442	45.475
44)	L9 AR-1268-4	9.214	8.096	14340449	13380470	50.797	47.773
45)	L9 AR-1268-5	9.609	8.387	90370817	85884170	45.510	47.229


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
Data File : PO114898.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Nov 2025 18:31
Operator : YP/AJ
Sample : AR1268ICC050
Misc :
ALS Vial : 30 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1268ICC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 05 01:23:46 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Nov 05 01:21:38 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
Data File : PO114899.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Nov 2025 19:08
Operator : YP/AJ
Sample : PO110425ICV500
Misc :
ALS Vial : 31 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
ICVPO110425

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 05 02:03:07 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Nov 05 02:02:35 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.348	3.649	350.6E6	216.0E6	49.817	48.358
2)	SA Decachlor...	9.932	8.640	217.2E6	202.0E6	49.145	51.742
	Target Compounds						
3)	L1 AR-1016-1	5.487	4.728	119.9E6	92647170	487.281	486.962
4)	L1 AR-1016-2	5.508	4.747	193.0E6	132.4E6	485.938	488.627
5)	L1 AR-1016-3	5.570	4.921	110.7E6	72980433	496.242	493.200
6)	L1 AR-1016-4	5.666	4.963	92089690	57444659	487.569	489.233
7)	L1 AR-1016-5	5.957	5.176	84866211	74982175	476.249	490.999
31)	L7 AR-1260-1	7.071	6.205	144.7E6	141.9E6	473.335	479.180
32)	L7 AR-1260-2	7.325	6.393	200.2E6	209.0E6	505.658	486.103
33)	L7 AR-1260-3	7.682	6.546	167.5E6	170.3E6	504.732	484.991
34)	L7 AR-1260-4	7.904	7.017	152.7E6	150.0E6	500.629	495.013
35)	L7 AR-1260-5	8.210	7.257	350.1E6	404.4E6	490.903	492.224

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


ICAL Form


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
Data File : PO114899.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Nov 2025 19:08
Operator : YP/AJ
Sample : PO110425ICV500
Misc :
ALS Vial : 31 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
ICVPO110425

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 05 02:03:07 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Nov 05 02:02:35 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
Data File : PO114900.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Nov 2025 20:21
Operator : YP/AJ
Sample : AR1242ICV500
Misc :
ALS Vial : 32 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
ICVPO110425AR1242

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 11/05/2025
Supervised By :Yogesh Patel 11/05/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 05 01:57:53 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Nov 05 01:57:32 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.348	3.649	369.9E6	227.0E6	50.980	48.485
2)	SA Decachlor...	9.932	8.639	237.1E6	216.0E6	50.883	50.180
	Target Compounds						
16)	L4 AR-1242-1	5.486	4.727	91685645	72678150	492.296m	497.041
17)	L4 AR-1242-2	5.508	4.747	146.5E6	102.4E6	490.193m	494.324
18)	L4 AR-1242-3	5.570	4.921	84401159	56402697	492.512m	503.697
19)	L4 AR-1242-4	5.666	5.005	70483519	53471210	492.260m	499.119
20)	L4 AR-1242-5	6.393	5.525	75170815	70192007	533.288	491.136

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
Data File : PO114900.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Nov 2025 20:21
Operator : YP/AJ
Sample : AR1242ICV500
Misc :
ALS Vial : 32 Sample Multiplier: 1


Instrument :
ECD_O
ClientSampleId :
ICVPO110425AR1242


Manual Integrations APPROVED

Reviewed By :Rahul Chavli 11/05/2025
Supervised By :Yogesh Patel 11/05/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 05 01:57:53 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Nov 05 01:57:32 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
Data File : PO114901.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Nov 2025 20:58
Operator : YP/AJ
Sample : AR1248ICV500
Misc :
ALS Vial : 33 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
ICVPO110425AR1248

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 05 01:49:34 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Nov 05 01:49:02 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.348	3.649	364.4E6	233.5E6	49.370	48.616
2)	SA Decachlor...	9.931	8.640	231.0E6	213.4E6	50.545	50.607
	Target Compounds						
21)	L5 AR-1248-1	5.487	4.727	67941811	53066852	485.834	485.472
22)	L5 AR-1248-2	5.758	4.963	96189175	72201389	494.605	484.897
23)	L5 AR-1248-3	5.958	5.005	104.8E6	75367544	497.247	484.966
24)	L5 AR-1248-4	6.354	5.175	125.7E6	89630325	497.795	489.109
25)	L5 AR-1248-5	6.393	5.567	118.0E6	89347145	500.480	484.717

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


ICAL Form


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
Data File : PO114901.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Nov 2025 20:58
Operator : YP/AJ
Sample : AR1248ICV500
Misc :
ALS Vial : 33 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
ICVPO110425AR1248

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 05 01:49:34 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Nov 05 01:49:02 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
Data File : PO114902.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Nov 2025 21:35
Operator : YP/AJ
Sample : AR1254ICV500
Misc :
ALS Vial : 34 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
ICVPO110425AR1254

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 05 05:47:46 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Nov 05 05:47:04 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.348	3.649	365.7E6	233.3E6	48.598	47.717
2)	SA Decachlor...	9.931	8.640	237.7E6	218.5E6	49.015	49.554
	Target Compounds						
26)	L6 AR-1254-1	6.331	5.526	122.7E6	141.2E6	484.213	470.374
27)	L6 AR-1254-2	6.546	5.673	176.1E6	124.6E6	475.284	462.124
28)	L6 AR-1254-3	6.906	6.075	193.7E6	200.1E6	503.340	473.848
29)	L6 AR-1254-4	7.188	6.302	159.9E6	143.9E6	486.428	484.000
30)	L6 AR-1254-5	7.603	6.719	147.9E6	190.6E6	490.023	479.436

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


ICAL Form


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
Data File : PO114902.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Nov 2025 21:35
Operator : YP/AJ
Sample : AR1254ICV500
Misc :
ALS Vial : 34 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
ICVPO110425AR1254

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 05 05:47:46 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Nov 05 05:47:04 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
Data File : PO114903.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Nov 2025 22:12
Operator : YP/AJ
Sample : AR1268ICV500
Misc :
ALS Vial : 35 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
ICVPO110425AR1268

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 05 01:39:52 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Nov 05 01:38:15 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.347	3.647	376.8E6	240.7E6	50.252	49.688
2)	SA Decachlor...	9.929	8.640	416.0E6	381.0E6	50.328	49.259
	Target Compounds						
41)	L9 AR-1268-1	8.509	7.541	437.5E6	476.9E6	502.163	488.581
42)	L9 AR-1268-2	8.599	7.605	374.7E6	393.6E6	501.949	494.742
43)	L9 AR-1268-3	8.819	7.814	331.8E6	336.4E6	502.049	497.748
44)	L9 AR-1268-4	9.214	8.096	138.2E6	136.4E6	492.596	495.687
45)	L9 AR-1268-5	9.608	8.385	968.0E6	889.7E6	499.171	498.578


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
Data File : PO114903.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Nov 2025 22:12
Operator : YP/AJ
Sample : AR1268ICV500
Misc :
ALS Vial : 35 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
ICVPO110425AR1268

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 05 01:39:52 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Nov 05 01:38:15 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





Raw Data

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

RETENTION TIMES OF INITIAL CALIBRATION

Lab Name: Alliance **Contract:** MATR02

Lab Code: ACE **SDG NO.:** Q3450

Instrument ID: ECD_P **Calibration Date(s):** 10/21/2025 10/21/2025

Calibration Times: 09:06 17:01

GC Column: ZB-MR1 ID: 0.32 (mm)

LAB FILE ID: RT 1000 = PP075888.D RT 750 = PP075889.D

RT 500 = PP075890.D RT 250 = PP075891.D RT 050 = PP075892.D

COMPOUND		RT 1000	RT 750	RT 500	RT 250	RT 050	MEAN RT	RT WINDOW FROM	RT WINDOW TO
Aroclor-1016-1	(1)	5.80	5.80	5.80	5.80	5.80	5.80	5.70	5.90
Aroclor-1016-2	(2)	5.82	5.82	5.82	5.82	5.82	5.82	5.72	5.92
Aroclor-1016-3	(3)	5.88	5.88	5.88	5.88	5.88	5.88	5.78	5.98
Aroclor-1016-4	(4)	5.98	5.98	5.98	5.98	5.98	5.98	5.88	6.08
Aroclor-1016-5	(5)	6.27	6.27	6.27	6.27	6.27	6.27	6.17	6.37
Aroclor-1260-1	(1)	7.39	7.39	7.39	7.39	7.39	7.39	7.29	7.49
Aroclor-1260-2	(2)	7.64	7.64	7.64	7.64	7.64	7.64	7.54	7.74
Aroclor-1260-3	(3)	8.00	8.00	8.00	8.00	8.00	8.00	7.90	8.10
Aroclor-1260-4	(4)	8.23	8.23	8.23	8.23	8.23	8.23	8.13	8.33
Aroclor-1260-5	(5)	8.56	8.55	8.55	8.55	8.56	8.55	8.45	8.65
Decachlorobiphenyl		10.42	10.42	10.42	10.42	10.42	10.42	10.32	10.52
Tetrachloro-m-xylene		4.65	4.64	4.64	4.65	4.65	4.64	4.54	4.74
Aroclor-1242-1	(1)	5.80	5.80	5.80	5.80	5.80	5.80	5.70	5.90
Aroclor-1242-2	(2)	5.82	5.82	5.82	5.82	5.82	5.82	5.72	5.92
Aroclor-1242-3	(3)	5.88	5.88	5.88	5.88	5.88	5.88	5.78	5.98
Aroclor-1242-4	(4)	5.98	5.98	5.98	5.98	5.98	5.98	5.88	6.08
Aroclor-1242-5	(5)	6.71	6.71	6.71	6.71	6.71	6.71	6.61	6.81
Decachlorobiphenyl		10.42	10.42	10.42	10.42	10.42	10.42	10.32	10.52
Tetrachloro-m-xylene		4.65	4.64	4.65	4.65	4.65	4.65	4.55	4.75
Aroclor-1248-1	(1)	5.80	5.80	5.80	5.80	5.80	5.80	5.70	5.90
Aroclor-1248-2	(2)	6.07	6.07	6.07	6.07	6.07	6.07	5.97	6.17
Aroclor-1248-3	(3)	6.27	6.27	6.27	6.27	6.27	6.27	6.17	6.37
Aroclor-1248-4	(4)	6.67	6.67	6.67	6.67	6.67	6.67	6.57	6.77
Aroclor-1248-5	(5)	6.71	6.71	6.71	6.71	6.71	6.71	6.61	6.81
Decachlorobiphenyl		10.43	10.42	10.42	10.42	10.42	10.42	10.32	10.52
Tetrachloro-m-xylene		4.65	4.64	4.65	4.65	4.65	4.65	4.55	4.75
Aroclor-1254-1	(1)	6.65	6.65	6.65	6.65	6.64	6.65	6.55	6.75
Aroclor-1254-2	(2)	6.86	6.86	6.87	6.86	6.86	6.86	6.76	6.96
Aroclor-1254-3	(3)	7.23	7.23	7.23	7.23	7.22	7.23	7.13	7.33
Aroclor-1254-4	(4)	7.51	7.51	7.51	7.51	7.50	7.51	7.41	7.61
Aroclor-1254-5	(5)	7.93	7.92	7.93	7.93	7.92	7.92	7.82	8.02
Decachlorobiphenyl		10.42	10.42	10.42	10.42	10.42	10.42	10.32	10.52
Tetrachloro-m-xylene		4.65	4.64	4.65	4.65	4.64	4.65	4.55	4.75
Aroclor-1268-1	(1)	8.88	8.88	8.88	8.88	8.88	8.88	8.78	8.98
Aroclor-1268-2	(2)	8.97	8.97	8.97	8.98	8.97	8.97	8.87	9.07
Aroclor-1268-3	(3)	9.21	9.21	9.21	9.21	9.21	9.21	9.11	9.31
Aroclor-1268-4	(4)	9.64	9.64	9.64	9.64	9.64	9.64	9.54	9.74
Aroclor-1268-5	(5)	10.07	10.07	10.07	10.07	10.07	10.07	9.97	10.17

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

RETENTION TIMES OF INITIAL CALIBRATION

Decachlorobiphenyl	10.42	10.42	10.42	10.43	10.42	10.42	10.32	10.52
Tetrachloro-m-xylene	4.65	4.65	4.65	4.65	4.65	4.65	4.55	4.75

Raw Data

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

RETENTION TIMES OF INITIAL CALIBRATION

Lab Name: Alliance **Contract:** MATR02

Lab Code: ACE **SDG NO.:** Q3450

Instrument ID: ECD_P **Calibration Date(s):** 10/21/2025 10/21/2025

Calibration Times: 09:06 17:01

GC Column: ZB-MR2 ID: 0.32 (mm)

LAB FILE ID: RT 1000 = PP075888.D RT 750 = PP075889.D

RT 500 = PP075890.D RT 250 = PP075891.D RT 050 = PP075892.D

COMPOUND		RT 1000	RT 750	RT 500	RT 250	RT 050	MEAN RT	RT WINDOW FROM	RT WINDOW TO
Aroclor-1016-1	(1)	4.88	4.88	4.88	4.88	4.88	4.88	4.78	4.98
Aroclor-1016-2	(2)	4.94	4.94	4.94	4.94	4.94	4.94	4.84	5.04
Aroclor-1016-3	(3)	5.06	5.06	5.05	5.06	5.05	5.06	4.96	5.16
Aroclor-1016-4	(4)	5.10	5.10	5.10	5.10	5.09	5.10	5.00	5.20
Aroclor-1016-5	(5)	5.31	5.31	5.31	5.31	5.31	5.31	5.21	5.41
Aroclor-1260-1	(1)	6.34	6.34	6.34	6.34	6.34	6.34	6.24	6.44
Aroclor-1260-2	(2)	6.53	6.53	6.53	6.53	6.53	6.53	6.43	6.63
Aroclor-1260-3	(3)	6.68	6.68	6.68	6.68	6.68	6.68	6.58	6.78
Aroclor-1260-4	(4)	7.15	7.15	7.15	7.15	7.15	7.15	7.05	7.25
Aroclor-1260-5	(5)	7.39	7.39	7.39	7.39	7.39	7.39	7.29	7.49
Decachlorobiphenyl		8.78	8.78	8.78	8.78	8.78	8.78	8.68	8.88
Tetrachloro-m-xylene		3.78	3.78	3.78	3.78	3.78	3.78	3.68	3.88
Aroclor-1242-1	(1)	4.88	4.88	4.88	4.88	4.88	4.88	4.78	4.98
Aroclor-1242-2	(2)	4.94	4.94	4.94	4.94	4.94	4.94	4.84	5.04
Aroclor-1242-3	(3)	5.06	5.06	5.06	5.06	5.05	5.06	4.96	5.16
Aroclor-1242-4	(4)	5.14	5.14	5.14	5.14	5.14	5.14	5.04	5.24
Aroclor-1242-5	(5)	5.66	5.66	5.66	5.66	5.66	5.66	5.56	5.76
Decachlorobiphenyl		8.78	8.78	8.79	8.78	8.78	8.78	8.68	8.88
Tetrachloro-m-xylene		3.78	3.78	3.78	3.78	3.78	3.78	3.68	3.88
Aroclor-1248-1	(1)	4.88	4.88	4.88	4.88	4.87	4.88	4.78	4.98
Aroclor-1248-2	(2)	5.10	5.10	5.10	5.10	5.10	5.10	5.00	5.20
Aroclor-1248-3	(3)	5.14	5.14	5.14	5.14	5.14	5.14	5.04	5.24
Aroclor-1248-4	(4)	5.31	5.31	5.31	5.31	5.31	5.31	5.21	5.41
Aroclor-1248-5	(5)	5.70	5.70	5.70	5.70	5.73	5.71	5.61	5.81
Decachlorobiphenyl		8.78	8.78	8.78	8.78	8.78	8.78	8.68	8.88
Tetrachloro-m-xylene		3.78	3.78	3.78	3.78	3.78	3.78	3.68	3.88
Aroclor-1254-1	(1)	5.66	5.66	5.66	5.66	5.66	5.66	5.56	5.76
Aroclor-1254-2	(2)	5.81	5.81	5.81	5.81	5.81	5.81	5.71	5.91
Aroclor-1254-3	(3)	6.21	6.21	6.21	6.21	6.21	6.21	6.11	6.31
Aroclor-1254-4	(4)	6.44	6.44	6.44	6.44	6.43	6.44	6.34	6.54
Aroclor-1254-5	(5)	6.85	6.85	6.85	6.85	6.85	6.85	6.75	6.95
Decachlorobiphenyl		8.78	8.78	8.79	8.78	8.78	8.78	8.68	8.88
Tetrachloro-m-xylene		3.78	3.78	3.79	3.78	3.78	3.78	3.68	3.88
Aroclor-1268-1	(1)	7.67	7.67	7.67	7.67	7.67	7.67	7.57	7.77
Aroclor-1268-2	(2)	7.74	7.74	7.74	7.74	7.74	7.74	7.64	7.84
Aroclor-1268-3	(3)	7.94	7.94	7.94	7.94	7.94	7.94	7.84	8.04
Aroclor-1268-4	(4)	8.23	8.23	8.23	8.23	8.23	8.23	8.13	8.33
Aroclor-1268-5	(5)	8.53	8.53	8.53	8.53	8.53	8.53	8.43	8.63

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

RETENTION TIMES OF INITIAL CALIBRATION

Decachlorobiphenyl	8.79	8.78	8.79	8.79	8.79	8.79	8.69	8.89
Tetrachloro-m-xylene	3.78	3.78	3.79	3.79	3.79	3.78	3.68	3.88

Raw Data

Alliance TECHNICAL GROUP

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION FACTOR OF INITIAL CALIBRATION

Lab Name: Alliance
Lab Code: ACE
Instrument ID: ECD_P

Contract: MATR02
SDG NO.: Q3450
Calibration Date(s): 10/21/2025 10/21/2025
Calibration Times: 09:06 17:01

GC Column: ZB-MR1 **ID:** 0.32 (mm)

LAB FILE ID: CF 1000 = PP075888.D CF 750 = PP075889.D
CF 500 = PP075890.D CF 250 = PP075891.D CF 050 = PP075892.D

COMPOUND		CF 1000	CF 750	CF 500	CF 250	CF 050	CF	% RSD
Aroclor-1016-1	(1)	53893845	57848439	60228768	63056856	60602640	59126110	6
Aroclor-1016-2	(2)	81387326	86593457	91038396	94931280	88270120	88444116	6
Aroclor-1016-3	(3)	50025937	53493744	56764402	59544472	56926600	55351031	7
Aroclor-1016-4	(4)	41815325	44617539	46982858	48544784	45240520	45440205	6
Aroclor-1016-5	(5)	39075321	41377448	43150568	44782124	48642160	43405524	8
Aroclor-1260-1	(1)	64655691	67908073	70553690	74100180	66565440	68756615	5
Aroclor-1260-2	(2)	80744063	85234851	88594752	93987136	99325680	89577296	8
Aroclor-1260-3	(3)	64964687	67552583	69003280	72715708	72196540	69286560	5
Aroclor-1260-4	(4)	65448766	70711301	72542976	74553116	59600740	68571380	9
Aroclor-1260-5	(5)	139384380	146230631	152452738	155039368	146650940	147951611	4
Decachlorobiphenyl		1012132950	1071302787	1124785220	1137889120	1036523400	1076526695	5
Tetrachloro-m-xylene		1370219590	1453116893	1490571040	1483020960	1348537200	1429093137	5
Aroclor-1242-1	(1)	44019933	46315467	47674416	50255084	51268840	47906748	6
Aroclor-1242-2	(2)	64942800	67715737	70979556	73767248	76553360	70791740	7
Aroclor-1242-3	(3)	40140584	42044764	44589646	47135032	54298900	45641785	12
Aroclor-1242-4	(4)	33543618	35162207	36420600	36967860	39702420	36359341	6
Aroclor-1242-5	(5)	36405140	38662289	41888622	39296052	48467620	40943945	11
Decachlorobiphenyl		1136280900	1209279307	1231278960	1289202560	1357807000	1244769745	7
Tetrachloro-m-xylene		1466601570	1512648573	1545825660	1595207720	1573043800	1538665465	3
Aroclor-1248-1	(1)	33471805	35785521	36963520	39121904	34010240	35870598	6
Aroclor-1248-2	(2)	44816839	47584043	49402910	52441868	49476340	48744400	6
Aroclor-1248-3	(3)	49477903	52446567	61105936	57335280	50958020	54264741	9
Aroclor-1248-4	(4)	60435029	64167197	67505182	72174088	67539280	66364155	7
Aroclor-1248-5	(5)	58134534	61713151	65347636	71297512	67810300	64860627	8
Decachlorobiphenyl		1140637530	1190681093	1243623480	1306886720	1250686400	1226503045	5
Tetrachloro-m-xylene		1468701040	1529744213	1558324940	1626496320	1450854600	1526824223	5
Aroclor-1254-1	(1)	66261960	71898345	78270020	82464920	100572060	79893461	16
Aroclor-1254-2	(2)	87851253	92570580	96669370	102809560	120020480	99984249	12
Aroclor-1254-3	(3)	92035771	96900691	99787992	105988040	125061300	103954759	12
Aroclor-1254-4	(4)	73136868	76689867	79390286	83967352	104884540	83613783	15
Aroclor-1254-5	(5)	85440842	89720104	92252192	95013656	95427040	91570767	5
Decachlorobiphenyl		1194970230	1245235800	1296058720	1332579960	1514859800	1316740902	9
Tetrachloro-m-xylene		1499217980	1559170693	1583264260	1614158480	1720822800	1595326843	5
Aroclor-1268-1	(1)	186733090	195066952	205606148	210022232	235253460	206536376	9

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION FACTOR OF INITIAL CALIBRATION

Aroclor-1268-2	(2)	172486766	179135464	188184104	190078376	211424800	188261902	8
Aroclor-1268-3	(3)	145245892	150281380	156894280	158090072	165408840	155184093	5
Aroclor-1268-4	(4)	71036137	74805685	76392484	76627212	94373740	78647052	12
Aroclor-1268-5	(5)	446420076	458917541	475352492	479939416	541304600	480386825	8
Decachlorobiphenyl		1988394870	2075434813	2173906840	2211711440	2356033600	2161096313	6
Tetrachloro-m-xylene		1515083040	1549265533	1603737900	1591085800	1659449000	1583724255	3

Raw Data

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION FACTOR OF INITIAL CALIBRATION

Lab Name: Alliance
Lab Code: ACE
Instrument ID: ECD_P

Contract: MATR02
SDG NO.: Q3450
Calibration Date(s): 10/21/2025 10/21/2025
Calibration Times: 09:06 17:01

GC Column: ZB-MR2 **ID:** 0.32 (mm)

LAB FILE ID: CF 1000 = PP075888.D CF 750 = PP075889.D
CF 500 = PP075890.D CF 250 = PP075891.D CF 050 = PP075892.D

COMPOUND		CF 1000	CF 750	CF 500	CF 250	CF 050	CF	% RSD
Aroclor-1016-1	(1)	708597432	755130533	750168220	768814916	718149060	740172032	3
Aroclor-1016-2	(2)	344688013	362255933	349033076	368003724	285681000	341932349	10
Aroclor-1016-3	(3)	184756021	198805969	201689770	204443556	237017160	205342495	9
Aroclor-1016-4	(4)	185429480	195612065	193130110	203643812	156612400	186885573	10
Aroclor-1016-5	(5)	202087874	223186380	221264342	249936892	243439240	227982946	8
Aroclor-1260-1	(1)	519115007	547143451	551523748	550034516	570658900	547695124	3
Aroclor-1260-2	(2)	816933203	847057133	836099344	841169992	727408040	813733542	6
Aroclor-1260-3	(3)	571186815	591919587	603061842	595667512	451948620	562756875	11
Aroclor-1260-4	(4)	562542796	557220415	556309062	533352684	387232680	519331527	14
Aroclor-1260-5	(5)	1443277342	1495374137	1448146362	1379476524	1011499060	1355554685	15
Decachlorobiphenyl		7970697090	8448130240	8416417320	8375691800	7903993200	8222985930	3
Tetrachloro-m-xylene		6687145430	7018651333	6804923780	6758234840	5822487600	6618288597	7
Aroclor-1242-1	(1)	571147464	588341980	597723718	569760388	546790280	574752766	3
Aroclor-1242-2	(2)	271622570	270496133	289123856	291348184	264965740	277511297	4
Aroclor-1242-3	(3)	148417684	150685240	162169734	157714788	168071120	157411713	5
Aroclor-1242-4	(4)	195275059	191450107	210824584	191270784	136740280	185112163	15
Aroclor-1242-5	(5)	229588702	236959084	244202752	234962408	205474700	230237529	6
Decachlorobiphenyl		8905080650	8993413907	9094156860	9113318400	9309338200	9083061603	2
Tetrachloro-m-xylene		7556133750	7417192413	7366207320	7201170040	6123421000	7132824905	8
Aroclor-1248-1	(1)	362937087	375551145	389160858	390611340	359825360	375617158	4
Aroclor-1248-2	(2)	226906631	241863020	243838414	257801160	259130280	245907901	5
Aroclor-1248-3	(3)	258884180	267447340	286041186	300772772	321318420	286892780	9
Aroclor-1248-4	(4)	264944699	288582168	281543548	284810560	293675780	282711351	4
Aroclor-1248-5	(5)	531974990	530976477	538260328	580030476	417051840	519658822	12
Decachlorobiphenyl		8995204870	9175718960	9269974040	9050625160	7344080200	8767120646	9
Tetrachloro-m-xylene		7343926730	7496283907	7432157100	7125361600	5416154400	6962776747	13
Aroclor-1254-1	(1)	570889099	597066183	605466528	586480320	460097860	563999998	11
Aroclor-1254-2	(2)	484331632	496825463	519240264	512723972	637282480	530080762	12
Aroclor-1254-3	(3)	866771846	893592493	892235862	873540968	729313040	851090842	8
Aroclor-1254-4	(4)	656603428	687178148	673132272	636725912	495267720	629781496	12
Aroclor-1254-5	(5)	714919063	793859644	718111150	679323440	693198580	719882375	6
Decachlorobiphenyl		9294247170	9652211800	9737198260	9430246600	10010120400	9624804846	3
Tetrachloro-m-xylene		7467434200	7759548133	7675903540	7263970760	7005463400	7434464007	4
Aroclor-1268-1	(1)	1831610079	1813456776	1858863308	1732000476	1817331660	1810652460	3

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION FACTOR OF INITIAL CALIBRATION

Aroclor-1268-2	(2)	1800363506	1772458256	1795415328	1662166904	1645167940	1735114387	4
Aroclor-1268-3	(3)	1369797429	1358701644	1402746042	1358715304	1345500120	1367092108	2
Aroclor-1268-4	(4)	542764588	536784571	540251440	501746924	465672060	517443917	6
Aroclor-1268-5	(5)	3855476951	3900359425	4013489614	3707995788	3381331260	3771730608	6
Decachlorobiphenyl		16718332370	16659738347	17037476180	16197346520	15850998600	16492778403	3
Tetrachloro-m-xylene		7499400880	7531972947	7553431460	7073074800	6855756800	7302727377	4

Raw Data

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Lab Name: Alliance **Contract:** MATR02

Lab Code: ACE **SDG NO.:** Q3450

Instrument ID: ECD_P **Date(s) Analyzed:** 10/21/2025 10/21/2025

GC Column: ZB-MR1 **ID:** 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW FROM	RT WINDOW TO	CALIBRATION FACTOR
Aroclor-1221	500	1	4.84	4.74	4.94	20176600
		2	4.93	4.83	5.03	15061400
		3	5.01	4.91	5.11	46915600
		4	0.00			0
		5	0.00			0
Aroclor-1232	500	1	5.01	4.91	5.11	39314800
		2	5.53	5.43	5.63	20627600
		3	5.82	5.72	5.92	38639600
		4	5.98	5.88	6.08	20172600
		5	6.07	5.97	6.17	15662900
Aroclor-1262	500	1	8.23	8.13	8.33	84026000
		2	8.56	8.46	8.66	163669000
		3	8.88	8.78	8.98	121787000
		4	8.97	8.87	9.07	92111600
		5	9.64	9.54	9.74	67824600

Raw Data

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Lab Name: Alliance **Contract:** MATR02

Lab Code: ACE **SDG NO.:** Q3450

Instrument ID: ECD_P **Date(s) Analyzed:** 10/21/2025 10/21/2025

GC Column: ZB-MR2 **ID:** 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW FROM	RT WINDOW TO	CALIBRATION FACTOR
Aroclor-1221	500	1	3.99	3.89	4.09	90009000
		2	4.08	3.98	4.18	63177800
		3	4.15	4.05	4.25	212036000
		4	0.00			0
		5	0.00			0
Aroclor-1232	500	1	4.15	4.05	4.25	168044000
		2	4.88	4.78	4.98	297516000
		3	5.06	4.96	5.16	81766800
		4	5.14	5.04	5.24	84943400
		5	5.31	5.21	5.41	100651000
Aroclor-1262	500	1	6.89	6.79	6.99	986952000
		2	7.15	7.05	7.25	703208000
		3	7.67	7.57	7.77	606960000
		4	7.74	7.64	7.84	1101340000
		5	8.23	8.13	8.33	461492000

Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075888.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2025 09:06
Operator : YP\AJ
Sample : AR1660ICC1000
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660ICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 17:20:08 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 21 17:14:53 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.646	3.784	137.0E6	668.7E6	95.793	99.127
2)	SA Decachlor...	10.422	8.784	101.2E6	797.1E6	94.728	97.280
	Target Compounds						
3)	L1 AR-1016-1	5.799	4.880	53893845	708.6E6	944.490	971.503
4)	L1 AR-1016-2	5.820	4.938	81387326	344.7E6	944.028	993.737
5)	L1 AR-1016-3	5.883	5.058	50025937	184.8E6	936.900	956.181
6)	L1 AR-1016-4	5.980	5.099	41815325	185.4E6	941.806	979.658
7)	L1 AR-1016-5	6.272	5.313	39075321	202.1E6	950.438	954.703
31)	L7 AR-1260-1	7.391	6.341	64655691	519.1E6	956.379	969.730
32)	L7 AR-1260-2	7.643	6.527	80744063	816.9E6	953.639	988.405
33)	L7 AR-1260-3	8.001	6.681	64964687	571.2E6	969.854	972.855
34)	L7 AR-1260-4	8.229	7.150	65448766	562.5E6	948.590	1005.572
35)	L7 AR-1260-5	8.555	7.390	139.4E6	1443.3E6	955.220	998.316

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


ICAL Form


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075888.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2025 09:06
Operator : YP\AJ
Sample : AR1660ICC1000
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660ICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 17:20:08 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 21 17:14:53 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075889.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2025 09:22
Operator : YP\AJ
Sample : AR1660ICC750
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 17:23:54 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 21 17:14:53 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.642	3.781	109.0E6	526.4E6	75.790	76.994
2)	SA Decachlor...	10.419	8.782	80347709	633.6E6	75.133	76.538
	Target Compounds						
3)	L1 AR-1016-1	5.795	4.877	43386329	566.3E6	756.866	767.445
4)	L1 AR-1016-2	5.817	4.935	64945093	271.7E6	752.204	771.869
5)	L1 AR-1016-3	5.879	5.055	40120308	149.1E6	750.922	764.309
6)	L1 AR-1016-4	5.977	5.096	33463154	146.7E6	752.456	766.543
7)	L1 AR-1016-5	6.270	5.309	31033086	167.4E6	753.210	776.704
31)	L7 AR-1260-1	7.387	6.338	50931055	410.4E6	752.240	760.963
32)	L7 AR-1260-2	7.640	6.525	63926138	635.3E6	753.332	762.324
33)	L7 AR-1260-3	7.999	6.678	50664437	443.9E6	754.232	754.073
34)	L7 AR-1260-4	8.226	7.147	53033476	417.9E6	762.329	748.026
35)	L7 AR-1260-5	8.552	7.387	109.7E6	1121.5E6	751.069	766.981

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


ICAL Form


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075889.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2025 09:22
Operator : YP\AJ
Sample : AR1660ICC750
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 17:23:54 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 21 17:14:53 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075890.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2025 09:38
Operator : YP\AJ
Sample : AR1660ICC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 17:17:19 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 21 17:14:53 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.642	3.781	74528552	340.2E6	50.000	50.000
2)	SA Decachlor...	10.419	8.782	56239261	420.8E6	50.000	50.000
	Target Compounds						
3)	L1 AR-1016-1	5.795	4.878	30114384	375.1E6	500.000	500.000
4)	L1 AR-1016-2	5.817	4.936	45519198	174.5E6	500.000	500.000
5)	L1 AR-1016-3	5.880	5.054	28382201	100.8E6	500.000	500.000
6)	L1 AR-1016-4	5.977	5.096	23491429	96565055	500.000	500.000
7)	L1 AR-1016-5	6.269	5.310	21575284	110.6E6	500.000	500.000
31)	L7 AR-1260-1	7.387	6.338	35276845	275.8E6	500.000	500.000
32)	L7 AR-1260-2	7.640	6.525	44297376	418.0E6	500.000	500.000
33)	L7 AR-1260-3	7.999	6.678	34501640	301.5E6	500.000	500.000
34)	L7 AR-1260-4	8.227	7.147	36271488	278.2E6	500.000	500.000
35)	L7 AR-1260-5	8.552	7.387	76226369	724.1E6	500.000	500.000

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


ICAL Form


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075890.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2025 09:38
Operator : YP\AJ
Sample : AR1660ICC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 17:17:19 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 21 17:14:53 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075891.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2025 09:54
Operator : YP\AJ
Sample : AR1660ICC250
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660ICC250

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 17:27:12 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 21 17:14:53 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.645	3.784	37075524	169.0E6	25.583	24.784
2)	SA Decachlor...	10.422	8.784	28447228	209.4E6	26.182	25.220
	Target Compounds						
3)	L1 AR-1016-1	5.798	4.879	15764214	192.2E6	268.295	257.757
4)	L1 AR-1016-2	5.819	4.938	23732820	92000931	268.205	258.433
5)	L1 AR-1016-3	5.882	5.057	14886118	51110889	270.868	258.889
6)	L1 AR-1016-4	5.979	5.098	12136196	50910953	266.787	261.815
7)	L1 AR-1016-5	6.272	5.311	11195531	62484223	265.950	278.799
31)	L7 AR-1260-1	7.390	6.340	18525045	137.5E6	267.300	253.727
32)	L7 AR-1260-2	7.643	6.526	23496784	210.3E6	269.643	251.752
33)	L7 AR-1260-3	8.001	6.680	18178927	148.9E6	265.157	252.205
34)	L7 AR-1260-4	8.229	7.150	18638279	133.3E6	263.200	241.399
35)	L7 AR-1260-5	8.554	7.389	38759842	344.9E6	261.402	239.232

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


ICAL Form


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075891.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2025 09:54
Operator : YP\AJ
Sample : AR1660ICC250
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660ICC250

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 17:27:12 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 21 17:14:53 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075892.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2025 10:11
Operator : YP\AJ
Sample : AR1660ICC050
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660ICC050

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/22/2025
Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 17:33:37 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 21 17:14:53 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.645	3.781	6742686	29112438	4.718	4.278m
2)	SA Decachlor...	10.422	8.783	5182617	39519966	4.814	4.806
	Target Compounds						
3)	L1 AR-1016-1	5.798	4.880	3030132	35907453	51.249	48.512
4)	L1 AR-1016-2	5.819	4.943	4413506	14284050	49.902	41.774
5)	L1 AR-1016-3	5.880	5.052	2846330	11850858	50.609m	58.316m
6)	L1 AR-1016-4	5.978	5.094	2262026	7830620	48.653m	41.996m
7)	L1 AR-1016-5	6.272	5.308	2432108	12171962	56.032	53.433m
31)	L7 AR-1260-1	7.390	6.339	3328272	28532945	48.407	52.096
32)	L7 AR-1260-2	7.643	6.528	4966284	36370402	55.441	44.696
33)	L7 AR-1260-3	8.001	6.681	3609827	22597431	52.100	40.155
34)	L7 AR-1260-4	8.229	7.150	2980037	19361634	43.459	37.282
35)	L7 AR-1260-5	8.555	7.388	7332547	50574953	49.560	37.309

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075892.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2025 10:11
Operator : YP\AJ
Sample : AR1660ICC050
Misc :
ALS Vial : 7 Sample Multiplier: 1


Instrument :
ECD_P
ClientSampleId :
AR1660ICC050


Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/22/2025
Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 17:33:37 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 21 17:14:53 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075893.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2025 10:27
Operator : YP\AJ
Sample : AR1221ICC500
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1221ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 16:37:58 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 21 16:37:41 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.643	3.782	75268417	347.9E6	50.000	50.000
2)	SA Decachlor...	10.419	8.781	55765598	413.6E6	50.000	50.000
	Target Compounds						
8)	L2 AR-1221-1	4.843	3.993	10088314	45004530	500.000	500.000
9)	L2 AR-1221-2	4.929	4.076	7530715	31588869	500.000	500.000
10)	L2 AR-1221-3	5.005	4.154	23457783	106.0E6	500.000	500.000

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


ICAL Form


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075893.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2025 10:27
Operator : YP\AJ
Sample : AR1221ICC500
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1221ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 16:37:58 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 21 16:37:41 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075894.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2025 10:43
Operator : YP\AJ
Sample : AR1232ICC500
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1232ICC500

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/22/2025
Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 16:45:39 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 21 16:45:19 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.644	3.783	76762692	367.2E6	50.000	50.000
2)	SA Decachlor...	10.418	8.783	57573853	421.9E6	50.000	50.000
	Target Compounds						
11)	L3 AR-1232-1	5.006	4.154	19657396	84022049	500.000	500.000
12)	L3 AR-1232-2	5.532	4.877	10313847	148.8E6	500.000	500.000
13)	L3 AR-1232-3	5.818	5.056	19319843	40883369	500.000	500.000
14)	L3 AR-1232-4	5.978	5.136	10086255	42471740	500.000	480.688m
15)	L3 AR-1232-5	6.068	5.311	7831441	50325540	500.000	500.000

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075894.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2025 10:43
Operator : YP\AJ
Sample : AR1232ICC500
Misc :
ALS Vial : 9 Sample Multiplier: 1


Instrument :
ECD_P
ClientSampleId :
AR1232ICC500


Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/22/2025
Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 16:45:39 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 21 16:45:19 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075895.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2025 10:59
Operator : YP\AJ
Sample : AR1242ICC1000
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1242ICC1000

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/22/2025
Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 16:57:47 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 21 16:53:53 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.646	3.784	146.7E6	755.6E6	97.370	101.273
2)	SA Decachlor...	10.422	8.784	113.6E6	890.5E6	95.988	98.950
	Target Compounds						
16)	L4 AR-1242-1	5.799	4.880	44019933	571.1E6	960.145	977.263
17)	L4 AR-1242-2	5.820	4.938	64942800	271.6E6	955.587	968.789
18)	L4 AR-1242-3	5.882	5.058	40140584	148.4E6	947.491	955.722
19)	L4 AR-1242-4	5.980	5.140	33543618	195.3E6	958.879	946.968m
20)	L4 AR-1242-5	6.710	5.660	36405140	229.6E6	929.963	969.155

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075895.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2025 10:59
Operator : YP\AJ
Sample : AR1242ICC1000
Misc :
ALS Vial : 10 Sample Multiplier: 1


Instrument :
ECD_P
ClientSampleId :
AR1242ICC1000


Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/22/2025
Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 16:57:47 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 21 16:53:53 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075896.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2025 11:16
Operator : YP\AJ
Sample : AR1242ICC750
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1242ICC750

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/22/2025
Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 17:03:07 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 21 16:53:53 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.644	3.782	113.4E6	556.3E6	75.213	74.705
2)	SA Decachlor...	10.420	8.782	90695948	674.5E6	76.069	74.966
	Target Compounds						
16)	L4 AR-1242-1	5.797	4.877	34736600	441.3E6	755.090	753.335
17)	L4 AR-1242-2	5.818	4.936	50786803	202.9E6	748.192	732.177
18)	L4 AR-1242-3	5.881	5.056	31533573	113.0E6	746.210	735.014
19)	L4 AR-1242-4	5.978	5.138	26371655	143.6E6	752.570	715.813m
20)	L4 AR-1242-5	6.708	5.658	28996717	177.7E6	743.785	750.134

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075896.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2025 11:16
Operator : YP\AJ
Sample : AR1242ICC750
Misc :
ALS Vial : 11 Sample Multiplier: 1


Instrument :
ECD_P
ClientSampleId :
AR1242ICC750


Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/22/2025
Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 17:03:07 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 21 16:53:53 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075897.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2025 11:48
Operator : YP\AJ
Sample : AR1242ICC500
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1242ICC500

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/22/2025
Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 16:54:12 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 21 16:53:53 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.647	3.784	77291283	368.3E6	50.000	50.000
2)	SA Decachlor...	10.422	8.785	61563948	454.7E6	50.000	50.000
	Target Compounds						
16)	L4 AR-1242-1	5.799	4.880	23837208	298.9E6	498.805	500.000
17)	L4 AR-1242-2	5.821	4.938	35489778	144.6E6	500.000	500.000
18)	L4 AR-1242-3	5.884	5.059	22294823	81084867	500.000	500.000
19)	L4 AR-1242-4	5.981	5.141	18210300	105.4E6	500.000	515.181m
20)	L4 AR-1242-5	6.710	5.660	20944311	122.1E6	500.000	500.000

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075897.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2025 11:48
Operator : YP\AJ
Sample : AR1242ICC500
Misc :
ALS Vial : 12 Sample Multiplier: 1


Instrument :
ECD_P
ClientSampleId :
AR1242ICC500


Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/22/2025
Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 16:54:12 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 21 16:53:53 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075898.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2025 12:04
Operator : YP\AJ
Sample : AR1242ICC250
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1242ICC250

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/22/2025
Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 17:06:44 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 21 16:53:53 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.645	3.783	39880193	180.0E6	26.064	24.377
2)	SA Decachlor...	10.420	8.782	32230064	227.8E6	26.494	25.240
	Target Compounds						
16)	L4 AR-1242-1	5.797	4.880	12563771	142.4E6	266.938	244.850
17)	L4 AR-1242-2	5.819	4.936	18441812	72837046	265.919	259.532
18)	L4 AR-1242-3	5.881	5.057	11783758	39428697	271.031	254.795
19)	L4 AR-1242-4	5.979	5.139	9241965	47817696	260.164	239.194m
20)	L4 AR-1242-5	6.707	5.659	9824013	58740602	248.605m	248.450

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075898.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2025 12:04
Operator : YP\AJ
Sample : AR1242ICC250
Misc :
ALS Vial : 13 Sample Multiplier: 1


Instrument :
ECD_P
ClientSampleId :
AR1242ICC250


Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/22/2025
Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 17:06:44 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 21 16:53:53 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075899.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2025 12:21
Operator : YP\AJ
Sample : AR1242ICC050
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1242ICC050

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/22/2025
Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 17:10:30 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 21 16:53:53 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.646	3.783	7865219	30617105	5.112	4.301m
2)	SA Decachlor...	10.422	8.783	6789035	46546691	5.454	5.125
	Target Compounds						
16)	L4 AR-1242-1	5.798	4.882	2563442	27339514	53.509	47.567
17)	L4 AR-1242-2	5.820	4.939	3827668	13248287	54.069	47.740
18)	L4 AR-1242-3	5.883	5.054	2714945	8403556	59.484	53.386
19)	L4 AR-1242-4	5.979	5.138	1985121	6837014	54.597	36.934 #
20)	L4 AR-1242-5	6.708	5.660	2423381	10273735	59.725m	44.622 #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075899.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2025 12:21
Operator : YP\AJ
Sample : AR1242ICC050
Misc :
ALS Vial : 14 Sample Multiplier: 1


Instrument :
ECD_P
ClientSampleId :
AR1242ICC050


Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/22/2025
Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 17:10:30 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 21 16:53:53 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075900.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2025 12:41
Operator : YP\AJ
Sample : AR1248ICC1000
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1248ICC1000

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/22/2025
Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 14:17:09 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 21 14:13:23 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.647	3.780	146.9E6	734.4E6	97.039	99.403
2)	SA Decachlor...	10.426	8.783	114.1E6	899.5E6	95.681	98.496
	Target Compounds						
21)	L5 AR-1248-1	5.799	4.876	33471805	362.9E6	950.427	965.133
22)	L5 AR-1248-2	6.071	5.095	44816839	226.9E6	951.326	964.032
23)	L5 AR-1248-3	6.274	5.136	49477903	258.9E6	894.849	947.454m
24)	L5 AR-1248-4	6.672	5.309	60435029	264.9E6	944.739	969.626
25)	L5 AR-1248-5	6.711	5.698	58134534	532.0E6	941.586	994.127

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075900.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2025 12:41
Operator : YP\AJ
Sample : AR1248ICC1000
Misc :
ALS Vial : 15 Sample Multiplier: 1


Instrument :
ECD_P
ClientSampleId :
AR1248ICC1000


Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/22/2025
Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 14:17:09 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 21 14:13:23 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075901.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2025 12:57
Operator : YP\AJ
Sample : AR1248ICC750
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1248ICC750

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/22/2025
Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 14:20:31 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 21 14:13:23 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.644	3.784	114.7E6	562.2E6	75.534	75.729
2)	SA Decachlor...	10.422	8.784	89301082	688.2E6	74.939	75.236
	Target Compounds						
21)	L5 AR-1248-1	5.797	4.879	26839141	281.7E6	758.019	749.338
22)	L5 AR-1248-2	6.068	5.099	35688032	181.4E6	755.016	763.662
23)	L5 AR-1248-3	6.271	5.139	39334925	200.6E6	723.821	728.506m
24)	L5 AR-1248-4	6.670	5.311	48125398	216.4E6	751.539	777.551
25)	L5 AR-1248-5	6.709	5.701	46284863	398.2E6	749.774	746.121

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075901.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2025 12:57
Operator : YP\AJ
Sample : AR1248ICC750
Misc :
ALS Vial : 16 Sample Multiplier: 1


Instrument :
ECD_P
ClientSampleId :
AR1248ICC750


Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/22/2025
Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 14:20:31 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 21 14:13:23 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075902.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2025 13:13
Operator : YP\AJ
Sample : AR1248ICC500
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1248ICC500

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/22/2025
Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 14:13:37 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 21 14:13:23 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.645	3.784	77916247	371.6E6	50.000	50.000
2)	SA Decachlor...	10.424	8.784	62181174	463.5E6	50.000	50.000
	Target Compounds						
21)	L5 AR-1248-1	5.798	4.877	18481760	194.6E6	500.000	500.000
22)	L5 AR-1248-2	6.069	5.098	24701455	121.9E6	500.000	500.000
23)	L5 AR-1248-3	6.273	5.138	30552968	143.0E6	500.000	491.112m
24)	L5 AR-1248-4	6.671	5.312	33752591	140.8E6	500.000	500.000
25)	L5 AR-1248-5	6.710	5.701	32673818	269.1E6	500.000	500.000

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075902.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2025 13:13
Operator : YP\AJ
Sample : AR1248ICC500
Misc :
ALS Vial : 17 Sample Multiplier: 1


Instrument :
ECD_P
ClientSampleId :
AR1248ICC500


Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/22/2025
Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 14:13:37 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 21 14:13:23 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075903.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2025 13:30
Operator : YP\AJ
Sample : AR1248ICC250
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1248ICC250

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/22/2025
Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 14:24:34 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 21 14:13:23 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.646	3.783	40662408	178.1E6	26.305	24.238
2)	SA Decachlor...	10.423	8.784	32672168	226.3E6	26.770	24.802
	Target Compounds						
21)	L5 AR-1248-1	5.799	4.878	9780476	97652835	269.170	257.276
22)	L5 AR-1248-2	6.070	5.099	13110467	64450290	269.977	265.662
23)	L5 AR-1248-3	6.273	5.138	14333820	75193193	260.182	273.023m
24)	L5 AR-1248-4	6.671	5.312	18043522	71202640	273.096	254.322
25)	L5 AR-1248-5	6.710	5.701	17824378	145.0E6	277.971	265.917

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075903.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2025 13:30
Operator : YP\AJ
Sample : AR1248ICC250
Misc :
ALS Vial : 18 Sample Multiplier: 1


Instrument :
ECD_P
ClientSampleId :
AR1248ICC250


Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/22/2025
Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 14:24:34 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 21 14:13:23 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075904.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2025 13:46
Operator : YP\AJ
Sample : AR1248ICC050
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1248ICC050

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/22/2025
Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 14:29:55 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 21 14:13:23 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.645	3.780	7254273	27080772	4.751	3.914m
2)	SA Decachlor...	10.422	8.783	6253432	36720401	5.099	4.188
	Target Compounds						
21)	L5 AR-1248-1	5.798	4.873	1700512	17991268	47.407	47.898
22)	L5 AR-1248-2	6.069	5.101	2473817	12956514	50.751	52.688
23)	L5 AR-1248-3	6.273	5.138	2547901	16065921	46.953	56.000
24)	L5 AR-1248-4	6.669	5.313	3376964	14683789	49.218m	51.704m
25)	L5 AR-1248-5	6.708	5.730	3390515	20852592	50.206m	38.783m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075904.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2025 13:46
Operator : YP\AJ
Sample : AR1248ICC050
Misc :
ALS Vial : 19 Sample Multiplier: 1


Instrument :
ECD_P
ClientSampleId :
AR1248ICC050


Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/22/2025
Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 14:29:55 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 21 14:13:23 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075905.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2025 14:02
Operator : YP\AJ
Sample : AR1254ICC1000
Misc :
ALS Vial : 20 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1254ICC1000

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/22/2025
Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 15:57:57 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 21 15:24:42 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.646	3.783	149.9E6	746.7E6	97.273	98.623
2)	SA Decachlor...	10.424	8.784	119.5E6	929.4E6	95.942	97.673
	Target Compounds						
26)	L6 AR-1254-1	6.646	5.662	66261960	570.9E6	872.670m	970.606
27)	L6 AR-1254-2	6.864	5.809	87851253	484.3E6	952.211	965.216
28)	L6 AR-1254-3	7.226	6.211	92035771	866.8E6	959.587	985.524
29)	L6 AR-1254-4	7.509	6.437	73136868	656.6E6	959.001	987.570
30)	L6 AR-1254-5	7.925	6.854	85440842	714.9E6	961.668	997.772

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075905.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2025 14:02
Operator : YP\AJ
Sample : AR1254ICC1000
Misc :
ALS Vial : 20 Sample Multiplier: 1


Instrument :
ECD_P
ClientSampleId :
AR1254ICC1000


Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/22/2025
Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 15:57:57 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 21 15:24:42 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075906.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2025 14:18
Operator : YP\AJ
Sample : AR1254ICC750
Misc :
ALS Vial : 21 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1254ICC750

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/22/2025
Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 16:03:45 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 21 15:24:42 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.644	3.783	116.9E6	582.0E6	75.579	76.230
2)	SA Decachlor...	10.423	8.784	93392685	723.9E6	74.989	75.714
	Target Compounds						
26)	L6 AR-1254-1	6.645	5.661	53923759	447.8E6	731.053m	757.518
27)	L6 AR-1254-2	6.863	5.808	69427935	372.6E6	751.680	745.041
28)	L6 AR-1254-3	7.225	6.210	72675518	670.2E6	755.137	757.967
29)	L6 AR-1254-4	7.507	6.436	57517400	515.4E6	752.790	766.592
30)	L6 AR-1254-5	7.923	6.851	67290078	595.4E6	754.900	785.342m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075906.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2025 14:18
Operator : YP\AJ
Sample : AR1254ICC750
Misc :
ALS Vial : 21 Sample Multiplier: 1


Instrument :
ECD_P
ClientSampleId :
AR1254ICC750


Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/22/2025
Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 16:03:45 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 21 15:24:42 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075907.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2025 14:35
Operator : YP\AJ
Sample : AR1254ICC500
Misc :
ALS Vial : 22 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1254ICC500

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/22/2025
Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 15:24:59 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 21 15:24:42 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.647	3.785	79163213	383.8E6	50.000	50.000
2)	SA Decachlor...	10.424	8.785	64802936	486.9E6	50.000	50.000
	Target Compounds						
26)	L6 AR-1254-1	6.648	5.663	39135010	302.7E6	480.703m	500.000
27)	L6 AR-1254-2	6.866	5.810	48334685	259.6E6	500.000	500.000
28)	L6 AR-1254-3	7.228	6.212	49893996	446.1E6	500.000	500.000
29)	L6 AR-1254-4	7.510	6.438	39695143	336.6E6	500.000	500.000
30)	L6 AR-1254-5	7.926	6.854	46126096	359.1E6	500.000	500.000

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075907.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2025 14:35
Operator : YP\AJ
Sample : AR1254ICC500
Misc :
ALS Vial : 22 Sample Multiplier: 1


Instrument :
ECD_P
ClientSampleId :
AR1254ICC500


Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/22/2025
Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 15:24:59 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 21 15:24:42 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075908.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2025 14:51
Operator : YP\AJ
Sample : AR1254ICC250
Misc :
ALS Vial : 23 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1254ICC250

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/22/2025
Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 16:07:49 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 21 15:24:42 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.646	3.784	40353962	181.6E6	25.803	24.079
2)	SA Decachlor...	10.422	8.784	33314499	235.8E6	26.290	24.742
	Target Compounds						
26)	L6 AR-1254-1	6.647	5.662	20616230	146.6E6	278.580m	248.519
27)	L6 AR-1254-2	6.864	5.809	25702390	128.2E6	270.622	254.691
28)	L6 AR-1254-3	7.226	6.210	26497010	218.4E6	268.520	247.733
29)	L6 AR-1254-4	7.508	6.436	20991838	159.2E6	268.108	239.944
30)	L6 AR-1254-5	7.925	6.853	23753414	169.8E6	262.160	233.749

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075908.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2025 14:51
Operator : YP\AJ
Sample : AR1254ICC250
Misc :
ALS Vial : 23 Sample Multiplier: 1


Instrument :
ECD_P
ClientSampleId :
AR1254ICC250


Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/22/2025
Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 16:07:49 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 21 15:24:42 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075909.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2025 15:07
Operator : YP\AJ
Sample : AR1254ICC050
Misc :
ALS Vial : 24 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1254ICC050

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/22/2025
Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 16:12:03 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 21 15:24:42 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.642	3.780	8604114	35027317	5.393	4.744m
2)	SA Decachlor...	10.419	8.782	7574299	50050602	5.752	5.200
	Target Compounds						
26)	L6 AR-1254-1	6.644	5.659	5028603	23004893	64.906	40.789 #
27)	L6 AR-1254-2	6.860	5.807	6001024	31864124	60.020	60.390
28)	L6 AR-1254-3	7.222	6.206	6253065	36465652	60.152	42.846 #
29)	L6 AR-1254-4	7.504	6.433	5244227	24763386	62.720	39.321 #
30)	L6 AR-1254-5	7.921	6.850	4771352	34659929	52.106	48.147

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075909.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2025 15:07
Operator : YP\AJ
Sample : AR1254ICC050
Misc :
ALS Vial : 24 Sample Multiplier: 1


Instrument :
ECD_P
ClientSampleId :
AR1254ICC050


Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/22/2025
Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 16:12:03 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 21 15:24:42 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075910.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2025 15:40
Operator : YP\AJ
Sample : AR1262ICC500
Misc :
ALS Vial : 25 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1262ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 16:21:11 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 21 16:20:51 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.647	3.785	76058633	352.0E6	50.000	50.000
2)	SA Decachlor...	10.425	8.786	61011705	448.1E6	50.000	50.000
	Target Compounds						
36)	L8 AR-1262-1	8.232	6.893	42013007	493.5E6	500.000	500.000
37)	L8 AR-1262-2	8.557	7.150	81834582	351.6E6	500.000	500.000
38)	L8 AR-1262-3	8.883	7.672	60893486	303.5E6	500.000	500.000
39)	L8 AR-1262-4	8.971	7.736	46055772	550.7E6	500.000	500.000
40)	L8 AR-1262-5	9.642	8.231	33912301	230.7E6	500.000	500.000

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


ICAL Form


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075910.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2025 15:40
Operator : YP\AJ
Sample : AR1262ICC500
Misc :
ALS Vial : 25 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1262ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 16:21:11 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 21 16:20:51 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075911.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2025 15:56
Operator : YP\AJ
Sample : AR1268ICC1000
Misc :
ALS Vial : 26 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1268ICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 18:02:42 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 21 17:59:29 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.646	3.784	151.5E6	749.9E6	97.157	99.641
2)	SA Decachlor...	10.424	8.785	198.8E6	1671.8E6	95.543	99.055
	Target Compounds						
41)	L9 AR-1268-1	8.878	7.673	186.7E6	1831.6E6	951.896	992.615
42)	L9 AR-1268-2	8.973	7.737	172.5E6	1800.4E6	956.477	1001.376
43)	L9 AR-1268-3	9.213	7.944	145.2E6	1369.8E6	961.447	988.116
44)	L9 AR-1268-4	9.640	8.230	71036137	542.8E6	963.668	1002.321
45)	L9 AR-1268-5	10.071	8.526	446.4E6	3855.5E6	968.612	979.920

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


ICAL Form


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075911.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2025 15:56
Operator : YP\AJ
Sample : AR1268ICC1000
Misc :
ALS Vial : 26 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1268ICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 18:02:42 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 21 17:59:29 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075912.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2025 16:12
Operator : YP\AJ
Sample : AR1268ICC750
Misc :
ALS Vial : 27 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1268ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 18:06:09 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 21 17:59:29 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.646	3.783	116.2E6	564.9E6	74.674	75.037
2)	SA Decachlor...	10.423	8.784	155.7E6	1249.5E6	74.863	74.351
	Target Compounds						
41)	L9 AR-1268-1	8.877	7.671	146.3E6	1360.1E6	747.184	741.339
42)	L9 AR-1268-2	8.973	7.736	134.4E6	1329.3E6	746.666	742.894
43)	L9 AR-1268-3	9.211	7.943	112.7E6	1019.0E6	747.385	739.990
44)	L9 AR-1268-4	9.640	8.229	56104264	402.6E6	757.366	745.626
45)	L9 AR-1268-5	10.070	8.525	344.2E6	2925.3E6	747.861	745.651

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


ICAL Form


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075912.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2025 16:12
Operator : YP\AJ
Sample : AR1268ICC750
Misc :
ALS Vial : 27 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1268ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 18:06:09 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 21 17:59:29 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075913.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2025 16:29
Operator : YP\AJ
Sample : AR1268ICC500
Misc :
ALS Vial : 28 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1268ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 17:59:51 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 21 17:59:29 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.647	3.785	80186895	377.7E6	50.000	50.000
2)	SA Decachlor...	10.424	8.785	108.7E6	851.9E6	50.000	50.000
	Target Compounds						
41)	L9 AR-1268-1	8.878	7.673	102.8E6	929.4E6	500.000	500.000
42)	L9 AR-1268-2	8.973	7.737	94092052	897.7E6	500.000	500.000
43)	L9 AR-1268-3	9.213	7.943	78447140	701.4E6	500.000	500.000
44)	L9 AR-1268-4	9.641	8.229	38196242	270.1E6	500.000	500.000
45)	L9 AR-1268-5	10.070	8.526	237.7E6	2006.7E6	500.000	500.000

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


ICAL Form


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075913.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2025 16:29
Operator : YP\AJ
Sample : AR1268ICC500
Misc :
ALS Vial : 28 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1268ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 17:59:51 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 21 17:59:29 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075914.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2025 16:45
Operator : YP\AJ
Sample : AR1268ICC250
Misc :
ALS Vial : 29 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1268ICC250

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 18:11:19 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 21 17:59:29 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.648	3.785	39777145	176.8E6	25.420	23.849
2)	SA Decachlor...	10.427	8.786	55292786	404.9E6	26.176	24.316
	Target Compounds						
41)	L9 AR-1268-1	8.880	7.673	52505558	433.0E6	263.374	239.361
42)	L9 AR-1268-2	8.975	7.737	47519594	415.5E6	260.422	236.426
43)	L9 AR-1268-3	9.214	7.943	39522518	339.7E6	258.947	247.491
44)	L9 AR-1268-4	9.642	8.231	19156803	125.4E6	256.397	236.500
45)	L9 AR-1268-5	10.073	8.527	120.0E6	927.0E6	257.945	239.576

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


ICAL Form


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075914.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2025 16:45
Operator : YP\AJ
Sample : AR1268ICC250
Misc :
ALS Vial : 29 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1268ICC250

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 18:11:19 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 21 17:59:29 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075915.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2025 17:01
Operator : YP\AJ
Sample : AR1268ICC050
Misc :
ALS Vial : 30 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1268ICC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 18:15:07 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 21 17:59:29 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.646	3.785	8297245	34278784	5.239	4.694
2)	SA Decachlor...	10.424	8.786	11780168	79254993	5.451	4.805
	Target Compounds						
41)	L9 AR-1268-1	8.878	7.671	11762673	90866583	56.952	50.184
42)	L9 AR-1268-2	8.973	7.736	10571240	82258397	56.152	47.408
43)	L9 AR-1268-3	9.212	7.944	8270442	67275006	53.294	49.210
44)	L9 AR-1268-4	9.641	8.228	4718687	23283603	59.998	44.997 #
45)	L9 AR-1268-5	10.071	8.525	27065230	169.1E6	56.340	44.825

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


ICAL Form


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075915.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2025 17:01
Operator : YP\AJ
Sample : AR1268ICC050
Misc :
ALS Vial : 30 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1268ICC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 18:15:07 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 21 17:59:29 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075916.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2025 17:34
Operator : YP\AJ
Sample : PP102125ICV500
Misc :
ALS Vial : 31 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
ICVPP102125

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 18:26:54 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 21 18:24:24 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.646	3.784	73895870	340.9E6	51.708	51.507
2)	SA Decachlor...	10.423	8.784	58947888	432.2E6	54.757	52.559
	Target Compounds						
3)	L1 AR-1016-1	5.799	4.879	30579076	390.9E6	517.184	528.086
4)	L1 AR-1016-2	5.820	4.937	46283521	183.5E6	523.308	536.724
5)	L1 AR-1016-3	5.883	5.057	29014100	102.8E6	524.184	500.470
6)	L1 AR-1016-4	5.980	5.097	23843232	102.7E6	524.717	549.750
7)	L1 AR-1016-5	6.273	5.312	22384300	107.6E6	515.702	472.171
31)	L7 AR-1260-1	7.391	6.340	38313477	286.2E6	557.233	522.496
32)	L7 AR-1260-2	7.644	6.526	47189702	452.3E6	526.804	555.781
33)	L7 AR-1260-3	8.002	6.680	37186118	328.9E6	536.700	584.366
34)	L7 AR-1260-4	8.230	7.149	40498377	308.8E6	590.602	594.624
35)	L7 AR-1260-5	8.556	7.389	79566551	796.8E6	537.788	587.781


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075916.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2025 17:34
Operator : YP\AJ
Sample : PP102125ICV500
Misc :
ALS Vial : 31 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
ICVPP102125

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 18:26:54 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 21 18:24:24 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075917.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2025 18:06
Operator : YP\AJ
Sample : AR1242ICV500
Misc :
ALS Vial : 32 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
ICVPP102125AR1242

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/24/2025
Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 22 08:16:42 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 21 18:24:24 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.646	3.784	79486082	365.2E6	55.620	55.173
2)	SA Decachlor...	10.423	8.785	64567530	484.9E6	59.978m	58.969
	Target Compounds						
16)	L4 AR-1242-1	5.799	4.879	24532478	301.7E6	512.088	524.891
17)	L4 AR-1242-2	5.820	4.937	37086619	138.8E6	523.883	499.981
18)	L4 AR-1242-3	5.883	5.057	22990427	78773054	503.714	500.427
19)	L4 AR-1242-4	5.980	5.139	18909181	99357590	520.064	536.743m
20)	L4 AR-1242-5	6.710	5.660	21676788	123.9E6	529.426	538.030

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075917.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2025 18:06
Operator : YP\AJ
Sample : AR1242ICV500
Misc :
ALS Vial : 32 Sample Multiplier: 1


Instrument :
ECD_P
ClientSampleId :
ICVPP102125AR1242


Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/24/2025
Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 22 08:16:42 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 21 18:24:24 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075918.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2025 18:39
Operator : YP\AJ
Sample : AR1248ICV500
Misc :
ALS Vial : 33 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
ICVPP102125AR1248

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/24/2025
Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 22 08:17:44 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 21 18:24:24 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.647	3.784	77354608	375.4E6	54.128	56.717
2)	SA Decachlor...	10.425	8.785	62090346	474.3E6	57.677	57.676
	Target Compounds						
21)	L5 AR-1248-1	5.799	4.878	18454423	193.8E6	514.472	515.973
22)	L5 AR-1248-2	6.070	5.099	24718037	124.2E6	507.095	505.124
23)	L5 AR-1248-3	6.273	5.138	27349933	132.7E6	504.009	462.616m
24)	L5 AR-1248-4	6.671	5.312	33802442	144.9E6	509.348	512.423
25)	L5 AR-1248-5	6.711	5.701	33065578	279.8E6	509.794	538.456

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075918.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2025 18:39
Operator : YP\AJ
Sample : AR1248ICV500
Misc :
ALS Vial : 33 Sample Multiplier: 1


Instrument :
ECD_P
ClientSampleId :
ICVPP102125AR1248


Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/24/2025
Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 22 08:17:44 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 21 18:24:24 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075919.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Oct 2025 08:49
Operator : YP\AJ
Sample : AR1254ICV500
Misc :
ALS Vial : 34 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
ICVPP102125AR1254

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/24/2025
Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 22 09:02:43 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 21 18:24:24 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.654	3.784	73513288	361.3E6	51.441	54.591
2)	SA Decachlor...	10.439	8.789	61994754	468.2E6	57.588	56.937
	Target Compounds						
26)	L6 AR-1254-1	6.657	5.662	33305659	337.3E6	416.876	597.997m#
27)	L6 AR-1254-2	6.873	5.811	45460472	248.5E6	454.676	468.721
28)	L6 AR-1254-3	7.235	6.214	47667437	420.5E6	458.540	494.093
29)	L6 AR-1254-4	7.517	6.439	37597676	337.3E6	449.659	535.605
30)	L6 AR-1254-5	7.934	6.855	44185805	348.6E6	482.532	484.302

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075919.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Oct 2025 08:49
Operator : YP\AJ
Sample : AR1254ICV500
Misc :
ALS Vial : 34 Sample Multiplier: 1


Instrument :
ECD_P
ClientSampleId :
ICVPP102125AR1254


Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/24/2025
Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 22 09:02:43 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 21 18:24:24 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075920.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Oct 2025 09:07
Operator : YP\AJ
Sample : AR1268ICV500
Misc :
ALS Vial : 35 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
ICVPP102125AR1268

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 24 15:40:49 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 24 15:38:49 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.648	3.782	80645176	382.7E6	50.921	52.402
2)	SA Decachlor...	10.430	8.784	112.8E6	913.8E6	52.199	55.404
	Target Compounds						
41)	L9 AR-1268-1	8.881	7.671	106.5E6	989.5E6	515.825	546.466
42)	L9 AR-1268-2	8.976	7.736	98184801	960.8E6	521.533	553.714
43)	L9 AR-1268-3	9.216	7.942	81566200	749.4E6	525.609	548.157
44)	L9 AR-1268-4	9.645	8.228	39672536	280.8E6	504.438	542.576
45)	L9 AR-1268-5	10.076	8.525	246.0E6	2111.9E6	512.030	559.930

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


ICAL Form


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075920.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Oct 2025 09:07
Operator : YP\AJ
Sample : AR1268ICV500
Misc :
ALS Vial : 35 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
ICVPP102125AR1268

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 24 15:40:49 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 24 15:38:49 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





Raw Data

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

RETENTION TIMES OF INITIAL CALIBRATION

Lab Name: Alliance **Contract:** MATR02

Lab Code: ACE **SDG NO.:** Q3450

Instrument ID: ECD_Q **Calibration Date(s):** 10/10/2025 10/10/2025

Calibration Times: 08:57 16:43

GC Column: ZB-MR1 ID: 0.32 (mm)

LAB FILE ID: RT 1000 = PQ070973.D RT 750 = PQ070974.D

RT 500 = PQ070975.D RT 250 = PQ070976.D RT 050 = PQ070977.D

COMPOUND		RT 1000	RT 750	RT 500	RT 250	RT 050	MEAN RT	RT WINDOW FROM	RT WINDOW TO
Aroclor-1016-1	(1)	4.48	4.48	4.48	4.48	4.48	4.48	4.38	4.58
Aroclor-1016-2	(2)	4.50	4.50	4.50	4.50	4.50	4.50	4.40	4.60
Aroclor-1016-3	(3)	4.56	4.56	4.56	4.56	4.56	4.56	4.46	4.66
Aroclor-1016-4	(4)	4.65	4.65	4.65	4.65	4.65	4.65	4.55	4.75
Aroclor-1016-5	(5)	4.93	4.93	4.93	4.93	4.93	4.93	4.83	5.03
Aroclor-1260-1	(1)	6.02	6.02	6.02	6.02	6.02	6.02	5.92	6.12
Aroclor-1260-2	(2)	6.28	6.28	6.28	6.28	6.27	6.28	6.18	6.38
Aroclor-1260-3	(3)	6.63	6.63	6.63	6.63	6.63	6.63	6.53	6.73
Aroclor-1260-4	(4)	6.84	6.84	6.84	6.84	6.84	6.84	6.74	6.94
Aroclor-1260-5	(5)	7.15	7.15	7.15	7.15	7.15	7.15	7.05	7.25
Decachlorobiphenyl		8.50	8.50	8.50	8.50	8.50	8.50	8.40	8.60
Tetrachloro-m-xylene		3.40	3.40	3.40	3.40	3.40	3.40	3.30	3.50
Aroclor-1242-1	(1)	4.48	4.48	4.48	4.48	4.48	4.48	4.38	4.58
Aroclor-1242-2	(2)	4.50	4.50	4.50	4.50	4.50	4.50	4.40	4.60
Aroclor-1242-3	(3)	4.56	4.56	4.56	4.56	4.56	4.56	4.46	4.66
Aroclor-1242-4	(4)	4.65	4.65	4.65	4.65	4.65	4.65	4.55	4.75
Aroclor-1242-5	(5)	5.35	5.35	5.35	5.35	5.32	5.34	5.24	5.44
Decachlorobiphenyl		8.50	8.50	8.50	8.50	8.50	8.50	8.40	8.60
Tetrachloro-m-xylene		3.40	3.40	3.40	3.40	3.40	3.40	3.30	3.50
Aroclor-1248-1	(1)	4.48	4.48	4.48	4.48	4.48	4.48	4.38	4.58
Aroclor-1248-2	(2)	4.74	4.74	4.74	4.74	4.74	4.74	4.64	4.84
Aroclor-1248-3	(3)	4.93	4.93	4.93	4.93	4.93	4.93	4.83	5.03
Aroclor-1248-4	(4)	5.32	5.32	5.32	5.32	5.32	5.32	5.22	5.42
Aroclor-1248-5	(5)	5.35	5.35	5.35	5.35	5.35	5.35	5.25	5.45
Decachlorobiphenyl		8.50	8.50	8.50	8.50	8.50	8.50	8.40	8.60
Tetrachloro-m-xylene		3.40	3.40	3.40	3.40	3.40	3.40	3.30	3.50
Aroclor-1254-1	(1)	5.29	5.29	5.29	5.29	5.29	5.29	5.19	5.39
Aroclor-1254-2	(2)	5.51	5.51	5.51	5.51	5.51	5.51	5.41	5.61
Aroclor-1254-3	(3)	5.86	5.85	5.86	5.86	5.86	5.86	5.76	5.96
Aroclor-1254-4	(4)	6.14	6.13	6.14	6.14	6.14	6.14	6.04	6.24
Aroclor-1254-5	(5)	6.54	6.54	6.54	6.54	6.54	6.54	6.44	6.64
Decachlorobiphenyl		8.50	8.50	8.50	8.50	8.50	8.50	8.40	8.60
Tetrachloro-m-xylene		3.40	3.40	3.40	3.40	3.40	3.40	3.30	3.50
Aroclor-1268-1	(1)	7.42	7.42	7.42	7.42	7.42	7.42	7.32	7.52
Aroclor-1268-2	(2)	7.49	7.50	7.50	7.50	7.50	7.50	7.40	7.60
Aroclor-1268-3	(3)	7.68	7.68	7.68	7.68	7.68	7.68	7.58	7.78
Aroclor-1268-4	(4)	7.99	7.99	7.99	7.99	7.99	7.99	7.89	8.09
Aroclor-1268-5	(5)	8.27	8.27	8.27	8.27	8.27	8.27	8.17	8.37

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

RETENTION TIMES OF INITIAL CALIBRATION

Decachlorobiphenyl	8.50	8.50	8.50	8.50	8.50	8.50	8.40	8.60
Tetrachloro-m-xylene	3.40	3.40	3.40	3.40	3.40	3.40	3.30	3.50

Raw Data

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

RETENTION TIMES OF INITIAL CALIBRATION

Lab Name: Alliance **Contract:** MATR02

Lab Code: ACE **SDG NO.:** Q3450

Instrument ID: ECD_Q **Calibration Date(s):** 10/10/2025 10/10/2025

Calibration Times: 08:57 16:43

GC Column: ZB-MR2 ID: 0.32 (mm)

LAB FILE ID: RT 1000 = PQ070973.D RT 750 = PQ070974.D

RT 500 = PQ070975.D RT 250 = PQ070976.D RT 050 = PQ070977.D

COMPOUND		RT 1000	RT 750	RT 500	RT 250	RT 050	MEAN RT	RT WINDOW FROM	RT WINDOW TO
Aroclor-1016-1	(1)	3.78	3.78	3.78	3.78	3.78	3.78	3.68	3.88
Aroclor-1016-2	(2)	3.79	3.79	3.79	3.79	3.79	3.79	3.69	3.89
Aroclor-1016-3	(3)	3.95	3.95	3.95	3.95	3.95	3.95	3.85	4.05
Aroclor-1016-4	(4)	4.00	4.00	4.00	4.00	4.00	4.00	3.90	4.10
Aroclor-1016-5	(5)	4.20	4.20	4.20	4.20	4.20	4.20	4.10	4.30
Aroclor-1260-1	(1)	5.19	5.19	5.19	5.19	5.19	5.19	5.09	5.29
Aroclor-1260-2	(2)	5.38	5.38	5.38	5.38	5.38	5.38	5.28	5.48
Aroclor-1260-3	(3)	5.52	5.52	5.52	5.52	5.52	5.52	5.42	5.62
Aroclor-1260-4	(4)	5.98	5.98	5.98	5.98	5.98	5.98	5.88	6.08
Aroclor-1260-5	(5)	6.22	6.22	6.22	6.22	6.22	6.22	6.12	6.32
Decachlorobiphenyl		7.53	7.53	7.53	7.53	7.53	7.53	7.43	7.63
Tetrachloro-m-xylene		2.78	2.78	2.78	2.78	2.78	2.78	2.68	2.88
Aroclor-1242-1	(1)	3.78	3.78	3.78	3.78	3.78	3.78	3.68	3.88
Aroclor-1242-2	(2)	3.79	3.79	3.79	3.79	3.79	3.79	3.69	3.89
Aroclor-1242-3	(3)	3.95	3.95	3.95	3.95	3.95	3.95	3.85	4.05
Aroclor-1242-4	(4)	4.04	4.04	4.04	4.04	4.04	4.04	3.94	4.14
Aroclor-1242-5	(5)	4.53	4.53	4.53	4.53	4.53	4.53	4.43	4.63
Decachlorobiphenyl		7.53	7.53	7.53	7.53	7.53	7.53	7.43	7.63
Tetrachloro-m-xylene		2.78	2.78	2.78	2.78	2.78	2.78	2.68	2.88
Aroclor-1248-1	(1)	3.78	3.78	3.78	3.78	3.78	3.78	3.68	3.88
Aroclor-1248-2	(2)	4.00	4.00	4.00	4.00	4.00	4.00	3.90	4.10
Aroclor-1248-3	(3)	4.04	4.04	4.04	4.04	4.04	4.04	3.94	4.14
Aroclor-1248-4	(4)	4.20	4.20	4.20	4.20	4.20	4.20	4.10	4.30
Aroclor-1248-5	(5)	4.57	4.57	4.57	4.57	4.57	4.57	4.47	4.67
Decachlorobiphenyl		7.53	7.53	7.53	7.53	7.53	7.53	7.43	7.63
Tetrachloro-m-xylene		2.78	2.78	2.78	2.78	2.78	2.78	2.68	2.88
Aroclor-1254-1	(1)	4.53	4.53	4.53	4.53	4.53	4.53	4.43	4.63
Aroclor-1254-2	(2)	4.68	4.68	4.68	4.68	4.68	4.68	4.58	4.78
Aroclor-1254-3	(3)	5.06	5.06	5.06	5.06	5.06	5.06	4.96	5.16
Aroclor-1254-4	(4)	5.29	5.29	5.29	5.29	5.29	5.29	5.19	5.39
Aroclor-1254-5	(5)	5.69	5.69	5.69	5.69	5.69	5.69	5.59	5.79
Decachlorobiphenyl		7.53	7.53	7.53	7.53	7.53	7.53	7.43	7.63
Tetrachloro-m-xylene		2.78	2.78	2.78	2.78	2.78	2.78	2.68	2.88
Aroclor-1268-1	(1)	6.50	6.50	6.50	6.50	6.50	6.50	6.40	6.60
Aroclor-1268-2	(2)	6.56	6.56	6.56	6.56	6.56	6.56	6.46	6.66
Aroclor-1268-3	(3)	6.76	6.76	6.76	6.76	6.76	6.76	6.66	6.86
Aroclor-1268-4	(4)	7.05	7.05	7.05	7.05	7.05	7.05	6.95	7.15
Aroclor-1268-5	(5)	7.32	7.32	7.32	7.32	7.32	7.32	7.22	7.42

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

RETENTION TIMES OF INITIAL CALIBRATION

Decachlorobiphenyl	7.53	7.53	7.53	7.53	7.53	7.53	7.43	7.63
Tetrachloro-m-xylene	2.78	2.78	2.78	2.78	2.78	2.78	2.68	2.88

Raw Data

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION FACTOR OF INITIAL CALIBRATION

Lab Name: Alliance **Contract:** MATR02

Lab Code: ACE **SDG NO.:** Q3450

Instrument ID: ECD_Q **Calibration Date(s):** 10/10/2025 10/10/2025

Calibration Times: 08:57 16:43

GC Column: ZB-MR1 **ID:** 0.32 (mm)

LAB FILE ID: **CF 1000 =** PQ070973.D **CF 750 =** PQ070974.D

CF 500 = PQ070975.D **CF 250 =** PQ070976.D **CF 050 =** PQ070977.D

COMPOUND		CF 1000	CF 750	CF 500	CF 250	CF 050	CF	% RSD
Aroclor-1016-1	(1)	152632258	161381460	169566884	175492100	161647220	164143984	5
Aroclor-1016-2	(2)	235274296	256838307	264605432	271518004	253235720	256294352	5
Aroclor-1016-3	(3)	145675938	156782704	164416996	169596776	159119400	159118363	6
Aroclor-1016-4	(4)	120162951	129378400	134848252	138303760	126953020	129929277	5
Aroclor-1016-5	(5)	117231883	128545740	133336348	136306768	128593380	128802824	6
Aroclor-1260-1	(1)	208928914	234195365	241877518	247859876	239493660	234471067	6
Aroclor-1260-2	(2)	243695794	275266251	287710272	296720404	279652200	276608984	7
Aroclor-1260-3	(3)	178017013	196424459	204653606	209590628	207855920	199308325	6
Aroclor-1260-4	(4)	191556315	214263107	223690164	228814336	216092480	214883280	7
Aroclor-1260-5	(5)	400986941	436898331	447507618	451939420	431553500	433777162	5
Decachlorobiphenyl		2603997920	2888430733	3017037660	3097201720	2913248400	2903983287	6
Tetrachloro-m-xylene		4048991050	4420325093	4447782420	4497118040	4337178200	4350278961	4
Aroclor-1242-1	(1)	117929900	120501580	121917368	124177888	110233240	118951995	5
Aroclor-1242-2	(2)	179238730	188762288	194471730	197719720	193371880	190712870	4
Aroclor-1242-3	(3)	111298009	117265485	120728326	123754708	121434700	118896246	4
Aroclor-1242-4	(4)	91702396	97376287	98944784	101196952	108558560	99555796	6
Aroclor-1242-5	(5)	63322787	67097277	82718820	102665812	77804800	78721899	20
Decachlorobiphenyl		3082568640	3243406933	3352805760	3359769360	3286790800	3265068299	3
Tetrachloro-m-xylene		4441893010	4577290613	4623065280	4605855840	4702516200	4590124189	2
Aroclor-1248-1	(1)	89951061	92802913	99278242	97399904	67891380	89464700	14
Aroclor-1248-2	(2)	126849870	131764987	138900528	142257364	142462060	136446962	5
Aroclor-1248-3	(3)	152980867	157544657	166146850	167481096	162764480	161383590	4
Aroclor-1248-4	(4)	167077290	173126060	182827440	186713196	157721500	173493097	7
Aroclor-1248-5	(5)	163667698	168136284	132211082	121389384	159185140	148917918	14
Decachlorobiphenyl		3177348440	3248518907	3402766540	3434790520	3353916400	3323468161	3
Tetrachloro-m-xylene		4546508980	4599154480	4739413180	4685828360	4518221600	4617825320	2
Aroclor-1254-1	(1)	186317660	198212097	200308250	222421948	174131700	196278331	9
Aroclor-1254-2	(2)	272465681	301191412	300558232	318683640	259516480	290483089	8
Aroclor-1254-3	(3)	294493285	320137152	318530722	335107472	267769020	307207530	9
Aroclor-1254-4	(4)	212636724	232647637	229798130	237628972	184991000	219540493	10
Aroclor-1254-5	(5)	229095618	241104957	242576814	249415576	197285560	231895705	9
Decachlorobiphenyl		3135993520	3237057933	3300114540	3432442200	2724290800	3165979799	9
Tetrachloro-m-xylene		4449551480	4584042107	4604497820	4625575480	3701892800	4393111937	9
Aroclor-1268-1	(1)	555593291	572438855	583365466	598015136	603747960	582632142	3

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION FACTOR OF INITIAL CALIBRATION

Aroclor-1268-2	(2)	510325576	524057337	534075228	546100468	539884140	530888550	3
Aroclor-1268-3	(3)	416538407	427532203	436646098	443325192	442459640	433300308	3
Aroclor-1268-4	(4)	179373252	185531188	189624920	193669728	190155160	187670850	3
Aroclor-1268-5	(5)	1283167757	1308010297	1321380362	1326514116	1318957680	1311606042	1
Decachlorobiphenyl		5757514290	5913646120	6037365900	6192444040	6223042400	6024802550	3
Tetrachloro-m-xylene		4734874420	4871269707	4916225920	4964521920	4932484000	4883875193	2

Raw Data

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION FACTOR OF INITIAL CALIBRATION

Lab Name: Alliance
Lab Code: ACE
Instrument ID: ECD_Q

Contract: MATR02
SDG NO.: Q3450
Calibration Date(s): 10/10/2025 10/10/2025
Calibration Times: 08:57 16:43

GC Column: ZB-MR2 **ID:** 0.32 (mm)

LAB FILE ID: CF 1000 = PQ070973.D CF 750 = PQ070974.D
CF 500 = PQ070975.D CF 250 = PQ070976.D CF 050 = PQ070977.D

COMPOUND		CF 1000	CF 750	CF 500	CF 250	CF 050	CF	% RSD
Aroclor-1016-1	(1)	198094739	217218485	228189888	237783008	238723760	224001976	8
Aroclor-1016-2	(2)	306974778	336928535	345611618	359075648	352024660	340123048	6
Aroclor-1016-3	(3)	164197912	180535471	187035142	195263012	182262860	181858879	6
Aroclor-1016-4	(4)	139678855	154088441	161162796	170890484	161242760	157412667	7
Aroclor-1016-5	(5)	171310856	187991799	196963118	205633060	197310680	191841903	7
Aroclor-1260-1	(1)	322104344	352779140	365742242	382158164	379245720	360405922	7
Aroclor-1260-2	(2)	392921251	432078265	450339936	466790948	457838380	439993756	7
Aroclor-1260-3	(3)	374441100	409997701	423595070	436050024	430913160	414999411	6
Aroclor-1260-4	(4)	324169048	356082777	368301826	380117112	370143240	359762801	6
Aroclor-1260-5	(5)	773923407	822481391	873089514	880740576	851175200	840282018	5
Decachlorobiphenyl		4511855410	4959881947	5165429800	5359918520	5250729800	5049563095	7
Tetrachloro-m-xylene		4790824740	5183285133	5313707940	5302586120	4809696600	5080020107	5
Aroclor-1242-1	(1)	149120136	160366853	160219200	165386980	179219740	162862582	7
Aroclor-1242-2	(2)	228406508	237737935	244055020	248772624	247903320	241375081	4
Aroclor-1242-3	(3)	123267302	131780304	132089372	135429308	136535900	131820437	4
Aroclor-1242-4	(4)	107929318	113409847	120761922	128344628	102401880	114569519	9
Aroclor-1242-5	(5)	175313049	196019013	175387368	167275268	156704980	174139936	8
Decachlorobiphenyl		5194557400	5483323560	5616478880	5698756520	6104877800	5619598832	6
Tetrachloro-m-xylene		5511103540	5712989200	5725543240	5446452960	5436162600	5566450308	3
Aroclor-1248-1	(1)	118147108	120641916	127443132	132200888	134031720	126492953	6
Aroclor-1248-2	(2)	172958749	177705395	189325858	196101732	206725040	188563355	7
Aroclor-1248-3	(3)	167593672	172934769	178216680	182062792	197467520	179655087	6
Aroclor-1248-4	(4)	207432918	214191484	230649658	232116576	237264740	224331075	6
Aroclor-1248-5	(5)	211078475	216720032	243430792	250826296	242157620	232842643	8
Decachlorobiphenyl		5357284330	5460780160	5730809660	5844375240	6130897000	5704829278	5
Tetrachloro-m-xylene		5499268270	5571294053	5722936200	5664492400	5408993600	5573396905	2
Aroclor-1254-1	(1)	326099704	341297420	345505280	377987212	295169840	337211891	9
Aroclor-1254-2	(2)	283781194	302336180	304849560	326229636	268575420	297154398	7
Aroclor-1254-3	(3)	455841488	477545104	484627186	501619228	415693040	467065209	7
Aroclor-1254-4	(4)	291449906	303039177	308578712	317073776	256115840	295251482	8
Aroclor-1254-5	(5)	417470468	427079459	439495106	452352324	365350180	420349507	8
Decachlorobiphenyl		5360763820	5577066440	5716894620	5972893640	5256420800	5576807864	5
Tetrachloro-m-xylene		5311856950	5451428933	5484489360	5522984680	4238156400	5201783265	10
Aroclor-1268-1	(1)	931455367	961712979	973652708	974423320	1030986200	974446115	4

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION FACTOR OF INITIAL CALIBRATION

Aroclor-1268-2	(2)	858872879	891158693	900594916	906701084	936258500	898717214	3
Aroclor-1268-3	(3)	746016914	769745548	779615614	791820108	821422560	781724149	4
Aroclor-1268-4	(4)	330162118	339473573	345636662	356393996	350955500	344524370	3
Aroclor-1268-5	(5)	2374681378	2422263161	2432472106	2427996872	2444443520	2420371407	1
Decachlorobiphenyl		10401085910	10736618840	10941521580	11269357960	11896046400	11048926138	5
Tetrachloro-m-xylene		5581659480	5729850200	5733063640	5828463760	5669436400	5708494696	2

Raw Data

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Lab Name: Alliance **Contract:** MATR02

Lab Code: ACE **SDG NO.:** Q3450

Instrument ID: ECD_Q **Date(s) Analyzed:** 10/10/2025 10/10/2025

GC Column: ZB-MR1 **ID:** 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW FROM	RT WINDOW TO	CALIBRATION FACTOR
Aroclor-1221	500	1	3.59	3.49	3.69	59190800
		2	3.67	3.57	3.77	43146800
		3	3.74	3.64	3.84	133548000
		4	0.00			0
		5	0.00			0
Aroclor-1232	500	1	3.74	3.64	3.84	108948000
		2	4.23	4.13	4.33	57502800
		3	4.50	4.40	4.60	107099000
		4	4.65	4.55	4.75	54014200
		5	4.79	4.69	4.89	43713200
Aroclor-1262	500	1	6.84	6.74	6.94	274018000
		2	7.15	7.05	7.25	498476000
		3	7.42	7.32	7.52	328660000
		4	7.49	7.39	7.59	252340000
		5	7.99	7.89	8.09	164350000

Raw Data

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Lab Name: Alliance **Contract:** MATR02

Lab Code: ACE **SDG NO.:** Q3450

Instrument ID: ECD_Q **Date(s) Analyzed:** 10/10/2025 10/10/2025

GC Column: ZB-MR2 **ID:** 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW FROM	RT WINDOW TO	CALIBRATION FACTOR
Aroclor-1221	500	1	2.98	2.88	3.08	73819200
		2	3.05	2.95	3.15	60003200
		3	3.12	3.02	3.22	171184000
		4	0.00			0
		5	0.00			0
Aroclor-1232	500	1	3.12	3.02	3.22	139342000
		2	3.79	3.69	3.89	138900000
		3	3.95	3.85	4.05	73450200
		4	4.04	3.94	4.14	63034400
		5	4.20	4.10	4.30	71607200
Aroclor-1262	500	1	5.73	5.63	5.83	452486000
		2	5.98	5.88	6.08	427702000
		3	6.50	6.40	6.60	332708000
		4	6.56	6.46	6.66	616820000
		5	7.05	6.95	7.15	302986000

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ070973.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 08:57
Operator : YP\AJ
Sample : AR1660ICC1000
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1660ICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 10:54:32 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 10 10:49:25 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	3.401	2.778	404.9E6	479.1E6	95.307	94.825
2)	SA Decachlor...	8.496	7.534	260.4E6	451.2E6	92.652	93.246
	Target Compounds						
3)	L1 AR-1016-1	4.480	3.778	152.6E6	198.1E6	947.441	929.401
4)	L1 AR-1016-2	4.499	3.793	235.3E6	307.0E6	941.324	940.794
5)	L1 AR-1016-3	4.556	3.954	145.7E6	164.2E6	939.563	934.980
6)	L1 AR-1016-4	4.646	4.002	120.2E6	139.7E6	942.413	928.587
7)	L1 AR-1016-5	4.931	4.197	117.2E6	171.3E6	935.728	930.345
31)	L7 AR-1260-1	6.020	5.190	208.9E6	322.1E6	926.912	936.559
32)	L7 AR-1260-2	6.275	5.379	243.7E6	392.9E6	917.174	931.909
33)	L7 AR-1260-3	6.626	5.520	178.0E6	374.4E6	930.393	938.406
34)	L7 AR-1260-4	6.840	5.980	191.6E6	324.2E6	922.615	936.268
35)	L7 AR-1260-5	7.146	6.224	401.0E6	773.9E6	945.173	939.790


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ070973.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 08:57
Operator : YP\AJ
Sample : AR1660ICC1000
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1660ICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 10:54:32 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 10 10:49:25 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ070974.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 09:11
Operator : YP\AJ
Sample : AR1660ICC750
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1660ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 10:57:39 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 10 10:49:25 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	3.400	2.778	331.5E6	388.7E6	76.997	76.286
2)	SA Decachlor...	8.496	7.534	216.6E6	372.0E6	76.373	76.242
	Target Compounds						
3)	L1 AR-1016-1	4.480	3.778	121.0E6	162.9E6	750.874	759.502
4)	L1 AR-1016-2	4.499	3.794	192.6E6	252.7E6	763.674	766.122
5)	L1 AR-1016-3	4.556	3.954	117.6E6	135.4E6	755.578	763.875
6)	L1 AR-1016-4	4.646	4.002	97033800	115.6E6	757.308	762.093
7)	L1 AR-1016-5	4.930	4.197	96409305	141.0E6	762.905	760.395
31)	L7 AR-1260-1	6.020	5.189	175.6E6	264.6E6	769.253	762.765
32)	L7 AR-1260-2	6.276	5.378	206.4E6	324.1E6	767.783	762.288
33)	L7 AR-1260-3	6.627	5.519	147.3E6	307.5E6	763.182	763.633
34)	L7 AR-1260-4	6.840	5.980	160.7E6	267.1E6	765.822	764.087
35)	L7 AR-1260-5	7.146	6.224	327.7E6	616.9E6	764.763	749.377

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


ICAL Form


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ070974.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 09:11
Operator : YP\AJ
Sample : AR1660ICC750
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1660ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 10:57:39 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 10 10:49:25 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ070975.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 09:26
Operator : YP\AJ
Sample : AR1660ICC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1660ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 10:49:52 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 10 10:49:25 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	3.400	2.778	222.4E6	265.7E6	50.000	50.000
2)	SA Decachlor...	8.496	7.533	150.9E6	258.3E6	50.000	50.000
	Target Compounds						
3)	L1 AR-1016-1	4.480	3.778	84783442	114.1E6	500.000	500.000
4)	L1 AR-1016-2	4.499	3.793	132.3E6	172.8E6	500.000	500.000
5)	L1 AR-1016-3	4.556	3.953	82208498	93517571	500.000	500.000
6)	L1 AR-1016-4	4.646	4.002	67424126	80581398	500.000	500.000
7)	L1 AR-1016-5	4.931	4.197	66668174	98481559	500.000	500.000
31)	L7 AR-1260-1	6.019	5.189	120.9E6	182.9E6	500.000	500.000
32)	L7 AR-1260-2	6.275	5.378	143.9E6	225.2E6	500.000	500.000
33)	L7 AR-1260-3	6.626	5.519	102.3E6	211.8E6	500.000	500.000
34)	L7 AR-1260-4	6.840	5.979	111.8E6	184.2E6	500.000	500.000
35)	L7 AR-1260-5	7.145	6.224	223.8E6	436.5E6	500.000	500.000

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


ICAL Form


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ070975.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 09:26
Operator : YP\AJ
Sample : AR1660ICC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1660ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 10:49:52 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 10 10:49:25 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ070976.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 09:40
Operator : YP\AJ
Sample : AR1660ICC250
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1660ICC250

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 11:01:33 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 10 10:49:25 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	3.400	2.778	112.4E6	132.6E6	25.824	25.753
2)	SA Decachlor...	8.495	7.534	77430043	134.0E6	26.685	26.803
	Target Compounds						
3)	L1 AR-1016-1	4.480	3.778	43873025	59445752	266.271	269.814
4)	L1 AR-1016-2	4.499	3.794	67879501	89768912	264.062	266.260
5)	L1 AR-1016-3	4.557	3.954	42399194	48815753	266.464	268.576
6)	L1 AR-1016-4	4.646	4.003	34575940	42722621	264.598	273.066
7)	L1 AR-1016-5	4.931	4.197	34076692	51408265	264.457	269.895
31)	L7 AR-1260-1	6.019	5.190	61964969	95539541	265.698	268.599
32)	L7 AR-1260-2	6.275	5.379	74180101	116.7E6	268.916	267.943
33)	L7 AR-1260-3	6.626	5.520	52397657	109.0E6	265.747	265.224
34)	L7 AR-1260-4	6.840	5.980	57203584	95029278	266.583	266.063
35)	L7 AR-1260-5	7.146	6.224	113.0E6	220.2E6	260.134	262.889

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


ICAL Form


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ070976.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 09:40
Operator : YP\AJ
Sample : AR1660ICC250
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1660ICC250

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 11:01:33 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 10 10:49:25 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ070977.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 09:55
Operator : YP\AJ
Sample : AR1660ICC050
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1660ICC050

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/13/2025
Supervised By :mohammad ahmed 10/14/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 11:05:13 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 10 10:49:25 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	3.400	2.778	21685891	24048483	4.985	4.734
2)	SA Decachlor...	8.495	7.533	14566242	26253649	5.016	5.199
	Target Compounds						
3)	L1 AR-1016-1	4.480	3.778	8082361	11936188	49.239	53.286
4)	L1 AR-1016-2	4.498	3.793	12661786	17601233	49.403	51.750
5)	L1 AR-1016-3	4.557	3.953	7955970	9113143	50.000	50.111
6)	L1 AR-1016-4	4.645	4.002	6347651	8062138	48.855	51.217
7)	L1 AR-1016-5	4.930	4.197	6429669	9865534	49.919	51.425
31)	L7 AR-1260-1	6.019	5.190	11974683	18962286	51.162m	52.614
32)	L7 AR-1260-2	6.274	5.378	13982610	22891919	50.322m	52.028
33)	L7 AR-1260-3	6.626	5.519	10392796	21545658	52.144	51.917
34)	L7 AR-1260-4	6.840	5.980	10804624	18507162	50.281	51.443
35)	L7 AR-1260-5	7.145	6.224	21577675	42558760	49.744	50.648

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ070977.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 09:55
Operator : YP\AJ
Sample : AR1660ICC050
Misc :
ALS Vial : 7 Sample Multiplier: 1


Instrument :
ECD_Q
ClientSampleId :
AR1660ICC050


Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/13/2025
Supervised By :mohammad ahmed 10/14/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 11:05:13 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 10 10:49:25 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ070978.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 10:09
Operator : YP\AJ
Sample : AR1221ICC500
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1221ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 11:55:54 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 10 11:55:37 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	3.400	2.778	220.4E6	255.3E6	50.000	50.000
2)	SA Decachlor...	8.496	7.534	149.9E6	262.1E6	50.000	50.000
	Target Compounds						
8)	L2 AR-1221-1	3.594	2.977	29595379	36909578	500.000	500.000
9)	L2 AR-1221-2	3.672	3.053	21573372	30001568	500.000	500.000
10)	L2 AR-1221-3	3.739	3.122	66773962	85591891	500.000	500.000

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


ICAL Form


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ070978.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 10:09
Operator : YP\AJ
Sample : AR1221ICC500
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1221ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 11:55:54 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 10 11:55:37 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ070979.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 10:24
Operator : YP\AJ
Sample : AR1232ICC500
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1232ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 12:00:39 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 10 11:55:37 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	3.400	2.778	234.5E6	276.3E6	50.000	50.000
2)	SA Decachlor...	8.497	7.534	163.6E6	280.6E6	50.000	50.000
	Target Compounds						
11)	L3 AR-1232-1	3.739	3.121	54474047	69671226	500.000	500.000
12)	L3 AR-1232-2	4.226	3.793	28751418	69449831	500.000	500.000
13)	L3 AR-1232-3	4.499	3.953	53549480	36725094	500.000	500.000
14)	L3 AR-1232-4	4.646	4.039	27007118	31517168	500.000	500.000
15)	L3 AR-1232-5	4.785	4.197	21856633	35803624	500.000	500.000


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ070979.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 10:24
Operator : YP\AJ
Sample : AR1232ICC500
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1232ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 12:00:39 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 10 11:55:37 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ070980.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 10:38
Operator : YP\AJ
Sample : AR1242ICC1000
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1242ICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 12:08:51 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 10 12:05:14 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	3.400	2.778	444.2E6	551.1E6	98.001	98.092
2)	SA Decachlor...	8.495	7.534	308.3E6	519.5E6	95.801	96.097
	Target Compounds						
16)	L4 AR-1242-1	4.480	3.778	117.9E6	149.1E6	983.375	964.120
17)	L4 AR-1242-2	4.499	3.793	179.2E6	228.4E6	959.238	966.879
18)	L4 AR-1242-3	4.556	3.953	111.3E6	123.3E6	959.357	965.452
19)	L4 AR-1242-4	4.646	4.039	91702396	107.9E6	962.012	943.887
20)	L4 AR-1242-5	5.349	4.534	63322787	175.3E6	867.188	999.788


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ070980.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 10:38
Operator : YP\AJ
Sample : AR1242ICC1000
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1242ICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 12:08:51 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 10 12:05:14 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ070981.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 10:53
Operator : YP\AJ
Sample : AR1242ICC750
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1242ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 12:12:26 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 10 12:10:08 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	3.400	2.778	343.3E6	428.5E6	75.493	75.838
2)	SA Decachlor...	8.496	7.534	243.3E6	411.2E6	75.399	75.716
	Target Compounds						
16)	L4 AR-1242-1	4.480	3.778	90376185	120.3E6	752.406	768.194
17)	L4 AR-1242-2	4.498	3.794	141.6E6	178.3E6	755.086	753.183
18)	L4 AR-1242-3	4.556	3.954	87949114	98835228	755.378	765.893
19)	L4 AR-1242-4	4.646	4.039	73032215	85057385	760.690	745.897
20)	L4 AR-1242-5	5.348	4.534	50322958	147.0E6	708.312	806.708


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ070981.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 10:53
Operator : YP\AJ
Sample : AR1242ICC750
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1242ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 12:12:26 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 10 12:10:08 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ070982.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 11:08
Operator : YP\AJ
Sample : AR1242ICC500
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1242ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 12:05:29 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 10 12:05:14 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	3.400	2.778	231.2E6	286.3E6	50.000	50.000
2)	SA Decachlor...	8.496	7.533	167.6E6	280.8E6	50.000	50.000
	Target Compounds						
16)	L4 AR-1242-1	4.480	3.778	60958684	80109600	500.000	500.000
17)	L4 AR-1242-2	4.499	3.793	97235865	122.0E6	500.000	500.000
18)	L4 AR-1242-3	4.556	3.954	60364163	66044686	500.000	500.000
19)	L4 AR-1242-4	4.646	4.040	49472392	60380961	500.000	500.000
20)	L4 AR-1242-5	5.351	4.534	41359410	87693684	500.000	500.000

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


ICAL Form


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ070982.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 11:08
Operator : YP\AJ
Sample : AR1242ICC500
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1242ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 12:05:29 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 10 12:05:14 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ070983.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 11:22
Operator : YP\AJ
Sample : AR1242ICC250
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1242ICC250

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 12:15:55 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 10 12:10:08 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	3.401	2.778	115.1E6	136.2E6	25.240	24.319
2)	SA Decachlor...	8.497	7.534	83994234	142.5E6	25.768	25.912
	Target Compounds						
16)	L4 AR-1242-1	4.481	3.778	31044472	41346745	256.287	260.414
17)	L4 AR-1242-2	4.499	3.794	49429930	62193156	260.092	259.416
18)	L4 AR-1242-3	4.557	3.954	30938677	33857327	261.612	259.162
19)	L4 AR-1242-4	4.647	4.040	25299238	32086157	259.999	272.815
20)	L4 AR-1242-5	5.352	4.534	25666453	41818817	325.093	234.281 #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


ICAL Form


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ070983.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 11:22
Operator : YP\AJ
Sample : AR1242ICC250
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1242ICC250

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 12:15:55 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 10 12:10:08 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ070984.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 11:37
Operator : YP\AJ
Sample : AR1242ICC050
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1242ICC050

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/13/2025
Supervised By :mohammad ahmed 10/14/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 12:20:34 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 10 12:10:08 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	3.400	2.778	23512581	27180813	5.122	4.883
2)	SA Decachlor...	8.496	7.534	16433954	30524389	5.033	5.432
	Target Compounds						
16)	L4 AR-1242-1	4.481	3.777	5511662	8960987	46.335	57.042m
17)	L4 AR-1242-2	4.499	3.793	9668594	12395166	50.697	51.291m
18)	L4 AR-1242-3	4.556	3.953	6071735	6826795	51.068	51.803m
19)	L4 AR-1242-4	4.646	4.039	5427928	5120094	54.521	44.579m
20)	L4 AR-1242-5	5.317	4.534	3890240	7835249	49.488m	45.137m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ070984.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 11:37
Operator : YP\AJ
Sample : AR1242ICC050
Misc :
ALS Vial : 14 Sample Multiplier: 1


Instrument :
ECD_Q
ClientSampleId :
AR1242ICC050


Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/13/2025
Supervised By :mohammad ahmed 10/14/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 12:20:34 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 10 12:10:08 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ070985.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 11:51
Operator : YP\AJ
Sample : AR1248ICC1000
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1248ICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 12:40:50 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 10 12:37:43 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	3.400	2.778	454.7E6	549.9E6	97.923	98.007
2)	SA Decachlor...	8.496	7.534	317.7E6	535.7E6	96.574	96.631
	Target Compounds						
21)	L5 AR-1248-1	4.480	3.778	89951061	118.1E6	950.710	962.148
22)	L5 AR-1248-2	4.743	4.002	126.8E6	173.0E6	954.654	954.823
23)	L5 AR-1248-3	4.930	4.039	153.0E6	167.6E6	958.744	969.281
24)	L5 AR-1248-4	5.318	4.197	167.1E6	207.4E6	954.987	947.004
25)	L5 AR-1248-5	5.353	4.572	163.7E6	211.1E6	1106.316	928.819

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


ICAL Form


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ070985.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 11:51
Operator : YP\AJ
Sample : AR1248ICC1000
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1248ICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 12:40:50 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 10 12:37:43 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ070986.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 12:06
Operator : YP\AJ
Sample : AR1248ICC750
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1248ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 12:46:34 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 10 12:37:43 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	3.400	2.778	344.9E6	417.8E6	74.527	74.644
2)	SA Decachlor...	8.495	7.533	243.6E6	409.6E6	74.366	74.245
	Target Compounds						
21)	L5 AR-1248-1	4.480	3.777	69602185	90481437	740.364	741.181
22)	L5 AR-1248-2	4.744	4.002	98823740	133.3E6	745.811	740.453
23)	L5 AR-1248-3	4.931	4.039	118.2E6	129.7E6	743.646	750.086
24)	L5 AR-1248-4	5.318	4.197	129.8E6	160.6E6	744.762	738.847
25)	L5 AR-1248-5	5.352	4.571	126.1E6	162.5E6	815.290	726.458

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


ICAL Form


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ070986.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 12:06
Operator : YP\AJ
Sample : AR1248ICC750
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1248ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 12:46:34 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 10 12:37:43 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ070987.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 12:20
Operator : YP\AJ
Sample : AR1248ICC500
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1248ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 12:37:58 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 10 12:37:43 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	3.400	2.778	237.0E6	286.1E6	50.000	50.000
2)	SA Decachlor...	8.496	7.532	170.1E6	286.5E6	50.000	50.000
	Target Compounds						
21)	L5 AR-1248-1	4.480	3.778	49639121	63721566	500.000	500.000
22)	L5 AR-1248-2	4.744	4.002	69450264	94662929	500.000	500.000
23)	L5 AR-1248-3	4.931	4.039	83073425	89108340	500.000	500.000
24)	L5 AR-1248-4	5.318	4.197	91413720	115.3E6	500.000	500.000
25)	L5 AR-1248-5	5.351	4.571	66105541	121.7E6	500.000	500.000

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


ICAL Form


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ070987.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 12:20
Operator : YP\AJ
Sample : AR1248ICC500
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1248ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 12:37:58 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 10 12:37:43 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ070988.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 12:35
Operator : YP\AJ
Sample : AR1248ICC250
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1248ICC250

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/13/2025
Supervised By :mohammad ahmed 10/14/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 12:50:45 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 10 12:37:43 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	3.401	2.778	117.1E6	141.6E6	25.232	25.223
2)	SA Decachlor...	8.497	7.534	85869763	146.1E6	25.897	26.099
	Target Compounds						
21)	L5 AR-1248-1	4.481	3.778	24349976	33050222	256.699	265.233
22)	L5 AR-1248-2	4.744	4.002	35564341	49025433	263.550	266.409
23)	L5 AR-1248-3	4.931	4.039	41870274	45515698	260.002	259.790
24)	L5 AR-1248-4	5.318	4.196	46678299	58029144	263.071	262.398m
25)	L5 AR-1248-5	5.351	4.570	30347346	62706574	207.360	273.971m#

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ070988.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 12:35
Operator : YP\AJ
Sample : AR1248ICC250
Misc :
ALS Vial : 18 Sample Multiplier: 1


Instrument :
ECD_Q
ClientSampleId :
AR1248ICC250


Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/13/2025
Supervised By :mohammad ahmed 10/14/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 12:50:45 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 10 12:37:43 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ070989.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 12:50
Operator : YP\AJ
Sample : AR1248ICC050
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1248ICC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 14:31:10 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 10 14:30:55 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	3.400	2.778	22591108	27044968	4.892	4.853
2)	SA Decachlor...	8.497	7.534	16769582	30654485	5.046	5.373
	Target Compounds						
21)	L5 AR-1248-1	4.481	3.778	3394569	6701586	37.943	52.980 #
22)	L5 AR-1248-2	4.743	4.002	7123103	10336252	52.204	54.816
23)	L5 AR-1248-3	4.931	4.039	8138224	9873376	50.428	54.957
24)	L5 AR-1248-4	5.318	4.197	7886075	11863237	45.455	52.883
25)	L5 AR-1248-5	5.352	4.571	7959257	12107881	53.447	52.000

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


ICAL Form


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ070989.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 12:50
Operator : YP\AJ
Sample : AR1248ICC050
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1248ICC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 14:31:10 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 10 14:30:55 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ070990.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 13:04
Operator : YP\AJ
Sample : AR1254ICC1000
Misc :
ALS Vial : 20 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1254ICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 14:53:43 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 10 14:52:12 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	3.400	2.778	445.0E6	531.2E6	98.289	98.401
2)	SA Decachlor...	8.496	7.533	313.6E6	536.1E6	97.450	96.785
	Target Compounds						
26)	L6 AR-1254-1	5.292	4.533	186.3E6	326.1E6	963.814	971.106
27)	L6 AR-1254-2	5.507	4.680	272.5E6	283.8E6	950.975	964.208
28)	L6 AR-1254-3	5.855	5.060	294.5E6	455.8E6	960.789	969.392
29)	L6 AR-1254-4	6.135	5.286	212.6E6	291.4E6	961.211	971.453
30)	L6 AR-1254-5	6.543	5.690	229.1E6	417.5E6	971.418	974.299

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


ICAL Form


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ070990.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 13:04
Operator : YP\AJ
Sample : AR1254ICC1000
Misc :
ALS Vial : 20 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1254ICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 14:53:43 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 10 14:52:12 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ070991.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 13:19
Operator : YP\AJ
Sample : AR1254ICC750
Misc :
ALS Vial : 21 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1254ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 14:59:55 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 10 14:59:41 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	3.401	2.778	343.8E6	408.9E6	75.627	75.492
2)	SA Decachlor...	8.496	7.534	242.8E6	418.3E6	75.295	75.344
	Target Compounds						
26)	L6 AR-1254-1	5.293	4.533	148.7E6	256.0E6	762.565	758.137
27)	L6 AR-1254-2	5.507	4.680	225.9E6	226.8E6	775.187	763.504
28)	L6 AR-1254-3	5.854	5.060	240.1E6	358.2E6	771.902	757.733
29)	L6 AR-1254-4	6.134	5.286	174.5E6	227.3E6	775.397	755.024
30)	L6 AR-1254-5	6.544	5.690	180.8E6	320.3E6	761.088	748.361

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


ICAL Form


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ070991.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 13:19
Operator : YP\AJ
Sample : AR1254ICC750
Misc :
ALS Vial : 21 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1254ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 14:59:55 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 10 14:59:41 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ070992.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 13:33
Operator : YP\AJ
Sample : AR1254ICC500
Misc :
ALS Vial : 22 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1254ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 14:49:59 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 10 14:49:02 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	3.400	2.778	230.2E6	274.2E6	50.000	50.000
2)	SA Decachlor...	8.497	7.534	165.0E6	285.8E6	50.000	50.000
	Target Compounds						
26)	L6 AR-1254-1	5.293	4.532	100.2E6	172.8E6	500.000	500.000
27)	L6 AR-1254-2	5.507	4.680	150.3E6	152.4E6	500.000	500.000
28)	L6 AR-1254-3	5.855	5.060	159.3E6	242.3E6	500.000	500.000
29)	L6 AR-1254-4	6.136	5.286	114.9E6	154.3E6	500.000	500.000
30)	L6 AR-1254-5	6.543	5.689	121.3E6	219.7E6	500.000	500.000

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


ICAL Form


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ070992.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 13:33
Operator : YP\AJ
Sample : AR1254ICC500
Misc :
ALS Vial : 22 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1254ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 14:49:59 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 10 14:49:02 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ070993.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 13:48
Operator : YP\AJ
Sample : AR1254ICC250
Misc :
ALS Vial : 23 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1254ICC250

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 15:02:42 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 10 15:02:26 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	3.401	2.778	115.6E6	138.1E6	25.327	25.369
2)	SA Decachlor...	8.497	7.534	85811055	149.3E6	26.191	26.396
	Target Compounds						
26)	L6 AR-1254-1	5.293	4.533	55605487	94496803	275.527	271.759
27)	L6 AR-1254-2	5.507	4.679	79670910	81557409	267.151	268.017
28)	L6 AR-1254-3	5.855	5.060	83776868	125.4E6	264.224	261.310
29)	L6 AR-1254-4	6.135	5.286	59407243	79268444	260.355	259.866
30)	L6 AR-1254-5	6.544	5.690	62353894	113.1E6	259.216	260.512

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


ICAL Form


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ070993.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 13:48
Operator : YP\AJ
Sample : AR1254ICC250
Misc :
ALS Vial : 23 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1254ICC250

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 15:02:42 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 10 15:02:26 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ070994.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 15:15
Operator : YP\AJ
Sample : AR1254ICC050
Misc :
ALS Vial : 24 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1254ICC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 15:31:28 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 10 15:31:02 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	3.400	2.777	18509464	21190782	4.213	4.074
2)	SA Decachlor...	8.497	7.533	13621454	26282104	4.302	4.713
	Target Compounds						
26)	L6 AR-1254-1	5.294	4.533	8706585	14758492	44.358	43.766
27)	L6 AR-1254-2	5.507	4.679	12975824	13428771	44.670	45.191
28)	L6 AR-1254-3	5.856	5.060	13388451	20784652	43.581	44.501
29)	L6 AR-1254-4	6.135	5.286	9249550	12805792	42.131	43.372
30)	L6 AR-1254-5	6.544	5.690	9864278	18267509	42.538	43.458

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


ICAL Form


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ070994.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 15:15
Operator : YP\AJ
Sample : AR1254ICC050
Misc :
ALS Vial : 24 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1254ICC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 15:31:28 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 10 15:31:02 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ070995.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 15:30
Operator : YP\AJ
Sample : AR1262ICC500
Misc :
ALS Vial : 25 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1262ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 15:45:51 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 10 15:45:32 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	3.400	2.778	233.7E6	278.1E6	50.000	50.000
2)	SA Decachlor...	8.497	7.533	168.1E6	302.5E6	50.000	50.000
	Target Compounds						
36)	L8 AR-1262-1	6.841	5.733	137.0E6	226.2E6	500.000	500.000
37)	L8 AR-1262-2	7.146	5.979	249.2E6	213.9E6	500.000	500.000
38)	L8 AR-1262-3	7.419	6.500	164.3E6	166.4E6	500.000	500.000
39)	L8 AR-1262-4	7.492	6.561	126.2E6	308.4E6	500.000	500.000
40)	L8 AR-1262-5	7.988	7.051	82174920	151.5E6	500.000	500.000

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


ICAL Form


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ070995.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 15:30
Operator : YP\AJ
Sample : AR1262ICC500
Misc :
ALS Vial : 25 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1262ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 15:45:51 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 10 15:45:32 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ070996.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 15:44
Operator : YP\AJ
Sample : AR1268ICC1000
Misc :
ALS Vial : 26 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1268ICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 16:44:44 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 10 16:41:42 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	3.400	2.778	473.5E6	558.2E6	98.121	98.662
2)	SA Decachlor...	8.496	7.533	575.8E6	1040.1E6	97.627	97.468
	Target Compounds						
41)	L9 AR-1268-1	7.418	6.501	555.6E6	931.5E6	975.616	977.850
42)	L9 AR-1268-2	7.494	6.563	510.3E6	858.9E6	977.260	976.287
43)	L9 AR-1268-3	7.676	6.763	416.5E6	746.0E6	976.432	977.977
44)	L9 AR-1268-4	7.988	7.051	179.4E6	330.2E6	972.218	977.102
45)	L9 AR-1268-5	8.271	7.318	1283.2E6	2374.7E6	985.329	987.978

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


ICAL Form


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ070996.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 15:44
Operator : YP\AJ
Sample : AR1268ICC1000
Misc :
ALS Vial : 26 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1268ICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 16:44:44 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 10 16:41:42 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ070997.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 15:59
Operator : YP\AJ
Sample : AR1268ICC750
Misc :
ALS Vial : 27 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1268ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 16:47:10 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 10 16:41:42 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	3.400	2.778	365.3E6	429.7E6	75.472	75.638
2)	SA Decachlor...	8.496	7.533	443.5E6	805.2E6	75.137	75.305
	Target Compounds						
41)	L9 AR-1268-1	7.417	6.501	429.3E6	721.3E6	752.594	754.792
42)	L9 AR-1268-2	7.495	6.564	393.0E6	668.4E6	751.776	756.465
43)	L9 AR-1268-3	7.677	6.763	320.6E6	577.3E6	751.101	754.528
44)	L9 AR-1268-4	7.988	7.051	139.1E6	254.6E6	752.792	752.326
45)	L9 AR-1268-5	8.270	7.318	981.0E6	1816.7E6	752.199	753.877

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


ICAL Form


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ070997.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 15:59
Operator : YP\AJ
Sample : AR1268ICC750
Misc :
ALS Vial : 27 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1268ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 16:47:10 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 10 16:41:42 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ070998.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 16:13
Operator : YP\AJ
Sample : AR1268ICC500
Misc :
ALS Vial : 28 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1268ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 16:41:57 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 10 16:41:42 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	3.401	2.778	245.8E6	286.7E6	50.000	50.000
2)	SA Decachlor...	8.497	7.533	301.9E6	547.1E6	50.000	50.000
	Target Compounds						
41)	L9 AR-1268-1	7.418	6.501	291.7E6	486.8E6	500.000	500.000
42)	L9 AR-1268-2	7.496	6.564	267.0E6	450.3E6	500.000	500.000
43)	L9 AR-1268-3	7.677	6.763	218.3E6	389.8E6	500.000	500.000
44)	L9 AR-1268-4	7.988	7.051	94812460	172.8E6	500.000	500.000
45)	L9 AR-1268-5	8.271	7.318	660.7E6	1216.2E6	500.000	500.000

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


ICAL Form


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ070998.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 16:13
Operator : YP\AJ
Sample : AR1268ICC500
Misc :
ALS Vial : 28 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1268ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 16:41:57 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 10 16:41:42 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ070999.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 16:28
Operator : YP\AJ
Sample : AR1268ICC250
Misc :
ALS Vial : 29 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1268ICC250

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 16:50:34 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 10 16:41:42 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	3.400	2.778	124.1E6	145.7E6	25.476	25.482
2)	SA Decachlor...	8.496	7.533	154.8E6	281.7E6	25.909	25.997
	Target Compounds						
41)	L9 AR-1268-1	7.418	6.500	149.5E6	243.6E6	258.947	253.674
42)	L9 AR-1268-2	7.495	6.563	136.5E6	226.7E6	258.257	254.883
43)	L9 AR-1268-3	7.677	6.763	110.8E6	198.0E6	257.143	256.485
44)	L9 AR-1268-4	7.988	7.051	48417432	89098499	258.848	259.826
45)	L9 AR-1268-5	8.270	7.317	331.6E6	607.0E6	253.196	251.413

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


ICAL Form


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ070999.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 16:28
Operator : YP\AJ
Sample : AR1268ICC250
Misc :
ALS Vial : 29 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1268ICC250

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 16:50:34 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 10 16:41:42 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ071000.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 16:43
Operator : YP\AJ
Sample : AR1268ICC050
Misc :
ALS Vial : 30 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1268ICC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 17:16:32 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 10 17:16:19 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	3.400	2.778	24662420	28347182	5.050	4.966
2)	SA Decachlor...	8.497	7.533	31115212	59480232	5.165	5.383
	Target Compounds						
41)	L9 AR-1268-1	7.418	6.500	30187398	51549310	51.812	52.901
42)	L9 AR-1268-2	7.495	6.563	26994207	46812925	50.847	52.089
43)	L9 AR-1268-3	7.676	6.763	22122982	41071128	51.057	52.539
44)	L9 AR-1268-4	7.989	7.050	9507758	17547775	50.662	50.933
45)	L9 AR-1268-5	8.271	7.317	65947884	122.2E6	50.280	50.497


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ071000.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 16:43
Operator : YP\AJ
Sample : AR1268ICC050
Misc :
ALS Vial : 30 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1268ICC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 17:16:32 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 10 17:16:19 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ071001.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 16:57
Operator : YP\AJ
Sample : PQ101025ICV
Misc :
ALS Vial : 31 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
ICVPQ101025

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 10/31/2025
Supervised By :Yogesh Patel 10/31/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 17:29:38 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 10 17:24:38 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	3.400	2.778	237.3E6	265.1E6	54.558	52.183
2)	SA Decachlor...	8.497	7.533	168.1E6	302.1E6	57.872	59.825
	Target Compounds						
3)	L1 AR-1016-1	4.480	3.777	84536347	121.7E6	515.013	543.285m
4)	L1 AR-1016-2	4.499	3.792	137.9E6	157.0E6	537.936	461.544m
5)	L1 AR-1016-3	4.556	3.953	86370882	89605925	542.809	492.722m
6)	L1 AR-1016-4	4.646	4.001	72647096	77616175	559.128	493.075m
7)	L1 AR-1016-5	4.931	4.196	71678763	95334887	556.500	496.945m
31)	L7 AR-1260-1	6.019	5.188	124.6E6	179.0E6	531.360m	496.698m
32)	L7 AR-1260-2	6.275	5.377	157.7E6	216.4E6	569.990m	491.786m
33)	L7 AR-1260-3	6.627	5.518	115.0E6	195.1E6	576.989	470.083
34)	L7 AR-1260-4	6.840	5.979	126.1E6	186.2E6	586.767	517.623
35)	L7 AR-1260-5	7.146	6.224	259.3E6	433.0E6	597.827	515.349

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ071001.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 16:57
Operator : YP\AJ
Sample : PQ101025ICV
Misc :
ALS Vial : 31 Sample Multiplier: 1


Instrument :
ECD_Q
ClientSampleId :
ICVPQ101025


Manual Integrations APPROVED

Reviewed By :Rahul Chavli 10/31/2025
Supervised By :Yogesh Patel 10/31/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 17:29:38 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 10 17:24:38 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ071002.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 17:41
Operator : YP\AJ
Sample : AR1242ICV500
Misc :
ALS Vial : 32 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
ICVPQ101025AR1242

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/13/2025
Supervised By :mohammad ahmed 10/14/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 11 01:19:16 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Oct 11 01:16:43 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	3.400	2.778	255.6E6	301.0E6	55.692	54.076
2)	SA Decachlor...	8.496	7.534	182.2E6	328.4E6	55.793	58.440
	Target Compounds						
16)	L4 AR-1242-1	4.480	3.778	67446984	88802083	567.010	545.258
17)	L4 AR-1242-2	4.499	3.793	106.5E6	130.9E6	558.407	542.137
18)	L4 AR-1242-3	4.556	3.952	65846450	69528801	553.814	527.451m
19)	L4 AR-1242-4	4.646	4.039	55324188	66246241	555.710	578.219
20)	L4 AR-1242-5	5.349	4.534	35965359	88786139	456.866	509.855

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ071002.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 17:41
Operator : YP\AJ
Sample : AR1242ICV500
Misc :
ALS Vial : 32 Sample Multiplier: 1


Instrument :
ECD_Q
ClientSampleId :
ICVPQ101025AR1242


Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/13/2025
Supervised By :mohammad ahmed 10/14/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 11 01:19:16 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Oct 11 01:16:43 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ071003.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 18:25
Operator : YP\AJ
Sample : AR1248ICV500
Misc :
ALS Vial : 33 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
ICVPQ101025AR1248

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/13/2025
Supervised By :mohammad ahmed 10/14/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 11 01:26:29 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Oct 11 01:23:25 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	3.400	2.778	253.0E6	292.6E6	54.792	52.505
2)	SA Decachlor...	8.497	7.534	179.8E6	324.3E6	54.114	56.854
	Target Compounds						
21)	L5 AR-1248-1	4.480	3.777	47223685	65282112	527.847	516.093
22)	L5 AR-1248-2	4.743	4.002	73173417	97020820	536.277	514.526
23)	L5 AR-1248-3	4.931	4.039	87732864	87534945	543.629	487.239
24)	L5 AR-1248-4	5.317	4.196	98470443	117.2E6	567.576m	522.551m
25)	L5 AR-1248-5	5.349	4.571	59958766	126.5E6	402.630m	543.081 #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ071003.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 18:25
Operator : YP\AJ
Sample : AR1248ICV500
Misc :
ALS Vial : 33 Sample Multiplier: 1


Instrument :
ECD_Q
ClientSampleId :
ICVPQ101025AR1248


Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/13/2025
Supervised By :mohammad ahmed 10/14/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 11 01:26:29 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Oct 11 01:23:25 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ071004.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 19:08
Operator : YP\AJ
Sample : AR1254ICV500
Misc :
ALS Vial : 34 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
ICVPQ101025AR1254

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 11 01:34:52 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Oct 11 01:30:57 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	3.400	2.778	247.0E6	287.2E6	56.214	55.217
2)	SA Decachlor...	8.498	7.533	175.2E6	316.3E6	55.323	56.717
	Target Compounds						
26)	L6 AR-1254-1	5.293	4.533	115.3E6	191.1E6	587.461	566.643
27)	L6 AR-1254-2	5.507	4.680	167.9E6	162.6E6	578.152	547.137
28)	L6 AR-1254-3	5.855	5.060	176.3E6	253.6E6	573.790	542.980
29)	L6 AR-1254-4	6.134	5.286	124.7E6	161.4E6	568.016	546.628
30)	L6 AR-1254-5	6.544	5.690	130.6E6	228.4E6	563.253	543.396

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


ICAL Form


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ071004.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 19:08
Operator : YP\AJ
Sample : AR1254ICV500
Misc :
ALS Vial : 34 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
ICVPQ101025AR1254

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 11 01:34:52 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Oct 11 01:30:57 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ071005.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 19:52
Operator : YP\AJ
Sample : AR1268ICV500
Misc :
ALS Vial : 35 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
ICVPQ101025AR1268

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 11 01:39:57 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Oct 11 01:37:12 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	3.401	2.777	256.6E6	296.1E6	52.549	51.872
2)	SA Decachlor...	8.497	7.533	317.0E6	560.4E6	52.621	50.721
	Target Compounds						
41)	L9 AR-1268-1	7.418	6.500	307.3E6	498.7E6	527.367	511.825
42)	L9 AR-1268-2	7.495	6.563	283.4E6	464.0E6	533.785	516.248
43)	L9 AR-1268-3	7.677	6.763	233.9E6	395.9E6	539.709	506.489
44)	L9 AR-1268-4	7.989	7.051	100.4E6	175.1E6	534.727	508.315
45)	L9 AR-1268-5	8.271	7.318	696.8E6	1236.0E6	531.266	510.673

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


ICAL Form


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ071005.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 19:52
Operator : YP\AJ
Sample : AR1268ICV500
Misc :
ALS Vial : 35 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
ICVPQ101025AR1268

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 11 01:39:57 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Oct 11 01:37:12 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





Raw Data: **PO115031.D**

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** MATR02

Lab Code: ACE **SDG NO.:** Q3450

Continuing Calib Date: 11/12/2025 **Initial Calibration Date(s):** 11/04/2025 11/04/2025

Continuing Calib Time: 09:49 **Initial Calibration Time(s):** 09:39 18:31

GC Column: ZB-MR1 **ID:** 0.32 **(mm)**

COMPOUND		CCAL RT	AVG RT	RT WINDOW FROM	RT WINDOW TO	DIFF RT
Aroclor-1016-1	(1)	5.49	5.49	5.39	5.59	0.00
Aroclor-1016-2	(2)	5.51	5.51	5.41	5.61	0.00
Aroclor-1016-3	(3)	5.57	5.57	5.47	5.67	0.00
Aroclor-1016-4	(4)	5.67	5.67	5.57	5.77	0.01
Aroclor-1016-5	(5)	5.96	5.96	5.86	6.06	0.00
Aroclor-1260-1	(1)	7.07	7.07	6.97	7.17	0.00
Aroclor-1260-2	(2)	7.32	7.33	7.23	7.43	0.01
Aroclor-1260-3	(3)	7.68	7.68	7.58	7.78	0.00
Aroclor-1260-4	(4)	7.90	7.90	7.80	8.00	0.00
Aroclor-1260-5	(5)	8.21	8.21	8.11	8.31	0.00
Tetrachloro-m-xylene		4.35	4.35	4.25	4.45	0.00
Decachlorobiphenyl		9.93	9.93	9.83	10.03	0.00

Raw Data: **PO115031.D**

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** MATR02

Lab Code: ACE **SDG NO.:** Q3450

Continuing Calib Date: 11/12/2025 **Initial Calibration Date(s):** 11/04/2025 11/04/2025

Continuing Calib Time: 09:49 **Initial Calibration Time(s):** 09:39 18:31

GC Column: ZB-MR2 **ID:** 0.32 **(mm)**

COMPOUND		CCAL RT	AVG RT	RT WINDOW FROM	TO	DIFF RT
Aroclor-1016-1	(1)	4.73	4.73	4.63	4.83	0.00
Aroclor-1016-2	(2)	4.75	4.75	4.65	4.85	0.01
Aroclor-1016-3	(3)	4.92	4.92	4.82	5.02	0.00
Aroclor-1016-4	(4)	4.96	4.96	4.86	5.06	0.00
Aroclor-1016-5	(5)	5.17	5.18	5.08	5.28	0.01
Aroclor-1260-1	(1)	6.21	6.21	6.11	6.31	0.01
Aroclor-1260-2	(2)	6.39	6.39	6.29	6.49	0.00
Aroclor-1260-3	(3)	6.55	6.55	6.45	6.65	0.00
Aroclor-1260-4	(4)	7.02	7.02	6.92	7.12	0.01
Aroclor-1260-5	(5)	7.26	7.26	7.16	7.36	0.00
Tetrachloro-m-xylene		3.65	3.65	3.55	3.75	0.00
Decachlorobiphenyl		8.64	8.64	8.54	8.74	0.00

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** MATR02

Lab Code: ACE **SDG NO.:** Q3450

GC Column: ZB-MR1 **ID:** 0.32 **(mm)** **Initi. Calib. Date(s):** 11/04/2025 11/04/2025

Client Sample No.: CCAL01 **Date Analyzed:** 11/12/2025

Lab Sample No.: AR1660CCC500 **Data File :** PO115031.D **Time Analyzed:** 09:49

COMPOUND	RT	RT WINDOW FROM	RT WINDOW TO	CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	5.486	5.387	5.587	502.390	500.000	0.5
Aroclor-1016-2	5.507	5.408	5.608	500.300	500.000	0.1
Aroclor-1016-3	5.569	5.470	5.670	514.140	500.000	2.8
Aroclor-1016-4	5.665	5.566	5.766	503.350	500.000	0.7
Aroclor-1016-5	5.957	5.858	6.058	492.960	500.000	-1.4
Aroclor-1260-1	7.070	6.971	7.171	480.960	500.000	-3.8
Aroclor-1260-2	7.324	7.225	7.425	516.090	500.000	3.2
Aroclor-1260-3	7.682	7.582	7.782	515.890	500.000	3.2
Aroclor-1260-4	7.903	7.803	8.003	499.810	500.000	0.0
Aroclor-1260-5	8.210	8.110	8.310	497.030	500.000	-0.6
Decachlorobiphenyl	9.932	9.830	10.030	49.750	50.000	-0.5
Tetrachloro-m-xylene	4.346	4.248	4.448	51.820	50.000	3.6

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** MATR02

Lab Code: ACE **SDG NO.:** Q3450

GC Column: ZB-MR2 **ID:** 0.32 **(mm)** **Initi. Calib. Date(s):** 11/04/2025 11/04/2025

Client Sample No.: CCAL01 **Date Analyzed:** 11/12/2025

Lab Sample No.: AR1660CCC500 **Data File :** PO115031.D **Time Analyzed:** 09:49

COMPOUND	RT	RT WINDOW FROM	RT WINDOW TO	CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.726	4.628	4.828	494.480	500.000	-1.1
Aroclor-1016-2	4.745	4.647	4.847	491.490	500.000	-1.7
Aroclor-1016-3	4.919	4.821	5.021	501.010	500.000	0.2
Aroclor-1016-4	4.962	4.864	5.064	502.510	500.000	0.5
Aroclor-1016-5	5.174	5.076	5.276	500.560	500.000	0.1
Aroclor-1260-1	6.205	6.106	6.306	472.430	500.000	-5.5
Aroclor-1260-2	6.392	6.294	6.494	465.770	500.000	-6.8
Aroclor-1260-3	6.545	6.446	6.646	471.710	500.000	-5.7
Aroclor-1260-4	7.015	6.917	7.117	480.570	500.000	-3.9
Aroclor-1260-5	7.257	7.158	7.358	476.520	500.000	-4.7
Decachlorobiphenyl	8.639	8.541	8.741	51.660	50.000	3.3
Tetrachloro-m-xylene	3.648	3.548	3.748	49.940	50.000	-0.1

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO111225\
Data File : PO115031.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Nov 2025 09:49
Operator : YP/AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 12 21:38:21 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Nov 05 06:00:07 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.346	3.648	364.7E6	223.1E6	51.822	49.942
2)	SA Decachlor...	9.932	8.639	219.9E6	201.7E6	49.750	51.659
	Target Compounds						
3)	L1 AR-1016-1	5.486	4.726	123.6E6	94077877	502.391	494.482
4)	L1 AR-1016-2	5.507	4.745	198.7E6	133.1E6	500.298	491.494
5)	L1 AR-1016-3	5.569	4.919	114.7E6	74136345	514.136	501.012
6)	L1 AR-1016-4	5.665	4.962	95070428	59003676	503.351	502.510
7)	L1 AR-1016-5	5.957	5.174	87843698	76442839	492.958	500.564
31)	L7 AR-1260-1	7.070	6.205	147.0E6	139.9E6	480.964	472.433
32)	L7 AR-1260-2	7.324	6.392	204.3E6	200.3E6	516.087	465.774
33)	L7 AR-1260-3	7.682	6.545	171.2E6	165.6E6	515.885	471.710
34)	L7 AR-1260-4	7.903	7.015	152.5E6	145.6E6	499.811	480.571
35)	L7 AR-1260-5	8.210	7.257	354.5E6	391.5E6	497.032	476.521


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO111225\
Data File : PO115031.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Nov 2025 09:49
Operator : YP/AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 12 21:38:21 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Nov 05 06:00:07 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





Raw Data: **PO115046.D**

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** MATR02

Lab Code: ACE **SDG NO.:** Q3450

Continuing Calib Date: 11/12/2025 **Initial Calibration Date(s):** 11/04/2025 11/04/2025

Continuing Calib Time: 15:55 **Initial Calibration Time(s):** 09:39 18:31

GC Column: ZB-MR1 **ID:** 0.32 **(mm)**

COMPOUND		CCAL RT	AVG RT	RT WINDOW FROM	RT WINDOW TO	DIFF RT
Aroclor-1016-1	(1)	5.49	5.49	5.39	5.59	0.00
Aroclor-1016-2	(2)	5.51	5.51	5.41	5.61	0.00
Aroclor-1016-3	(3)	5.57	5.57	5.47	5.67	0.00
Aroclor-1016-4	(4)	5.67	5.67	5.57	5.77	0.00
Aroclor-1016-5	(5)	5.96	5.96	5.86	6.06	0.00
Aroclor-1260-1	(1)	7.07	7.07	6.97	7.17	0.00
Aroclor-1260-2	(2)	7.33	7.33	7.23	7.43	0.01
Aroclor-1260-3	(3)	7.68	7.68	7.58	7.78	0.00
Aroclor-1260-4	(4)	7.90	7.90	7.80	8.00	0.00
Aroclor-1260-5	(5)	8.21	8.21	8.11	8.31	0.00
Tetrachloro-m-xylene		4.35	4.35	4.25	4.45	0.00
Decachlorobiphenyl		9.93	9.93	9.83	10.03	0.00

Raw Data: **PO115046.D**

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** MATR02

Lab Code: ACE **SDG NO.:** Q3450

Continuing Calib Date: 11/12/2025 **Initial Calibration Date(s):** 11/04/2025 11/04/2025

Continuing Calib Time: 15:55 **Initial Calibration Time(s):** 09:39 18:31

GC Column: ZB-MR2 **ID:** 0.32 **(mm)**

COMPOUND		CCAL RT	AVG RT	RT WINDOW FROM	RT WINDOW TO	DIFF RT
Aroclor-1016-1	(1)	4.73	4.73	4.63	4.83	0.00
Aroclor-1016-2	(2)	4.75	4.75	4.65	4.85	0.00
Aroclor-1016-3	(3)	4.92	4.92	4.82	5.02	0.00
Aroclor-1016-4	(4)	4.96	4.96	4.86	5.06	0.00
Aroclor-1016-5	(5)	5.18	5.18	5.08	5.28	0.00
Aroclor-1260-1	(1)	6.20	6.21	6.11	6.31	0.01
Aroclor-1260-2	(2)	6.39	6.39	6.29	6.49	0.00
Aroclor-1260-3	(3)	6.55	6.55	6.45	6.65	0.00
Aroclor-1260-4	(4)	7.02	7.02	6.92	7.12	0.00
Aroclor-1260-5	(5)	7.26	7.26	7.16	7.36	0.00
Tetrachloro-m-xylene		3.65	3.65	3.55	3.75	0.00
Decachlorobiphenyl		8.64	8.64	8.54	8.74	0.00

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** MATR02

Lab Code: ACE **SDG NO.:** Q3450

GC Column: ZB-MR1 **ID:** 0.32 **(mm)** **Initi. Calib. Date(s):** 11/04/2025 11/04/2025

Client Sample No.: CCAL02 **Date Analyzed:** 11/12/2025

Lab Sample No.: AR1660CCC500 **Data File :** PO115046.D **Time Analyzed:** 15:55

COMPOUND	RT	RT WINDOW FROM	RT WINDOW TO	CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	5.487	5.387	5.587	486.740	500.000	-2.7
Aroclor-1016-2	5.509	5.408	5.608	467.970	500.000	-6.4
Aroclor-1016-3	5.570	5.470	5.670	485.340	500.000	-2.9
Aroclor-1016-4	5.666	5.566	5.766	460.290	500.000	-7.9
Aroclor-1016-5	5.958	5.858	6.058	456.330	500.000	-8.7
Aroclor-1260-1	7.071	6.971	7.171	467.020	500.000	-6.6
Aroclor-1260-2	7.325	7.225	7.425	495.090	500.000	-1.0
Aroclor-1260-3	7.682	7.582	7.782	497.620	500.000	-0.5
Aroclor-1260-4	7.904	7.803	8.003	484.630	500.000	-3.1
Aroclor-1260-5	8.212	8.110	8.310	487.270	500.000	-2.5
Decachlorobiphenyl	9.932	9.830	10.030	48.700	50.000	-2.6
Tetrachloro-m-xylene	4.347	4.248	4.448	49.330	50.000	-1.3

Raw Data: **PO115046.D**

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** MATR02

Lab Code: ACE **SDG NO.:** Q3450

GC Column: ZB-MR2 **ID:** 0.32 **(mm)** **Initi. Calib. Date(s):** 11/04/2025 11/04/2025

Client Sample No.: CCAL02 **Date Analyzed:** 11/12/2025

Lab Sample No.: AR1660CCC500 **Data File :** PO115046.D **Time Analyzed:** 15:55

COMPOUND	RT	RT WINDOW FROM	RT WINDOW TO	CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.726	4.628	4.828	463.320	500.000	-7.3
Aroclor-1016-2	4.746	4.647	4.847	469.030	500.000	-6.2
Aroclor-1016-3	4.920	4.821	5.021	468.660	500.000	-6.3
Aroclor-1016-4	4.962	4.864	5.064	475.910	500.000	-4.8
Aroclor-1016-5	5.175	5.076	5.276	483.800	500.000	-3.2
Aroclor-1260-1	6.204	6.106	6.306	453.970	500.000	-9.2
Aroclor-1260-2	6.392	6.294	6.494	457.030	500.000	-8.6
Aroclor-1260-3	6.545	6.446	6.646	463.220	500.000	-7.4
Aroclor-1260-4	7.016	6.917	7.117	464.740	500.000	-7.1
Aroclor-1260-5	7.257	7.158	7.358	467.270	500.000	-6.5
Decachlorobiphenyl	8.640	8.541	8.741	50.800	50.000	1.6
Tetrachloro-m-xylene	3.648	3.548	3.748	47.940	50.000	-4.1

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO111225\
Data File : PO115046.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Nov 2025 15:55
Operator : YP/AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 12 21:50:17 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Nov 05 06:00:07 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.347	3.648	347.2E6	214.2E6	49.332	47.945m
2)	SA Decachlor...	9.932	8.640	215.3E6	198.3E6	48.698	50.798
	Target Compounds						
3)	L1 AR-1016-1	5.487	4.726	119.7E6	88148279	486.736	463.315
4)	L1 AR-1016-2	5.509	4.746	185.8E6	127.0E6	467.973	469.026
5)	L1 AR-1016-3	5.570	4.920	108.2E6	69349031	485.344	468.659
6)	L1 AR-1016-4	5.666	4.962	86937096	55880539	460.289	475.912
7)	L1 AR-1016-5	5.958	5.175	81315893	73882104	456.325	483.796
31)	L7 AR-1260-1	7.071	6.204	142.7E6	134.4E6	467.022	453.965
32)	L7 AR-1260-2	7.325	6.392	196.0E6	196.5E6	495.092	457.026
33)	L7 AR-1260-3	7.682	6.545	165.1E6	162.7E6	497.618	463.225
34)	L7 AR-1260-4	7.904	7.016	147.8E6	140.8E6	484.625	464.742
35)	L7 AR-1260-5	8.212	7.257	347.5E6	383.9E6	487.265	467.269


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO111225\
Data File : PO115046.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Nov 2025 15:55
Operator : YP/AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 12 21:50:17 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Nov 05 06:00:07 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





Raw Data: **PP076042.D**

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** MATR02

Lab Code: ACE **SDG NO.:** Q3450

Continuing Calib Date: 10/24/2025 **Initial Calibration Date(s):** 10/21/2025 10/21/2025

Continuing Calib Time: 18:58 **Initial Calibration Time(s):** 09:06 17:01

GC Column: ZB-MR1 **ID:** 0.32 **(mm)**

COMPOUND		CCAL RT	AVG RT	RT WINDOW FROM	RT WINDOW TO	DIFF RT
Aroclor-1016-1	(1)	5.79	5.80	5.70	5.90	0.01
Aroclor-1016-2	(2)	5.81	5.82	5.72	5.92	0.01
Aroclor-1016-3	(3)	5.88	5.88	5.78	5.98	0.00
Aroclor-1016-4	(4)	5.97	5.98	5.88	6.08	0.01
Aroclor-1016-5	(5)	6.27	6.27	6.17	6.37	0.00
Aroclor-1260-1	(1)	7.38	7.39	7.29	7.49	0.01
Aroclor-1260-2	(2)	7.64	7.64	7.54	7.74	0.00
Aroclor-1260-3	(3)	8.00	8.00	7.90	8.10	0.01
Aroclor-1260-4	(4)	8.22	8.23	8.13	8.33	0.01
Aroclor-1260-5	(5)	8.55	8.55	8.45	8.65	0.00
Tetrachloro-m-xylene		4.64	4.64	4.54	4.74	0.00
Decachlorobiphenyl		10.41	10.42	10.32	10.52	0.01

Raw Data: **PP076042.D**

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** MATR02

Lab Code: ACE **SDG NO.:** Q3450

Continuing Calib Date: 10/24/2025 **Initial Calibration Date(s):** 10/21/2025 10/21/2025

Continuing Calib Time: 18:58 **Initial Calibration Time(s):** 09:06 17:01

GC Column: ZB-MR2 **ID:** 0.32 **(mm)**

COMPOUND		CCAL RT	AVG RT	RT WINDOW FROM	TO	DIFF RT
Aroclor-1016-1	(1)	4.87	4.88	4.78	4.98	0.01
Aroclor-1016-2	(2)	4.93	4.94	4.84	5.04	0.01
Aroclor-1016-3	(3)	5.05	5.05	4.95	5.15	0.00
Aroclor-1016-4	(4)	5.09	5.10	5.00	5.20	0.01
Aroclor-1016-5	(5)	5.31	5.31	5.21	5.41	0.00
Aroclor-1260-1	(1)	6.33	6.34	6.24	6.44	0.01
Aroclor-1260-2	(2)	6.52	6.53	6.43	6.63	0.01
Aroclor-1260-3	(3)	6.67	6.68	6.58	6.78	0.01
Aroclor-1260-4	(4)	7.14	7.15	7.05	7.25	0.01
Aroclor-1260-5	(5)	7.38	7.39	7.29	7.49	0.01
Tetrachloro-m-xylene		3.78	3.78	3.68	3.88	0.00
Decachlorobiphenyl		8.78	8.78	8.68	8.88	0.00

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** MATR02

Lab Code: ACE **SDG NO.:** Q3450

GC Column: ZB-MR1 **ID:** 0.32 **(mm)** **Initi. Calib. Date(s):** 10/21/2025 10/21/2025

Client Sample No.: CCAL03 **Date Analyzed:** 10/24/2025

Lab Sample No.: AR1660CCC500 **Data File :** PP076042.D **Time Analyzed:** 18:58

COMPOUND	RT	RT WINDOW FROM	RT WINDOW TO	CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	5.792	5.695	5.895	466.180	500.000	-6.8
Aroclor-1016-2	5.814	5.717	5.917	470.850	500.000	-5.8
Aroclor-1016-3	5.876	5.780	5.980	471.910	500.000	-5.6
Aroclor-1016-4	5.974	5.877	6.077	475.240	500.000	-5.0
Aroclor-1016-5	6.266	6.169	6.369	472.190	500.000	-5.6
Aroclor-1260-1	7.383	7.287	7.487	508.760	500.000	1.8
Aroclor-1260-2	7.637	7.540	7.740	477.450	500.000	-4.5
Aroclor-1260-3	7.995	7.899	8.099	498.370	500.000	-0.3
Aroclor-1260-4	8.222	8.127	8.327	531.900	500.000	6.4
Aroclor-1260-5	8.547	8.452	8.652	490.190	500.000	-2.0
Decachlorobiphenyl	10.411	10.319	10.519	49.890	50.000	-0.2
Tetrachloro-m-xylene	4.640	4.542	4.742	45.590	50.000	-8.8

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** MATR02

Lab Code: ACE **SDG NO.:** Q3450

GC Column: ZB-MR2 **ID:** 0.32 **(mm)** **Initi. Calib. Date(s):** 10/21/2025 10/21/2025

Client Sample No.: CCAL03 **Date Analyzed:** 10/24/2025

Lab Sample No.: AR1660CCC500 **Data File :** PP076042.D **Time Analyzed:** 18:58

COMPOUND	RT	RT WINDOW FROM	RT WINDOW TO	CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.874	4.778	4.978	465.970	500.000	-6.8
Aroclor-1016-2	4.933	4.836	5.036	486.720	500.000	-2.7
Aroclor-1016-3	5.052	4.954	5.154	475.940	500.000	-4.8
Aroclor-1016-4	5.093	4.996	5.196	498.570	500.000	-0.3
Aroclor-1016-5	5.306	5.210	5.410	466.880	500.000	-6.6
Aroclor-1260-1	6.334	6.238	6.438	497.200	500.000	-0.6
Aroclor-1260-2	6.521	6.425	6.625	510.950	500.000	2.2
Aroclor-1260-3	6.674	6.578	6.778	514.410	500.000	2.9
Aroclor-1260-4	7.142	7.047	7.247	563.770	500.000	12.8
Aroclor-1260-5	7.383	7.287	7.487	531.590	500.000	6.3
Decachlorobiphenyl	8.776	8.682	8.882	47.410	50.000	-5.2
Tetrachloro-m-xylene	3.779	3.681	3.881	46.610	50.000	-6.8

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102425\
Data File : PP076042.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 24 Oct 2025 18:58
Operator : YP\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 24 23:38:56 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 24 15:40:32 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.640	3.779	65154729	308.5E6	45.592	46.609
2)	SA Decachlor...	10.411	8.776	53708533	389.8E6	49.891	47.409
	Target Compounds						
3)	L1 AR-1016-1	5.792	4.874	27563278	344.9E6	466.178	465.975
4)	L1 AR-1016-2	5.814	4.933	41643707	166.4E6	470.848	486.715
5)	L1 AR-1016-3	5.876	5.052	26120781	97730819	471.911	475.941
6)	L1 AR-1016-4	5.974	5.093	21594789	93175401	475.235	498.569
7)	L1 AR-1016-5	6.266	5.306	20495798	106.4E6	472.193	466.880
31)	L7 AR-1260-1	7.383	6.334	34980707	272.3E6	508.761	497.196
32)	L7 AR-1260-2	7.637	6.521	42768798	415.8E6	477.451	510.946
33)	L7 AR-1260-3	7.995	6.674	34530147	289.5E6	498.367	514.409
34)	L7 AR-1260-4	8.222	7.142	36473340	292.8E6	531.903	563.770
35)	L7 AR-1260-5	8.547	7.383	72524197	720.6E6	490.189	531.589


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102425\
Data File : PP076042.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 24 Oct 2025 18:58
Operator : YP\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 24 23:38:56 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 24 15:40:32 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





Raw Data: **PP076057.D**

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** MATR02

Lab Code: ACE **SDG NO.:** Q3450

Continuing Calib Date: 10/25/2025 **Initial Calibration Date(s):** 10/21/2025 10/21/2025

Continuing Calib Time: 00:22 **Initial Calibration Time(s):** 09:06 17:01

GC Column: ZB-MR1 **ID:** 0.32 **(mm)**

COMPOUND		CCAL RT	AVG RT	RT WINDOW FROM	RT WINDOW TO	DIFF RT
Aroclor-1016-1	(1)	5.80	5.80	5.70	5.90	0.00
Aroclor-1016-2	(2)	5.82	5.82	5.72	5.92	0.00
Aroclor-1016-3	(3)	5.88	5.88	5.78	5.98	0.00
Aroclor-1016-4	(4)	5.98	5.98	5.88	6.08	0.00
Aroclor-1016-5	(5)	6.27	6.27	6.17	6.37	0.00
Aroclor-1260-1	(1)	7.39	7.39	7.29	7.49	0.00
Aroclor-1260-2	(2)	7.64	7.64	7.54	7.74	0.00
Aroclor-1260-3	(3)	8.00	8.00	7.90	8.10	0.00
Aroclor-1260-4	(4)	8.23	8.23	8.13	8.33	0.00
Aroclor-1260-5	(5)	8.55	8.55	8.45	8.65	0.00
Tetrachloro-m-xylene		4.64	4.64	4.54	4.74	0.00
Decachlorobiphenyl		10.41	10.42	10.32	10.52	0.01

Raw Data: **PP076057.D**

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** MATR02

Lab Code: ACE **SDG NO.:** Q3450

Continuing Calib Date: 10/25/2025 **Initial Calibration Date(s):** 10/21/2025 10/21/2025

Continuing Calib Time: 00:22 **Initial Calibration Time(s):** 09:06 17:01

GC Column: ZB-MR2 **ID:** 0.32 **(mm)**

COMPOUND		CCAL RT	AVG RT	RT WINDOW FROM	TO	DIFF RT
Aroclor-1016-1	(1)	4.88	4.88	4.78	4.98	0.00
Aroclor-1016-2	(2)	4.94	4.94	4.84	5.04	0.00
Aroclor-1016-3	(3)	5.06	5.05	4.95	5.15	-0.01
Aroclor-1016-4	(4)	5.10	5.10	5.00	5.20	0.00
Aroclor-1016-5	(5)	5.31	5.31	5.21	5.41	0.00
Aroclor-1260-1	(1)	6.34	6.34	6.24	6.44	0.00
Aroclor-1260-2	(2)	6.52	6.53	6.43	6.63	0.01
Aroclor-1260-3	(3)	6.68	6.68	6.58	6.78	0.00
Aroclor-1260-4	(4)	7.15	7.15	7.05	7.25	0.00
Aroclor-1260-5	(5)	7.39	7.39	7.29	7.49	0.00
Tetrachloro-m-xylene		3.78	3.78	3.68	3.88	0.00
Decachlorobiphenyl		8.78	8.78	8.68	8.88	0.00

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** MATR02

Lab Code: ACE **SDG NO.:** Q3450

GC Column: ZB-MR1 **ID:** 0.32 **(mm)** **Initi. Calib. Date(s):** 10/21/2025 10/21/2025

Client Sample No.: CCAL04 **Date Analyzed:** 10/25/2025

Lab Sample No.: AR1660CCC500 **Data File :** PP076057.D **Time Analyzed:** 00:22

COMPOUND	RT	RT WINDOW FROM	RT WINDOW TO	CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	5.795	5.695	5.895	471.470	500.000	-5.7
Aroclor-1016-2	5.816	5.717	5.917	468.300	500.000	-6.3
Aroclor-1016-3	5.879	5.780	5.980	468.920	500.000	-6.2
Aroclor-1016-4	5.976	5.877	6.077	474.310	500.000	-5.1
Aroclor-1016-5	6.268	6.169	6.369	471.100	500.000	-5.8
Aroclor-1260-1	7.386	7.287	7.487	527.420	500.000	5.5
Aroclor-1260-2	7.639	7.540	7.740	488.330	500.000	-2.3
Aroclor-1260-3	7.997	7.899	8.099	510.250	500.000	2.1
Aroclor-1260-4	8.225	8.127	8.327	539.530	500.000	7.9
Aroclor-1260-5	8.549	8.452	8.652	497.530	500.000	-0.5
Decachlorobiphenyl	10.412	10.319	10.519	50.390	50.000	0.8
Tetrachloro-m-xylene	4.643	4.542	4.742	46.220	50.000	-7.6

Raw Data: **PP076057.D**

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** MATR02

Lab Code: ACE **SDG NO.:** Q3450

GC Column: ZB-MR2 **ID:** 0.32 **(mm)** **Initi. Calib. Date(s):** 10/21/2025 10/21/2025

Client Sample No.: CCAL04 **Date Analyzed:** 10/25/2025

Lab Sample No.: AR1660CCC500 **Data File :** PP076057.D **Time Analyzed:** 00:22

COMPOUND	RT	RT WINDOW FROM	RT WINDOW TO	CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.878	4.778	4.978	475.300	500.000	-4.9
Aroclor-1016-2	4.937	4.836	5.036	506.010	500.000	1.2
Aroclor-1016-3	5.056	4.954	5.154	486.710	500.000	-2.7
Aroclor-1016-4	5.097	4.996	5.196	501.120	500.000	0.2
Aroclor-1016-5	5.311	5.210	5.410	470.680	500.000	-5.9
Aroclor-1260-1	6.337	6.238	6.438	513.790	500.000	2.8
Aroclor-1260-2	6.524	6.425	6.625	515.860	500.000	3.2
Aroclor-1260-3	6.678	6.578	6.778	518.110	500.000	3.6
Aroclor-1260-4	7.146	7.047	7.247	568.000	500.000	13.6
Aroclor-1260-5	7.386	7.287	7.487	533.710	500.000	6.7
Decachlorobiphenyl	8.779	8.682	8.882	46.800	50.000	-6.4
Tetrachloro-m-xylene	3.782	3.681	3.881	46.920	50.000	-6.2

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102425\
Data File : PP076057.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 25 Oct 2025 00:22
Operator : YP\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 25 05:33:15 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 24 15:40:32 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.643	3.782	66049864	310.6E6	46.218	46.925
2)	SA Decachlor...	10.412	8.779	54250357	384.8E6	50.394	46.799
	Target Compounds						
3)	L1 AR-1016-1	5.795	4.878	27876164	351.8E6	471.470	475.300
4)	L1 AR-1016-2	5.816	4.937	41418265	173.0E6	468.299	506.011
5)	L1 AR-1016-3	5.879	5.056	25955144	99942279	468.919	486.710
6)	L1 AR-1016-4	5.976	5.097	21552666	93652976	474.308	501.125
7)	L1 AR-1016-5	6.268	5.311	20448172	107.3E6	471.096	470.680
31)	L7 AR-1260-1	7.386	6.337	36263643	281.4E6	527.420	513.786
32)	L7 AR-1260-2	7.639	6.524	43743655	419.8E6	488.334	515.864
33)	L7 AR-1260-3	7.997	6.678	35353730	291.6E6	510.254	518.109
34)	L7 AR-1260-4	8.225	7.146	36996195	295.0E6	539.528	568.004
35)	L7 AR-1260-5	8.549	7.386	73609646	723.5E6	497.525	533.711


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102425\
Data File : PP076057.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 25 Oct 2025 00:22
Operator : YP\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 25 05:33:15 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 24 15:40:32 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





Raw Data: **PP076063.D**

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** MATR02

Lab Code: ACE **SDG NO.:** Q3450

Continuing Calib Date: 10/27/2025 **Initial Calibration Date(s):** 10/21/2025 10/21/2025

Continuing Calib Time: 08:56 **Initial Calibration Time(s):** 09:06 17:01

GC Column: ZB-MR1 **ID:** 0.32 **(mm)**

COMPOUND		CCAL RT	AVG RT	RT WINDOW FROM	RT WINDOW TO	DIFF RT
Aroclor-1016-1	(1)	5.79	5.80	5.70	5.90	0.01
Aroclor-1016-2	(2)	5.81	5.82	5.72	5.92	0.01
Aroclor-1016-3	(3)	5.87	5.88	5.78	5.98	0.01
Aroclor-1016-4	(4)	5.97	5.98	5.88	6.08	0.01
Aroclor-1016-5	(5)	6.26	6.27	6.17	6.37	0.01
Aroclor-1260-1	(1)	7.38	7.39	7.29	7.49	0.01
Aroclor-1260-2	(2)	7.63	7.64	7.54	7.74	0.01
Aroclor-1260-3	(3)	7.99	8.00	7.90	8.10	0.01
Aroclor-1260-4	(4)	8.22	8.23	8.13	8.33	0.01
Aroclor-1260-5	(5)	8.54	8.55	8.45	8.65	0.01
Tetrachloro-m-xylene		4.64	4.64	4.54	4.74	0.00
Decachlorobiphenyl		10.41	10.42	10.32	10.52	0.02

Raw Data: **PP076063.D**

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** MATR02

Lab Code: ACE **SDG NO.:** Q3450

Continuing Calib Date: 10/27/2025 **Initial Calibration Date(s):** 10/21/2025 10/21/2025

Continuing Calib Time: 08:56 **Initial Calibration Time(s):** 09:06 17:01

GC Column: ZB-MR2 **ID:** 0.32 **(mm)**

COMPOUND		CCAL RT	AVG RT	RT WINDOW FROM	TO	DIFF RT
Aroclor-1016-1	(1)	4.87	4.88	4.78	4.98	0.01
Aroclor-1016-2	(2)	4.93	4.94	4.84	5.04	0.01
Aroclor-1016-3	(3)	5.05	5.05	4.95	5.15	0.00
Aroclor-1016-4	(4)	5.09	5.10	5.00	5.20	0.01
Aroclor-1016-5	(5)	5.31	5.31	5.21	5.41	0.00
Aroclor-1260-1	(1)	6.33	6.34	6.24	6.44	0.01
Aroclor-1260-2	(2)	6.52	6.53	6.43	6.63	0.01
Aroclor-1260-3	(3)	6.67	6.68	6.58	6.78	0.01
Aroclor-1260-4	(4)	7.14	7.15	7.05	7.25	0.01
Aroclor-1260-5	(5)	7.38	7.39	7.29	7.49	0.01
Tetrachloro-m-xylene		3.78	3.78	3.68	3.88	0.00
Decachlorobiphenyl		8.78	8.78	8.68	8.88	0.01

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** MATR02

Lab Code: ACE **SDG NO.:** Q3450

GC Column: ZB-MR1 **ID:** 0.32 **(mm)** **Initi. Calib. Date(s):** 10/21/2025 10/21/2025

Client Sample No.: CCAL05 **Date Analyzed:** 10/27/2025

Lab Sample No.: AR1660CCC500 **Data File :** PP076063.D **Time Analyzed:** 08:56

COMPOUND	RT	RT WINDOW FROM	RT WINDOW TO	CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	5.789	5.695	5.895	527.890	500.000	5.6
Aroclor-1016-2	5.811	5.717	5.917	519.780	500.000	4.0
Aroclor-1016-3	5.873	5.780	5.980	525.330	500.000	5.1
Aroclor-1016-4	5.971	5.877	6.077	527.600	500.000	5.5
Aroclor-1016-5	6.263	6.169	6.369	513.510	500.000	2.7
Aroclor-1260-1	7.380	7.287	7.487	555.310	500.000	11.1
Aroclor-1260-2	7.633	7.540	7.740	520.510	500.000	4.1
Aroclor-1260-3	7.991	7.899	8.099	543.790	500.000	8.8
Aroclor-1260-4	8.219	8.127	8.327	568.050	500.000	13.6
Aroclor-1260-5	8.544	8.452	8.652	527.260	500.000	5.5
Decachlorobiphenyl	10.405	10.319	10.519	52.540	50.000	5.1
Tetrachloro-m-xylene	4.637	4.542	4.742	53.280	50.000	6.6

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** MATR02

Lab Code: ACE **SDG NO.:** Q3450

GC Column: ZB-MR2 **ID:** 0.32 **(mm)** **Initi. Calib. Date(s):** 10/21/2025 10/21/2025

Client Sample No.: CCAL05 **Date Analyzed:** 10/27/2025

Lab Sample No.: AR1660CCC500 **Data File :** PP076063.D **Time Analyzed:** 08:56

COMPOUND	RT	RT WINDOW FROM	RT WINDOW TO	CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.873	4.778	4.978	528.360	500.000	5.7
Aroclor-1016-2	4.930	4.836	5.036	524.210	500.000	4.8
Aroclor-1016-3	5.051	4.954	5.154	458.440	500.000	-8.3
Aroclor-1016-4	5.092	4.996	5.196	553.310	500.000	10.7
Aroclor-1016-5	5.306	5.210	5.410	471.040	500.000	-5.8
Aroclor-1260-1	6.333	6.238	6.438	534.930	500.000	7.0
Aroclor-1260-2	6.520	6.425	6.625	523.920	500.000	4.8
Aroclor-1260-3	6.673	6.578	6.778	544.970	500.000	9.0
Aroclor-1260-4	7.142	7.047	7.247	557.220	500.000	11.4
Aroclor-1260-5	7.381	7.287	7.487	542.600	500.000	8.5
Decachlorobiphenyl	8.775	8.682	8.882	47.580	50.000	-4.8
Tetrachloro-m-xylene	3.779	3.681	3.881	52.390	50.000	4.8

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102725\
Data File : PP076063.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Oct 2025 08:56
Operator : YP\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/28/2025
Supervised By :mohammad ahmed 10/29/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 27 11:00:04 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 24 15:40:32 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.637	3.779	76140554	346.7E6	53.279	52.387
2)	SA Decachlor...	10.405	8.775	56558628	391.2E6	52.538	47.576
	Target Compounds						
3)	L1 AR-1016-1	5.789	4.873	31211797	391.1E6	527.885	528.362m
4)	L1 AR-1016-2	5.811	4.930	45971798	179.2E6	519.784	524.212m
5)	L1 AR-1016-3	5.873	5.051	29077501	94136772	525.329	458.438
6)	L1 AR-1016-4	5.971	5.092	23974413	103.4E6	527.604	553.305
7)	L1 AR-1016-5	6.263	5.306	22289187	107.4E6	513.510	471.035
31)	L7 AR-1260-1	7.380	6.333	38181326	293.0E6	555.311	534.927
32)	L7 AR-1260-2	7.633	6.520	46626204	426.3E6	520.514	523.919
33)	L7 AR-1260-3	7.991	6.673	37677160	306.7E6	543.787	544.973
34)	L7 AR-1260-4	8.219	7.142	38952287	289.4E6	568.055	557.215
35)	L7 AR-1260-5	8.544	7.381	78009133	735.5E6	527.261	542.599

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102725\
Data File : PP076063.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Oct 2025 08:56
Operator : YP\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1


Instrument :
ECD_P
ClientSampleId :
AR1660CCC500


Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/28/2025
Supervised By :mohammad ahmed 10/29/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 27 11:00:04 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 24 15:40:32 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





Raw Data: **PP076078.D**

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** MATR02

Lab Code: ACE **SDG NO.:** Q3450

Continuing Calib Date: 10/27/2025 **Initial Calibration Date(s):** 10/21/2025 10/21/2025

Continuing Calib Time: 14:08 **Initial Calibration Time(s):** 09:06 17:01

GC Column: ZB-MR1 **ID:** 0.32 **(mm)**

COMPOUND		CCAL RT	AVG RT	RT WINDOW FROM	RT WINDOW TO	DIFF RT
Aroclor-1016-1	(1)	5.79	5.80	5.70	5.90	0.01
Aroclor-1016-2	(2)	5.82	5.82	5.72	5.92	0.00
Aroclor-1016-3	(3)	5.88	5.88	5.78	5.98	0.00
Aroclor-1016-4	(4)	5.98	5.98	5.88	6.08	0.01
Aroclor-1016-5	(5)	6.27	6.27	6.17	6.37	0.00
Aroclor-1260-1	(1)	7.38	7.39	7.29	7.49	0.01
Aroclor-1260-2	(2)	7.64	7.64	7.54	7.74	0.00
Aroclor-1260-3	(3)	8.00	8.00	7.90	8.10	0.01
Aroclor-1260-4	(4)	8.22	8.23	8.13	8.33	0.01
Aroclor-1260-5	(5)	8.55	8.55	8.45	8.65	0.00
Tetrachloro-m-xylene		4.64	4.64	4.54	4.74	0.00
Decachlorobiphenyl		10.41	10.42	10.32	10.52	0.01

Raw Data: **PP076078.D**

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** MATR02

Lab Code: ACE **SDG NO.:** Q3450

Continuing Calib Date: 10/27/2025 **Initial Calibration Date(s):** 10/21/2025 10/21/2025

Continuing Calib Time: 14:08 **Initial Calibration Time(s):** 09:06 17:01

GC Column: ZB-MR2 **ID:** 0.32 **(mm)**

COMPOUND		CCAL RT	AVG RT	RT WINDOW FROM	TO	DIFF RT
Aroclor-1016-1	(1)	4.88	4.88	4.78	4.98	0.00
Aroclor-1016-2	(2)	4.94	4.94	4.84	5.04	0.01
Aroclor-1016-3	(3)	5.05	5.05	4.95	5.15	0.00
Aroclor-1016-4	(4)	5.10	5.10	5.00	5.20	0.01
Aroclor-1016-5	(5)	5.31	5.31	5.21	5.41	0.00
Aroclor-1260-1	(1)	6.34	6.34	6.24	6.44	0.00
Aroclor-1260-2	(2)	6.52	6.53	6.43	6.63	0.01
Aroclor-1260-3	(3)	6.68	6.68	6.58	6.78	0.00
Aroclor-1260-4	(4)	7.14	7.15	7.05	7.25	0.01
Aroclor-1260-5	(5)	7.38	7.39	7.29	7.49	0.01
Tetrachloro-m-xylene		3.78	3.78	3.68	3.88	0.00
Decachlorobiphenyl		8.78	8.78	8.68	8.88	0.00

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** MATR02

Lab Code: ACE **SDG NO.:** Q3450

GC Column: ZB-MR1 **ID:** 0.32 **(mm)** **Initi. Calib. Date(s):** 10/21/2025 10/21/2025

Client Sample No.: CCAL06 **Date Analyzed:** 10/27/2025

Lab Sample No.: AR1660CCC500 **Data File :** PP076078.D **Time Analyzed:** 14:08

COMPOUND	RT	RT WINDOW FROM	RT WINDOW TO	CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	5.793	5.695	5.895	489.480	500.000	-2.1
Aroclor-1016-2	5.815	5.717	5.917	493.840	500.000	-1.2
Aroclor-1016-3	5.877	5.780	5.980	492.480	500.000	-1.5
Aroclor-1016-4	5.975	5.877	6.077	498.560	500.000	-0.3
Aroclor-1016-5	6.267	6.169	6.369	482.900	500.000	-3.4
Aroclor-1260-1	7.384	7.287	7.487	535.750	500.000	7.2
Aroclor-1260-2	7.636	7.540	7.740	493.780	500.000	-1.2
Aroclor-1260-3	7.995	7.899	8.099	512.820	500.000	2.6
Aroclor-1260-4	8.222	8.127	8.327	550.900	500.000	10.2
Aroclor-1260-5	8.547	8.452	8.652	502.490	500.000	0.5
Decachlorobiphenyl	10.409	10.319	10.519	50.530	50.000	1.1
Tetrachloro-m-xylene	4.642	4.542	4.742	49.760	50.000	-0.5

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** MATR02

Lab Code: ACE **SDG NO.:** Q3450

GC Column: ZB-MR2 **ID:** 0.32 **(mm)** **Initi. Calib. Date(s):** 10/21/2025 10/21/2025

Client Sample No.: CCAL06 **Date Analyzed:** 10/27/2025

Lab Sample No.: AR1660CCC500 **Data File :** PP076078.D **Time Analyzed:** 14:08

COMPOUND	RT	RT WINDOW FROM	RT WINDOW TO	CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.876	4.778	4.978	518.620	500.000	3.7
Aroclor-1016-2	4.935	4.836	5.036	531.450	500.000	6.3
Aroclor-1016-3	5.054	4.954	5.154	501.290	500.000	0.3
Aroclor-1016-4	5.095	4.996	5.196	541.930	500.000	8.4
Aroclor-1016-5	5.308	5.210	5.410	547.650	500.000	9.5
Aroclor-1260-1	6.336	6.238	6.438	486.190	500.000	-2.8
Aroclor-1260-2	6.522	6.425	6.625	481.180	500.000	-3.8
Aroclor-1260-3	6.675	6.578	6.778	504.200	500.000	0.8
Aroclor-1260-4	7.144	7.047	7.247	522.340	500.000	4.5
Aroclor-1260-5	7.384	7.287	7.487	484.740	500.000	-3.1
Decachlorobiphenyl	8.777	8.682	8.882	45.370	50.000	-9.3
Tetrachloro-m-xylene	3.782	3.681	3.881	53.460	50.000	6.9

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102725\
Data File : PP076078.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Oct 2025 14:08
Operator : YP\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 28 01:02:38 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 24 15:40:32 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.642	3.782	71108591	353.8E6	49.758	53.460
2)	SA Decachlor...	10.409	8.777	54396935	373.1E6	50.530	45.368
	Target Compounds						
3)	L1 AR-1016-1	5.793	4.876	28941333	383.9E6	489.485	518.615
4)	L1 AR-1016-2	5.815	4.935	43677020	181.7E6	493.837	531.450
5)	L1 AR-1016-3	5.877	5.054	27259037	102.9E6	492.476	501.291
6)	L1 AR-1016-4	5.975	5.095	22654846	101.3E6	498.564	541.928
7)	L1 AR-1016-5	6.267	5.308	20960588	124.9E6	482.901	547.651
31)	L7 AR-1260-1	7.384	6.336	36836407	266.3E6	535.751	486.193
32)	L7 AR-1260-2	7.636	6.522	44231121	391.6E6	493.776	481.177
33)	L7 AR-1260-3	7.995	6.675	35531662	283.7E6	512.822	504.203
34)	L7 AR-1260-4	8.222	7.144	37775708	271.3E6	550.896	522.340
35)	L7 AR-1260-5	8.547	7.384	74344445	657.1E6	502.492	484.740


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102725\
Data File : PP076078.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Oct 2025 14:08
Operator : YP\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 28 01:02:38 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 24 15:40:32 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





Raw Data: **PP076271.D**

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** MATR02

Lab Code: ACE **SDG NO.:** Q3450

Continuing Calib Date: 10/31/2025 **Initial Calibration Date(s):** 10/21/2025 10/21/2025

Continuing Calib Time: 20:57 **Initial Calibration Time(s):** 09:06 17:01

GC Column: ZB-MR1 **ID:** 0.32 **(mm)**

COMPOUND		CCAL RT	AVG RT	RT WINDOW FROM	TO	DIFF RT
Aroclor-1016-1	(1)	5.80	5.80	5.70	5.90	0.00
Aroclor-1016-2	(2)	5.82	5.82	5.72	5.92	0.00
Aroclor-1016-3	(3)	5.88	5.88	5.78	5.98	0.00
Aroclor-1016-4	(4)	5.98	5.98	5.88	6.08	0.00
Aroclor-1016-5	(5)	6.27	6.27	6.17	6.37	0.00
Aroclor-1260-1	(1)	7.39	7.39	7.29	7.49	0.00
Aroclor-1260-2	(2)	7.64	7.64	7.54	7.74	0.00
Aroclor-1260-3	(3)	8.00	8.00	7.90	8.10	0.00
Aroclor-1260-4	(4)	8.23	8.23	8.13	8.33	0.00
Aroclor-1260-5	(5)	8.56	8.55	8.45	8.65	-0.01
Tetrachloro-m-xylene		4.65	4.64	4.54	4.74	0.00
Decachlorobiphenyl		10.42	10.42	10.32	10.52	0.00

Raw Data: **PP076271.D**

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** MATR02

Lab Code: ACE **SDG NO.:** Q3450

Continuing Calib Date: 10/31/2025 **Initial Calibration Date(s):** 10/21/2025 10/21/2025

Continuing Calib Time: 20:57 **Initial Calibration Time(s):** 09:06 17:01

GC Column: ZB-MR2 **ID:** 0.32 **(mm)**

COMPOUND		CCAL RT	AVG RT	RT WINDOW FROM	TO	DIFF RT
Aroclor-1016-1	(1)	4.88	4.88	4.78	4.98	0.00
Aroclor-1016-2	(2)	4.93	4.94	4.84	5.04	0.01
Aroclor-1016-3	(3)	5.05	5.05	4.95	5.15	0.00
Aroclor-1016-4	(4)	5.09	5.10	5.00	5.20	0.01
Aroclor-1016-5	(5)	5.31	5.31	5.21	5.41	0.00
Aroclor-1260-1	(1)	6.34	6.34	6.24	6.44	0.00
Aroclor-1260-2	(2)	6.52	6.53	6.43	6.63	0.01
Aroclor-1260-3	(3)	6.68	6.68	6.58	6.78	0.00
Aroclor-1260-4	(4)	7.14	7.15	7.05	7.25	0.01
Aroclor-1260-5	(5)	7.39	7.39	7.29	7.49	0.00
Tetrachloro-m-xylene		3.78	3.78	3.68	3.88	0.00
Decachlorobiphenyl		8.78	8.78	8.68	8.88	0.00

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** MATR02

Lab Code: ACE **SDG NO.:** Q3450

GC Column: ZB-MR1 **ID:** 0.32 **(mm)** **Initi. Calib. Date(s):** 10/21/2025 10/21/2025

Client Sample No.: CCAL07 **Date Analyzed:** 10/31/2025

Lab Sample No.: AR1660CCC500 **Data File :** PP076271.D **Time Analyzed:** 20:57

COMPOUND	RT	RT WINDOW FROM	RT WINDOW TO	CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	5.798	5.695	5.895	525.900	500.000	5.2
Aroclor-1016-2	5.819	5.717	5.917	526.230	500.000	5.2
Aroclor-1016-3	5.881	5.780	5.980	523.540	500.000	4.7
Aroclor-1016-4	5.979	5.877	6.077	529.160	500.000	5.8
Aroclor-1016-5	6.271	6.169	6.369	517.380	500.000	3.5
Aroclor-1260-1	7.390	7.287	7.487	557.240	500.000	11.4
Aroclor-1260-2	7.643	7.540	7.740	529.160	500.000	5.8
Aroclor-1260-3	8.001	7.899	8.099	545.500	500.000	9.1
Aroclor-1260-4	8.229	8.127	8.327	577.430	500.000	15.5
Aroclor-1260-5	8.555	8.452	8.652	526.510	500.000	5.3
Decachlorobiphenyl	10.423	10.319	10.519	50.300	50.000	0.6
Tetrachloro-m-xylene	4.645	4.542	4.742	53.750	50.000	7.5

Raw Data: **PP076271.D**

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** MATR02

Lab Code: ACE **SDG NO.:** Q3450

GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/21/2025 10/21/2025

Client Sample No.: CCAL07 **Date Analyzed:** 10/31/2025

Lab Sample No.: AR1660CCC500 **Data File :** PP076271.D **Time Analyzed:** 20:57

COMPOUND	RT	RT WINDOW FROM	RT WINDOW TO	CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.876	4.778	4.978	532.360	500.000	6.5
Aroclor-1016-2	4.933	4.836	5.036	519.120	500.000	3.8
Aroclor-1016-3	5.054	4.954	5.154	468.650	500.000	-6.3
Aroclor-1016-4	5.093	4.996	5.196	569.830	500.000	14.0
Aroclor-1016-5	5.308	5.210	5.410	488.380	500.000	-2.3
Aroclor-1260-1	6.336	6.238	6.438	525.820	500.000	5.2
Aroclor-1260-2	6.522	6.425	6.625	535.740	500.000	7.1
Aroclor-1260-3	6.675	6.578	6.778	554.600	500.000	10.9
Aroclor-1260-4	7.144	7.047	7.247	597.580	500.000	19.5
Aroclor-1260-5	7.385	7.287	7.487	554.130	500.000	10.8
Decachlorobiphenyl	8.779	8.682	8.882	44.880	50.000	-10.2
Tetrachloro-m-xylene	3.781	3.681	3.881	55.450	50.000	10.9

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP103125\
Data File : PP076271.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 31 Oct 2025 20:57
Operator : YP\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 31 23:27:05 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 24 15:40:32 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.645	3.781	76807876	367.0E6	53.746	55.454
2)	SA Decachlor...	10.423	8.779	54144204	369.1E6	50.295	44.883
	Target Compounds						
3)	L1 AR-1016-1	5.798	4.876	31094442	394.0E6	525.900	532.360
4)	L1 AR-1016-2	5.819	4.933	46541761	177.5E6	526.228	519.115
5)	L1 AR-1016-3	5.881	5.054	28978262	96234315	523.536	468.653
6)	L1 AR-1016-4	5.979	5.093	24045136	106.5E6	529.160	569.830
7)	L1 AR-1016-5	6.271	5.308	22457200	111.3E6	517.381	488.384
31)	L7 AR-1260-1	7.390	6.336	38313648	288.0E6	557.236	525.820
32)	L7 AR-1260-2	7.643	6.522	47400407	436.0E6	529.156	535.744
33)	L7 AR-1260-3	8.001	6.675	37796081	312.1E6	545.504	554.598
34)	L7 AR-1260-4	8.229	7.144	39595146	310.3E6	577.430	597.575
35)	L7 AR-1260-5	8.555	7.385	77897600	751.2E6	526.507	554.133


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP103125\
Data File : PP076271.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 31 Oct 2025 20:57
Operator : YP\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 31 23:27:05 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 24 15:40:32 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





Raw Data: **PP076286.D**

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** MATR02

Lab Code: ACE **SDG NO.:** Q3450

Continuing Calib Date: 11/01/2025 **Initial Calibration Date(s):** 10/21/2025 10/21/2025

Continuing Calib Time: 02:22 **Initial Calibration Time(s):** 09:06 17:01

GC Column: ZB-MR1 **ID:** 0.32 **(mm)**

COMPOUND		CCAL RT	AVG RT	RT WINDOW FROM	RT WINDOW TO	DIFF RT
Aroclor-1016-1	(1)	5.80	5.80	5.70	5.90	0.00
Aroclor-1016-2	(2)	5.82	5.82	5.72	5.92	0.00
Aroclor-1016-3	(3)	5.88	5.88	5.78	5.98	0.00
Aroclor-1016-4	(4)	5.98	5.98	5.88	6.08	0.00
Aroclor-1016-5	(5)	6.27	6.27	6.17	6.37	0.00
Aroclor-1260-1	(1)	7.39	7.39	7.29	7.49	0.00
Aroclor-1260-2	(2)	7.64	7.64	7.54	7.74	0.00
Aroclor-1260-3	(3)	8.00	8.00	7.90	8.10	0.00
Aroclor-1260-4	(4)	8.23	8.23	8.13	8.33	0.00
Aroclor-1260-5	(5)	8.55	8.55	8.45	8.65	0.00
Tetrachloro-m-xylene		4.64	4.64	4.54	4.74	0.00
Decachlorobiphenyl		10.42	10.42	10.32	10.52	0.00

Raw Data: **PP076286.D**

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** MATR02

Lab Code: ACE **SDG NO.:** Q3450

Continuing Calib Date: 11/01/2025 **Initial Calibration Date(s):** 10/21/2025 10/21/2025

Continuing Calib Time: 02:22 **Initial Calibration Time(s):** 09:06 17:01

GC Column: ZB-MR2 **ID:** 0.32 **(mm)**

COMPOUND		CCAL RT	AVG RT	RT WINDOW FROM	TO	DIFF RT
Aroclor-1016-1	(1)	4.88	4.88	4.78	4.98	0.00
Aroclor-1016-2	(2)	4.93	4.94	4.84	5.04	0.01
Aroclor-1016-3	(3)	5.05	5.05	4.95	5.15	0.00
Aroclor-1016-4	(4)	5.09	5.10	5.00	5.20	0.01
Aroclor-1016-5	(5)	5.31	5.31	5.21	5.41	0.00
Aroclor-1260-1	(1)	6.33	6.34	6.24	6.44	0.01
Aroclor-1260-2	(2)	6.52	6.53	6.43	6.63	0.01
Aroclor-1260-3	(3)	6.68	6.68	6.58	6.78	0.00
Aroclor-1260-4	(4)	7.14	7.15	7.05	7.25	0.01
Aroclor-1260-5	(5)	7.39	7.39	7.29	7.49	0.00
Tetrachloro-m-xylene		3.78	3.78	3.68	3.88	0.00
Decachlorobiphenyl		8.78	8.78	8.68	8.88	0.00

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** MATR02

Lab Code: ACE **SDG NO.:** Q3450

GC Column: ZB-MR1 **ID:** 0.32 **(mm)** **Initi. Calib. Date(s):** 10/21/2025 10/21/2025

Client Sample No.: CCAL08 **Date Analyzed:** 11/01/2025

Lab Sample No.: AR1660CCC500 **Data File :** PP076286.D **Time Analyzed:** 02:22

COMPOUND	RT	RT WINDOW FROM	RT WINDOW TO	CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	5.796	5.695	5.895	536.390	500.000	7.3
Aroclor-1016-2	5.817	5.717	5.917	532.910	500.000	6.6
Aroclor-1016-3	5.880	5.780	5.980	541.690	500.000	8.3
Aroclor-1016-4	5.977	5.877	6.077	547.810	500.000	9.6
Aroclor-1016-5	6.270	6.169	6.369	533.130	500.000	6.6
Aroclor-1260-1	7.388	7.287	7.487	567.020	500.000	13.4
Aroclor-1260-2	7.641	7.540	7.740	531.700	500.000	6.3
Aroclor-1260-3	7.999	7.899	8.099	550.460	500.000	10.1
Aroclor-1260-4	8.227	8.127	8.327	578.750	500.000	15.8
Aroclor-1260-5	8.553	8.452	8.652	525.730	500.000	5.1
Decachlorobiphenyl	10.419	10.319	10.519	50.740	50.000	1.5
Tetrachloro-m-xylene	4.643	4.542	4.742	54.070	50.000	8.1

Raw Data: **PP076286.D**

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** MATR02

Lab Code: ACE **SDG NO.:** Q3450

GC Column: ZB-MR2 **ID:** 0.32 **(mm)** **Initi. Calib. Date(s):** 10/21/2025 10/21/2025

Client Sample No.: CCAL08 **Date Analyzed:** 11/01/2025

Lab Sample No.: AR1660CCC500 **Data File :** PP076286.D **Time Analyzed:** 02:22

COMPOUND	RT	RT WINDOW FROM	RT WINDOW TO	CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.875	4.778	4.978	529.650	500.000	5.9
Aroclor-1016-2	4.934	4.836	5.036	521.320	500.000	4.3
Aroclor-1016-3	5.052	4.954	5.154	504.990	500.000	1.0
Aroclor-1016-4	5.093	4.996	5.196	539.690	500.000	7.9
Aroclor-1016-5	5.307	5.210	5.410	491.620	500.000	-1.7
Aroclor-1260-1	6.334	6.238	6.438	540.420	500.000	8.1
Aroclor-1260-2	6.522	6.425	6.625	552.530	500.000	10.5
Aroclor-1260-3	6.675	6.578	6.778	557.040	500.000	11.4
Aroclor-1260-4	7.143	7.047	7.247	577.020	500.000	15.4
Aroclor-1260-5	7.385	7.287	7.487	543.820	500.000	8.8
Decachlorobiphenyl	8.778	8.682	8.882	44.780	50.000	-10.4
Tetrachloro-m-xylene	3.781	3.681	3.881	55.710	50.000	11.4

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP103125\
Data File : PP076286.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 01 Nov 2025 02:22
Operator : YP\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 11/03/2025
Supervised By :Yogesh Patel 11/03/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 01 02:50:01 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 24 15:40:32 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.643	3.781	77264896	368.7E6	54.066	55.712
2)	SA Decachlor...	10.419	8.778	54620619	368.3E6	50.738	44.783
	Target Compounds						
3)	L1 AR-1016-1	5.796	4.875	31714622	392.0E6	536.389	529.646
4)	L1 AR-1016-2	5.817	4.934	47133127	178.3E6	532.914	521.317
5)	L1 AR-1016-3	5.880	5.052	29983370	103.7E6	541.695	504.988
6)	L1 AR-1016-4	5.977	5.093	24892492	100.9E6	547.808	539.689
7)	L1 AR-1016-5	6.270	5.307	23140795	112.1E6	533.130	491.617
31)	L7 AR-1260-1	7.388	6.334	38986107	296.0E6	567.016	540.418
32)	L7 AR-1260-2	7.641	6.522	47628374	449.6E6	531.701	552.531
33)	L7 AR-1260-3	7.999	6.675	38139718	313.5E6	550.463	557.040
34)	L7 AR-1260-4	8.227	7.143	39685668	299.7E6	578.750	577.025m
35)	L7 AR-1260-5	8.553	7.385	77782336	737.2E6	525.728	543.823

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Sample Results: **PP076286.D**

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP103125\
Data File : PP076286.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 01 Nov 2025 02:22
Operator : YP\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1


Instrument :
ECD_P
ClientSampleId :
AR1660CCC500


Manual Integrations APPROVED

Reviewed By :Rahul Chavli 11/03/2025
Supervised By :Yogesh Patel 11/03/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 01 02:50:01 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 24 15:40:32 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





Raw Data: **PP076292.D**

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** MATR02

Lab Code: ACE **SDG NO.:** Q3450

Continuing Calib Date: 11/03/2025 **Initial Calibration Date(s):** 10/21/2025 10/21/2025

Continuing Calib Time: 08:44 **Initial Calibration Time(s):** 09:06 17:01

GC Column: ZB-MR1 **ID:** 0.32 **(mm)**

COMPOUND		CCAL RT	AVG RT	RT WINDOW FROM	TO	DIFF RT
Aroclor-1016-1	(1)	5.79	5.80	5.70	5.90	0.01
Aroclor-1016-2	(2)	5.82	5.82	5.72	5.92	0.00
Aroclor-1016-3	(3)	5.88	5.88	5.78	5.98	0.00
Aroclor-1016-4	(4)	5.97	5.98	5.88	6.08	0.01
Aroclor-1016-5	(5)	6.27	6.27	6.17	6.37	0.00
Aroclor-1260-1	(1)	7.38	7.39	7.29	7.49	0.01
Aroclor-1260-2	(2)	7.64	7.64	7.54	7.74	0.00
Aroclor-1260-3	(3)	8.00	8.00	7.90	8.10	0.01
Aroclor-1260-4	(4)	8.22	8.23	8.13	8.33	0.01
Aroclor-1260-5	(5)	8.55	8.55	8.45	8.65	0.00
Tetrachloro-m-xylene		4.64	4.64	4.54	4.74	0.00
Decachlorobiphenyl		10.41	10.42	10.32	10.52	0.01

Raw Data: **PP076292.D**

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** MATR02

Lab Code: ACE **SDG NO.:** Q3450

Continuing Calib Date: 11/03/2025 **Initial Calibration Date(s):** 10/21/2025 10/21/2025

Continuing Calib Time: 08:44 **Initial Calibration Time(s):** 09:06 17:01

GC Column: ZB-MR2 **ID:** 0.32 **(mm)**

COMPOUND		CCAL RT	AVG RT	RT WINDOW FROM	TO	DIFF RT
Aroclor-1016-1	(1)	4.88	4.88	4.78	4.98	0.00
Aroclor-1016-2	(2)	4.93	4.94	4.84	5.04	0.01
Aroclor-1016-3	(3)	5.05	5.05	4.95	5.15	0.00
Aroclor-1016-4	(4)	5.09	5.10	5.00	5.20	0.01
Aroclor-1016-5	(5)	5.31	5.31	5.21	5.41	0.00
Aroclor-1260-1	(1)	6.33	6.34	6.24	6.44	0.01
Aroclor-1260-2	(2)	6.52	6.53	6.43	6.63	0.01
Aroclor-1260-3	(3)	6.67	6.68	6.58	6.78	0.01
Aroclor-1260-4	(4)	7.14	7.15	7.05	7.25	0.01
Aroclor-1260-5	(5)	7.38	7.39	7.29	7.49	0.01
Tetrachloro-m-xylene		3.78	3.78	3.68	3.88	0.00
Decachlorobiphenyl		8.77	8.78	8.68	8.88	0.01

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** MATR02

Lab Code: ACE **SDG NO.:** Q3450

GC Column: ZB-MR1 **ID:** 0.32 **(mm)** **Initi. Calib. Date(s):** 10/21/2025 10/21/2025

Client Sample No.: CCAL09 **Date Analyzed:** 11/03/2025

Lab Sample No.: AR1660CCC500 **Data File :** PP076292.D **Time Analyzed:** 08:44

COMPOUND	RT	RT WINDOW FROM	RT WINDOW TO	CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	5.793	5.695	5.895	549.830	500.000	10.0
Aroclor-1016-2	5.815	5.717	5.917	540.140	500.000	8.0
Aroclor-1016-3	5.877	5.780	5.980	546.540	500.000	9.3
Aroclor-1016-4	5.974	5.877	6.077	550.230	500.000	10.0
Aroclor-1016-5	6.267	6.169	6.369	532.730	500.000	6.5
Aroclor-1260-1	7.384	7.287	7.487	580.700	500.000	16.1
Aroclor-1260-2	7.637	7.540	7.740	549.670	500.000	9.9
Aroclor-1260-3	7.995	7.899	8.099	576.350	500.000	15.3
Aroclor-1260-4	8.223	8.127	8.327	597.300	500.000	19.5
Aroclor-1260-5	8.547	8.452	8.652	550.880	500.000	10.2
Decachlorobiphenyl	10.411	10.319	10.519	52.840	50.000	5.7
Tetrachloro-m-xylene	4.641	4.542	4.742	54.140	50.000	8.3

Raw Data: **PP076292.D**

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** MATR02

Lab Code: ACE **SDG NO.:** Q3450

GC Column: ZB-MR2 **ID:** 0.32 **(mm)** **Initi. Calib. Date(s):** 10/21/2025 10/21/2025

Client Sample No.: CCAL09 **Date Analyzed:** 11/03/2025

Lab Sample No.: AR1660CCC500 **Data File :** PP076292.D **Time Analyzed:** 08:44

COMPOUND	RT	RT WINDOW FROM	RT WINDOW TO	CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.875	4.778	4.978	535.300	500.000	7.1
Aroclor-1016-2	4.932	4.836	5.036	535.190	500.000	7.0
Aroclor-1016-3	5.052	4.954	5.154	481.690	500.000	-3.7
Aroclor-1016-4	5.092	4.996	5.196	565.650	500.000	13.1
Aroclor-1016-5	5.307	5.210	5.410	469.810	500.000	-6.0
Aroclor-1260-1	6.334	6.238	6.438	553.960	500.000	10.8
Aroclor-1260-2	6.520	6.425	6.625	558.660	500.000	11.7
Aroclor-1260-3	6.673	6.578	6.778	580.770	500.000	16.2
Aroclor-1260-4	7.141	7.047	7.247	631.580	500.000	26.3
Aroclor-1260-5	7.382	7.287	7.487	589.330	500.000	17.9
Decachlorobiphenyl	8.774	8.682	8.882	51.320	50.000	2.6
Tetrachloro-m-xylene	3.780	3.681	3.881	57.770	50.000	15.5

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP110325\
Data File : PP076292.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 03 Nov 2025 08:44
Operator : YP\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 03 11:20:37 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 24 15:40:32 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.641	3.780	77374612	382.3E6	54.142	57.771
2)	SA Decachlor...	10.411	8.774	56883294	422.0E6	52.840	51.318
	Target Compounds						
3)	L1 AR-1016-1	5.793	4.875	32509355	396.2E6	549.831	535.300
4)	L1 AR-1016-2	5.815	4.932	47771850	183.0E6	540.136	535.192
5)	L1 AR-1016-3	5.877	5.052	30251676	98910823	546.542	481.687
6)	L1 AR-1016-4	5.974	5.092	25002503	105.7E6	550.229	565.651
7)	L1 AR-1016-5	6.267	5.307	23123317	107.1E6	532.727	469.811
31)	L7 AR-1260-1	7.384	6.334	39927293	303.4E6	580.705	553.962
32)	L7 AR-1260-2	7.637	6.520	49238333	454.6E6	549.674	558.660
33)	L7 AR-1260-3	7.995	6.673	39933186	326.8E6	576.348	580.766
34)	L7 AR-1260-4	8.223	7.141	40957589	328.0E6	597.299	631.579
35)	L7 AR-1260-5	8.547	7.382	81502906	798.9E6	550.875	589.333

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Sample Results: **PP076292.D**


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP110325\
Data File : PP076292.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 03 Nov 2025 08:44
Operator : YP\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 03 11:20:37 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 24 15:40:32 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





Raw Data: **PP076304.D**

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** MATR02

Lab Code: ACE **SDG NO.:** Q3450

Continuing Calib Date: 11/03/2025 **Initial Calibration Date(s):** 10/21/2025 10/21/2025

Continuing Calib Time: 14:19 **Initial Calibration Time(s):** 09:06 17:01

GC Column: ZB-MR1 **ID:** 0.32 **(mm)**

COMPOUND		CCAL RT	AVG RT	RT WINDOW FROM	RT WINDOW TO	DIFF RT
Aroclor-1016-1	(1)	5.79	5.80	5.70	5.90	0.01
Aroclor-1016-2	(2)	5.82	5.82	5.72	5.92	0.00
Aroclor-1016-3	(3)	5.88	5.88	5.78	5.98	0.00
Aroclor-1016-4	(4)	5.97	5.98	5.88	6.08	0.01
Aroclor-1016-5	(5)	6.27	6.27	6.17	6.37	0.00
Aroclor-1260-1	(1)	7.38	7.39	7.29	7.49	0.01
Aroclor-1260-2	(2)	7.64	7.64	7.54	7.74	0.00
Aroclor-1260-3	(3)	8.00	8.00	7.90	8.10	0.01
Aroclor-1260-4	(4)	8.22	8.23	8.13	8.33	0.01
Aroclor-1260-5	(5)	8.55	8.55	8.45	8.65	0.00
Tetrachloro-m-xylene		4.64	4.64	4.54	4.74	0.00
Decachlorobiphenyl		10.41	10.42	10.32	10.52	0.01

Raw Data: **PP076304.D**

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** MATR02

Lab Code: ACE **SDG NO.:** Q3450

Continuing Calib Date: 11/03/2025 **Initial Calibration Date(s):** 10/21/2025 10/21/2025

Continuing Calib Time: 14:19 **Initial Calibration Time(s):** 09:06 17:01

GC Column: ZB-MR2 **ID:** 0.32 **(mm)**

COMPOUND		CCAL RT	AVG RT	RT WINDOW FROM	TO	DIFF RT
Aroclor-1016-1	(1)	4.87	4.88	4.78	4.98	0.01
Aroclor-1016-2	(2)	4.93	4.94	4.84	5.04	0.01
Aroclor-1016-3	(3)	5.05	5.05	4.95	5.15	0.00
Aroclor-1016-4	(4)	5.09	5.10	5.00	5.20	0.01
Aroclor-1016-5	(5)	5.31	5.31	5.21	5.41	0.00
Aroclor-1260-1	(1)	6.33	6.34	6.24	6.44	0.01
Aroclor-1260-2	(2)	6.52	6.53	6.43	6.63	0.01
Aroclor-1260-3	(3)	6.67	6.68	6.58	6.78	0.01
Aroclor-1260-4	(4)	7.14	7.15	7.05	7.25	0.01
Aroclor-1260-5	(5)	7.38	7.39	7.29	7.49	0.01
Tetrachloro-m-xylene		3.78	3.78	3.68	3.88	0.00
Decachlorobiphenyl		8.78	8.78	8.68	8.88	0.01

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** MATR02

Lab Code: ACE **SDG NO.:** Q3450

GC Column: ZB-MR1 **ID:** 0.32 **(mm)** **Initi. Calib. Date(s):** 10/21/2025 10/21/2025

Client Sample No.: CCAL10 **Date Analyzed:** 11/03/2025

Lab Sample No.: AR1660CCC500 **Data File :** PP076304.D **Time Analyzed:** 14:19

COMPOUND	RT	RT WINDOW FROM	RT WINDOW TO	CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	5.793	5.695	5.895	500.340	500.000	0.1
Aroclor-1016-2	5.815	5.717	5.917	502.560	500.000	0.5
Aroclor-1016-3	5.878	5.780	5.980	506.050	500.000	1.2
Aroclor-1016-4	5.974	5.877	6.077	509.790	500.000	2.0
Aroclor-1016-5	6.267	6.169	6.369	470.310	500.000	-5.9
Aroclor-1260-1	7.384	7.287	7.487	514.570	500.000	2.9
Aroclor-1260-2	7.637	7.540	7.740	507.870	500.000	1.6
Aroclor-1260-3	7.995	7.899	8.099	507.650	500.000	1.5
Aroclor-1260-4	8.223	8.127	8.327	545.660	500.000	9.1
Aroclor-1260-5	8.547	8.452	8.652	501.480	500.000	0.3
Decachlorobiphenyl	10.412	10.319	10.519	48.500	50.000	-3.0
Tetrachloro-m-xylene	4.641	4.542	4.742	53.160	50.000	6.3

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** MATR02

Lab Code: ACE **SDG NO.:** Q3450

GC Column: ZB-MR2 **ID:** 0.32 **(mm)** **Initi. Calib. Date(s):** 10/21/2025 10/21/2025

Client Sample No.: CCAL10 **Date Analyzed:** 11/03/2025

Lab Sample No.: AR1660CCC500 **Data File :** PP076304.D **Time Analyzed:** 14:19

COMPOUND	RT	RT WINDOW FROM	RT WINDOW TO	CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.874	4.778	4.978	532.670	500.000	6.5
Aroclor-1016-2	4.933	4.836	5.036	523.230	500.000	4.6
Aroclor-1016-3	5.051	4.954	5.154	511.130	500.000	2.2
Aroclor-1016-4	5.092	4.996	5.196	562.590	500.000	12.5
Aroclor-1016-5	5.306	5.210	5.410	502.290	500.000	0.5
Aroclor-1260-1	6.334	6.238	6.438	461.510	500.000	-7.7
Aroclor-1260-2	6.521	6.425	6.625	458.300	500.000	-8.3
Aroclor-1260-3	6.674	6.578	6.778	473.640	500.000	-5.3
Aroclor-1260-4	7.143	7.047	7.247	481.260	500.000	-3.7
Aroclor-1260-5	7.383	7.287	7.487	435.790	500.000	-12.8
Decachlorobiphenyl	8.775	8.682	8.882	41.250	50.000	-17.5
Tetrachloro-m-xylene	3.780	3.681	3.881	54.460	50.000	8.9

Quantitation Report (QT Reviewed)

Sample Results: **PP076304.D**

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP110325\
Data File : PP076304.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 03 Nov 2025 14:19
Operator : YP\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 04 01:11:12 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 24 15:40:32 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.641	3.780	75974510	360.5E6	53.163	54.463
2)	SA Decachlor...	10.412	8.775	52208092	339.2E6	48.497	41.254
	Target Compounds						
3)	L1 AR-1016-1	5.793	4.874	29582959	394.3E6	500.337	532.668
4)	L1 AR-1016-2	5.815	4.933	44448700	178.9E6	502.563	523.230
5)	L1 AR-1016-3	5.878	5.051	28010222	105.0E6	506.047	511.127
6)	L1 AR-1016-4	5.974	5.092	23164974	105.1E6	509.790	562.590
7)	L1 AR-1016-5	6.267	5.306	20413894	114.5E6	470.306	502.289
31)	L7 AR-1260-1	7.384	6.334	35379920	252.8E6	514.568	461.507
32)	L7 AR-1260-2	7.637	6.521	45493205	372.9E6	507.865	458.297
33)	L7 AR-1260-3	7.995	6.674	35173285	266.5E6	507.649	473.637
34)	L7 AR-1260-4	8.223	7.143	37416901	249.9E6	545.664	481.260
35)	L7 AR-1260-5	8.547	7.383	74195290	590.7E6	501.483	435.795


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP110325\
Data File : PP076304.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 03 Nov 2025 14:19
Operator : YP\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 04 01:11:12 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 24 15:40:32 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





Raw Data: **PQ071149.D**

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** MATR02

Lab Code: ACE **SDG NO.:** Q3450

Continuing Calib Date: 10/27/2025 **Initial Calibration Date(s):** 10/10/2025 10/10/2025

Continuing Calib Time: 08:57 **Initial Calibration Time(s):** 08:57 16:43

GC Column: ZB-MR1 **ID:** 0.32 **(mm)**

COMPOUND		CCAL RT	AVG RT	RT WINDOW FROM	TO	DIFF RT
Aroclor-1016-1	(1)	4.48	4.48	4.38	4.58	0.00
Aroclor-1016-2	(2)	4.50	4.50	4.40	4.60	0.00
Aroclor-1016-3	(3)	4.56	4.56	4.46	4.66	0.00
Aroclor-1016-4	(4)	4.65	4.65	4.55	4.75	0.00
Aroclor-1016-5	(5)	4.93	4.93	4.83	5.03	0.00
Aroclor-1260-1	(1)	6.02	6.02	5.92	6.12	0.00
Aroclor-1260-2	(2)	6.28	6.28	6.18	6.38	0.00
Aroclor-1260-3	(3)	6.63	6.63	6.53	6.73	0.00
Aroclor-1260-4	(4)	6.84	6.84	6.74	6.94	0.00
Aroclor-1260-5	(5)	7.15	7.15	7.05	7.25	0.00
Tetrachloro-m-xylene		3.40	3.40	3.30	3.50	0.00
Decachlorobiphenyl		8.50	8.50	8.40	8.60	0.00

Raw Data: **PQ071149.D**

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** MATR02

Lab Code: ACE **SDG NO.:** Q3450

Continuing Calib Date: 10/27/2025 **Initial Calibration Date(s):** 10/10/2025 10/10/2025

Continuing Calib Time: 08:57 **Initial Calibration Time(s):** 08:57 16:43

GC Column: ZB-MR2 **ID:** 0.32 **(mm)**

COMPOUND		CCAL RT	AVG RT	RT WINDOW FROM	TO	DIFF RT
Aroclor-1016-1	(1)	3.78	3.78	3.68	3.88	0.00
Aroclor-1016-2	(2)	3.79	3.79	3.69	3.89	0.00
Aroclor-1016-3	(3)	3.95	3.95	3.85	4.05	0.00
Aroclor-1016-4	(4)	4.00	4.00	3.90	4.10	0.00
Aroclor-1016-5	(5)	4.19	4.20	4.10	4.30	0.01
Aroclor-1260-1	(1)	5.19	5.19	5.09	5.29	0.00
Aroclor-1260-2	(2)	5.38	5.38	5.28	5.48	0.00
Aroclor-1260-3	(3)	5.52	5.52	5.42	5.62	0.00
Aroclor-1260-4	(4)	5.98	5.98	5.88	6.08	0.00
Aroclor-1260-5	(5)	6.22	6.22	6.12	6.32	0.00
Tetrachloro-m-xylene		2.78	2.78	2.68	2.88	0.00
Decachlorobiphenyl		7.53	7.53	7.43	7.63	0.00

Raw Data: **PQ071149.D**

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** MATR02

Lab Code: ACE **SDG NO.:** Q3450

GC Column: ZB-MR1 **ID:** 0.32 **(mm)** **Initi. Calib. Date(s):** 10/10/2025 10/10/2025

Client Sample No.: CCAL11 **Date Analyzed:** 10/27/2025

Lab Sample No.: AR1660CCC500 **Data File :** PQ071149.D **Time Analyzed:** 08:57

COMPOUND	RT	RT WINDOW FROM	RT WINDOW TO	CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.480	4.380	4.580	582.660	500.000	16.5
Aroclor-1016-2	4.499	4.399	4.599	562.340	500.000	12.5
Aroclor-1016-3	4.556	4.456	4.656	584.520	500.000	16.9
Aroclor-1016-4	4.646	4.546	4.746	585.390	500.000	17.1
Aroclor-1016-5	4.931	4.831	5.031	576.280	500.000	15.3
Aroclor-1260-1	6.020	5.919	6.119	558.050	500.000	11.6
Aroclor-1260-2	6.276	6.175	6.375	557.500	500.000	11.5
Aroclor-1260-3	6.626	6.526	6.726	583.750	500.000	16.8
Aroclor-1260-4	6.841	6.740	6.940	592.250	500.000	18.5
Aroclor-1260-5	7.146	7.045	7.245	579.210	500.000	15.8
Decachlorobiphenyl	8.497	8.396	8.596	59.340	50.000	18.7
Tetrachloro-m-xylene	3.400	3.300	3.500	57.140	50.000	14.3

Raw Data: **PQ071149.D**

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** MATR02

Lab Code: ACE **SDG NO.:** Q3450

GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/10/2025 10/10/2025

Client Sample No.: CCAL11 **Date Analyzed:** 10/27/2025

Lab Sample No.: AR1660CCC500 **Data File :** PQ071149.D **Time Analyzed:** 08:57

COMPOUND	RT	RT WINDOW FROM	RT WINDOW TO	CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	3.776	3.678	3.878	461.700	500.000	-7.7
Aroclor-1016-2	3.791	3.693	3.893	465.260	500.000	-6.9
Aroclor-1016-3	3.951	3.853	4.053	480.210	500.000	-4.0
Aroclor-1016-4	4.000	3.902	4.102	488.300	500.000	-2.3
Aroclor-1016-5	4.194	4.097	4.297	480.390	500.000	-3.9
Aroclor-1260-1	5.187	5.089	5.289	477.270	500.000	-4.5
Aroclor-1260-2	5.376	5.278	5.478	476.660	500.000	-4.7
Aroclor-1260-3	5.517	5.419	5.619	480.230	500.000	-4.0
Aroclor-1260-4	5.977	5.879	6.079	481.030	500.000	-3.8
Aroclor-1260-5	6.222	6.124	6.324	480.940	500.000	-3.8
Decachlorobiphenyl	7.531	7.433	7.633	56.210	50.000	12.4
Tetrachloro-m-xylene	2.776	2.678	2.878	49.690	50.000	-0.6

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ102725\
Data File : PQ071149.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Oct 2025 08:57
Operator : YP\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1660CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 27 11:27:45 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Oct 11 01:55:41 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	3.400	2.776	248.6E6	252.4E6	57.145	49.691
2)	SA Decachlor...	8.497	7.531	172.3E6	283.8E6	59.344	56.205
	Target Compounds						
3)	L1 AR-1016-1	4.480	3.776	95639952	103.4E6	582.659	461.704
4)	L1 AR-1016-2	4.499	3.791	144.1E6	158.2E6	562.335	465.258
5)	L1 AR-1016-3	4.556	3.951	93008322	87331257	584.523	480.214
6)	L1 AR-1016-4	4.646	4.000	76059348	76865356	585.390	488.305
7)	L1 AR-1016-5	4.931	4.194	74226869	92158770	576.283	480.389
31)	L7 AR-1260-1	6.020	5.187	130.8E6	172.0E6	558.047	477.273
32)	L7 AR-1260-2	6.276	5.376	154.2E6	209.7E6	557.496	476.659
33)	L7 AR-1260-3	6.626	5.517	116.3E6	199.3E6	583.746	480.230
34)	L7 AR-1260-4	6.841	5.977	127.3E6	173.1E6	592.247	481.031
35)	L7 AR-1260-5	7.146	6.222	251.2E6	404.1E6	579.210	480.935


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ102725\
Data File : PQ071149.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Oct 2025 08:57
Operator : YP\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1660CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 27 11:27:45 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Oct 11 01:55:41 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





Raw Data: **PQ071161.D**

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** MATR02

Lab Code: ACE **SDG NO.:** Q3450

Continuing Calib Date: 10/27/2025 **Initial Calibration Date(s):** 10/10/2025 10/10/2025

Continuing Calib Time: 12:35 **Initial Calibration Time(s):** 08:57 16:43

GC Column: ZB-MR1 **ID:** 0.32 **(mm)**

COMPOUND		CCAL RT	AVG RT	RT WINDOW FROM	TO	DIFF RT
Aroclor-1016-1	(1)	4.48	4.48	4.38	4.58	0.00
Aroclor-1016-2	(2)	4.50	4.50	4.40	4.60	0.00
Aroclor-1016-3	(3)	4.56	4.56	4.46	4.66	0.00
Aroclor-1016-4	(4)	4.65	4.65	4.55	4.75	0.01
Aroclor-1016-5	(5)	4.93	4.93	4.83	5.03	0.00
Aroclor-1260-1	(1)	6.02	6.02	5.92	6.12	0.00
Aroclor-1260-2	(2)	6.27	6.28	6.18	6.38	0.01
Aroclor-1260-3	(3)	6.63	6.63	6.53	6.73	0.00
Aroclor-1260-4	(4)	6.84	6.84	6.74	6.94	0.00
Aroclor-1260-5	(5)	7.15	7.15	7.05	7.25	0.01
Tetrachloro-m-xylene		3.40	3.40	3.30	3.50	0.00
Decachlorobiphenyl		8.50	8.50	8.40	8.60	0.01

Raw Data: **PQ071161.D**

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** MATR02

Lab Code: ACE **SDG NO.:** Q3450

Continuing Calib Date: 10/27/2025 **Initial Calibration Date(s):** 10/10/2025 10/10/2025

Continuing Calib Time: 12:35 **Initial Calibration Time(s):** 08:57 16:43

GC Column: ZB-MR2 **ID:** 0.32 **(mm)**

COMPOUND		CCAL RT	AVG RT	RT WINDOW FROM	TO	DIFF RT
Aroclor-1016-1	(1)	3.78	3.78	3.68	3.88	0.00
Aroclor-1016-2	(2)	3.79	3.79	3.69	3.89	0.00
Aroclor-1016-3	(3)	3.95	3.95	3.85	4.05	0.00
Aroclor-1016-4	(4)	4.00	4.00	3.90	4.10	0.00
Aroclor-1016-5	(5)	4.20	4.20	4.10	4.30	0.00
Aroclor-1260-1	(1)	5.19	5.19	5.09	5.29	0.00
Aroclor-1260-2	(2)	5.38	5.38	5.28	5.48	0.00
Aroclor-1260-3	(3)	5.52	5.52	5.42	5.62	0.00
Aroclor-1260-4	(4)	5.98	5.98	5.88	6.08	0.00
Aroclor-1260-5	(5)	6.22	6.22	6.12	6.32	0.00
Tetrachloro-m-xylene		2.78	2.78	2.68	2.88	0.00
Decachlorobiphenyl		7.53	7.53	7.43	7.63	0.00

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** MATR02

Lab Code: ACE **SDG NO.:** Q3450

GC Column: ZB-MR1 **ID:** 0.32 **(mm)** **Initi. Calib. Date(s):** 10/10/2025 10/10/2025

Client Sample No.: CCAL12 **Date Analyzed:** 10/27/2025

Lab Sample No.: AR1660CCC500 **Data File :** PQ071161.D **Time Analyzed:** 12:35

COMPOUND	RT	RT WINDOW FROM	RT WINDOW TO	CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.479	4.380	4.580	550.580	500.000	10.1
Aroclor-1016-2	4.498	4.399	4.599	559.440	500.000	11.9
Aroclor-1016-3	4.556	4.456	4.656	571.120	500.000	14.2
Aroclor-1016-4	4.645	4.546	4.746	570.660	500.000	14.1
Aroclor-1016-5	4.930	4.831	5.031	565.870	500.000	13.2
Aroclor-1260-1	6.019	5.919	6.119	557.440	500.000	11.5
Aroclor-1260-2	6.274	6.175	6.375	551.940	500.000	10.4
Aroclor-1260-3	6.626	6.526	6.726	579.440	500.000	15.9
Aroclor-1260-4	6.839	6.740	6.940	591.430	500.000	18.3
Aroclor-1260-5	7.145	7.045	7.245	581.000	500.000	16.2
Decachlorobiphenyl	8.495	8.396	8.596	59.520	50.000	19.0
Tetrachloro-m-xylene	3.399	3.300	3.500	56.350	50.000	12.7

Raw Data: **PQ071161.D**

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** MATR02

Lab Code: ACE **SDG NO.:** Q3450

GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/10/2025 10/10/2025

Client Sample No.: CCAL12 **Date Analyzed:** 10/27/2025

Lab Sample No.: AR1660CCC500 **Data File :** PQ071161.D **Time Analyzed:** 12:35

COMPOUND	RT	RT WINDOW FROM	RT WINDOW TO	CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	3.776	3.678	3.878	459.340	500.000	-8.1
Aroclor-1016-2	3.791	3.693	3.893	461.520	500.000	-7.7
Aroclor-1016-3	3.951	3.853	4.053	473.930	500.000	-5.2
Aroclor-1016-4	4.000	3.902	4.102	480.760	500.000	-3.8
Aroclor-1016-5	4.195	4.097	4.297	472.690	500.000	-5.5
Aroclor-1260-1	5.186	5.089	5.289	469.950	500.000	-6.0
Aroclor-1260-2	5.376	5.278	5.478	469.430	500.000	-6.1
Aroclor-1260-3	5.517	5.419	5.619	470.970	500.000	-5.8
Aroclor-1260-4	5.977	5.879	6.079	477.010	500.000	-4.6
Aroclor-1260-5	6.221	6.124	6.324	476.200	500.000	-4.8
Decachlorobiphenyl	7.531	7.433	7.633	55.990	50.000	12.0
Tetrachloro-m-xylene	2.776	2.678	2.878	50.180	50.000	0.4

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ102725\
Data File : PQ071161.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Oct 2025 12:35
Operator : YP\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1660CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 27 12:58:57 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Oct 11 01:55:41 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	3.399	2.776	245.1E6	254.9E6	56.346	50.181
2)	SA Decachlor...	8.495	7.531	172.8E6	282.7E6	59.521	55.992
	Target Compounds						
3)	L1 AR-1016-1	4.479	3.776	90374544	102.9E6	550.581	459.336
4)	L1 AR-1016-2	4.498	3.791	143.4E6	157.0E6	559.440	461.520
5)	L1 AR-1016-3	4.556	3.951	90876424	86188973	571.125	473.933
6)	L1 AR-1016-4	4.645	4.000	74144880	75678013	570.656	480.762
7)	L1 AR-1016-5	4.930	4.195	72885811	90681279	565.871	472.688
31)	L7 AR-1260-1	6.019	5.186	130.7E6	169.4E6	557.445	469.948
32)	L7 AR-1260-2	6.274	5.376	152.7E6	206.5E6	551.943	469.434
33)	L7 AR-1260-3	6.626	5.517	115.5E6	195.5E6	579.442	470.969
34)	L7 AR-1260-4	6.839	5.977	127.1E6	171.6E6	591.425	477.012
35)	L7 AR-1260-5	7.145	6.221	252.0E6	400.1E6	580.996	476.203


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ102725\
Data File : PQ071161.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Oct 2025 12:35
Operator : YP\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1660CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 27 12:58:57 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Oct 11 01:55:41 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Analytical Sequence

Client: Matrix New World Engineering **SDG No.:** Q3450

Project: PSEG Athenia Substation **Instrument ID:** ECD_O

GC Column: ZB-MR1 **ID:** 0.32 (mm) **Inst. Calib. Date(s):** 11/04/2025 11/04/2025

THE ANALYTICAL SEQUENCE OF PERFORMANCE EVALUATION MIXTURES, BLANKS, SAMPLES, AND STANDARDS IS GIVEN BELOW:

CLIENT ID	LAB SAMPLE ID	DATE ANALYZED	TIME ANALYZED	DATAFILE	DCB RT #	TCX RT #
I.BLK	I.BLK	11/04/2025	09:21	PO114870.D	9.93	4.35
AR1660ICC1000	AR1660ICC1000	11/04/2025	09:39	PO114871.D	9.93	4.35
AR1660ICC750	AR1660ICC750	11/04/2025	09:58	PO114872.D	9.93	4.35
AR1660ICC500	AR1660ICC500	11/04/2025	10:16	PO114873.D	9.93	4.35
AR1660ICC250	AR1660ICC250	11/04/2025	10:34	PO114874.D	9.93	4.35
AR1660ICC050	AR1660ICC050	11/04/2025	10:53	PO114875.D	9.93	4.35
AR1221ICC500	AR1221ICC500	11/04/2025	11:11	PO114876.D	9.93	4.35
AR1232ICC500	AR1232ICC500	11/04/2025	11:30	PO114877.D	9.93	4.35
AR1242ICC1000	AR1242ICC1000	11/04/2025	11:48	PO114878.D	9.93	4.35
AR1242ICC750	AR1242ICC750	11/04/2025	12:06	PO114879.D	9.93	4.35
AR1242ICC500	AR1242ICC500	11/04/2025	12:25	PO114880.D	9.93	4.35
AR1242ICC250	AR1242ICC250	11/04/2025	12:42	PO114881.D	9.93	4.35
AR1242ICC050	AR1242ICC050	11/04/2025	13:00	PO114882.D	9.93	4.35
AR1248ICC1000	AR1248ICC1000	11/04/2025	13:19	PO114883.D	9.93	4.35
AR1248ICC750	AR1248ICC750	11/04/2025	13:37	PO114884.D	9.93	4.35
AR1248ICC500	AR1248ICC500	11/04/2025	13:56	PO114885.D	9.93	4.35
AR1248ICC250	AR1248ICC250	11/04/2025	14:14	PO114886.D	9.93	4.35
AR1248ICC050	AR1248ICC050	11/04/2025	14:33	PO114887.D	9.93	4.35
AR1254ICC1000	AR1254ICC1000	11/04/2025	14:51	PO114888.D	9.93	4.35
AR1254ICC750	AR1254ICC750	11/04/2025	15:09	PO114889.D	9.93	4.35
AR1254ICC500	AR1254ICC500	11/04/2025	15:28	PO114890.D	9.93	4.35
AR1254ICC250	AR1254ICC250	11/04/2025	16:04	PO114891.D	9.93	4.35
AR1254ICC050	AR1254ICC050	11/04/2025	16:23	PO114892.D	9.93	4.35
AR1262ICC500	AR1262ICC500	11/04/2025	17:00	PO114893.D	9.95	4.37
AR1268ICC1000	AR1268ICC1000	11/04/2025	17:18	PO114894.D	9.93	4.35
AR1268ICC750	AR1268ICC750	11/04/2025	17:36	PO114895.D	9.93	4.35
AR1268ICC500	AR1268ICC500	11/04/2025	17:55	PO114896.D	9.93	4.35
AR1268ICC250	AR1268ICC250	11/04/2025	18:13	PO114897.D	9.93	4.35
AR1268ICC050	AR1268ICC050	11/04/2025	18:31	PO114898.D	9.93	4.35
AR1660CCC500	AR1660CCC500	11/12/2025	09:49	PO115031.D	9.93	4.35
I.BLK	I.BLK	11/12/2025	11:03	PO115035.D	9.93	4.35
PB170497BL	PB170497BL	11/12/2025	11:58	PO115037.D	9.93	4.35
PB170497BS	PB170497BS	11/12/2025	12:16	PO115038.D	9.93	4.35
TP-1MS	Q3597-01MS	11/12/2025	12:53	PO115040.D	9.93	4.35
TP-1MSD	Q3597-01MSD	11/12/2025	13:11	PO115041.D	9.94	4.35
SS-13R-(4.5-5.0)	Q3450-07	11/12/2025	15:36	PO115045.D	9.93	4.35
AR1660CCC500	AR1660CCC500	11/12/2025	15:55	PO115046.D	9.93	4.35
I.BLK	I.BLK	11/12/2025	17:07	PO115050.D	9.93	4.35
I.BLK	I.BLK	10/21/2025	08:49	PP075887.D	10.42	4.64
AR1660ICC1000	AR1660ICC1000	10/21/2025	09:06	PP075888.D	10.42	4.65
AR1660ICC750	AR1660ICC750	10/21/2025	09:22	PP075889.D	10.42	4.64
AR1660ICC500	AR1660ICC500	10/21/2025	09:38	PP075890.D	10.42	4.64

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Analytical Sequence

AR1660ICC250	AR1660ICC250	10/21/2025	09:54	PP075891.D	10.42	4.65
AR1660ICC050	AR1660ICC050	10/21/2025	10:11	PP075892.D	10.42	4.65
AR1221ICC500	AR1221ICC500	10/21/2025	10:27	PP075893.D	10.42	4.64
AR1232ICC500	AR1232ICC500	10/21/2025	10:43	PP075894.D	10.42	4.64
AR1242ICC1000	AR1242ICC1000	10/21/2025	10:59	PP075895.D	10.42	4.65
AR1242ICC750	AR1242ICC750	10/21/2025	11:16	PP075896.D	10.42	4.64
AR1242ICC500	AR1242ICC500	10/21/2025	11:48	PP075897.D	10.42	4.65
AR1242ICC250	AR1242ICC250	10/21/2025	12:04	PP075898.D	10.42	4.65
AR1242ICC050	AR1242ICC050	10/21/2025	12:21	PP075899.D	10.42	4.65
AR1248ICC1000	AR1248ICC1000	10/21/2025	12:41	PP075900.D	10.43	4.65
AR1248ICC750	AR1248ICC750	10/21/2025	12:57	PP075901.D	10.42	4.64
AR1248ICC500	AR1248ICC500	10/21/2025	13:13	PP075902.D	10.42	4.65
AR1248ICC250	AR1248ICC250	10/21/2025	13:30	PP075903.D	10.42	4.65
AR1248ICC050	AR1248ICC050	10/21/2025	13:46	PP075904.D	10.42	4.65
AR1254ICC1000	AR1254ICC1000	10/21/2025	14:02	PP075905.D	10.42	4.65
AR1254ICC750	AR1254ICC750	10/21/2025	14:18	PP075906.D	10.42	4.64
AR1254ICC500	AR1254ICC500	10/21/2025	14:35	PP075907.D	10.42	4.65
AR1254ICC250	AR1254ICC250	10/21/2025	14:51	PP075908.D	10.42	4.65
AR1254ICC050	AR1254ICC050	10/21/2025	15:07	PP075909.D	10.42	4.64
AR1262ICC500	AR1262ICC500	10/21/2025	15:40	PP075910.D	10.43	4.65
AR1268ICC1000	AR1268ICC1000	10/21/2025	15:56	PP075911.D	10.42	4.65
AR1268ICC750	AR1268ICC750	10/21/2025	16:12	PP075912.D	10.42	4.65
AR1268ICC500	AR1268ICC500	10/21/2025	16:29	PP075913.D	10.42	4.65
AR1268ICC250	AR1268ICC250	10/21/2025	16:45	PP075914.D	10.43	4.65
AR1268ICC050	AR1268ICC050	10/21/2025	17:01	PP075915.D	10.42	4.65
AR1660CCC500	AR1660CCC500	10/24/2025	18:58	PP076042.D	10.41	4.64
I.BLK	I.BLK	10/24/2025	20:36	PP076046.D	10.41	4.64
SS-20-(3.5-4.0)	Q3450-10	10/24/2025	20:52	PP076047.D	10.41	4.64
SS-28-(2.0-2.5)	Q3450-11	10/24/2025	21:08	PP076048.D	10.41	4.64
SS-21-(0.0-0.5)	Q3450-20	10/24/2025	21:24	PP076049.D	10.41	4.64
SS-21-(2.0-2.5)	Q3450-21	10/24/2025	21:41	PP076050.D	10.41	4.64
SS-21-(3.5-4.0)	Q3450-22	10/24/2025	21:57	PP076051.D	10.41	4.64
AR1660CCC500	AR1660CCC500	10/25/2025	00:22	PP076057.D	10.41	4.64
I.BLK	I.BLK	10/25/2025	02:16	PP076061.D	10.41	4.64
AR1660CCC500	AR1660CCC500	10/27/2025	08:56	PP076063.D	10.41	4.64
I.BLK	I.BLK	10/27/2025	10:02	PP076067.D	10.40	4.64
PB170236BL	PB170236BL	10/27/2025	10:18	PP076068.D	10.40	4.64
PB170236BS	PB170236BS	10/27/2025	10:34	PP076069.D	10.41	4.64
SS-21-(0.0-0.5)DL	Q3450-20DL	10/27/2025	11:42	PP076072.D	10.41	4.64
SS-21-(3.5-4.0)DL	Q3450-22DL	10/27/2025	11:58	PP076073.D	10.40	4.64
AR1660CCC500	AR1660CCC500	10/27/2025	14:08	PP076078.D	10.41	4.64
I.BLK	I.BLK	10/27/2025	15:13	PP076082.D	10.41	4.64
AR1660CCC500	AR1660CCC500	10/31/2025	20:57	PP076271.D	10.42	4.65
I.BLK	I.BLK	10/31/2025	22:34	PP076275.D	10.42	4.65
PB170358BL	PB170358BL	10/31/2025	22:51	PP076276.D	10.42	4.65
PB170358BS	PB170358BS	10/31/2025	23:07	PP076277.D	10.42	4.65
SS-10R-(3.5-4.0)	Q3450-04	10/31/2025	23:40	PP076279.D	10.42	4.64
SS-10R-(3.5-4.0)MS	Q3450-04MS	10/31/2025	23:56	PP076280.D	10.42	4.64
SS-10R-(3.5-4.0)MSD	Q3450-04MSD	11/01/2025	00:12	PP076281.D	10.42	4.65
SS-13R-(3.5-4.0)	Q3450-05	11/01/2025	00:28	PP076282.D	10.42	4.65
SS-13R-(5.0-5.5)	Q3450-06	11/01/2025	00:45	PP076283.D	10.42	4.64
SS-22-(0.0-0.5)	Q3450-23	11/01/2025	01:01	PP076284.D	10.42	4.65

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Analytical Sequence

SS-22-(3.5-4.0)	Q3450-25	11/01/2025	01:17	PP076285.D	10.42	4.64
AR1660CCC500	AR1660CCC500	11/01/2025	02:22	PP076286.D	10.42	4.64
I.BLK	I.BLK	11/01/2025	04:00	PP076290.D	10.42	4.64
AR1660CCC500	AR1660CCC500	11/03/2025	08:44	PP076292.D	10.41	4.64
I.BLK	I.BLK	11/03/2025	09:53	PP076296.D	10.41	4.64
SS-22-(0.0-0.5)DL	Q3450-23DL	11/03/2025	10:09	PP076297.D	10.41	4.64
AR1660CCC500	AR1660CCC500	11/03/2025	14:19	PP076304.D	10.41	4.64
I.BLK	I.BLK	11/03/2025	15:24	PP076308.D	10.41	4.64
I.BLK	I.BLK	10/10/2025	08:42	PQ070972.D	8.50	3.40
AR1660ICC1000	AR1660ICC1000	10/10/2025	08:57	PQ070973.D	8.50	3.40
AR1660ICC750	AR1660ICC750	10/10/2025	09:11	PQ070974.D	8.50	3.40
AR1660ICC500	AR1660ICC500	10/10/2025	09:26	PQ070975.D	8.50	3.40
AR1660ICC250	AR1660ICC250	10/10/2025	09:40	PQ070976.D	8.50	3.40
AR1660ICC050	AR1660ICC050	10/10/2025	09:55	PQ070977.D	8.50	3.40
AR1221ICC500	AR1221ICC500	10/10/2025	10:09	PQ070978.D	8.50	3.40
AR1232ICC500	AR1232ICC500	10/10/2025	10:24	PQ070979.D	8.50	3.40
AR1242ICC1000	AR1242ICC1000	10/10/2025	10:38	PQ070980.D	8.50	3.40
AR1242ICC750	AR1242ICC750	10/10/2025	10:53	PQ070981.D	8.50	3.40
AR1242ICC500	AR1242ICC500	10/10/2025	11:08	PQ070982.D	8.50	3.40
AR1242ICC250	AR1242ICC250	10/10/2025	11:22	PQ070983.D	8.50	3.40
AR1242ICC050	AR1242ICC050	10/10/2025	11:37	PQ070984.D	8.50	3.40
AR1248ICC1000	AR1248ICC1000	10/10/2025	11:51	PQ070985.D	8.50	3.40
AR1248ICC750	AR1248ICC750	10/10/2025	12:06	PQ070986.D	8.50	3.40
AR1248ICC500	AR1248ICC500	10/10/2025	12:20	PQ070987.D	8.50	3.40
AR1248ICC250	AR1248ICC250	10/10/2025	12:35	PQ070988.D	8.50	3.40
AR1248ICC050	AR1248ICC050	10/10/2025	12:50	PQ070989.D	8.50	3.40
AR1254ICC1000	AR1254ICC1000	10/10/2025	13:04	PQ070990.D	8.50	3.40
AR1254ICC750	AR1254ICC750	10/10/2025	13:19	PQ070991.D	8.50	3.40
AR1254ICC500	AR1254ICC500	10/10/2025	13:33	PQ070992.D	8.50	3.40
AR1254ICC250	AR1254ICC250	10/10/2025	13:48	PQ070993.D	8.50	3.40
AR1254ICC050	AR1254ICC050	10/10/2025	15:15	PQ070994.D	8.50	3.40
AR1262ICC500	AR1262ICC500	10/10/2025	15:30	PQ070995.D	8.50	3.40
AR1268ICC1000	AR1268ICC1000	10/10/2025	15:44	PQ070996.D	8.50	3.40
AR1268ICC750	AR1268ICC750	10/10/2025	15:59	PQ070997.D	8.50	3.40
AR1268ICC500	AR1268ICC500	10/10/2025	16:13	PQ070998.D	8.50	3.40
AR1268ICC250	AR1268ICC250	10/10/2025	16:28	PQ070999.D	8.50	3.40
AR1268ICC050	AR1268ICC050	10/10/2025	16:43	PQ071000.D	8.50	3.40
AR1660CCC500	AR1660CCC500	10/27/2025	08:57	PQ071149.D	8.50	3.40
I.BLK	I.BLK	10/27/2025	09:11	PQ071150.D	8.50	3.40
SS-5R-(2.0-2.5)	Q3450-01	10/27/2025	09:26	PQ071151.D	8.50	3.40
SS-5R-(3.5-4.0)	Q3450-02	10/27/2025	09:41	PQ071152.D	8.50	3.40
SS-5R-(3.5-4.0)MS	Q3450-02MS	10/27/2025	09:55	PQ071153.D	8.50	3.40
SS-5R-(3.5-4.0)MSD	Q3450-02MSD	10/27/2025	10:10	PQ071154.D	8.50	3.40
SS-20-(0.0-0.5)	Q3450-08	10/27/2025	10:24	PQ071155.D	8.50	3.40
SS-20-(2.0-2.5)	Q3450-09	10/27/2025	10:39	PQ071156.D	8.50	3.40
AR1660CCC500	AR1660CCC500	10/27/2025	12:35	PQ071161.D	8.50	3.40
I.BLK	I.BLK	10/27/2025	14:18	PQ071165.D	8.50	3.40

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Analytical Sequence

Client: Matrix New World Engineering **SDG No.:** Q3450

Project: PSEG Athenia Substation **Instrument ID:** ECD_O

GC Column: ZB-MR2 **ID:** 0.32 (mm) **Inst. Calib. Date(s):** 11/04/2025 11/04/2025

THE ANALYTICAL SEQUENCE OF PERFORMANCE EVALUATION MIXTURES, BLANKS, SAMPLES, AND STANDARDS IS GIVEN BELOW:

CLIENT ID	LAB SAMPLE ID	DATE ANALYZED	TIME ANALYZED	DATAFILE	DCB RT #	TCX RT #
I.BLK	I.BLK	11/04/2025	09:21	PO114870.D	8.64	3.65
AR1660ICC1000	AR1660ICC1000	11/04/2025	09:39	PO114871.D	8.64	3.65
AR1660ICC750	AR1660ICC750	11/04/2025	09:58	PO114872.D	8.64	3.65
AR1660ICC500	AR1660ICC500	11/04/2025	10:16	PO114873.D	8.64	3.65
AR1660ICC250	AR1660ICC250	11/04/2025	10:34	PO114874.D	8.64	3.65
AR1660ICC050	AR1660ICC050	11/04/2025	10:53	PO114875.D	8.64	3.65
AR1221ICC500	AR1221ICC500	11/04/2025	11:11	PO114876.D	8.64	3.65
AR1232ICC500	AR1232ICC500	11/04/2025	11:30	PO114877.D	8.64	3.65
AR1242ICC1000	AR1242ICC1000	11/04/2025	11:48	PO114878.D	8.64	3.65
AR1242ICC750	AR1242ICC750	11/04/2025	12:06	PO114879.D	8.64	3.65
AR1242ICC500	AR1242ICC500	11/04/2025	12:25	PO114880.D	8.64	3.65
AR1242ICC250	AR1242ICC250	11/04/2025	12:42	PO114881.D	8.64	3.65
AR1242ICC050	AR1242ICC050	11/04/2025	13:00	PO114882.D	8.64	3.65
AR1248ICC1000	AR1248ICC1000	11/04/2025	13:19	PO114883.D	8.64	3.65
AR1248ICC750	AR1248ICC750	11/04/2025	13:37	PO114884.D	8.64	3.65
AR1248ICC500	AR1248ICC500	11/04/2025	13:56	PO114885.D	8.64	3.65
AR1248ICC250	AR1248ICC250	11/04/2025	14:14	PO114886.D	8.64	3.65
AR1248ICC050	AR1248ICC050	11/04/2025	14:33	PO114887.D	8.64	3.65
AR1254ICC1000	AR1254ICC1000	11/04/2025	14:51	PO114888.D	8.64	3.65
AR1254ICC750	AR1254ICC750	11/04/2025	15:09	PO114889.D	8.64	3.65
AR1254ICC500	AR1254ICC500	11/04/2025	15:28	PO114890.D	8.64	3.65
AR1254ICC250	AR1254ICC250	11/04/2025	16:04	PO114891.D	8.64	3.65
AR1254ICC050	AR1254ICC050	11/04/2025	16:23	PO114892.D	8.64	3.65
AR1262ICC500	AR1262ICC500	11/04/2025	17:00	PO114893.D	8.64	3.65
AR1268ICC1000	AR1268ICC1000	11/04/2025	17:18	PO114894.D	8.64	3.65
AR1268ICC750	AR1268ICC750	11/04/2025	17:36	PO114895.D	8.64	3.65
AR1268ICC500	AR1268ICC500	11/04/2025	17:55	PO114896.D	8.64	3.65
AR1268ICC250	AR1268ICC250	11/04/2025	18:13	PO114897.D	8.64	3.65
AR1268ICC050	AR1268ICC050	11/04/2025	18:31	PO114898.D	8.64	3.65
AR1660CCC500	AR1660CCC500	11/12/2025	09:49	PO115031.D	8.64	3.65
I.BLK	I.BLK	11/12/2025	11:03	PO115035.D	8.64	3.65
PB170497BL	PB170497BL	11/12/2025	11:58	PO115037.D	8.64	3.65
PB170497BS	PB170497BS	11/12/2025	12:16	PO115038.D	8.64	3.65
TP-1MS	Q3597-01MS	11/12/2025	12:53	PO115040.D	8.64	3.65
TP-1MSD	Q3597-01MSD	11/12/2025	13:11	PO115041.D	8.64	3.65
SS-13R-(4.5-5.0)	Q3450-07	11/12/2025	15:36	PO115045.D	8.64	3.65
AR1660CCC500	AR1660CCC500	11/12/2025	15:55	PO115046.D	8.64	3.65
I.BLK	I.BLK	11/12/2025	17:07	PO115050.D	8.64	3.65
I.BLK	I.BLK	10/21/2025	08:49	PP075887.D	8.78	3.78
AR1660ICC1000	AR1660ICC1000	10/21/2025	09:06	PP075888.D	8.78	3.78
AR1660ICC750	AR1660ICC750	10/21/2025	09:22	PP075889.D	8.78	3.78
AR1660ICC500	AR1660ICC500	10/21/2025	09:38	PP075890.D	8.78	3.78

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Analytical Sequence

AR1660ICC250	AR1660ICC250	10/21/2025	09:54	PP075891.D	8.78	3.78
AR1660ICC050	AR1660ICC050	10/21/2025	10:11	PP075892.D	8.78	3.78
AR1221ICC500	AR1221ICC500	10/21/2025	10:27	PP075893.D	8.78	3.78
AR1232ICC500	AR1232ICC500	10/21/2025	10:43	PP075894.D	8.78	3.78
AR1242ICC1000	AR1242ICC1000	10/21/2025	10:59	PP075895.D	8.78	3.78
AR1242ICC750	AR1242ICC750	10/21/2025	11:16	PP075896.D	8.78	3.78
AR1242ICC500	AR1242ICC500	10/21/2025	11:48	PP075897.D	8.79	3.78
AR1242ICC250	AR1242ICC250	10/21/2025	12:04	PP075898.D	8.78	3.78
AR1242ICC050	AR1242ICC050	10/21/2025	12:21	PP075899.D	8.78	3.78
AR1248ICC1000	AR1248ICC1000	10/21/2025	12:41	PP075900.D	8.78	3.78
AR1248ICC750	AR1248ICC750	10/21/2025	12:57	PP075901.D	8.78	3.78
AR1248ICC500	AR1248ICC500	10/21/2025	13:13	PP075902.D	8.78	3.78
AR1248ICC250	AR1248ICC250	10/21/2025	13:30	PP075903.D	8.78	3.78
AR1248ICC050	AR1248ICC050	10/21/2025	13:46	PP075904.D	8.78	3.78
AR1254ICC1000	AR1254ICC1000	10/21/2025	14:02	PP075905.D	8.78	3.78
AR1254ICC750	AR1254ICC750	10/21/2025	14:18	PP075906.D	8.78	3.78
AR1254ICC500	AR1254ICC500	10/21/2025	14:35	PP075907.D	8.79	3.79
AR1254ICC250	AR1254ICC250	10/21/2025	14:51	PP075908.D	8.78	3.78
AR1254ICC050	AR1254ICC050	10/21/2025	15:07	PP075909.D	8.78	3.78
AR1262ICC500	AR1262ICC500	10/21/2025	15:40	PP075910.D	8.79	3.79
AR1268ICC1000	AR1268ICC1000	10/21/2025	15:56	PP075911.D	8.79	3.78
AR1268ICC750	AR1268ICC750	10/21/2025	16:12	PP075912.D	8.78	3.78
AR1268ICC500	AR1268ICC500	10/21/2025	16:29	PP075913.D	8.79	3.79
AR1268ICC250	AR1268ICC250	10/21/2025	16:45	PP075914.D	8.79	3.79
AR1268ICC050	AR1268ICC050	10/21/2025	17:01	PP075915.D	8.79	3.79
AR1660CCC500	AR1660CCC500	10/24/2025	18:58	PP076042.D	8.78	3.78
I.BLK	I.BLK	10/24/2025	20:36	PP076046.D	8.78	3.78
SS-20-(3.5-4.0)	Q3450-10	10/24/2025	20:52	PP076047.D	8.78	3.78
SS-28-(2.0-2.5)	Q3450-11	10/24/2025	21:08	PP076048.D	8.78	3.78
SS-21-(0.0-0.5)	Q3450-20	10/24/2025	21:24	PP076049.D	8.78	3.78
SS-21-(2.0-2.5)	Q3450-21	10/24/2025	21:41	PP076050.D	8.78	3.78
SS-21-(3.5-4.0)	Q3450-22	10/24/2025	21:57	PP076051.D	8.78	3.78
AR1660CCC500	AR1660CCC500	10/25/2025	00:22	PP076057.D	8.78	3.78
I.BLK	I.BLK	10/25/2025	02:16	PP076061.D	8.78	3.78
AR1660CCC500	AR1660CCC500	10/27/2025	08:56	PP076063.D	8.78	3.78
I.BLK	I.BLK	10/27/2025	10:02	PP076067.D	8.77	3.78
PB170236BL	PB170236BL	10/27/2025	10:18	PP076068.D	8.77	3.78
PB170236BS	PB170236BS	10/27/2025	10:34	PP076069.D	8.78	3.78
SS-21-(0.0-0.5)DL	Q3450-20DL	10/27/2025	11:42	PP076072.D	8.78	3.78
SS-21-(3.5-4.0)DL	Q3450-22DL	10/27/2025	11:58	PP076073.D	8.77	3.78
AR1660CCC500	AR1660CCC500	10/27/2025	14:08	PP076078.D	8.78	3.78
I.BLK	I.BLK	10/27/2025	15:13	PP076082.D	8.77	3.78
AR1660CCC500	AR1660CCC500	10/31/2025	20:57	PP076271.D	8.78	3.78
I.BLK	I.BLK	10/31/2025	22:34	PP076275.D	8.78	3.78
PB170358BL	PB170358BL	10/31/2025	22:51	PP076276.D	8.78	3.78
PB170358BS	PB170358BS	10/31/2025	23:07	PP076277.D	8.78	3.78
SS-10R-(3.5-4.0)	Q3450-04	10/31/2025	23:40	PP076279.D	8.78	3.78
SS-10R-(3.5-4.0)MS	Q3450-04MS	10/31/2025	23:56	PP076280.D	8.78	3.78
SS-10R-(3.5-4.0)MSD	Q3450-04MSD	11/01/2025	00:12	PP076281.D	8.78	3.78
SS-13R-(3.5-4.0)	Q3450-05	11/01/2025	00:28	PP076282.D	8.78	3.78
SS-13R-(5.0-5.5)	Q3450-06	11/01/2025	00:45	PP076283.D	8.78	3.78
SS-22-(0.0-0.5)	Q3450-23	11/01/2025	01:01	PP076284.D	8.78	3.78

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Analytical Sequence

SS-22-(3.5-4.0)	Q3450-25	11/01/2025	01:17	PP076285.D	8.78	3.78
AR1660CCC500	AR1660CCC500	11/01/2025	02:22	PP076286.D	8.78	3.78
I.BLK	I.BLK	11/01/2025	04:00	PP076290.D	8.78	3.78
AR1660CCC500	AR1660CCC500	11/03/2025	08:44	PP076292.D	8.77	3.78
I.BLK	I.BLK	11/03/2025	09:53	PP076296.D	8.77	3.78
SS-22-(0.0-0.5)DL	Q3450-23DL	11/03/2025	10:09	PP076297.D	8.77	3.77
AR1660CCC500	AR1660CCC500	11/03/2025	14:19	PP076304.D	8.78	3.78
I.BLK	I.BLK	11/03/2025	15:24	PP076308.D	8.78	3.78
I.BLK	I.BLK	10/10/2025	08:42	PQ070972.D	7.53	2.78
AR1660ICC1000	AR1660ICC1000	10/10/2025	08:57	PQ070973.D	7.53	2.78
AR1660ICC750	AR1660ICC750	10/10/2025	09:11	PQ070974.D	7.53	2.78
AR1660ICC500	AR1660ICC500	10/10/2025	09:26	PQ070975.D	7.53	2.78
AR1660ICC250	AR1660ICC250	10/10/2025	09:40	PQ070976.D	7.53	2.78
AR1660ICC050	AR1660ICC050	10/10/2025	09:55	PQ070977.D	7.53	2.78
AR1221ICC500	AR1221ICC500	10/10/2025	10:09	PQ070978.D	7.53	2.78
AR1232ICC500	AR1232ICC500	10/10/2025	10:24	PQ070979.D	7.53	2.78
AR1242ICC1000	AR1242ICC1000	10/10/2025	10:38	PQ070980.D	7.53	2.78
AR1242ICC750	AR1242ICC750	10/10/2025	10:53	PQ070981.D	7.53	2.78
AR1242ICC500	AR1242ICC500	10/10/2025	11:08	PQ070982.D	7.53	2.78
AR1242ICC250	AR1242ICC250	10/10/2025	11:22	PQ070983.D	7.53	2.78
AR1242ICC050	AR1242ICC050	10/10/2025	11:37	PQ070984.D	7.53	2.78
AR1248ICC1000	AR1248ICC1000	10/10/2025	11:51	PQ070985.D	7.53	2.78
AR1248ICC750	AR1248ICC750	10/10/2025	12:06	PQ070986.D	7.53	2.78
AR1248ICC500	AR1248ICC500	10/10/2025	12:20	PQ070987.D	7.53	2.78
AR1248ICC250	AR1248ICC250	10/10/2025	12:35	PQ070988.D	7.53	2.78
AR1248ICC050	AR1248ICC050	10/10/2025	12:50	PQ070989.D	7.53	2.78
AR1254ICC1000	AR1254ICC1000	10/10/2025	13:04	PQ070990.D	7.53	2.78
AR1254ICC750	AR1254ICC750	10/10/2025	13:19	PQ070991.D	7.53	2.78
AR1254ICC500	AR1254ICC500	10/10/2025	13:33	PQ070992.D	7.53	2.78
AR1254ICC250	AR1254ICC250	10/10/2025	13:48	PQ070993.D	7.53	2.78
AR1254ICC050	AR1254ICC050	10/10/2025	15:15	PQ070994.D	7.53	2.78
AR1262ICC500	AR1262ICC500	10/10/2025	15:30	PQ070995.D	7.53	2.78
AR1268ICC1000	AR1268ICC1000	10/10/2025	15:44	PQ070996.D	7.53	2.78
AR1268ICC750	AR1268ICC750	10/10/2025	15:59	PQ070997.D	7.53	2.78
AR1268ICC500	AR1268ICC500	10/10/2025	16:13	PQ070998.D	7.53	2.78
AR1268ICC250	AR1268ICC250	10/10/2025	16:28	PQ070999.D	7.53	2.78
AR1268ICC050	AR1268ICC050	10/10/2025	16:43	PQ071000.D	7.53	2.78
AR1660CCC500	AR1660CCC500	10/27/2025	08:57	PQ071149.D	7.53	2.78
I.BLK	I.BLK	10/27/2025	09:11	PQ071150.D	7.53	2.78
SS-5R-(2.0-2.5)	Q3450-01	10/27/2025	09:26	PQ071151.D	7.53	2.78
SS-5R-(3.5-4.0)	Q3450-02	10/27/2025	09:41	PQ071152.D	7.53	2.78
SS-5R-(3.5-4.0)MS	Q3450-02MS	10/27/2025	09:55	PQ071153.D	7.53	2.78
SS-5R-(3.5-4.0)MSD	Q3450-02MSD	10/27/2025	10:10	PQ071154.D	7.53	2.78
SS-20-(0.0-0.5)	Q3450-08	10/27/2025	10:24	PQ071155.D	7.53	2.78
SS-20-(2.0-2.5)	Q3450-09	10/27/2025	10:39	PQ071156.D	7.53	2.78
AR1660CCC500	AR1660CCC500	10/27/2025	12:35	PQ071161.D	7.53	2.78
I.BLK	I.BLK	10/27/2025	14:18	PQ071165.D	7.53	2.78

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

IDENTIFICATION SUMMARY
FOR MULTICOMPONENT ANALYTES

SAMPLE NO.

PB170497BS

Lab Name: Alliance Contract: MATR02

Lab Code: ACE SDG NO.: Q3450

Lab Sample ID: PB170497BS Date(s) Analyzed: 11/12/2025 11/12/2025

Instrument ID (1): ECD_O Instrument ID (2): ECD_O

GC Column: (1): ZB-MR1 ID: 0.32 (mm) GC Column:(2): ZB-MR2 ID: 0.32 (mm)

Data file PO115038.D

ANALYTE	COL	RT	RT WINDOW FROM	RT WINDOW TO	CONCENTRATION	MEAN CONCENTRATION	%RPD
Aroclor-1016	1	5.486	5.436	5.536	147		
	2	5.508	5.458	5.558	145		
	3	5.57	5.52	5.62	149		
COLUMN 1	4	5.666	5.616	5.716	144		
	5	5.957	5.907	6.007	138	145	
	1	4.726	4.676	4.776	144		
	2	4.745	4.695	4.795	142		
	3	4.92	4.87	4.97	143		
	4	4.963	4.913	5.013	140		
COLUMN 2	5	5.175	5.125	5.225	138	142	2.09
Aroclor-1260	1	7.072	7.022	7.122	143		
	2	7.325	7.275	7.375	154		
	3	7.683	7.633	7.733	127		
COLUMN 1	4	7.904	7.854	7.954	133		
	5	8.211	8.161	8.261	128	137	
	1	6.205	6.155	6.255	140		
	2	6.393	6.343	6.443	138		
	3	6.545	6.495	6.595	140		
	4	7.016	6.966	7.066	123		
COLUMN 2	5	7.258	7.208	7.308	124	133	2.96

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

IDENTIFICATION SUMMARY
FOR MULTICOMPONENT ANALYTES

SAMPLE NO.: TP-1MS

Lab Name: Alliance | Contract: MATR02
Lab Code: ACE | SDG NO.: Q3450
Lab Sample ID: Q3597-01MS | Date(s) Analyzed: 11/12/2025 11/12/2025
Instrument ID (1): ECD_O | Instrument ID (2): ECD_O
GC Column: (1): ZB-MR1 ID: 0.32 (mm) | GC Column:(2): ZB-MR2 ID: 0.32 (mm)
Data file: PO115040.D

ANALYTE	COL	RT	RT WINDOW FROM	RT WINDOW TO	CONCENTRATION	MEAN CONCENTRATION	%RPD
Aroclor-1016	1	5.486	5.436	5.536	162		
	2	5.508	5.458	5.558	159		
	3	5.57	5.52	5.62	161		
COLUMN 1	4	5.666	5.616	5.716	164		
	5	5.958	5.908	6.008	153	160	
	1	4.727	4.677	4.777	159		
	2	4.746	4.696	4.796	160		
	3	4.92	4.87	4.97	160		
	4	4.963	4.913	5.013	155		
COLUMN 2	5	5.175	5.125	5.225	155	157	1.89
Aroclor-1260	1	7.071	7.021	7.121	154		
	2	7.325	7.275	7.375	162		
	3	7.683	7.633	7.733	134		
COLUMN 1	4	7.906	7.856	7.956	141		
	5	8.211	8.161	8.261	141	146	
	1	6.206	6.156	6.256	154		
	2	6.393	6.343	6.443	151		
	3	6.546	6.496	6.596	152		
	4	7.017	6.967	7.067	134		
COLUMN 2	5	7.258	7.208	7.308	135	145	0.69

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

IDENTIFICATION SUMMARY
FOR MULTICOMPONENT ANALYTES

SAMPLE NO.

TP-1MSD

Lab Name: Alliance Contract: MATR02

Lab Code: ACE SDG NO.: Q3450

Lab Sample ID: Q3597-01MSD Date(s) Analyzed: 11/12/2025 11/12/2025

Instrument ID (1): ECD_O Instrument ID (2): ECD_O

GC Column: (1): ZB-MR1 ID: 0.32 (mm) GC Column:(2): ZB-MR2 ID: 0.32 (mm)

Data file PO115041.D

ANALYTE	COL	RT	RT WINDOW FROM	RT WINDOW TO	CONCENTRATION	MEAN CONCENTRATION	%RPD
Aroclor-1016	1	5.487	5.437	5.537	161		
	2	5.508	5.458	5.558	159		
	3	5.57	5.52	5.62	160		
COLUMN 1	4	5.666	5.616	5.716	160		
	5	5.958	5.908	6.008	151	158	
	1	4.727	4.677	4.777	161		
	2	4.746	4.696	4.796	159		
	3	4.921	4.871	4.971	160		
	4	4.963	4.913	5.013	156		
COLUMN 2	5	5.175	5.125	5.225	155	158	0
Aroclor-1260	1	7.072	7.022	7.122	157		
	2	7.326	7.276	7.376	161		
	3	7.683	7.633	7.733	137		
COLUMN 1	4	7.905	7.855	7.955	144		
	5	8.212	8.162	8.262	139	147	
	1	6.206	6.156	6.256	153		
	2	6.393	6.343	6.443	150		
	3	6.546	6.496	6.596	152		
	4	7.017	6.967	7.067	134		
COLUMN 2	5	7.258	7.208	7.308	134	145	1.37

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

IDENTIFICATION SUMMARY
FOR MULTICOMPONENT ANALYTES

SAMPLE NO.

SS-13R-(4.5-5.0)

Lab Name: Alliance Contract: MATR02

Lab Code: ACE SDG NO.: Q3450

Lab Sample ID: Q3450-07 Date(s) Analyzed: 11/12/2025 11/12/2025

Instrument ID (1): ECD_O Instrument ID (2): ECD_O

GC Column: (1): ZB-MR1 ID: 0.32 (mm) GC Column:(2): ZB-MR2 ID: 0.32 (mm)

Data file PO115045.D

ANALYTE	COL	RT	RT WINDOW FROM	RT WINDOW TO	CONCENTRATION	MEAN CONCENTRATION	%RPD
Aroclor-1254	1	6.326	6.276	6.376	5.61		
	2	6.547	6.497	6.597	6.14		
	3	6.905	6.855	6.955	13.4		
COLUMN 1	4	7.189	7.139	7.239	12.7		
	5	7.604	7.554	7.654	31.0	13.8	
	1	5.528	5.478	5.578	4.22		
	2	5.672	5.622	5.722	5.51		
	3	6.075	6.025	6.125	10.6		
	4	6.303	6.253	6.353	11.6		
COLUMN 2	5	6.719	6.669	6.769	26.3	11.7	16.47

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

IDENTIFICATION SUMMARY
FOR MULTICOMPONENT ANALYTES

SAMPLE NO.: SS-20-(3.5-4.0)

Lab Name: Alliance | Contract: MATR02
Lab Code: ACE | SDG NO.: Q3450
Lab Sample ID: Q3450-10 | Date(s) Analyzed: 10/24/2025 10/24/2025
Instrument ID (1): ECD_P | Instrument ID (2): ECD_P
GC Column: (1): ZB-MR1 ID: 0.32 (mm) | GC Column:(2): ZB-MR2 ID: 0.32 (mm)
Data file: PP076047.D

ANALYTE	COL	RT	RT WINDOW FROM	RT WINDOW TO	CONCENTRATION	MEAN CONCENTRATION	%RPD
Aroclor-1254	1	6.639	6.589	6.689	11.5		
	2	6.856	6.806	6.906	25.9		
	3	7.218	7.168	7.268	49.4		
COLUMN 1	4	7.5	7.45	7.55	75.2		
	5	7.917	7.867	7.967	114	55.2	
	1	5.653	5.603	5.703	15.4		
	2	5.801	5.751	5.851	29.7		
	3	6.204	6.154	6.254	46.1		
	4	6.43	6.38	6.48	65.1		
COLUMN 2	5	6.847	6.797	6.897	105	52.2	5.59

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

IDENTIFICATION SUMMARY
FOR MULTICOMPONENT ANALYTES

SAMPLE NO.

SS-21-(0.0-0.5)

Lab Name: Alliance Contract: MATR02

Lab Code: ACE SDG NO.: Q3450

Lab Sample ID: Q3450-20 Date(s) Analyzed: 10/24/2025 10/24/2025

Instrument ID (1): ECD_P Instrument ID (2): ECD_P

GC Column: (1): ZB-MR1 ID: 0.32 (mm) GC Column:(2): ZB-MR2 ID: 0.32 (mm)

Data file PP076049.D

ANALYTE	COL	RT	RT WINDOW FROM	RT WINDOW TO	CONCENTRATION	MEAN CONCENTRATION	%RPD
Aroclor-1254	1	6.641	6.591	6.691	201		
	2	6.856	6.806	6.906	353		
	3	7.219	7.169	7.269	609		
COLUMN 1	4	7.501	7.451	7.551	463		
	5	7.917	7.867	7.967	615	448	
	1	5.656	5.606	5.706	296		
	2	5.804	5.754	5.854	406		
	3	6.205	6.155	6.255	721		
	4	6.431	6.381	6.481	596		
COLUMN 2	5	6.848	6.798	6.898	944	593	27.86

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

IDENTIFICATION SUMMARY
FOR MULTICOMPONENT ANALYTES

SAMPLE NO.

SS-21-(3.5-4.0)

Lab Name: Alliance Contract: MATR02

Lab Code: ACE SDG NO.: Q3450

Lab Sample ID: Q3450-22 Date(s) Analyzed: 10/24/2025 10/24/2025

Instrument ID (1): ECD_P Instrument ID (2): ECD_P

GC Column: (1): ZB-MR1 ID: 0.32 (mm) GC Column:(2): ZB-MR2 ID: 0.32 (mm)

Data file PP076051.D

ANALYTE	COL	RT	RT WINDOW FROM	RT WINDOW TO	CONCENTRATION	MEAN CONCENTRATION	%RPD
Aroclor-1254	1	6.64	6.59	6.69	59.4		
	2	6.856	6.806	6.906	164		
	3	7.218	7.168	7.268	372		
COLUMN 1	4	7.5	7.45	7.55	531		
	5	7.916	7.866	7.966	888	403	
	1	5.657	5.607	5.707	89.0		
	2	5.804	5.754	5.854	170		
	3	6.204	6.154	6.254	435		
	4	6.431	6.381	6.481	673		
COLUMN 2	5	6.848	6.798	6.898	1220	516	24.59

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

IDENTIFICATION SUMMARY
FOR MULTICOMPONENT ANALYTES

SAMPLE NO.

PB170236BS

Lab Name: Alliance Contract: MATR02

Lab Code: ACE SDG NO.: Q3450

Lab Sample ID: PB170236BS Date(s) Analyzed: 10/27/2025 10/27/2025

Instrument ID (1): ECD_P Instrument ID (2): ECD_P

GC Column: (1): ZB-MR1 ID: 0.32 (mm) GC Column:(2): ZB-MR2 ID: 0.32 (mm)

Data file PP076069.D

ANALYTE	COL	RT	RT WINDOW FROM	RT WINDOW TO	CONCENTRATION	MEAN CONCENTRATION	%RPD
Aroclor-1016	1	5.793	5.743	5.843	146		
	2	5.814	5.764	5.864	147		
	3	5.877	5.827	5.927	145		
COLUMN 1	4	5.974	5.924	6.024	146		
	5	6.266	6.216	6.316	140	145	
	1	4.877	4.827	4.927	132		
	2	4.935	4.885	4.985	131		
	3	5.054	5.004	5.104	122		
	4	5.095	5.045	5.145	129		
COLUMN 2	5	5.309	5.259	5.359	120	127	13.24
Aroclor-1260	1	7.383	7.333	7.433	167		
	2	7.636	7.586	7.686	153		
	3	7.994	7.944	8.044	134		
COLUMN 1	4	8.22	8.17	8.27	157		
	5	8.546	8.496	8.596	135	149	
	1	6.335	6.285	6.385	147		
	2	6.522	6.472	6.572	147		
	3	6.675	6.625	6.725	152		
	4	7.144	7.094	7.194	142		
COLUMN 2	5	7.384	7.334	7.434	129	143	4.11

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

IDENTIFICATION SUMMARY
FOR MULTICOMPONENT ANALYTES

SAMPLE NO.: SS-21-(0.0-0.5)DL

Lab Name: Alliance | Contract: MATR02
Lab Code: ACE | SDG NO.: Q3450
Lab Sample ID: Q3450-20DL | Date(s) Analyzed: 10/27/2025 10/27/2025
Instrument ID (1): ECD_P | Instrument ID (2): ECD_P
GC Column: (1): ZB-MR1 ID: 0.32 (mm) | GC Column:(2): ZB-MR2 ID: 0.32 (mm)
Data file: PP076072.D

ANALYTE	COL	RT	RT WINDOW FROM	RT WINDOW TO	CONCENTRATION	MEAN CONCENTRATION	%RPD
Aroclor-1254	1	6.641	6.591	6.691	262		
	2	6.856	6.806	6.906	401		
	3	7.219	7.169	7.269	657		
COLUMN 1	4	7.5	7.45	7.55	524		
	5	7.917	7.867	7.967	699	509	
	1	5.657	5.607	5.707	308		
	2	5.804	5.754	5.854	391		
	3	6.205	6.155	6.255	588		
	4	6.431	6.381	6.481	497		
COLUMN 2	5	6.848	6.798	6.898	590	475	6.91

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

IDENTIFICATION SUMMARY
FOR MULTICOMPONENT ANALYTES

SAMPLE NO.

SS-21-(3.5-4.0)DL

Lab Name: Alliance Contract: MATR02

Lab Code: ACE SDG NO.: Q3450

Lab Sample ID: Q3450-22DL Date(s) Analyzed: 10/27/2025 10/27/2025

Instrument ID (1): ECD_P Instrument ID (2): ECD_P

GC Column: (1): ZB-MR1 ID: 0.32 (mm) GC Column:(2): ZB-MR2 ID: 0.32 (mm)

Data file PP076073.D

ANALYTE	COL	RT	RT WINDOW FROM	RT WINDOW TO	CONCENTRATION	MEAN CONCENTRATION	%RPD
Aroclor-1254	1	6.635	6.585	6.685	95.0		
	2	6.851	6.801	6.901	184		
	3	7.214	7.164	7.264	474		
COLUMN 1	4	7.497	7.447	7.547	581		
	5	7.913	7.863	7.963	916	450	
	1	5.667	5.617	5.717	143		
	2	5.799	5.749	5.849	172		
	3	6.202	6.152	6.252	434		
	4	6.428	6.378	6.478	672		
COLUMN 2	5	6.844	6.794	6.894	1100	503	11.12

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

IDENTIFICATION SUMMARY
FOR MULTICOMPONENT ANALYTES

SAMPLE NO.

PB170358BS

Lab Name: Alliance Contract: MATR02

Lab Code: ACE SDG NO.: Q3450

Lab Sample ID: PB170358BS Date(s) Analyzed: 10/31/2025 10/31/2025

Instrument ID (1): ECD_P Instrument ID (2): ECD_P

GC Column: (1): ZB-MR1 ID: 0.32 (mm) GC Column:(2): ZB-MR2 ID: 0.32 (mm)

Data file PP076277.D

ANALYTE	COL	RT	RT WINDOW FROM	RT WINDOW TO	CONCENTRATION	MEAN CONCENTRATION	%RPD
Aroclor-1016	1	5.798	5.748	5.848	172		
	2	5.819	5.769	5.869	170		
	3	5.882	5.832	5.932	172		
COLUMN 1	4	5.979	5.929	6.029	174		
	5	6.272	6.222	6.322	164	170	
	1	4.875	4.825	4.925	156		
	2	4.933	4.883	4.983	151		
	3	5.052	5.002	5.102	140		
	4	5.096	5.046	5.146	169		
COLUMN 2	5	5.308	5.258	5.358	140	151	11.84
Aroclor-1260	1	7.39	7.34	7.44	184		
	2	7.643	7.593	7.693	173		
	3	8.001	7.951	8.051	152		
COLUMN 1	4	8.228	8.178	8.278	173		
	5	8.555	8.505	8.605	150	166	
	1	6.336	6.286	6.386	168		
	2	6.523	6.473	6.573	167		
	3	6.676	6.626	6.726	174		
	4	7.145	7.095	7.195	179		
COLUMN 2	5	7.385	7.335	7.435	155	168	1.2

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

IDENTIFICATION SUMMARY
FOR MULTICOMPONENT ANALYTES

SAMPLE NO.: SS-10R-(3.5-4.0)MS

Lab Name: Alliance
Contract: MATR02
Lab Code: ACE
SDG NO.: Q3450
Lab Sample ID: Q3450-04MS
Date(s) Analyzed: 10/31/2025 10/31/2025
Instrument ID (1): ECD_P
Instrument ID (2): ECD_P
GC Column: (1): ZB-MR1 ID: 0.32 (mm)
GC Column:(2): ZB-MR2 ID: 0.32 (mm)
Data file: PP076280.D

ANALYTE	COL	RT	RT WINDOW FROM	RT WINDOW TO	CONCENTRATION	MEAN CONCENTRATION	%RPD
Aroclor-1016	1	5.797	5.747	5.847	195		
	2	5.819	5.769	5.869	195		
	3	5.881	5.831	5.931	192		
COLUMN 1	4	5.979	5.929	6.029	197		
	5	6.271	6.221	6.321	184	193	
	1	4.876	4.826	4.926	185		
	2	4.933	4.883	4.983	179		
	3	5.054	5.004	5.104	161		
	4	5.095	5.045	5.145	182		
COLUMN 2	5	5.308	5.258	5.358	169	175	9.78
Aroclor-1260	1	7.389	7.339	7.439	230		
	2	7.642	7.592	7.692	230		
	3	8	7.95	8.05	184		
COLUMN 1	4	8.227	8.177	8.277	278		
	5	8.554	8.504	8.604	184	221	
	1	6.337	6.287	6.387	222		
	2	6.523	6.473	6.573	233		
	3	6.675	6.625	6.725	234		
	4	7.145	7.095	7.195	158		
COLUMN 2	5	7.385	7.335	7.435	196	209	5.58

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

IDENTIFICATION SUMMARY
FOR MULTICOMPONENT ANALYTES

SAMPLE NO.

SS-10R-(3.5-4.0)MSD

Lab Name: Alliance Contract: MATR02

Lab Code: ACE SDG NO.: Q3450

Lab Sample ID: Q3450-04MSD Date(s) Analyzed: 11/01/2025 11/01/2025

Instrument ID (1): ECD_P Instrument ID (2): ECD_P

GC Column: (1): ZB-MR1 ID: 0.32 (mm) GC Column:(2): ZB-MR2 ID: 0.32 (mm)

Data file PP076281.D

ANALYTE	COL	RT	RT WINDOW FROM	RT WINDOW TO	CONCENTRATION	MEAN CONCENTRATION	%RPD
Aroclor-1016	1	5.798	5.748	5.848	191		
	2	5.819	5.769	5.869	190		
	3	5.882	5.832	5.932	187		
COLUMN 1	4	5.979	5.929	6.029	195		
	5	6.272	6.222	6.322	181	189	
	1	4.876	4.826	4.926	180		
	2	4.933	4.883	4.983	178		
	3	5.054	5.004	5.104	163		
	4	5.095	5.045	5.145	179		
COLUMN 2	5	5.308	5.258	5.358	166	173	8.84
Aroclor-1260	1	7.39	7.34	7.44	224		
	2	7.643	7.593	7.693	224		
	3	8	7.95	8.05	179		
COLUMN 1	4	8.227	8.177	8.277	270		
	5	8.554	8.504	8.604	179	215	
	1	6.336	6.286	6.386	224		
	2	6.523	6.473	6.573	231		
	3	6.676	6.626	6.726	226		
	4	7.145	7.095	7.195	156		
COLUMN 2	5	7.385	7.335	7.435	191	206	4.28

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

IDENTIFICATION SUMMARY
FOR MULTICOMPONENT ANALYTES

SAMPLE NO.: SS-22-(0.0-0.5)

Lab Name: Alliance | Contract: MATR02
Lab Code: ACE | SDG NO.: Q3450
Lab Sample ID: Q3450-23 | Date(s) Analyzed: 11/01/2025 11/01/2025
Instrument ID (1): ECD_P | Instrument ID (2): ECD_P
GC Column: (1): ZB-MR1 ID: 0.32 (mm) | GC Column:(2): ZB-MR2 ID: 0.32 (mm)
Data file: PP076284.D

ANALYTE	COL	RT	RT WINDOW FROM	RT WINDOW TO	CONCENTRATION	MEAN CONCENTRATION	%RPD
Aroclor-1254	1	6.646	6.596	6.696	564		
	2	6.863	6.813	6.913	739		
	3	7.226	7.176	7.276	994		
COLUMN 1	4	7.508	7.458	7.558	1030		
	5	7.924	7.874	7.974	1460	956	
	1	5.659	5.609	5.709	743		
	2	5.805	5.755	5.855	888		
	3	6.208	6.158	6.258	1240		
	4	6.434	6.384	6.484	1350		
COLUMN 2	5	6.851	6.801	6.901	1980	1200	22.63

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

IDENTIFICATION SUMMARY
FOR MULTICOMPONENT ANALYTES

SAMPLE NO.: SS-22-(3.5-4.0)

Lab Name: Alliance | Contract: MATR02
Lab Code: ACE | SDG NO.: Q3450
Lab Sample ID: Q3450-25 | Date(s) Analyzed: 11/01/2025 11/01/2025
Instrument ID (1): ECD_P | Instrument ID (2): ECD_P
GC Column: (1): ZB-MR1 ID: 0.32 (mm) | GC Column:(2): ZB-MR2 ID: 0.32 (mm)
Data file: PP076285.D

ANALYTE	COL	RT	RT WINDOW FROM	RT WINDOW TO	CONCENTRATION	MEAN CONCENTRATION	%RPD
Aroclor-1254	1	6.642	6.592	6.692	118		
	2	6.859	6.809	6.909	196		
	3	7.221	7.171	7.271	303		
COLUMN 1	4	7.503	7.453	7.553	407		
	5	7.92	7.87	7.97	581	321	
	1	5.654	5.604	5.704	152		
	2	5.801	5.751	5.851	208		
	3	6.203	6.153	6.253	342		
	4	6.43	6.38	6.48	487		
COLUMN 2	5	6.847	6.797	6.897	715	381	17.09

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

IDENTIFICATION SUMMARY
FOR MULTICOMPONENT ANALYTES

SAMPLE NO.

SS-22-(0.0-0.5)DL

Lab Name:	Alliance	**Contract:**	MATR02
Lab Code:	ACE	**SDG NO.:**	Q3450
Lab Sample ID:	Q3450-23DL	**Date(s) Analyzed:**	11/03/2025 11/03/2025
Instrument ID (1):	ECD_P	**Instrument ID (2):**	ECD_P
GC Column: (1):	ZB-MR1 **ID:** 0.32 (mm)	**GC Column:(2):**	ZB-MR2 **ID:** 0.32 (mm)
Data file	PP076297.D		

ANALYTE	COL	RT	RT WINDOW FROM	RT WINDOW TO	CONCENTRATION	MEAN CONCENTRATION	%RPD
Aroclor-1254	1	6.637	6.587	6.687	479		
	2	6.853	6.803	6.903	633		
	3	7.216	7.166	7.266	827		
COLUMN 1	4	7.498	7.448	7.548	866		
	5	7.914	7.864	7.964	1210	802	
	1	5.652	5.602	5.702	527		
	2	5.798	5.748	5.848	658		
	3	6.2	6.15	6.25	868		
	4	6.426	6.376	6.476	976		
COLUMN 2	5	6.842	6.792	6.892	1380	881	9.39

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

IDENTIFICATION SUMMARY
FOR MULTICOMPONENT ANALYTES

SAMPLE NO.

SS-5R-(2.0-2.5)

Lab Name: Alliance Contract: MATR02

Lab Code: ACE SDG NO.: Q3450

Lab Sample ID: Q3450-01 Date(s) Analyzed: 10/27/2025 10/27/2025

Instrument ID (1): ECD_Q Instrument ID (2): ECD_Q

GC Column: (1): ZB-MR1 ID: 0.32 (mm) GC Column:(2): ZB-MR2 ID: 0.32 (mm)

Data file PQ071151.D

ANALYTE	COL	RT	RT WINDOW FROM	RT WINDOW TO	CONCENTRATION	MEAN CONCENTRATION	%RPD
Aroclor-1254	1	5.293	5.243	5.343	8.30		
	2	5.507	5.457	5.557	23.0		
	3	5.855	5.805	5.905	49.1		
COLUMN 1	4	6.136	6.086	6.186	55.7		
	5	6.544	6.494	6.594	91.2	45.5	
	1	4.531	4.481	4.581	10.5		
	2	4.677	4.627	4.727	23.7		
	3	5.057	5.007	5.107	45.6		
	4	5.284	5.234	5.334	55.2		
COLUMN 2	5	5.688	5.638	5.738	85.7	44.1	3.12

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

IDENTIFICATION SUMMARY
FOR MULTICOMPONENT ANALYTES

SAMPLE NO.: SS-5R-(3.5-4.0)

Lab Name: Alliance | Contract: MATR02
Lab Code: ACE | SDG NO.: Q3450
Lab Sample ID: Q3450-02 | Date(s) Analyzed: 10/27/2025 10/27/2025
Instrument ID (1): ECD_Q | Instrument ID (2): ECD_Q
GC Column: (1): ZB-MR1 ID: 0.32 (mm) | GC Column:(2): ZB-MR2 ID: 0.32 (mm)
Data file: PQ071152.D

ANALYTE	COL	RT	RT WINDOW FROM	RT WINDOW TO	CONCENTRATION	MEAN CONCENTRATION	%RPD
Aroclor-1254	1	5.291	5.241	5.341	3.69		
	2	5.507	5.457	5.557	7.74		
	3	5.854	5.804	5.904	12.4		
COLUMN 1	4	6.134	6.084	6.184	12.9		
	5	6.543	6.493	6.593	20.0	11.3	
	1	4.531	4.481	4.581	2.91		
	2	4.678	4.628	4.728	7.27		
	3	5.058	5.008	5.108	11.5		
	4	5.285	5.235	5.335	12.9		
COLUMN 2	5	5.688	5.638	5.738	19.9	10.9	3.6

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

IDENTIFICATION SUMMARY
FOR MULTICOMPONENT ANALYTES

SAMPLE NO.: SS-5R-(3.5-4.0)MS

Lab Name: Alliance
Contract: MATR02
Lab Code: ACE
SDG NO.: Q3450
Lab Sample ID: Q3450-02MS
Date(s) Analyzed: 10/27/2025 10/27/2025
Instrument ID (1): ECD_Q
Instrument ID (2): ECD_Q
GC Column: (1): ZB-MR1 ID: 0.32 (mm)
GC Column:(2): ZB-MR2 ID: 0.32 (mm)
Data file: PQ071153.D

ANALYTE	COL	RT	RT WINDOW FROM	RT WINDOW TO	CONCENTRATION	MEAN CONCENTRATION	%RPD
Aroclor-1016	1	4.481	4.431	4.531	59.9		
	2	4.498	4.448	4.548	123		
	3	4.555	4.505	4.605	140		
COLUMN 1	4	4.645	4.595	4.695	148		
	5	4.93	4.88	4.98	135	121	
	1	3.775	3.725	3.825	131		
	2	3.791	3.741	3.841	110		
	3	3.951	3.901	4.001	119		
	4	3.999	3.949	4.049	115		
COLUMN 2	5	4.195	4.145	4.245	114	118	2.51
Aroclor-1254	1	5.291	5.241	5.341	73.6		
	2	5.506	5.456	5.556	65.1		
	3	5.854	5.804	5.904	52.0		
COLUMN 1	4	6.132	6.082	6.182	65.2		
	5	6.542	6.492	6.592	271	105	
	1	4.531	4.481	4.581	57.2		
	2	4.677	4.627	4.727	82.7		
	3	5.058	5.008	5.108	47.0		
	4	5.284	5.234	5.334	62.9		
COLUMN 2	5	5.687	5.637	5.737	209	91.7	13.52

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

IDENTIFICATION SUMMARY

FOR MULTICOMPONENT ANALYTES

Aroclor-1260	1	6.018	5.968	6.068	162		
	2	6.275	6.225	6.325	162		
	3	6.626	6.576	6.676	125		
COLUMN 1	4	6.84	6.79	6.89	137		
	5	7.146	7.096	7.196	139	145	
	1	5.186	5.136	5.236	135		
	2	5.377	5.327	5.427	123		
	3	5.516	5.466	5.566	141		
	4	5.978	5.928	6.028	109		
COLUMN 2	5	6.222	6.172	6.272	107	123	16.42

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

IDENTIFICATION SUMMARY
FOR MULTICOMPONENT ANALYTES

SAMPLE NO.

SS-5R-(3.5-4.0)MSD

Lab Name: Alliance
Contract: MATR02
Lab Code: ACE
SDG NO.: Q3450
Lab Sample ID: Q3450-02MSD
Date(s) Analyzed: 10/27/2025 10/27/2025
Instrument ID (1): ECD_Q
Instrument ID (2): ECD_Q
GC Column: (1): ZB-MR1 ID: 0.32 (mm)
GC Column:(2): ZB-MR2 ID: 0.32 (mm)
Data file PQ071154.D

ANALYTE	COL	RT	RT WINDOW FROM	RT WINDOW TO	CONCENTRATION	MEAN CONCENTRATION	%RPD
Aroclor-1016	1	4.478	4.428	4.528	155		
	2	4.497	4.447	4.547	160		
	3	4.555	4.505	4.605	166		
COLUMN 1	4	4.645	4.595	4.695	166		
	5	4.93	4.88	4.98	161	162	
	1	3.775	3.725	3.825	150		
	2	3.791	3.741	3.841	128		
	3	3.952	3.902	4.002	153		
	4	4	3.95	4.05	150		
COLUMN 2	5	4.195	4.145	4.245	136	143	12.46
Aroclor-1254	1	5.292	5.242	5.342	91.0		
	2	5.507	5.457	5.557	89.9		
	3	5.854	5.804	5.904	99.6		
COLUMN 1	4	6.134	6.084	6.184	114		
	5	6.543	6.493	6.593	347	148	
	1	4.531	4.481	4.581	74.2		
	2	4.678	4.628	4.728	97.9		
	3	5.058	5.008	5.108	90.8		
	4	5.284	5.234	5.334	102		
COLUMN 2	5	5.687	5.637	5.737	280	129	13.72

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

IDENTIFICATION SUMMARY

FOR MULTICOMPONENT ANALYTES

Aroclor-1260	1	6.019	5.969	6.069	196		
	2	6.275	6.225	6.325	197		
	3	6.625	6.575	6.675	137		
COLUMN 1	4	6.84	6.79	6.89	153		
	5	7.145	7.095	7.195	147	166	
	1	5.187	5.137	5.237	183		
	2	5.376	5.326	5.426	165		
	3	5.516	5.466	5.566	176		
	4	5.977	5.927	6.027	121		
COLUMN 2	5	6.221	6.171	6.271	120	153	8.15

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

IDENTIFICATION SUMMARY
FOR MULTICOMPONENT ANALYTES

SAMPLE NO.

SS-20-(0.0-0.5)

Lab Name: Alliance Contract: MATR02

Lab Code: ACE SDG NO.: Q3450

Lab Sample ID: Q3450-08 Date(s) Analyzed: 10/27/2025 10/27/2025

Instrument ID (1): ECD_Q Instrument ID (2): ECD_Q

GC Column: (1): ZB-MR1 ID: 0.32 (mm) GC Column:(2): ZB-MR2 ID: 0.32 (mm)

Data file PQ071155.D

ANALYTE	COL	RT	RT WINDOW FROM	RT WINDOW TO	CONCENTRATION	MEAN CONCENTRATION	%RPD
Aroclor-1254	1	5.292	5.242	5.342	97.3		
	2	5.506	5.456	5.556	127		
	3	5.853	5.803	5.903	201		
COLUMN 1	4	6.134	6.084	6.184	168		
	5	6.542	6.492	6.592	277	174	
	1	4.531	4.481	4.581	92.8		
	2	4.677	4.627	4.727	138		
	3	5.058	5.008	5.108	182		
	4	5.284	5.234	5.334	165		
COLUMN 2	5	5.687	5.637	5.737	253	166	4.71

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

IDENTIFICATION SUMMARY
FOR MULTICOMPONENT ANALYTES

SAMPLE NO.

SS-20-(2.0-2.5)

Lab Name: Alliance Contract: MATR02

Lab Code: ACE SDG NO.: Q3450

Lab Sample ID: Q3450-09 Date(s) Analyzed: 10/27/2025 10/27/2025

Instrument ID (1): ECD_Q Instrument ID (2): ECD_Q

GC Column: (1): ZB-MR1 ID: 0.32 (mm) GC Column:(2): ZB-MR2 ID: 0.32 (mm)

Data file PQ071156.D

ANALYTE	COL	RT	RT WINDOW FROM	RT WINDOW TO	CONCENTRATION	MEAN CONCENTRATION	%RPD
Aroclor-1254	1	5.292	5.242	5.342	4.87		
	2	5.506	5.456	5.556	6.00		
	3	5.854	5.804	5.904	12.7		
COLUMN 1	4	6.134	6.084	6.184	10.3		
	5	6.542	6.492	6.592	22.0	11.2	
	1	4.53	4.48	4.58	3.99		
	2	4.677	4.627	4.727	8.08		
	3	5.057	5.007	5.107	10.7		
	4	5.283	5.233	5.333	11.4		
COLUMN 2	5	5.687	5.637	5.737	22.1	11.3	0.89

QC SAMPLE DATA

Raw Data: **PP076068.D**

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Matrix New World Engineering	Date Collected:	
Project:	PSEG Athenia Substation	Date Received:	
Client Sample ID:	PB170236BL	SDG No.:	Q3450
Lab Sample ID:	PB170236BL	Matrix:	SOIL
Analytical Method:	8082A	% Solid:	100
Sample Wt/Vol:	30.02 g	Final Vol: 10000 uL	Test: PCB
Prep Method:	SW3541B	Prep Date: 10/24/25	

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
12674-11-2	Aroclor-1016	3.90	U	1	3.90	17.0	ug/kg	10/27/25 10:18	PB170236
11104-28-2	Aroclor-1221	4.00	U	1	4.00	17.0	ug/kg	10/27/25 10:18	PB170236
11141-16-5	Aroclor-1232	3.70	U	1	3.70	17.0	ug/kg	10/27/25 10:18	PB170236
53469-21-9	Aroclor-1242	4.00	U	1	4.00	17.0	ug/kg	10/27/25 10:18	PB170236
12672-29-6	Aroclor-1248	5.90	U	1	5.90	17.0	ug/kg	10/27/25 10:18	PB170236
11097-69-1	Aroclor-1254	3.20	U	1	3.20	17.0	ug/kg	10/27/25 10:18	PB170236
37324-23-5	Aroclor-1262	5.00	U	1	5.00	17.0	ug/kg	10/27/25 10:18	PB170236
11100-14-4	Aroclor-1268	3.60	U	1	3.60	17.0	ug/kg	10/27/25 10:18	PB170236
11096-82-5	Aroclor-1260	3.20	U	1	3.20	17.0	ug/kg	10/27/25 10:18	PB170236
SURROGATES									
877-09-8	Tetrachloro-m-xylene	27.3			30 (21) - 150 (165)	137%	SPK: 20		
2051-24-3	Decachlorobiphenyl	29.0			30 (10) - 150 (170)	145%	SPK: 20		

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
E = Value Exceeds Calibration Range
P = Indicates >25% difference for detected concentrations between the two GC columns
Q = indicates LCS control criteria did not meet requirements
M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
B = Analyte Found in Associated Method Blank
N = Presumptive Evidence of a Compound
* = Values outside of QC limits
D = Dilution
S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.
() = Laboratory InHouse Limit

Sample Results: **PP076068.D**

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102725\
Data File : PP076068.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Oct 2025 10:18
Operator : YP\AJ
Sample : PB170236BL
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
PB170236BL

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/28/2025
Supervised By :mohammad ahmed 10/29/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 27 11:02:59 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 24 15:40:32 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.637	3.780	39070596	172.6E6	27.339m	26.077
2)	SA Decachlor...	10.403	8.773	31178050	207.4E6	28.962m	25.220
	Target Compounds						

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Sample Results: **PP076068.D**

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102725\
Data File : PP076068.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Oct 2025 10:18
Operator : YP\AJ
Sample : PB170236BL
Misc :
ALS Vial : 7 Sample Multiplier: 1


Instrument :
ECD_P
ClientSampleId :
PB170236BL


Manual Integrations APPROVED


Reviewed By :Yogesh Patel 10/28/2025
Supervised By :mohammad ahmed 10/29/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 27 11:02:59 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 24 15:40:32 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm








#1 Tetrachloro-m-xylene

R.T.: 4.637 min
Delta R.T.: -0.005 min
Response: 39070596
Conc: 27.34 ng/ml

Instrument :
ECD_P
ClientSampleId :
PB170236BL


Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/28/2025
Supervised By :mohammad ahmed 10/29/2025




#1 Tetrachloro-m-xylene

R.T.: 3.780 min
Delta R.T.: -0.002 min
Response: 172582422
Conc: 26.08 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.403 min
Delta R.T.: -0.015 min
Response: 31178050
Conc: 28.96 ng/ml m



#2 Decachlorobiphenyl

R.T.: 8.773 min
Delta R.T.: -0.009 min
Response: 207383403
Conc: 25.22 ng/ml

Raw Data: **PP076276.D**

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Matrix New World Engineering	Date Collected:	
Project:	PSEG Athenia Substation	Date Received:	
Client Sample ID:	PB170358BL	SDG No.:	Q3450
Lab Sample ID:	PB170358BL	Matrix:	SOIL
Analytical Method:	8082A	% Solid:	100
Sample Wt/Vol:	30.01 g	Final Vol: 10000 uL	Test: PCB
Prep Method:	SW3541B	Prep Date: 10/31/25	

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
12674-11-2	Aroclor-1016	3.90	U	1	3.90	17.0	ug/kg	10/31/25 22:51	PB170358
11104-28-2	Aroclor-1221	4.00	U	1	4.00	17.0	ug/kg	10/31/25 22:51	PB170358
11141-16-5	Aroclor-1232	3.70	U	1	3.70	17.0	ug/kg	10/31/25 22:51	PB170358
53469-21-9	Aroclor-1242	4.00	U	1	4.00	17.0	ug/kg	10/31/25 22:51	PB170358
12672-29-6	Aroclor-1248	5.90	U	1	5.90	17.0	ug/kg	10/31/25 22:51	PB170358
11097-69-1	Aroclor-1254	3.20	U	1	3.20	17.0	ug/kg	10/31/25 22:51	PB170358
37324-23-5	Aroclor-1262	5.00	U	1	5.00	17.0	ug/kg	10/31/25 22:51	PB170358
11100-14-4	Aroclor-1268	3.60	U	1	3.60	17.0	ug/kg	10/31/25 22:51	PB170358
11096-82-5	Aroclor-1260	3.20	U	1	3.20	17.0	ug/kg	10/31/25 22:51	PB170358
SURROGATES									
877-09-8	Tetrachloro-m-xylene	28.4			30 (21) - 150 (165)	142%	SPK: 20		
2051-24-3	Decachlorobiphenyl	27.5			30 (10) - 150 (170)	138%	SPK: 20		

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
E = Value Exceeds Calibration Range
P = Indicates >25% difference for detected concentrations between the two GC columns
Q = indicates LCS control criteria did not meet requirements
M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
B = Analyte Found in Associated Method Blank
N = Presumptive Evidence of a Compound
* = Values outside of QC limits
D = Dilution
S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.
() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP103125\
Data File : PP076276.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 31 Oct 2025 22:51
Operator : YP\AJ
Sample : PB170358BL
Misc :
ALS Vial : 26 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
PB170358BL

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 31 23:28:33 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 24 15:40:32 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.646	3.782	40546452	170.9E6	28.372	25.825
2)	SA Decachlor...	10.423	8.779	29622897	189.6E6	27.517	23.059
	Target Compounds						

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Sample Results: **PP076276.D**


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP103125\
Data File : PP076276.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 31 Oct 2025 22:51
Operator : YP\AJ
Sample : PB170358BL
Misc :
ALS Vial : 26 Sample Multiplier: 1


Instrument :
ECD_P
ClientSampleId :
PB170358BL

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 31 23:28:33 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 24 15:40:32 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm








#1 Tetrachloro-m-xylene


R.T.: 4.646 min
Delta R.T.: 0.004 min
Response: 40546452
Conc: 28.37 ng/ml

Instrument :
ECD_P
ClientSampleId :
PB170358BL




#1 Tetrachloro-m-xylene

R.T.: 3.782 min
Delta R.T.: 0.000 min
Response: 170917150
Conc: 25.82 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.423 min
Delta R.T.: 0.004 min
Response: 29622897
Conc: 27.52 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.779 min
Delta R.T.: -0.003 min
Response: 189616356
Conc: 23.06 ng/ml

Raw Data: **PO115037.D**

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Matrix New World Engineering	Date Collected:	
Project:	PSEG Athenia Substation	Date Received:	
Client Sample ID:	PB170497BL	SDG No.:	Q3450
Lab Sample ID:	PB170497BL	Matrix:	SOIL
Analytical Method:	8082A	% Solid:	100
Sample Wt/Vol:	30.01 g	Final Vol: 10000 uL	Test: PCB
Prep Method:	SW3541B	Prep Date: 11/12/25	

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
12674-11-2	Aroclor-1016	3.90	U	1	3.90	17.0	ug/kg	11/12/25 11:58	PB170497
11104-28-2	Aroclor-1221	4.00	U	1	4.00	17.0	ug/kg	11/12/25 11:58	PB170497
11141-16-5	Aroclor-1232	3.70	U	1	3.70	17.0	ug/kg	11/12/25 11:58	PB170497
53469-21-9	Aroclor-1242	4.00	U	1	4.00	17.0	ug/kg	11/12/25 11:58	PB170497
12672-29-6	Aroclor-1248	5.90	U	1	5.90	17.0	ug/kg	11/12/25 11:58	PB170497
11097-69-1	Aroclor-1254	3.20	U	1	3.20	17.0	ug/kg	11/12/25 11:58	PB170497
37324-23-5	Aroclor-1262	5.00	U	1	5.00	17.0	ug/kg	11/12/25 11:58	PB170497
11100-14-4	Aroclor-1268	3.60	U	1	3.60	17.0	ug/kg	11/12/25 11:58	PB170497
11096-82-5	Aroclor-1260	3.20	U	1	3.20	17.0	ug/kg	11/12/25 11:58	PB170497
SURROGATES									
877-09-8	Tetrachloro-m-xylene	22.0			30 (21) - 150 (165)	110%	SPK: 20		
2051-24-3	Decachlorobiphenyl	23.0			30 (10) - 150 (170)	115%	SPK: 20		

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
E = Value Exceeds Calibration Range
P = Indicates >25% difference for detected concentrations between the two GC columns
Q = indicates LCS control criteria did not meet requirements
M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
B = Analyte Found in Associated Method Blank
N = Presumptive Evidence of a Compound
* = Values outside of QC limits
D = Dilution
S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.
() = Laboratory InHouse Limit

Sample Results: **PO115037.D**

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO111225\
Data File : PO115037.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Nov 2025 11:58
Operator : YP/AJ
Sample : PB170497BL
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
PB170497BL

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 12 21:43:39 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Nov 05 06:00:07 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.345	3.648	154.6E6	95492976	21.964	21.378
2)	SA Decachlor...	9.931	8.640	95337342	89842725	21.568	23.014
	Target Compounds						


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO111225\
Data File : PO115037.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Nov 2025 11:58
Operator : YP/AJ
Sample : PB170497BL
Misc :
ALS Vial : 8 Sample Multiplier: 1


Instrument :
ECD_O
ClientSampleId :
PB170497BL

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 12 21:43:39 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Nov 05 06:00:07 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm








#1 Tetrachloro-m-xylene


R.T.: 4.345 min
Delta R.T.: -0.003 min
Response: 154584742
Conc: 21.96 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB170497BL




#1 Tetrachloro-m-xylene

R.T.: 3.648 min
Delta R.T.: 0.000 min
Response: 95492976
Conc: 21.38 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.931 min
Delta R.T.: 0.000 min
Response: 95337342
Conc: 21.57 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.640 min
Delta R.T.: 0.000 min
Response: 89842725
Conc: 23.01 ng/ml

Raw Data: **PO114870.D**

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Matrix New World Engineering	Date Collected:	11/04/25
Project:	PSEG Athenia Substation	Date Received:	11/04/25
Client Sample ID:	PIBLK-PO114870.D	SDG No.:	Q3450
Lab Sample ID:	I.BLK-PO114870.D	Matrix:	WATER
Analytical Method:	8082A	% Solid:	0
Sample Wt/Vol:	1000 mL Final Vol: 10000 uL	Test:	PCB
Prep Method:	5030 Prep Date:		

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
12674-11-2	Aroclor-1016	0.097	U	1	0.097	0.50	ug/L	11/04/25	po110425
11104-28-2	Aroclor-1221	0.13	U	1	0.13	0.50	ug/L	11/04/25	po110425
11141-16-5	Aroclor-1232	0.096	U	1	0.096	0.50	ug/L	11/04/25	po110425
53469-21-9	Aroclor-1242	0.12	U	1	0.12	0.50	ug/L	11/04/25	po110425
12672-29-6	Aroclor-1248	0.071	U	1	0.071	0.50	ug/L	11/04/25	po110425
11097-69-1	Aroclor-1254	0.094	U	1	0.094	0.50	ug/L	11/04/25	po110425
11096-82-5	Aroclor-1260	0.081	U	1	0.081	0.50	ug/L	11/04/25	po110425
37324-23-5	Aroclor-1262	0.14	U	1	0.14	0.50	ug/L	11/04/25	po110425
11100-14-4	Aroclor-1268	0.11	U	1	0.11	0.50	ug/L	11/04/25	po110425
SURROGATES									
877-09-8	Tetrachloro-m-xylene	15.5			70 (60) - 130 (140)	77%	SPK: 20		
2051-24-3	Decachlorobiphenyl	16.8			70 (60) - 130 (140)	84%	SPK: 20		

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
E = Value Exceeds Calibration Range
P = Indicates >25% difference for detected concentrations between the two GC columns
Q = indicates LCS control criteria did not meet requirements
M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
B = Analyte Found in Associated Method Blank
N = Presumptive Evidence of a Compound
* = Values outside of QC limits
D = Dilution
S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.
() = Laboratory InHouse Limit

Sample Results: **PO114870.D**

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
Data File : PO114870.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Nov 2025 09:21
Operator : YP/AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 05 02:16:53 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Nov 05 02:14:41 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.348	3.649	108.8E6	72481407	15.462	16.227
2)	SA Decachlor...	9.930	8.641	74180497	67480996	16.782	17.286
	Target Compounds						


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO110425\
Data File : PO114870.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Nov 2025 09:21
Operator : YP/AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1


Instrument :
ECD_O
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 05 02:16:53 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Nov 05 02:14:41 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm








#1 Tetrachloro-m-xylene


R.T.: 4.348 min
Delta R.T.: 0.000 min
Response: 108823881
Conc: 15.46 ng/ml

Instrument :
ECD_O
ClientSampleId :
I.BLK




#1 Tetrachloro-m-xylene

R.T.: 3.649 min
Delta R.T.: 0.000 min
Response: 72481407
Conc: 16.23 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.930 min
Delta R.T.: 0.000 min
Response: 74180497
Conc: 16.78 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.641 min
Delta R.T.: 0.000 min
Response: 67480996
Conc: 17.29 ng/ml

Raw Data: **PO115035.D**

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Matrix New World Engineering	Date Collected:	11/12/25
Project:	PSEG Athenia Substation	Date Received:	11/12/25
Client Sample ID:	PIBLK-PO115035.D	SDG No.:	Q3450
Lab Sample ID:	I.BLK-PO115035.D	Matrix:	WATER
Analytical Method:	8082A	% Solid:	0
Sample Wt/Vol:	1000 mL Final Vol: 10000 uL	Test:	PCB
Prep Method:	5030 Prep Date:		

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
12674-11-2	Aroclor-1016	0.097	U	1	0.097	0.50	ug/L	11/12/25	PO111225
11104-28-2	Aroclor-1221	0.13	U	1	0.13	0.50	ug/L	11/12/25	PO111225
11141-16-5	Aroclor-1232	0.096	U	1	0.096	0.50	ug/L	11/12/25	PO111225
53469-21-9	Aroclor-1242	0.12	U	1	0.12	0.50	ug/L	11/12/25	PO111225
12672-29-6	Aroclor-1248	0.071	U	1	0.071	0.50	ug/L	11/12/25	PO111225
11097-69-1	Aroclor-1254	0.094	U	1	0.094	0.50	ug/L	11/12/25	PO111225
11096-82-5	Aroclor-1260	0.081	U	1	0.081	0.50	ug/L	11/12/25	PO111225
37324-23-5	Aroclor-1262	0.14	U	1	0.14	0.50	ug/L	11/12/25	PO111225
11100-14-4	Aroclor-1268	0.11	U	1	0.11	0.50	ug/L	11/12/25	PO111225
SURROGATES									
877-09-8	Tetrachloro-m-xylene	19.6			70 (60) - 130 (140)	98%	SPK: 20		
2051-24-3	Decachlorobiphenyl	19.5			70 (60) - 130 (140)	98%	SPK: 20		

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
E = Value Exceeds Calibration Range
P = Indicates >25% difference for detected concentrations between the two GC columns
Q = indicates LCS control criteria did not meet requirements
M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
B = Analyte Found in Associated Method Blank
N = Presumptive Evidence of a Compound
* = Values outside of QC limits
D = Dilution
S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.
() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO111225\
Data File : PO115035.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Nov 2025 11:03
Operator : YP/AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 12 21:41:52 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Nov 05 06:00:07 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.346	3.647	137.9E6	88346115	19.586	19.778
2)	SA Decachlor...	9.933	8.641	86258252	82739043	19.514	21.194
	Target Compounds						


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO111225\
Data File : PO115035.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Nov 2025 11:03
Operator : YP/AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1


Instrument :
ECD_O
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 12 21:41:52 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Nov 05 06:00:07 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm








#1 Tetrachloro-m-xylene


R.T.: 4.346 min
Delta R.T.: -0.002 min
Response: 137851673
Conc: 19.59 ng/ml

Instrument :
ECD_O
ClientSampleId :
I.BLK




#1 Tetrachloro-m-xylene

R.T.: 3.647 min
Delta R.T.: 0.000 min
Response: 88346115
Conc: 19.78 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.933 min
Delta R.T.: 0.003 min
Response: 86258252
Conc: 19.51 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.641 min
Delta R.T.: 0.000 min
Response: 82739043
Conc: 21.19 ng/ml

Raw Data: **PO115050.D**

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Matrix New World Engineering	Date Collected:	11/12/25
Project:	PSEG Athenia Substation	Date Received:	11/12/25
Client Sample ID:	PIBLK-PO115050.D	SDG No.:	Q3450
Lab Sample ID:	I.BLK-PO115050.D	Matrix:	WATER
Analytical Method:	8082A	% Solid:	0
Sample Wt/Vol:	1000 mL Final Vol: 10000 uL	Test:	PCB
Prep Method:	5030 Prep Date:		

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
12674-11-2	Aroclor-1016	0.097	U	1	0.097	0.50	ug/L	11/12/25	PO111225
11104-28-2	Aroclor-1221	0.13	U	1	0.13	0.50	ug/L	11/12/25	PO111225
11141-16-5	Aroclor-1232	0.096	U	1	0.096	0.50	ug/L	11/12/25	PO111225
53469-21-9	Aroclor-1242	0.12	U	1	0.12	0.50	ug/L	11/12/25	PO111225
12672-29-6	Aroclor-1248	0.071	U	1	0.071	0.50	ug/L	11/12/25	PO111225
11097-69-1	Aroclor-1254	0.094	U	1	0.094	0.50	ug/L	11/12/25	PO111225
11096-82-5	Aroclor-1260	0.081	U	1	0.081	0.50	ug/L	11/12/25	PO111225
37324-23-5	Aroclor-1262	0.14	U	1	0.14	0.50	ug/L	11/12/25	PO111225
11100-14-4	Aroclor-1268	0.11	U	1	0.11	0.50	ug/L	11/12/25	PO111225
SURROGATES									
877-09-8	Tetrachloro-m-xylene	19.6			70 (60) - 130 (140)	98%	SPK: 20		
2051-24-3	Decachlorobiphenyl	19.6			70 (60) - 130 (140)	98%	SPK: 20		

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
E = Value Exceeds Calibration Range
P = Indicates >25% difference for detected concentrations between the two GC columns
Q = indicates LCS control criteria did not meet requirements
M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
B = Analyte Found in Associated Method Blank
N = Presumptive Evidence of a Compound
* = Values outside of QC limits
D = Dilution
S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.
() = Laboratory InHouse Limit

Raw Data: **PP075887.D**

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Matrix New World Engineering	Date Collected:	10/21/25
Project:	PSEG Athenia Substation	Date Received:	10/21/25
Client Sample ID:	PIBLK-PP075887.D	SDG No.:	Q3450
Lab Sample ID:	I.BLK-PP075887.D	Matrix:	WATER
Analytical Method:	8082A	% Solid:	0
Sample Wt/Vol:	1000 mL Final Vol: 10000 uL	Test:	PCB
Prep Method:	5030 Prep Date:		

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
12674-11-2	Aroclor-1016	0.097	U	1	0.097	0.50	ug/L	10/21/25	PP102125
11104-28-2	Aroclor-1221	0.13	U	1	0.13	0.50	ug/L	10/21/25	PP102125
11141-16-5	Aroclor-1232	0.096	U	1	0.096	0.50	ug/L	10/21/25	PP102125
53469-21-9	Aroclor-1242	0.12	U	1	0.12	0.50	ug/L	10/21/25	PP102125
12672-29-6	Aroclor-1248	0.071	U	1	0.071	0.50	ug/L	10/21/25	PP102125
11097-69-1	Aroclor-1254	0.094	U	1	0.094	0.50	ug/L	10/21/25	PP102125
11096-82-5	Aroclor-1260	0.081	U	1	0.081	0.50	ug/L	10/21/25	PP102125
37324-23-5	Aroclor-1262	0.14	U	1	0.14	0.50	ug/L	10/21/25	PP102125
11100-14-4	Aroclor-1268	0.11	U	1	0.11	0.50	ug/L	10/21/25	PP102125
SURROGATES									
877-09-8	Tetrachloro-m-xylene	18.8			70 (60) - 130 (140)	94%	SPK: 20		
2051-24-3	Decachlorobiphenyl	19.3			70 (60) - 130 (140)	96%	SPK: 20		

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
E = Value Exceeds Calibration Range
P = Indicates >25% difference for detected concentrations between the two GC columns
Q = indicates LCS control criteria did not meet requirements
M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
B = Analyte Found in Associated Method Blank
N = Presumptive Evidence of a Compound
* = Values outside of QC limits
D = Dilution
S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.
() = Laboratory InHouse Limit

Sample Results: **PP075887.D**

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075887.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2025 08:49
Operator : YP\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
I.BLK

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/22/2025
Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 18:26:13 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 21 18:24:24 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.644	3.783	28503905	124.7E6	19.945	18.843
2)	SA Decachlor...	10.419	8.784	21974278	158.6E6	20.412m	19.281
	Target Compounds						

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Sample Results: **PP075887.D**

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102125\
Data File : PP075887.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2025 08:49
Operator : YP\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1


Instrument :
ECD_P
ClientSampleId :
I.BLK


Manual Integrations APPROVED


Reviewed By :Yogesh Patel 10/22/2025
Supervised By :mohammad ahmed 10/25/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 18:26:13 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 21 18:24:24 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm








#1 Tetrachloro-m-xylene

R.T.: 4.644 min
Delta R.T.: 0.002 min
Response: 28503905
Conc: 19.95 ng/ml

Instrument :
ECD_P
ClientSampleId :
I.BLK


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/22/2025
Supervised By :mohammad ahmed 10/25/2025




#1 Tetrachloro-m-xylene

R.T.: 3.783 min
Delta R.T.: 0.002 min
Response: 124709992
Conc: 18.84 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.419 min
Delta R.T.: 0.000 min
Response: 21974278
Conc: 20.41 ng/ml m



#2 Decachlorobiphenyl

R.T.: 8.784 min
Delta R.T.: 0.002 min
Response: 158550080
Conc: 19.28 ng/ml

Raw Data: **PP076046.D**

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Matrix New World Engineering	Date Collected:	10/24/25
Project:	PSEG Athenia Substation	Date Received:	10/24/25
Client Sample ID:	PIBLK-PP076046.D	SDG No.:	Q3450
Lab Sample ID:	I.BLK-PP076046.D	Matrix:	WATER
Analytical Method:	8082A	% Solid:	0
Sample Wt/Vol:	1000 mL Final Vol: 10000 uL	Test:	PCB
Prep Method:	5030 Prep Date:		

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
12674-11-2	Aroclor-1016	0.097	U	1	0.097	0.50	ug/L	10/24/25	pp102425
11104-28-2	Aroclor-1221	0.13	U	1	0.13	0.50	ug/L	10/24/25	pp102425
11141-16-5	Aroclor-1232	0.096	U	1	0.096	0.50	ug/L	10/24/25	pp102425
53469-21-9	Aroclor-1242	0.12	U	1	0.12	0.50	ug/L	10/24/25	pp102425
12672-29-6	Aroclor-1248	0.071	U	1	0.071	0.50	ug/L	10/24/25	pp102425
11097-69-1	Aroclor-1254	0.094	U	1	0.094	0.50	ug/L	10/24/25	pp102425
11096-82-5	Aroclor-1260	0.081	U	1	0.081	0.50	ug/L	10/24/25	pp102425
37324-23-5	Aroclor-1262	0.14	U	1	0.14	0.50	ug/L	10/24/25	pp102425
11100-14-4	Aroclor-1268	0.11	U	1	0.11	0.50	ug/L	10/24/25	pp102425
SURROGATES									
877-09-8	Tetrachloro-m-xylene	17.6			70 (60) - 130 (140)	88%	SPK: 20		
2051-24-3	Decachlorobiphenyl	20.1			70 (60) - 130 (140)	101%	SPK: 20		

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
E = Value Exceeds Calibration Range
P = Indicates >25% difference for detected concentrations between the two GC columns
Q = indicates LCS control criteria did not meet requirements
M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
B = Analyte Found in Associated Method Blank
N = Presumptive Evidence of a Compound
* = Values outside of QC limits
D = Dilution
S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.
() = Laboratory InHouse Limit

Sample Results: **PP076046.D**

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102425\
Data File : PP076046.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 24 Oct 2025 20:36
Operator : YP\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 24 23:40:03 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 24 15:40:32 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.642	3.782	29442757	116.3E6	20.602	17.572
2)	SA Decachlor...	10.413	8.778	25427395	165.5E6	23.620	20.126
	Target Compounds						


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102425\
Data File : PP076046.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 24 Oct 2025 20:36
Operator : YP\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1


Instrument :
ECD_P
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 24 23:40:03 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 24 15:40:32 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm








#1 Tetrachloro-m-xylene


R.T.: 4.642 min
Delta R.T.: 0.000 min
Response: 29442757
Conc: 20.60 ng/ml

Instrument :
ECD_P
ClientSampleId :
I.BLK




#1 Tetrachloro-m-xylene

R.T.: 3.782 min
Delta R.T.: 0.000 min
Response: 116298499
Conc: 17.57 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.413 min
Delta R.T.: -0.006 min
Response: 25427395
Conc: 23.62 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.778 min
Delta R.T.: -0.004 min
Response: 165496707
Conc: 20.13 ng/ml

Raw Data: **PP076061.D**

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Matrix New World Engineering	Date Collected:	10/25/25
Project:	PSEG Athenia Substation	Date Received:	10/25/25
Client Sample ID:	PIBLK-PP076061.D	SDG No.:	Q3450
Lab Sample ID:	I.BLK-PP076061.D	Matrix:	WATER
Analytical Method:	8082A	% Solid:	0
Sample Wt/Vol:	1000 mL Final Vol: 10000 uL	Test:	PCB
Prep Method:	5030 Prep Date:		

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
12674-11-2	Aroclor-1016	0.097	U	1	0.097	0.50	ug/L	10/25/25	pp102425
11104-28-2	Aroclor-1221	0.13	U	1	0.13	0.50	ug/L	10/25/25	pp102425
11141-16-5	Aroclor-1232	0.096	U	1	0.096	0.50	ug/L	10/25/25	pp102425
53469-21-9	Aroclor-1242	0.12	U	1	0.12	0.50	ug/L	10/25/25	pp102425
12672-29-6	Aroclor-1248	0.071	U	1	0.071	0.50	ug/L	10/25/25	pp102425
11097-69-1	Aroclor-1254	0.094	U	1	0.094	0.50	ug/L	10/25/25	pp102425
11096-82-5	Aroclor-1260	0.081	U	1	0.081	0.50	ug/L	10/25/25	pp102425
37324-23-5	Aroclor-1262	0.14	U	1	0.14	0.50	ug/L	10/25/25	pp102425
11100-14-4	Aroclor-1268	0.11	U	1	0.11	0.50	ug/L	10/25/25	pp102425
SURROGATES									
877-09-8	Tetrachloro-m-xylene	18.2			70 (60) - 130 (140)	91%	SPK: 20		
2051-24-3	Decachlorobiphenyl	21.0			70 (60) - 130 (140)	105%	SPK: 20		

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
E = Value Exceeds Calibration Range
P = Indicates >25% difference for detected concentrations between the two GC columns
Q = indicates LCS control criteria did not meet requirements
M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
B = Analyte Found in Associated Method Blank
N = Presumptive Evidence of a Compound
* = Values outside of QC limits
D = Dilution
S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.
() = Laboratory InHouse Limit

Sample Results: **PP076061.D**

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102425\
Data File : PP076061.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 25 Oct 2025 02:16
Operator : YP\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 25 05:34:19 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 24 15:40:32 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.643	3.783	29833113	120.5E6	20.876	18.205
2)	SA Decachlor...	10.411	8.778	25539926	172.5E6	23.724	20.975
	Target Compounds						


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102425\
Data File : PP076061.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 25 Oct 2025 02:16
Operator : YP\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1


Instrument :
ECD_P
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 25 05:34:19 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 24 15:40:32 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm








#1 Tetrachloro-m-xylene


R.T.: 4.643 min
Delta R.T.: 0.000 min
Response: 29833113
Conc: 20.88 ng/ml

Instrument :
ECD_P
ClientSampleId :
I.BLK




#1 Tetrachloro-m-xylene

R.T.: 3.783 min
Delta R.T.: 0.001 min
Response: 120487723
Conc: 18.21 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.411 min
Delta R.T.: -0.008 min
Response: 25539926
Conc: 23.72 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.778 min
Delta R.T.: -0.004 min
Response: 172478139
Conc: 20.98 ng/ml

Raw Data: **PP076067.D**

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Matrix New World Engineering	Date Collected:	10/27/25
Project:	PSEG Athenia Substation	Date Received:	10/27/25
Client Sample ID:	PIBLK-PP076067.D	SDG No.:	Q3450
Lab Sample ID:	I.BLK-PP076067.D	Matrix:	WATER
Analytical Method:	8082A	% Solid:	0
Sample Wt/Vol:	1000 mL Final Vol: 10000 uL	Test:	PCB
Prep Method:	5030 Prep Date:		

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
12674-11-2	Aroclor-1016	0.097	U	1	0.097	0.50	ug/L	10/27/25	PP102725
11104-28-2	Aroclor-1221	0.13	U	1	0.13	0.50	ug/L	10/27/25	PP102725
11141-16-5	Aroclor-1232	0.096	U	1	0.096	0.50	ug/L	10/27/25	PP102725
53469-21-9	Aroclor-1242	0.12	U	1	0.12	0.50	ug/L	10/27/25	PP102725
12672-29-6	Aroclor-1248	0.071	U	1	0.071	0.50	ug/L	10/27/25	PP102725
11097-69-1	Aroclor-1254	0.094	U	1	0.094	0.50	ug/L	10/27/25	PP102725
11096-82-5	Aroclor-1260	0.081	U	1	0.081	0.50	ug/L	10/27/25	PP102725
37324-23-5	Aroclor-1262	0.14	U	1	0.14	0.50	ug/L	10/27/25	PP102725
11100-14-4	Aroclor-1268	0.11	U	1	0.11	0.50	ug/L	10/27/25	PP102725
SURROGATES									
877-09-8	Tetrachloro-m-xylene	18.9			70 (60) - 130 (140)	95%	SPK: 20		
2051-24-3	Decachlorobiphenyl	18.2			70 (60) - 130 (140)	91%	SPK: 20		

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
E = Value Exceeds Calibration Range
P = Indicates >25% difference for detected concentrations between the two GC columns
Q = indicates LCS control criteria did not meet requirements
M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
B = Analyte Found in Associated Method Blank
N = Presumptive Evidence of a Compound
* = Values outside of QC limits
D = Dilution
S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.
() = Laboratory InHouse Limit

Sample Results: **PP076067.D**

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102725\
Data File : PP076067.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Oct 2025 10:02
Operator : YP\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 27 11:02:24 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 24 15:40:32 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.638	3.778	31506659	125.1E6	22.047	18.907
2)	SA Decachlor...	10.404	8.773	24084542	150.0E6	22.372	18.238
	Target Compounds						

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Sample Results: **PP076067.D**


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102725\
Data File : PP076067.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Oct 2025 10:02
Operator : YP\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1


Instrument :
ECD_P
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 27 11:02:24 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 24 15:40:32 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm








#1 Tetrachloro-m-xylene


R.T.: 4.638 min
Delta R.T.: -0.004 min
Response: 31506659
Conc: 22.05 ng/ml

Instrument :
ECD_P
ClientSampleId :
I.BLK




#1 Tetrachloro-m-xylene

R.T.: 3.778 min
Delta R.T.: -0.003 min
Response: 125130938
Conc: 18.91 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.404 min
Delta R.T.: -0.015 min
Response: 24084542
Conc: 22.37 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.773 min
Delta R.T.: -0.009 min
Response: 149972909
Conc: 18.24 ng/ml

Raw Data: **PP076082.D**

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Matrix New World Engineering	Date Collected:	10/27/25
Project:	PSEG Athenia Substation	Date Received:	10/27/25
Client Sample ID:	PIBLK-PP076082.D	SDG No.:	Q3450
Lab Sample ID:	I.BLK-PP076082.D	Matrix:	WATER
Analytical Method:	8082A	% Solid:	0
Sample Wt/Vol:	1000 mL Final Vol: 10000 uL	Test:	PCB
Prep Method:	5030 Prep Date:		

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
12674-11-2	Aroclor-1016	0.097	U	1	0.097	0.50	ug/L	10/27/25	pp102725
11104-28-2	Aroclor-1221	0.13	U	1	0.13	0.50	ug/L	10/27/25	pp102725
11141-16-5	Aroclor-1232	0.096	U	1	0.096	0.50	ug/L	10/27/25	pp102725
53469-21-9	Aroclor-1242	0.12	U	1	0.12	0.50	ug/L	10/27/25	pp102725
12672-29-6	Aroclor-1248	0.071	U	1	0.071	0.50	ug/L	10/27/25	pp102725
11097-69-1	Aroclor-1254	0.094	U	1	0.094	0.50	ug/L	10/27/25	pp102725
11096-82-5	Aroclor-1260	0.081	U	1	0.081	0.50	ug/L	10/27/25	pp102725
37324-23-5	Aroclor-1262	0.14	U	1	0.14	0.50	ug/L	10/27/25	pp102725
11100-14-4	Aroclor-1268	0.11	U	1	0.11	0.50	ug/L	10/27/25	pp102725
SURROGATES									
877-09-8	Tetrachloro-m-xylene	17.1			70 (60) - 130 (140)	86%	SPK: 20		
2051-24-3	Decachlorobiphenyl	16.7			70 (60) - 130 (140)	83%	SPK: 20		

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
E = Value Exceeds Calibration Range
P = Indicates >25% difference for detected concentrations between the two GC columns
Q = indicates LCS control criteria did not meet requirements
M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
B = Analyte Found in Associated Method Blank
N = Presumptive Evidence of a Compound
* = Values outside of QC limits
D = Dilution
S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.
() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102725\
Data File : PP076082.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Oct 2025 15:13
Operator : YP\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 28 01:03:28 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 24 15:40:32 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.639	3.778	27067383	113.2E6	18.940	17.111
2)	SA Decachlor...	10.406	8.774	21916774	137.1E6	20.359	16.668
	Target Compounds						

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Sample Results: **PP076082.D**


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102725\
Data File : PP076082.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Oct 2025 15:13
Operator : YP\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1


Instrument :
ECD_P
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 28 01:03:28 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 24 15:40:32 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm








#1 Tetrachloro-m-xylene


R.T.: 4.639 min
Delta R.T.: -0.003 min
Response: 27067383
Conc: 18.94 ng/ml

Instrument :
ECD_P
ClientSampleId :
I.BLK




#1 Tetrachloro-m-xylene

R.T.: 3.778 min
Delta R.T.: -0.003 min
Response: 113248293
Conc: 17.11 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.406 min
Delta R.T.: -0.013 min
Response: 21916774
Conc: 20.36 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.774 min
Delta R.T.: -0.008 min
Response: 137057682
Conc: 16.67 ng/ml

Raw Data: **PP076275.D**

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Matrix New World Engineering	Date Collected:	10/31/25
Project:	PSEG Athenia Substation	Date Received:	10/31/25
Client Sample ID:	PIBLK-PP076275.D	SDG No.:	Q3450
Lab Sample ID:	I.BLK-PP076275.D	Matrix:	WATER
Analytical Method:	8082A	% Solid:	0
Sample Wt/Vol:	1000 mL Final Vol: 10000 uL	Test:	PCB
Prep Method:	5030 Prep Date:		

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
12674-11-2	Aroclor-1016	0.097	U	1	0.097	0.50	ug/L	10/31/25	pp103125
11104-28-2	Aroclor-1221	0.13	U	1	0.13	0.50	ug/L	10/31/25	pp103125
11141-16-5	Aroclor-1232	0.096	U	1	0.096	0.50	ug/L	10/31/25	pp103125
53469-21-9	Aroclor-1242	0.12	U	1	0.12	0.50	ug/L	10/31/25	pp103125
12672-29-6	Aroclor-1248	0.071	U	1	0.071	0.50	ug/L	10/31/25	pp103125
11097-69-1	Aroclor-1254	0.094	U	1	0.094	0.50	ug/L	10/31/25	pp103125
11096-82-5	Aroclor-1260	0.081	U	1	0.081	0.50	ug/L	10/31/25	pp103125
37324-23-5	Aroclor-1262	0.14	U	1	0.14	0.50	ug/L	10/31/25	pp103125
11100-14-4	Aroclor-1268	0.11	U	1	0.11	0.50	ug/L	10/31/25	pp103125
SURROGATES									
877-09-8	Tetrachloro-m-xylene	19.3			70 (60) - 130 (140)	97%	SPK: 20		
2051-24-3	Decachlorobiphenyl	17.5			70 (60) - 130 (140)	87%	SPK: 20		

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
E = Value Exceeds Calibration Range
P = Indicates >25% difference for detected concentrations between the two GC columns
Q = indicates LCS control criteria did not meet requirements
M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
B = Analyte Found in Associated Method Blank
N = Presumptive Evidence of a Compound
* = Values outside of QC limits
D = Dilution
S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.
() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP103125\
Data File : PP076275.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 31 Oct 2025 22:34
Operator : YP\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 31 23:28:12 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 24 15:40:32 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.645	3.780	30851530	127.7E6	21.588	19.299
2)	SA Decachlor...	10.423	8.781	22654969	143.6E6	21.045	17.462
	Target Compounds						


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP103125\
Data File : PP076275.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 31 Oct 2025 22:34
Operator : YP\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1


Instrument :
ECD_P
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 31 23:28:12 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 24 15:40:32 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm








#1 Tetrachloro-m-xylene


R.T.: 4.645 min
Delta R.T.: 0.003 min
Response: 30851530
Conc: 21.59 ng/ml

Instrument :
ECD_P
ClientSampleId :
I.BLK




#1 Tetrachloro-m-xylene

R.T.: 3.780 min
Delta R.T.: 0.000 min
Response: 127726179
Conc: 19.30 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.423 min
Delta R.T.: 0.004 min
Response: 22654969
Conc: 21.04 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.781 min
Delta R.T.: -0.001 min
Response: 143592542
Conc: 17.46 ng/ml

Raw Data: **PP076290.D**

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Matrix New World Engineering	Date Collected:	11/01/25
Project:	PSEG Athenia Substation	Date Received:	11/01/25
Client Sample ID:	PIBLK-PP076290.D	SDG No.:	Q3450
Lab Sample ID:	I.BLK-PP076290.D	Matrix:	WATER
Analytical Method:	8082A	% Solid:	0
Sample Wt/Vol:	1000 mL Final Vol: 10000 uL	Test:	PCB
Prep Method:	5030 Prep Date:		

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
12674-11-2	Aroclor-1016	0.097	U	1	0.097	0.50	ug/L	11/01/25	pp103125
11104-28-2	Aroclor-1221	0.13	U	1	0.13	0.50	ug/L	11/01/25	pp103125
11141-16-5	Aroclor-1232	0.096	U	1	0.096	0.50	ug/L	11/01/25	pp103125
53469-21-9	Aroclor-1242	0.12	U	1	0.12	0.50	ug/L	11/01/25	pp103125
12672-29-6	Aroclor-1248	0.071	U	1	0.071	0.50	ug/L	11/01/25	pp103125
11097-69-1	Aroclor-1254	0.094	U	1	0.094	0.50	ug/L	11/01/25	pp103125
11096-82-5	Aroclor-1260	0.081	U	1	0.081	0.50	ug/L	11/01/25	pp103125
37324-23-5	Aroclor-1262	0.14	U	1	0.14	0.50	ug/L	11/01/25	pp103125
11100-14-4	Aroclor-1268	0.11	U	1	0.11	0.50	ug/L	11/01/25	pp103125
SURROGATES									
877-09-8	Tetrachloro-m-xylene	19.2			70 (60) - 130 (140)	96%	SPK: 20		
2051-24-3	Decachlorobiphenyl	17.5			70 (60) - 130 (140)	88%	SPK: 20		

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
E = Value Exceeds Calibration Range
P = Indicates >25% difference for detected concentrations between the two GC columns
Q = indicates LCS control criteria did not meet requirements
M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
B = Analyte Found in Associated Method Blank
N = Presumptive Evidence of a Compound
* = Values outside of QC limits
D = Dilution
S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.
() = Laboratory InHouse Limit

Sample Results: **PP076290.D**

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP103125\
Data File : PP076290.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 01 Nov 2025 04:00
Operator : YP\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 01 04:47:25 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 24 15:40:32 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.641	3.778	31437079	126.9E6	21.998	19.173
2)	SA Decachlor...	10.416	8.775	22759458	144.0E6	21.142	17.506
	Target Compounds						

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Sample Results: **PP076290.D**


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP103125\
Data File : PP076290.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 01 Nov 2025 04:00
Operator : YP\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1


Instrument :
ECD_P
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 01 04:47:25 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 24 15:40:32 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm








#1 Tetrachloro-m-xylene


R.T.: 4.641 min
Delta R.T.: 0.000 min
Response: 31437079
Conc: 22.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
I.BLK




#1 Tetrachloro-m-xylene

R.T.: 3.778 min
Delta R.T.: -0.003 min
Response: 126890520
Conc: 19.17 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.416 min
Delta R.T.: -0.002 min
Response: 22759458
Conc: 21.14 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.775 min
Delta R.T.: -0.007 min
Response: 143955049
Conc: 17.51 ng/ml

Raw Data: **PP076296.D**

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Matrix New World Engineering	Date Collected:	11/03/25
Project:	PSEG Athenia Substation	Date Received:	11/03/25
Client Sample ID:	PIBLK-PP076296.D	SDG No.:	Q3450
Lab Sample ID:	I.BLK-PP076296.D	Matrix:	WATER
Analytical Method:	8082A	% Solid:	0
Sample Wt/Vol:	1000 mL Final Vol: 10000 uL	Test:	PCB
Prep Method:	5030 Prep Date:		

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
12674-11-2	Aroclor-1016	0.097	U	1	0.097	0.50	ug/L	11/03/25	Pp110325
11104-28-2	Aroclor-1221	0.13	U	1	0.13	0.50	ug/L	11/03/25	Pp110325
11141-16-5	Aroclor-1232	0.096	U	1	0.096	0.50	ug/L	11/03/25	Pp110325
53469-21-9	Aroclor-1242	0.12	U	1	0.12	0.50	ug/L	11/03/25	Pp110325
12672-29-6	Aroclor-1248	0.071	U	1	0.071	0.50	ug/L	11/03/25	Pp110325
11097-69-1	Aroclor-1254	0.094	U	1	0.094	0.50	ug/L	11/03/25	Pp110325
11096-82-5	Aroclor-1260	0.081	U	1	0.081	0.50	ug/L	11/03/25	Pp110325
37324-23-5	Aroclor-1262	0.14	U	1	0.14	0.50	ug/L	11/03/25	Pp110325
11100-14-4	Aroclor-1268	0.11	U	1	0.11	0.50	ug/L	11/03/25	Pp110325
SURROGATES									
877-09-8	Tetrachloro-m-xylene	17.7			70 (60) - 130 (140)	88%	SPK: 20		
2051-24-3	Decachlorobiphenyl	18.8			70 (60) - 130 (140)	94%	SPK: 20		

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
E = Value Exceeds Calibration Range
P = Indicates >25% difference for detected concentrations between the two GC columns
Q = indicates LCS control criteria did not meet requirements
M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
B = Analyte Found in Associated Method Blank
N = Presumptive Evidence of a Compound
* = Values outside of QC limits
D = Dilution
S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.
() = Laboratory InHouse Limit

Quantitation Report (QT Reviewed)

Sample Results: **PP076296.D**

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP110325\
Data File : PP076296.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 03 Nov 2025 09:53
Operator : YP\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 03 11:22:55 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 24 15:40:32 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.640	3.780	28795383	117.0E6	20.149	17.683
2)	SA Decachlor...	10.408	8.773	22690741	154.2E6	21.078	18.756
	Target Compounds						


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP110325\
Data File : PP076296.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 03 Nov 2025 09:53
Operator : YP\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1


Instrument :
ECD_P
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 03 11:22:55 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 24 15:40:32 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm








#1 Tetrachloro-m-xylene


R.T.: 4.640 min
Delta R.T.: -0.002 min
Response: 28795383
Conc: 20.15 ng/ml

Instrument :
ECD_P
ClientSampleId :
I.BLK




#1 Tetrachloro-m-xylene

R.T.: 3.780 min
Delta R.T.: -0.002 min
Response: 117030847
Conc: 17.68 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.408 min
Delta R.T.: -0.010 min
Response: 22690741
Conc: 21.08 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.773 min
Delta R.T.: -0.009 min
Response: 154231444
Conc: 18.76 ng/ml

Raw Data: **PP076308.D**

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Matrix New World Engineering	Date Collected:	11/03/25
Project:	PSEG Athenia Substation	Date Received:	11/03/25
Client Sample ID:	PIBLK-PP076308.D	SDG No.:	Q3450
Lab Sample ID:	I.BLK-PP076308.D	Matrix:	WATER
Analytical Method:	8082A	% Solid:	0
Sample Wt/Vol:	1000 mL Final Vol: 10000 uL	Test:	PCB
Prep Method:	5030 Prep Date:		

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
12674-11-2	Aroclor-1016	0.097	U	1	0.097	0.50	ug/L	11/03/25	pP110325
11104-28-2	Aroclor-1221	0.13	U	1	0.13	0.50	ug/L	11/03/25	pP110325
11141-16-5	Aroclor-1232	0.096	U	1	0.096	0.50	ug/L	11/03/25	pP110325
53469-21-9	Aroclor-1242	0.12	U	1	0.12	0.50	ug/L	11/03/25	pP110325
12672-29-6	Aroclor-1248	0.071	U	1	0.071	0.50	ug/L	11/03/25	pP110325
11097-69-1	Aroclor-1254	0.094	U	1	0.094	0.50	ug/L	11/03/25	pP110325
11096-82-5	Aroclor-1260	0.081	U	1	0.081	0.50	ug/L	11/03/25	pP110325
37324-23-5	Aroclor-1262	0.14	U	1	0.14	0.50	ug/L	11/03/25	pP110325
11100-14-4	Aroclor-1268	0.11	U	1	0.11	0.50	ug/L	11/03/25	pP110325
SURROGATES									
877-09-8	Tetrachloro-m-xylene	19.0			70 (60) - 130 (140)	95%	SPK: 20		
2051-24-3	Decachlorobiphenyl	15.1			70 (60) - 130 (140)	76%	SPK: 20		

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
E = Value Exceeds Calibration Range
P = Indicates >25% difference for detected concentrations between the two GC columns
Q = indicates LCS control criteria did not meet requirements
M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
B = Analyte Found in Associated Method Blank
N = Presumptive Evidence of a Compound
* = Values outside of QC limits
D = Dilution
S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.
() = Laboratory InHouse Limit

Quantitation Report (QT Reviewed)

Sample Results: **PP076308.D**

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP110325\
Data File : PP076308.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 03 Nov 2025 15:24
Operator : YP\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 04 01:12:04 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 24 15:40:32 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.641	3.780	29202180	125.5E6	20.434	18.963
2)	SA Decachlor...	10.413	8.775	21375736	124.5E6	19.856	15.143
	Target Compounds						


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP110325\
Data File : PP076308.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 03 Nov 2025 15:24
Operator : YP\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1


Instrument :
ECD_P
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 04 01:12:04 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 24 15:40:32 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm








#1 Tetrachloro-m-xylene


R.T.: 4.641 min
Delta R.T.: -0.001 min
Response: 29202180
Conc: 20.43 ng/ml

Instrument :
ECD_P
ClientSampleId :
I.BLK




#1 Tetrachloro-m-xylene

R.T.: 3.780 min
Delta R.T.: -0.001 min
Response: 125504738
Conc: 18.96 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.413 min
Delta R.T.: -0.006 min
Response: 21375736
Conc: 19.86 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.775 min
Delta R.T.: -0.007 min
Response: 124520195
Conc: 15.14 ng/ml

Raw Data: **PQ070972.D**

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Matrix New World Engineering	Date Collected:	10/10/25
Project:	PSEG Athenia Substation	Date Received:	10/10/25
Client Sample ID:	PIBLK-PQ070972.D	SDG No.:	Q3450
Lab Sample ID:	I.BLK-PQ070972.D	Matrix:	WATER
Analytical Method:	8082A	% Solid:	0
Sample Wt/Vol:	1000 mL Final Vol: 10000 uL	Test:	PCB
Prep Method:	5030 Prep Date:		

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
12674-11-2	Aroclor-1016	0.097	U	1	0.097	0.50	ug/L	10/10/25	pq101025
11104-28-2	Aroclor-1221	0.13	U	1	0.13	0.50	ug/L	10/10/25	pq101025
11141-16-5	Aroclor-1232	0.096	U	1	0.096	0.50	ug/L	10/10/25	pq101025
53469-21-9	Aroclor-1242	0.12	U	1	0.12	0.50	ug/L	10/10/25	pq101025
12672-29-6	Aroclor-1248	0.071	U	1	0.071	0.50	ug/L	10/10/25	pq101025
11097-69-1	Aroclor-1254	0.094	U	1	0.094	0.50	ug/L	10/10/25	pq101025
11096-82-5	Aroclor-1260	0.081	U	1	0.081	0.50	ug/L	10/10/25	pq101025
37324-23-5	Aroclor-1262	0.14	U	1	0.14	0.50	ug/L	10/10/25	pq101025
11100-14-4	Aroclor-1268	0.11	U	1	0.11	0.50	ug/L	10/10/25	pq101025
SURROGATES									
877-09-8	Tetrachloro-m-xylene	18.3			70 (60) - 130 (140)	91%	SPK: 20		
2051-24-3	Decachlorobiphenyl	19.0			70 (60) - 130 (140)	95%	SPK: 20		

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
E = Value Exceeds Calibration Range
P = Indicates >25% difference for detected concentrations between the two GC columns
Q = indicates LCS control criteria did not meet requirements
M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
B = Analyte Found in Associated Method Blank
N = Presumptive Evidence of a Compound
* = Values outside of QC limits
D = Dilution
S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.
() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ070972.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 08:42
Operator : YP\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 17:32:33 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 10 17:24:38 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	3.400	2.778	79509611	97022264	18.277	19.099
2)	SA Decachlor...	8.496	7.533	55246960	97962147	19.025	19.400
	Target Compounds						


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ101025\
Data File : PQ070972.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2025 08:42
Operator : YP\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1


Instrument :
ECD_Q
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 10 17:32:33 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 10 17:24:38 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm








#1 Tetrachloro-m-xylene


R.T.: 3.400 min
Delta R.T.: 0.000 min
Response: 79509611
Conc: 18.28 ng/ml

Instrument :
ECD_Q
ClientSampleId :
I.BLK




#1 Tetrachloro-m-xylene

R.T.: 2.778 min
Delta R.T.: 0.000 min
Response: 97022264
Conc: 19.10 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.496 min
Delta R.T.: 0.000 min
Response: 55246960
Conc: 19.02 ng/ml



#2 Decachlorobiphenyl

R.T.: 7.533 min
Delta R.T.: 0.000 min
Response: 97962147
Conc: 19.40 ng/ml

Raw Data: **PQ071150.D**

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Matrix New World Engineering	Date Collected:	10/27/25
Project:	PSEG Athenia Substation	Date Received:	10/27/25
Client Sample ID:	PIBLK-PQ071150.D	SDG No.:	Q3450
Lab Sample ID:	I.BLK-PQ071150.D	Matrix:	WATER
Analytical Method:	8082A	% Solid:	0
Sample Wt/Vol:	1000 mL Final Vol: 10000 uL	Test:	PCB
Prep Method:	5030 Prep Date:		

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
12674-11-2	Aroclor-1016	0.097	U	1	0.097	0.50	ug/L	10/27/25	PQ102725
11104-28-2	Aroclor-1221	0.13	U	1	0.13	0.50	ug/L	10/27/25	PQ102725
11141-16-5	Aroclor-1232	0.096	U	1	0.096	0.50	ug/L	10/27/25	PQ102725
53469-21-9	Aroclor-1242	0.12	U	1	0.12	0.50	ug/L	10/27/25	PQ102725
12672-29-6	Aroclor-1248	0.071	U	1	0.071	0.50	ug/L	10/27/25	PQ102725
11097-69-1	Aroclor-1254	0.094	U	1	0.094	0.50	ug/L	10/27/25	PQ102725
11096-82-5	Aroclor-1260	0.081	U	1	0.081	0.50	ug/L	10/27/25	PQ102725
37324-23-5	Aroclor-1262	0.14	U	1	0.14	0.50	ug/L	10/27/25	PQ102725
11100-14-4	Aroclor-1268	0.11	U	1	0.11	0.50	ug/L	10/27/25	PQ102725
SURROGATES									
877-09-8	Tetrachloro-m-xylene	19.9			70 (60) - 130 (140)	100%	SPK: 20		
2051-24-3	Decachlorobiphenyl	24.5			70 (60) - 130 (140)	122%	SPK: 20		

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
E = Value Exceeds Calibration Range
P = Indicates >25% difference for detected concentrations between the two GC columns
Q = indicates LCS control criteria did not meet requirements
M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
B = Analyte Found in Associated Method Blank
N = Presumptive Evidence of a Compound
* = Values outside of QC limits
D = Dilution
S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.
() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ102725\
Data File : PQ071150.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Oct 2025 09:11
Operator : YP\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 27 11:28:19 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Oct 11 01:55:41 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	3.401	2.776	97916618	101.1E6	22.508	19.909
2)	SA Decachlor...	8.497	7.532	74642604	123.6E6	25.704	24.472
	Target Compounds						

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Sample Results: **PQ071150.D**


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ102725\
Data File : PQ071150.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Oct 2025 09:11
Operator : YP\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1


Instrument :
ECD_Q
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 27 11:28:19 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Oct 11 01:55:41 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm








#1 Tetrachloro-m-xylene


R.T.: 3.401 min
Delta R.T.: 0.000 min
Response: 97916618
Conc: 22.51 ng/ml

Instrument :
ECD_Q
ClientSampleId :
I.BLK




#1 Tetrachloro-m-xylene

R.T.: 2.776 min
Delta R.T.: -0.002 min
Response: 101136067
Conc: 19.91 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.497 min
Delta R.T.: -0.003 min
Response: 74642604
Conc: 25.70 ng/ml



#2 Decachlorobiphenyl

R.T.: 7.532 min
Delta R.T.: -0.001 min
Response: 123572351
Conc: 24.47 ng/ml

Raw Data: **PQ071165.D**

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Matrix New World Engineering	Date Collected:	10/27/25
Project:	PSEG Athenia Substation	Date Received:	10/27/25
Client Sample ID:	PIBLK-PQ071165.D	SDG No.:	Q3450
Lab Sample ID:	I.BLK-PQ071165.D	Matrix:	WATER
Analytical Method:	8082A	% Solid:	0
Sample Wt/Vol:	1000 mL Final Vol: 10000 uL	Test:	PCB
Prep Method:	5030 Prep Date:		

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
12674-11-2	Aroclor-1016	0.097	U	1	0.097	0.50	ug/L	10/27/25	PQ102725
11104-28-2	Aroclor-1221	0.13	U	1	0.13	0.50	ug/L	10/27/25	PQ102725
11141-16-5	Aroclor-1232	0.096	U	1	0.096	0.50	ug/L	10/27/25	PQ102725
53469-21-9	Aroclor-1242	0.12	U	1	0.12	0.50	ug/L	10/27/25	PQ102725
12672-29-6	Aroclor-1248	0.071	U	1	0.071	0.50	ug/L	10/27/25	PQ102725
11097-69-1	Aroclor-1254	0.094	U	1	0.094	0.50	ug/L	10/27/25	PQ102725
11096-82-5	Aroclor-1260	0.081	U	1	0.081	0.50	ug/L	10/27/25	PQ102725
37324-23-5	Aroclor-1262	0.14	U	1	0.14	0.50	ug/L	10/27/25	PQ102725
11100-14-4	Aroclor-1268	0.11	U	1	0.11	0.50	ug/L	10/27/25	PQ102725
SURROGATES									
877-09-8	Tetrachloro-m-xylene	18.4			70 (60) - 130 (140)	92%	SPK: 20		
2051-24-3	Decachlorobiphenyl	22.2			70 (60) - 130 (140)	111%	SPK: 20		

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
E = Value Exceeds Calibration Range
P = Indicates >25% difference for detected concentrations between the two GC columns
Q = indicates LCS control criteria did not meet requirements
M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
B = Analyte Found in Associated Method Blank
N = Presumptive Evidence of a Compound
* = Values outside of QC limits
D = Dilution
S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.
() = Laboratory InHouse Limit

Quantitation Report (QT Reviewed)

Sample Results: **PQ071165.D**

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ102725\
Data File : PQ071165.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Oct 2025 14:18
Operator : YP\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 28 00:23:56 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Oct 11 01:55:41 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	3.399	2.776	88339084	93561979	20.307	18.418
2)	SA Decachlor...	8.497	7.530	68135285	112.1E6	23.463	22.205
	Target Compounds						


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ102725\
Data File : PQ071165.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Oct 2025 14:18
Operator : YP\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1


Instrument :
ECD_Q
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 28 00:23:56 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Oct 11 01:55:41 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm








#1 Tetrachloro-m-xylene


R.T.: 3.399 min
Delta R.T.: -0.001 min
Response: 88339084
Conc: 20.31 ng/ml

Instrument :
ECD_Q
ClientSampleId :
I.BLK




#1 Tetrachloro-m-xylene

R.T.: 2.776 min
Delta R.T.: -0.002 min
Response: 93561979
Conc: 18.42 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.497 min
Delta R.T.: -0.003 min
Response: 68135285
Conc: 23.46 ng/ml



#2 Decachlorobiphenyl

R.T.: 7.530 min
Delta R.T.: -0.003 min
Response: 112123551
Conc: 22.20 ng/ml

Raw Data: **PP076069.D**

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Matrix New World Engineering	Date Collected:	
Project:	PSEG Athenia Substation	Date Received:	
Client Sample ID:	PB170236BS	SDG No.:	Q3450
Lab Sample ID:	PB170236BS	Matrix:	SOIL
Analytical Method:	8082A	% Solid:	100
Sample Wt/Vol:	30.01 g	Final Vol: 10000 uL	Test: PCB
Prep Method:	SW3541B	Prep Date: 10/24/25	

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
12674-11-2	Aroclor-1016	145		1	3.90	17.0	ug/kg	10/27/25 10:34	PB170236
11104-28-2	Aroclor-1221	4.00	U	1	4.00	17.0	ug/kg	10/27/25 10:34	PB170236
11141-16-5	Aroclor-1232	3.70	U	1	3.70	17.0	ug/kg	10/27/25 10:34	PB170236
53469-21-9	Aroclor-1242	4.00	U	1	4.00	17.0	ug/kg	10/27/25 10:34	PB170236
12672-29-6	Aroclor-1248	5.90	U	1	5.90	17.0	ug/kg	10/27/25 10:34	PB170236
11097-69-1	Aroclor-1254	3.20	U	1	3.20	17.0	ug/kg	10/27/25 10:34	PB170236
37324-23-5	Aroclor-1262	5.00	U	1	5.00	17.0	ug/kg	10/27/25 10:34	PB170236
11100-14-4	Aroclor-1268	3.60	U	1	3.60	17.0	ug/kg	10/27/25 10:34	PB170236
11096-82-5	Aroclor-1260	149		1	3.20	17.0	ug/kg	10/27/25 10:34	PB170236
SURROGATES									
877-09-8	Tetrachloro-m-xylene	26.7			30 (21) - 150 (165)	134%	SPK: 20		
2051-24-3	Decachlorobiphenyl	28.6			30 (10) - 150 (170)	143%	SPK: 20		

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
E = Value Exceeds Calibration Range
P = Indicates >25% difference for detected concentrations between the two GC columns
Q = indicates LCS control criteria did not meet requirements
M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
B = Analyte Found in Associated Method Blank
N = Presumptive Evidence of a Compound
* = Values outside of QC limits
D = Dilution
S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.
() = Laboratory InHouse Limit

Sample Results: **PP076069.D**

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102725\
Data File : PP076069.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Oct 2025 10:34
Operator : YP\AJ
Sample : PB170236BS
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
PB170236BS

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/28/2025
Supervised By :mohammad ahmed 10/29/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 27 11:18:04 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 24 15:40:32 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.640	3.780	38149794	141.0E6	26.695m	21.306m
2)	SA Decachlor...	10.408	8.775	30747417	199.5E6	28.562	24.264
	Target Compounds						
3)	L1 AR-1016-1	5.793	4.877	25825913	292.1E6	436.794	394.687
4)	L1 AR-1016-2	5.814	4.935	39096018	134.3E6	442.042	392.888
5)	L1 AR-1016-3	5.877	5.054	24165535	75316147	436.587	366.783
6)	L1 AR-1016-4	5.974	5.095	19941002	72452629	438.840	387.684
7)	L1 AR-1016-5	6.266	5.309	18220008	81872028	419.762	359.115
31)	L7 AR-1260-1	7.383	6.335	34475825	242.2E6	501.418	442.275
32)	L7 AR-1260-2	7.636	6.522	41196225	358.0E6	459.896	440.000
33)	L7 AR-1260-3	7.994	6.675	27871115	256.4E6	402.259	455.573
34)	L7 AR-1260-4	8.220	7.144	32220662	221.2E6	469.885m	425.862
35)	L7 AR-1260-5	8.546	7.384	60142007	524.0E6	406.498	386.522

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102725\
Data File : PP076069.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Oct 2025 10:34
Operator : YP\AJ
Sample : PB170236BS
Misc :
ALS Vial : 8 Sample Multiplier: 1


Instrument :
ECD_P
ClientSampleId :
PB170236BS


Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/28/2025
Supervised By :mohammad ahmed 10/29/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 27 11:18:04 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 24 15:40:32 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





Raw Data: **PP076277.D**

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Matrix New World Engineering	Date Collected:	
Project:	PSEG Athenia Substation	Date Received:	
Client Sample ID:	PB170358BS	SDG No.:	Q3450
Lab Sample ID:	PB170358BS	Matrix:	SOIL
Analytical Method:	8082A	% Solid:	100
Sample Wt/Vol:	30.03 g	Final Vol: 10000 uL	Test: PCB
Prep Method:	SW3541B	Prep Date: 10/31/25	

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
12674-11-2	Aroclor-1016	170		1	3.90	17.0	ug/kg	10/31/25 23:07	PB170358
11104-28-2	Aroclor-1221	4.00	U	1	4.00	17.0	ug/kg	10/31/25 23:07	PB170358
11141-16-5	Aroclor-1232	3.70	U	1	3.70	17.0	ug/kg	10/31/25 23:07	PB170358
53469-21-9	Aroclor-1242	4.00	U	1	4.00	17.0	ug/kg	10/31/25 23:07	PB170358
12672-29-6	Aroclor-1248	5.90	U	1	5.90	17.0	ug/kg	10/31/25 23:07	PB170358
11097-69-1	Aroclor-1254	3.20	U	1	3.20	17.0	ug/kg	10/31/25 23:07	PB170358
37324-23-5	Aroclor-1262	5.00	U	1	5.00	17.0	ug/kg	10/31/25 23:07	PB170358
11100-14-4	Aroclor-1268	3.60	U	1	3.60	17.0	ug/kg	10/31/25 23:07	PB170358
11096-82-5	Aroclor-1260	168		1	3.20	17.0	ug/kg	10/31/25 23:07	PB170358
SURROGATES									
877-09-8	Tetrachloro-m-xylene	27.9			30 (21) - 150 (165)	140%	SPK: 20		
2051-24-3	Decachlorobiphenyl	25.9			30 (10) - 150 (170)	130%	SPK: 20		

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
E = Value Exceeds Calibration Range
P = Indicates >25% difference for detected concentrations between the two GC columns
Q = indicates LCS control criteria did not meet requirements
M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
B = Analyte Found in Associated Method Blank
N = Presumptive Evidence of a Compound
* = Values outside of QC limits
D = Dilution
S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.
() = Laboratory InHouse Limit

Sample Results: **PP076277.D**

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP103125\
Data File : PP076277.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 31 Oct 2025 23:07
Operator : YP\AJ
Sample : PB170358BS
Misc :
ALS Vial : 27 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
PB170358BS

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 11/03/2025
Supervised By :Yogesh Patel 11/03/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 01 02:47:06 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 24 15:40:32 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.645	3.780	39882244	163.4E6	27.907	24.687
2)	SA Decachlor...	10.424	8.780	27906416	179.6E6	25.923	21.838
	Target Compounds						
3)	L1 AR-1016-1	5.798	4.875	30590838	346.4E6	517.383	467.986
4)	L1 AR-1016-2	5.819	4.933	45242453	154.9E6	511.537	453.018
5)	L1 AR-1016-3	5.882	5.052	28659013	86475288	517.768	421.127
6)	L1 AR-1016-4	5.979	5.096	23687449	94999053	521.288	508.327
7)	L1 AR-1016-5	6.272	5.308	21357053	95671893	492.035	419.645
31)	L7 AR-1260-1	7.390	6.336	38013381	276.1E6	552.869	504.087
32)	L7 AR-1260-2	7.643	6.523	46511458	407.0E6	519.233	500.120
33)	L7 AR-1260-3	8.001	6.676	31545963	293.4E6	455.297	521.330
34)	L7 AR-1260-4	8.228	7.145	35697607	279.2E6	520.590m	537.528
35)	L7 AR-1260-5	8.555	7.385	66465542	631.8E6	449.238	466.101

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP103125\
Data File : PP076277.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 31 Oct 2025 23:07
Operator : YP\AJ
Sample : PB170358BS
Misc :
ALS Vial : 27 Sample Multiplier: 1


Instrument :
ECD_P
ClientSampleId :
PB170358BS


Manual Integrations APPROVED

Reviewed By :Rahul Chavli 11/03/2025
Supervised By :Yogesh Patel 11/03/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 01 02:47:06 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 24 15:40:32 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





Raw Data: **PO115038.D**

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Matrix New World Engineering	Date Collected:	
Project:	PSEG Athenia Substation	Date Received:	
Client Sample ID:	PB170497BS	SDG No.:	Q3450
Lab Sample ID:	PB170497BS	Matrix:	SOIL
Analytical Method:	8082A	% Solid:	100
Sample Wt/Vol:	30.03 g	Final Vol: 10000 uL	Test: PCB
Prep Method:	SW3541B	Prep Date: 11/12/25	

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
12674-11-2	Aroclor-1016	145		1	3.90	17.0	ug/kg	11/12/25 12:16	PB170497
11104-28-2	Aroclor-1221	4.00	U	1	4.00	17.0	ug/kg	11/12/25 12:16	PB170497
11141-16-5	Aroclor-1232	3.70	U	1	3.70	17.0	ug/kg	11/12/25 12:16	PB170497
53469-21-9	Aroclor-1242	4.00	U	1	4.00	17.0	ug/kg	11/12/25 12:16	PB170497
12672-29-6	Aroclor-1248	5.90	U	1	5.90	17.0	ug/kg	11/12/25 12:16	PB170497
11097-69-1	Aroclor-1254	3.20	U	1	3.20	17.0	ug/kg	11/12/25 12:16	PB170497
37324-23-5	Aroclor-1262	5.00	U	1	5.00	17.0	ug/kg	11/12/25 12:16	PB170497
11100-14-4	Aroclor-1268	3.60	U	1	3.60	17.0	ug/kg	11/12/25 12:16	PB170497
11096-82-5	Aroclor-1260	137		1	3.20	17.0	ug/kg	11/12/25 12:16	PB170497
SURROGATES									
877-09-8	Tetrachloro-m-xylene	21.8			30 (21) - 150 (165)	109%	SPK: 20		
2051-24-3	Decachlorobiphenyl	22.6			30 (10) - 150 (170)	113%	SPK: 20		

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
E = Value Exceeds Calibration Range
P = Indicates >25% difference for detected concentrations between the two GC columns
Q = indicates LCS control criteria did not meet requirements
M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
B = Analyte Found in Associated Method Blank
N = Presumptive Evidence of a Compound
* = Values outside of QC limits
D = Dilution
S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.
() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO111225\
Data File : PO115038.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Nov 2025 12:16
Operator : YP/AJ
Sample : PB170497BS
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
PB170497BS

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 12 21:44:21 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Nov 05 06:00:07 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.347	3.647	153.4E6	90414118	21.790	20.241
2)	SA Decachlor...	9.932	8.641	94195741	88344683	21.310	22.630
	Target Compounds						
3)	L1 AR-1016-1	5.486	4.726	108.9E6	82503476	442.611	433.646
4)	L1 AR-1016-2	5.508	4.745	173.0E6	115.3E6	435.574	425.843
5)	L1 AR-1016-3	5.570	4.920	99809915	63712360	447.576	430.567
6)	L1 AR-1016-4	5.666	4.963	81815202	49421919	433.171	420.907
7)	L1 AR-1016-5	5.957	5.175	73719815	63506278	413.698	415.852
31)	L7 AR-1260-1	7.072	6.205	131.4E6	124.7E6	429.845	421.020
32)	L7 AR-1260-2	7.325	6.393	182.5E6	178.8E6	461.047	415.759
33)	L7 AR-1260-3	7.683	6.545	126.2E6	148.0E6	380.261	421.539
34)	L7 AR-1260-4	7.904	7.016	121.9E6	112.3E6	399.487	370.666
35)	L7 AR-1260-5	8.211	7.258	273.3E6	306.0E6	383.257	372.392


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO111225\
Data File : PO115038.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Nov 2025 12:16
Operator : YP/AJ
Sample : PB170497BS
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
PB170497BS

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 12 21:44:21 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Nov 05 06:00:07 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





Raw Data: **PQ071153.D**

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Matrix New World Engineering	Date Collected:	10/23/25
Project:	PSEG Athenia Substation	Date Received:	10/23/25
Client Sample ID:	SS-5R-(3.5-4.0)MS	SDG No.:	Q3450
Lab Sample ID:	Q3450-02MS	Matrix:	SOIL
Analytical Method:	8082A	% Solid:	86.3
Sample Wt/Vol:	30.06 g	Final Vol: 10000 uL	Test: PCB
Prep Method:	SW3541B	Prep Date: 10/24/25	

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
12674-11-2	Aroclor-1016	121		1	4.60	19.7	ug/kg	10/27/25 09:55	PB170236
11104-28-2	Aroclor-1221	4.70	U	1	4.70	19.7	ug/kg	10/27/25 09:55	PB170236
11141-16-5	Aroclor-1232	4.30	U	1	4.30	19.7	ug/kg	10/27/25 09:55	PB170236
53469-21-9	Aroclor-1242	4.60	U	1	4.60	19.7	ug/kg	10/27/25 09:55	PB170236
12672-29-6	Aroclor-1248	6.80	U	1	6.80	19.7	ug/kg	10/27/25 09:55	PB170236
11097-69-1	Aroclor-1254	105		1	3.70	19.7	ug/kg	10/27/25 09:55	PB170236
37324-23-5	Aroclor-1262	5.80	U	1	5.80	19.7	ug/kg	10/27/25 09:55	PB170236
11100-14-4	Aroclor-1268	4.20	U	1	4.20	19.7	ug/kg	10/27/25 09:55	PB170236
11096-82-5	Aroclor-1260	145		1	3.70	19.7	ug/kg	10/27/25 09:55	PB170236
SURROGATES									
877-09-8	Tetrachloro-m-xylene	23.4			30 (21) - 150 (165)	117%	SPK: 20		
2051-24-3	Decachlorobiphenyl	22.4			30 (10) - 150 (170)	112%	SPK: 20		

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
E = Value Exceeds Calibration Range
P = Indicates >25% difference for detected concentrations between the two GC columns
Q = indicates LCS control criteria did not meet requirements
M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
B = Analyte Found in Associated Method Blank
N = Presumptive Evidence of a Compound
* = Values outside of QC limits
D = Dilution
S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.
() = Laboratory InHouse Limit

Sample Results: **PQ071153.D**

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ102725\
Data File : PQ071153.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Oct 2025 09:55
Operator : YP\AJ
Sample : Q3450-02MS
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
SS-5R-(3.5-4.0)MS

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/27/2025
Supervised By :mohammad ahmed 10/29/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 27 11:30:04 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Oct 11 01:55:41 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	3.399	2.776	101.9E6	95077060	23.434	18.716m
2)	SA Decachlor...	8.495	7.531	64988561	104.5E6	22.379	20.702
	Target Compounds						
3)	L1 AR-1016-1	4.481	3.775	25519004	75904857	155.467m	338.858m#
4)	L1 AR-1016-2	4.498	3.791	81532605	96925792	318.121m	284.973m
5)	L1 AR-1016-3	4.555	3.951	57737086	56328796	362.856m	309.739m
6)	L1 AR-1016-4	4.645	3.999	49897243	46929761	384.034m	298.132m
7)	L1 AR-1016-5	4.930	4.195	44988750	56961892	349.284	296.921
26)	L6 AR-1254-1	5.291	4.531	37478695	50045658	190.947m	148.410
27)	L6 AR-1254-2	5.506	4.677	49023240	63782741	168.765m	214.645 #
28)	L6 AR-1254-3	5.854	5.058	41414456	56889515	134.809m	121.802m
29)	L6 AR-1254-4	6.132	5.284	37113579	48175061	169.051m	163.166
30)	L6 AR-1254-5	6.542	5.687	163.1E6	227.7E6	703.418	541.766
31)	L7 AR-1260-1	6.018	5.186	98581616	125.9E6	420.443m	349.310m
32)	L7 AR-1260-2	6.275	5.377	116.0E6	140.7E6	419.237m	319.751
33)	L7 AR-1260-3	6.626	5.516	64602779	151.4E6	324.135	364.829
34)	L7 AR-1260-4	6.840	5.978	76642151	101.8E6	356.669	282.828
35)	L7 AR-1260-5	7.146	6.222	156.2E6	234.0E6	360.118	278.474

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Sample Results: **PQ071153.D**

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ102725\
Data File : PQ071153.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Oct 2025 09:55
Operator : YP\AJ
Sample : Q3450-02MS
Misc :
ALS Vial : 13 Sample Multiplier: 1


Instrument :
ECD_Q
ClientSampleId :
SS-5R-(3.5-4.0)MS


Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/27/2025
Supervised By :mohammad ahmed 10/29/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 27 11:30:04 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Oct 11 01:55:41 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





Raw Data: **PP076280.D**

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Matrix New World Engineering	Date Collected:	10/23/25
Project:	PSEG Athenia Substation	Date Received:	10/23/25
Client Sample ID:	SS-10R-(3.5-4.0)MS	SDG No.:	Q3450
Lab Sample ID:	Q3450-04MS	Matrix:	SOIL
Analytical Method:	8082A	% Solid:	86.6
Sample Wt/Vol:	30.02 g Final Vol: 10000 uL	Test:	PCB
Prep Method:	SW3541B Prep Date: 10/31/25		

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
12674-11-2	Aroclor-1016	193		1	4.60	19.6	ug/kg	10/31/25 23:56	PB170358
11104-28-2	Aroclor-1221	4.70	U	1	4.70	19.6	ug/kg	10/31/25 23:56	PB170358
11141-16-5	Aroclor-1232	4.30	U	1	4.30	19.6	ug/kg	10/31/25 23:56	PB170358
53469-21-9	Aroclor-1242	4.60	U	1	4.60	19.6	ug/kg	10/31/25 23:56	PB170358
12672-29-6	Aroclor-1248	6.80	U	1	6.80	19.6	ug/kg	10/31/25 23:56	PB170358
11097-69-1	Aroclor-1254	3.70	U	1	3.70	19.6	ug/kg	10/31/25 23:56	PB170358
37324-23-5	Aroclor-1262	5.80	U	1	5.80	19.6	ug/kg	10/31/25 23:56	PB170358
11100-14-4	Aroclor-1268	4.20	U	1	4.20	19.6	ug/kg	10/31/25 23:56	PB170358
11096-82-5	Aroclor-1260	221		1	3.70	19.6	ug/kg	10/31/25 23:56	PB170358
SURROGATES									
877-09-8	Tetrachloro-m-xylene	25.4			30 (21) - 150 (165)	127%	SPK: 20		
2051-24-3	Decachlorobiphenyl	22.3			30 (10) - 150 (170)	112%	SPK: 20		

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
E = Value Exceeds Calibration Range
P = Indicates >25% difference for detected concentrations between the two GC columns
Q = indicates LCS control criteria did not meet requirements
M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
B = Analyte Found in Associated Method Blank
N = Presumptive Evidence of a Compound
* = Values outside of QC limits
D = Dilution
S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.
() = Laboratory InHouse Limit

Sample Results: **PP076280.D**

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP103125\
Data File : PP076280.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 31 Oct 2025 23:56
Operator : YP\AJ
Sample : Q3450-04MS
Misc :
ALS Vial : 30 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
SS-10R-(3.5-4.0)MS

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 11/03/2025
Supervised By :Yogesh Patel 11/03/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 01 02:48:04 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 24 15:40:32 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.644	3.781	36250473	150.8E6	25.366	22.782
2)	SA Decachlor...	10.421	8.778	24051909	159.7E6	22.342	19.416
	Target Compounds						
3)	L1 AR-1016-1	5.797	4.876	30046314	355.6E6	508.173	480.364
4)	L1 AR-1016-2	5.819	4.933	44808665	158.9E6	506.633	464.680
5)	L1 AR-1016-3	5.881	5.054	27633553	86159548	499.242	419.589
6)	L1 AR-1016-4	5.979	5.095	23248808	88287315	511.635	472.414
7)	L1 AR-1016-5	6.271	5.308	20784720	99976121	478.850	438.525
31)	L7 AR-1260-1	7.389	6.337	41152966	316.8E6	598.531	578.363
32)	L7 AR-1260-2	7.642	6.523	53659231	493.0E6	599.027	605.803
33)	L7 AR-1260-3	8.000	6.675	33149951	342.1E6	478.447	607.951 #
34)	L7 AR-1260-4	8.227	7.145	49589358	213.9E6	723.179	411.851 #
35)	L7 AR-1260-5	8.554	7.385	70962810	690.7E6	479.635	509.542

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP103125\
Data File : PP076280.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 31 Oct 2025 23:56
Operator : YP\AJ
Sample : Q3450-04MS
Misc :
ALS Vial : 30 Sample Multiplier: 1


Instrument :
ECD_P
ClientSampleId :
SS-10R-(3.5-4.0)MS


Manual Integrations APPROVED

Reviewed By :Rahul Chavli 11/03/2025
Supervised By :Yogesh Patel 11/03/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 01 02:48:04 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 24 15:40:32 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





Raw Data: **PO115040.D**

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Matrix New World Engineering	Date Collected:	11/10/25
Project:	PSEG Athenia Substation	Date Received:	11/10/25
Client Sample ID:	TP-1MS	SDG No.:	Q3450
Lab Sample ID:	Q3597-01MS	Matrix:	SOIL
Analytical Method:	8082A	% Solid:	97.4
Sample Wt/Vol:	30.04 g	Final Vol: 10000 uL	Test: PCB
Prep Method:	SW3541B	Prep Date: 11/12/25	

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
12674-11-2	Aroclor-1016	157		1	4.10	17.4	ug/kg	11/12/25 12:53	PB170497
11104-28-2	Aroclor-1221	4.10	U	1	4.10	17.4	ug/kg	11/12/25 12:53	PB170497
11141-16-5	Aroclor-1232	3.80	U	1	3.80	17.4	ug/kg	11/12/25 12:53	PB170497
53469-21-9	Aroclor-1242	4.10	U	1	4.10	17.4	ug/kg	11/12/25 12:53	PB170497
12672-29-6	Aroclor-1248	6.10	U	1	6.10	17.4	ug/kg	11/12/25 12:53	PB170497
11097-69-1	Aroclor-1254	3.30	U	1	3.30	17.4	ug/kg	11/12/25 12:53	PB170497
37324-23-5	Aroclor-1262	5.10	U	1	5.10	17.4	ug/kg	11/12/25 12:53	PB170497
11100-14-4	Aroclor-1268	3.70	U	1	3.70	17.4	ug/kg	11/12/25 12:53	PB170497
11096-82-5	Aroclor-1260	145		1	3.30	17.4	ug/kg	11/12/25 12:53	PB170497
SURROGATES									
877-09-8	Tetrachloro-m-xylene	25.0			30 (30) - 150 (150)	125%	SPK: 20		
2051-24-3	Decachlorobiphenyl	24.3			30 (30) - 150 (150)	122%	SPK: 20		

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
E = Value Exceeds Calibration Range
P = Indicates >25% difference for detected concentrations between the two GC columns
Q = indicates LCS control criteria did not meet requirements
M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
B = Analyte Found in Associated Method Blank
N = Presumptive Evidence of a Compound
* = Values outside of QC limits
D = Dilution
S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.
() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO111225\
Data File : PO115040.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Nov 2025 12:53
Operator : YP/AJ
Sample : Q3597-01MS
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
TP-1MS

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 12 21:45:56 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Nov 05 06:00:07 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.347	3.648	182.6E6	111.7E6	25.939	25.016
2)	SA Decachlor...	9.933	8.642	107.5E6	101.1E6	24.319	25.897
	Target Compounds						
3)	L1 AR-1016-1	5.486	4.727	116.4E6	88301400	473.130	464.120
4)	L1 AR-1016-2	5.508	4.746	185.0E6	126.8E6	465.858	467.990
5)	L1 AR-1016-3	5.570	4.920	105.0E6	69077775	471.008	466.826
6)	L1 AR-1016-4	5.666	4.963	90425410	53116247	478.758	452.370
7)	L1 AR-1016-5	5.958	5.175	79821811	69122104	447.941	452.626
31)	L7 AR-1260-1	7.071	6.206	137.9E6	133.3E6	451.102	450.266
32)	L7 AR-1260-2	7.325	6.393	187.7E6	189.6E6	474.156	440.969
33)	L7 AR-1260-3	7.683	6.546	130.1E6	156.3E6	392.211	444.971
34)	L7 AR-1260-4	7.906	7.017	125.7E6	118.6E6	412.124	391.471
35)	L7 AR-1260-5	8.211	7.258	294.7E6	324.6E6	413.167	395.039

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Sample Results: **PO115040.D**


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO111225\
Data File : PO115040.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Nov 2025 12:53
Operator : YP/AJ
Sample : Q3597-01MS
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
TP-1MS

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 12 21:45:56 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Nov 05 06:00:07 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





Raw Data: **PQ071154.D**

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Matrix New World Engineering	Date Collected:	10/23/25
Project:	PSEG Athenia Substation	Date Received:	10/23/25
Client Sample ID:	SS-5R-(3.5-4.0)MSD	SDG No.:	Q3450
Lab Sample ID:	Q3450-02MSD	Matrix:	SOIL
Analytical Method:	8082A	% Solid:	86.3
Sample Wt/Vol:	30.04 g	Final Vol: 10000 uL	Test: PCB
Prep Method:	SW3541B	Prep Date: 10/24/25	

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
12674-11-2	Aroclor-1016	162		1	4.60	19.7	ug/kg	10/27/25 10:10	PB170236
11104-28-2	Aroclor-1221	4.70	U	1	4.70	19.7	ug/kg	10/27/25 10:10	PB170236
11141-16-5	Aroclor-1232	4.30	U	1	4.30	19.7	ug/kg	10/27/25 10:10	PB170236
53469-21-9	Aroclor-1242	4.60	U	1	4.60	19.7	ug/kg	10/27/25 10:10	PB170236
12672-29-6	Aroclor-1248	6.90	U	1	6.90	19.7	ug/kg	10/27/25 10:10	PB170236
11097-69-1	Aroclor-1254	148		1	3.70	19.7	ug/kg	10/27/25 10:10	PB170236
37324-23-5	Aroclor-1262	5.80	U	1	5.80	19.7	ug/kg	10/27/25 10:10	PB170236
11100-14-4	Aroclor-1268	4.20	U	1	4.20	19.7	ug/kg	10/27/25 10:10	PB170236
11096-82-5	Aroclor-1260	166		1	3.70	19.7	ug/kg	10/27/25 10:10	PB170236
SURROGATES									
877-09-8	Tetrachloro-m-xylene	25.7			30 (21) - 150 (165)	128%	SPK: 20		
2051-24-3	Decachlorobiphenyl	24.0			30 (10) - 150 (170)	120%	SPK: 20		

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
E = Value Exceeds Calibration Range
P = Indicates >25% difference for detected concentrations between the two GC columns
Q = indicates LCS control criteria did not meet requirements
M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
B = Analyte Found in Associated Method Blank
N = Presumptive Evidence of a Compound
* = Values outside of QC limits
D = Dilution
S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.
() = Laboratory InHouse Limit

Sample Results: **PQ071154.D**

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ102725\
Data File : PQ071154.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Oct 2025 10:10
Operator : YP\AJ
Sample : Q3450-02MSD
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
SS-5R-(3.5-4.0)MSD

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/27/2025
Supervised By :mohammad ahmed 10/29/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 27 11:30:42 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Oct 11 01:55:41 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	3.399	2.776	111.8E6	112.9E6	25.689	22.228
2)	SA Decachlor...	8.496	7.531	69730856	115.6E6	24.012	22.895
	Target Compounds						
3)	L1 AR-1016-1	4.478	3.775	66133318	86839134	402.898m	387.671m
4)	L1 AR-1016-2	4.497	3.791	106.6E6	113.2E6	415.843m	332.751m
5)	L1 AR-1016-3	4.555	3.952	68384934	72254323	429.774m	397.310
6)	L1 AR-1016-4	4.645	4.000	55794798	61092237	429.424m	388.102
7)	L1 AR-1016-5	4.930	4.195	53701684	67804394	416.929	353.439
26)	L6 AR-1254-1	5.292	4.531	46316796	64856412	235.975	192.331
27)	L6 AR-1254-2	5.507	4.678	67669054	75430352	232.954	253.842
28)	L6 AR-1254-3	5.854	5.058	79307585	109.9E6	258.156	235.279
29)	L6 AR-1254-4	6.134	5.284	64884968	78163021	295.549	264.734
30)	L6 AR-1254-5	6.543	5.687	208.6E6	305.6E6	899.407	726.931
31)	L7 AR-1260-1	6.019	5.187	119.0E6	170.6E6	507.455	473.416
32)	L7 AR-1260-2	6.275	5.376	141.2E6	188.2E6	510.578	427.657
33)	L7 AR-1260-3	6.625	5.516	71042700	189.8E6	356.446	457.353 #
34)	L7 AR-1260-4	6.840	5.977	85195497	113.2E6	396.473	314.727
35)	L7 AR-1260-5	7.145	6.221	165.8E6	260.8E6	382.125	310.420

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Sample Results: **PQ071154.D**

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ102725\
Data File : PQ071154.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Oct 2025 10:10
Operator : YP\AJ
Sample : Q3450-02MSD
Misc :
ALS Vial : 14 Sample Multiplier: 1


Instrument :
ECD_Q
ClientSampleId :
SS-5R-(3.5-4.0)MSD


Manual Integrations APPROVED

Reviewed By :Yogesh Patel 10/27/2025
Supervised By :mohammad ahmed 10/29/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 27 11:30:42 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ101025.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Oct 11 01:55:41 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





Raw Data: **PP076281.D**

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Matrix New World Engineering	Date Collected:	10/23/25
Project:	PSEG Athenia Substation	Date Received:	10/23/25
Client Sample ID:	SS-10R-(3.5-4.0)MSD	SDG No.:	Q3450
Lab Sample ID:	Q3450-04MSD	Matrix:	SOIL
Analytical Method:	8082A	% Solid:	86.6
Sample Wt/Vol:	30.04 g	Final Vol: 10000 uL	Test: PCB
Prep Method:	SW3541B	Prep Date: 10/31/25	

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
12674-11-2	Aroclor-1016	189		1	4.60	19.6	ug/kg	11/01/25 00:12	PB170358
11104-28-2	Aroclor-1221	4.60	U	1	4.60	19.6	ug/kg	11/01/25 00:12	PB170358
11141-16-5	Aroclor-1232	4.30	U	1	4.30	19.6	ug/kg	11/01/25 00:12	PB170358
53469-21-9	Aroclor-1242	4.60	U	1	4.60	19.6	ug/kg	11/01/25 00:12	PB170358
12672-29-6	Aroclor-1248	6.80	U	1	6.80	19.6	ug/kg	11/01/25 00:12	PB170358
11097-69-1	Aroclor-1254	3.70	U	1	3.70	19.6	ug/kg	11/01/25 00:12	PB170358
37324-23-5	Aroclor-1262	5.80	U	1	5.80	19.6	ug/kg	11/01/25 00:12	PB170358
11100-14-4	Aroclor-1268	4.20	U	1	4.20	19.6	ug/kg	11/01/25 00:12	PB170358
11096-82-5	Aroclor-1260	215		1	3.70	19.6	ug/kg	11/01/25 00:12	PB170358
SURROGATES									
877-09-8	Tetrachloro-m-xylene	24.5			30 (21) - 150 (165)	123%	SPK: 20		
2051-24-3	Decachlorobiphenyl	21.6			30 (10) - 150 (170)	108%	SPK: 20		

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
E = Value Exceeds Calibration Range
P = Indicates >25% difference for detected concentrations between the two GC columns
Q = indicates LCS control criteria did not meet requirements
M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
B = Analyte Found in Associated Method Blank
N = Presumptive Evidence of a Compound
* = Values outside of QC limits
D = Dilution
S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.
() = Laboratory InHouse Limit

Sample Results: **PP076281.D**

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP103125\
Data File : PP076281.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 01 Nov 2025 00:12
Operator : YP\AJ
Sample : Q3450-04MSD
Misc :
ALS Vial : 31 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
SS-10R-(3.5-4.0)MSD

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 11/03/2025
Supervised By :Yogesh Patel 11/03/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 01 02:48:23 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 24 15:40:32 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml	
	System Monitoring Compounds							
1)	SA Tetrachlo...	4.645	3.781	35006523	145.2E6	24.496	21.940	
2)	SA Decachlor...	10.422	8.780	23293010	150.0E6	21.637	18.244	
	Target Compounds							
3)	L1 AR-1016-1	5.798	4.876	29301588	347.3E6	495.578	469.171	
4)	L1 AR-1016-2	5.819	4.933	43612529	158.4E6	493.108	463.261	
5)	L1 AR-1016-3	5.882	5.054	26982644	87031207	487.482	423.834	
6)	L1 AR-1016-4	5.979	5.095	23017921	87217966	506.554	466.692	
7)	L1 AR-1016-5	6.272	5.308	20484744	98194985	471.939	430.712	
31)	L7 AR-1260-1	7.390	6.336	40083418	319.3E6	582.975	582.996	
32)	L7 AR-1260-2	7.643	6.523	52157765	488.1E6	582.265	599.767	
33)	L7 AR-1260-3	8.000	6.676	32210765	331.4E6	464.892	588.926	#
34)	L7 AR-1260-4	8.227	7.145	48140524	210.7E6	702.050	405.749	#
35)	L7 AR-1260-5	8.554	7.385	68980673	672.2E6	466.238	495.893	

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Sample Results: **PP076281.D**

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP103125\
Data File : PP076281.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 01 Nov 2025 00:12
Operator : YP\AJ
Sample : Q3450-04MSD
Misc :
ALS Vial : 31 Sample Multiplier: 1


Instrument :
ECD_P
ClientSampleId :
SS-10R-(3.5-4.0)MSD


Manual Integrations APPROVED

Reviewed By :Rahul Chavli 11/03/2025
Supervised By :Yogesh Patel 11/03/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 01 02:48:23 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102125.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 24 15:40:32 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





Raw Data: **PO115041.D**

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Matrix New World Engineering	Date Collected:	11/10/25
Project:	PSEG Athenia Substation	Date Received:	11/10/25
Client Sample ID:	TP-1MSD	SDG No.:	Q3450
Lab Sample ID:	Q3597-01MSD	Matrix:	SOIL
Analytical Method:	8082A	% Solid:	97.4
Sample Wt/Vol:	30.08 g	Final Vol: 10000 uL	Test: PCB
Prep Method:	SW3541B	Prep Date: 11/12/25	

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
12674-11-2	Aroclor-1016	158		1	4.00	17.4	ug/kg	11/12/25 13:11	PB170497
11104-28-2	Aroclor-1221	4.10	U	1	4.10	17.4	ug/kg	11/12/25 13:11	PB170497
11141-16-5	Aroclor-1232	3.80	U	1	3.80	17.4	ug/kg	11/12/25 13:11	PB170497
53469-21-9	Aroclor-1242	4.10	U	1	4.10	17.4	ug/kg	11/12/25 13:11	PB170497
12672-29-6	Aroclor-1248	6.10	U	1	6.10	17.4	ug/kg	11/12/25 13:11	PB170497
11097-69-1	Aroclor-1254	3.30	U	1	3.30	17.4	ug/kg	11/12/25 13:11	PB170497
37324-23-5	Aroclor-1262	5.10	U	1	5.10	17.4	ug/kg	11/12/25 13:11	PB170497
11100-14-4	Aroclor-1268	3.70	U	1	3.70	17.4	ug/kg	11/12/25 13:11	PB170497
11096-82-5	Aroclor-1260	145		1	3.30	17.4	ug/kg	11/12/25 13:11	PB170497
SURROGATES									
877-09-8	Tetrachloro-m-xylene	25.1			30 (30) - 150 (150)	125%	SPK: 20		
2051-24-3	Decachlorobiphenyl	24.2			30 (30) - 150 (150)	121%	SPK: 20		

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
E = Value Exceeds Calibration Range
P = Indicates >25% difference for detected concentrations between the two GC columns
Q = indicates LCS control criteria did not meet requirements
M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
B = Analyte Found in Associated Method Blank
N = Presumptive Evidence of a Compound
* = Values outside of QC limits
D = Dilution
S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.
() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO111225\
Data File : PO115041.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Nov 2025 13:11
Operator : YP/AJ
Sample : Q3597-01MSD
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
TP-1MSD

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 12 21:46:38 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Nov 05 06:00:07 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.347	3.648	182.9E6	111.9E6	25.984	25.054
2)	SA Decachlor...	9.936	8.642	107.1E6	101.8E6	24.223	26.080
	Target Compounds						
3)	L1 AR-1016-1	5.487	4.727	115.8E6	89586614	470.882	470.875
4)	L1 AR-1016-2	5.508	4.746	185.3E6	126.2E6	466.650	465.755
5)	L1 AR-1016-3	5.570	4.921	104.6E6	69446667	469.080	469.319
6)	L1 AR-1016-4	5.666	4.963	88808256	53754844	470.196	457.808
7)	L1 AR-1016-5	5.958	5.175	78889531	69295929	442.709	453.764
31)	L7 AR-1260-1	7.072	6.206	140.5E6	132.9E6	459.774	448.812
32)	L7 AR-1260-2	7.326	6.393	186.2E6	188.9E6	470.291	439.328
33)	L7 AR-1260-3	7.683	6.546	133.6E6	156.5E6	402.683	445.777
34)	L7 AR-1260-4	7.905	7.017	128.4E6	119.2E6	420.994	393.265
35)	L7 AR-1260-5	8.212	7.258	289.7E6	322.7E6	406.256	392.763


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO111225\
Data File : PO115041.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Nov 2025 13:11
Operator : YP/AJ
Sample : Q3597-01MSD
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
TP-1MSD

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 12 21:46:38 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Nov 05 06:00:07 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

Manual Integration Report

Sequence:	po110425	Instrument	ECD_o

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
AR1660ICC250	PO114874.D	AR-1260-1 #2	rahul	11/5/2025 8:39:44 AM	yogesh	11/5/2025 10:22:06	Peak Integrated by Software
AR1660ICC250	PO114874.D	AR-1260-2 #2	rahul	11/5/2025 8:39:44 AM	yogesh	11/5/2025 10:22:06	Peak Integrated by Software
AR1660ICC050	PO114875.D	AR-1016-3 #2	rahul	11/5/2025 8:39:48 AM	yogesh	11/5/2025 10:22:08	Peak Integrated by Software
AR1660ICC050	PO114875.D	AR-1016-5 #2	rahul	11/5/2025 8:39:48 AM	yogesh	11/5/2025 10:22:08	Peak Integrated by Software
AR1660ICC050	PO114875.D	AR-1260-1	rahul	11/5/2025 8:39:48 AM	yogesh	11/5/2025 10:22:08	Peak Integrated by Software
AR1660ICC050	PO114875.D	AR-1260-1 #2	rahul	11/5/2025 8:39:48 AM	yogesh	11/5/2025 10:22:08	Peak Integrated by Software
AR1660ICC050	PO114875.D	AR-1260-2	rahul	11/5/2025 8:39:48 AM	yogesh	11/5/2025 10:22:08	Peak Integrated by Software
AR1660ICC050	PO114875.D	AR-1260-2 #2	rahul	11/5/2025 8:39:48 AM	yogesh	11/5/2025 10:22:08	Peak Integrated by Software
AR1660ICC050	PO114875.D	AR-1260-3	rahul	11/5/2025 8:39:48 AM	yogesh	11/5/2025 10:22:08	Peak Integrated by Software
AR1660ICC050	PO114875.D	AR-1260-4	rahul	11/5/2025 8:39:48 AM	yogesh	11/5/2025 10:22:08	Peak Integrated by Software
AR1242ICC050	PO114882.D	AR-1242-5 #2	rahul	11/5/2025 8:39:51 AM	yogesh	11/5/2025 10:22:11	Peak Integrated by Software
AR1254ICC050	PO114892.D	AR-1254-1	rahul	11/5/2025 8:40:06 AM	yogesh	11/5/2025 10:22:11	Peak Integrated by Software
AR1254ICC050	PO114892.D	AR-1254-2	rahul	11/5/2025 8:40:06 AM	yogesh	11/5/2025 10:22:11	Peak Integrated by Software

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

Manual Integration Report

Sequence:	po110425	Instrument	ECD_o

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
AR1254ICC050	PO114892.D	AR-1254-4	rahul	11/5/2025 8:40:06 AM	yogesh	11/5/2025 10:22:11	Peak Integrated by Software
AR1254ICC050	PO114892.D	Decachlorobiphenyl	rahul	11/5/2025 8:40:06 AM	yogesh	11/5/2025 10:22:11	Peak Integrated by Software
AR1268ICC250	PO114897.D	AR-1268-1 #2	rahul	11/5/2025 8:40:09 AM	yogesh	11/5/2025 10:22:13	Peak Integrated by Software
AR1242ICV500	PO114900.D	AR-1242-1	rahul	11/5/2025 8:40:12 AM	yogesh	11/5/2025 10:22:14	Peak Integrated by Software
AR1242ICV500	PO114900.D	AR-1242-2	rahul	11/5/2025 8:40:12 AM	yogesh	11/5/2025 10:22:14	Peak Integrated by Software
AR1242ICV500	PO114900.D	AR-1242-3	rahul	11/5/2025 8:40:12 AM	yogesh	11/5/2025 10:22:14	Peak Integrated by Software
AR1242ICV500	PO114900.D	AR-1242-4	rahul	11/5/2025 8:40:12 AM	yogesh	11/5/2025 10:22:14	Peak Integrated by Software

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

Manual Integration Report

Sequence:	PO111225	Instrument	ECD_o

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

Manual Integration Report

Sequence:	PP102125	Instrument	ECD_p

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
I.BLK	PP075887.D	Decachlorobiphenyl	yogesh	10/22/2025 12:24:41 PM	mohammad	10/25/2025 7:28:37	Peak Integrated by Software
AR1660ICC050	PP075892.D	AR-1016-3	yogesh	10/22/2025 12:25:25 PM	mohammad	10/25/2025 7:28:37	Peak Integrated by Software
AR1660ICC050	PP075892.D	AR-1016-3 #2	yogesh	10/22/2025 12:25:25 PM	mohammad	10/25/2025 7:28:37	Peak Integrated by Software
AR1660ICC050	PP075892.D	AR-1016-4	yogesh	10/22/2025 12:25:25 PM	mohammad	10/25/2025 7:28:37	Peak Integrated by Software
AR1660ICC050	PP075892.D	AR-1016-4 #2	yogesh	10/22/2025 12:25:25 PM	mohammad	10/25/2025 7:28:37	Peak Integrated by Software
AR1660ICC050	PP075892.D	AR-1016-5 #2	yogesh	10/22/2025 12:25:25 PM	mohammad	10/25/2025 7:28:37	Peak Integrated by Software
AR1660ICC050	PP075892.D	Tetrachloro-m-xylene #2	yogesh	10/22/2025 12:25:25 PM	mohammad	10/25/2025 7:28:37	Peak Integrated by Software
AR1232ICC500	PP075894.D	AR-1232-4 #2	yogesh	10/22/2025 12:24:44 PM	mohammad	10/25/2025 7:28:37	Peak Integrated by Software
AR1242ICC1000	PP075895.D	AR-1242-4 #2	yogesh	10/22/2025 12:24:45 PM	mohammad	10/25/2025 7:28:37	Peak Integrated by Software
AR1242ICC750	PP075896.D	AR-1242-4 #2	yogesh	10/22/2025 12:24:46 PM	mohammad	10/25/2025 7:28:37	Peak Integrated by Software
AR1242ICC500	PP075897.D	AR-1242-4 #2	yogesh	10/22/2025 12:24:48 PM	mohammad	10/25/2025 7:28:37	Peak Integrated by Software
AR1242ICC250	PP075898.D	AR-1242-4 #2	yogesh	10/22/2025 12:24:50 PM	mohammad	10/25/2025 7:28:37	Peak Integrated by Software
AR1242ICC250	PP075898.D	AR-1242-5	yogesh	10/22/2025 12:24:50 PM	mohammad	10/25/2025 7:28:37	Peak Integrated by Software

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

Manual Integration Report

Sequence:	PP102125	Instrument	ECD_p

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
AR1242ICC050	PP075899.D	AR-1242-5	yogesh	10/22/2025 12:24:52 PM	mohammad	10/25/2025 7:28:37	Peak Integrated by Software
AR1242ICC050	PP075899.D	Tetrachloro-m-xylene #2	yogesh	10/22/2025 12:24:52 PM	mohammad	10/25/2025 7:28:37	Peak Integrated by Software
AR1248ICC1000	PP075900.D	AR-1248-3 #2	yogesh	10/22/2025 12:24:53 PM	mohammad	10/25/2025 7:28:37	Peak Integrated by Software
AR1248ICC750	PP075901.D	AR-1248-3 #2	yogesh	10/22/2025 12:24:55 PM	mohammad	10/25/2025 7:28:37	Peak Integrated by Software
AR1248ICC500	PP075902.D	AR-1248-3 #2	yogesh	10/22/2025 12:24:57 PM	mohammad	10/25/2025 7:28:37	Peak Integrated by Software
AR1248ICC250	PP075903.D	AR-1248-3 #2	yogesh	10/22/2025 12:24:58 PM	mohammad	10/25/2025 7:28:37	Peak Integrated by Software
AR1248ICC050	PP075904.D	AR-1248-4	yogesh	10/22/2025 12:25:00 PM	mohammad	10/25/2025 7:28:37	Peak Integrated by Software
AR1248ICC050	PP075904.D	AR-1248-4 #2	yogesh	10/22/2025 12:25:00 PM	mohammad	10/25/2025 7:28:37	Peak Integrated by Software
AR1248ICC050	PP075904.D	AR-1248-5	yogesh	10/22/2025 12:25:00 PM	mohammad	10/25/2025 7:28:37	Peak Integrated by Software
AR1248ICC050	PP075904.D	AR-1248-5 #2	yogesh	10/22/2025 12:25:00 PM	mohammad	10/25/2025 7:28:37	Peak Integrated by Software
AR1248ICC050	PP075904.D	Tetrachloro-m-xylene #2	yogesh	10/22/2025 12:25:00 PM	mohammad	10/25/2025 7:28:37	Peak Integrated by Software
AR1254ICC1000	PP075905.D	AR-1254-1	yogesh	10/22/2025 12:25:02 PM	mohammad	10/25/2025 7:28:37	Peak Integrated by Software
AR1254ICC750	PP075906.D	AR-1254-1	yogesh	10/22/2025 12:25:03 PM	mohammad	10/25/2025 7:28:37	Peak Integrated by Software

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

Manual Integration Report

Sequence:	PP102125	Instrument	ECD_p

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
AR1254ICC750	PP075906.D	AR-1254-5 #2	yogesh	10/22/2025 12:25:03 PM	mohammad	10/25/2025 7:28:37	Peak Integrated by Software
AR1254ICC500	PP075907.D	AR-1254-1	yogesh	10/22/2025 12:25:05 PM	mohammad	10/25/2025 7:28:37	Peak Integrated by Software
AR1254ICC250	PP075908.D	AR-1254-1	yogesh	10/22/2025 12:25:07 PM	mohammad	10/25/2025 7:28:37	Peak Integrated by Software
AR1254ICC050	PP075909.D	Tetrachloro-m-xylene #2	yogesh	10/22/2025 12:25:09 PM	mohammad	10/25/2025 7:28:37	Peak Integrated by Software
AR1242ICV500	PP075917.D	AR-1242-4 #2	yogesh	10/24/2025 4:12:30 PM	mohammad	10/25/2025 7:28:37	Peak Integrated by Software
AR1242ICV500	PP075917.D	Decachlorobiphenyl	yogesh	10/24/2025 4:12:30 PM	mohammad	10/25/2025 7:28:37	Peak Integrated by Software
AR1248ICV500	PP075918.D	AR-1248-3 #2	yogesh	10/24/2025 4:12:35 PM	mohammad	10/25/2025 7:28:37	Peak Integrated by Software
AR1254ICV500	PP075919.D	AR-1254-1 #2	yogesh	10/24/2025 4:12:33 PM	mohammad	10/25/2025 7:28:37	Peak Integrated by Software

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

Manual Integration Report

Sequence:	PP102425	Instrument	ECD_p

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
AR1242CCC500	PP076013.D	AR-1242-1 #2	yogesh	10/24/2025 4:09:50 PM	mohammad	10/28/2025 6:10:31	Peak Integrated by Software
AR1242CCC500	PP076013.D	AR-1242-2 #2	yogesh	10/24/2025 4:09:50 PM	mohammad	10/28/2025 6:10:31	Peak Integrated by Software
AR1242CCC500	PP076013.D	AR-1242-3 #2	yogesh	10/24/2025 4:09:50 PM	mohammad	10/28/2025 6:10:31	Peak Integrated by Software
AR1242CCC500	PP076013.D	AR-1242-4 #2	yogesh	10/24/2025 4:09:50 PM	mohammad	10/28/2025 6:10:31	Peak Integrated by Software
AR1242CCC500	PP076013.D	AR-1242-5 #2	yogesh	10/24/2025 4:09:50 PM	mohammad	10/28/2025 6:10:31	Peak Integrated by Software
AR1248CCC500	PP076014.D	AR-1248-3 #2	yogesh	10/24/2025 4:09:51 PM	mohammad	10/28/2025 6:10:31	Peak Integrated by Software
AR1248CCC500	PP076014.D	AR-1248-5 #2	yogesh	10/24/2025 4:09:51 PM	mohammad	10/28/2025 6:10:31	Peak Integrated by Software
AR1254CCC500	PP076015.D	AR-1254-1	yogesh	10/24/2025 4:09:57 PM	mohammad	10/28/2025 6:10:31	Peak Integrated by Software
AR1242CCC500	PP076028.D	AR-1242-4 #2	yogesh	10/24/2025 4:10:02 PM	mohammad	10/28/2025 6:10:31	Peak Integrated by Software
AR1248CCC500	PP076029.D	AR-1248-3 #2	yogesh	10/24/2025 4:10:04 PM	mohammad	10/28/2025 6:10:31	Peak Integrated by Software
AR1254CCC500	PP076030.D	AR-1254-1	yogesh	10/27/2025 8:23:51 AM	mohammad	10/28/2025 6:10:31	Peak Integrated by Software
AR1242CCC500	PP076043.D	AR-1242-4 #2	yogesh	10/27/2025 8:24:14 AM	mohammad	10/28/2025 6:10:31	Peak Integrated by Software
AR1248CCC500	PP076044.D	AR-1248-3 #2	yogesh	10/27/2025 8:24:15 AM	mohammad	10/28/2025 6:10:31	Peak Integrated by Software

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

Manual Integration Report

Sequence:	PP102425	Instrument	ECD_p

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
AR1254CCC500	PP076045.D	AR-1254-1	yogesh	10/27/2025 8:24:17 AM	mohammad	10/28/2025 6:10:31	Peak Integrated by Software
AR1254CCC500	PP076045.D	AR-1254-1 #2	yogesh	10/27/2025 8:24:17 AM	mohammad	10/28/2025 6:10:31	Peak Integrated by Software
AR1254CCC500	PP076045.D	Tetrachloro-m-xylene #2	yogesh	10/27/2025 8:24:17 AM	mohammad	10/28/2025 6:10:31	Peak Integrated by Software
Q3450-10	PP076047.D	AR-1254-1	yogesh	10/27/2025 8:24:19 AM	mohammad	10/28/2025 6:10:31	Peak Integrated by Software
AR1242CCC500	PP076058.D	AR-1242-4 #2	yogesh	10/27/2025 8:24:32 AM	mohammad	10/28/2025 6:10:31	Peak Integrated by Software
AR1248CCC500	PP076059.D	AR-1248-3 #2	yogesh	10/27/2025 8:24:34 AM	mohammad	10/28/2025 6:10:31	Peak Integrated by Software
AR1254CCC500	PP076060.D	AR-1254-1	yogesh	10/27/2025 8:24:36 AM	mohammad	10/28/2025 6:10:31	Peak Integrated by Software
AR1254CCC500	PP076060.D	AR-1254-1 #2	yogesh	10/27/2025 8:24:36 AM	mohammad	10/28/2025 6:10:31	Peak Integrated by Software
AR1254CCC500	PP076060.D	AR-1254-2 #2	yogesh	10/27/2025 8:24:36 AM	mohammad	10/28/2025 6:10:31	Peak Integrated by Software
AR1254CCC500	PP076060.D	AR-1254-3 #2	yogesh	10/27/2025 8:24:36 AM	mohammad	10/28/2025 6:10:31	Peak Integrated by Software
AR1254CCC500	PP076060.D	AR-1254-4 #2	yogesh	10/27/2025 8:24:36 AM	mohammad	10/28/2025 6:10:31	Peak Integrated by Software
AR1254CCC500	PP076060.D	AR-1254-5 #2	yogesh	10/27/2025 8:24:36 AM	mohammad	10/28/2025 6:10:31	Peak Integrated by Software

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

Manual Integration Report

Sequence:	PP102425	Instrument	ECD_p

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

Manual Integration Report

Sequence:	PP102725	Instrument	ECD_p

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
AR1660CCC50 0	PP076063.D	AR-1016-1 #2	yogesh	10/28/2025 7:45:04 AM	mohammad	10/29/2025 5:27:46	Peak Integrated by Software
AR1660CCC50 0	PP076063.D	AR-1016-2 #2	yogesh	10/28/2025 7:45:04 AM	mohammad	10/29/2025 5:27:46	Peak Integrated by Software
AR1242CCC50 0	PP076064.D	AR-1242-4 #2	yogesh	10/28/2025 7:45:06 AM	mohammad	10/29/2025 5:27:46	Peak Integrated by Software
AR1248CCC50 0	PP076065.D	AR-1248-3 #2	yogesh	10/28/2025 7:45:08 AM	mohammad	10/29/2025 5:27:46	Peak Integrated by Software
AR1254CCC50 0	PP076066.D	AR-1254-1	yogesh	10/28/2025 7:45:20 AM	mohammad	10/29/2025 5:27:46	Peak Integrated by Software
PB170236BL	PP076068.D	Decachlorobiphenyl	yogesh	10/28/2025 7:45:22 AM	mohammad	10/29/2025 5:27:46	Peak Integrated by Software
PB170236BL	PP076068.D	Tetrachloro-m-xylene	yogesh	10/28/2025 7:45:22 AM	mohammad	10/29/2025 5:27:46	Peak Integrated by Software
PB170236BS	PP076069.D	AR-1260-4	yogesh	10/28/2025 7:45:24 AM	mohammad	10/29/2025 5:27:46	Peak Integrated by Software
PB170236BS	PP076069.D	Tetrachloro-m-xylene	yogesh	10/28/2025 7:45:24 AM	mohammad	10/29/2025 5:27:46	Peak Integrated by Software
PB170236BS	PP076069.D	Tetrachloro-m-xylene #2	yogesh	10/28/2025 7:45:24 AM	mohammad	10/29/2025 5:27:46	Peak Integrated by Software
Q3450-20DL	PP076072.D	Tetrachloro-m-xylene	yogesh	10/28/2025 7:45:26 AM	mohammad	10/29/2025 5:27:46	Peak Integrated by Software
Q3450-22DL	PP076073.D	AR-1254-1	yogesh	10/28/2025 7:45:27 AM	mohammad	10/29/2025 5:27:46	Peak Integrated by Software
Q3450-22DL	PP076073.D	AR-1254-1 #2	yogesh	10/28/2025 7:45:27 AM	mohammad	10/29/2025 5:27:46	Peak Integrated by Software

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

Manual Integration Report

Sequence:	PP102725	Instrument	ECD_p

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
Q3450-22DL	PP076073.D	AR-1254-2	yogesh	10/28/2025 7:45:27 AM	mohammad	10/29/2025 5:27:46	Peak Integrated by Software
Q3450-22DL	PP076073.D	AR-1254-2 #2	yogesh	10/28/2025 7:45:27 AM	mohammad	10/29/2025 5:27:46	Peak Integrated by Software
AR1242CCC500	PP076079.D	AR-1242-4 #2	yogesh	10/28/2025 7:45:37 AM	mohammad	10/29/2025 5:27:46	Peak Integrated by Software
AR1248CCC500	PP076080.D	AR-1248-3 #2	yogesh	10/28/2025 7:45:40 AM	mohammad	10/29/2025 5:27:46	Peak Integrated by Software
AR1248CCC500	PP076080.D	AR-1248-5 #2	yogesh	10/28/2025 7:45:40 AM	mohammad	10/29/2025 5:27:46	Peak Integrated by Software
AR1254CCC500	PP076081.D	AR-1254-1	yogesh	10/28/2025 7:45:42 AM	mohammad	10/29/2025 5:27:46	Peak Integrated by Software
AR1242CCC500	PP076094.D	AR-1242-4 #2	yogesh	10/28/2025 7:46:25 AM	mohammad	10/29/2025 5:27:46	Peak Integrated by Software
AR1248CCC500	PP076095.D	AR-1248-3 #2	yogesh	10/28/2025 7:46:27 AM	mohammad	10/29/2025 5:27:46	Peak Integrated by Software
AR1254CCC500	PP076096.D	AR-1254-1	yogesh	10/28/2025 7:46:29 AM	mohammad	10/29/2025 5:27:46	Peak Integrated by Software
AR1254CCC500	PP076096.D	AR-1254-1 #2	yogesh	10/28/2025 7:46:29 AM	mohammad	10/29/2025 5:27:46	Peak Integrated by Software
AR1242CCC500	PP076105.D	AR-1242-4 #2	yogesh	10/28/2025 7:46:39 AM	mohammad	10/29/2025 5:27:46	Peak Integrated by Software
AR1242CCC500	PP076105.D	AR-1242-5 #2	yogesh	10/28/2025 7:46:39 AM	mohammad	10/29/2025 5:27:46	Peak Integrated by Software
AR1248CCC500	PP076106.D	AR-1248-3 #2	yogesh	10/28/2025 7:46:41 AM	mohammad	10/29/2025 5:27:46	Peak Integrated by Software

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

Manual Integration Report

Sequence:	PP102725	Instrument	ECD_p

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

Manual Integration Report

Sequence:	pp103125	Instrument	ECD_p

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
AR1242CCC500	PP076243.D	AR-1242-4 #2	rahul	11/3/2025 8:15:40 AM	yogesh	11/3/2025 8:47:04	Peak Integrated by Software
AR1242CCC500	PP076243.D	AR-1242-5 #2	rahul	11/3/2025 8:15:40 AM	yogesh	11/3/2025 8:47:04	Peak Integrated by Software
AR1248CCC500	PP076244.D	AR-1248-3 #2	rahul	11/3/2025 8:15:43 AM	yogesh	11/3/2025 8:47:07	Peak Integrated by Software
AR1248CCC500	PP076244.D	AR-1248-5 #2	rahul	11/3/2025 8:15:43 AM	yogesh	11/3/2025 8:47:07	Peak Integrated by Software
AR1254CCC500	PP076245.D	AR-1254-1	rahul	11/3/2025 8:15:46 AM	yogesh	11/3/2025 8:47:10	Peak Integrated by Software
AR1254CCC500	PP076245.D	AR-1254-1 #2	rahul	11/3/2025 8:15:46 AM	yogesh	11/3/2025 8:47:10	Peak Integrated by Software
AR1254CCC500	PP076245.D	Tetrachloro-m-xylene #2	rahul	11/3/2025 8:15:46 AM	yogesh	11/3/2025 8:47:10	Peak Integrated by Software
AR1660CCC500	PP076256.D	AR-1260-4 #2	rahul	11/3/2025 8:16:13 AM	yogesh	11/3/2025 8:47:36	Peak Integrated by Software
AR1242CCC500	PP076257.D	AR-1242-4 #2	rahul	11/3/2025 8:16:16 AM	yogesh	11/3/2025 8:47:39	Peak Integrated by Software
AR1242CCC500	PP076257.D	AR-1242-5 #2	rahul	11/3/2025 8:16:16 AM	yogesh	11/3/2025 8:47:39	Peak Integrated by Software
AR1242CCC500	PP076257.D	Tetrachloro-m-xylene	rahul	11/3/2025 8:16:16 AM	yogesh	11/3/2025 8:47:39	Peak Integrated by Software
AR1242CCC500	PP076257.D	Tetrachloro-m-xylene #2	rahul	11/3/2025 8:16:16 AM	yogesh	11/3/2025 8:47:39	Peak Integrated by Software
AR1248CCC500	PP076258.D	AR-1248-3 #2	rahul	11/3/2025 8:16:22 AM	yogesh	11/3/2025 8:47:41	Peak Integrated by Software

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

Manual Integration Report

Sequence:	pp103125	Instrument	ECD_p

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
AR1248CCC500	PP076258.D	AR-1248-5 #2	rahul	11/3/2025 8:16:22 AM	yogesh	11/3/2025 8:47:41	Peak Integrated by Software
AR1248CCC500	PP076258.D	Tetrachloro-m-xylene #2	rahul	11/3/2025 8:16:22 AM	yogesh	11/3/2025 8:47:41	Peak Integrated by Software
AR1254CCC500	PP076259.D	AR-1254-1	rahul	11/3/2025 8:16:24 AM	yogesh	11/3/2025 8:47:43	Peak Integrated by Software
AR1254CCC500	PP076259.D	AR-1254-1 #2	rahul	11/3/2025 8:16:24 AM	yogesh	11/3/2025 8:47:43	Peak Integrated by Software
AR1254CCC500	PP076259.D	AR-1254-5 #2	rahul	11/3/2025 8:16:24 AM	yogesh	11/3/2025 8:47:43	Peak Integrated by Software
AR1254CCC500	PP076259.D	Tetrachloro-m-xylene #2	rahul	11/3/2025 8:16:24 AM	yogesh	11/3/2025 8:47:43	Peak Integrated by Software
AR1242CCC500	PP076272.D	AR-1242-1 #2	rahul	11/3/2025 8:17:16 AM	yogesh	11/3/2025 8:48:28	Peak Integrated by Software
AR1242CCC500	PP076272.D	AR-1242-4 #2	rahul	11/3/2025 8:17:16 AM	yogesh	11/3/2025 8:48:28	Peak Integrated by Software
AR1242CCC500	PP076272.D	AR-1242-5 #2	rahul	11/3/2025 8:17:16 AM	yogesh	11/3/2025 8:48:28	Peak Integrated by Software
AR1248CCC500	PP076273.D	AR-1248-3 #2	rahul	11/3/2025 8:17:34 AM	yogesh	11/3/2025 8:48:30	Peak Integrated by Software
AR1248CCC500	PP076273.D	Tetrachloro-m-xylene #2	rahul	11/3/2025 8:17:34 AM	yogesh	11/3/2025 8:48:30	Peak Integrated by Software
AR1254CCC500	PP076274.D	AR-1254-1	rahul	11/3/2025 8:17:37 AM	yogesh	11/3/2025 8:48:33	Peak Integrated by Software
AR1254CCC500	PP076274.D	AR-1254-1 #2	rahul	11/3/2025 8:17:37 AM	yogesh	11/3/2025 8:48:33	Peak Integrated by Software

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

Manual Integration Report

Sequence:	pp103125	Instrument	ECD_p

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
AR1254CCC500	PP076274.D	Tetrachloro-m-xylene #2	rahul	11/3/2025 8:17:37 AM	yogesh	11/3/2025 8:48:33	Peak Integrated by Software
Q3450-04MS	PP076280.D	AR-1254-1	rahul	11/3/2025 8:17:47 AM	yogesh	11/3/2025 8:48:41	Peak Integrated by Software
Q3450-04MSD	PP076281.D	AR-1254-1	rahul	11/3/2025 8:17:50 AM	yogesh	11/3/2025 8:48:48	Peak Integrated by Software
Q3450-04MSD	PP076281.D	AR-1254-1 #2	rahul	11/3/2025 8:17:50 AM	yogesh	11/3/2025 8:48:48	Peak Integrated by Software
Q3450-06	PP076283.D	Tetrachloro-m-xylene	rahul	11/3/2025 8:17:52 AM	yogesh	11/3/2025 8:48:50	Peak Integrated by Software
Q3450-23	PP076284.D	AR-1254-1	rahul	11/3/2025 8:17:55 AM	yogesh	11/3/2025 8:48:52	Peak Integrated by Software
Q3450-23	PP076284.D	AR-1260-4	rahul	11/3/2025 8:17:55 AM	yogesh	11/3/2025 8:48:52	Peak Integrated by Software
Q3450-25	PP076285.D	AR-1260-4	rahul	11/3/2025 8:17:57 AM	yogesh	11/3/2025 8:48:54	Peak Integrated by Software
AR1660CCC500	PP076286.D	AR-1260-4 #2	rahul	11/3/2025 8:18:01 AM	yogesh	11/3/2025 8:48:56	Peak Integrated by Software
AR1242CCC500	PP076287.D	AR-1242-4 #2	rahul	11/3/2025 8:18:04 AM	yogesh	11/3/2025 8:48:58	Peak Integrated by Software
AR1242CCC500	PP076287.D	AR-1242-5 #2	rahul	11/3/2025 8:18:04 AM	yogesh	11/3/2025 8:48:58	Peak Integrated by Software
AR1248CCC500	PP076288.D	AR-1248-3 #2	rahul	11/3/2025 8:18:07 AM	yogesh	11/3/2025 8:49:00	Peak Integrated by Software
AR1248CCC500	PP076288.D	AR-1248-5 #2	rahul	11/3/2025 8:18:07 AM	yogesh	11/3/2025 8:49:00	Peak Integrated by Software

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

Manual Integration Report

Sequence:	pp103125	Instrument	ECD_p

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
AR1254CCC500	PP076289.D	AR-1254-1	rahul	11/3/2025 8:18:10 AM	yogesh	11/3/2025 8:49:02	Peak Integrated by Software
AR1254CCC500	PP076289.D	AR-1254-1 #2	rahul	11/3/2025 8:18:10 AM	yogesh	11/3/2025 8:49:02	Peak Integrated by Software
AR1254CCC500	PP076289.D	AR-1254-2 #2	rahul	11/3/2025 8:18:10 AM	yogesh	11/3/2025 8:49:02	Peak Integrated by Software
AR1254CCC500	PP076289.D	AR-1254-3 #2	rahul	11/3/2025 8:18:10 AM	yogesh	11/3/2025 8:49:02	Peak Integrated by Software
Q3450-04	PP076279.D	AR-1254-1	rahul	11/3/2025 8:17:45 AM	yogesh	11/3/2025 8:48:39	Peak Integrated by Software
Q3450-04	PP076279.D	AR-1254-1 #2	rahul	11/3/2025 8:17:45 AM	yogesh	11/3/2025 8:48:39	Peak Integrated by Software
Q3450-04	PP076279.D	AR-1254-2	rahul	11/3/2025 8:17:45 AM	yogesh	11/3/2025 8:48:39	Peak Integrated by Software
Q3450-04	PP076279.D	AR-1254-2 #2	rahul	11/3/2025 8:17:45 AM	yogesh	11/3/2025 8:48:39	Peak Integrated by Software
Q3450-04	PP076279.D	AR-1254-3 #2	rahul	11/3/2025 8:17:45 AM	yogesh	11/3/2025 8:48:39	Peak Integrated by Software
Q3450-04	PP076279.D	AR-1254-4	rahul	11/3/2025 8:17:45 AM	yogesh	11/3/2025 8:48:39	Peak Integrated by Software
Q3450-04	PP076279.D	AR-1254-5 #2	rahul	11/3/2025 8:17:45 AM	yogesh	11/3/2025 8:48:39	Peak Integrated by Software
Q3450-04	PP076279.D	AR-1260-2 #2	rahul	11/3/2025 8:17:45 AM	yogesh	11/3/2025 8:48:39	Peak Integrated by Software
Q3450-04	PP076279.D	AR-1260-3 #2	rahul	11/3/2025 8:17:45 AM	yogesh	11/3/2025 8:48:39	Peak Integrated by Software

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

Manual Integration Report

Sequence:	pp103125	Instrument	ECD_p

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
Q3450-04	PP076279.D	AR-1260-4 #2	rahul	11/3/2025 8:17:45 AM	yogesh	11/3/2025 8:48:39	Peak Integrated by Software
Q3450-04	PP076279.D	AR-1260-5 #2	rahul	11/3/2025 8:17:45 AM	yogesh	11/3/2025 8:48:39	Peak Integrated by Software
PB170358BS	PP076277.D	AR-1260-4	rahul	11/3/2025 8:17:40 AM	yogesh	11/3/2025 8:48:35	Peak Integrated by Software

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

Manual Integration Report

Sequence:	Pp110325	Instrument	ECD_p

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
AR1242CCC500	PP076293.D	AR-1242-4 #2	rahul	11/4/2025 8:40:48 AM	yogesh	11/4/2025 10:10:08	Peak Integrated by Software
AR1248CCC500	PP076294.D	AR-1248-3 #2	rahul	11/4/2025 8:40:51 AM	yogesh	11/4/2025 10:10:09	Peak Integrated by Software
AR1254CCC500	PP076295.D	AR-1254-1	rahul	11/4/2025 8:40:53 AM	yogesh	11/4/2025 10:10:12	Peak Integrated by Software
Q3450-23DL	PP076297.D	AR-1254-1	rahul	11/4/2025 8:40:56 AM	yogesh	11/4/2025 10:10:13	Peak Integrated by Software
Q3450-23DL	PP076297.D	Tetrachloro-m-xylene #2	rahul	11/4/2025 8:40:56 AM	yogesh	11/4/2025 10:10:13	Peak Integrated by Software
AR1242CCC500	PP076305.D	AR-1242-4 #2	rahul	11/4/2025 8:41:16 AM	yogesh	11/4/2025 10:10:27	Peak Integrated by Software
AR1248CCC500	PP076306.D	AR-1248-2 #2	rahul	11/4/2025 8:41:19 AM	yogesh	11/4/2025 10:10:29	Peak Integrated by Software
AR1248CCC500	PP076306.D	AR-1248-3 #2	rahul	11/4/2025 8:41:19 AM	yogesh	11/4/2025 10:10:29	Peak Integrated by Software
AR1248CCC500	PP076306.D	AR-1248-4 #2	rahul	11/4/2025 8:41:19 AM	yogesh	11/4/2025 10:10:29	Peak Integrated by Software
AR1248CCC500	PP076306.D	AR-1248-5 #2	rahul	11/4/2025 8:41:19 AM	yogesh	11/4/2025 10:10:29	Peak Integrated by Software
AR1254CCC500	PP076307.D	AR-1254-1	rahul	11/4/2025 8:41:21 AM	yogesh	11/4/2025 10:10:31	Peak Integrated by Software
AR1254CCC500	PP076307.D	AR-1254-1 #2	rahul	11/4/2025 8:41:21 AM	yogesh	11/4/2025 10:10:31	Peak Integrated by Software
AR1254CCC500	PP076307.D	AR-1254-5 #2	rahul	11/4/2025 8:41:21 AM	yogesh	11/4/2025 10:10:31	Peak Integrated by Software

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

Manual Integration Report

Sequence:	Pp110325	Instrument	ECD_p

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

Manual Integration Report

Sequence:	pq101025	Instrument	ECD_q

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
AR1660ICC050	PQ070977.D	AR-1260-1	yogesh	10/13/2025 8:00:35 AM	mohammad	10/14/2025 2:48:24	Peak Integrated by Software
AR1660ICC050	PQ070977.D	AR-1260-2	yogesh	10/13/2025 8:00:35 AM	mohammad	10/14/2025 2:48:24	Peak Integrated by Software
AR1242ICC050	PQ070984.D	AR-1242-1 #2	yogesh	10/13/2025 8:00:37 AM	mohammad	10/14/2025 2:48:24	Peak Integrated by Software
AR1242ICC050	PQ070984.D	AR-1242-2 #2	yogesh	10/13/2025 8:00:37 AM	mohammad	10/14/2025 2:48:24	Peak Integrated by Software
AR1242ICC050	PQ070984.D	AR-1242-3 #2	yogesh	10/13/2025 8:00:37 AM	mohammad	10/14/2025 2:48:24	Peak Integrated by Software
AR1242ICC050	PQ070984.D	AR-1242-4 #2	yogesh	10/13/2025 8:00:37 AM	mohammad	10/14/2025 2:48:24	Peak Integrated by Software
AR1242ICC050	PQ070984.D	AR-1242-5	yogesh	10/13/2025 8:00:37 AM	mohammad	10/14/2025 2:48:24	Peak Integrated by Software
AR1242ICC050	PQ070984.D	AR-1242-5 #2	yogesh	10/13/2025 8:00:37 AM	mohammad	10/14/2025 2:48:24	Peak Integrated by Software
AR1248ICC250	PQ070988.D	AR-1248-4 #2	yogesh	10/13/2025 8:00:39 AM	mohammad	10/14/2025 2:48:24	Peak Integrated by Software
AR1248ICC250	PQ070988.D	AR-1248-5 #2	yogesh	10/13/2025 8:00:39 AM	mohammad	10/14/2025 2:48:24	Peak Integrated by Software
PQ101025ICV500	PQ071001.D	AR-1016-1 #2	rahul	10/31/2025 8:33:47 AM	yogesh	10/31/2025 8:34:28	Peak Integrated by Software
PQ101025ICV500	PQ071001.D	AR-1016-2 #2	rahul	10/31/2025 8:33:47 AM	yogesh	10/31/2025 8:34:28	Peak Integrated by Software
PQ101025ICV500	PQ071001.D	AR-1016-3 #2	rahul	10/31/2025 8:33:47 AM	yogesh	10/31/2025 8:34:28	Peak Integrated by Software

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

Manual Integration Report

Sequence:	pq101025	Instrument	ECD_q

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
PQ101025ICV500	PQ071001.D	AR-1016-4 #2	rahul	10/31/2025 8:33:47 AM	yogesh	10/31/2025 8:34:28	Peak Integrated by Software
PQ101025ICV500	PQ071001.D	AR-1016-5 #2	rahul	10/31/2025 8:33:47 AM	yogesh	10/31/2025 8:34:28	Peak Integrated by Software
PQ101025ICV500	PQ071001.D	AR-1260-1	rahul	10/31/2025 8:33:47 AM	yogesh	10/31/2025 8:34:28	Peak Integrated by Software
PQ101025ICV500	PQ071001.D	AR-1260-1 #2	rahul	10/31/2025 8:33:47 AM	yogesh	10/31/2025 8:34:28	Peak Integrated by Software
PQ101025ICV500	PQ071001.D	AR-1260-2	rahul	10/31/2025 8:33:47 AM	yogesh	10/31/2025 8:34:28	Peak Integrated by Software
PQ101025ICV500	PQ071001.D	AR-1260-2 #2	rahul	10/31/2025 8:33:47 AM	yogesh	10/31/2025 8:34:28	Peak Integrated by Software
AR1242ICV500	PQ071002.D	AR-1242-3 #2	yogesh	10/13/2025 8:00:43 AM	mohammad	10/14/2025 2:48:24	Peak Integrated by Software
AR1248ICV500	PQ071003.D	AR-1248-4	yogesh	10/13/2025 8:00:45 AM	mohammad	10/14/2025 2:48:24	Peak Integrated by Software
AR1248ICV500	PQ071003.D	AR-1248-4 #2	yogesh	10/13/2025 8:00:45 AM	mohammad	10/14/2025 2:48:24	Peak Integrated by Software
AR1248ICV500	PQ071003.D	AR-1248-5	yogesh	10/13/2025 8:00:45 AM	mohammad	10/14/2025 2:48:24	Peak Integrated by Software

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

Manual Integration Report

Sequence:	PQ102725	Instrument	ECD_q

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
Q3450-02MS	PQ071153.D	AR-1016-1	yogesh	10/27/2025 1:46:49 PM	mohammad	10/29/2025 5:27:51	Peak Integrated by Software
Q3450-02MS	PQ071153.D	AR-1016-1 #2	yogesh	10/27/2025 1:46:49 PM	mohammad	10/29/2025 5:27:51	Peak Integrated by Software
Q3450-02MS	PQ071153.D	AR-1016-2	yogesh	10/27/2025 1:46:49 PM	mohammad	10/29/2025 5:27:51	Peak Integrated by Software
Q3450-02MS	PQ071153.D	AR-1016-2 #2	yogesh	10/27/2025 1:46:49 PM	mohammad	10/29/2025 5:27:51	Peak Integrated by Software
Q3450-02MS	PQ071153.D	AR-1016-3	yogesh	10/27/2025 1:46:49 PM	mohammad	10/29/2025 5:27:51	Peak Integrated by Software
Q3450-02MS	PQ071153.D	AR-1016-3 #2	yogesh	10/27/2025 1:46:49 PM	mohammad	10/29/2025 5:27:51	Peak Integrated by Software
Q3450-02MS	PQ071153.D	AR-1016-4	yogesh	10/27/2025 1:46:49 PM	mohammad	10/29/2025 5:27:51	Peak Integrated by Software
Q3450-02MS	PQ071153.D	AR-1016-4 #2	yogesh	10/27/2025 1:46:49 PM	mohammad	10/29/2025 5:27:51	Peak Integrated by Software
Q3450-02MS	PQ071153.D	AR-1254-1	yogesh	10/27/2025 1:46:49 PM	mohammad	10/29/2025 5:27:51	Peak Integrated by Software
Q3450-02MS	PQ071153.D	AR-1254-2	yogesh	10/27/2025 1:46:49 PM	mohammad	10/29/2025 5:27:51	Peak Integrated by Software
Q3450-02MS	PQ071153.D	AR-1254-3	yogesh	10/27/2025 1:46:49 PM	mohammad	10/29/2025 5:27:51	Peak Integrated by Software
Q3450-02MS	PQ071153.D	AR-1254-3 #2	yogesh	10/27/2025 1:46:49 PM	mohammad	10/29/2025 5:27:51	Peak Integrated by Software
Q3450-02MS	PQ071153.D	AR-1254-4	yogesh	10/27/2025 1:46:49 PM	mohammad	10/29/2025 5:27:51	Peak Integrated by Software

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

Manual Integration Report

Sequence:	PQ102725	Instrument	ECD_q

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
Q3450-02MS	PQ071153.D	AR-1260-1	yogesh	10/27/2025 1:46:49 PM	mohammad	10/29/2025 5:27:51	Peak Integrated by Software
Q3450-02MS	PQ071153.D	AR-1260-1 #2	yogesh	10/27/2025 1:46:49 PM	mohammad	10/29/2025 5:27:51	Peak Integrated by Software
Q3450-02MS	PQ071153.D	AR-1260-2	yogesh	10/27/2025 1:46:49 PM	mohammad	10/29/2025 5:27:51	Peak Integrated by Software
Q3450-02MS	PQ071153.D	Tetrachloro-m-xylene #2	yogesh	10/27/2025 1:46:49 PM	mohammad	10/29/2025 5:27:51	Peak Integrated by Software
Q3450-02MSD	PQ071154.D	AR-1016-1	yogesh	10/27/2025 1:46:51 PM	mohammad	10/29/2025 5:27:51	Peak Integrated by Software
Q3450-02MSD	PQ071154.D	AR-1016-1 #2	yogesh	10/27/2025 1:46:51 PM	mohammad	10/29/2025 5:27:51	Peak Integrated by Software
Q3450-02MSD	PQ071154.D	AR-1016-2	yogesh	10/27/2025 1:46:51 PM	mohammad	10/29/2025 5:27:51	Peak Integrated by Software
Q3450-02MSD	PQ071154.D	AR-1016-2 #2	yogesh	10/27/2025 1:46:51 PM	mohammad	10/29/2025 5:27:51	Peak Integrated by Software
Q3450-02MSD	PQ071154.D	AR-1016-3	yogesh	10/27/2025 1:46:51 PM	mohammad	10/29/2025 5:27:51	Peak Integrated by Software
Q3450-02MSD	PQ071154.D	AR-1016-4	yogesh	10/27/2025 1:46:51 PM	mohammad	10/29/2025 5:27:51	Peak Integrated by Software
Q3450-09	PQ071156.D	AR-1254-3 #2	yogesh	10/27/2025 1:46:53 PM	mohammad	10/29/2025 5:27:51	Peak Integrated by Software
Q3450-09	PQ071156.D	AR-1254-4 #2	yogesh	10/27/2025 1:46:53 PM	mohammad	10/29/2025 5:27:51	Peak Integrated by Software
AR1254CCC500	PQ071179.D	AR-1254-1	yogesh	10/28/2025 7:20:03 AM	mohammad	10/29/2025 5:27:51	Peak Integrated by Software

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

Manual Integration Report

Sequence:	PQ102725	Instrument	ECD_q

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
AR1254CCC500	PQ071179.D	AR-1254-2	yogesh	10/28/2025 7:20:03 AM	mohammad	10/29/2025 5:27:51	Peak Integrated by Software
AR1254CCC500	PQ071179.D	AR-1254-3	yogesh	10/28/2025 7:20:03 AM	mohammad	10/29/2025 5:27:51	Peak Integrated by Software
AR1254CCC500	PQ071179.D	AR-1254-4	yogesh	10/28/2025 7:20:03 AM	mohammad	10/29/2025 5:27:51	Peak Integrated by Software
AR1254CCC500	PQ071179.D	AR-1254-5	yogesh	10/28/2025 7:20:03 AM	mohammad	10/29/2025 5:27:51	Peak Integrated by Software
AR1254CCC500	PQ071179.D	Decachlorobiphenyl	yogesh	10/28/2025 7:20:03 AM	mohammad	10/29/2025 5:27:51	Peak Integrated by Software
AR1660CCC500	PQ071176.D	AR-1016-3	yogesh	10/28/2025 7:20:00 AM	mohammad	10/29/2025 5:27:51	Peak Integrated by Software
AR1660CCC500	PQ071176.D	AR-1016-4	yogesh	10/28/2025 7:20:00 AM	mohammad	10/29/2025 5:27:51	Peak Integrated by Software
AR1660CCC500	PQ071176.D	AR-1260-1	yogesh	10/28/2025 7:20:00 AM	mohammad	10/29/2025 5:27:51	Peak Integrated by Software
AR1660CCC500	PQ071176.D	AR-1260-3	yogesh	10/28/2025 7:20:00 AM	mohammad	10/29/2025 5:27:51	Peak Integrated by Software
AR1660CCC500	PQ071176.D	AR-1260-4	yogesh	10/28/2025 7:20:00 AM	mohammad	10/29/2025 5:27:51	Peak Integrated by Software
AR1660CCC500	PQ071176.D	AR-1260-5	yogesh	10/28/2025 7:20:00 AM	mohammad	10/29/2025 5:27:51	Peak Integrated by Software

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QCBatch ID # PO110425

Review By	yogesh	Review On	11/4/2025 12:15:43 PM		
Supervise By	rahul	Supervise On	11/5/2025 8:40:26 AM		
SubDirectory	PO110425	HP Acquire Method		HP Processing Method	PO110425

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829 ,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854
CCC	
Internal Standard/PEM	
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	HEXANE	PO114869.D	04 Nov 2025 09:03	YP/AJ	Ok
2	I.BLK	PO114870.D	04 Nov 2025 09:21	YP/AJ	Ok
3	AR1660ICC1000	PO114871.D	04 Nov 2025 09:39	YP/AJ	Ok
4	AR1660ICC750	PO114872.D	04 Nov 2025 09:58	YP/AJ	Ok
5	AR1660ICC500	PO114873.D	04 Nov 2025 10:16	YP/AJ	Ok
6	AR1660ICC250	PO114874.D	04 Nov 2025 10:34	YP/AJ	Ok,M
7	AR1660ICC050	PO114875.D	04 Nov 2025 10:53	YP/AJ	Ok,M
8	AR1221ICC500	PO114876.D	04 Nov 2025 11:11	YP/AJ	Ok
9	AR1232ICC500	PO114877.D	04 Nov 2025 11:30	YP/AJ	Ok
10	AR1242ICC1000	PO114878.D	04 Nov 2025 11:48	YP/AJ	Ok
11	AR1242ICC750	PO114879.D	04 Nov 2025 12:06	YP/AJ	Ok
12	AR1242ICC500	PO114880.D	04 Nov 2025 12:25	YP/AJ	Ok
13	AR1242ICC250	PO114881.D	04 Nov 2025 12:42	YP/AJ	Ok
14	AR1242ICC050	PO114882.D	04 Nov 2025 13:00	YP/AJ	Ok,M
15	AR1248ICC1000	PO114883.D	04 Nov 2025 13:19	YP/AJ	Ok
16	AR1248ICC750	PO114884.D	04 Nov 2025 13:37	YP/AJ	Ok
17	AR1248ICC500	PO114885.D	04 Nov 2025 13:56	YP/AJ	Ok
18	AR1248ICC250	PO114886.D	04 Nov 2025 14:14	YP/AJ	Ok
19	AR1248ICC050	PO114887.D	04 Nov 2025 14:33	YP/AJ	Ok
20	AR1254ICC1000	PO114888.D	04 Nov 2025 14:51	YP/AJ	Ok
21	AR1254ICC750	PO114889.D	04 Nov 2025 15:09	YP/AJ	Ok

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QCBatch ID # PO110425

Review By	yogesh	Review On	11/4/2025 12:15:43 PM		
Supervise By	rahul	Supervise On	11/5/2025 8:40:26 AM		
SubDirectory	PO110425	HP Acquire Method		HP Processing Method	PO110425

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854
CCC	
Internal Standard/PEM	
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

22	AR1254ICC500	PO114890.D	04 Nov 2025 15:28	YP/AJ	Ok
23	AR1254ICC250	PO114891.D	04 Nov 2025 16:04	YP/AJ	Ok
24	AR1254ICC050	PO114892.D	04 Nov 2025 16:23	YP/AJ	Ok,M
25	AR1262ICC500	PO114893.D	04 Nov 2025 17:00	YP/AJ	Ok
26	AR1268ICC1000	PO114894.D	04 Nov 2025 17:18	YP/AJ	Ok
27	AR1268ICC750	PO114895.D	04 Nov 2025 17:36	YP/AJ	Ok
28	AR1268ICC500	PO114896.D	04 Nov 2025 17:55	YP/AJ	Ok
29	AR1268ICC250	PO114897.D	04 Nov 2025 18:13	YP/AJ	Ok,M
30	AR1268ICC050	PO114898.D	04 Nov 2025 18:31	YP/AJ	Ok
31	PO110425ICV500	PO114899.D	04 Nov 2025 19:08	YP/AJ	Ok
32	AR1242ICV500	PO114900.D	04 Nov 2025 20:21	YP/AJ	Ok,M
33	AR1248ICV500	PO114901.D	04 Nov 2025 20:58	YP/AJ	Ok
34	AR1254ICV500	PO114902.D	04 Nov 2025 21:35	YP/AJ	Ok
35	AR1268ICV500	PO114903.D	04 Nov 2025 22:12	YP/AJ	Ok
36	DDT ANALOG	PO114904.D	04 Nov 2025 22:48	YP/AJ	Ok

M : Manual Integration

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QCBatch ID # PO111225

Review By	rahul	Review On	11/13/2025 1:15:59 PM		
Supervise By	yogesh	Supervise On	11/13/2025 1:22:24 PM		
SubDirectory	PO111225	HP Acquire Method		HP Processing Method	PP102125

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829 ,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854
CCC	PP24824,PP24828,PP24832,PP24836,PP24840,PP24844,PP24852
Internal Standard/PEM	
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	HEXANE	PO115030.D	12 Nov 2025 09:31	YP/AJ	Ok
2	AR1660CCC500	PO115031.D	12 Nov 2025 09:49	YP/AJ	Ok
3	AR1242CCC500	PO115032.D	12 Nov 2025 10:08	YP/AJ	Ok
4	AR1248CCC500	PO115033.D	12 Nov 2025 10:27	YP/AJ	Ok
5	AR1254CCC500	PO115034.D	12 Nov 2025 10:45	YP/AJ	Ok
6	I.BLK	PO115035.D	12 Nov 2025 11:03	YP/AJ	Ok
7	Q3561-01DL	PO115036.D	12 Nov 2025 11:22	YP/AJ	Ok,M
8	PB170497BL	PO115037.D	12 Nov 2025 11:58	YP/AJ	Ok
9	PB170497BS	PO115038.D	12 Nov 2025 12:16	YP/AJ	Ok
10	Q3597-01	PO115039.D	12 Nov 2025 12:34	YP/AJ	Ok
11	Q3597-01MS	PO115040.D	12 Nov 2025 12:53	YP/AJ	Ok
12	Q3597-01MSD	PO115041.D	12 Nov 2025 13:11	YP/AJ	Ok
13	Q3597-04	PO115042.D	12 Nov 2025 13:29	YP/AJ	Ok
14	Q3597-07	PO115043.D	12 Nov 2025 13:48	YP/AJ	Ok
15	Q3597-10	PO115044.D	12 Nov 2025 14:06	YP/AJ	Ok
16	Q3450-07	PO115045.D	12 Nov 2025 15:36	YP/AJ	Ok
17	AR1660CCC500	PO115046.D	12 Nov 2025 15:55	YP/AJ	Ok,NR
18	AR1242CCC500	PO115047.D	12 Nov 2025 16:13	YP/AJ	Ok
19	AR1248CCC500	PO115048.D	12 Nov 2025 16:31	YP/AJ	Ok
20	AR1254CCC500	PO115049.D	12 Nov 2025 16:50	YP/AJ	Ok
21	I.BLK	PO115050.D	12 Nov 2025 17:07	YP/AJ	Ok

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QCBatch ID # PO111225

Review By	rahul	Review On	11/13/2025 1:15:59 PM		
Supervise By	yogesh	Supervise On	11/13/2025 1:22:24 PM		
SubDirectory	PO111225	HP Acquire Method		HP Processing Method	PP102125

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829 ,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854
CCC	PP24824,PP24828,PP24832,PP24836,PP24840,PP24844,PP24852
Internal Standard/PEM	
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

22	PB170499BL	PO115051.D	12 Nov 2025 17:26	YP/AJ	Ok
23	PB170499BS	PO115052.D	12 Nov 2025 17:44	YP/AJ	Ok
24	Q3591-01	PO115053.D	12 Nov 2025 18:02	YP/AJ	Ok,M
25	Q3591-02	PO115054.D	12 Nov 2025 18:21	YP/AJ	Ok
26	Q3591-03	PO115055.D	12 Nov 2025 18:39	YP/AJ	Ok,M
27	Q3591-04	PO115056.D	12 Nov 2025 18:58	YP/AJ	Ok,M
28	Q3591-05	PO115057.D	12 Nov 2025 19:16	YP/AJ	Ok,M
29	Q3591-06	PO115058.D	12 Nov 2025 19:34	YP/AJ	Ok,M
30	Q3596-01	PO115059.D	12 Nov 2025 19:53	YP/AJ	Ok,M
31	Q3596-02	PO115060.D	12 Nov 2025 20:11	YP/AJ	Ok,M
32	AR1660CCC500	PO115061.D	12 Nov 2025 21:25	YP/AJ	Ok
33	AR1242CCC500	PO115062.D	12 Nov 2025 22:19	YP/AJ	Ok
34	AR1248CCC500	PO115063.D	12 Nov 2025 22:38	YP/AJ	Ok
35	AR1254CCC500	PO115064.D	12 Nov 2025 22:56	YP/AJ	Ok
36	I.BLK	PO115065.D	12 Nov 2025 23:13	YP/AJ	Ok

M : Manual Integration

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QCBatch ID # PP102125

Review By	yogesh	Review On	10/21/2025 12:01:15 PM		
Supervise By	mohammad	Supervise On	10/25/2025 7:28:37 AM		
SubDirectory	PP102125	HP Acquire Method		HP Processing Method	PP102125

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829 ,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854
CCC	
Internal Standard/PEM	
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	HEXANE	PP075886.D	21 Oct 2025 08:33	YP\AJ	Ok
2	I.BLK	PP075887.D	21 Oct 2025 08:49	YP\AJ	Ok,M
3	AR1660ICC1000	PP075888.D	21 Oct 2025 09:06	YP\AJ	Ok
4	AR1660ICC750	PP075889.D	21 Oct 2025 09:22	YP\AJ	Ok
5	AR1660ICC500	PP075890.D	21 Oct 2025 09:38	YP\AJ	Ok
6	AR1660ICC250	PP075891.D	21 Oct 2025 09:54	YP\AJ	Ok
7	AR1660ICC050	PP075892.D	21 Oct 2025 10:11	YP\AJ	Ok,M
8	AR1221ICC500	PP075893.D	21 Oct 2025 10:27	YP\AJ	Ok
9	AR1232ICC500	PP075894.D	21 Oct 2025 10:43	YP\AJ	Ok,M
10	AR1242ICC1000	PP075895.D	21 Oct 2025 10:59	YP\AJ	Ok,M
11	AR1242ICC750	PP075896.D	21 Oct 2025 11:16	YP\AJ	Ok,M
12	AR1242ICC500	PP075897.D	21 Oct 2025 11:48	YP\AJ	Ok,M
13	AR1242ICC250	PP075898.D	21 Oct 2025 12:04	YP\AJ	Ok,M
14	AR1242ICC050	PP075899.D	21 Oct 2025 12:21	YP\AJ	Ok,M
15	AR1248ICC1000	PP075900.D	21 Oct 2025 12:41	YP\AJ	Ok,M
16	AR1248ICC750	PP075901.D	21 Oct 2025 12:57	YP\AJ	Ok,M
17	AR1248ICC500	PP075902.D	21 Oct 2025 13:13	YP\AJ	Ok,M
18	AR1248ICC250	PP075903.D	21 Oct 2025 13:30	YP\AJ	Ok,M
19	AR1248ICC050	PP075904.D	21 Oct 2025 13:46	YP\AJ	Ok,M
20	AR1254ICC1000	PP075905.D	21 Oct 2025 14:02	YP\AJ	Ok,M
21	AR1254ICC750	PP075906.D	21 Oct 2025 14:18	YP\AJ	Ok,M

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QCBatch ID # PP102125

Review By	yogesh	Review On	10/21/2025 12:01:15 PM		
Supervise By	mohammad	Supervise On	10/25/2025 7:28:37 AM		
SubDirectory	PP102125	HP Acquire Method		HP Processing Method	PP102125

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854
CCC	
Internal Standard/PEM	
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

22	AR1254ICC500	PP075907.D	21 Oct 2025 14:35	YP\AJ	Ok,M
23	AR1254ICC250	PP075908.D	21 Oct 2025 14:51	YP\AJ	Ok,M
24	AR1254ICC050	PP075909.D	21 Oct 2025 15:07	YP\AJ	Ok,M
25	AR1262ICC500	PP075910.D	21 Oct 2025 15:40	YP\AJ	Ok
26	AR1268ICC1000	PP075911.D	21 Oct 2025 15:56	YP\AJ	Ok
27	AR1268ICC750	PP075912.D	21 Oct 2025 16:12	YP\AJ	Ok
28	AR1268ICC500	PP075913.D	21 Oct 2025 16:29	YP\AJ	Ok
29	AR1268ICC250	PP075914.D	21 Oct 2025 16:45	YP\AJ	Ok
30	AR1268ICC050	PP075915.D	21 Oct 2025 17:01	YP\AJ	Ok
31	PP102125ICV500	PP075916.D	21 Oct 2025 17:34	YP\AJ	Ok
32	AR1242ICV500	PP075917.D	21 Oct 2025 18:06	YP\AJ	Ok,M
33	AR1248ICV500	PP075918.D	21 Oct 2025 18:39	YP\AJ	Ok,M
34	AR1254ICV500	PP075919.D	22 Oct 2025 08:49	YP\AJ	Ok,M
35	AR1268ICV500	PP075920.D	22 Oct 2025 09:07	YP\AJ	Ok
36	DDT ANALOG	PP075921.D	22 Oct 2025 09:25	YP\AJ	Ok

M : Manual Integration

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QCBatch ID # PP102425

Review By	yogesh	Review On	10/24/2025 11:09:51 AM		
Supervise By	mohammad	Supervise On	10/28/2025 6:10:31 AM		
SubDirectory	PP102425	HP Acquire Method		HP Processing Method	PP102125

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829 ,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP2484 4,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854
CCC	PP24824,PP24828,PP24832,PP24836,PP24840,PP24844,PP24852
Internal Standard/PEM	
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	HEXANE	PP076011.D	24 Oct 2025 08:41	YP\AJ	Ok
2	AR1660CCC500	PP076012.D	24 Oct 2025 08:57	YP\AJ	Ok
3	AR1242CCC500	PP076013.D	24 Oct 2025 09:29	YP\AJ	Ok,M
4	AR1248CCC500	PP076014.D	24 Oct 2025 09:45	YP\AJ	Ok,M
5	AR1254CCC500	PP076015.D	24 Oct 2025 10:02	YP\AJ	Ok,M
6	I.BLK	PP076016.D	24 Oct 2025 10:18	YP\AJ	Ok
7	PB170234BL	PP076017.D	24 Oct 2025 10:34	YP\AJ	Ok
8	PB170234BS	PP076018.D	24 Oct 2025 10:51	YP\AJ	Ok
9	Q3437-03	PP076019.D	24 Oct 2025 11:07	YP\AJ	Ok
10	Q3437-05	PP076020.D	24 Oct 2025 11:23	YP\AJ	Ok,M
11	Q3437-11	PP076021.D	24 Oct 2025 11:39	YP\AJ	Ok,M
12	Q3445-01	PP076022.D	24 Oct 2025 11:56	YP\AJ	Ok
13	Q3446-01	PP076023.D	24 Oct 2025 12:12	YP\AJ	Dilution
14	Q3446-02	PP076024.D	24 Oct 2025 12:28	YP\AJ	Dilution
15	Q3414-12DL2	PP076025.D	24 Oct 2025 12:44	YP\AJ	Ok
16	Q3414-13DL2	PP076026.D	24 Oct 2025 13:01	YP\AJ	Ok
17	AR1660CCC500	PP076027.D	24 Oct 2025 14:06	YP\AJ	Ok
18	AR1242CCC500	PP076028.D	24 Oct 2025 14:22	YP\AJ	Ok,M
19	AR1248CCC500	PP076029.D	24 Oct 2025 14:38	YP\AJ	Ok,M
20	AR1254CCC500	PP076030.D	24 Oct 2025 14:54	YP\AJ	Ok,M
21	I.BLK	PP076031.D	24 Oct 2025 15:11	YP\AJ	Ok

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QCBatch ID # PP102425

Review By	yogesh	Review On	10/24/2025 11:09:51 AM		
Supervise By	mohammad	Supervise On	10/28/2025 6:10:31 AM		
SubDirectory	PP102425	HP Acquire Method		HP Processing Method	PP102125

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829 ,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854
CCC	PP24824,PP24828,PP24832,PP24836,PP24840,PP24844,PP24852
Internal Standard/PEM	
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

22	Q3414-07DL2	PP076032.D	24 Oct 2025 15:27	YP\AJ	Ok
23	Q3446-06	PP076033.D	24 Oct 2025 15:43	YP\AJ	Dilution
24	Q3446-07	PP076034.D	24 Oct 2025 16:00	YP\AJ	Dilution
25	Q3446-08	PP076035.D	24 Oct 2025 16:16	YP\AJ	Dilution
26	Q3446-03	PP076036.D	24 Oct 2025 16:32	YP\AJ	Dilution
27	Q3446-04	PP076037.D	24 Oct 2025 16:48	YP\AJ	Dilution
28	Q3446-01DL	PP076038.D	24 Oct 2025 17:05	YP\AJ	Ok,M
29	Q3446-02DL	PP076039.D	24 Oct 2025 17:21	YP\AJ	Ok,M
30	Q3446-05DL	PP076040.D	24 Oct 2025 17:37	YP\AJ	Ok,M
31	Q3446-06DL	PP076041.D	24 Oct 2025 17:53	YP\AJ	Ok,M
32	AR1660CCC500	PP076042.D	24 Oct 2025 18:58	YP\AJ	Ok
33	AR1242CCC500	PP076043.D	24 Oct 2025 19:31	YP\AJ	Ok,M
34	AR1248CCC500	PP076044.D	24 Oct 2025 20:03	YP\AJ	Ok,M
35	AR1254CCC500	PP076045.D	24 Oct 2025 20:19	YP\AJ	Ok,M
36	I.BLK	PP076046.D	24 Oct 2025 20:36	YP\AJ	Ok
37	Q3450-10	PP076047.D	24 Oct 2025 20:52	YP\AJ	Ok,M
38	Q3450-11	PP076048.D	24 Oct 2025 21:08	YP\AJ	Ok
39	Q3450-20	PP076049.D	24 Oct 2025 21:24	YP\AJ	Dilution
40	Q3450-21	PP076050.D	24 Oct 2025 21:41	YP\AJ	Ok
41	Q3450-22	PP076051.D	24 Oct 2025 21:57	YP\AJ	Dilution
42	Q3427-01	PP076052.D	24 Oct 2025 22:13	YP\AJ	Ok,M
43	Q3446-07DL	PP076053.D	24 Oct 2025 22:29	YP\AJ	Ok,M
44	Q3446-08DL	PP076054.D	24 Oct 2025 22:45	YP\AJ	Ok,M

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QCBatch ID # PP102425

Review By	yogesh	Review On	10/24/2025 11:09:51 AM		
Supervise By	mohammad	Supervise On	10/28/2025 6:10:31 AM		
SubDirectory	PP102425	HP Acquire Method		HP Processing Method	PP102125

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829 ,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP2484 4,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854
CCC	PP24824,PP24828,PP24832,PP24836,PP24840,PP24844,PP24852
Internal Standard/PEM	
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

45	Q3446-03DL	PP076055.D	24 Oct 2025 23:01	YP\AJ	Ok,M
46	Q3446-04DL	PP076056.D	24 Oct 2025 23:17	YP\AJ	Ok,M
47	AR1660CCC500	PP076057.D	25 Oct 2025 00:22	YP\AJ	Ok
48	AR1242CCC500	PP076058.D	25 Oct 2025 01:11	YP\AJ	Ok,M
49	AR1248CCC500	PP076059.D	25 Oct 2025 01:43	YP\AJ	Ok,M
50	AR1254CCC500	PP076060.D	25 Oct 2025 01:59	YP\AJ	Ok,M
51	I.BLK	PP076061.D	25 Oct 2025 02:16	YP\AJ	Ok

M : Manual Integration

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QCBatch ID # PP102725

Review By	yogesh	Review On	10/27/2025 10:53:48 AM		
Supervise By	mohammad	Supervise On	10/29/2025 5:27:46 AM		
SubDirectory	PP102725	HP Acquire Method		HP Processing Method	PP102125

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829 ,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP2484 4,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854
CCC	PP24824,PP24828,PP24832,PP24836,PP24840,PP24844,PP24852
Internal Standard/PEM	
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	HEXANE	PP076062.D	27 Oct 2025 08:40	YP\AJ	Ok
2	AR1660CCC500	PP076063.D	27 Oct 2025 08:56	YP\AJ	Ok,M
3	AR1242CCC500	PP076064.D	27 Oct 2025 09:13	YP\AJ	Ok,M
4	AR1248CCC500	PP076065.D	27 Oct 2025 09:29	YP\AJ	Ok,M
5	AR1254CCC500	PP076066.D	27 Oct 2025 09:45	YP\AJ	Ok,M
6	I.BLK	PP076067.D	27 Oct 2025 10:02	YP\AJ	Ok
7	PB170236BL	PP076068.D	27 Oct 2025 10:18	YP\AJ	Ok,M
8	PB170236BS	PP076069.D	27 Oct 2025 10:34	YP\AJ	Ok,M
9	Q3439-01	PP076070.D	27 Oct 2025 11:06	YP\AJ	ReRun
10	Q3449-01	PP076071.D	27 Oct 2025 11:26	YP\AJ	Ok
11	Q3450-20DL	PP076072.D	27 Oct 2025 11:42	YP\AJ	Ok,M
12	Q3450-22DL	PP076073.D	27 Oct 2025 11:58	YP\AJ	Ok,M
13	Q3446-05	PP076074.D	27 Oct 2025 12:14	YP\AJ	Not Ok
14	Q3447-03	PP076075.D	27 Oct 2025 12:30	YP\AJ	ReRun
15	Q3447-07	PP076076.D	27 Oct 2025 12:47	YP\AJ	Ok,M
16	Q3446-05	PP076077.D	27 Oct 2025 13:03	YP\AJ	Dilution
17	AR1660CCC500	PP076078.D	27 Oct 2025 14:08	YP\AJ	Ok
18	AR1242CCC500	PP076079.D	27 Oct 2025 14:24	YP\AJ	Ok,M
19	AR1248CCC500	PP076080.D	27 Oct 2025 14:40	YP\AJ	Ok,M
20	AR1254CCC500	PP076081.D	27 Oct 2025 14:57	YP\AJ	Ok,M
21	I.BLK	PP076082.D	27 Oct 2025 15:13	YP\AJ	Ok

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QCBatch ID # PP102725

Review By	yogesh	Review On	10/27/2025 10:53:48 AM		
Supervise By	mohammad	Supervise On	10/29/2025 5:27:46 AM		
SubDirectory	PP102725	HP Acquire Method		HP Processing Method	PP102125

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829 ,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854
CCC	PP24824,PP24828,PP24832,PP24836,PP24840,PP24844,PP24852
Internal Standard/PEM	
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

22	Q3456-01	PP076083.D	27 Oct 2025 15:29	YP\AJ	Ok,M
23	Q3457-01	PP076084.D	27 Oct 2025 15:45	YP\AJ	Ok,M
24	Q3460-01	PP076085.D	27 Oct 2025 16:02	YP\AJ	Ok
25	Q3460-17	PP076086.D	27 Oct 2025 16:18	YP\AJ	Ok,M
26	Q3447-05	PP076087.D	27 Oct 2025 16:34	YP\AJ	Ok,M
27	Q3447-01	PP076088.D	27 Oct 2025 16:50	YP\AJ	ReRun
28	Q3459-01	PP076089.D	27 Oct 2025 17:07	YP\AJ	Ok,M
29	Q3459-02	PP076090.D	27 Oct 2025 17:23	YP\AJ	Ok,M
30	Q3459-03	PP076091.D	27 Oct 2025 17:39	YP\AJ	Ok,M
31	Q3459-04	PP076092.D	27 Oct 2025 17:55	YP\AJ	Ok
32	AR1660CCC500	PP076093.D	27 Oct 2025 19:00	YP\AJ	Ok
33	AR1242CCC500	PP076094.D	27 Oct 2025 19:49	YP\AJ	Ok,M
34	AR1248CCC500	PP076095.D	27 Oct 2025 20:05	YP\AJ	Ok,M
35	AR1254CCC500	PP076096.D	27 Oct 2025 20:21	YP\AJ	Ok,M
36	I.BLK	PP076097.D	27 Oct 2025 20:38	YP\AJ	Ok
37	PB170272BL	PP076098.D	27 Oct 2025 20:54	YP\AJ	Ok
38	PB170272BS	PP076099.D	27 Oct 2025 21:10	YP\AJ	Ok,M
39	Q3459-05	PP076100.D	27 Oct 2025 21:26	YP\AJ	Ok,M
40	Q3459-06	PP076101.D	27 Oct 2025 21:42	YP\AJ	Ok,M
41	Q3459-07	PP076102.D	27 Oct 2025 21:59	YP\AJ	Ok,M
42	Q3459-08	PP076103.D	27 Oct 2025 22:15	YP\AJ	Ok,M
43	AR1660CCC500	PP076104.D	27 Oct 2025 23:20	YP\AJ	Ok
44	AR1242CCC500	PP076105.D	28 Oct 2025 00:08	YP\AJ	Ok,M

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QCBatch ID # PP102725

Review By	yogesh	Review On	10/27/2025 10:53:48 AM		
Supervise By	mohammad	Supervise On	10/29/2025 5:27:46 AM		
SubDirectory	PP102725	HP Acquire Method		HP Processing Method	PP102125

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829 ,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854
CCC	PP24824,PP24828,PP24832,PP24836,PP24840,PP24844,PP24852
Internal Standard/PEM	
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

45	AR1248CCC500	PP076106.D	28 Oct 2025 00:25	YP\AJ	Ok,M
46	AR1254CCC500	PP076107.D	28 Oct 2025 00:41	YP\AJ	Ok
47	I.BLK	PP076108.D	28 Oct 2025 00:57	YP\AJ	Ok

M : Manual Integration

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QCBatch ID # PP103125

Review By	rahul	Review On	11/3/2025 8:18:24 AM		
Supervise By	yogesh	Supervise On	11/3/2025 8:49:19 AM		
SubDirectory	PP103125	HP Acquire Method		HP Processing Method	PP102125

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829 ,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854
CCC	PP24824,PP24828,PP24832,PP24836,PP24840,PP24844,PP24852
Internal Standard/PEM	
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	HEXANE	PP076241.D	31 Oct 2025 09:26	YP\AJ	Ok
2	AR1660CCC500	PP076242.D	31 Oct 2025 09:43	YP\AJ	Ok
3	AR1242CCC500	PP076243.D	31 Oct 2025 09:59	YP\AJ	Ok,M
4	AR1248CCC500	PP076244.D	31 Oct 2025 10:15	YP\AJ	Ok,M
5	AR1254CCC500	PP076245.D	31 Oct 2025 10:32	YP\AJ	Ok,M
6	I.BLK	PP076246.D	31 Oct 2025 10:48	YP\AJ	Ok
7	Q3482-04	PP076247.D	31 Oct 2025 11:09	YP\AJ	Ok,M
8	Q3500-01	PP076248.D	31 Oct 2025 13:21	YP\AJ	Ok,M
9	Q3501-01	PP076249.D	31 Oct 2025 13:38	YP\AJ	Not Ok
10	Q3502-01	PP076250.D	31 Oct 2025 13:54	YP\AJ	Ok,M
11	Q3503-01	PP076251.D	31 Oct 2025 14:10	YP\AJ	Ok,M
12	Q3503-03	PP076252.D	31 Oct 2025 14:26	YP\AJ	Ok,M
13	Q3503-05	PP076253.D	31 Oct 2025 14:43	YP\AJ	Ok,M
14	Q3505-02	PP076254.D	31 Oct 2025 14:59	YP\AJ	Ok,M
15	Q3505-03	PP076255.D	31 Oct 2025 15:15	YP\AJ	Ok,M
16	AR1660CCC500	PP076256.D	31 Oct 2025 16:04	YP\AJ	Ok,M
17	AR1242CCC500	PP076257.D	31 Oct 2025 16:20	YP\AJ	Ok,M
18	AR1248CCC500	PP076258.D	31 Oct 2025 16:37	YP\AJ	Ok,M
19	AR1254CCC500	PP076259.D	31 Oct 2025 16:53	YP\AJ	Ok,M
20	I.BLK	PP076260.D	31 Oct 2025 17:09	YP\AJ	Ok
21	PB170347BL	PP076261.D	31 Oct 2025 17:26	YP\AJ	Ok

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QCBatch ID # PP103125

Review By	rahul	Review On	11/3/2025 8:18:24 AM		
Supervise By	yogesh	Supervise On	11/3/2025 8:49:19 AM		
SubDirectory	PP103125	HP Acquire Method		HP Processing Method	PP102125

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829 ,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854
CCC	PP24824,PP24828,PP24832,PP24836,PP24840,PP24844,PP24852
Internal Standard/PEM	
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

22	PB170347BS	PP076262.D	31 Oct 2025 17:42	YP\AJ	Ok,M
23	PB170347BSD	PP076263.D	31 Oct 2025 17:58	YP\AJ	Ok,M
24	Q3484-01	PP076264.D	31 Oct 2025 18:14	YP\AJ	Ok,M
25	Q3492-01	PP076265.D	31 Oct 2025 18:31	YP\AJ	Ok,M
26	Q3495-03	PP076266.D	31 Oct 2025 18:47	YP\AJ	Ok,M
27	Q3500-02	PP076267.D	31 Oct 2025 19:03	YP\AJ	Ok,M
28	Q3500-03	PP076268.D	31 Oct 2025 19:19	YP\AJ	Ok,M
29	Q3500-04	PP076269.D	31 Oct 2025 19:36	YP\AJ	Ok,M
30	Q3503-07	PP076270.D	31 Oct 2025 19:52	YP\AJ	ReRun
31	AR1660CCC500	PP076271.D	31 Oct 2025 20:57	YP\AJ	Ok
32	AR1242CCC500	PP076272.D	31 Oct 2025 21:46	YP\AJ	Ok,M
33	AR1248CCC500	PP076273.D	31 Oct 2025 22:02	YP\AJ	Ok,M
34	AR1254CCC500	PP076274.D	31 Oct 2025 22:18	YP\AJ	Ok,M
35	I.BLK	PP076275.D	31 Oct 2025 22:34	YP\AJ	Ok
36	PB170358BL	PP076276.D	31 Oct 2025 22:51	YP\AJ	Ok
37	PB170358BS	PP076277.D	31 Oct 2025 23:07	YP\AJ	Ok,M
38	Q3437-15	PP076278.D	31 Oct 2025 23:23	YP\AJ	Ok,M
39	Q3450-04	PP076279.D	31 Oct 2025 23:40	YP\AJ	Ok,M
40	Q3450-04MS	PP076280.D	31 Oct 2025 23:56	YP\AJ	Ok,M
41	Q3450-04MSD	PP076281.D	01 Nov 2025 00:12	YP\AJ	Ok,M
42	Q3450-05	PP076282.D	01 Nov 2025 00:28	YP\AJ	Ok
43	Q3450-06	PP076283.D	01 Nov 2025 00:45	YP\AJ	Ok,M
44	Q3450-23	PP076284.D	01 Nov 2025 01:01	YP\AJ	Dilution

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QCBatch ID # PP103125

Review By	rahul	Review On	11/3/2025 8:18:24 AM		
Supervise By	yogesh	Supervise On	11/3/2025 8:49:19 AM		
SubDirectory	PP103125	HP Acquire Method		HP Processing Method	PP102125

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829 ,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP2484 4,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854
CCC	PP24824,PP24828,PP24832,PP24836,PP24840,PP24844,PP24852
Internal Standard/PEM	
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

45	Q3450-25	PP076285.D	01 Nov 2025 01:17	YP\AJ	Ok,M
46	AR1660CCC500	PP076286.D	01 Nov 2025 02:22	YP\AJ	Ok,M
47	AR1242CCC500	PP076287.D	01 Nov 2025 03:11	YP\AJ	Ok,M
48	AR1248CCC500	PP076288.D	01 Nov 2025 03:28	YP\AJ	Ok,M
49	AR1254CCC500	PP076289.D	01 Nov 2025 03:44	YP\AJ	Ok,M
50	I.BLK	PP076290.D	01 Nov 2025 04:00	YP\AJ	Ok

M : Manual Integration

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QCBatch ID # PP110325

Review By	rahul	Review On	11/4/2025 8:41:32 AM		
Supervise By	yogesh	Supervise On	11/4/2025 10:10:47 AM		
SubDirectory	PP110325	HP Acquire Method		HP Processing Method	PP102125

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829 ,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854
CCC	PP24824,PP24828,PP24832,PP24836,PP24840,PP24844,PP24852
Internal Standard/PEM	
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	HEXANE	PP076291.D	03 Nov 2025 08:28	YP\AJ	Ok
2	AR1660CCC500	PP076292.D	03 Nov 2025 08:44	YP\AJ	Ok
3	AR1242CCC500	PP076293.D	03 Nov 2025 09:04	YP\AJ	Ok,M
4	AR1248CCC500	PP076294.D	03 Nov 2025 09:21	YP\AJ	Ok,M
5	AR1254CCC500	PP076295.D	03 Nov 2025 09:37	YP\AJ	Ok,M
6	I.BLK	PP076296.D	03 Nov 2025 09:53	YP\AJ	Ok
7	Q3450-23DL	PP076297.D	03 Nov 2025 10:09	YP\AJ	Ok,M
8	PB170370BL	PP076298.D	03 Nov 2025 11:53	YP\AJ	Ok,M
9	PB170370BS	PP076299.D	03 Nov 2025 12:09	YP\AJ	Ok,M
10	Q3515-01	PP076300.D	03 Nov 2025 12:25	YP\AJ	Ok,M
11	Q3515-01MS	PP076301.D	03 Nov 2025 12:41	YP\AJ	Ok,M
12	Q3515-01MSD	PP076302.D	03 Nov 2025 12:58	YP\AJ	Ok,M
13	Q3516-01	PP076303.D	03 Nov 2025 13:14	YP\AJ	Ok,M
14	AR1660CCC500	PP076304.D	03 Nov 2025 14:19	YP\AJ	Ok
15	AR1242CCC500	PP076305.D	03 Nov 2025 14:35	YP\AJ	Ok,M
16	AR1248CCC500	PP076306.D	03 Nov 2025 14:51	YP\AJ	Ok,M
17	AR1254CCC500	PP076307.D	03 Nov 2025 15:08	YP\AJ	Ok,M
18	I.BLK	PP076308.D	03 Nov 2025 15:24	YP\AJ	Ok

M : Manual Integration

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: ECD_Q

Daily Analysis Runlog For Sequence/QCBatch ID # PQ101025

Review By	yogesh	Review On	10/10/2025 10:11:25 AM		
Supervise By	mohammad	Supervise On	10/14/2025 2:48:24 AM		
SubDirectory	PQ101025	HP Acquire Method		HP Processing Method	PQ101025

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854
CCC	
Internal Standard/PEM	
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	HEXANE	PQ070971.D	10 Oct 2025 08:27	YP\AJ	Ok
2	I.BLK	PQ070972.D	10 Oct 2025 08:42	YP\AJ	Ok
3	AR1660ICC1000	PQ070973.D	10 Oct 2025 08:57	YP\AJ	Ok
4	AR1660ICC750	PQ070974.D	10 Oct 2025 09:11	YP\AJ	Ok
5	AR1660ICC500	PQ070975.D	10 Oct 2025 09:26	YP\AJ	Ok
6	AR1660ICC250	PQ070976.D	10 Oct 2025 09:40	YP\AJ	Ok
7	AR1660ICC050	PQ070977.D	10 Oct 2025 09:55	YP\AJ	Ok,M
8	AR1221ICC500	PQ070978.D	10 Oct 2025 10:09	YP\AJ	Ok
9	AR1232ICC500	PQ070979.D	10 Oct 2025 10:24	YP\AJ	Ok
10	AR1242ICC1000	PQ070980.D	10 Oct 2025 10:38	YP\AJ	Ok
11	AR1242ICC750	PQ070981.D	10 Oct 2025 10:53	YP\AJ	Ok
12	AR1242ICC500	PQ070982.D	10 Oct 2025 11:08	YP\AJ	Ok
13	AR1242ICC250	PQ070983.D	10 Oct 2025 11:22	YP\AJ	Ok
14	AR1242ICC050	PQ070984.D	10 Oct 2025 11:37	YP\AJ	Ok,M
15	AR1248ICC1000	PQ070985.D	10 Oct 2025 11:51	YP\AJ	Ok
16	AR1248ICC750	PQ070986.D	10 Oct 2025 12:06	YP\AJ	Ok
17	AR1248ICC500	PQ070987.D	10 Oct 2025 12:20	YP\AJ	Ok
18	AR1248ICC250	PQ070988.D	10 Oct 2025 12:35	YP\AJ	Ok,M
19	AR1248ICC050	PQ070989.D	10 Oct 2025 12:50	YP\AJ	Ok
20	AR1254ICC1000	PQ070990.D	10 Oct 2025 13:04	YP\AJ	Ok
21	AR1254ICC750	PQ070991.D	10 Oct 2025 13:19	YP\AJ	Ok

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: ECD_Q

Daily Analysis Runlog For Sequence/QCBatch ID # PQ101025

Review By	yogesh	Review On	10/10/2025 10:11:25 AM		
Supervise By	mohammad	Supervise On	10/14/2025 2:48:24 AM		
SubDirectory	PQ101025	HP Acquire Method		HP Processing Method	PQ101025

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854
CCC	
Internal Standard/PEM	
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

22	AR1254ICC500	PQ070992.D	10 Oct 2025 13:33	YP\AJ	Ok
23	AR1254ICC250	PQ070993.D	10 Oct 2025 13:48	YP\AJ	Ok
24	AR1254ICC050	PQ070994.D	10 Oct 2025 15:15	YP\AJ	Ok
25	AR1262ICC500	PQ070995.D	10 Oct 2025 15:30	YP\AJ	Ok
26	AR1268ICC1000	PQ070996.D	10 Oct 2025 15:44	YP\AJ	Ok
27	AR1268ICC750	PQ070997.D	10 Oct 2025 15:59	YP\AJ	Ok
28	AR1268ICC500	PQ070998.D	10 Oct 2025 16:13	YP\AJ	Ok
29	AR1268ICC250	PQ070999.D	10 Oct 2025 16:28	YP\AJ	Ok
30	AR1268ICC050	PQ071000.D	10 Oct 2025 16:43	YP\AJ	Ok
31	PQ101025ICV500	PQ071001.D	10 Oct 2025 16:57	YP\AJ	Ok,M
32	AR1242ICV500	PQ071002.D	10 Oct 2025 17:41	YP\AJ	Ok,M
33	AR1248ICV500	PQ071003.D	10 Oct 2025 18:25	YP\AJ	Ok,M
34	AR1254ICV500	PQ071004.D	10 Oct 2025 19:08	YP\AJ	Ok
35	AR1268ICV500	PQ071005.D	10 Oct 2025 19:52	YP\AJ	Ok
36	DDT ANALOG	PQ071006.D	10 Oct 2025 20:36	YP\AJ	Ok

M : Manual Integration

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: ECD_Q

Daily Analysis Runlog For Sequence/QCBatch ID # PQ102725

Review By	yogesh	Review On	10/27/2025 10:54:57 AM		
Supervise By	mohammad	Supervise On	10/29/2025 5:27:51 AM		
SubDirectory	PQ102725	HP Acquire Method		HP Processing Method	PQ101025

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829 ,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP2484 4,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854
CCC	PP24824,PP24828,PP24832,PP24836,PP24840,PP24844,PP24852
Internal Standard/PEM	
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	HEXANE	PQ071148.D	27 Oct 2025 08:42	YP\AJ	Ok
2	AR1660CCC500	PQ071149.D	27 Oct 2025 08:57	YP\AJ	Ok
3	I.BLK	PQ071150.D	27 Oct 2025 09:11	YP\AJ	Ok
4	Q3450-01	PQ071151.D	27 Oct 2025 09:26	YP\AJ	Ok
5	Q3450-02	PQ071152.D	27 Oct 2025 09:41	YP\AJ	Ok
6	Q3450-02MS	PQ071153.D	27 Oct 2025 09:55	YP\AJ	Ok,M
7	Q3450-02MSD	PQ071154.D	27 Oct 2025 10:10	YP\AJ	Ok,M
8	Q3450-08	PQ071155.D	27 Oct 2025 10:24	YP\AJ	Ok
9	Q3450-09	PQ071156.D	27 Oct 2025 10:39	YP\AJ	Ok,M
10	Q3437-16	PQ071157.D	27 Oct 2025 10:54	YP\AJ	Ok
11	Q3419-01	PQ071158.D	27 Oct 2025 11:08	YP\AJ	Ok,M
12	Q3419-02MS	PQ071159.D	27 Oct 2025 11:23	YP\AJ	Ok
13	Q3419-03MSD	PQ071160.D	27 Oct 2025 11:37	YP\AJ	Ok
14	AR1660CCC500	PQ071161.D	27 Oct 2025 12:35	YP\AJ	Ok
15	AR1242CCC500	PQ071162.D	27 Oct 2025 13:34	YP\AJ	Ok
16	AR1248CCC500	PQ071163.D	27 Oct 2025 13:49	YP\AJ	Ok
17	AR1254CCC500	PQ071164.D	27 Oct 2025 14:03	YP\AJ	Ok
18	I.BLK	PQ071165.D	27 Oct 2025 14:18	YP\AJ	Ok
19	PB170255BL	PQ071166.D	27 Oct 2025 14:32	YP\AJ	Not Ok
20	PB170255BS	PQ071167.D	27 Oct 2025 14:47	YP\AJ	Not Ok
21	Q3451-01	PQ071168.D	27 Oct 2025 15:01	YP\AJ	Ok

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: ECD_Q

Daily Analysis Runlog For Sequence/QCBatch ID # PQ102725

Review By	yogesh	Review On	10/27/2025 10:54:57 AM		
Supervise By	mohammad	Supervise On	10/29/2025 5:27:51 AM		
SubDirectory	PQ102725	HP Acquire Method		HP Processing Method	PQ101025

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829 ,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854
CCC	PP24824,PP24828,PP24832,PP24836,PP24840,PP24844,PP24852
Internal Standard/PEM	
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

22	Q3451-05	PQ071169.D	27 Oct 2025 15:16	YP\AJ	Ok
23	Q3451-05MS	PQ071170.D	27 Oct 2025 15:31	YP\AJ	Ok
24	Q3451-05MSD	PQ071171.D	27 Oct 2025 15:45	YP\AJ	Ok
25	Q3452-01	PQ071172.D	27 Oct 2025 16:00	YP\AJ	Ok
26	Q3460-05	PQ071173.D	27 Oct 2025 16:14	YP\AJ	ReRun
27	Q3460-09	PQ071174.D	27 Oct 2025 16:29	YP\AJ	ReRun
28	Q3460-13	PQ071175.D	27 Oct 2025 16:43	YP\AJ	Ok,M
29	AR1660CCC500	PQ071176.D	27 Oct 2025 17:42	YP\AJ	Ok,M
30	AR1242CCC500	PQ071177.D	27 Oct 2025 18:40	YP\AJ	Ok
31	AR1248CCC500	PQ071178.D	27 Oct 2025 18:55	YP\AJ	Ok
32	AR1254CCC500	PQ071179.D	27 Oct 2025 19:09	YP\AJ	Ok,M
33	I.BLK	PQ071180.D	27 Oct 2025 19:24	YP\AJ	Ok

M : Manual Integration

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QCBatch ID # PO110425

Review By	yogesh	Review On	11/4/2025 12:15:43 PM		
Supervise By	rahul	Supervise On	11/5/2025 8:40:26 AM		
SubDirectory	PO110425	HP Acquire Method		HP Processing Method	PO110425

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854
CCC	
Internal Standard/PEM	
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PO114869.D	04 Nov 2025 09:03		YP/AJ	Ok
2	I.BLK	I.BLK	PO114870.D	04 Nov 2025 09:21		YP/AJ	Ok
3	AR1660ICC1000	AR1660ICC1000	PO114871.D	04 Nov 2025 09:39		YP/AJ	Ok
4	AR1660ICC750	AR1660ICC750	PO114872.D	04 Nov 2025 09:58		YP/AJ	Ok
5	AR1660ICC500	AR1660ICC500	PO114873.D	04 Nov 2025 10:16		YP/AJ	Ok
6	AR1660ICC250	AR1660ICC250	PO114874.D	04 Nov 2025 10:34		YP/AJ	Ok,M
7	AR1660ICC050	AR1660ICC050	PO114875.D	04 Nov 2025 10:53		YP/AJ	Ok,M
8	AR1221ICC500	AR1221ICC500	PO114876.D	04 Nov 2025 11:11		YP/AJ	Ok
9	AR1232ICC500	AR1232ICC500	PO114877.D	04 Nov 2025 11:30		YP/AJ	Ok
10	AR1242ICC1000	AR1242ICC1000	PO114878.D	04 Nov 2025 11:48		YP/AJ	Ok
11	AR1242ICC750	AR1242ICC750	PO114879.D	04 Nov 2025 12:06		YP/AJ	Ok
12	AR1242ICC500	AR1242ICC500	PO114880.D	04 Nov 2025 12:25		YP/AJ	Ok
13	AR1242ICC250	AR1242ICC250	PO114881.D	04 Nov 2025 12:42		YP/AJ	Ok
14	AR1242ICC050	AR1242ICC050	PO114882.D	04 Nov 2025 13:00		YP/AJ	Ok,M
15	AR1248ICC1000	AR1248ICC1000	PO114883.D	04 Nov 2025 13:19		YP/AJ	Ok
16	AR1248ICC750	AR1248ICC750	PO114884.D	04 Nov 2025 13:37		YP/AJ	Ok
17	AR1248ICC500	AR1248ICC500	PO114885.D	04 Nov 2025 13:56		YP/AJ	Ok
18	AR1248ICC250	AR1248ICC250	PO114886.D	04 Nov 2025 14:14		YP/AJ	Ok

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QCBatch ID # PO110425

Review By	yogesh	Review On	11/4/2025 12:15:43 PM		
Supervise By	rahul	Supervise On	11/5/2025 8:40:26 AM		
SubDirectory	PO110425	HP Acquire Method		HP Processing Method	PO110425

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854
CCC	
Internal Standard/PEM	
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

19	AR1248ICC050	AR1248ICC050	PO114887.D	04 Nov 2025 14:33		YP/AJ	Ok
20	AR1254ICC1000	AR1254ICC1000	PO114888.D	04 Nov 2025 14:51		YP/AJ	Ok
21	AR1254ICC750	AR1254ICC750	PO114889.D	04 Nov 2025 15:09		YP/AJ	Ok
22	AR1254ICC500	AR1254ICC500	PO114890.D	04 Nov 2025 15:28		YP/AJ	Ok
23	AR1254ICC250	AR1254ICC250	PO114891.D	04 Nov 2025 16:04		YP/AJ	Ok
24	AR1254ICC050	AR1254ICC050	PO114892.D	04 Nov 2025 16:23		YP/AJ	Ok,M
25	AR1262ICC500	AR1262ICC500	PO114893.D	04 Nov 2025 17:00		YP/AJ	Ok
26	AR1268ICC1000	AR1268ICC1000	PO114894.D	04 Nov 2025 17:18		YP/AJ	Ok
27	AR1268ICC750	AR1268ICC750	PO114895.D	04 Nov 2025 17:36		YP/AJ	Ok
28	AR1268ICC500	AR1268ICC500	PO114896.D	04 Nov 2025 17:55		YP/AJ	Ok
29	AR1268ICC250	AR1268ICC250	PO114897.D	04 Nov 2025 18:13		YP/AJ	Ok,M
30	AR1268ICC050	AR1268ICC050	PO114898.D	04 Nov 2025 18:31		YP/AJ	Ok
31	PO110425ICV500	ICVPO110425	PO114899.D	04 Nov 2025 19:08		YP/AJ	Ok
32	AR1242ICV500	ICVPO110425AR1242	PO114900.D	04 Nov 2025 20:21		YP/AJ	Ok,M
33	AR1248ICV500	ICVPO110425AR1248	PO114901.D	04 Nov 2025 20:58		YP/AJ	Ok
34	AR1254ICV500	ICVPO110425AR1254	PO114902.D	04 Nov 2025 21:35		YP/AJ	Ok
35	AR1268ICV500	ICVPO110425AR1268	PO114903.D	04 Nov 2025 22:12		YP/AJ	Ok
36	DDT ANALOG	DDT ANALOG	PO114904.D	04 Nov 2025 22:48		YP/AJ	Ok

M : Manual Integration

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QCBatch ID # PO111225

Review By	rahul	Review On	11/13/2025 1:15:59 PM		
Supervise By	yogesh	Supervise On	11/13/2025 1:22:24 PM		
SubDirectory	PO111225	HP Acquire Method		HP Processing Method	PP102125

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854
CCC	PP24824,PP24828,PP24832,PP24836,PP24840,PP24844,PP24852
Internal Standard/PEM	
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PO115030.D	12 Nov 2025 09:31		YP/AJ	Ok
2	AR1660CCC500	AR1660CCC500	PO115031.D	12 Nov 2025 09:49		YP/AJ	Ok
3	AR1242CCC500	AR1242CCC500	PO115032.D	12 Nov 2025 10:08		YP/AJ	Ok
4	AR1248CCC500	AR1248CCC500	PO115033.D	12 Nov 2025 10:27		YP/AJ	Ok
5	AR1254CCC500	AR1254CCC500	PO115034.D	12 Nov 2025 10:45		YP/AJ	Ok
6	I.BLK	I.BLK	PO115035.D	12 Nov 2025 11:03		YP/AJ	Ok
7	Q3561-01DL	1027-A-DDL	PO115036.D	12 Nov 2025 11:22	TCMX high in 2nd column , AR1248 hit	YP/AJ	Ok,M
8	PB170497BL	PB170497BL	PO115037.D	12 Nov 2025 11:58		YP/AJ	Ok
9	PB170497BS	PB170497BS	PO115038.D	12 Nov 2025 12:16		YP/AJ	Ok
10	Q3597-01	TP-1	PO115039.D	12 Nov 2025 12:34		YP/AJ	Ok
11	Q3597-01MS	TP-1MS	PO115040.D	12 Nov 2025 12:53		YP/AJ	Ok
12	Q3597-01MSD	TP-1MSD	PO115041.D	12 Nov 2025 13:11		YP/AJ	Ok
13	Q3597-04	TP-2	PO115042.D	12 Nov 2025 13:29		YP/AJ	Ok
14	Q3597-07	TP-3	PO115043.D	12 Nov 2025 13:48		YP/AJ	Ok
15	Q3597-10	TP-4	PO115044.D	12 Nov 2025 14:06		YP/AJ	Ok
16	Q3450-07	SS-13R-(4.5-5.0)	PO115045.D	12 Nov 2025 15:36	AR1254 Hit	YP/AJ	Ok
17	AR1660CCC500	AR1660CCC500	PO115046.D	12 Nov 2025 15:55		YP/AJ	Ok,NR

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QCBatch ID # PO111225

Review By	rahul	Review On	11/13/2025 1:15:59 PM		
Supervise By	yogesh	Supervise On	11/13/2025 1:22:24 PM		
SubDirectory	PO111225	HP Acquire Method		HP Processing Method	PP102125

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854
CCC	PP24824,PP24828,PP24832,PP24836,PP24840,PP24844,PP24852
Internal Standard/PEM	
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

18	AR1242CCC500	AR1242CCC500	PO115047.D	12 Nov 2025 16:13		YP/AJ	Ok
19	AR1248CCC500	AR1248CCC500	PO115048.D	12 Nov 2025 16:31		YP/AJ	Ok
20	AR1254CCC500	AR1254CCC500	PO115049.D	12 Nov 2025 16:50		YP/AJ	Ok
21	I.BLK	I.BLK	PO115050.D	12 Nov 2025 17:07		YP/AJ	Ok
22	PB170499BL	PB170499BL	PO115051.D	12 Nov 2025 17:26		YP/AJ	Ok
23	PB170499BS	PB170499BS	PO115052.D	12 Nov 2025 17:44		YP/AJ	Ok
24	Q3591-01	KMH4680-END-1-1	PO115053.D	12 Nov 2025 18:02		YP/AJ	Ok,M
25	Q3591-02	KMH4680-END-1-2	PO115054.D	12 Nov 2025 18:21		YP/AJ	Ok
26	Q3591-03	M4-END-A	PO115055.D	12 Nov 2025 18:39	AR1254 Hit	YP/AJ	Ok,M
27	Q3591-04	M4-END-B	PO115056.D	12 Nov 2025 18:58	AR1254 Hit	YP/AJ	Ok,M
28	Q3591-05	M5-END-A	PO115057.D	12 Nov 2025 19:16	AR1254 Hit	YP/AJ	Ok,M
29	Q3591-06	M5-END-B	PO115058.D	12 Nov 2025 19:34	AR1254 Hit	YP/AJ	Ok,M
30	Q3596-01	11-7-25-A	PO115059.D	12 Nov 2025 19:53	AR1242+1254 Hit	YP/AJ	Ok,M
31	Q3596-02	11-7-25-B	PO115060.D	12 Nov 2025 20:11	AR1242 Hit	YP/AJ	Ok,M
32	AR1660CCC500	AR1660CCC500	PO115061.D	12 Nov 2025 21:25		YP/AJ	Ok
33	AR1242CCC500	AR1242CCC500	PO115062.D	12 Nov 2025 22:19		YP/AJ	Ok
34	AR1248CCC500	AR1248CCC500	PO115063.D	12 Nov 2025 22:38		YP/AJ	Ok
35	AR1254CCC500	AR1254CCC500	PO115064.D	12 Nov 2025 22:56		YP/AJ	Ok
36	I.BLK	I.BLK	PO115065.D	12 Nov 2025 23:13		YP/AJ	Ok

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QCBatch ID # PO111225

Review By	rahul	Review On	11/13/2025 1:15:59 PM		
Supervise By	yogesh	Supervise On	11/13/2025 1:22:24 PM		
SubDirectory	PO111225	HP Acquire Method		HP Processing Method	PP102125

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854
CCC	PP24824,PP24828,PP24832,PP24836,PP24840,PP24844,PP24852
Internal Standard/PEM	
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

M : Manual Integration

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QCBatch ID # PP102125

Review By	yogesh	Review On	10/21/2025 12:01:15 PM		
Supervise By	mohammad	Supervise On	10/25/2025 7:28:37 AM		
SubDirectory	PP102125	HP Acquire Method		HP Processing Method	PP102125

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854
CCC	
Internal Standard/PEM	
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PP075886.D	21 Oct 2025 08:33		YP\AJ	Ok
2	I.BLK	I.BLK	PP075887.D	21 Oct 2025 08:49		YP\AJ	Ok,M
3	AR1660ICC1000	AR1660ICC1000	PP075888.D	21 Oct 2025 09:06		YP\AJ	Ok
4	AR1660ICC750	AR1660ICC750	PP075889.D	21 Oct 2025 09:22		YP\AJ	Ok
5	AR1660ICC500	AR1660ICC500	PP075890.D	21 Oct 2025 09:38		YP\AJ	Ok
6	AR1660ICC250	AR1660ICC250	PP075891.D	21 Oct 2025 09:54		YP\AJ	Ok
7	AR1660ICC050	AR1660ICC050	PP075892.D	21 Oct 2025 10:11		YP\AJ	Ok,M
8	AR1221ICC500	AR1221ICC500	PP075893.D	21 Oct 2025 10:27		YP\AJ	Ok
9	AR1232ICC500	AR1232ICC500	PP075894.D	21 Oct 2025 10:43		YP\AJ	Ok,M
10	AR1242ICC1000	AR1242ICC1000	PP075895.D	21 Oct 2025 10:59		YP\AJ	Ok,M
11	AR1242ICC750	AR1242ICC750	PP075896.D	21 Oct 2025 11:16		YP\AJ	Ok,M
12	AR1242ICC500	AR1242ICC500	PP075897.D	21 Oct 2025 11:48		YP\AJ	Ok,M
13	AR1242ICC250	AR1242ICC250	PP075898.D	21 Oct 2025 12:04		YP\AJ	Ok,M
14	AR1242ICC050	AR1242ICC050	PP075899.D	21 Oct 2025 12:21		YP\AJ	Ok,M
15	AR1248ICC1000	AR1248ICC1000	PP075900.D	21 Oct 2025 12:41		YP\AJ	Ok,M
16	AR1248ICC750	AR1248ICC750	PP075901.D	21 Oct 2025 12:57		YP\AJ	Ok,M
17	AR1248ICC500	AR1248ICC500	PP075902.D	21 Oct 2025 13:13		YP\AJ	Ok,M
18	AR1248ICC250	AR1248ICC250	PP075903.D	21 Oct 2025 13:30		YP\AJ	Ok,M

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QCBatch ID # PP102125

Review By	yogesh	Review On	10/21/2025 12:01:15 PM		
Supervise By	mohammad	Supervise On	10/25/2025 7:28:37 AM		
SubDirectory	PP102125	HP Acquire Method		HP Processing Method	PP102125

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854
CCC	
Internal Standard/PEM	
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

19	AR1248ICC050	AR1248ICC050	PP075904.D	21 Oct 2025 13:46		YP\AJ	Ok,M
20	AR1254ICC1000	AR1254ICC1000	PP075905.D	21 Oct 2025 14:02		YP\AJ	Ok,M
21	AR1254ICC750	AR1254ICC750	PP075906.D	21 Oct 2025 14:18		YP\AJ	Ok,M
22	AR1254ICC500	AR1254ICC500	PP075907.D	21 Oct 2025 14:35		YP\AJ	Ok,M
23	AR1254ICC250	AR1254ICC250	PP075908.D	21 Oct 2025 14:51		YP\AJ	Ok,M
24	AR1254ICC050	AR1254ICC050	PP075909.D	21 Oct 2025 15:07		YP\AJ	Ok,M
25	AR1262ICC500	AR1262ICC500	PP075910.D	21 Oct 2025 15:40		YP\AJ	Ok
26	AR1268ICC1000	AR1268ICC1000	PP075911.D	21 Oct 2025 15:56		YP\AJ	Ok
27	AR1268ICC750	AR1268ICC750	PP075912.D	21 Oct 2025 16:12		YP\AJ	Ok
28	AR1268ICC500	AR1268ICC500	PP075913.D	21 Oct 2025 16:29		YP\AJ	Ok
29	AR1268ICC250	AR1268ICC250	PP075914.D	21 Oct 2025 16:45		YP\AJ	Ok
30	AR1268ICC050	AR1268ICC050	PP075915.D	21 Oct 2025 17:01		YP\AJ	Ok
31	PP102125ICV500	ICVPP102125	PP075916.D	21 Oct 2025 17:34		YP\AJ	Ok
32	AR1242ICV500	ICVPP102125AR1242	PP075917.D	21 Oct 2025 18:06		YP\AJ	Ok,M
33	AR1248ICV500	ICVPP102125AR1248	PP075918.D	21 Oct 2025 18:39		YP\AJ	Ok,M
34	AR1254ICV500	ICVPP102125AR1254	PP075919.D	22 Oct 2025 08:49		YP\AJ	Ok,M
35	AR1268ICV500	ICVPP102125AR1268	PP075920.D	22 Oct 2025 09:07		YP\AJ	Ok
36	DDT ANALOG	DDT ANALOG	PP075921.D	22 Oct 2025 09:25		YP\AJ	Ok

M : Manual Integration

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QCBatch ID # PP102425

Review By	yogesh	Review On	10/24/2025 11:09:51 AM		
Supervise By	mohammad	Supervise On	10/28/2025 6:10:31 AM		
SubDirectory	PP102425	HP Acquire Method		HP Processing Method	PP102125

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854
CCC	PP24824,PP24828,PP24832,PP24836,PP24840,PP24844,PP24852
Internal Standard/PEM	
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PP076011.D	24 Oct 2025 08:41		YP\AJ	Ok
2	AR1660CCC500	AR1660CCC500	PP076012.D	24 Oct 2025 08:57		YP\AJ	Ok
3	AR1242CCC500	AR1242CCC500	PP076013.D	24 Oct 2025 09:29		YP\AJ	Ok,M
4	AR1248CCC500	AR1248CCC500	PP076014.D	24 Oct 2025 09:45		YP\AJ	Ok,M
5	AR1254CCC500	AR1254CCC500	PP076015.D	24 Oct 2025 10:02		YP\AJ	Ok,M
6	I.BLK	I.BLK	PP076016.D	24 Oct 2025 10:18		YP\AJ	Ok
7	PB170234BL	PB170234BL	PP076017.D	24 Oct 2025 10:34		YP\AJ	Ok
8	PB170234BS	PB170234BS	PP076018.D	24 Oct 2025 10:51		YP\AJ	Ok
9	Q3437-03	SS-15R(2.0-2.5)	PP076019.D	24 Oct 2025 11:07	AR1254 hits	YP\AJ	Ok
10	Q3437-05	SS-26(0-0.5)	PP076020.D	24 Oct 2025 11:23	AR1254 hits	YP\AJ	Ok,M
11	Q3437-11	SS-27(0-0.5)	PP076021.D	24 Oct 2025 11:39	AR1254 hits	YP\AJ	Ok,M
12	Q3445-01	EO-01-102325	PP076022.D	24 Oct 2025 11:56		YP\AJ	Ok
13	Q3446-01	SB-14	PP076023.D	24 Oct 2025 12:12	F Flag in TCMX in 1st column, All surrogate fail, AR1242 hits, Need 20X	YP\AJ	Dilution
14	Q3446-02	SB-15	PP076024.D	24 Oct 2025 12:28	AR1242 hits, Need 20X	YP\AJ	Dilution
15	Q3414-12DL2	SB-12DL2	PP076025.D	24 Oct 2025 12:44	AR1242 hit	YP\AJ	Ok
16	Q3414-13DL2	SB-13DL2	PP076026.D	24 Oct 2025 13:01	AR1242 hit	YP\AJ	Ok
17	AR1660CCC500	AR1660CCC500	PP076027.D	24 Oct 2025 14:06		YP\AJ	Ok

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QCBatch ID # PP102425

Review By	yogesh	Review On	10/24/2025 11:09:51 AM		
Supervise By	mohammad	Supervise On	10/28/2025 6:10:31 AM		
SubDirectory	PP102425	HP Acquire Method		HP Processing Method	PP102125

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854
CCC	PP24824,PP24828,PP24832,PP24836,PP24840,PP24844,PP24852
Internal Standard/PEM	
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

18	AR1242CCC500	AR1242CCC500	PP076028.D	24 Oct 2025 14:22		YP\AJ	Ok,M
19	AR1248CCC500	AR1248CCC500	PP076029.D	24 Oct 2025 14:38		YP\AJ	Ok,M
20	AR1254CCC500	AR1254CCC500	PP076030.D	24 Oct 2025 14:54		YP\AJ	Ok,M
21	I.BLK	I.BLK	PP076031.D	24 Oct 2025 15:11		YP\AJ	Ok
22	Q3414-07DL2	SB-7DL2	PP076032.D	24 Oct 2025 15:27	AR1242 hit	YP\AJ	Ok
23	Q3446-06	SB-19	PP076033.D	24 Oct 2025 15:43	F Flag in TCMX in 1st column , AR1242+1260 Hit , TCMX high in both column , DCB low in 1st column , Need dilution	YP\AJ	Dilution
24	Q3446-07	SB-20	PP076034.D	24 Oct 2025 16:00	AR1260 Hit , Need dilution , TCMX high in 2nd column	YP\AJ	Dilution
25	Q3446-08	SB-21	PP076035.D	24 Oct 2025 16:16	AR1242+1260 Hit , DCB low in both column , TCMX high in both column , need dilution	YP\AJ	Dilution
26	Q3446-03	SB-16	PP076036.D	24 Oct 2025 16:32	AR1260 Hit , TCMX high in 2nd column , need dilution	YP\AJ	Dilution
27	Q3446-04	SB-17	PP076037.D	24 Oct 2025 16:48	AR1260 Hit , TCMX high in 2nd column , need dilution	YP\AJ	Dilution
28	Q3446-01DL	SB-14DL	PP076038.D	24 Oct 2025 17:05	AR1242 hit	YP\AJ	Ok,M
29	Q3446-02DL	SB-15DL	PP076039.D	24 Oct 2025 17:21	AR1242 hit	YP\AJ	Ok,M
30	Q3446-05DL	SB-18DL	PP076040.D	24 Oct 2025 17:37	AR1242 Hit	YP\AJ	Ok,M
31	Q3446-06DL	SB-19DL	PP076041.D	24 Oct 2025 17:53	AR1242+1260 Hit	YP\AJ	Ok,M
32	AR1660CCC500	AR1660CCC500	PP076042.D	24 Oct 2025 18:58		YP\AJ	Ok
33	AR1242CCC500	AR1242CCC500	PP076043.D	24 Oct 2025 19:31		YP\AJ	Ok,M
34	AR1248CCC500	AR1248CCC500	PP076044.D	24 Oct 2025 20:03		YP\AJ	Ok,M

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QCBatch ID # PP102425

Review By	yogesh	Review On	10/24/2025 11:09:51 AM		
Supervise By	mohammad	Supervise On	10/28/2025 6:10:31 AM		
SubDirectory	PP102425	HP Acquire Method		HP Processing Method	PP102125

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854
CCC	PP24824,PP24828,PP24832,PP24836,PP24840,PP24844,PP24852
Internal Standard/PEM	
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

35	AR1254CCC500	AR1254CCC500	PP076045.D	24 Oct 2025 20:19		YP\AJ	Ok,M
36	I.BLK	I.BLK	PP076046.D	24 Oct 2025 20:36		YP\AJ	Ok
37	Q3450-10	SS-20-(3.5-4.0)	PP076047.D	24 Oct 2025 20:52	AR1254 Hit	YP\AJ	Ok,M
38	Q3450-11	SS-28-(2.0-2.5)	PP076048.D	24 Oct 2025 21:08		YP\AJ	Ok
39	Q3450-20	SS-21-(0.0-0.5)	PP076049.D	24 Oct 2025 21:24	AR1254 Hit , Need 5x dilution	YP\AJ	Dilution
40	Q3450-21	SS-21-(2.0-2.5)	PP076050.D	24 Oct 2025 21:41		YP\AJ	Ok
41	Q3450-22	SS-21-(3.5-4.0)	PP076051.D	24 Oct 2025 21:57	AR1254 Hit , Need 2x dilution	YP\AJ	Dilution
42	Q3427-01	TP-8	PP076052.D	24 Oct 2025 22:13		YP\AJ	Ok,M
43	Q3446-07DL	SB-20DL	PP076053.D	24 Oct 2025 22:29	AR1242 Hit	YP\AJ	Ok,M
44	Q3446-08DL	SB-21DL	PP076054.D	24 Oct 2025 22:45	AR1242+1260 Hit	YP\AJ	Ok,M
45	Q3446-03DL	SB-16DL	PP076055.D	24 Oct 2025 23:01	AR1260 Hit	YP\AJ	Ok,M
46	Q3446-04DL	SB-17DL	PP076056.D	24 Oct 2025 23:17	AR1260 Hit	YP\AJ	Ok,M
47	AR1660CCC500	AR1660CCC500	PP076057.D	25 Oct 2025 00:22		YP\AJ	Ok
48	AR1242CCC500	AR1242CCC500	PP076058.D	25 Oct 2025 01:11		YP\AJ	Ok,M
49	AR1248CCC500	AR1248CCC500	PP076059.D	25 Oct 2025 01:43		YP\AJ	Ok,M
50	AR1254CCC500	AR1254CCC500	PP076060.D	25 Oct 2025 01:59		YP\AJ	Ok,M
51	I.BLK	I.BLK	PP076061.D	25 Oct 2025 02:16		YP\AJ	Ok

M : Manual Integration

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QCBatch ID # PP102725

Review By	yogesh	Review On	10/27/2025 10:53:48 AM		
Supervise By	mohammad	Supervise On	10/29/2025 5:27:46 AM		
SubDirectory	PP102725	HP Acquire Method		HP Processing Method	PP102125

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854
CCC	PP24824,PP24828,PP24832,PP24836,PP24840,PP24844,PP24852
Internal Standard/PEM	
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PP076062.D	27 Oct 2025 08:40		YP\AJ	Ok
2	AR1660CCC500	AR1660CCC500	PP076063.D	27 Oct 2025 08:56		YP\AJ	Ok,M
3	AR1242CCC500	AR1242CCC500	PP076064.D	27 Oct 2025 09:13		YP\AJ	Ok,M
4	AR1248CCC500	AR1248CCC500	PP076065.D	27 Oct 2025 09:29		YP\AJ	Ok,M
5	AR1254CCC500	AR1254CCC500	PP076066.D	27 Oct 2025 09:45		YP\AJ	Ok,M
6	I.BLK	I.BLK	PP076067.D	27 Oct 2025 10:02		YP\AJ	Ok
7	PB170236BL	PB170236BL	PP076068.D	27 Oct 2025 10:18		YP\AJ	Ok,M
8	PB170236BS	PB170236BS	PP076069.D	27 Oct 2025 10:34		YP\AJ	Ok,M
9	Q3439-01	TP-9	PP076070.D	27 Oct 2025 11:06	DCB,TCX high 1st column	YP\AJ	ReRun
10	Q3449-01	CF-620-COMP-41	PP076071.D	27 Oct 2025 11:26		YP\AJ	Ok
11	Q3450-20DL	SS-21-(0.0-0.5)DL	PP076072.D	27 Oct 2025 11:42	AR1254 Hit	YP\AJ	Ok,M
12	Q3450-22DL	SS-21-(3.5-4.0)DL	PP076073.D	27 Oct 2025 11:58	AR1254 Hit	YP\AJ	Ok,M
13	Q3446-05	SB-18	PP076074.D	27 Oct 2025 12:14	AR1242 Hit, Need Dilution, f flag	YP\AJ	Not Ok
14	Q3447-03	P001-TW2-251023-01	PP076075.D	27 Oct 2025 12:30	F Flag in TCMX in 2nd column , TCMX high in both column , AR1254+1260 Hit	YP\AJ	ReRun
15	Q3447-07	P001-TW4-251023-01	PP076076.D	27 Oct 2025 12:47	TCMX high in 2nd column	YP\AJ	Ok,M
16	Q3446-05	SB-18	PP076077.D	27 Oct 2025 13:03	TCMX high in 2nd column , AR1242 hit , need 100x dilution	YP\AJ	Dilution
17	AR1660CCC500	AR1660CCC500	PP076078.D	27 Oct 2025 14:08		YP\AJ	Ok

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QCBatch ID # PP102725

Review By	yogesh	Review On	10/27/2025 10:53:48 AM		
Supervise By	mohammad	Supervise On	10/29/2025 5:27:46 AM		
SubDirectory	PP102725	HP Acquire Method		HP Processing Method	PP102125

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854
CCC	PP24824,PP24828,PP24832,PP24836,PP24840,PP24844,PP24852
Internal Standard/PEM	
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

18	AR1242CCC500	AR1242CCC500	PP076079.D	27 Oct 2025 14:24		YP\AJ	Ok,M
19	AR1248CCC500	AR1248CCC500	PP076080.D	27 Oct 2025 14:40		YP\AJ	Ok,M
20	AR1254CCC500	AR1254CCC500	PP076081.D	27 Oct 2025 14:57		YP\AJ	Ok,M
21	I.BLK	I.BLK	PP076082.D	27 Oct 2025 15:13		YP\AJ	Ok
22	Q3456-01	OILY-DEBRIS	PP076083.D	27 Oct 2025 15:29		YP\AJ	Ok,M
23	Q3457-01	TR-05-102425	PP076084.D	27 Oct 2025 15:45		YP\AJ	Ok,M
24	Q3460-01	WC-1	PP076085.D	27 Oct 2025 16:02		YP\AJ	Ok
25	Q3460-17	WC-5	PP076086.D	27 Oct 2025 16:18		YP\AJ	Ok,M
26	Q3447-05	P001-TW3-251023-01	PP076087.D	27 Oct 2025 16:34	TCMX high in 2nd column	YP\AJ	Ok,M
27	Q3447-01	P001-TW1-251023-01	PP076088.D	27 Oct 2025 16:50	F Flag in TCMX in 2nd column , TCMX high in both column	YP\AJ	ReRun
28	Q3459-01	Y2404-0071-END-1-1	PP076089.D	27 Oct 2025 17:07	AR1254 Hit	YP\AJ	Ok,M
29	Q3459-02	Y2404-0071-END-1-2	PP076090.D	27 Oct 2025 17:23		YP\AJ	Ok,M
30	Q3459-03	Y2404-0071-END-2-1	PP076091.D	27 Oct 2025 17:39	AR1254 Hit	YP\AJ	Ok,M
31	Q3459-04	Y2404-0071-END-2-2	PP076092.D	27 Oct 2025 17:55		YP\AJ	Ok
32	AR1660CCC500	AR1660CCC500	PP076093.D	27 Oct 2025 19:00		YP\AJ	Ok
33	AR1242CCC500	AR1242CCC500	PP076094.D	27 Oct 2025 19:49		YP\AJ	Ok,M
34	AR1248CCC500	AR1248CCC500	PP076095.D	27 Oct 2025 20:05		YP\AJ	Ok,M
35	AR1254CCC500	AR1254CCC500	PP076096.D	27 Oct 2025 20:21		YP\AJ	Ok,M
36	I.BLK	I.BLK	PP076097.D	27 Oct 2025 20:38		YP\AJ	Ok

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QCBatch ID # PP102725

Review By	yogesh	Review On	10/27/2025 10:53:48 AM		
Supervise By	mohammad	Supervise On	10/29/2025 5:27:46 AM		
SubDirectory	PP102725	HP Acquire Method		HP Processing Method	PP102125

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854
CCC	PP24824,PP24828,PP24832,PP24836,PP24840,PP24844,PP24852
Internal Standard/PEM	
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

37	PB170272BL	PB170272BL	PP076098.D	27 Oct 2025 20:54		YP\AJ	Ok
38	PB170272BS	PB170272BS	PP076099.D	27 Oct 2025 21:10		YP\AJ	Ok,M
39	Q3459-05	BC259240LOK-END-1-	PP076100.D	27 Oct 2025 21:26		YP\AJ	Ok,M
40	Q3459-06	BC259240LOK-END-1-	PP076101.D	27 Oct 2025 21:42		YP\AJ	Ok,M
41	Q3459-07	BC271327-END-1-1	PP076102.D	27 Oct 2025 21:59		YP\AJ	Ok,M
42	Q3459-08	BC271327-END-1-2	PP076103.D	27 Oct 2025 22:15		YP\AJ	Ok,M
43	AR1660CCC500	AR1660CCC500	PP076104.D	27 Oct 2025 23:20		YP\AJ	Ok
44	AR1242CCC500	AR1242CCC500	PP076105.D	28 Oct 2025 00:08		YP\AJ	Ok,M
45	AR1248CCC500	AR1248CCC500	PP076106.D	28 Oct 2025 00:25		YP\AJ	Ok,M
46	AR1254CCC500	AR1254CCC500	PP076107.D	28 Oct 2025 00:41		YP\AJ	Ok
47	I.BLK	I.BLK	PP076108.D	28 Oct 2025 00:57		YP\AJ	Ok

M : Manual Integration

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QCBatch ID # PP103125

Review By	rahul	Review On	11/3/2025 8:18:24 AM		
Supervise By	yogesh	Supervise On	11/3/2025 8:49:19 AM		
SubDirectory	PP103125	HP Acquire Method		HP Processing Method	PP102125

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854
CCC	PP24824,PP24828,PP24832,PP24836,PP24840,PP24844,PP24852
Internal Standard/PEM	
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PP076241.D	31 Oct 2025 09:26		YP\AJ	Ok
2	AR1660CCC500	AR1660CCC500	PP076242.D	31 Oct 2025 09:43	AR1260-04 high in second column	YP\AJ	Ok
3	AR1242CCC500	AR1242CCC500	PP076243.D	31 Oct 2025 09:59		YP\AJ	Ok,M
4	AR1248CCC500	AR1248CCC500	PP076244.D	31 Oct 2025 10:15		YP\AJ	Ok,M
5	AR1254CCC500	AR1254CCC500	PP076245.D	31 Oct 2025 10:32		YP\AJ	Ok,M
6	I.BLK	I.BLK	PP076246.D	31 Oct 2025 10:48		YP\AJ	Ok
7	Q3482-04	WC-4	PP076247.D	31 Oct 2025 11:09		YP\AJ	Ok,M
8	Q3500-01	459	PP076248.D	31 Oct 2025 13:21	TCMX high in 2nd column ,	YP\AJ	Ok,M
9	Q3501-01	461	PP076249.D	31 Oct 2025 13:38	DCB low in 2nd and high in 1st column , TCMX low in 2nd column	YP\AJ	Not Ok
10	Q3502-01	CATCH-CAN-USED-OI	PP076250.D	31 Oct 2025 13:54		YP\AJ	Ok,M
11	Q3503-01	322	PP076251.D	31 Oct 2025 14:10		YP\AJ	Ok,M
12	Q3503-03	323	PP076252.D	31 Oct 2025 14:26	DCB Low in 2nd column , AR1260 Hit	YP\AJ	Ok,M
13	Q3503-05	324	PP076253.D	31 Oct 2025 14:43		YP\AJ	Ok,M
14	Q3505-02	10-10-25-PLASTIC	PP076254.D	31 Oct 2025 14:59	DCB Low in 2nd column	YP\AJ	Ok,M
15	Q3505-03	10-17-25-RAGS	PP076255.D	31 Oct 2025 15:15		YP\AJ	Ok,M
16	AR1660CCC500	AR1660CCC500	PP076256.D	31 Oct 2025 16:04		YP\AJ	Ok,M
17	AR1242CCC500	AR1242CCC500	PP076257.D	31 Oct 2025 16:20		YP\AJ	Ok,M

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QCBatch ID # PP103125

Review By	rahul	Review On	11/3/2025 8:18:24 AM		
Supervise By	yogesh	Supervise On	11/3/2025 8:49:19 AM		
SubDirectory	PP103125	HP Acquire Method		HP Processing Method	PP102125

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854
CCC	PP24824,PP24828,PP24832,PP24836,PP24840,PP24844,PP24852
Internal Standard/PEM	
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

18	AR1248CCC500	AR1248CCC500	PP076258.D	31 Oct 2025 16:37		YP\AJ	Ok,M
19	AR1254CCC500	AR1254CCC500	PP076259.D	31 Oct 2025 16:53		YP\AJ	Ok,M
20	I.BLK	I.BLK	PP076260.D	31 Oct 2025 17:09		YP\AJ	Ok
21	PB170347BL	PB170347BL	PP076261.D	31 Oct 2025 17:26		YP\AJ	Ok
22	PB170347BS	PB170347BS	PP076262.D	31 Oct 2025 17:42		YP\AJ	Ok,M
23	PB170347BSD	PB170347BSD	PP076263.D	31 Oct 2025 17:58		YP\AJ	Ok,M
24	Q3484-01	TRE-25-0118	PP076264.D	31 Oct 2025 18:14	AR1254 hit	YP\AJ	Ok,M
25	Q3492-01	252611	PP076265.D	31 Oct 2025 18:31		YP\AJ	Ok,M
26	Q3495-03	TW-1025-1	PP076266.D	31 Oct 2025 18:47		YP\AJ	Ok,M
27	Q3500-02	462	PP076267.D	31 Oct 2025 19:03	AR1248+1260 Hit	YP\AJ	Ok,M
28	Q3500-03	463	PP076268.D	31 Oct 2025 19:19	AR1248+1260 Hit	YP\AJ	Ok,M
29	Q3500-04	465	PP076269.D	31 Oct 2025 19:36	DCB Low in 2nd column	YP\AJ	Ok,M
30	Q3503-07	320	PP076270.D	31 Oct 2025 19:52	DCB low and TCMX high in 2nd column	YP\AJ	ReRun
31	AR1660CCC500	AR1660CCC500	PP076271.D	31 Oct 2025 20:57		YP\AJ	Ok
32	AR1242CCC500	AR1242CCC500	PP076272.D	31 Oct 2025 21:46	TCMX high in 2nd column	YP\AJ	Ok,M
33	AR1248CCC500	AR1248CCC500	PP076273.D	31 Oct 2025 22:02		YP\AJ	Ok,M
34	AR1254CCC500	AR1254CCC500	PP076274.D	31 Oct 2025 22:18		YP\AJ	Ok,M
35	I.BLK	I.BLK	PP076275.D	31 Oct 2025 22:34		YP\AJ	Ok
36	PB170358BL	PB170358BL	PP076276.D	31 Oct 2025 22:51		YP\AJ	Ok

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QCBatch ID # PP103125

Review By	rahul	Review On	11/3/2025 8:18:24 AM		
Supervise By	yogesh	Supervise On	11/3/2025 8:49:19 AM		
SubDirectory	PP103125	HP Acquire Method		HP Processing Method	PP102125

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854
CCC	PP24824,PP24828,PP24832,PP24836,PP24840,PP24844,PP24852
Internal Standard/PEM	
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

37	PB170358BS	PB170358BS	PP076277.D	31 Oct 2025 23:07		YP\AJ	Ok,M
38	Q3437-15	SS-32(0-0.5)	PP076278.D	31 Oct 2025 23:23	AR1254 Hit	YP\AJ	Ok,M
39	Q3450-04	SS-10R-(3.5-4.0)	PP076279.D	31 Oct 2025 23:40		YP\AJ	Ok,M
40	Q3450-04MS	SS-10R-(3.5-4.0)MS	PP076280.D	31 Oct 2025 23:56		YP\AJ	Ok,M
41	Q3450-04MSD	SS-10R-(3.5-4.0)MSD	PP076281.D	01 Nov 2025 00:12		YP\AJ	Ok,M
42	Q3450-05	SS-13R-(3.5-4.0)	PP076282.D	01 Nov 2025 00:28		YP\AJ	Ok
43	Q3450-06	SS-13R-(5.0-5.5)	PP076283.D	01 Nov 2025 00:45		YP\AJ	Ok,M
44	Q3450-23	SS-22-(0.0-0.5)	PP076284.D	01 Nov 2025 01:01	AR1254 Hit , Need 5x dilution	YP\AJ	Dilution
45	Q3450-25	SS-22-(3.5-4.0)	PP076285.D	01 Nov 2025 01:17	AR1254 Hit	YP\AJ	Ok,M
46	AR1660CCC500	AR1660CCC500	PP076286.D	01 Nov 2025 02:22		YP\AJ	Ok,M
47	AR1242CCC500	AR1242CCC500	PP076287.D	01 Nov 2025 03:11	TCMX high in 2nd column	YP\AJ	Ok,M
48	AR1248CCC500	AR1248CCC500	PP076288.D	01 Nov 2025 03:28	TCMX high in 2nd column	YP\AJ	Ok,M
49	AR1254CCC500	AR1254CCC500	PP076289.D	01 Nov 2025 03:44	TCMX high in 2nd column	YP\AJ	Ok,M
50	I.BLK	I.BLK	PP076290.D	01 Nov 2025 04:00		YP\AJ	Ok

M : Manual Integration

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QCBatch ID # PP110325

Review By	rahul	Review On	11/4/2025 8:41:32 AM		
Supervise By	yogesh	Supervise On	11/4/2025 10:10:47 AM		
SubDirectory	PP110325	HP Acquire Method		HP Processing Method	PP102125

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854
CCC	PP24824,PP24828,PP24832,PP24836,PP24840,PP24844,PP24852
Internal Standard/PEM	
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PP076291.D	03 Nov 2025 08:28		YP\AJ	Ok
2	AR1660CCC500	AR1660CCC500	PP076292.D	03 Nov 2025 08:44	AR1260-04 high in second column	YP\AJ	Ok
3	AR1242CCC500	AR1242CCC500	PP076293.D	03 Nov 2025 09:04		YP\AJ	Ok,M
4	AR1248CCC500	AR1248CCC500	PP076294.D	03 Nov 2025 09:21		YP\AJ	Ok,M
5	AR1254CCC500	AR1254CCC500	PP076295.D	03 Nov 2025 09:37		YP\AJ	Ok,M
6	I.BLK	I.BLK	PP076296.D	03 Nov 2025 09:53		YP\AJ	Ok
7	Q3450-23DL	SS-22-(0.0-0.5)DL	PP076297.D	03 Nov 2025 10:09	AR1254 Hit	YP\AJ	Ok,M
8	PB170370BL	PB170370BL	PP076298.D	03 Nov 2025 11:53		YP\AJ	Ok,M
9	PB170370BS	PB170370BS	PP076299.D	03 Nov 2025 12:09		YP\AJ	Ok,M
10	Q3515-01	SU-04-103125	PP076300.D	03 Nov 2025 12:25	TCMX high in 2nd column	YP\AJ	Ok,M
11	Q3515-01MS	SU-04-103125MS	PP076301.D	03 Nov 2025 12:41	TCMX high in 2nd column	YP\AJ	Ok,M
12	Q3515-01MSD	SU-04-103125MSD	PP076302.D	03 Nov 2025 12:58	TCMX high in 2nd column	YP\AJ	Ok,M
13	Q3516-01	HD-01-103125	PP076303.D	03 Nov 2025 13:14		YP\AJ	Ok,M
14	AR1660CCC500	AR1660CCC500	PP076304.D	03 Nov 2025 14:19		YP\AJ	Ok
15	AR1242CCC500	AR1242CCC500	PP076305.D	03 Nov 2025 14:35		YP\AJ	Ok,M
16	AR1248CCC500	AR1248CCC500	PP076306.D	03 Nov 2025 14:51		YP\AJ	Ok,M
17	AR1254CCC500	AR1254CCC500	PP076307.D	03 Nov 2025 15:08		YP\AJ	Ok,M

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QCBatch ID # PP110325

Review By	rahul	Review On	11/4/2025 8:41:32 AM		
Supervise By	yogesh	Supervise On	11/4/2025 10:10:47 AM		
SubDirectory	PP110325	HP Acquire Method		HP Processing Method	PP102125

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854
CCC	PP24824,PP24828,PP24832,PP24836,PP24840,PP24844,PP24852
Internal Standard/PEM	
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

18	I.BLK	I.BLK	PP076308.D	03 Nov 2025 15:24		YP\AJ	Ok

M : Manual Integration

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: ECD_Q

Daily Analysis Runlog For Sequence/QCBatch ID # PQ101025

Review By	yogesh	Review On	10/10/2025 10:11:25 AM		
Supervise By	mohammad	Supervise On	10/14/2025 2:48:24 AM		
SubDirectory	PQ101025	HP Acquire Method		HP Processing Method	PQ101025

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854
CCC	
Internal Standard/PEM	
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PQ070971.D	10 Oct 2025 08:27		YP\AJ	Ok
2	I.BLK	I.BLK	PQ070972.D	10 Oct 2025 08:42		YP\AJ	Ok
3	AR1660ICC1000	AR1660ICC1000	PQ070973.D	10 Oct 2025 08:57		YP\AJ	Ok
4	AR1660ICC750	AR1660ICC750	PQ070974.D	10 Oct 2025 09:11		YP\AJ	Ok
5	AR1660ICC500	AR1660ICC500	PQ070975.D	10 Oct 2025 09:26		YP\AJ	Ok
6	AR1660ICC250	AR1660ICC250	PQ070976.D	10 Oct 2025 09:40		YP\AJ	Ok
7	AR1660ICC050	AR1660ICC050	PQ070977.D	10 Oct 2025 09:55		YP\AJ	Ok,M
8	AR1221ICC500	AR1221ICC500	PQ070978.D	10 Oct 2025 10:09		YP\AJ	Ok
9	AR1232ICC500	AR1232ICC500	PQ070979.D	10 Oct 2025 10:24		YP\AJ	Ok
10	AR1242ICC1000	AR1242ICC1000	PQ070980.D	10 Oct 2025 10:38		YP\AJ	Ok
11	AR1242ICC750	AR1242ICC750	PQ070981.D	10 Oct 2025 10:53		YP\AJ	Ok
12	AR1242ICC500	AR1242ICC500	PQ070982.D	10 Oct 2025 11:08		YP\AJ	Ok
13	AR1242ICC250	AR1242ICC250	PQ070983.D	10 Oct 2025 11:22		YP\AJ	Ok
14	AR1242ICC050	AR1242ICC050	PQ070984.D	10 Oct 2025 11:37		YP\AJ	Ok,M
15	AR1248ICC1000	AR1248ICC1000	PQ070985.D	10 Oct 2025 11:51		YP\AJ	Ok
16	AR1248ICC750	AR1248ICC750	PQ070986.D	10 Oct 2025 12:06		YP\AJ	Ok
17	AR1248ICC500	AR1248ICC500	PQ070987.D	10 Oct 2025 12:20		YP\AJ	Ok
18	AR1248ICC250	AR1248ICC250	PQ070988.D	10 Oct 2025 12:35		YP\AJ	Ok,M

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: ECD_Q

Daily Analysis Runlog For Sequence/QCBatch ID # PQ101025

Review By	yogesh	Review On	10/10/2025 10:11:25 AM		
Supervise By	mohammad	Supervise On	10/14/2025 2:48:24 AM		
SubDirectory	PQ101025	HP Acquire Method		HP Processing Method	PQ101025

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854
CCC	
Internal Standard/PEM	
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

19	AR1248ICC050	AR1248ICC050	PQ070989.D	10 Oct 2025 12:50		YP\AJ	Ok
20	AR1254ICC1000	AR1254ICC1000	PQ070990.D	10 Oct 2025 13:04		YP\AJ	Ok
21	AR1254ICC750	AR1254ICC750	PQ070991.D	10 Oct 2025 13:19		YP\AJ	Ok
22	AR1254ICC500	AR1254ICC500	PQ070992.D	10 Oct 2025 13:33		YP\AJ	Ok
23	AR1254ICC250	AR1254ICC250	PQ070993.D	10 Oct 2025 13:48		YP\AJ	Ok
24	AR1254ICC050	AR1254ICC050	PQ070994.D	10 Oct 2025 15:15		YP\AJ	Ok
25	AR1262ICC500	AR1262ICC500	PQ070995.D	10 Oct 2025 15:30		YP\AJ	Ok
26	AR1268ICC1000	AR1268ICC1000	PQ070996.D	10 Oct 2025 15:44		YP\AJ	Ok
27	AR1268ICC750	AR1268ICC750	PQ070997.D	10 Oct 2025 15:59		YP\AJ	Ok
28	AR1268ICC500	AR1268ICC500	PQ070998.D	10 Oct 2025 16:13		YP\AJ	Ok
29	AR1268ICC250	AR1268ICC250	PQ070999.D	10 Oct 2025 16:28		YP\AJ	Ok
30	AR1268ICC050	AR1268ICC050	PQ071000.D	10 Oct 2025 16:43		YP\AJ	Ok
31	PQ101025ICV500	ICVPQ101025	PQ071001.D	10 Oct 2025 16:57		YP\AJ	Ok,M
32	AR1242ICV500	ICVPQ101025AR1242	PQ071002.D	10 Oct 2025 17:41		YP\AJ	Ok,M
33	AR1248ICV500	ICVPQ101025AR1248	PQ071003.D	10 Oct 2025 18:25		YP\AJ	Ok,M
34	AR1254ICV500	ICVPQ101025AR1254	PQ071004.D	10 Oct 2025 19:08		YP\AJ	Ok
35	AR1268ICV500	ICVPQ101025AR1268	PQ071005.D	10 Oct 2025 19:52		YP\AJ	Ok
36	DDT ANALOG	DDT ANALOG	PQ071006.D	10 Oct 2025 20:36		YP\AJ	Ok

M : Manual Integration

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: ECD_Q

Daily Analysis Runlog For Sequence/QCBatch ID # PQ102725

Review By	yogesh	Review On	10/27/2025 10:54:57 AM		
Supervise By	mohammad	Supervise On	10/29/2025 5:27:51 AM		
SubDirectory	PQ102725	HP Acquire Method		HP Processing Method	PQ101025

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854
CCC	PP24824,PP24828,PP24832,PP24836,PP24840,PP24844,PP24852
Internal Standard/PEM	
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PQ071148.D	27 Oct 2025 08:42		YP\AJ	Ok
2	AR1660CCC500	AR1660CCC500	PQ071149.D	27 Oct 2025 08:57		YP\AJ	Ok
3	I.BLK	I.BLK	PQ071150.D	27 Oct 2025 09:11		YP\AJ	Ok
4	Q3450-01	SS-5R-(2.0-2.5)	PQ071151.D	27 Oct 2025 09:26	AR1254 Hit	YP\AJ	Ok
5	Q3450-02	SS-5R-(3.5-4.0)	PQ071152.D	27 Oct 2025 09:41	AR1254 Hit	YP\AJ	Ok
6	Q3450-02MS	SS-5R-(3.5-4.0)MS	PQ071153.D	27 Oct 2025 09:55		YP\AJ	Ok,M
7	Q3450-02MSD	SS-5R-(3.5-4.0)MSD	PQ071154.D	27 Oct 2025 10:10		YP\AJ	Ok,M
8	Q3450-08	SS-20-(0.0-0.5)	PQ071155.D	27 Oct 2025 10:24	AR1254 Hit	YP\AJ	Ok
9	Q3450-09	SS-20-(2.0-2.5)	PQ071156.D	27 Oct 2025 10:39	AR1254 Hit	YP\AJ	Ok,M
10	Q3437-16	SS-23(0-0.5)	PQ071157.D	27 Oct 2025 10:54	AR1254 Hit, TCX high 1st column	YP\AJ	Ok
11	Q3419-01	PR-132-S16-015094-20	PQ071158.D	27 Oct 2025 11:08	AR1242 hits	YP\AJ	Ok,M
12	Q3419-02MS	PR-132-S16-015094-20	PQ071159.D	27 Oct 2025 11:23		YP\AJ	Ok
13	Q3419-03MSD	PR-132-S16-015094-20	PQ071160.D	27 Oct 2025 11:37		YP\AJ	Ok
14	AR1660CCC500	AR1660CCC500	PQ071161.D	27 Oct 2025 12:35		YP\AJ	Ok
15	AR1242CCC500	AR1242CCC500	PQ071162.D	27 Oct 2025 13:34		YP\AJ	Ok
16	AR1248CCC500	AR1248CCC500	PQ071163.D	27 Oct 2025 13:49		YP\AJ	Ok
17	AR1254CCC500	AR1254CCC500	PQ071164.D	27 Oct 2025 14:03		YP\AJ	Ok

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: ECD_Q

Daily Analysis Runlog For Sequence/QCBatch ID # PQ102725

Review By	yogesh	Review On	10/27/2025 10:54:57 AM		
Supervise By	mohammad	Supervise On	10/29/2025 5:27:51 AM		
SubDirectory	PQ102725	HP Acquire Method		HP Processing Method	PQ101025

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854
CCC	PP24824,PP24828,PP24832,PP24836,PP24840,PP24844,PP24852
Internal Standard/PEM	
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

18	I.BLK	I.BLK	PQ071165.D	27 Oct 2025 14:18		YP\AJ	Ok
19	PB170255BL	PB170255BL	PQ071166.D	27 Oct 2025 14:32	DCB high 1st column for PSEG	YP\AJ	Not Ok
20	PB170255BS	PB170255BS	PQ071167.D	27 Oct 2025 14:47	DCB high 1st column for PSEG	YP\AJ	Not Ok
21	Q3451-01	TP-11	PQ071168.D	27 Oct 2025 15:01	Tcmx high in 1st column	YP\AJ	Ok
22	Q3451-05	TP-12	PQ071169.D	27 Oct 2025 15:16		YP\AJ	Ok
23	Q3451-05MS	TP-12MS	PQ071170.D	27 Oct 2025 15:31	Tcmx high in 1st column	YP\AJ	Ok
24	Q3451-05MSD	TP-12MSD	PQ071171.D	27 Oct 2025 15:45	Tcmx high in 1st column	YP\AJ	Ok
25	Q3452-01	JB-1	PQ071172.D	27 Oct 2025 16:00		YP\AJ	Ok
26	Q3460-05	WC-2	PQ071173.D	27 Oct 2025 16:14	TCMX & DCB high in 1st column	YP\AJ	ReRun
27	Q3460-09	WC-3	PQ071174.D	27 Oct 2025 16:29	TCMX & DCB high in 1st column	YP\AJ	ReRun
28	Q3460-13	WC-4	PQ071175.D	27 Oct 2025 16:43	DCB high in 1st column, AR1260 hits	YP\AJ	Ok,M
29	AR1660CCC500	AR1660CCC500	PQ071176.D	27 Oct 2025 17:42		YP\AJ	Ok,M
30	AR1242CCC500	AR1242CCC500	PQ071177.D	27 Oct 2025 18:40		YP\AJ	Ok
31	AR1248CCC500	AR1248CCC500	PQ071178.D	27 Oct 2025 18:55		YP\AJ	Ok
32	AR1254CCC500	AR1254CCC500	PQ071179.D	27 Oct 2025 19:09		YP\AJ	Ok,M
33	I.BLK	I.BLK	PQ071180.D	27 Oct 2025 19:24		YP\AJ	Ok

M : Manual Integration

PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 10/27/2025

OVENTEMP IN Celsius(°C): 107
Time IN: 17:25
In Date: 10/24/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 104
Time OUT: 08:27
Out Date: 10/25/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID-OVEN

QC: LB137635

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
Q3449-01	CF-620-COMP-41	1	1.14	10.13	11.27	10.24	89.8	
Q3449-02	CF-620-VOC-41	2	1.15	10.74	11.89	11.28	94.3	
Q3449-03	CF-620-121	3	1.16	10.81	11.97	11.05	91.5	
Q3449-04	CF-620-122	4	1.14	10.40	11.54	10.35	88.6	
Q3449-05	CF-620-123	5	1.19	10.37	11.56	9.88	83.8	
Q3450-01	SS-5R-(2.0-2.5)	6	1.19	10.00	11.19	10.48	92.9	
Q3450-02	SS-5R-(3.5-4.0)	7	1.15	10.95	12.1	10.6	86.3	
Q3450-03	SS-10R-(2.0-2.5)	8	1.18	10.53	11.71	11.00	93.3	
Q3450-04	SS-10R-(3.5-4.0)	9	1.12	10.86	11.98	10.52	86.6	
Q3450-05	SS-13R-(3.5-4.0)	10	1.15	10.58	11.73	10.88	92.0	
Q3450-06	SS-13R-(5.0-5.5)	11	1.17	10.15	11.32	9.71	84.1	
Q3450-07	SS-13R-(5.0-5.5)	12	1.13	10.68	11.81	10.24	85.3	
Q3450-08	SS-20-(0.0-0.5)	13	1.15	9.72	10.87	10.24	93.5	
Q3450-09	SS-20-(2.0-2.5)	14	1.14	10.79	11.93	10.45	86.3	
Q3450-10	SS-20-(3.5-4.0)	15	1.17	10.15	11.32	9.7	84.0	
Q3450-11	SS-28-(2.0-2.5)	16	1.18	10.44	11.62	10.89	93.0	
Q3450-12	SS-29-(20-2.5)	17	1.13	11.27	12.4	10.24	80.8	
Q3450-13	SS-30-(0.0-0.5)	18	1.18	10.14	11.32	10.42	91.1	
Q3450-14	SS-30-(2.0-2.5)	19	1.18	10.12	11.3	10.04	87.5	
Q3450-15	SS-30-(3.5-4.0)	20	1.15	10.14	11.29	9.74	84.7	
Q3450-16	SS-31-(0.0-0.5)	21	1.17	11.01	12.18	10.93	88.6	
Q3450-17	SS-34-(0.0-0.5)	22	1.11	10.41	11.52	10.14	86.7	
Q3450-18	SS-34-(2.0-2.5)	23	1.18	10.67	11.85	11.00	92.0	
Q3450-19	SS-34-(3.5-4.0)	24	1.14	10.04	11.18	10.79	96.1	
Q3450-20	SS-21-(0.0-0.5)	25	1.14	10.28	11.42	10.57	91.7	
Q3450-21	SS-21-(2.0-2.5)	26	1.14	9.99	11.13	10.45	93.2	
Q3450-22	SS-21-(3.5-4.0)	27	1.18	10.56	11.74	10.12	84.7	
Q3450-23	SS-22-(0.0-0.5)	28	1.18	10.29	11.47	10.59	91.4	

PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 10/27/2025

OVENTEMP IN Celsius(°C): 107
Time IN: 17:25
In Date: 10/24/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 104
Time OUT: 08:27
Out Date: 10/25/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID-OVEN

QC: LB137635

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
Q3450-24	SS-22-(2.0-2.5)	29	1.14	10.09	11.23	10.32	91.0	
Q3450-25	SS-22-(3.5-4.0)	30	1.13	10.66	11.79	9.95	82.7	
Q3451-01	TP-11	31	1.18	10.88	12.06	10.95	89.8	
Q3451-02	TP-11-EPH	32	1.13	10.40	11.53	10.07	86.0	
Q3451-03	TP-11-VOC	33	1.16	10.54	11.7	10.99	93.3	
Q3451-05	TP-12	34	1.19	10.28	11.47	10.35	89.1	
Q3451-06	TP-12-EPH	35	1.15	11.08	12.23	10.62	85.5	
Q3451-07	TP-12-VOC	36	1.17	10.05	11.22	10.24	90.2	
Q3452-01	JB-1	37	1.12	10.26	11.38	8.46	71.5	
Q3452-02	JB-1-EPH	38	1.15	10.80	11.95	9.15	74.1	
Q3452-03	JB-1-VOC	39	1.16	10.37	11.53	10.32	88.3	
Q3453-01	102225-A	40	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
Q3453-02	102225-B	41	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
Q3453-03	102325-A	42	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
Q3453-04	102325-B	43	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
Q3456-01	OILY-DEBRIS	50	1.00	1.00	2.00	2.00	100.0	oily-debris
Q3457-01	TR-05-102425	51	1.14	10.12	11.26	10.47	92.2	
Q3457-02	TR-05-102425-E2	52	1.16	10.37	11.53	10.88	93.7	
Q3458-01	TRE-25-0116	53	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
Q3460-01	WC-1	54	1.18	10.42	11.6	10.9	93.3	
Q3460-02	WC-1-EPH	55	1.19	10.51	11.7	11.00	93.3	
Q3460-03	WC-1-VOC	56	1.19	10.04	11.23	10.52	92.9	
Q3460-05	WC-2	57	1.14	10.24	11.38	11.04	96.7	
Q3460-06	WC-2-EPH	58	1.18	10.48	11.66	11.3	96.6	
Q3460-07	WC-2-VOC	59	1.13	9.83	10.96	10.6	96.3	
Q3460-09	WC-3	60	1.19	10.78	11.97	11.36	94.3	
Q3460-10	WC-3-EPH	61	1.11	10.30	11.41	10.81	94.2	
Q3460-11	WC-3-VOC	62	1.16	10.12	11.28	10.74	94.7	

PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 10/27/2025

OVENTEMP IN Celsius(°C): 107
Time IN: 17:25
In Date: 10/24/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 104
Time OUT: 08:27
Out Date: 10/25/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID-OVEN

QC: LB137635

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
Q3460-13	WC-4	63	1.13	10.80	11.93	11.41	95.2	
Q3460-14	WC-4-EPH	64	1.18	10.43	11.61	11.02	94.3	
Q3460-15	WC-4-VOC	65	1.18	9.96	11.14	10.6	94.6	
Q3460-17	WC-5	66	1.17	10.38	11.55	10.99	94.6	
Q3460-18	WC-5-EPH	67	1.19	10.50	11.69	11.14	94.8	
Q3460-19	WC-5-VOC	68	1.16	10.44	11.6	11.09	95.1	
Q3463-05	SVOC-GPC-BLANK	44	1.00	1.00	2.00	2.00	100.0	
Q3463-06	PEST-GPC-BLANK	45	1.00	1.00	2.00	2.00	100.0	
Q3463-07	PEST-GPC-BLANK-SPIKE	46	1.00	1.00	2.00	2.00	100.0	
Q3463-08	SVOC-GPC2-BLANK	47	1.00	1.00	2.00	2.00	100.0	
Q3463-09	PEST-GPC2-BLANK	48	1.00	1.00	2.00	2.00	100.0	
Q3463-10	PEST -GPC2-BLANK-SPIKE	49	1.00	1.00	2.00	2.00	100.0	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$

WORKLIST(Hardcopy Internal Chain)

VB 137635

WorkList Name : %1-102425 **WorkList ID :** 192653 **Department :** Wet-Chemistry **Date :** 10-24-2025 08:07:25

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q3449-01	CF-620-COMP-41	Solid	Percent Solids	Cool 4 deg C	PSEG03	D21	10/23/2025	Chemtech -SO
Q3449-02	CF-620-VOC-41	Solid	Percent Solids	Cool 4 deg C	PSEG03	D21	10/23/2025	Chemtech -SO
Q3449-03	CF-620-121	Solid	Percent Solids	Cool 4 deg C	PSEG03	D21	10/23/2025	Chemtech -SO
Q3449-04	CF-620-122	Solid	Percent Solids	Cool 4 deg C	PSEG03	D21	10/23/2025	Chemtech -SO
Q3449-05	CF-620-123	Solid	Percent Solids	Cool 4 deg C	PSEG03	D21	10/23/2025	Chemtech -SO
Q3450-01	SS-5R-(2.0-2.5)	Solid	Percent Solids	Cool 4 deg C	PSEG03	D21	10/23/2025	Chemtech -SO
Q3450-02	SS-5R-(3.5-4.0)	Solid	Percent Solids	Cool 4 deg C	MATR02	D41	10/23/2025	Chemtech -SO
Q3450-03	SS-10R-(2.0-2.5)	Solid	Percent Solids	Cool 4 deg C	MATR02	D41	10/23/2025	Chemtech -SO
Q3450-04	SS-10R-(3.5-4.0)	Solid	Percent Solids	Cool 4 deg C	MATR02	D41	10/23/2025	Chemtech -SO
Q3450-05	SS-13R-(3.5-4.0)	Solid	Percent Solids	Cool 4 deg C	MATR02	D41	10/23/2025	Chemtech -SO
Q3450-06	SS-13R-(5.0-5.5)	Solid	Percent Solids	Cool 4 deg C	MATR02	D41	10/23/2025	Chemtech -SO
Q3450-07	SS-13R-(5.0-5.5)	Solid	Percent Solids	Cool 4 deg C	MATR02	D41	10/23/2025	Chemtech -SO
Q3450-08	SS-20-(0.0-0.5)	Solid	Percent Solids	Cool 4 deg C	MATR02	D41	10/23/2025	Chemtech -SO
Q3450-09	SS-20-(2.0-2.5)	Solid	Percent Solids	Cool 4 deg C	MATR02	D41	10/23/2025	Chemtech -SO
Q3450-10	SS-20-(3.5-4.0)	Solid	Percent Solids	Cool 4 deg C	MATR02	D41	10/23/2025	Chemtech -SO
Q3450-11	SS-28-(2.0-2.5)	Solid	Percent Solids	Cool 4 deg C	MATR02	D41	10/23/2025	Chemtech -SO
Q3450-12	SS-29-(20-2.5)	Solid	Percent Solids	Cool 4 deg C	MATR02	D41	10/23/2025	Chemtech -SO
Q3450-13	SS-30-(0.0-0.5)	Solid	Percent Solids	Cool 4 deg C	MATR02	D41	10/23/2025	Chemtech -SO
Q3450-14	SS-30-(2.0-2.5)	Solid	Percent Solids	Cool 4 deg C	MATR02	D41	10/23/2025	Chemtech -SO
Q3450-15	SS-30-(3.5-4.0)	Solid	Percent Solids	Cool 4 deg C	MATR02	D41	10/23/2025	Chemtech -SO
Q3450-16	SS-31-(0.0-0.5)	Solid	Percent Solids	Cool 4 deg C	MATR02	D41	10/23/2025	Chemtech -SO

Date/Time 10/24/25 15:00

Raw Sample Received by: JB (WC)

Raw Sample Relinquished by: SH

Date/Time 10/24/25 17:35

Raw Sample Received by: SH

Raw Sample Relinquished by: JB (WC)

WORKLIST(Hardcopy Internal Chain)

VR 137135

WorkList Name : %1-102425 **WorkList ID :** 192653 **Department :** Wet-Chemistry **Date :** 10-24-2025 08:07:25

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q3450-17	SS-34-(0.0-0.5)	Solid	Percent Solids	Cool 4 deg C	MATR02	D41	10/23/2025	Chemtech -SO
Q3450-18	SS-34-(2.0-2.5)	Solid	Percent Solids	Cool 4 deg C	MATR02	D41	10/23/2025	Chemtech -SO
Q3450-19	SS-34-(3.5-4.0)	Solid	Percent Solids	Cool 4 deg C	MATR02	D41	10/23/2025	Chemtech -SO
Q3450-20	SS-21-(0.0-0.5)	Solid	Percent Solids	Cool 4 deg C	MATR02	D41	10/23/2025	Chemtech -SO
Q3450-21	SS-21-(2.0-2.5)	Solid	Percent Solids	Cool 4 deg C	MATR02	D41	10/23/2025	Chemtech -SO
Q3450-22	SS-21-(3.5-4.0)	Solid	Percent Solids	Cool 4 deg C	MATR02	D41	10/23/2025	Chemtech -SO
Q3450-23	SS-22-(0.0-0.5)	Solid	Percent Solids	Cool 4 deg C	MATR02	D41	10/23/2025	Chemtech -SO
Q3450-24	SS-22-(2.0-2.5)	Solid	Percent Solids	Cool 4 deg C	MATR02	D41	10/23/2025	Chemtech -SO
Q3450-25	SS-22-(3.5-4.0)	Solid	Percent Solids	Cool 4 deg C	MATR02	D41	10/23/2025	Chemtech -SO
Q3451-01	TP-11	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	10/23/2025	Chemtech -SO
Q3451-02	TP-11-EPH	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	10/23/2025	Chemtech -SO
Q3451-03	TP-11-VOC	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	10/23/2025	Chemtech -SO
Q3451-05	TP-12	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	10/23/2025	Chemtech -SO
Q3451-06	TP-12-EPH	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	10/23/2025	Chemtech -SO
Q3451-07	TP-12-VOC	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	10/23/2025	Chemtech -SO
Q3452-01	JB-1	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	10/23/2025	Chemtech -SO
Q3452-02	JB-1-EPH	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	10/24/2025	Chemtech -SO
Q3452-03	JB-1-VOC	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	10/24/2025	Chemtech -SO
Q3453-01	102225-A	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	10/24/2025	Chemtech -SO
Q3453-02	102225-B	Solid	Percent Solids	Cool 4 deg C	PSEG03	D41	10/24/2025	Chemtech -SO
Q3453-03	102325-A	Solid	Percent Solids	Cool 4 deg C	PSEG03	D41	10/24/2025	Chemtech -SO

Date/Time 10/24/25 15:00

Raw Sample Received by: TB WEC

Raw Sample Relinquished by: SH

Date/Time 10/24/25 17:35

Raw Sample Received by: SH

Raw Sample Relinquished by: TB WEC

WORKLIST(Hardcopy Internal Chain)

SR137635

WorkList Name : %1-102425 **WorkList ID :** 192653 **Department :** Wet-Chemistry **Date :** 10-24-2025 08:07:25

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q3453-04	102325-B	Solid	Percent Solids	Cool 4 deg C	PSEG03	D41	10/24/2025	Chemtech -SO
Q3456-01	OILY-DEBRIS	Solid	Percent Solids	Cool 4 deg C	PSEG03	D41	10/24/2025	Chemtech -SO
Q3457-01	TR-05-102425	Solid	Percent Solids	Cool 4 deg C	PSEG05	D31	10/24/2025	Chemtech -SO
Q3457-02	TR-05-102425-E2	Solid	Percent Solids	Cool 4 deg C	PSEG05	D31	10/24/2025	Chemtech -SO
Q3458-01	TRE-25-0116	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	10/24/2025	Chemtech -SO
Q3460-01	WC-1	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	10/24/2025	Chemtech -SO
Q3460-02	WC-1-EPH	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	10/24/2025	Chemtech -SO
Q3460-03	WC-1-VOC	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	10/24/2025	Chemtech -SO
Q3460-05	WC-2	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	10/24/2025	Chemtech -SO
Q3460-06	WC-2-EPH	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	10/24/2025	Chemtech -SO
Q3460-07	WC-2-VOC	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	10/24/2025	Chemtech -SO
Q3460-09	WC-3	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	10/24/2025	Chemtech -SO
Q3460-10	WC-3-EPH	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	10/24/2025	Chemtech -SO
Q3460-11	WC-3-VOC	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	10/24/2025	Chemtech -SO
Q3460-13	WC-4	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	10/24/2025	Chemtech -SO
Q3460-14	WC-4-EPH	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	10/24/2025	Chemtech -SO
Q3460-15	WC-4-VOC	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	10/24/2025	Chemtech -SO
Q3460-17	WC-5	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	10/24/2025	Chemtech -SO
Q3460-18	WC-5-EPH	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	10/24/2025	Chemtech -SO
Q3460-19	WC-5-VOC	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	10/24/2025	Chemtech -SO
Q3463-05	SVOC-GPC-BLANK	Solid	Percent Solids	Cool 4 deg C	CHEM02	J32	10/17/2025	Chemtech -SO

Date/Time 10/24/25 1510

Raw Sample Received by: TB WC

Raw Sample Relinquished by: SH

Date/Time 10/24/25 1735

Raw Sample Received by: SH

Raw Sample Relinquished by: TB Coc

WORKLIST(Hardcopy Internal Chain)

13:35

WorkList Name : %1-102425 **WorkList ID :** 192653 **Department :** Wet-Chemistry **Date :** 10-24-2025 08:07:25

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q3463-06	PEST-GPC-BLANK	Solid	Percent Solids	Cool 4 deg C	CHEM02	J32	10/17/2025	Chemtech -SO
Q3463-07	PEST-GPC-BLANK-SPIKE	Solid	Percent Solids	Cool 4 deg C	CHEM02	J32	10/17/2025	Chemtech -SO
Q3463-08	SVOC-GPC2-BLANK	Solid	Percent Solids	Cool 4 deg C	CHEM02	J32	10/17/2025	Chemtech -SO
Q3463-09	PEST-GPC2-BLANK	Solid	Percent Solids	Cool 4 deg C	CHEM02	J32	10/17/2025	Chemtech -SO
Q3463-10	PEST -GPC2-BLANK-SPIKE	Solid	Percent Solids	Cool 4 deg C	CHEM02	J32	10/17/2025	Chemtech -SO

Date/Time 10/24/25 13:00

Raw Sample Received by:

Raw Sample Relinquished by:

Date/Time 10/24/25 17:35

Raw Sample Received by:

Raw Sample Relinquished by:

Alliance TECHNICAL GROUP

EXTRACTION LOGPAGE

PB170236

SOP ID: M3541-ASE Extraction-15

Clean Up SOP #: Acid Cleanup

Matrix : Solid

Weigh By: EH

Balance check: EH

Balance ID: EX-SC-2

pH Strip Lot#: N/A

Extraction By: RJ

Filter By: RJ

pH Meter ID: N/A

Hood ID: 3,7

Extraction Start Date : 10/24/2025

Extraction Start Time : 12:05

Extraction End Date : 10/24/2025

Extraction End Time : 15:25

Concentration By: EH

Supervisor By : RUPESH

Extraction Method: ☐ Seperatory Funnel ☐ Continious Liquid/Liquid ☐ Sonication ☐ Waste Dilution ☑ Soxhlet

Standared Name	MLS USED	Concentration ug/mL	STD REF. # FROM LOG
Spike Sol 1	1.0ML	5000 PPB	PP24930
Surrogate	1.0ML	200 PPB	PP24663
N/A	N/A	N/A	N/A
N/A	N/A	N/A	N/A
N/A	N/A	N/A	N/A

Chemical Used	ML/SAMPLE USED	Lot Number
Hexane/Acetone/1:1	N/A	EP2643
Baked Na2SO4	N/A	EP2652
Sand	N/A	E3951
Hexane	N/A	E3978
H2SO4 1:1	N/A	EP2610
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

Extraction Conformance/Non-Conformance Comments:

40ML Vial Lot # 03-40BTS721. Q3447-07 used Limited volume as sample is tar matrix.

KD Bath ID: N/A

KD Bath Temperature: N/A

Envap ID: N-EVAP-02

Envap Temperature: 40 °C

Date / Time	Prepped Sample Relinquished By/Location	Received By/Location
10/24/25	RS (Ext Lab)	
15:30	Preparation Group	Analysis Group

Alliance TECHNICAL GROUP

EXTRACTION LOGPAGE

PrepBatch ID : PB170236

Analytical Method: M3541-ASE Extraction-15

Concentration Date: 10/24/2025

Sample ID	Client Sample ID	Test	g/mL	PH	Surr/Spike By:		Final Vol. (mL)	JarID	Comments	Prep Pos
					AddedBy	VerifiedBy				
PB170236BL	ABLK236	PCB	30.02	N/A	ritesh	Evelyn	10			U1-1
PB170236BS	ALCS236	PCB	30.01	N/A	ritesh	Evelyn	10			2
Q3427-01	TP-8	PCB	30.06	N/A	ritesh	Evelyn	10	E	Oily Mat.	3
Q3447-01	P001-TW1-251023-01	PCB	30.07	N/A	ritesh	Evelyn	10	E	Oily Mat.	4
Q3447-03	P001-TW2-251023-01	PCB	30.04	N/A	ritesh	Evelyn	10	E	Tar Mat.	5
Q3447-05	P001-TW3-251023-01	PCB	30.09	N/A	ritesh	Evelyn	10	E	Tar Mat.	6
Q3447-07	P001-TW4-251023-01	PCB	10.05	N/A	ritesh	Evelyn	10	E		U2-1
Q3449-01	CF-620-COMP-41	PCB	30.02	N/A	ritesh	Evelyn	10	E		2
Q3450-01	SS-5R-(2.0-2.5)	PCB	30.03	N/A	ritesh	Evelyn	10			3
Q3450-02	SS-5R-(3.5-4.0)	PCB	30.08	N/A	ritesh	Evelyn	10			4
Q3450-02MS	SS-5R-(3.5-4.0)MS	PCB	30.06	N/A	ritesh	Evelyn	10			5
Q3450-02MSD	SS-5R-(3.5-4.0)MSD	PCB	30.04	N/A	ritesh	Evelyn	10			6
Q3450-08	SS-20-(0.0-0.5)	PCB	30.01	N/A	ritesh	Evelyn	10			U3-1
Q3450-09	SS-20-(2.0-2.5)	PCB	30.09	N/A	ritesh	Evelyn	10			2
Q3450-10	SS-20-(3.5-4.0)	PCB	30.07	N/A	ritesh	Evelyn	10			3
Q3450-11	SS-28-(2.0-2.5)	PCB	30.02	N/A	ritesh	Evelyn	10			4
Q3450-20	SS-21-(0.0-0.5)	PCB	30.05	N/A	ritesh	Evelyn	10			5
Q3450-21	SS-21-(2.0-2.5)	PCB	30.08	N/A	ritesh	Evelyn	10			6
Q3450-22	SS-21-(3.5-4.0)	PCB	30.04	N/A	ritesh	Evelyn	10			U6-1

RS
10/24

* **Extracts relinquished on the same date as received.**

1702 10:45

WORKLIST(Hardcopy Internal Chain)

WorkList Name : Q3450 **WorkList ID :** 192676 **Department :** Extraction **Date :** 10-24-2025 10:40:05

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q3427-01	TP-8	Solid	PCB	Cool 4 deg C	PSEG03	D41	10/21/2025	8082A
Q3447-01	P001-TW1-251023-01	Solid	PCB	Cool 4 deg C	ROYF02	D41	10/23/2025	8082A
Q3447-03	P001-TW2-251023-01	Solid	PCB	Cool 4 deg C	ROYF02	D41	10/23/2025	8082A
Q3447-05	P001-TW3-251023-01	Solid	PCB	Cool 4 deg C	ROYF02	D41	10/23/2025	8082A
Q3447-07	P001-TW4-251023-01	Solid	PCB	Cool 4 deg C	ROYF02	D41	10/23/2025	8082A
Q3449-01	CF-620-COMP-41	Solid	PCB	Cool 4 deg C	PSEG03	D21	10/23/2025	8082A
Q3450-01	SS-5R-(2.0-2.5)	Solid	PCB	Cool 4 deg C	MATR02	D41	10/23/2025	8082A
Q3450-02	SS-5R-(3.5-4.0)	Solid	PCB	Cool 4 deg C	MATR02	D41	10/23/2025	8082A
Q3450-08	SS-20-(0.0-0.5)	Solid	PCB	Cool 4 deg C	MATR02	D41	10/23/2025	8082A
Q3450-09	SS-20-(2.0-2.5)	Solid	PCB	Cool 4 deg C	MATR02	D41	10/23/2025	8082A
Q3450-10	SS-20-(3.5-4.0)	Solid	PCB	Cool 4 deg C	MATR02	D41	10/23/2025	8082A
Q3450-11	SS-28-(2.0-2.5)	Solid	PCB	Cool 4 deg C	MATR02	D41	10/23/2025	8082A
Q3450-20	SS-21-(0.0-0.5)	Solid	PCB	Cool 4 deg C	MATR02	D41	10/23/2025	8082A
Q3450-21	SS-21-(2.0-2.5)	Solid	PCB	Cool 4 deg C	MATR02	D41	10/23/2025	8082A
Q3450-22	SS-21-(3.5-4.0)	Solid	PCB	Cool 4 deg C	MATR02	D41	10/23/2025	8082A

Date/Time 10/24/25 10:40

Raw Sample Received by: RJ (Ext-lab)

Raw Sample Relinquished by:

Date/Time 10/24/25 11:10

Raw Sample Received by:

Raw Sample Relinquished by: RJ (Ext-lab)

Alliance TECHNICAL GROUP

EXTRACTION LOGPAGE

PB170358

SOP ID:	M3541-ASE Extraction-15				
Clean Up SOP #:	Acid Cleanup			**Extraction Start Date :**	10/31/2025
Matrix :	Solid			**Extraction Start Time :**	12:10
Weigh By:	EH	**Extraction By:**	RJ	**Extraction End Date :**	10/31/2025
Balance check:	EH	**Filter By:**	RJ	**Extraction End Time :**	15:10
Balance ID:	EX-SC-2	**pH Meter ID:**	N/A	**Concentration By:**	EH
pH Strip Lot#:	N/A	**Hood ID:**	3,7	**Supervisor By :**	RUPESH

Extraction Method: [ ] Seperatory Funnel [ ] Continious Liquid/Liquid [ ] Sonication [ ] Waste Dilution [√] Soxhlet

Standared Name	MLS USED	Concentration ug/mL	STD REF. # FROM LOG
Spike Sol 1	1.0ML	5000 PPB	PP24930
Surrogate	1.0ML	200 PPB	PP24663
N/A	N/A	N/A	N/A
N/A	N/A	N/A	N/A
N/A	N/A	N/A	N/A

Chemical Used	ML/SAMPLE USED	Lot Number
Hexane/Acetone/1:1	N/A	EP2643
Baked Na2SO4	N/A	EP2655
Hexane	N/A	E3981
H2SO4 1:1	N/A	EP2610
Sand	N/A	E3951
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

Extraction Conformance/Non-Conformance Comments:

40ML Vial Lot # 03-40BTS721. Q3502-04 used Limited volume as sample is oil.

KD Bath ID:	N/A	**Envap ID:**	N-EVAP-02
KD Bath Temperature:	N/A	**Envap Temperature:**	40 °C

Date / Time	Prepped Sample Relinquished By/Location	Received By/Location
10/31/25	RS (Ext-lab)	Y-P. Pest/PCB.
15:15	**Preparation Group**	**Analysis Group**


Alliance TECHNICAL GROUP

EXTRACTION LOGPAGE

PrepBatch ID : PB170358

Analytical Method: M3541-ASE Extraction-15 **Concentration Date:** 10/31/2025

Sample ID	Client Sample ID	Test	g/ mL	PH	Surr/Spike By:		Final Vol. (mL)	JarID	Comments	Prep Pos
					AddedBy	VerifiedBy				
PB170358BL	ABLK358	PCB	30.01	N/A	ritesh	Evelyn	10			U2-1
PB170358BS	ALCS358	PCB	30.03	N/A	ritesh	Evelyn	10			2
Q3437-15	SS-32(0-0.5)	PCB	30.06	N/A	ritesh	Evelyn	10			3
Q3450-04	SS-10R-(3.5-4.0)	PCB	30.09	N/A	ritesh	Evelyn	10			4
Q3450-04MS	SS-10R-(3.5-4.0)MS	PCB	30.02	N/A	ritesh	Evelyn	10			5
Q3450-04MSD	SS-10R-(3.5-4.0)MSD	PCB	30.04	N/A	ritesh	Evelyn	10			6
Q3450-05	SS-13R-(3.5-4.0)	PCB	30.07	N/A	ritesh	Evelyn	10			U3-1
Q3450-06	SS-13R-(5.0-5.5)	PCB	30.03	N/A	ritesh	Evelyn	10			2
Q3450-23	SS-22-(0.0-0.5)	PCB	30.05	N/A	ritesh	Evelyn	10			3
Q3450-25	SS-22-(3.5-4.0)	PCB	30.04	N/A	ritesh	Evelyn	10			4
Q3502-04	PARTS-CLEANER	PCB	1.09	N/A	ritesh	Evelyn	10		Oil	



RS
10/31

*** Extracts relinquished on the same date as received.**

1703358
12:10

WORKLIST(Hardcopy Internal Chain)

WorkList Name : Q3437P **WorkList ID :** 192830 **Department :** Extraction **Date :** 10-31-2025 12:04:12

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q3437-15	SS-32(0-0.5)	Solid	PCB	Cool 4 deg C	MATR02	D31	10/22/2025	8082A
Q3450-04	SS-10R-(3.5-4.0)	Solid	PCB	Cool 4 deg C	MATR02	D41	10/23/2025	8082A
Q3450-05	SS-13R-(3.5-4.0)	Solid	PCB	Cool 4 deg C	MATR02	D41	10/23/2025	8082A
Q3450-06	SS-13R-(5.0-5.5)	Solid	PCB	Cool 4 deg C	MATR02	D41	10/23/2025	8082A
Q3450-23	SS-22-(0.0-0.5)	Solid	PCB	Cool 4 deg C	MATR02	D41	10/23/2025	8082A
Q3450-25	SS-22-(3.5-4.0)	Solid	PCB	Cool 4 deg C	MATR02	D41	10/23/2025	8082A
Q3502-04	PARTS-CLEANER	Solid	PCB	Cool 4 deg C	PSEG03	D41	10/30/2025	8082A

Date/Time 10/31/25 12:05
Raw Sample Received by: RJ (Ext-lab)
Raw Sample Relinquished by: SH

Date/Time 10/31/25 12:20
Raw Sample Received by: SH
Raw Sample Relinquished by: RJ (Ext-lab)

Alliance TECHNICAL GROUP

EXTRACTION LOGPAGE

PB170497

SOP ID: M3541-ASE Extraction-15

Clean Up SOP #: Acid Cleanup

Matrix : Solid

Weigh By: EH

Balance check: EH

Balance ID: EX-SC-2

pH Strip Lot#: N/A

Extraction By: RJ

Filter By: RJ

pH Meter ID: N/A

Hood ID: 3,7

Extraction Start Date : 11/12/2025

Extraction Start Time : 08:35

Extraction End Date : 11/12/2025

Extraction End Time : 11:45

Concentration By: EH

Supervisor By : RUPESH

Extraction Method: [ ] Seperatory Funnel [ ] Continious Liquid/Liquid [ ] Sonication [ ] Waste Dilution [√] Soxhlet

Standared Name	MLS USED	Concentration ug/mL	STD REF. # FROM LOG
Spike Sol 1	1.0ML	5000 PPB	PP24930
Surrogate	1.0ML	200 PPB	PP24663
N/A	N/A	N/A	N/A
N/A	N/A	N/A	N/A
N/A	N/A	N/A	N/A

Chemical Used	ML/SAMPLE USED	Lot Number
Hexane/Acetone/1:1	N/A	EP2660
Baked Na2SO4	N/A	EP2655
Hexane	N/A	E3981
H2SO4 1:1	N/A	EP2657
Sand	N/A	E3951
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

Extraction Conformance/Non-Conformance Comments:

40ML Vial Lot # 03-40BTS721. Q3450-07 Extracted out of holding time.

KD Bath ID: N/A

KD Bath Temperature: N/A

Envap ID: N-EVAP-02

Envap Temperature: 40 °C

Date / Time	Prepped Sample Relinquished By/Location	Received By/Location
11/12/25	RS (EX-06)	T.P. PREP-IP-3
11:50	Preparation Group	Analysis Group

Alliance TECHNICAL GROUP

EXTRACTION LOGPAGE

PrepBatch ID : PB170497

Analytical Method: M3541-ASE Extraction-15 **Concentration Date:** 11/12/2025

Sample ID	Client Sample ID	Test	g mL	PH	Surr/Spike By:		Final Vol. (mL)	JarID	Comments	Prep Pos
					AddedBy	VerifiedBy				
PB170497BL	ABLK497	PCB	30.01	N/A	ritesh	Evelyn	10			U3-1
PB170497BS	ALCS497	PCB	30.03	N/A	ritesh	Evelyn	10			2
Q3450-07	SS-13R-(4.5-5.0)	PCB	30.08	N/A	ritesh	Evelyn	10			3
Q3597-01	TP-1	PCB	30.06	N/A	ritesh	Evelyn	10	E		4
Q3597-01MS	TP-1MS	PCB	30.04	N/A	ritesh	Evelyn	10	E		5
Q3597-01MSD	TP-1MSD	PCB	30.08	N/A	ritesh	Evelyn	10	E		6
Q3597-04	TP-2	PCB	30.04	N/A	ritesh	Evelyn	10	E		U4-1
Q3597-07	TP-3	PCB	30.05	N/A	ritesh	Evelyn	10	E		2
Q3597-10	TP-4	PCB	30.03	N/A	ritesh	Evelyn	10	E		3
Q3598-01	EO-CONCRETE	PCB	30.07	N/A	ritesh	Evelyn	10	D	Concrete	4
Q3599-01	LIVINGSTON-SOIL	PCB	30.06	N/A	ritesh	Evelyn	10	E	Concrete	5
Q3600-01	CENTRAL-ROW-CONCRETE	PCB	30.09	N/A	ritesh	Evelyn	10	B	Concrete	6
Q3601-01	TP-1	PCB	30.04	N/A	ritesh	Evelyn	10	E		U6-1
Q3601-05	TP-2	PCB	30.08	N/A	ritesh	Evelyn	10	E		2
Q3601-09	TP-3	PCB	30.09	N/A	ritesh	Evelyn	10	E		3
Q3601-13	TP-4	PCB	30.05	N/A	ritesh	Evelyn	10	E		4
Q3601-17	TP-5	PCB	30.04	N/A	ritesh	Evelyn	10	E		5
Q3604-01	S-1	PCB	30.03	N/A	ritesh	Evelyn	10	A		6
Q3604-02	S-2	PCB	30.06	N/A	ritesh	Evelyn	10	A		U1-1

RS
11/12

* **Extracts relinquished on the same date as received.**

170497
8:35

WORKLIST(Hardcopy Internal Chain)

WorkList Name : Q3598 **WorkList ID :** 193057 **Department :** Extraction **Date :** 11-12-2025 08:29:40

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q3450-07	SS-13R-(4.5-5.0)	Solid	PCB	Cool 4 deg C	MATR02	D41	10/23/2025	8082A
Q3604-01	S-1	Solid	PCB	Cool 4 deg C	ROMA02	D41	11/10/2025	8082A
Q3604-02	S-2	Solid	PCB	Cool 4 deg C	ROMA02	D41	11/10/2025	8082A
Q3597-01	TP-1	Solid	PCB	Cool 4 deg C	PSEG03	D41	11/10/2025	8082A
Q3597-04	TP-2	Solid	PCB	Cool 4 deg C	PSEG03	D41	11/10/2025	8082A
Q3597-07	TP-3	Solid	PCB	Cool 4 deg C	PSEG03	D41	11/10/2025	8082A
Q3597-10	TP-4	Solid	PCB	Cool 4 deg C	PSEG03	D41	11/10/2025	8082A
Q3598-01	EO-CONCRETE	Solid	PCB	Cool 4 deg C	PSEG03	D41	11/10/2025	8082A
Q3599-01	LIVINGSTON-SOIL	Solid	PCB	Cool 4 deg C	PSEG03	D41	11/10/2025	8082A
Q3600-01	CENTRAL-ROW-CONCRETE	Solid	PCB	Cool 4 deg C	PSEG03	D41	11/10/2025	8082A
Q3601-01	TP-1	Solid	PCB	Cool 4 deg C	PSEG03	D41	11/10/2025	8082A
Q3601-05	TP-2	Solid	PCB	Cool 4 deg C	PSEG03	D41	11/10/2025	8082A
Q3601-09	TP-3	Solid	PCB	Cool 4 deg C	PSEG03	D41	11/10/2025	8082A
Q3601-13	TP-4	Solid	PCB	Cool 4 deg C	PSEG03	D41	11/10/2025	8082A
Q3601-17	TP-5	Solid	PCB	Cool 4 deg C	PSEG03	D41	11/10/2025	8082A

Date/Time 11/12/25 8:30
Raw Sample Received by: RJ (EA-lab)
Raw Sample Relinquished by: JDCSM

Date/Time 11/12/25 8:55
Raw Sample Received by: JDCSM
Raw Sample Relinquished by: RJ (EA lab)

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

Prep Standard - Chemical Standard Summary

Order ID : Q3450

Test : PCB

Prepbatch ID : PB170236,PB170358,

Sequence ID/Qc Batch ID: pp102425,PP102725,pp103125,Pp110325,PQ102725,

Standard ID :

EP2610,EP2643,EP2652,EP2655,PP24329,PP24330,PP24381,PP24663,PP24805,PP24806,PP24807,PP24808,PP24809,PP24810,PP24811,PP24812,PP24814,PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862,PP24930,

Chemical ID :

E3804,E3875,E3944,E3951,E3962,E3963,E3971,E3972,E3978,E3981,M6157,P 11522,P11592,P12700,P12703,P12952,P12953,P12957,P13356,P13373,P13375,P13590,P13592,P13695,P13697,P13706,P13786,P13789,P13833,P13879,P13881,P13885,W3112,W3177,W3234,

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Extractions STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
314	1.1 H2SO4 SOLN	EP2610	05/07/2025	11/07/2025	RUPESHKUMAR SHAH	Extraction_SCALE_2 (EX-SC-2)	None	Riteshkumar Patel 05/07/2025

FROM 1000.00000ml of M6157 + 1000.00000ml of W3112 = Final Quantity: 2000.000 ml

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
230	1:1ACETONE/HEXANE	EP2643	09/16/2025	03/16/2026	Evelyn Huang	None	None	Riteshkumar Patel 09/16/2025

FROM 4000.00000ml of E3971 + 4000.00000ml of E3972 = Final Quantity: 8000.000 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Extractions STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
3923	Baked Sodium Sulfate	EP2652	10/10/2025	01/28/2026	Evelyn Huang	Extraction_SCALE_2 (EX-SC-2)	None	Riteshkumar Patel 10/10/2025

FROM 4000.00000gram of E3875 = Final Quantity: 4000.000 gram

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
3923	Baked Sodium Sulfate	EP2655	10/24/2025	01/28/2026	RUPESHKUMAR SHAH	Extraction_SCALE_2 (EX-SC-2)	None	Riteshkumar Patel 10/24/2025

FROM 4000.00000gram of E3875 = Final Quantity: 4000.000 gram

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Pest/Pcb STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
84	Pest/PCB Surrogate Stock 20 PPM	PP24329	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza 04/03/2025

FROM 1.00000ml of P13356 + 9.00000ml of W3177 = Final Quantity: 10.000 ml

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
202	AR1660 1000/100 ppb working solution 1st source	PP24330	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza 04/03/2025

FROM 0.10000ml of P13697 + 99.40000ml of W3177 + 0.50000ml of PP24329 = Final Quantity: 100.000 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Pest/Pcb STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
1893	AR1254 1000 PPB Working Sol. 2nd Source	PP24381	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza 04/03/2025

FROM 1.00000ml of P12957 + 98.50000ml of W3177 + 0.50000ml of PP24329 = Final Quantity: 100.000 ml

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
465	200 PPB Pest/PCB Surrogate Spike	PP24663	06/24/2025	12/24/2025	Abdul Mirza	None	None	Yogesh Patel 07/21/2025

FROM 1.00000ml of P13786 + 999.00000ml of E3944 = Final Quantity: 1000.000 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Pest/Pcb STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
84	Pest/PCB Surrogate Stock 20 PPM	PP24805	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza 08/19/2025

FROM 1.00000ml of P13789 + 9.00000ml of E3962 = Final Quantity: 10.000 ml

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
213	AR1221 1000 PPB WORKING SOLUTION	PP24806	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza 08/19/2025

FROM 0.10000ml of P13375 + 99.40000ml of E3962 + 0.50000ml of PP24805 = Final Quantity: 100.000 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Pest/Pcb STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
214	AR1232 1000 PPB WORKING SOLUTION	PP24807	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza 08/19/2025

FROM 0.10000ml of P13879 + 99.40000ml of E3962 + 0.50000ml of PP24805 = Final Quantity: 100.000 ml

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
215	AR1242 1000 PPB WORKING STD	PP24808	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza 08/19/2025

FROM 0.10000ml of P13706 + 99.40000ml of E3962 + 0.50000ml of PP24805 = Final Quantity: 100.000 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Pest/Pcb STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
216	AR1248 1000 PPB WORKING STD	PP24809	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza 08/19/2025

FROM 0.10000ml of P13695 + 99.40000ml of E3962 + 0.50000ml of PP24805 = Final Quantity: 100.000 ml

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
217	AR1254 1000 PPB WORKING STD	PP24810	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza 08/19/2025

FROM 0.10000ml of P11592 + 99.40000ml of E3962 + 0.50000ml of PP24805 = Final Quantity: 100.000 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Pest/Pcb STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
1529	AR1262 1000 PPB Working Solution	PP24811	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza 08/19/2025

FROM 0.10000ml of P13881 + 99.40000ml of E3962 + 0.50000ml of PP24805 = Final Quantity: 100.000 ml

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
1532	AR1268 1000 PPB Working Solution	PP24812	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza 08/19/2025

FROM 0.10000ml of P13885 + 99.40000ml of E3962 + 0.50000ml of PP24805 = Final Quantity: 100.000 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Pest/Pcb STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
404	AR1660 100 PPM Stock Solution 2nd Source	PP24814	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza 08/19/2025

FROM 1.00000ml of P12952 + 9.00000ml of E3804 = Final Quantity: 10.000 ml

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
405	AR1660 1000/100 PPB ICV STD	PP24815	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza 08/19/2025

FROM 98.50000ml of E3962 + 0.50000ml of PP24805 + 1.00000ml of PP24814 = Final Quantity: 100.000 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Pest/Pcb STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
3789	AR1221 1000 PPB WORKING SOL.2ND SOURCE(AGILENT)	PP24816	08/18/2025	02/15/2026	Yogesh Patel	None	None	Abdul Mirza 08/19/2025

FROM 1.00000ml of P13373 + 98.50000ml of E3962 + 0.50000ml of PP24805 = Final Quantity: 100.000 ml

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
1887	AR1232 1000 PPB Working Sol. 2nd Source	PP24817	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza 08/19/2025

FROM 1.00000ml of P12700 + 98.50000ml of E3962 + 0.50000ml of PP24805 = Final Quantity: 100.000 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Pest/Pcb STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
1889	AR1242 1000 PPB Working Sol. 2nd Source	PP24818	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza 08/19/2025

FROM 1.00000ml of P13590 + 98.50000ml of E3962 + 0.50000ml of PP24805 = Final Quantity: 100.000 ml

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
1890	AR1248 1000 PPB Working Sol. 2nd Source	PP24819	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza 08/19/2025

FROM 1.00000ml of P13592 + 98.50000ml of E3962 + 0.50000ml of PP24805 = Final Quantity: 100.000 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Pest/Pcb STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
2193	AR1254 100 PPM STOCK SOLUTION	PP24820	08/18/2025	02/15/2026	Yogesh Patel	None	None	Abdul Mirza 08/19/2025

FROM 1.00000ml of P13833 + 9.00000ml of E3963 = Final Quantity: 10.000 ml

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
3757	AR1262 1000 PPB Working Solution second source	PP24821	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza 08/19/2025

FROM 1.00000ml of P12703 + 98.50000ml of E3962 + 0.50000ml of PP24805 = Final Quantity: 100.000 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Pest/Pcb STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
3817	AR1268 1000 ppb Working Soln. 2nd source	PP24822	08/18/2025	02/15/2026	Yogesh Patel	None	None	Abdul Mirza 08/19/2025

FROM 1.00000ml of P11522 + 98.50000ml of E3962 + 0.50000ml of PP24805 = Final Quantity: 100.000 ml

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
203	AR1660 750 PPB STD	PP24823	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza 08/19/2025

FROM 0.25000ml of W3234 + 0.75000ml of PP24330 = Final Quantity: 1.000 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Pest/Pcb STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
204	AR1660 500 PPB STD	PP24824	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza 08/19/2025

FROM 0.50000ml of W3234 + 0.50000ml of PP24330 = Final Quantity: 1.000 ml

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
205	AR1660 250 PPB STD	PP24825	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza 08/19/2025

FROM 0.75000ml of W3234 + 0.25000ml of PP24330 = Final Quantity: 1.000 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Pest/Pcb STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
206	AR1660 50 PPB STD	PP24826	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza 08/19/2025

FROM 0.90000ml of W3234 + 0.10000ml of PP24824 = Final Quantity: 1.000 ml

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
1079	AR1221 750 PPB STD	PP24827	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza 08/19/2025

FROM 0.25000ml of W3234 + 0.75000ml of PP24806 = Final Quantity: 1.000 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Pest/Pcb STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
222	AR1221 500 PPB STD	PP24828	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza 08/19/2025

FROM 0.50000ml of W3234 + 0.50000ml of PP24806 = Final Quantity: 1.000 ml

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
1080	AR1221 250 PPB STD	PP24829	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza 08/19/2025

FROM 0.75000ml of W3234 + 0.25000ml of PP24806 = Final Quantity: 1.000 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Pest/Pcb STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
1081	AR1221 50 PPB STD	PP24830	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza 08/19/2025

FROM 0.90000ml of W3234 + 0.10000ml of PP24828 = Final Quantity: 1.000 ml

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
1063	AR1232 750 PPB STD	PP24831	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza 08/19/2025

FROM 0.25000ml of W3234 + 0.75000ml of PP24807 = Final Quantity: 1.000 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Pest/Pcb STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
223	AR1232 500 PPB STD	PP24832	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza 08/19/2025

FROM 0.50000ml of W3234 + 0.50000ml of PP24807 = Final Quantity: 1.000 ml

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
1064	AR1232 250 PPB STD	PP24833	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza 08/19/2025

FROM 0.75000ml of W3234 + 0.25000ml of PP24807 = Final Quantity: 1.000 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Pest/Pcb STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
1065	AR1232 50 PPB STD	PP24834	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza 08/19/2025

FROM 0.90000ml of W3234 + 0.10000ml of PP24832 = Final Quantity: 1.000 ml

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
1067	AR1242 750 PPB STD	PP24835	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza 08/19/2025

FROM 0.25000ml of W3234 + 0.75000ml of PP24808 = Final Quantity: 1.000 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Pest/Pcb STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
224	AR1242 500 PPB STD	PP24836	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza 08/19/2025

FROM 0.50000ml of W3234 + 0.50000ml of PP24808 = Final Quantity: 1.000 ml

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
1068	AR1242 250 PPB STD	PP24837	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza 08/19/2025

FROM 0.75000ml of W3234 + 0.25000ml of PP24808 = Final Quantity: 1.000 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Pest/Pcb STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
1069	AR1242 50 PPB STD	PP24838	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza 08/19/2025

FROM 0.90000ml of W3234 + 0.10000ml of PP24836 = Final Quantity: 1.000 ml

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
1075	AR1248 750 PPB STD	PP24839	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza 08/19/2025

FROM 0.25000ml of W3234 + 0.75000ml of PP24809 = Final Quantity: 1.000 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Pest/Pcb STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
225	AR1248 500 PPB STD	PP24840	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza 08/19/2025

FROM 0.50000ml of W3234 + 0.50000ml of PP24809 = Final Quantity: 1.000 ml

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
1076	AR1248 250 PPB STD	PP24841	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza 08/19/2025

FROM 0.75000ml of W3234 + 0.25000ml of PP24809 = Final Quantity: 1.000 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Pest/Pcb STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
1077	AR1248 50 PPB STD	PP24842	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza 08/19/2025

FROM 0.90000ml of W3234 + 0.10000ml of PP24840 = Final Quantity: 1.000 ml

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
1071	AR1254 750 PPB STD	PP24843	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza 08/19/2025

FROM 0.25000ml of W3234 + 0.75000ml of PP24810 = Final Quantity: 1.000 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Pest/Pcb STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
226	AR1254 500 PPB STD	PP24844	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza 08/19/2025

FROM 0.50000ml of W3234 + 0.50000ml of PP24810 = Final Quantity: 1.000 ml

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
1072	AR1254 250 PPB STD	PP24845	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza 08/19/2025

FROM 0.75000ml of W3234 + 0.25000ml of PP24810 = Final Quantity: 1.000 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Pest/Pcb STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
1073	AR1254 50 PPB STD	PP24846	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza 08/19/2025

FROM 0.90000ml of W3234 + 0.10000ml of PP24844 = Final Quantity: 1.000 ml

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
3753	AR1262 750 PPB STD	PP24847	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza 08/19/2025

FROM 0.25000ml of W3234 + 0.75000ml of PP24811 = Final Quantity: 1.000 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Pest/Pcb STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
1530	AR1262 500 PPB STD	PP24848	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza 08/19/2025

FROM 0.50000ml of W3234 + 0.50000ml of PP24811 = Final Quantity: 1.000 ml

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
3754	AR1262 250 PPB STD	PP24849	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza 08/19/2025

FROM 0.75000ml of W3234 + 0.25000ml of PP24811 = Final Quantity: 1.000 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Pest/Pcb STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
3755	AR1262 50 PPB STD	PP24850	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza 08/19/2025

FROM 0.90000ml of W3234 + 0.10000ml of PP24848 = Final Quantity: 1.000 ml

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
3820	AR1268 750 PPB STD	PP24851	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza 08/19/2025

FROM 0.25000ml of W3234 + 0.75000ml of PP24812 = Final Quantity: 1.000 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Pest/Pcb STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
1533	AR1268 500 PPB STD	PP24852	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza 08/19/2025

FROM 0.50000ml of W3234 + 0.50000ml of PP24812 = Final Quantity: 1.000 ml

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
3821	AR1268 250 PPB STD	PP24853	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza 08/19/2025

FROM 0.75000ml of W3234 + 0.25000ml of PP24812 = Final Quantity: 1.000 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Pest/Pcb STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
3822	AR1268 50 PPB STD	PP24854	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza 08/19/2025

FROM 0.90000ml of W3234 + 0.10000ml of PP24852 = Final Quantity: 1.000 ml

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
406	AR1660 500 PPB ICV	PP24855	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza 08/19/2025

FROM 0.50000ml of W3234 + 0.50000ml of PP24815 = Final Quantity: 1.000 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Pest/Pcb STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
3790	AR1221 500 PPB ICV(AGILENT)	PP24856	08/18/2025	02/15/2026	Yogesh Patel	None	None	Abdul Mirza 08/19/2025

FROM 0.50000ml of W3234 + 0.50000ml of PP24816 = Final Quantity: 1.000 ml

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
1888	AR1232 500 PPB ICV	PP24857	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza 08/19/2025

FROM 0.50000ml of W3234 + 0.50000ml of PP24817 = Final Quantity: 1.000 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Pest/Pcb STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
1891	AR1242 500 PPB ICV	PP24858	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza 08/19/2025

FROM 0.50000ml of W3234 + 0.50000ml of PP24818 = Final Quantity: 1.000 ml

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
1892	AR1248 500 PPB ICV	PP24859	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza 08/19/2025

FROM 0.50000ml of W3234 + 0.50000ml of PP24819 = Final Quantity: 1.000 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Pest/Pcb STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
1894	AR1254 500 PPB ICV	PP24860	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza 08/19/2025

FROM 0.50000ml of W3234 + 0.50000ml of PP24381 = Final Quantity: 1.000 ml

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
3758	AR1262 500 PPB STD ICV	PP24861	08/18/2025	02/18/2026	Yogesh Patel	None	None	Abdul Mirza 08/19/2025

FROM 0.50000ml of W3234 + 0.50000ml of PP24821 = Final Quantity: 1.000 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Pest/Pcb STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
3823	AR1268 500 PPB STD ICV	PP24862	08/18/2025	02/15/2026	Yogesh Patel	None	None	Abdul Mirza 08/19/2025

FROM 0.50000ml of W3234 + 0.50000ml of PP24822 = Final Quantity: 1.000 ml

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
3857	5000 PPB PCB SPIKE SOLUTION 2ND SOURCE	PP24930	09/18/2025	03/18/2026	Abdul Mirza	None	None	Yogesh Patel 09/19/2025

FROM 0.50000ml of P12953 + 99.50000ml of E3972 = Final Quantity: 100.000 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	9005-05 / Acetone Ultra (cs/4x4L)	24E0761004	02/18/2026	10/01/2024 / Rajesh	09/25/2024 / Rajesh	E3804

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	PC19631-100 / SODIUM SULFATE, ANHYDROUS, PEST GRADE, 1	417203	07/28/2026	07/28/2025 / RUPESH	01/29/2025 / Rajesh	E3875

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9254-03 / Acetone, Ultra Resi (cs/4x4L)	24H1462005	05/24/2027	06/20/2025 / RUPESH	05/14/2025 / RUPESH	E3944

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-3382-05 / Sand, Purified (cs/4x2.5kg)	25A2756718	12/31/2028	07/09/2025 / RUPESH	04/28/2020 / RUPESH	E3951

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9262-03 / Hexane, Ultra-Resi (cs/4x4L)	25C0362005	04/30/2026	08/05/2025 / RUPESH	07/30/2025 / RUPESH	E3962

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9254-03 / Acetone, Ultra Resi (cs/4x4L)	24H2762008	02/15/2026	08/15/2025 / RUPESH	08/07/2025 / RUPESH	E3963

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9262-03 / Hexane, Ultra-Resi (cs/4x4L)	25C0362005	04/30/2026	09/15/2025 / Evelyn	09/04/2025 / Riteshkumar	E3971

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9254-03 / Acetone, Ultra Resi (cs/4x4L)	24H1462005	05/24/2027	09/16/2025 / Evelyn	09/04/2025 / Riteshkumar	E3972

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9262-03 / Hexane, Ultra-Resi (cs/4x4L)	25C0362006	04/13/2026	10/13/2025 / RUPESH	10/08/2025 / Riteshkumar	E3978

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9262-03 / Hexane, Ultra-Resi (cs/4x4L)	25C0362006	04/30/2026	10/27/2025 / RUPESH	10/22/2025 / RUPESH	E3981

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9673-33 / Sulfuric Acid, Instra-Analyzed (cs/6c2.5L)	24i1262013	11/07/2025	05/07/2025 / RUPESH	02/18/2025 / Mohan	M6157

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Agilent Technologies	PP-382-1 / Aroclor 1268	0006587800	09/18/2025	03/18/2025 / yogesh	02/21/2022 / Ankita	P11522

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32011 / PCB Mix, Aroclor 1254, 1000ug/mL, Hexane, 1mL/ampul	A0175403	02/18/2026	08/18/2025 / yogesh	03/18/2022 / Abdul	P11592

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards,Inc	91867 / Aroclor 1232 100 ug/mL	020823	02/18/2026	08/18/2025 / yogesh	08/07/2023 / Ankita	P12700

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards,Inc	x9166 / Aroclor 1262 100 ug/mL	060523	02/18/2026	08/18/2025 / yogesh	08/07/2023 / Ankita	P12703

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	20064 / Aroclor 1016/1260	022023	02/18/2026	08/18/2025 / yogesh	12/20/2023 / Yogesh	P12952

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	20064 / Aroclor 1016/1260	022023	03/18/2026	09/18/2025 / Abdul	12/20/2023 / Yogesh	P12953

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	/ Arochlor 1254	121823	04/03/2025	10/03/2024 / Ankita	12/20/2023 / Yogesh	P12957

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32000 / Pesticide Mix, CLP method, Pesticide Surrogate Mix, 200ug/mL, Acetone, 1mL	A0206810	09/18/2025	03/18/2025 / yogesh	04/22/2024 / Abdul	P13356

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Agilent Technologies	PP-292-1 / Aroclor 1221	0006783205	09/18/2025	03/18/2025 / yogesh	05/02/2024 / Ankita	P13373

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32007 / PCB Mix, Aroclor 1221, 1000ug/mL, Hexane, 1mL/ampul	A0207511	02/18/2026	08/18/2025 / yogesh	05/03/2024 / Abdul	P13375

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Agilent Technologies	PP-312-1 / Aroclor 1242	0006665550	02/18/2026	08/18/2025 / yogesh	10/14/2024 / Ankita	P13590

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Agilent Technologies	PP-342-1 / Aroclor 1248	0006726317	02/18/2026	08/18/2025 / yogesh	10/14/2024 / Ankita	P13592

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32010 / PCB Mix, Aroclor 1248, 1000ug/mL, Hexane, 1mL/ampul	A0214733	02/18/2026	08/18/2025 / yogesh	10/17/2024 / yogesh	P13695

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32039 / PCB Mix, Aroclor 1016/1260, 1000ug/mL, hexane, 1mL/ampul	A0210629	09/18/2025	03/18/2025 / yogesh	10/17/2024 / yogesh	P13697

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32009 / PCB Mix, Aroclor 1242, 1000ug/mL, Hexane, 1mL/ampul	A0210232	02/18/2026	08/18/2025 / yogesh	10/17/2024 / yogesh	P13706

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32000 / Pesticide Mix, CLP method, Pesticide Surrogate Mix, 200ug/mL, Acetone, 1mL	A0214495	12/24/2025	06/24/2025 / Abdul	11/19/2024 / Ankita	P13786

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32000 / Pesticide Mix, CLP method, Pesticide Surrogate Mix, 200ug/mL, Acetone, 1mL	A0214495	02/18/2026	08/18/2025 / yogesh	11/19/2024 / Ankita	P13789

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32011 / PCB Mix, Aroclor 1254, 1000ug/mL, Hexane, 1mL/ampul	a0217391	02/18/2026	08/18/2025 / yogesh	12/09/2024 / Ankita	P13833

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32008 / PCB Mix, Aroclor 1232, 1000ug/mL, Hexane, 1mL/ampul	A0219655	02/18/2026	08/18/2025 / yogesh	01/23/2025 / Ankita	P13879

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32409 / PCB Stock Solution, Aroclor 1262 Std, 1mL, Hexane	A0217041	02/18/2026	08/18/2025 / yogesh	01/23/2025 / Ankita	P13881

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32410 / PCB Stock Solution, Aroclor 1268 Std, 1mL, Hexane	A0217264	02/18/2026	08/18/2025 / yogesh	01/23/2025 / Ankita	P13885

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	DIW / DI Water	Daily Lab-Certified	07/03/2029	07/03/2024 / Iwona	07/03/2024 / Iwona	W3112

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9262-03 / Hexane, Ultra-Resi (cs/4x4L)	24G1962003	08/22/2025	02/03/2025 / jignesh	01/31/2025 / jignesh	W3177

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9262-03 / Hexane, Ultra-Resi (cs/4x4L)	25C0362005	04/30/2026	07/28/2025 / jignesh	07/25/2025 / jignesh	W3234

Acetone
CMOS

avantor™ | J.T.Baker®

Material No.: 9005-05
Batch No.: 24E0761004
Manufactured Date: 2024-05-02
Retest Date: 2029-05-01
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
Assay ($(CH_3)_2CO$) (by GC, corrected for water)	≥ 99.5 %	99.8 %
Color (APHA)	≤ 10	< 5
Residue after Evaporation	≤ 5 ppm	< 1 ppm
Titrable Acid (µeq/g)	≤ 0.3	0.1
Titrable Base (µeq/g)	≤ 0.5	0.1
Water (H_2O)	≤ 0.5 %	0.1 %
Solubility in H_2O	Passes Test	Passes Test
Chloride (Cl)	≤ 0.2 ppm	< 0.2 ppm
Phosphate (PO_4)	≤ 0.05 ppm	< 0.05 ppm
Trace Impurities - Aluminum (Al)	≤ 50.0 ppb	< 5.0 ppb
Arsenic and Antimony (as As)	≤ 5.0 ppb	< 5.0 ppb
Trace Impurities - Barium (Ba)	≤ 20.0 ppb	< 1.0 ppb
Trace Impurities - Beryllium (Be)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities - Bismuth (Bi)	≤ 20.0 ppb	< 10.0 ppb
Trace Impurities - Boron (B)	≤ 10.0 ppb	< 5.0 ppb
Trace Impurities - Cadmium (Cd)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities - Calcium (Ca)	≤ 25.0 ppb	3.6 ppb
Trace Impurities - Chromium (Cr)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities - Cobalt (Co)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities - Copper (Cu)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities - Gallium (Ga)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities - Germanium (Ge)	≤ 10.0 ppb	< 10.0 ppb
Trace Impurities - Gold (Au)	≤ 20 ppb	< 5 ppb
Trace Impurities - Iron (Fe)	≤ 20.0 ppb	< 1.0 ppb
Trace Impurities - Lead (Pb)	≤ 10.0 ppb	< 10.0 ppb
Trace Impurities - Lithium (Li)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities - Magnesium (Mg)	≤ 20 ppb	< 1 ppb
Trace Impurities - Manganese (Mn)	≤ 10.0 ppb	< 1.0 ppb

>>> Continued on page 2 >>>

Recd by RP on 9/25/24

E 3804

Acetone
CMOS

avantor™ | J.T.Baker®

Material No.: 9005-05
Batch No.: 24E0761004

Test	Specification	Result
Trace Impurities – Molybdenum (Mo)	≤ 10.0 ppb	< 5.0 ppb
Trace Impurities – Nickel (Ni)	≤ 10.0 ppb	< 5.0 ppb
Trace Impurities – Niobium (Nb)	≤ 50.0 ppb	< 1.0 ppb
Trace Impurities – Potassium (K)	≤ 10.0 ppb	< 10.0 ppb
Trace Impurities – Silicon (Si)	≤ 50 ppb	< 10 ppb
Trace Impurities – Silver (Ag)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities – Sodium (Na)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities – Strontium (Sr)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities – Tantalum (Ta)	≤ 50.0 ppb	< 5.0 ppb
Trace Impurities – Thallium (Tl)	≤ 10.0 ppb	< 5.0 ppb
Trace Impurities – Tin (Sn)	≤ 20.0 ppb	< 10.0 ppb
Trace Impurities – Titanium (Ti)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities – Vanadium (V)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities – Zinc (Zn)	≤ 20.0 ppb	7.9 ppb
Trace Impurities – Zirconium (Zr)	≤ 10.0 ppb	< 1.0 ppb
Particle Count – 0.5 µm and greater (Rion KS42AF)	≤ 100 par/ml	8 par/ml
Particle Count – 1.0 µm and greater (Rion KS42AF)	≤ 8 par/ml	2 par/ml

>>> Continued on page 3 >>>

Acetone
CMOS

avantor™ | J.T.Baker®

Material No.: 9005-05
Batch No.: 24E0761004

Test	Specification	Result

For Microelectronic Use

Country of Origin: USA
Packaging Site: Paris Mfg Ctr & DC

Michelle Bales
Sr. Manager, Quality Assurance

PRODUCTOS QUIMICOS MONTERREY, S.A. DE C.V.

MIRADOR 201, COL. MIRADOR
MONTERREY, N.L. MÉXICO
CP 64070
TEL +52 81 13 52 57 57
www.pqm.com.mx

CERTIFICATE OF ANALYSIS

PRODUCT :	SODIUM SULFATE CRYSTALS ANHYDROUS		
QUALITY :	ACS (CODE **RMB3375)**	FORMULA :	Na_2SO_4
SPECIFICATION NUMBER:	6399	RELEASE DATE:	MAY/23/2024
LOT NUMBER :	417203		

TEST	SPECIFICATIONS	LOT VALUES
Assay (Na_2SO_4)	Min. 99.0%	99.8 %
pH of a 5% solution at 25°C	5.2 - 9.2	6.2
Insoluble matter	Max. 0.01%	0.001 %
Loss on ignition	Max. 0.5%	0.1 %
Chloride (Cl)	Max. 0.001%	<0.001 %
Nitrogen compounds (as N)	Max. 5 ppm	<5 ppm
Phosphate (PO_4)	Max. 0.001%	<0.001 %
Heavy metals (as Pb)	Max. 5 ppm	<5 ppm
Iron (Fe)	Max. 0.001%	<0.001 %
Calcium (Ca)	Max. 0.01%	0.001 %
Magnesium (Mg)	Max. 0.005%	0.001 %
Potassium (K)	Max. 0.008%	0.001 %
Extraction-concentration suitability	Passes test	Passes test
Appearance	Passes test	Passes test
Identification	Passes test	Passes test
Solubility and foreing matter	Passes test	Passes test
Retained on US Standard No. 10 sieve	Max. 1%	0.2 %
Retained on US Standard No. 60 sieve	Min. 94%	96.2 %
Through US Standard No. 60 sieve	Max. 5%	3.5 %
Through US Standard No. 100 sieve	Max. 10%	0.1 %

COMMENTS

QC: PhC Irma Belmares

If you need further details, please call our factory or contact our local distributor.

E 3875

Acetone
BAKER RESI-ANALYZED® Reagent
For Organic Residue Analysis

avantor™ | J.T.Baker

Material No.: 9254-03
Batch No.: 24H1462005
Manufactured Date: 2024-05-24
Expiration Date:2027-05-24
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
Assay ($(CH_3)_2CO$) (by GC, corrected forwater)	>= 99.4 %	99.8 %
Color (APHA)	<= 10	5
Residue after Evaporation	<= 1.0 ppm	0.2 ppm
Substances Reducing Permanganate	Passes Test	Passes Test
Titrable Acid (µeq/g)	<= 0.3	0.2
Titrable Base (µeq/g)	<= 0.6	<0.1
Water (H_2O)	<= 0.5 %	0.2 %
FID-Sensitive Impurities (as 2-Octanol)Single Impurity Peak (ng/mL)	<= 5	<1
ECD Sensitive Impurities (as HeptachlorEpoxide) Single Peak (pg/mL)	<= 10	1

For Laboratory,Research,or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: United States
Packaging Site: Phillipsburg Mfg Ctr & DC

E3944

JCroak

Jamie Croak
Director Quality Operations, Bioscience Production

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

VWR®
BDH® CHEMICALS

Certificate of Analysis

Material	BDH9274-2.5KG
Material Description	BDH SAND STDD OTTAWA W+I 2.5KG
Grade	NOT APPLICABLE
Batch	25A2756718
Reassay Date	12/31/2028
CAS Number	14808-60-7
Molecular Formula	SiO2
Molecular Mass	60.09
Date of Manufacture	12/05/2024
Storage	Room Temperature

Characteristics	Specifications	Measured Values
Appearance	Beige granules.	Beige granules.
Moisture	<= 0.1 %	0.1 %
Particle Size 30-40 mesh	>= 80 %	99 %
CUSTOMER PART # BDH9274-2.5KG		

Received on 7/9/25.

E3951

Internal ID #: 793

Signature	Additional Information
We certify that this batch conforms to the specifications listed above.	Analysis may have been rounded to significant digits in specification limits
This document has been electronically produced and is valid without a signature.	Product meets analytical specifications of the grades listed.
Leona Edwardson, Quality Control Sr. Manager - Solon VWR Chemicals, LLC. 28600 Fountain Parkway, Solon OH 44139 USA	

n-Hexane 95%
ULTRA RESI-ANALYZED
For Organic Residue Analysis

avantor™ | J.T.Baker

Material No.: 9262-03
Batch No.: 25C0362005
Manufactured Date: 2025-01-29
Expiration Date:2026-04-30
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
FID-Sensitive Impurities (as 2-Octanol)Single Impurity Peak (ng/mL)	<= 5	1
ECD Sensitive Impurities (as HeptachlorEpoxide) Single Peak (pg/mL)	<= 10	6
ECD-Sensitive Impurities (as EthyleneDibromide) - Single Impurity Peak (ng/mL)	<= 5	5
Assay (Total Saturated C_6 Isomers) (byGC, corrected for water)	>= 99.5 %	100.0 %
Assay (as n-Hexane) (by GC, correctedfor water)	>= 95 %	100 %
Color (APHA)	<= 10	10
Residue after Evaporation	<= 1.0 ppm	0.1 ppm
Substances Darkened by H_2SO_4	Passes Test	Passes Test
Water (by KF, coulometric)	<= 0.05 %	<0.01 %

For Laboratory,Research,or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Recieved on 7/30/25

Country of Origin: United States
Packaging Site: Phillipsburg Mfg Ctr & DC

E3962

JCroak

Jamie Croak
Director Quality Operations, Bioscience Production

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

PO: PO2-2575 PRODUCT CODE: SHIP DATE: 6/20/25

Acetone,
BAKER RESI-ANALYZED® Reagent
For Organic Residue Analysis

avantor™ | J.T.Baker

Material No.: 9254-03
Batch No.: 24H2762008
Manufactured Date: 2024-04-18
Expiration Date:2027-04-18
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
Assay ($(CH_3)_2CO$) (by GC, corrected forwater)	>= 99.4 %	100.0 %
Color (APHA)	<= 10	5
Residue after Evaporation	<= 1.0 ppm	0.0 ppm
Substances Reducing Permanganate	Passes Test	Passes Test
Titrable Acid (µeq/g)	<= 0.3	0.2
Titrable Base (µeq/g)	<= 0.6	<0.1
Water (H_2O)	<= 0.5 %	<0.1 %
FID-Sensitive Impurities (as 2-Octanol)Single Impurity Peak (ng/mL)	<= 5	1
ECD Sensitive Impurities (as HeptachlorEpoxide) Single Peak (pg/mL)	<= 10	1

For Laboratory,Research,or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: United States
Packaging Site: Phillipsburg Mfg Ctr & DC

Recieved on 8/6/25

E 3963

J Croak
Jamie Croak
Director Quality Operations, Bioscience Production

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

Avantor Performance Materials,LLC
100 Matsonford Rd, Suite 200,Radnor,PA,19087.U.S.A.Phone 610.386.1700

n-Hexane 95%
ULTRA RESI-ANALYZED
For Organic Residue Analysis

avantor™ | J.T.Baker

Material No.: 9262-03
Batch No.: 25C0362005
Manufactured Date: 2025-01-29
Expiration Date:2026-04-30
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
FID-Sensitive Impurities (as 2-Octanol)Single Impurity Peak (ng/mL)	<= 5	1
ECD Sensitive Impurities (as HeptachlorEpoxide) Single Peak (pg/mL)	<= 10	6
ECD-Sensitive Impurities (as EthyleneDibromide) - Single Impurity Peak (ng/mL)	<= 5	5
Assay (Total Saturated C_6 Isomers) (byGC, corrected for water)	>= 99.5 %	100.0 %
Assay (as n-Hexane) (by GC, correctedfor water)	>= 95 %	100 %
Color (APHA)	<= 10	10
Residue after Evaporation	<= 1.0 ppm	0.1 ppm
Substances Darkened by H_2SO_4	Passes Test	Passes Test
Water (by KF, coulometric)	<= 0.05 %	<0.01 %

For Laboratory,Research,or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: United States
Packaging Site: Phillipsburg Mfg Ctr & DC

E3971

Jamie Croak
Director Quality Operations, Bioscience Production

Acetone
BAKER RESI-ANALYZED® Reagent
For Organic Residue Analysis

avantor™ | J.T.Baker

Material No.: 9254-03
Batch No.: 24H1462005
Manufactured Date: 2024-05-24
Expiration Date:2027-05-24
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
Assay ($(CH_3)_2CO$) (by GC, corrected forwater)	>= 99.4 %	99.8 %
Color (APHA)	<= 10	5
Residue after Evaporation	<= 1.0 ppm	0.2 ppm
Substances Reducing Permanganate	Passes Test	Passes Test
Titrable Acid (µeq/g)	<= 0.3	0.2
Titrable Base (µeq/g)	<= 0.6	<0.1
Water (H_2O)	<= 0.5 %	0.2 %
FID-Sensitive Impurities (as 2-Octanol)Single Impurity Peak (ng/mL)	<= 5	<1
ECD Sensitive Impurities (as HeptachlorEpoxide) Single Peak (pg/mL)	<= 10	1

For Laboratory,Research,or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: United States
Packaging Site: Phillipsburg Mfg Ctr & DC

E3972

JCroak
Jamie Croak
Director Quality Operations, Bioscience Production

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

n-Hexane 95%
ULTRA RESI-ANALYZED
For Organic Residue Analysis

avantor™ | J.T.Baker

Material No.: 9262-03
Batch No.: 25C0362006
Manufactured Date: 2025-01-29
Expiration Date:2026-04-30
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
FID-Sensitive Impurities (as 2-Octanol)Single Impurity Peak (ng/mL)	<= 5	1
ECD Sensitive Impurities (as HeptachlorEpoxide) Single Peak (pg/mL)	<= 10	6
ECD-Sensitive Impurities (as EthyleneDibromide) - Single Impurity Peak (ng/mL)	<= 5	4
Assay (Total Saturated C_6 Isomers) (byGC, corrected for water)	>= 99.5 %	100.0 %
Assay (as n-Hexane) (by GC, correctedfor water)	>= 95 %	100 %
Color (APHA)	<= 10	10
Residue after Evaporation	<= 1.0 ppm	0.2 ppm
Substances Darkened by H_2SO_4	Passes Test	Passes Test
Water (by KF, coulometric)	<= 0.05 %	<0.01 %

For Laboratory,Research,or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: United States
Packaging Site: Phillipsburg Mfg Ctr & DC

E 3978

J Croak

Jamie Croak
Director Quality Operations, Bioscience Production

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

n-Hexane 95%
ULTRA RESI-ANALYZED
For Organic Residue Analysis

avantor™ | J.T.Baker

Material No.: 9262-03
Batch No.: 25C0362006
Manufactured Date: 2025-01-29
Expiration Date:2026-04-30
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
FID-Sensitive Impurities (as 2-Octanol)Single Impurity Peak (ng/mL)	<= 5	1
ECD Sensitive Impurities (as HeptachlorEpoxide) Single Peak (pg/mL)	<= 10	6
ECD-Sensitive Impurities (as EthyleneDibromide) - Single Impurity Peak (ng/mL)	<= 5	4
Assay (Total Saturated C_6 Isomers) (byGC, corrected for water)	>= 99.5 %	100.0 %
Assay (as n-Hexane) (by GC, correctedfor water)	>= 95 %	100 %
Color (APHA)	<= 10	10
Residue after Evaporation	<= 1.0 ppm	0.2 ppm
Substances Darkened by H_2SO_4	Passes Test	Passes Test
Water (by KF, coulometric)	<= 0.05 %	<0.01 %

For Laboratory,Research,or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Recieved on 10/22/25

Country of Origin: United States
Packaging Site: Phillipsburg Mfg Ctr & DC

E3981

JCroak
Jamie Croak
Director Quality Operations, Bioscience Production

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

Sulfuric Acid
BAKER INSTRA-ANALYZED® Reagent
For Trace Metal Analysis
Low Selenium

avantor | J.T.Baker

M615 MB

Material No.: 9673-33
Batch No.: 24I1262013
Manufactured Date: 2024-08-07
Retest Date:2029-08-06
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
ACS - Assay (H_2SO_4)	95.0 - 98.0 %	96.2 %
Appearance	Passes Test	Passes Test
ACS - Color (APHA)	<= 10	5
ACS - Residue after Ignition	<= 3 ppm	<1 ppm
ACS - Substances Reducing Permanganate(as SO2)	<= 2 ppm	<2 ppm
Ammonium (NH_4)	<= 1 ppm	<1 ppm
Chloride (Cl)	<= 0.1 ppm	<0.1 ppm
Nitrate (NO_3)	<= 0.2 ppm	0.1 ppm
Phosphate (PO_4)	<= 0.5 ppm	<0.1 ppm
Trace Impurities - Aluminum (Al)	<= 30.0 ppb	<5.0 ppb
Arsenic & Antimony (as As)	<= 4.0 ppb	<2.0 ppb
Trace Impurities - Boron (B)	<= 10.0 ppb	<5.0 ppb
Trace Impurities - Cadmium (Cd)	<= 2.0 ppb	<1.0 ppb
Trace Impurities - Chromium (Cr)	<= 6.0 ppb	<1.0 ppb
Trace Impurities - Cobalt (Co)	<= 0.5 ppb	<0.3 ppb
Trace Impurities - Copper (Cu)	<= 1.0 ppb	<1.0 ppb
Trace Impurities - Gold (Au)	<= 10.0 ppb	<5.0 ppb
Heavy Metals (as Pb)	<= 500.0 ppb	<100.0 ppb
Trace Impurities - Iron (Fe)	<= 50.0 ppb	<1.0 ppb
Trace Impurities - Lead (Pb)	<= 0.5 ppb	<0.5 ppb
Trace Impurities - Magnesium (Mg)	<= 7.0 ppb	<1.0 ppb
Trace Impurities - Manganese (Mn)	<= 1.0 ppb	<1.0 ppb
Trace Impurities - Mercury (Hg)	<= 0.5 ppb	<0.1 ppb
Trace Impurities - Nickel (Ni)	<= 2.0 ppb	<0.3 ppb
Trace Impurities - Potassium (K)	<= 500.0 ppb	<10.0 ppb
Trace Impurities - Selenium (Se)	<= 50.0 ppb	7.2 ppb
Trace Impurities - Silicon (Si)	<= 100.0 ppb	12.8 ppb
Trace Impurities - Silver (Ag)	<= 1.0 ppb	<1.0 ppb

Sulfuric Acid
BAKER INSTRA-ANALYZED® Reagent
For Trace Metal Analysis
Low Selenium

avantor™ | J.T.Baker

Material No.: 9673-33
Batch No.: 24I1262013

Test	Specification	Result
Trace Impurities - Sodium (Na)	<= 500.0 ppb	<5.0 ppb
Trace Impurities - Strontium (Sr)	<= 5.0 ppb	<1.0 ppb
Trace Impurities - Tin (Sn)	<= 5.0 ppb	1.1 ppb
Trace Impurities - Zinc (Zn)	<= 5.0 ppb	<1.0 ppb

For Laboratory,Research,or Manufacturing Use

Country of Origin: United States
Packaging Site: Phillipsburg Mfg Ctr & DC

J Croak

Jamie Croak
Director Quality Operations, Bioscience Production

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

P11518 ↓ P11522 AJ 02/21/22

Agilent

Certificate of Analysis

Product Name: Aroclor 1268 Standard

Product Number: PP-382-1 **Lot Issue Date:** 09-Feb-2021

Lot Number: 0006587800 **Expiration Date:** 31-Mar-2029

Description:
This analytical reference material (RM) was manufactured and verified in accordance with an ISO 9001 registered quality system and analyte concentrations were verified by an ISO 17025 accredited laboratory. The concentration and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed below.

Analyte	CAS#	Analyte Lot	Concentration ± Uncertainty
Aroclor 1268	011100-14-4	RM00937	100.0 ± 0.5 µg/mL

Matrix: isooctane (2,2,4-trimethylpentane)

Storage Conditions: Store at Room Temperature (15° to 30°C).

Traceability:
The balances used for these measurements are calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z540.3, ISO 9001, ISO 17025, and ISO 17034. Calibrated Class A glassware is used for volumetric measurements. Thermometers are calibrated against a NIST traceable thermometer in accordance with NIST Special Publication 1088.

Homogeneity:
This RM was unitized according to an in-house procedure and is guaranteed to be homogeneous. There is no minimum sub-sample size required.

Intended Use:
This RM is intended for the preparation of working reference samples for use in routine laboratory analyses, calibration of instruments, validation of analytical methods, assessments of measurement methods, and continuing calibration verification.

Instructions for Use:
Sample aliquots for analysis should be withdrawn at 20°C to 25°C immediately after opening the container and should be processed without delay for the certified values to be valid within the stated uncertainties.

Hazards:
Refer to the Safety Data Sheet on www.agilent.com for information regarding this RM.

Expiration of Certification:
The certification of this RM is valid until the expiration date specified above, provided the RM is handled and stored in accordance with the instructions given in this certificate. This certification is nullified if the RM is damaged, contaminated, or otherwise modified.

Maintenance of Certification:
If substantive changes are noted that affect the certification before the expiration of this certificate, Agilent will notify the purchaser.

Sample lot approver:

M Bourgeois

Monica Bourgeois
QMS Representative

ANAB ACCREDITED REFERENCE MATERIAL PRODUCER
ISO 17034 Cert No. AR-1936

RM was produced in accordance with TUV USA Inc registered ISO 9001 Quality Management System. Cert # 56 100 18560026

www.agilent.com/quality/
CSD-QA-015.1

ANAB ACCREDITED TESTING LABORATORY
ISO 17025 Cert No. AT-1937

250 Smith Street North Kingstown, Rhode Island 02852 www.agilent.com/quality

RESTEK®
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: (800)356-1688
Fax: (814)353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

ACCREDITED
ISO 17034 Accredited
Reference Material Producer
Certificate #3222.01

ACCREDITED
ISO/IEC 17025 Accredited
Testing Laboratory
Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32011 **Lot No.:** A0175403

Description : Aroclor® 1254 Standard

Aroclor® 1254 Standard 1,000 µg/mL, Hexane, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : November 30, 2027 **Storage:** 25°C nominal

Handling: This product contains PCBs. **Ship:** Ambient

CERTIFIED VALUES

Elution Order	Compound	Grav. Conc. (weight/volume)	Expanded Uncertainty (95% C.L.; K=2)	
1	Aroclor 1254	1,000.7 µg/mL	+/- 5.9437 µg/mL	Gravimetric
	CAS # 11097-69-1 (Lot 124-191-B)		+/- 31.7284 µg/mL	Unstressed
	Purity ----%		+/- 41.4406 µg/mL	Stressed

Solvent: Hexane
CAS # 110-54-3
Purity 99%

P11588 ↓ P11592 | 5

AR 04/30/2022

Column:
30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)


Carrier Gas:
helium-constant pressure 20 psi.

Temp. Program:
200°C to 300°C
@ 25°C/min. (hold 10 min.)

Inj. Temp:
250°C

Det. Temp:
300°C

Det. Type:
ECD



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Cathleen Soltis - Mix Technician

Date Mixed: 15-Aug-2021 **Balance:** 1128360905

Alexis Shelow - Operations Tech I

Date Passed: 17-Aug-2021

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

P11588 | (5)
1
P11592

AR
04/30/22

Absolute Standards, Inc.
800-368-1131
www.absolutestandards.com

Certified Refere

CERTIFIED WEIGHT REPORT

Part Number:	91867
Lot Number:	020823
Description:	WP 037 - Aroclor 1232 PCB Technical Mixture
Expiration Date:	020833
Recommended Storage:	Ambient (20 °C)
Nominal Concentration (µg/mL):	100
NIST Test ID#:	6UTB

Solvent(
Aceton


5E-05 Balance Uncertainty
0.057 Flask Uncertainty

Weight(s) shown below were combined and diluted to (mL): 100.0

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Uncertainty Purity	Target Weight (g
1. Aroclor 1232	17	45-6A	100	100	0.5	0.01000

- The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
- Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
- Standards are certified (+/-) 0.5% of the stated value, unless otherwise stated.
- All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
- Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measure Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).

Comments
GC3-M1 Analysis by Melissa Stonier
Column ID SPB-608 30 meter X 0.53mm X5µm film thickness
Flow rates: Helium (carrier) = 5mL/min, Helium (make-up) = 25mL/min
Hydrogen (make-up) = 30mL/min, Air (make-up) = 350mL/min
Oven Profile: Temp 1 = 150°C (Time 1 = 4 min), Temp 2 = 290°C (Time 2 = 13.5 min)
Rate = 8°C/min, Total run time = 35 min
Injector temp. = 200°C, FID Temp. = 300°C. FID Signal = Edaq Channel 1
Standard injection =1.5µL, Range=3



Absolute Standards, Inc.
800-368-1131
www.absolutestandards.com


Certified Referen

CERTIFIED WEIGHT REPORT

Part Number:	X9166	**Solvent(s):**	
Lot Number:	060523	Methanol	E
Description:	Aroclor 1262 Appendix IX Compound		
Expiration Date:	060533		
Recommended Storage:	Ambient (20 °C)		
Nominal Concentration (µg/mL):	100		
NIST Test ID#:	6UTB	5E-05	Balance Uncertainty
Volume(s) shown below were combined and diluted to (mL):	20.0	0.002	Flask Uncertainty

Compound	Part Number	Lot Number	Dil. Factor	Initial Vol. (mL)	Uncertainty Pipette (mL)	Co
1. Aroclor 1262	70444	080322	0.10	2.00	0.017	

Method GC7MSD-7.M: Column:(30m X 0.25mm ID X 0.25µm film thickn
min.), Rate = 8°C/min., Injector B= 200°C, Detector B = 290°C.



- The certified value is the concentration calculated from gravimetric and volumetric me
- Standards are prepared gravimetrically using balances that are calibrated with weight
- Standards are certifed (+/-) 0.5% of the stated value, unless otherwise stated.
- All Standards, after opening ampule, should be stored with caps tight and under appro
- Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).

Absolute Standards, Inc.
800-368-1131
www.absolutestandards.com

Certified Reference Material CRM

ANAB ISO 17034 Accredited
AR-1539 Certificate Number
https://Absolutestandards.com

CERTIFIED WEIGHT REPORT

Part Number:	20064
Lot Number:	022023
Description:	CLP PCB'S - Aroclor Mix
	Aroclors 1016 & 1260
Expiration Date:	022033
Recommended Storage:	Ambient (20 °C)
Nominal Concentration (µg/mL):	1000
NIST Test ID#:	6UTB

Solvent(s):	Lot#
Hexane	273615

5E-05 Balance Uncertainty

Weight(s) shown below were combined and diluted to (mL): 200.0 0.010 Flask Uncertainty


Formulated By:	Benson Chan	022023 DATE
Reviewed By:	Pedro L. Rentas	022023 DATE

P12946 ↓ P12955 — Y.P. 12/20/23

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Uncertainty Purity	Target Weight(g)	Actual Weight(g)	Actual Conc (µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.) CAS#	OSHA PEL (TWA)	LD50
1. Aroclor 1016	15	020491JC	1000	100	0.2	0.20004	0.20060	1002.8	4.0	12674-11-2	N/A	N/A
2. Aroclor 1260	21	020491JC	1000	100	0.2	0.20004	0.20081	1003.9	4.0	11096-82-5	0.5mg/m3	orl-rat 1315mg/kg

- The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
- Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
- Standards are certifed (+/-) 0.5% of the stated value, unless otherwise stated.
- All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
- Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).

Comments
GC3-M1 Analysis by Melissa Stonier
Column ID SPB-608 30 meter X 0.53mm X5µm film thickness
Flow rates: Helium (carrier) = 5mL/min, Helium (make-up) = 25mL/min
Hydogen (make-up) = 30mL/min, Air (make-up) = 350mL/min
Oven Profile: Temp 1 = 150°C (Time 1 = 4 min), Temp 2 = 290°C (Time 2 = 13.5 min)
Rate = 8°C/min, Total run time = 35 min
Injector temp. = 200°C, FID Temp. = 300°C. FID Signal = Edaq Channel 1
Standard injection =1.5µL, Range=3



Absolute Standards, Inc.
800-368-1131
www.absolutestandards.com

Certified Reference Material CRM

ANAB ISO 17034 Accredited
AR-1539 Certificate Number
https://Absolutestandards.com

CERTIFIED WEIGHT REPORT

Part Number:	20064
Lot Number:	022023
Description:	CLP PCB'S - Aroclor Mix
	Aroclors 1016 & 1260
Expiration Date:	022033
Recommended Storage:	Ambient (20 °C)
Nominal Concentration (µg/mL):	1000
NIST Test ID#:	6UTB

Solvent(s):	Lot#
Hexane	273615

5E-05 Balance Uncertainty

Weight(s) shown below were combined and diluted to (mL): 200.0 0.010 Flask Uncertainty


Formulated By:	Benson Chan	022023 DATE
Reviewed By:	Pedro L. Rentas	022023 DATE

P12946 ↓ P12955 — Y.P. 12/20/23

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Uncertainty Purity	Target Weight(g)	Actual Weight(g)	Actual Conc (µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.) CAS#	OSHA PEL (TWA)	LD50
1. Aroclor 1016	15	020491JC	1000	100	0.2	0.20004	0.20060	1002.8	4.0	12674-11-2	N/A	N/A
2. Aroclor 1260	21	020491JC	1000	100	0.2	0.20004	0.20081	1003.9	4.0	11096-82-5	0.5mg/m3	orl-rat 1315mg/kg

- The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
- Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
- Standards are certifed (+/-) 0.5% of the stated value, unless otherwise stated.
- All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
- Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).

Comments
GC3-M1 Analysis by Melissa Stonier
Column ID SPB-608 30 meter X 0.53mm X5µm film thickness
Flow rates: Helium (carrier) = 5mL/min, Helium (make-up) = 25mL/min
Hydogen (make-up) = 30mL/min, Air (make-up) = 350mL/min
Oven Profile: Temp 1 = 150°C (Time 1 = 4 min), Temp 2 = 290°C (Time 2 = 13.5 min)
Rate = 8°C/min, Total run time = 35 min
Injector temp. = 200°C, FID Temp. = 300°C. FID Signal = Edaq Channel 1
Standard injection =1.5µL, Range=3



Absolute Standards, Inc.
800-368-1131
www.absolutestandards.com

Certified Reference Material CRM

ANAB ISO 17034 Accredited
AR-1539 Certificate Number
https://Absolutestandards.com

CERTIFIED WEIGHT REPORT

Part Number:	99139	**Solvent(s):**	**Lot#**	
Lot Number:	121823	Iso-octane	82227	
Description:	Aroclor 1254			
Expiration Date:	121833			
Recommended Storage:	Ambient (20 °C)			
Nominal Concentration (µg/mL):	100			
NIST Test ID#:	6UTB	5E-05	Balance Uncertainty	
Volume(s) shown below were combined and diluted to (mL):		20.0	0.003	Flask Uncertainty

		121823
Formulated By:	Anthony Mahoney	DATE
		121823
Reviewed By:	Pedro L. Rentas	DATE

P12956 & P12957 Y.P. 12/20/23


Note: Aroclor 1254 is a mix of isomers.

Compound	Part Number	Lot Number	Dilution Factor	Initial Vol. (mL)	Uncertainty Pipette (mL)	Initial Conc.(µg/mL)	Final Conc.(µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.) CAS#	OSHA PEL (TWA)	LD50
1. Aroclor 1254	79100	121823	0.10	2.00	0.017	1003.3	100.1	1.8	11097-69-1	0.5mg/m3 (skin)	orl-rat 1295mg/kg

- The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
- Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
- Standards are certifed (+/-) 0.5% of the stated value, unless otherwise stated.
- All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
- Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).

Comments
GC3-M1 Analysis by Melissa Stonier
Column ID SPB-608 30 meter X 0.53mm X5µm film thickness
Flow rates: Helium (carrier) = 5mL/min, Helium (make-up) = 25mL/min
Hydrogen (make-up) = 30mL/min, Air (make-up) = 350mL/min
Oven Profile: Temp 1 = 150°C (Time 1 = 4 min), Temp 2 = 290°C (Time 2 = 13.5 min)
Rate = 8°C/min, Total run time = 35 min
Injector temp. = 200°C, FID Temp. = 300°C. FID Signal = Edaq Channel 1
Standard injection =1.5µL, Range=3

Peak No.	Name	FID RT (min.)
1	Total Aroclor 1254	18.12



RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus

ACCREDITED ISO 17034 Accredited Reference Material Producer Certificate #3222.01

ACCREDITED ISO/IEC 17025 Accredited Testing Laboratory Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

P13348 ↓ P13357 (10)

04/25/2024

Catalog No. : 32000 **Lot No.:** A0206810

Description : Pesticide Surrogate Mix

Pesticide Surrogate Mix 200 µg/mL, Acetone, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : April 30, 2030 **Storage:** 10°C or colder

Handling: Contains PCBs - sonicate prior to use. **Ship:** Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	2,4,5,6-Tetrachloro-m-xylene	877-09-8	RP220407	99%	200.3 µg/mL	+/- 11.1143
2	Decachlorobiphenyl (BZ# 209)	2051-24-3	30638	99%	200.6 µg/mL	+/- 11.1298

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Acetone
CAS # 67-64-1
Purity 99%

Tech Tips:

Decachlorobiphenyl has poor solubility in most organic solvents. The maximum concentration that can be prepared in acetone, hexane, or isooctane is 200µg/mL. Temperature will affect the solubility as well. Storing solutions at reduced temperatures will cause decachlorobiphenyl to precipitate.

Products containing decachlorobiphenyl must be sonicated for a minimum of 10 minutes prior to opening the ampul. Because each ultrasonic bath operates at a different energy level, 10 minutes is a guideline only. Longer sonication time will not affect product quality.

These precautions apply to working solutions prepared in your laboratory as well. The amount of compound that precipitates depends on concentration AND temperature. If you store your standards at a temperature lower than 4°C (even dilute solutions), allow extra sonication time.

Quality Confirmation Test

Column:
30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:
hellum-constant pressure 20 psi.

Temp. Program:
200°C to 300°C
@ 25°C/min. (hold 10 min.)


Inj. Temp:
250°C

Det. Temp:
300°C

Det. Type:
ECD

Split Vent:
10 ml/min.

Inj. Vol
1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Laith Clemente - Operations Technician I

Date Mixed: 22-Jan-2024 **Balance Serial #** 1128360905

Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 24-Jan-2024

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

P13348 (10)

↓

P13357

04/25/2025

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus

ACCREDITED
ISO 17034 Accredited
Reference Material Producer
Certificate #3222.01

ACCREDITED
ISO/IEC 17025 Accredited
Testing Laboratory
Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32007 **Lot No.:** A0207511

Description : Aroclor® 1221 Standard

Aroclor® 1221 Standard 1,000 µg/mL, Hexane, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : May 31, 2030 **Storage:** 25°C nominal

Handling: This product contains PCBs. **Ship:** Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Aroclor 1221	11104-28-2	13192500	----%	1,002.7 µg/mL	+/- 55.6424

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Hexane
CAS # 110-54-3
Purity 99%

P 13374 (2)
↓
P 13375
1

05/6/2024

Quality Confirmation Test

Column:
30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:
helium-constant pressure 20 psi.

Temp. Program:
200°C to 300°C
@ 25°C/min. (hold 10 min.)


Inj. Temp:
250°C

Det. Temp:
300°C

Det. Type:
ECD

Split Vent:
10 ml/min.

Inj. Vol
1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Laith Clemente - Operations Technician I **Date Mixed:** 06-Feb-2024 **Balance Serial #** 1128360905

Jennifer Pollino - Operations Tech III - ARM QC **Date Passed:** 08-Feb-2024

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

P13374 ②
↓
P13375

05/6/2024

ISO 17034

Agilent
Trusted Answers

Reference Material Certificate

Product Name:	Aroclor 1242 Standard	**Lot Number:**	0006665550
Product Number:	PP-312-1	**Lot Issue Date:**	08-Feb-2022
Storage Conditions:	Store at Room Temperature (15° to 30°C).	**Expiration Date:**	31-Jan-2027

	CERTIFIED VALUES				
Component Name	Concentration		Expanded Uncertainty	CAS#	Analyte Lot
Aroclor 1242	100.4	±	0.5 µg/mL	053469-21-9	NT01020

Matrix: isooctane (2,2,4-trimethylpentane)

Description:
This document is prepared in accordance with ISO 17034 and Guide 31. This analytical reference material standard was manufactured and verified in accordance with an ISO 9001 registered quality system and analyte concentrations were verified by an ISO 17025 accredited laboratory. The concentration and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed above.

Traceability:
The balances used for these measurements are calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z540.3, ISO 9001, ISO 17025, and ISO 17034. Calibrated Class A glassware is used for volumetric measurements. Thermometers are calibrated against a NIST traceable thermometer in accordance with NIST Special Publication 1088.

Homogeneity:
This analytical reference standard was unitized according to an in-house procedure and is guaranteed to be homogeneous. There is no minimum sub-sample size required.

Instructions for Use:
Sample aliquots for analysis should be withdrawn at 20°C to 25°C immediately after opening the container and should be processed without delay for the certified values to be valid within the stated uncertainties.

Safety:
Refer to the Safety Data Sheet on www.agilent.com for information regarding this analytical reference material.

Intended Use:
This analytical reference standard is intended for the preparation of working reference samples for use in routine laboratory analyses, calibration of instruments, validation of analytical methods, assessments of measurement methods, and continuing calibration verification.

Expiration of Certification:
The certification of this analytical reference standard is valid until the expiration date specified above, provided the material is handled and stored in accordance with the instructions given in this certificate. This certification is nullified if the material is damaged, contaminated, or otherwise modified.

P13589
↓
P13590

AJ
10/14/24

Maintenance of Certification:
If substantive changes are noted that affect the certification before the expiration of this certificate, Agilent will notify the purchaser.

Sample lot approver:

Monica Bourgeois
QMS Representative

Agilent
Trusted Answers

ISO 17034

Reference Material Certificate

Product Information Sheet

Product Name: Aroclor 1248 Standard
Product Number: PP-342-1
Storage Conditions: Store at Room Temperature (15° to 30°C).

Lot Number: 0006726317
Lot Issue Date: 27-Jan-2023
Expiration Date: 28-Feb-2031

Component Name	Concentration		Uncertainty	CAS#	Analyte Lot
Aroclor 1248	100.3	±	0.5 µg/mL	012672-29-6	NT01582

Matrix: isooctane (2,2,4-trimethylpentane)

Description:
This document is prepared in accordance with ISO 17034 and Guide 31. This analytical reference material (RM) standard was manufactured and verified in accordance with an ISO 9001 registered quality system and analyte concentrations were verified by an ISO 17025 accredited laboratory. The concentration and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed above. Purity values are taken from approved vendor raw material certificates.

Traceability:
The balances used for these measurements are calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z540.3, ISO 9001, ISO 17025, and ISO 17034. Calibrated Class A glassware is used for volumetric measurements. Thermometers are calibrated against a NIST traceable thermometer in accordance with NIST Special Publication 1088.

Homogeneity:
This analytical reference (RM) standard was unitized according to an in-house procedure and is guaranteed to be homogeneous. There is no minimum sub-sample size required.

Instructions for Use:
Sample aliquots for analysis should be withdrawn at 20°C to 25°C immediately after opening the container and should be processed without delay for the certified values to be valid within the stated uncertainties.

Safety:
Refer to the Safety Data Sheet on www.agilent.com for information regarding this analytical reference material.

Intended Use:
This analytical reference (RM) standard is intended for the preparation of working reference samples for use in routine laboratory analyses, calibration of instruments, validation of analytical methods, assessments of measurement methods, and continuing calibration verification.

Expiration of Certification:
The certification of this analytical reference standard (RM) is valid until the expiration date specified above, provided the material is handled and stored in accordance with the instructions given in this certificate. This certification is nullified if the material is damaged, contaminated, or otherwise modified.

P13591
↓
P13592

AJ
10/14/2024

Agilent

Trusted Answers

Maintenance of Certification:
If substantive changes are noted that affect the certification before the expiration of this certificate, Agilent will notify the purchaser.

Sample lot approver:

Monica Bourgeois
QMS Representative

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus

ACCREDITED
ISO 17034 Accredited
Reference Material Producer
Certificate #3222.01

ACCREDITED
ISO/IEC 17025 Accredited
Testing Laboratory
Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32010 **Lot No.:** A0214733

Description : Aroclor® 1248 Standard

Aroclor® 1248 Standard 1,000µg/mL, Hexane, 1mL/ampul

P13695
P13696
Y.P.
10/17/24

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : November 30, 2030 **Storage:** 25°C nominal

Handling: This product contains PCBs. **Ship:** Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Aroclor 1248	12672-29-6	15242800	----%	1,005.0 µg/mL	+/- 55.7700

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Hexane
CAS # 110-54-3
Purity 99%

Quality Confirmation Test

Column:
30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:
helium-constant pressure 20 psi.

Temp. Program:
200°C to 300°C
@ 25°C/min. (hold 10 min.)


Inj. Temp:
250°C

Det. Temp:
300°C

Det. Type:
ECD

Split Vent:
10 ml/min.

Inj. Vol
0.2µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Rebecca Gingerich - Operations Tech II

Date Mixed: 02-Aug-2024 **Balance Serial #** C322230531

Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 13-Aug-2024

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k\sqrt{u^2_{gravimetric} + u^2_{homogeneity} + u^2_{storage\ stability} + u^2_{shipping\ stability}}$$

  k is a coverage factor of 2, which gives a level of confidence of approximately 95%.
- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus

ACCREDITED ISO 17034 Accredited Reference Material Producer Certificate #3222.01

ACCREDITED ISO/IEC 17025 Accredited Testing Laboratory Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32039 **Lot No.:** A0210629

Description : Aroclor® 1016/1260 Mix

Aroclor® 1016/1260 Mix 1,000 µg/mL, Hexane, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : July 31, 2030 **Storage:** 25°C nominal

Handling: This product contains PCBs. **Ship:** Ambient

P13697 ↓ P13701 } Y.P. 10/19/24

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Aroclor 1016	12674-11-2	07	----%	1,005.3 µg/mL	+/- 55.7809
2	Aroclor 1260	11096-82-5	1320657	----%	1,000.0 µg/mL	+/- 55.4850

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Hexane
CAS # 110-54-3
Purity 99%

Quality Confirmation Test

Column:
30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:
helium-constant pressure 20 psi.

Temp. Program:
200°C to 300°C
@ 25°C/min. (hold 10 min.)


Inj. Temp:
250°C

Det. Temp:
300°C

Det. Type:
ECD

Split Vent:
10 ml/min.

Inj. Vol
0.2µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Laith Clemente - Operations Technician I

Date Mixed: 22-Apr-2024 **Balance Serial #** B442140311

Dillan Murphy - Operations Technician I

Date Passed: 24-Apr-2024

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k\sqrt{u^2_{gravimetric} + u^2_{homogeneity} + u^2_{storage\ stability} + u^2_{shipping\ stability}}$$

  k is a coverage factor of 2, which gives a level of confidence of approximately 95%.
- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.

RESTEK

110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus

ACCREDITED
ISO 17034 Accredited
Reference Material Producer
Certificate #3222.01

ACCREDITED
ISO/IEC 17025 Accredited
Testing Laboratory
Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32009 **Lot No.:** A0210232

Description : Aroclor® 1242 Standard

Aroclor® 1242 Standard 1,000 µg/mL, Hexane, 1mL/ampul

P13706
P13707
Y.P. 10/17/24

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : July 31, 2030 **Storage:** 25°C nominal

Handling: This product contains PCBs. **Ship:** Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Aroclor 1242	53469-21-9	01141	----%	1,000.7 µg/mL	+/- 55.5295

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Hexane
CAS # 110-54-3
Purity 99%

Quality Confirmation Test

Column:
30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:
helium-constant pressure 20 psi.

Temp. Program:
200°C to 300°C
@ 25°C/min. (hold 10 min.)


Inj. Temp:
250°C

Det. Temp:
300°C

Det. Type:
ECD

Split Vent:
10 ml/min.

Inj. Vol
0.2µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Amanda Miller - Operations Tech III - ARM QC

Date Mixed: 11-Apr-2024 **Balance Serial #** B442140311

Christie Mills - Operations Lead Tech - ARM QC

Date Passed: 17-Apr-2024

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k\sqrt{u^2_{gravimetric} + u^2_{homogeneity} + u^2_{storage\ stability} + u^2_{shipping\ stability}}$$

  k is a coverage factor of 2, which gives a level of confidence of approximately 95%.
- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.

RESTEK

110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus

ACCREDITED
ISO 17034 Accredited
Reference Material Producer
Certificate #3222.01

ACCREDITED
ISO/IEC 17025 Accredited
Testing Laboratory
Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32000 **Lot No.:** A0214495

Description : Pesticide Surrogate Mix

Pesticide Surrogate Mix 200 µg/mL, Acetone, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : October 31, 2030 **Storage:** 10°C or colder

Handling: Contains PCBs - sonicate prior to use. **Ship:** Ambient

P19785
↓
P19789
AJ
11/19/24

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	2,4,5,6-Tetrachloro-m-xylene	877-09-8	RP220407	99%	200.2 µg/mL	+/- 11.1087
2	Decachlorobiphenyl (BZ# 209)	2051-24-3	30679	99%	201.4 µg/mL	+/- 11.1753

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Acetone
CAS # 67-64-1
Purity 99%

Tech Tips:

Decachlorobiphenyl has poor solubility in most organic solvents. The maximum concentration that can be prepared in acetone, hexane, or isooctane is 200µg/mL. Temperature will affect the solubility as well. Storing solutions at reduced temperatures will cause decachlorobiphenyl to precipitate.

Products containing decachlorobiphenyl must be sonicated for a minimum of 10 minutes prior to opening the ampul. Because each ultrasonic bath operates at a different energy level, 10 minutes is a guideline only. Longer sonication time will not affect product quality.

These precautions apply to working solutions prepared in your laboratory as well. The amount of compound that precipitates depends on concentration AND temperature. If you store your standards at a temperature lower than 4°C (even dilute solutions), allow extra sonication time.

Quality Confirmation Test

Column:
30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:
helium-constant pressure 20 psi.

Temp. Program:
200°C to 300°C
@ 25°C/min. (hold 10 min.)


Inj. Temp:
250°C

Det. Temp:
300°C

Det. Type:
ECD

Split Vent:
10 ml/min.

Inj. Vol
1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Aaron Enyart - Operations Tech I

Date Mixed: 29-Jul-2024 **Balance Serial #** B345965662

Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 01-Aug-2024

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

RESTEK

110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus

ACCREDITED
ISO 17034 Accredited
Reference Material Producer
Certificate #3222.01

ACCREDITED
ISO/IEC 17025 Accredited
Testing Laboratory
Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32000 **Lot No.:** A0214495

Description : Pesticide Surrogate Mix

Pesticide Surrogate Mix 200 µg/mL, Acetone, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : October 31, 2030 **Storage:** 10°C or colder

Handling: Contains PCBs - sonicate prior to use. **Ship:** Ambient

P19785
↓
P19789
AJ
11/19/24

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	2,4,5,6-Tetrachloro-m-xylene	877-09-8	RP220407	99%	200.2 µg/mL	+/- 11.1087
2	Decachlorobiphenyl (BZ# 209)	2051-24-3	30679	99%	201.4 µg/mL	+/- 11.1753

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Acetone
CAS # 67-64-1
Purity 99%

Tech Tips:

Decachlorobiphenyl has poor solubility in most organic solvents. The maximum concentration that can be prepared in acetone, hexane, or isooctane is 200µg/mL. Temperature will affect the solubility as well. Storing solutions at reduced temperatures will cause decachlorobiphenyl to precipitate.

Products containing decachlorobiphenyl must be sonicated for a minimum of 10 minutes prior to opening the ampul. Because each ultrasonic bath operates at a different energy level, 10 minutes is a guideline only. Longer sonication time will not affect product quality.

These precautions apply to working solutions prepared in your laboratory as well. The amount of compound that precipitates depends on concentration AND temperature. If you store your standards at a temperature lower than 4°C (even dilute solutions), allow extra sonication time.

Quality Confirmation Test

Column:
30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:
helium-constant pressure 20 psi.

Temp. Program:
200°C to 300°C
@ 25°C/min. (hold 10 min.)


Inj. Temp:
250°C

Det. Temp:
300°C

Det. Type:
ECD

Split Vent:
10 ml/min.

Inj. Vol
1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Aaron Enyart - Operations Tech I

Date Mixed: 29-Jul-2024 **Balance Serial #** B345965662

Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 01-Aug-2024

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

RESTEK

110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus

ilac-MRA A2LA ACCREDITED ISO 17034 Accredited Reference Material Producer Certificate #3222.01

ilac-MRA A2LA ACCREDITED ISO/IEC 17025 Accredited Testing Laboratory Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32011 **Lot No.:** A0217391

Description : Aroclor® 1254 Standard

Aroclor® 1254 Standard 1,000 µg/mL, Hexane, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : January 31, 2031 **Storage:** 25°C nominal

Handling: This product contains PCBs. **Ship:** Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Aroclor 1254	11097-69-1	124-191-B	----%	1,004.7 µg/mL	+/- 55.7515

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Hexane
CAS # 110-54-3
Purity 99%

P13830
↓
P13832
↓
P13834

AJ
12/09/24

Quality Confirmation Test

Column:
30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:
helium-constant pressure 20 psi.

Temp. Program:
200°C to 300°C
@ 25°C/min. (hold 10 min.)


Inj. Temp:
250°C

Det. Temp:
300°C

Det. Type:
ECD

Split Vent:
300 ml/min.

Inj. Vol
1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Michael Maye - Operations Tech I

Date Mixed: 02-Oct-2024 **Balance Serial #** C322230531

Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 07-Oct-2024

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus

ACCREDITED ISO 17034 Accredited Reference Material Producer Certificate #3222.01

ACCREDITED ISO/IEC 17025 Accredited Testing Laboratory Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32008 **Lot No.:** A0219655

Description : Aroclor® 1232 Standard

Aroclor® 1232 Standard 1,000 µg/mL, Hexane, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : March 31, 2031 **Storage:** 25°C nominal

Handling: This product contains PCBs. **Ship:** Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Aroclor 1232	11141-16-5	15665-01	----%	1,007.0 µg/mL	+/- 55.8810

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Hexane
CAS # 110-54-3
Purity 99%

P13878
↓
P13880

AJ
01/28/25

Quality Confirmation Test

Column:
30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:
helium-constant pressure 20 psi.

Temp. Program:
200°C to 300°C
@ 25°C/min. (hold 10 min.)


Inj. Temp:
250°C

Det. Temp:
300°C

Det. Type:
ECD

Split Vent:
10 ml/min.

Inj. Vol
1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Michael Maye - Operations Tech I

Date Mixed: 02-Dec-2024 **Balance Serial #** C322230531

Brittany Federinko - Operations Tech I

Date Passed: 05-Dec-2024

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus

ACCREDITED
ISO 17034 Accredited
Reference Material Producer
Certificate #3222.01

ACCREDITED
ISO/IEC 17025 Accredited
Testing Laboratory
Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32409 **Lot No.:** A0217041

Description : Aroclor® 1262 Standard

Aroclor® 1262 Standard 1,000 µg/mL, 1mL/ampul, Hexane

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : December 31, 2030 **Storage:** 25°C nominal

Handling: This product contains PCBs. **Ship:** Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Aroclor 1262	37324-23-5	10849100	----%	1,000.0 µg/mL	+/- 55.4925

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Hexane
CAS # 110-54-3
Purity 99%

P13881

AJ
01/28/25

Quality Confirmation Test

Column:
30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:
helium-constant pressure 20 psi.

Temp. Program:
200°C to 300°C
@ 25°C/min. (hold 10 min.)


Inj. Temp:
250°C

Det. Temp:
300°C

Det. Type:
ECD

Split Vent:
300 ml/min.

Inj. Vol
0.2µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Rebecca Gingerich - Operations Tech II

Date Mixed: 25-Sep-2024 **Balance Serial #** C322230531

Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 30-Sep-2024

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus

ACCREDITED
ISO 17034 Accredited
Reference Material Producer
Certificate #3222.01

ACCREDITED
ISO/IEC 17025 Accredited
Testing Laboratory
Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32410 **Lot No.:** A0217264

Description : Aroclor® 1268 Standard

Aroclor® 1268 Standard 1,000 µg/mL, 1mL/ampul, Hexane

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : January 31, 2031 **Storage:** 25°C nominal

Handling: This product contains PCBs. **Ship:** Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Aroclor 1268	11100-14-4	12353400	----%	1,009.8 µg/mL	+/- 56.0364

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Hexane
CAS # 110-54-3
Purity 99%

P13884
↓
P13886

AJ
01/28/25

Quality Confirmation Test

Column:
30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:
helium-constant pressure 20 psi.

Temp. Program:
200°C to 300°C
@ 25°C/min. (hold 10 min.)


Inj. Temp:
250°C

Det. Temp:
300°C

Det. Type:
ECD

Split Vent:
Split ratio 500:1

Inj. Vol
0.2µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Stacey Wanner - Operations Technician I

Date Mixed: 01-Oct-2024 **Balance Serial #** 1128360905

Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 02-Oct-2024

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

n-Hexane 95%
ULTRA RESI-ANALYZED
For Organic Residue Analysis

avantor™ | J.T.Baker®

W3147

W3147 update. 02/03/2025 JP

Material No.: 9262-03
Batch No.: 24G1962003
Manufactured Date: 2024-05-23
Expiration Date: 2025-08-22
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL)	≤ 5	3
ECD Sensitive Impurities (as Heptachlor Epoxide) Single Peak (pg/mL)	≤ 10	1
ECD-Sensitive Impurities (as Ethylene Dibromide) - Single Impurity Peak (ng/mL)	≤ 5	1
Assay (Total Saturated C_6 Isomers) (by GC, corrected for water)	≥ 99.5 %	99.7 %
Assay (as n-Hexane) (by GC, corrected for water)	≥ 95 %	98 %
Color (APHA)	≤ 10	5
Residue after Evaporation	≤ 1.0 ppm	0.1 ppm
Substances Darkened by H_2SO_4	Passes Test	Passes Test
Water (by KF, coulometric)	≤ 0.05 %	< 0.01 %

For Laboratory,Research,or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: USA
Packaging Site: Phillipsburg Mfg Ctr & DC

JCroak

Jamie Croak
Director Quality Operations, Bioscience Production

n-Hexane 95%
ULTRA RESI-ANALYZED
For Organic Residue Analysis

avantor™ | J.T.Baker

W3234 JB
Update: 07/28/2025

Material No.: 9262-03
Batch No.: 25C0362005
Manufactured Date: 2025-01-29
Expiration Date:2026-04-30
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
FID-Sensitive Impurities (as 2-Octanol)Single Impurity Peak (ng/mL)	<= 5	1
ECD Sensitive Impurities (as HeptachlorEpoxide) Single Peak (pg/mL)	<= 10	6
ECD-Sensitive Impurities (as EthyleneDibromide) - Single Impurity Peak (ng/mL)	<= 5	5
Assay (Total Saturated C_6 Isomers) (byGC, corrected for water)	>= 99.5 %	100.0 %
Assay (as n-Hexane) (by GC, correctedfor water)	>= 95 %	100 %
Color (APHA)	<= 10	10
Residue after Evaporation	<= 1.0 ppm	0.1 ppm
Substances Darkened by H_2SO_4	Passes Test	Passes Test
Water (by KF, coulometric)	<= 0.05 %	<0.01 %

For Laboratory,Research,or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: United States
Packaging Site: Phillipsburg Mfg Ctr & DC

JCroak
Jamie Croak
Director Quality Operations, Bioscience Production

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700



SHIPPING DOCUMENTS

Alliance TECHNICAL GROUP

284 Sheffield Street, Mountainside, NJ 07092
(908) 789-8900 • Fax (908) 789-8922
www.chemtech.net

ALLIANCE PROJECT NO.
QUOTE NO. Q3450
COC Number 2046803

CLIENT INFORMATION

REPORT TO BE SENT TO:
COMPANY: Matrix New World Eng
ADDRESS: 26 Columbia Tpke
CITY Florham Park STATE: NJ ZIP: 07932
ATTENTION:
PHONE: 973-240-1800 FAX:

CLIENT PROJECT INFORMATION

PROJECT NAME: PSEG Athenia Substation
PROJECT NO.: 24-0935 LOCATION: Clifton
PROJECT MANAGER: Chris Pittarese
e-mail: cpittarese@mnwe.com
PHONE: 973-240-1800 FAX:

CLIENT BILLING INFORMATION

BILL TO: SAME AS CLIENT PO#: INFO
ADDRESS:
CITY STATE: ZIP:
ATTENTION: PHONE:

DATA TURNAROUND INFORMATION

FAX (RUSH) _____ DAYS*
HARDCOPY (DATA PACKAGE): _____ DAYS*
EDD: 24HR (1 DAY) DAYS*
*TO BE APPROVED BY CHEMTECH
STANDARD HARDCOPY TURNAROUND TIME IS 10 BUSINESS

DATA DELIVERABLE INFORMATION

- ☐ Level 1 (Results Only)
- ☐ Level 2 (Results + QC)
- ☐ Level 3 (Results + QC + Raw Data)
- ☐ Level 4 (QC + Full Raw Data)
- ☒ NJ Reduced ☐ US EPA CLP
- ☐ NYS ASP A ☐ NYS ASP B
- ☐ Other _____
- ☐ EDD FORMAT _____

ANALYSIS

PRESERVATIVES: 1 RCRA; 2; 3; 4; 5; 6; 7; 8; 9

COMMENTS: ← Specify Preservatives
A-HCl D-NaOH
B-HN03 E-ICE
C-H2SO4 F-OTHER

	ALLIANCE SAMPLE ID	PROJECT SAMPLE IDENTIFICATION	SAMPLE MATRIX	SAMPLE TYPE COMP	SAMPLE TYPE GRAB	SAMPLE COLLECTION DATE	SAMPLE COLLECTION TIME	# OF BOTTLES	1	2	3	4	5	6	7	8	9
1.		SS-5R (2.0-2.5)	S		X	10/23	1010	1	X								
2.		SS-5R (3.5-4.0)					1005	1	X								
3.		SS-10R (2.0-2.5)					1135	1	H								
4.		SS-10R (3.5-4.0)					1132	1	H								
5.		SS-13R (3.5-4.0)					0955	1	H								
6.		SS-13R (4.5-5.0)					0945	1	H								
7.		SS-13R (5.0-5.5)					0950	1	H								
8.		SS-20 (0.0-0.5)					1125	1	X								
9.		SS-20 (2.0-2.5)					1120	1	X								
10.		SS-20 (3.5-4.0)	↓		↓	↓	1117	1	X								

SAMPLE CUSTODY MUST BE DOCUMENTED BELOW EACH TIME SAMPLES CHANGE POSSESSION INCLUDING COURIER DELIVERY

RELINQUISHED BY SAMPLER:	DATE/TIME:	RECEIVED BY:
1. Madison Digati	10/23 1544	1.
2.		2.
3.		3.

Conditions of bottles or coolers at receipt: ☐ COMPLIANT ☐ NON COMPLIANT ☐ COOLER TEMP 2.3 °C
Comments: X=run H=hold
24 Count 1

Page _____ of

CLIENT: ☐ Hand Delivered ☐ Other _____

Shipment Complete ☐ YES ☐ NO

Copyright © 2024

WHITE - ALLIANCE COPY FOR RETURN TO CLIENT YELLOW - ALLIANCE COPY PINK - SAMPLER COPY

Alliance TECHNICAL GROUP

284 Sheffield Street, Mountainside, NJ 07092
(908) 789-8900 · Fax (908) 789-8922
www.chemtech.net

ALLIANCE PROJECT NO.
QUOTE NO. Q3450
COC Number 2044503

CLIENT INFORMATION

REPORT TO BE SENT TO:
COMPANY: Matrix New World Eng.
ADDRESS: 26 Columbia Tpke
CITY Florham Park STATE: NJ ZIP: 07932
ATTENTION:
PHONE: 973-240-1800 FAX:

CLIENT PROJECT INFORMATION

PROJECT NAME: PSEG Athenia Substation
PROJECT NO.: 24_0935 LOCATION: Clifton
PROJECT MANAGER: Chris Pittarese
e-mail: cpittarese@mnwe.com
PHONE: 973-240-1800 FAX:

CLIENT BILLING INFORMATION

BILL TO: SAME AS CLIENT PO#: INFO
ADDRESS:
CITY STATE: ZIP:
ATTENTION: PHONE:

DATA TURNAROUND INFORMATION

FAX (RUSH) _____ DAYS*
HARDCOPY (DATA PACKAGE): _____ DAYS*
EDD: 24HR (1DAY) DAYS*
*TO BE APPROVED BY CHEMTECH
STANDARD HARDCOPY TURNAROUND TIME IS 10 BUSINESS

DATA DELIVERABLE INFORMATION

- ☐ Level 1 (Results Only)
- ☐ Level 2 (Results + QC)
- ☐ Level 3 (Results + QC + Raw Data)
- ☐ EDD FORMAT _____
- ☐ Level 4 (QC + Full Raw Data)
- ☒ NJ Reduced ☐ US EPA CLP
- ☐ NYS ASP A ☐ NYS ASP B
- ☐ Other _____

ANALYSIS

1: PCBs 2 3 4 5 6 7 8 9

ALLIANCE SAMPLE ID	PROJECT SAMPLE IDENTIFICATION	SAMPLE MATRIX	SAMPLE TYPE COMP	SAMPLE TYPE GRAB	SAMPLE COLLECTION DATE	SAMPLE COLLECTION TIME	# OF BOTTLES	PRESERVATIVES 1	2	3	4	5	6	7	8	9	COMMENTS
1.	SS-28 (2.0-2.5)	S		X	10/23	1130	1	X									
2.	SS-29 (2.0-2.5)					1140	1	H									
3.	SS-30 (0.0-0.5)					1155	1	H									
4.	SS-30 (2.0-2.5)					1150	1	H									
5.	SS-30 (3.5-4.0)					1145	1	H									
6.	SS-31 (0.0-0.5)					1115	1	H									
7.	SS-34 (0.0-0.5)					0925	1	H									
8.	SS-34 (2.0-2.5)					0930	1	H									
9.	SS-34 (3.5-4.0)					0935	1	H									
10.	SS-21 (0.0-0.5)					1335	1	X									

Specify Preservatives: A-HCl D-NaOH B-HN03 E-ICE C-H2SO4 F-OTHER

SAMPLE CUSTODY MUST BE DOCUMENTED BELOW EACH TIME SAMPLES CHANGE POSSESSION INCLUDING COURIER DELIVERY

RELINQUISHED BY SAMPLER:	DATE/TIME:	RECEIVED BY:
1. Madison Igoff	10/23 15:44	1.
2.		2.
3.		3.

Conditions of bottles or coolers at receipt: ☐ COMPLIANT ☐ NON COMPLIANT ☐ COOLER TEMP 2.3°C °C
Comments: X=run H=hold
2 Containers

Page _____ of

CLIENT: ☐ Hand Delivered ☐ Other _____

Shipment Complete ☐ YES ☐ NO

Copyright © 2024

WHITE - ALLIANCE COPY FOR RETURN TO CLIENT YELLOW - ALLIANCE COPY PINK - SAMPLER COPY

Alliance TECHNICAL GROUP

284 Sheffield Street, Mountainside, NJ 07092
(908) 789-8900 • Fax (908) 789-8922
www.chemtech.net

ALLIANCE PROJECT NO.
QUOTE NO. Q3450
COC Number 2044504

CLIENT INFORMATION

REPORT TO BE SENT TO:
COMPANY: Matrix New World Eng
ADDRESS: 26 Columbia Tpke
CITY: Florham Park STATE: NJ ZIP: 07932
ATTENTION:
PHONE: 973-240-1800 FAX:

CLIENT PROJECT INFORMATION

PROJECT NAME: PSEG Athenia Substation
PROJECT NO: 24-0935 LOCATION: Clifton
PROJECT MANAGER: Chris Pitarese
e-mail: cpitarese@mnre.com
PHONE: 973-240-1800 FAX:

CLIENT BILLING INFORMATION

BILL TO: SAME AS CLIENT PO#: INFO
ADDRESS:
CITY STATE: ZIP:
ATTENTION: PHONE:

DATA TURNAROUND INFORMATION

FAX (RUSH) _____ DAYS*
HARDCOPY (DATA PACKAGE): _____ DAYS*
EDD: 24HR (1 DAY) DAYS*
*TO BE APPROVED BY CHEMTECH
STANDARD HARDCOPY TURNAROUND TIME IS 10 BUSINESS

DATA DELIVERABLE INFORMATION

- ☐ Level 1 (Results Only)
- ☐ Level 2 (Results + QC)
- ☐ Level 3 (Results + QC + Raw Data)
- ☐ Level 4 (QC + Full Raw Data)
- ☐ NJ Reduced ☐ US EPA CLP
- ☐ NYS ASP A ☐ NYS ASP B
- ☐ Other _____
- ☐ EDD FORMAT _____

ANALYSIS

1: PCBs 2 3 4 5 6 7 8 9

ALLIANCE SAMPLE ID	PROJECT SAMPLE IDENTIFICATION	SAMPLE MATRIX	SAMPLE TYPE COMP	SAMPLE TYPE GRAB	SAMPLE COLLECTION DATE	SAMPLE COLLECTION TIME	# OF BOTTLES	PRESERVATIVES 1	2	3	4	5	6	7	8	9	COMMENTS
1.	SS-21 (2.0-2.5)	S		X	10/23	1333	1	X									
2.	SS-21 (3.5-4.0)					1330	1	X									
3.	SS-22 (0.0-0.5)					1340	1	H									
4.	SS-22 (2.0-2.5)					1342	1	H									
5.	SS-22 (3.5-4.0)	↓		↓	↓	1338	1	H									
6.																	
7.																	
8.																	
9.																	
10.																	

← Specify Preservatives
A-HCl D-NaOH
B-HN03 E-ICE
C-H2SO4 F-OTHER

SAMPLE CUSTODY MUST BE DOCUMENTED BELOW EACH TIME SAMPLES CHANGE POSSESSION INCLUDING COURIER DELIVERY

RELINQUISHED BY SAMPLER:	DATE/TIME:	RECEIVED BY:
1. Madison Tyrell	10/23 1544	1. CR
2.		2.
3.		3.

Conditions of bottles or coolers at receipt: ☐ COMPLIANT ☐ NON COMPLIANT ☐ COOLER TEMP 2-3 °C
Comments: X=run H=hold
If Cooled

Page _____ of

CLIENT: ☐ Hand Delivered ☐ Other _____

Shipment Complete ☐ YES ☐ NO

Copyright © 2024

WHITE - ALLIANCE COPY FOR RETURN TO CLIENT YELLOW - ALLIANCE COPY PINK - SAMPLER COPY

From: Christopher Pittarese <cpittarese@mnwe.com>
Sent: Thursday, October 30, 2025 8:27 PM
To: Yazmeen Gomez
Cc: Melanie Racaza
Subject: PSEG Athenia Substation

EXTERNAL EMAIL - This email was sent by a person from outside your organization. Exercise caution when clicking links, opening attachments or taking further action, before validating its authenticity.

Secured by Check Point

Hi Yazmeen-

Would you please submit the following contingency samples for analysis for PCBs on the same TAT provided on the initial run of samples? Please see notes below regarding lab IDs and let me know if anything is unclear of if there are any questions. Thanks.

Matrix ID	Lab ID	Date Collected	Analysis	Note
SS-13R (4.5-5.0)	Q3450-06	10/23/2025	PCB	I believe this sample is on the login summary report with the same Matrix ID that is listed for Q3450-07; please double check IDs.
SS-13R (5.0-5.5)	Q3450-07	10/23/2025	PCB	
SS-10R (3.5-4.0)	Q3450-04	10/23/2025	PCB	
SS-22 (0.0-0.5)	Q3450-23	10/23/2025	PCB	
SS-22 (3.5-4.0)	Q3450-25	10/23/2025	PCB	
SS-32 (0-0.5)	Q3437-25	10/22/2025	PCB	Sample not shown on login summary, please double check lab ID

Chris Pittarese, LSRP
Senior Project Manager

Matrix New World Engineering
3600 Route 66, 4th Floor
Neptune, NJ 07753
P: 973.240.1800
D: 732.515.4942
C: 609.276.7217
F: 973.240.1818

MATRIXNEWORLD
A TRUE ENVIRONMENTAL COMPANY

www.mnwe.com



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Laboratory Certification

Certified By	License No.
Connecticut	PH-0830
DOD ELAP (ANAB)	L2219
Maine	2024021
Maryland	296
New Hampshire	255425
New Jersey	20012
New York	11376
Pennsylvania	68-00548
Soil Permit	525-24-234-08441
Texas	TX-C25-00189
Virginia	460312

QA Control Code: A2070148