

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q3450</u>	Analytical Method:	<u>8082A</u>
Client:	<u>Matrix New World Engineering</u>	Datafile :	<u>PP076069.D</u>

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD	Limits		RPD
								Qual	Low	High	
PB170236BS (Column 1)	AR1016	166.6	145	ug/kg	87				40 (71)	140 (120)	
	AR1260	166.6	149	ug/kg	89				40 (65)	140 (130)	
PB170236BS (Column 2)	AR1016	166.6	127	ug/kg	76				40 (71)	140 (120)	
	AR1260	166.6	143	ug/kg	86				40 (65)	140 (130)	