

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: MATR02

Lab Code: ACE

SDG NO.: Q3450

Continuing Calib Date: 10/24/2025

Initial Calibration Date(s): 10/21/2025

10/21/2025

Continuing Calib Time: 08:57

Initial Calibration Time(s): 09:06

17:01

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	5.79	5.80	5.70	5.90	0.01
Aroclor-1016-2	(2)	5.81	5.82	5.72	5.92	0.01
Aroclor-1016-3	(3)	5.88	5.88	5.78	5.98	0.00
Aroclor-1016-4	(4)	5.97	5.98	5.88	6.08	0.01
Aroclor-1016-5	(5)	6.27	6.27	6.17	6.37	0.00
Aroclor-1260-1	(1)	7.38	7.39	7.29	7.49	0.01
Aroclor-1260-2	(2)	7.64	7.64	7.54	7.74	0.00
Aroclor-1260-3	(3)	8.00	8.00	7.90	8.10	0.01
Aroclor-1260-4	(4)	8.22	8.23	8.13	8.33	0.01
Aroclor-1260-5	(5)	8.55	8.55	8.45	8.65	0.00
Tetrachloro-m-xylene		4.64	4.64	4.54	4.74	0.00
Decachlorobiphenyl		10.41	10.42	10.32	10.52	0.01

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: MATR02

Lab Code: ACE

SDG NO.: Q3450

Continuing Calib Date: 10/24/2025

Initial Calibration Date(s): 10/21/2025

10/21/2025

Continuing Calib Time: 08:57

Initial Calibration Time(s): 09:06

17:01

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	4.87	4.88	4.78	4.98	0.01
Aroclor-1016-2	(2)	4.93	4.94	4.84	5.04	0.01
Aroclor-1016-3	(3)	5.05	5.05	4.95	5.15	0.00
Aroclor-1016-4	(4)	5.09	5.10	5.00	5.20	0.01
Aroclor-1016-5	(5)	5.31	5.31	5.21	5.41	0.00
Aroclor-1260-1	(1)	6.33	6.34	6.24	6.44	0.01
Aroclor-1260-2	(2)	6.52	6.53	6.43	6.63	0.01
Aroclor-1260-3	(3)	6.68	6.68	6.58	6.78	0.00
Aroclor-1260-4	(4)	7.14	7.15	7.05	7.25	0.01
Aroclor-1260-5	(5)	7.38	7.39	7.29	7.49	0.01
Tetrachloro-m-xylene		3.78	3.78	3.68	3.88	0.00
Decachlorobiphenyl		8.78	8.78	8.68	8.88	0.00