

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: MATR02

Lab Code: ACE

SDG NO.: Q3450

Continuing Calib Date: 10/27/2025

Initial Calibration Date(s): 10/21/2025 10/21/2025

Continuing Calib Time: 20:21

Initial Calibration Time(s): 09:06 17:01

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Aroclor-1254-1 (1)	6.64	6.65	6.55	6.75	0.01
Aroclor-1254-2 (2)	6.86	6.87	6.77	6.97	0.01
Aroclor-1254-3 (3)	7.22	7.23	7.13	7.33	0.01
Aroclor-1254-4 (4)	7.50	7.51	7.41	7.61	0.01
Aroclor-1254-5 (5)	7.92	7.93	7.83	8.03	0.01
Tetrachloro-m-xylene	4.64	4.65	4.55	4.75	0.01
Decachlorobiphenyl	10.41	10.42	10.32	10.52	0.01

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: MATR02

Lab Code: ACE

SDG NO.: Q3450

Continuing Calib Date: 10/27/2025

Initial Calibration Date(s): 10/21/2025 10/21/2025

Continuing Calib Time: 20:21

Initial Calibration Time(s): 09:06 17:01

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1254-1	(1)	5.66	5.66	5.56	5.76	0.00
Aroclor-1254-2	(2)	5.81	5.81	5.71	5.91	0.01
Aroclor-1254-3	(3)	6.21	6.21	6.11	6.31	0.00
Aroclor-1254-4	(4)	6.43	6.44	6.34	6.54	0.01
Aroclor-1254-5	(5)	6.85	6.85	6.75	6.95	0.00
Tetrachloro-m-xylene		3.78	3.79	3.69	3.89	0.01
Decachlorobiphenyl		8.78	8.79	8.69	8.89	0.01