

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** MATR02
Lab Code: ACE **SDG NO.:** Q3450
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/21/2025 10/21/2025

Client Sample No.: CCAL07 **Date Analyzed:** 10/31/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP076271.D **Time Analyzed:** 20:57

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.798	5.695	5.895	525.900	500.000	5.2
Aroclor-1016-2	5.819	5.717	5.917	526.230	500.000	5.2
Aroclor-1016-3	5.881	5.780	5.980	523.540	500.000	4.7
Aroclor-1016-4	5.979	5.877	6.077	529.160	500.000	5.8
Aroclor-1016-5	6.271	6.169	6.369	517.380	500.000	3.5
Aroclor-1260-1	7.390	7.287	7.487	557.240	500.000	11.4
Aroclor-1260-2	7.643	7.540	7.740	529.160	500.000	5.8
Aroclor-1260-3	8.001	7.899	8.099	545.500	500.000	9.1
Aroclor-1260-4	8.229	8.127	8.327	577.430	500.000	15.5
Aroclor-1260-5	8.555	8.452	8.652	526.510	500.000	5.3
Decachlorobiphenyl	10.423	10.319	10.519	50.300	50.000	0.6
Tetrachloro-m-xylene	4.645	4.542	4.742	53.750	50.000	7.5

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** MATR02
Lab Code: ACE **SDG NO.:** Q3450
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/21/2025 10/21/2025

Client Sample No.: CCAL07 **Date Analyzed:** 10/31/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP076271.D **Time Analyzed:** 20:57

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.876	4.778	4.978	532.360	500.000	6.5
Aroclor-1016-2	4.933	4.836	5.036	519.120	500.000	3.8
Aroclor-1016-3	5.054	4.954	5.154	468.650	500.000	-6.3
Aroclor-1016-4	5.093	4.996	5.196	569.830	500.000	14.0
Aroclor-1016-5	5.308	5.210	5.410	488.380	500.000	-2.3
Aroclor-1260-1	6.336	6.238	6.438	525.820	500.000	5.2
Aroclor-1260-2	6.522	6.425	6.625	535.740	500.000	7.1
Aroclor-1260-3	6.675	6.578	6.778	554.600	500.000	10.9
Aroclor-1260-4	7.144	7.047	7.247	597.580	500.000	19.5
Aroclor-1260-5	7.385	7.287	7.487	554.130	500.000	10.8
Decachlorobiphenyl	8.779	8.682	8.882	44.880	50.000	-10.2
Tetrachloro-m-xylene	3.781	3.681	3.881	55.450	50.000	10.9