

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** MATR02
Lab Code: ACE **SDG NO.:** Q3450
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/21/2025 10/21/2025

Client Sample No.: CCAL10 **Date Analyzed:** 11/03/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP076304.D **Time Analyzed:** 14:19

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.793	5.695	5.895	500.340	500.000	0.1
Aroclor-1016-2	5.815	5.717	5.917	502.560	500.000	0.5
Aroclor-1016-3	5.878	5.780	5.980	506.050	500.000	1.2
Aroclor-1016-4	5.974	5.877	6.077	509.790	500.000	2.0
Aroclor-1016-5	6.267	6.169	6.369	470.310	500.000	-5.9
Aroclor-1260-1	7.384	7.287	7.487	514.570	500.000	2.9
Aroclor-1260-2	7.637	7.540	7.740	507.870	500.000	1.6
Aroclor-1260-3	7.995	7.899	8.099	507.650	500.000	1.5
Aroclor-1260-4	8.223	8.127	8.327	545.660	500.000	9.1
Aroclor-1260-5	8.547	8.452	8.652	501.480	500.000	0.3
Decachlorobiphenyl	10.412	10.319	10.519	48.500	50.000	-3.0
Tetrachloro-m-xylene	4.641	4.542	4.742	53.160	50.000	6.3

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** MATR02
Lab Code: ACE **SDG NO.:** Q3450
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/21/2025 10/21/2025

Client Sample No.: CCAL10 **Date Analyzed:** 11/03/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP076304.D **Time Analyzed:** 14:19

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.874	4.778	4.978	532.670	500.000	6.5
Aroclor-1016-2	4.933	4.836	5.036	523.230	500.000	4.6
Aroclor-1016-3	5.051	4.954	5.154	511.130	500.000	2.2
Aroclor-1016-4	5.092	4.996	5.196	562.590	500.000	12.5
Aroclor-1016-5	5.306	5.210	5.410	502.290	500.000	0.5
Aroclor-1260-1	6.334	6.238	6.438	461.510	500.000	-7.7
Aroclor-1260-2	6.521	6.425	6.625	458.300	500.000	-8.3
Aroclor-1260-3	6.674	6.578	6.778	473.640	500.000	-5.3
Aroclor-1260-4	7.143	7.047	7.247	481.260	500.000	-3.7
Aroclor-1260-5	7.383	7.287	7.487	435.790	500.000	-12.8
Decachlorobiphenyl	8.775	8.682	8.882	41.250	50.000	-17.5
Tetrachloro-m-xylene	3.780	3.681	3.881	54.460	50.000	8.9