

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** MATR02
Lab Code: ACE **SDG NO.:** Q3450
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/10/2025 10/10/2025

Client Sample No.: CCAL12 **Date Analyzed:** 10/27/2025
Lab Sample No.: AR1660CCC500 **Data File :** PQ071161.D **Time Analyzed:** 12:35

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.479	4.380	4.580	550.580	500.000	10.1
Aroclor-1016-2	4.498	4.399	4.599	559.440	500.000	11.9
Aroclor-1016-3	4.556	4.456	4.656	571.120	500.000	14.2
Aroclor-1016-4	4.645	4.546	4.746	570.660	500.000	14.1
Aroclor-1016-5	4.930	4.831	5.031	565.870	500.000	13.2
Aroclor-1260-1	6.019	5.919	6.119	557.440	500.000	11.5
Aroclor-1260-2	6.274	6.175	6.375	551.940	500.000	10.4
Aroclor-1260-3	6.626	6.526	6.726	579.440	500.000	15.9
Aroclor-1260-4	6.839	6.740	6.940	591.430	500.000	18.3
Aroclor-1260-5	7.145	7.045	7.245	581.000	500.000	16.2
Decachlorobiphenyl	8.495	8.396	8.596	59.520	50.000	19.0
Tetrachloro-m-xylene	3.399	3.300	3.500	56.350	50.000	12.7

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** MATR02
Lab Code: ACE **SDG NO.:** Q3450
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/10/2025 10/10/2025

Client Sample No.: CCAL12 **Date Analyzed:** 10/27/2025
Lab Sample No.: AR1660CCC500 **Data File :** PQ071161.D **Time Analyzed:** 12:35

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	3.776	3.678	3.878	459.340	500.000	-8.1
Aroclor-1016-2	3.791	3.693	3.893	461.520	500.000	-7.7
Aroclor-1016-3	3.951	3.853	4.053	473.930	500.000	-5.2
Aroclor-1016-4	4.000	3.902	4.102	480.760	500.000	-3.8
Aroclor-1016-5	4.195	4.097	4.297	472.690	500.000	-5.5
Aroclor-1260-1	5.186	5.089	5.289	469.950	500.000	-6.0
Aroclor-1260-2	5.376	5.278	5.478	469.430	500.000	-6.1
Aroclor-1260-3	5.517	5.419	5.619	470.970	500.000	-5.8
Aroclor-1260-4	5.977	5.879	6.079	477.010	500.000	-4.6
Aroclor-1260-5	6.221	6.124	6.324	476.200	500.000	-4.8
Decachlorobiphenyl	7.531	7.433	7.633	55.990	50.000	12.0
Tetrachloro-m-xylene	2.776	2.678	2.878	50.180	50.000	0.4