

DATA REPORTING QUALIFIERS- INORGANIC

For reporting results, the following “ Results Qualifiers” are used:

J	Indicates the reported value was obtained from a reading that was less than the Contract Required Detection Limit (CRDL), but greater than or equal to the Instrument Detection Limit (IDL).
U	Indicates the analyte was analyzed for, but not detected.
ND	Indicates the analyte was analyzed for, but not detected
E	Indicates the reported value is estimated because of the presence of interference
M	Indicates Duplicate injection precision not met.
N	Indicates the spiked sample recovery is not within control limits.
S	Indicates the reported value was determined by the Method of Standard Addition (MSA).
*	Indicates that the duplicate analysis is not within control limits.
+	Indicates the correlation coefficient for the MSA is less than 0.995.
D	Indicates the reported value is from a secondary analysis with a dilution factor. The original analysis exceeded the calibration range.
M	Method qualifiers “P” for ICP instrument “PM” for ICP when Microwave Digestion is used “CV” for Manual Cold Vapor AA “AV” for automated Cold Vapor AA “CA” for MIDI-Distillation Spectrophotometric “AS” for Semi -Automated Spectrophotometric “C” for Manual Spectrophotometric “T” for Titrimetric “NR” for analyte not required to be analyzed
OR	Indicates the analyte’s concentration exceeds the calibrated range of the instrument for that specific analysis.
Q	Indicates the LCS did not meet the control limits requirements
H	Sample Analysis Out Of Hold Time

LAB CHRONICLE

OrderID:	Q3455	OrderDate:	10/24/2025 12:10:00 PM
Client:	Europastray	Project:	MCUA Permit No 14241 - 571 Jersey Ave NB NJ
Contact:	Kevin Carlucci	Location:	D31

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q3455-01	MH-10242025	WATER			10/24/25 09:50			10/24/25
			Oil and Grease	1664A			10/27/25 10:00	
			TPH	1664A			10/27/25 10:30	



SAMPLE DATA

Report of Analysis

Client: Europastry
Project: MCUA Permit No 14241 - 571 Jersey Ave NB NJ
Client Sample ID: MH-10242025
Lab Sample ID: Q3455-01

Date Collected: 10/24/25 09:50
Date Received: 10/24/25
SDG No.: Q3455
Matrix: WATER
% Solid: 0

Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
Oil and Grease	267		1	0.29	5.00	mg/L		10/27/25 10:00	1664A
TPH	44.3		1	0.29	5.00	mg/L		10/27/25 10:30	1664A

Comments:

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
D = Dilution
Q = indicates LCS control criteria did not meet requirements
H = Sample Analysis Out Of Hold Time

J = Estimated Value
B = Analyte Found in Associated Method Blank
* = indicates the duplicate analysis is not within control limits.
E = Indicates the reported value is estimated because of the presence of interference.
OR = Over Range
N = Spiked sample recovery not within control limits



QC RESULT SUMMARY

Preparation Blank Summary

Client: Europastry

SDG No.: Q3455

Project: MCUA Permit No 14241 - 571 Jersey Ave NB NJ

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: LB137651BL							
TPH	mg/L	< 2.5000	2.5000	U	0.29	5.0	10/27/2025
Sample ID: LB137655BL							
Oil and Grease	mg/L	< 2.5000	2.5000	U	0.29	5.0	10/27/2025

Matrix Spike Summary

Client:	Europastry	SDG No.:	Q3455
Project:	MCUA Permit No 14241 - 571 Jersey Ave NB NJ	Sample ID:	Q3448-03
Client ID:	DSN001MS	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Oil and Grease	mg/L	78-114	22.4		2.30	J	20.0	1	101		10/27/2025

Matrix Spike Summary

Client:	Europastry	SDG No.:	Q3455
Project:	MCUA Permit No 14241 - 571 Jersey Ave NB NJ	Sample ID:	Q3448-03
Client ID:	DSN001MSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Oil and Grease	mg/L	78-114	22.6		2.30	J	20.0	1	102		10/27/2025

Matrix Spike Summary

Client:	Europastry	SDG No.:	Q3455
Project:	MCUA Permit No 14241 - 571 Jersey Ave NB NJ	Sample ID:	Q3455-01
Client ID:	Q3455 -01MS	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Oil and Grease	mg/L	78-114	287		267		20.0	1	101		10/27/2025

Matrix Spike Summary

Client:	Europastry	SDG No.:	Q3455
Project:	MCUA Permit No 14241 - 571 Jersey Ave NB NJ	Sample ID:	Q3455-01
Client ID:	Q3455 -01MSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Oil and Grease	mg/L	78-114	287		267		20.0	1	102		10/27/2025

Duplicate Sample Summary

Client: Europastry	SDG No.: Q3455
Project: MCUA Permit No 14241 - 571 Jersey Ave NB NJ	Sample ID: LB137651BS
Client ID: LB137651BSD	Percent Solids for Spike Sample: 0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
TPH	mg/L	+/-18	16.7		17.0		1	1.78		10/27/2025

Duplicate Sample Summary

Client: Europastry	SDG No.: Q3455
Project: MCUA Permit No 14241 - 571 Jersey Ave NB NJ	Sample ID: Q3448-03
Client ID: DSN001MSD	Percent Solids for Spike Sample: 0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Oil and Grease	mg/L	+/-18	22.4		22.6		1	0.89		10/27/2025

Duplicate Sample Summary

Client:	Europastry	SDG No.:	Q3455
Project:	MCUA Permit No 14241 - 571 Jersey Ave NB NJ	Sample ID:	Q3455-01
Client ID:	Q3455 -01MSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Oil and Grease	mg/L	+/-18	287		287		1	0.03		10/27/2025

Laboratory Control Sample Summary

Client: Europastry

SDG No.: Q3455

Project: MCUA Permit No 14241 - 571 Jersey Ave NB NJ

Run No.: LB137651

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB137651BS							
TPH	mg/L	20.0	16.7		84	1	78-114	10/27/2025

Laboratory Control Sample Summary

Client:	Europastry	SDG No.:	Q3455
Project:	MCUA Permit No 14241 - 571 Jersey Ave NB NJ	Run No.:	LB137651

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB137651BSD							
TPH	mg/L	20.0	17.0		85	1	78-114	10/27/2025

Laboratory Control Sample Summary

Client:	Europastry	SDG No.:	Q3455
Project:	MCUA Permit No 14241 - 571 Jersey Ave NB NJ	Run No.:	LB137655

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB137655BS							
Oil and Grease	mg/L	20.0	16.7		84	1	78-114	10/27/2025



RAW DATA

Extraction and Analytical Summary Report

Analysis Method: 1664A
Test: TPH
Run Number: LB137651
Analysis Date: 10/27/2025
BalanceID: WC SC-5
OvenID: EXT OVEN-3

ANALYST: jignesh
REVIEWED BY: Iwona
Extraction Date: 10/27/2025
Extraction IN Time: 09:00
Extraction OUT Time: 09:35
Thermometer ID: EXT OVEN#3

Dish #	Lab ID	Client ID	Matrix	pH	Sample Vol (ml)	Final Volume (ml)	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Silica Gel Weight (g)	Weight After Drying (g)	Final Weight After Drying (g)	Change Weight (g)	Result in ppm
1	LB137651BL	LB137651BL	WATER	1.3	1000	100	2.4416	2.4416	3.01	2.4417	2.4417	0.0001	0.1
2	LB137651BS	LB137651BS	WATER	1.3	1000	100	3.1258	3.1258	3.02	3.1425	3.1425	0.0167	16.7
3	LB137651BSD	LB137651BSD	WATER	1.3	1000	100	2.7013	2.7013	3.03	2.7183	2.7183	0.0170	17
4	Q3428-01	Grab	WATER	1.6	1000	100	3.0527	3.0527	3.05	3.0589	3.0589	0.0062	6.2
5	Q3434-02	PR132-WC2-20251022	WATER	1.3	1000	100	3.0375	3.0375	3.02	3.0409	3.0409	0.0034	3.4
6	Q3455-01	MH-10242025	WATER	1.6	1000	100	3.0349	3.0349	3.04	3.0792	3.0792	0.0443	44.3

QC Batch# LB137651

Test: TPH

Analysis Date: 10/27/2025

Chemicals Used:

Chemical Name	Chemical Lot #
HEXANE	W3240
pH Paper 0-14	M6069
Sodium Sulfate	EP2655
1:1 HCL	WP115016
Silica Gel	W3246
Sand	N/A

Standards Used:

Standard Name	Amount Used	Standard Lot #
LCSW	5.00 ML	WP115017
LCSWD	5.00 ML	WP115018
MS/MSD	N/A	N/A

BALANCE CALIBRATION / OVEN Dessicator Data

Analytical Balance ID # : WC SC-6

Before Analysis

0.0020 gram Balance: 0.0018 (0.0018-0.0022) In OVEN TEMP1 : 71 °C Dessicator Time In1 : 11:26

1.0000 gram Balance: 1.0004 (0.9950-1.0050) In Time1: 10:30

Bal Check Time: 09:10 Out OVEN TEMP1: 71 °C Dessicator Time Out1: 12:10

Out Time1: 11:25

After Analysis

0.0020 gram Balance: 0.0019 (0.0018-0.0022) In OVEN TEMP2 : 70 °C Dessicator Time In2 : 13:26

1.0000 gram Balance: 1.0005 (0.9950-1.0050) In Time2: 12:47

Bal Check Time: 14:02 Out OVEN TEMP2: 71 °C Dessicator Time Out2: 14:00

Out Time2: 13:25

WORKLIST(Hardcopy Internal Chain)

VB 13751

WorkList Name : tph q3455 WorkList ID : 192684 Department : Wet-Chemistry Date : 10-27-2025 07:53:22

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q3428-01	Grab	Water	TPH	Conc H2SO4 to pH < 2	ARAM01	J13	10/22/2025	1664A
Q3434-02	PR132-WC2-20251022	Water	TPH	Conc H2SO4 to pH < 2	AECO02	D41	10/22/2025	1664A
Q3455-01	MH-10242025	Water	TPH	Conc H2SO4 to pH < 2	EURO03	D31	10/24/2025	1664A

Date/Time 10/24/25 08:10
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]

Date/Time 10/24/25 15:30
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]

Extraction and Analytical Summary Report

Analysis Method: 1664A

Test: Oil and Grease

Run Number: LB137655

Analysis Date: 10/27/2025

BalanceID: WC SC-5

OvenID: EXT OVEN-3

ANALYST: jignesh

REVIEWED BY: Iwona

Extraction Date: 10/27/2025

Extraction IN Time: 08:15

Extraction OUT Time: 08:50

Thermometer ID: EXT OVEN#3

Dish #	Lab ID	Client ID	Matrix	pH	Sample Vol (ml)	Final Volume (ml)	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Silica Gel Weight (g)	Weight After Drying (g)	Final Weight After Drying (g)	Change Weight (g)	Result in ppm
1	LB137655BL	LB137655BL	WATER	1.3	1000	100	2.8563	2.8563	0	2.8564	2.8564	0.0001	0.1
2	LB137655BS	LB137655BS	WATER	1.3	1000	100	3.1474	3.1474	0	3.1641	3.1641	0.0167	16.7
3	Q3420-01	Oil and Grease #1	WATER	1.3	1000	100	3.0918	3.0918	0	3.0950	3.0950	0.0032	3.2
4	Q3420-02	Oil and Grease #2	WATER	1.3	1000	100	3.0632	3.0632	0	3.0672	3.0672	0.0040	4
5	Q3420-03	Oil and Grease #3	WATER	1.3	1000	100	3.0686	3.0686	0	3.0698	3.0698	0.0012	1.2
6	Q3441-01	001 Willets Pt Blvd (C	WATER	1.3	1000	100	3.0363	3.0363	0	3.0385	3.0385	0.0022	2.2
7	Q3441-02	002 35th Ave (Oct)	WATER	1.3	1000	100	3.0489	3.0489	0	3.0506	3.0506	0.0017	1.7
8	Q3448-03	DSN001	WATER	1.3	1000	100	3.0649	3.0649	0	3.0672	3.0672	0.0023	2.3
9	Q3448-05	Q3448-03MS	WATER	1.3	1000	100	2.7449	2.7449	0	2.7673	2.7673	0.0224	22.4
10	Q3448-06	Q3448-03MSD	WATER	1.3	1000	100	2.8903	2.8903	0	2.9129	2.9129	0.0226	22.6
11	Q3455-01	MH-10242025	WATER	1.6	1000	100	3.0228	3.0228	0	3.2893	3.2893	0.2665	266.5
12	Q3455-02	Q3455 -1MS	WATER	1.6	1000	100	2.4563	2.4563	0	2.7430	2.7430	0.2867	286.7
13	Q3455-03	Q3455 -1MSD	WATER	1.6	1000	100	2.9633	2.9633	0	3.2501	3.2501	0.2868	286.8
14	Q3462-01	SW-1	WATER	1.6	1000	100	3.0504	3.0504	0	3.0557	3.0557	0.0053	5.3

QC Batch# LB137655

Test: Oil and Grease

Analysis Date: 10/27/2025

Chemicals Used:

Chemical Name	Chemical Lot #
HEXANE	W3240
pH Paper 0-14	M6069
Sodium Sulfate	EP2655
1:1 HCL	WP115016
Silica Gel	W3246
Sand	N/A

Standards Used:

Standard Name	Amount Used	Standard Lot #
LCSW	2.5 ML	WP115017
LCSWD	N/A	N/A
MS/MSD	2.5 ML	WP115018

BALANCE CALIBRATION / OVEN Dessicator Data

Analytical Balance ID # : WC SC-6

Before Analysis

0.0020 gram Balance: 0.0018 (0.0018-0.0022) In OVEN TEMP1 : 70 °C Dessicator Time In1 : 10:46

1.0000 gram Balance: 1.0004 (0.9950-1.0050) In Time1: 10:00

Bal Check Time: 08:20 Out OVEN TEMP1: 70 °C Dessicator Time Out1: 11:25

Out Time1: 10:45

After Analysis

0.0020 gram Balance: 0.0019 (0.0018-0.0022) In OVEN TEMP2 : 71 °C Dessicator Time In2 : 12:31

1.0000 gram Balance: 1.0005 (0.9950-1.0050) In Time2: 12:00

Bal Check Time: 13:15 Out OVEN TEMP2: 71 °C Dessicator Time Out2: 13:10

Out Time2: 12:30

Instrument ID: WC SC-3

Daily Analysis Runlog For Sequence/QC Batch ID # LB137651

Review By	jignesh	Review On	10/27/2025 10:03:35 AM
Supervise By	Iwona	Supervise On	10/28/2025 10:55:05 AM
SubDirectory	LB137651	Test	TPH
STD. NAME	STD REF.#		
ICAL Standard	N/A		
ICV Standard	N/A		
CCV Standard	N/A		
ICSA Standard	N/A		
CRI Standard	N/A		
LCS Standard	N/A		
Chk Standard	W3240,M6069,EP2655,WP115016,W3246,N/A,WP115017,WP115018,N/A		

Sr#	SampleID	ClientID	QcType	Date	Comment	Operator	Status
1	LB137651BL	LB137651BL	MB	10/27/25 10:30		jignesh	OK
2	LB137651BS	LB137651BS	LCS	10/27/25 10:30		jignesh	OK
3	LB137651BSD	LB137651BSD	LCSD	10/27/25 10:30		jignesh	OK
4	Q3428-01	Grab	SAM	10/27/25 10:30		jignesh	OK
5	Q3434-02	PR132-WC2-2025102	SAM	10/27/25 10:30		jignesh	OK
6	Q3455-01	MH-10242025	SAM	10/27/25 10:30		jignesh	OK

Instrument ID: WC SC-3

Daily Analysis Runlog For Sequence/QC Batch ID # LB137655

Review By	jignesh	Review On	10/27/2025 11:57:24 AM
Supervise By	Iwona	Supervise On	10/27/2025 12:08:12 PM
SubDirectory	LB137655	Test	Oil and Grease
STD. NAME	STD REF.#		
ICAL Standard	N/A		
ICV Standard	N/A		
CCV Standard	N/A		
ICSA Standard	N/A		
CRI Standard	N/A		
LCS Standard	N/A		
Chk Standard	W3240,M6069,EP2655,WP115016,W3246,N/A,WP115017,N/A,WP115018		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	LB137655BL	LB137655BL	MB	10/27/25 10:00		jignesh	OK
2	LB137655BS	LB137655BS	LCS	10/27/25 10:00		jignesh	OK
3	Q3420-01	Oil and Grease #1	SAM	10/27/25 10:00		jignesh	OK
4	Q3420-02	Oil and Grease #2	SAM	10/27/25 10:00		jignesh	OK
5	Q3420-03	Oil and Grease #3	SAM	10/27/25 10:00		jignesh	OK
6	Q3441-01	001 Willets Pt Blvd (O	SAM	10/27/25 10:00		jignesh	OK
7	Q3441-02	002 35th Ave (Oct)	SAM	10/27/25 10:00		jignesh	OK
8	Q3448-03	DSN001	SAM	10/27/25 10:00		jignesh	OK
9	Q3448-05	Q3448-03MS	MS	10/27/25 10:00		jignesh	OK
10	Q3448-06	Q3448-03MSD	MSD	10/27/25 10:00		jignesh	OK
11	Q3455-01	MH-10242025	SAM	10/27/25 10:00		jignesh	OK
12	Q3455-02	Q3455 -01MS	MS	10/27/25 10:00		jignesh	OK
13	Q3455-03	Q3455 -01MSD	MSD	10/27/25 10:00		jignesh	OK
14	Q3462-01	SW-1	SAM	10/27/25 10:00		jignesh	OK

Prep Standard - Chemical Standard Summary

Order ID : Q3455

Test : Oil and Grease,TPH

Prepbatch ID :

Sequence ID/Qc Batch ID: LB137651, LB137655,

Standard ID :

EP2655, WP115016, WP115017, WP115018,

Chemical ID :

E3875, E3972, M6069, M6151, W2817, W2871, W3009, W3082, W3112, W3240, W3246,



<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3923	Baked Sodium Sulfate	EP2655	10/24/2025	01/28/2026	RUPESHKUMAR SHAH	Extraction_SCALE_2 (EX-SC-2)	None	Riteshkumar Patel 10/24/2025
<u>FROM</u> 4000.00000gram of E3875 = Final Quantity: 4000.000 gram								

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
229	1:1 HCL	WP115016	10/02/2025	02/17/2026	Jignesh Parikh	None	None	Iwona Zarych 10/02/2025
<u>FROM</u> 500.00000ml of M6151 + 500.00000ml of W3112 = Final Quantity: 1.000 L								



<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
2470	1664A SPIKING SOLN	WP115017	10/02/2025	04/02/2026	Jignesh Parikh	WETCHEM_SCALE_7 (WCS-6)	None	Iwona Zarych 10/02/2025
<u>FROM</u>	1000.00000ml of E3972 + 4.00000gram of W2817 + 4.00000gram of W2871 = Final Quantity: 1000.000 ml							

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3374	1664A QCS spiking solution-SS	WP115018	10/02/2025	04/02/2026	Jignesh Parikh	WETCHEM_SCALE_7 (WC-6)	None	Iwona Zarych 10/02/2025
<u>FROM</u>	1000.00000ml of E3972 + 4.00000gram of W3009 + 4.00000gram of W3082 = Final Quantity: 1000.000 ml							

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	PC19631-100 / SODIUM SULFATE, ANHYDROUS, PEST GRADE, 1	417203	01/28/2026	07/28/2025 / RUPESH	01/29/2025 / Rajesh	E3875

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9254-03 / Acetone, Ultra Resi (cs/4x4L)	24H1462005	05/24/2027	09/16/2025 / Evelyn	09/04/2025 / Riteshkumar	E3972

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	140440 / TEST PAPERS,PH,0-2.5,.2SENSI, 100PK	80A0441	02/29/2028	09/03/2024 / jignesh	08/19/2024 / Jaswal	M6069

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9530-33 / Hydrochloric Acid, Instra-Analyzed (cs/6x2.5L)	22G2862015	02/17/2026	02/18/2025 / Sagar	01/15/2025 / Sagar	M6151

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	A12244 / Stearic acid, 98%, 100 g	U20E006	04/02/2026	04/02/2021 / apatel	04/02/2021 / apatel	W2817

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	H223-57 / Hexadecane, 99.0%	0000266903	05/04/2027	09/07/2021 / apatel	08/26/2021 / apatel	W2871

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	H223-57 / Hexadecane, 99.0%	SHBP8192	02/27/2028	02/27/2023 / Iwona	02/27/2023 / Iwona	W3009

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	A12244 / Stearic acid, 98%, 100 g	U23E020	02/26/2029	02/26/2024 / Iwona	02/26/2024 / Iwona	W3082

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	DIW / DI Water	Daily Lab-Certified	07/03/2029	07/03/2024 / Iwona	07/03/2024 / Iwona	W3112

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9262-03 / Hexane, Ultra-Resi (cs/4x4L)	25C0362006	04/30/2026	09/15/2025 / JIGNESH	09/12/2025 / JIGNESH	W3240

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	04667-2.5 / Silica Gel (60-200 mesh), 2.5 KG	072154301	10/03/2030	10/03/2025 / Iwona	10/03/2025 / Iwona	W3246

Hexadecane, 99.0%



Material No.: H223-57
Batch No.: 0000266903
Manufactured Date: 2020/05/05
Retest Date: 2027/05/04
Revision No: 1

Certificate of Analysis

Test	Specification	Result
Assay ($\text{CH}_3(\text{CH}_2)_{14}\text{CH}_3$) (by GC)	$\geq 99.0 \%$	99.3
Infrared Spectrum	Passes Test	PT

For Laboratory, Research or Manufacturing Use

Country of Origin: US
Packaging Site: Paris Mfg Ctr & DC


Jamie Ethier
Vice President Global Quality

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700
Avantor Performance Materials, LLC
100 Matsonford Rd, Suite 200, Radnor, PA 19087. U.S.A. Phone: 610.386.1700

Product Name: Stearic acid, 98%, Thermo Scientific Chemicals
Catalog Number: A12244.14

CAS Number: 57-11-4
Molecular Formula: C₁₈H₃₆O₂
Molecular Weight: 284.48
InChI Key: QIQXTHQIDYTRH-UHFFFAOYSA-N
SMILES: CCCCCCCCCCCCCCCC(O)=O
Synonym: stearic acid acide stearique hydrofol acid 1855 hydrofol acid 1655 industrene 5016
stearic acid, ion(1-) (8Cl) glycon TP glycon DP acidum stearinicum hydrofol acid 150

Product Specification

Appearance (Color): White
Form: Crystals or powder or crystalline powder or flakes or waxy solid
Assay (Silylated GC): ≥97.5%
Melting Point (clear melt): 67.0-74.0°C

Date Of Print: 11/30/2023

Product Specifications are subject to amendment and may change over time. Data contained is accurate as of the date printed.

W3009
rec. 2/27/2023 12

Product Name:

Hexadecane - ReagentPlus®, 99%

Certificate of Analysis

Product Number:

H6703

Batch Number:

SHBP8192

 $\text{CH}_3(\text{CH}_2)_{14}\text{CH}_3$

Brand:

SIAL

CAS Number:

544-76-3

MDL Number:

MFCD00008998

Formula:

C16H34

Formula Weight:

226.44 g/mol

Quality Release Date:

04 AUG 2022

Test	Specification	Result
Appearance (Color)	Colorless or White	Colorless
Appearance (Form)	Liquid or Solid	Liquid
Infrared Spectrum	Conforms to Structure	Conforms
Refractive index at 20 ° C	1.432 - 1.436	1.435
Purity (GC)	≥ 98.5 %	99.3 %
Color Test	≤ 20 APHA	< 5 APHA


Larry Coers, Director

Quality Control

Sheboygan Falls, WI US

Sigma-Aldrich warrants, that at the time of the quality release or subsequent retest date this product conformed to the information contained in this publication. The current Specification sheet may be available at Sigma-Aldrich.com. For further inquiries, please contact Technical Service. Purchaser must determine the suitability of the product for its particular use. See reverse side of invoice or packing slip for additional terms and conditions of sale.





**PRODUCTOS
QUÍMICOS
MONTERREY, S.A. DE C.V.**

MIRADOR 201, COL. MIRADOR
MONTERREY, N.L. MÉXICO
CP 64070
TEL +52 81 13 52 67 67
www.pqm.com.mx

CERTIFICATE OF ANALYSIS

PRODUCT :	SODIUM SULFATE CRYSTALS ANHYDROUS		
QUALITY :	ACS (CODE RMB3375)	FORMULA :	Na ₂ SO ₄
SPECIFICATION NUMBER:	6399	RELEASE DATE:	MAY/23/2024
LOT NUMBER :	417203		

TEST	SPECIFICATIONS	LOT VALUES
Assay (Na ₂ SO ₄)	Min. 99.0%	99.8 %
pH of a 5% solution at 25°C	5.2 - 9.2	6.2
Insoluble matter	Max. 0.01%	0.001 %
Loss on ignition	Max. 0.5%	0.1 %
Chloride (Cl)	Max. 0.001%	<0.001 %
Nitrogen compounds (as N)	Max. 5 ppm	<5 ppm
Phosphate (PO ₄)	Max. 0.001%	<0.001 %
Heavy metals (as Pb)	Max. 5 ppm	<5 ppm
Iron (Fe)	Max. 0.001%	<0.001 %
Calcium (Ca)	Max. 0.01%	0.001 %
Magnesium (Mg)	Max. 0.005%	0.001 %
Potassium (K)	Max. 0.008%	0.001 %
Extraction-concentration suitability	Passes test	Passes test
Appearance	Passes test	Passes test
Identification	Passes test	Passes test
Solubility and foreign matter	Passes test	Passes test
Retained on US Standard No. 10 sieve	Max. 1%	0.2 %
Retained on US Standard No. 60 sieve	Min. 94%	96.2 %
Through US Standard No. 60 sieve	Max. 5%	3.5 %
Through US Standard No. 100 sieve	Max. 10%	0.1 %

COMMENTS

QC: PhC Irma Belmares

If you need further details, please call our factory or contact our local distributor.

RE-02-01, Ed. 3

E 3875

Acetone
BAKER RESI-ANALYZED® Reagent
For Organic Residue Analysis

avantor™



Material No.: 9254-03

Batch No.: 24H1462005

Manufactured Date: 2024-05-24

Expiration Date: 2027-05-24

Revision No.: 0

Certificate of Analysis

Test	Specification	Result
Assay ((CH ₃) ₂ CO) (by GC, corrected for water)	>= 99.4 %	99.8 %
Color (APHA)	<= 10	5
Residue after Evaporation	<= 1.0 ppm	0.2 ppm
Substances Reducing Permanganate	Passes Test	Passes Test
Titration Acid (µeq/g)	<= 0.3	0.2
Titration Base (µeq/g)	<= 0.6	<0.1
Water (H ₂ O)	<= 0.5 %	0.2 %
FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL)	<= 5	<1
ECD Sensitive Impurities (as Heptachlor Epoxide) Single Peak (pg/mL)	<= 10	1

For Laboratory, Research, or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: United States
Packaging Site: Phillipsburg Mfg Ctr & DC

E3972

Jamie Croak
Director Quality Operations, Bioscience Production

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

Avantor Performance Materials LLC



Certificate of Analysis

Product information

Product	pH-Fix 0.3-2.3
REF	92180
LOT	80A0441
Expiration date:	29.02.2028
Date of examination:	23.01.2024
Gradation:	pH 0.3-0.7-1.0-1.3-1.6-1.9-2.3

Confirmation

Hereby we confirm, that the above mentioned product has successfully passed our quality control system in accordance with ISO 9001 and meets the specific quality criteria.

This document has been produced electronically and is valid without a signature.



Hydrochloric Acid, 36.5–38.0%
BAKER INSTRA-ANALYZED® Reagent
For Trace Metal Analysis

 **avantor™**



M6151

R → 11/15/25

Material No.: 9530-33
Batch No.: 22G2862015
Manufactured Date: 2022-06-15
Retest Date: 2027-06-14
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
ACS – Assay (as HCl) (by acid–base titrn)	36.5 – 38.0 %	37.9 %
ACS – Color (APHA)	≤ 10	5
ACS – Residue after Ignition	≤ 3 ppm	< 1 ppm
ACS – Specific Gravity at 60°/60°F	1.185 – 1.192	1.191
ACS – Bromide (Br)	≤ 0.005 %	< 0.005 %
ACS – Extractable Organic Substances	≤ 5 ppm	< 1 ppm
ACS – Free Chlorine (as Cl ₂)	≤ 0.5 ppm	< 0.5 ppm
Phosphate (PO ₄)	≤ 0.05 ppm	< 0.03 ppm
Sulfate (SO ₄)	≤ 0.5 ppm	< 0.3 ppm
Sulfite (SO ₃)	≤ 0.8 ppm	0.3 ppm
Ammonium (NH ₄)	≤ 3 ppm	< 1 ppm
Trace Impurities – Arsenic (As)	≤ 0.010 ppm	< 0.003 ppm
Trace Impurities – Aluminum (Al)	≤ 10.0 ppb	1.3 ppb
Arsenic and Antimony (as As)	≤ 5.0 ppb	< 3.0 ppb
Trace Impurities – Barium (Ba)	≤ 1.0 ppb	0.2 ppb
Trace Impurities – Beryllium (Be)	≤ 1.0 ppb	< 0.2 ppb
Trace Impurities – Bismuth (Bi)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities – Boron (B)	≤ 20.0 ppb	< 5.0 ppb
Trace Impurities – Cadmium (Cd)	≤ 1.0 ppb	< 0.3 ppb
Trace Impurities – Calcium (Ca)	≤ 50.0 ppb	163.0 ppb
Trace Impurities – Chromium (Cr)	≤ 1.0 ppb	0.7 ppb
Trace Impurities – Cobalt (Co)	≤ 1.0 ppb	< 0.3 ppb
Trace Impurities – Copper (Cu)	≤ 1.0 ppb	< 0.1 ppb
Trace Impurities – Gallium (Ga)	≤ 1.0 ppb	< 0.2 ppb
Trace Impurities – Germanium (Ge)	≤ 3.0 ppb	< 2.0 ppb
Trace Impurities – Gold (Au)	≤ 4.0 ppb	0.6 ppb
Heavy Metals (as Pb)	≤ 100 ppb	< 50 ppb
Trace Impurities – Iron (Fe)	≤ 15 ppb	6 ppb

>>> Continued on page 2 >>>

Hydrochloric Acid, 36.5–38.0%
BAKER INSTRA-ANALYZED® Reagent
For Trace Metal Analysis

 **avantorsm**



Material No.: 9530-33
Batch No.: 22G2862015

Test	Specification	Result
Trace Impurities – Lead (Pb)	≤ 1.0 ppb	< 0.5 ppb
Trace Impurities – Lithium (Li)	≤ 1.0 ppb	< 0.2 ppb
Trace Impurities – Magnesium (Mg)	≤ 10.0 ppb	2.9 ppb
Trace Impurities – Manganese (Mn)	≤ 1.0 ppb	< 0.4 ppb
Trace Impurities – Mercury (Hg)	≤ 0.5 ppb	0.1 ppb
Trace Impurities – Molybdenum (Mo)	≤ 10.0 ppb	< 3.0 ppb
Trace Impurities – Nickel (Ni)	≤ 4.0 ppb	< 0.3 ppb
Trace Impurities – Niobium (Nb)	≤ 1.0 ppb	0.8 ppb
Trace Impurities – Potassium (K)	≤ 9.0 ppb	< 2.0 ppb
Trace Impurities – Selenium (Se), For Information Only		< 1.0 ppb
Trace Impurities – Silicon (Si)	≤ 100.0 ppb	< 10.0 ppb
Trace Impurities – Silver (Ag)	≤ 1.0 ppb	0.5 ppb
Trace Impurities – Sodium (Na)	≤ 100.0 ppb	2.3 ppb
Trace Impurities – Strontium (Sr)	≤ 1.0 ppb	< 0.2 ppb
Trace Impurities – Tantalum (Ta)	≤ 1.0 ppb	1.6 ppb
Trace Impurities – Thallium (Tl)	≤ 5.0 ppb	< 2.0 ppb
Trace Impurities – Tin (Sn)	≤ 5.0 ppb	4.0 ppb
Trace Impurities – Titanium (Ti)	≤ 1.0 ppb	1.5 ppb
Trace Impurities – Vanadium (V)	≤ 1.0 ppb	< 0.2 ppb
Trace Impurities – Zinc (Zn)	≤ 5.0 ppb	0.8 ppb
Trace Impurities – Zirconium (Zr)	≤ 1.0 ppb	0.3 ppb

>>> Continued on page 3 >>>

Hydrochloric Acid, 36.5–38.0%
BAKER INSTRA–ANALYZED® Reagent
For Trace Metal Analysis



Material No.: 9530-33
Batch No.: 22G2862015

Test	Specification	Result
------	---------------	--------

For Laboratory, Research, or Manufacturing Use
Product Information (not specifications):
Appearance (clear, fuming liquid)
Meets ACS Specifications
Storage Condition: Store below 25 °C.

Country of Origin: USA
Packaging Site: Phillipsburg Mfg Ctr & DC

A handwritten signature in cursive script that reads 'Jamie Ethier'.
Jamie Ethier
Vice President Global Quality

Certificate of analysis

W3082 Received on 2/26/2026 by IZ

Product No.: A12244
Product: Stearic acid, 98%
Lot No.: U23E020

Appearance White flakes
Assay 98.7 %

This document has been electronically generated and does not require a signature.

Order our products online www.alfa.com

ThermoFisher
S C I E N T I F I C

n-Hexane 95%
ULTRA RESI-ANALYZED
For Organic Residue Analysis

 **avantorsTM**



W3240
JP
07/15/2025
07/15/2025

Material No.: 9262-03
Batch No.: 25C0362006
Manufactured Date: 2025-01-29
Expiration Date: 2026-04-30
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL)	≤ 5	1
ECD Sensitive Impurities (as HeptachlorEpoxide) Single Peak (pg/mL)	≤ 10	6
ECD-Sensitive Impurities (as EthyleneDibromide) - Single Impurity Peak (ng/mL)	≤ 5	4
Assay (Total Saturated C ₆ Isomers) (by GC, corrected for water)	$\geq 99.5 \%$	100.0 %
Assay (as n-Hexane) (by GC, corrected for water)	$\geq 95 \%$	100 %
Color (APHA)	≤ 10	10
Residue after Evaporation	≤ 1.0 ppm	0.2 ppm
Substances Darkened by H ₂ SO ₄	Passes Test	Passes Test
Water (by KF, coulometric)	$\leq 0.05 \%$	$< 0.01 \%$

For Laboratory, Research, or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: United States
Packaging Site: Phillipsburg Mfg Ctr & DC



Jamie Croak
Director Quality Operations, Bioscience Production

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

Avantor Performance Materials LLC

Certificate of Analysis

Product information

Product: Silica 60, 0.063 - 0.200 mm
REF: 815330.25
LOT: 072154301

Technical data

Material: Synthetic amorphous silica (irregular shaped)
Description: White powder

Parameter	Specifications	Result
Specific surface (m ² /g, N2 adsorption) :	450 - 550	537
Particle size distribution (screen analysis) :	< 63 µm max. 5 %	0.3
	> 200 µm max. 5 %	0.1
pH value :	6.0 - 7.5	7
Water content (%) :	< 7	3.6
Pore volume (mL/g, N2 adsorption) :	0.65 - 0.85	0.82
Mean pore size (Å, N2 adsorption) :	50 - 70	62

Expiry

This product has no stated expiration date or shelf life.

We recommend to use the product within a time period of 5 years after date of QC release.

This time period is valid only if the product is stored under dry and frost-free conditions.

After 5 years we recommend retesting the adsorbent to make sure that the expected performance is still given.

Confirmation

Hereby we confirm, that the above mentioned product has successfully passed our quality control system in accordance with ISO 9001 and meets the specific quality criteria.

This document has been produced electronically and is valid without a signature.

Date of measurement: 16.02.2023 22:00



SHIPPING DOCUMENTS

CLIENT INFORMATION

CLIENT PROJECT INFORMATION

CLIENT BILLING INFORMATION

REPORT TO BE SENT TO:

COMPANY: Europastyl
ADDRESS: 571 Jersey Ave
CITY New Brunswick STATE: NJ ZIP: 08901
ATTENTION: Kenn Carducci
PHONE: (631) 563-6262 FAX:

PROJECT NAME: MCUA Permit No. 14241-571 Jersey Ave
PROJECT NO.: LOCATION:
PROJECT MANAGER:
e-mail:
PHONE: FAX:

BILL TO: PO#:
ADDRESS:
CITY STATE: ZIP:
ATTENTION: PHONE:

ANALYSIS

DATA TURNAROUND INFORMATION

DATA DELIVERABLE INFORMATION

FAX (RUSH) DAYS*
HARDCOPY (DATA PACKAGE): DAYS*
EDD: DAYS*

*TO BE APPROVED BY CHEMTECH

STANDARD HARDCOPY TURNAROUND TIME IS 10 BUSINESS

☐ Level 1 (Results Only) ☐ Level 4 (QC + Full Raw Data)
☐ Level 2 (Results + QC) ☐ NJ Reduced ☐ US EPA CLP
☐ Level 3 (Results + QC) ☐ NYS ASP A ☐ NYS ASP B
+ Raw Data ☐ Other
☐ EDD FORMAT

DPG
TBH

1 2 3 4 5 6 7 8 9

PRESERVATIVES

COMMENTS

ALLIANCE SAMPLE ID	PROJECT SAMPLE IDENTIFICATION	SAMPLE MATRIX	SAMPLE TYPE		SAMPLE COLLECTION		# OF BOTTLES										COMMENTS
			COMP	GRAB	DATE	TIME		1	2	3	4	5	6	7	8	9	
1.	MT-10242025	W.		X	10/24	0950	4	X	X								← Specify Preservatives A-HCl D-NaOH B-HNO3 E-ICE C-H2SO4 F-OTHER
2.																	
3.																	
4.																	
5.																	
6.																	
7.																	
8.																	
9.																	
10.																	

SAMPLE CUSTODY MUST BE DOCUMENTED BELOW EACH TIME SAMPLES CHANGE POSSESSION INCLUDING COURIER DELIVERY

RELINQUISHED BY SAMPLER: 1. <u>[Signature]</u>	DATE/TIME: <u>1010</u> <u>10/24/25</u>	RECEIVED BY: 1. <u>[Signature]</u>	Conditions of bottles or coolers at receipt: ☐ COMPLIANT ☐ NON COMPLIANT ☐ COOLER TEMP <u>21.9°C</u>
RELINQUISHED BY SAMPLER: 2. <u>[Signature]</u>	DATE/TIME:	RECEIVED BY: 2. <u>[Signature]</u>	Comments:
RELINQUISHED BY SAMPLER: 3. <u>[Signature]</u>	DATE/TIME: <u>1155</u> <u>10/24/25</u>	RECEIVED BY: 3. <u>[Signature]</u>	Page <u>1</u> of <u>1</u>

CLIENT: ☐ Hand Delivered ☐ Other

Shipment Complete

☐ YES ☐ NO

Laboratory Certification

Certified By	License No.
Connecticut	PH-0830
DOD ELAP (ANAB)	L2219
Maine	2024021
Maryland	296
New Hampshire	255425
New Jersey	20012
New York	11376
Pennsylvania	68-00548
Soil Permit	525-24-234-08441
Texas	TX-C25-00189
Virginia	460312