

Matrix Spike Summary

Client:	Europastry	SDG No.:	Q3455
Project:	MCUA Permit No 14241 - 571 Jersey Ave NB NJ	Sample ID:	Q3448-03
Client ID:	DSN001MS	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Oil and Grease	mg/L	78-114	22.4		2.30	J	20.0	1	101		10/27/2025

Matrix Spike Summary

Client:	Europastry	SDG No.:	Q3455
Project:	MCUA Permit No 14241 - 571 Jersey Ave NB NJ	Sample ID:	Q3448-03
Client ID:	DSN001MSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Oil and Grease	mg/L	78-114	22.6		2.30	J	20.0	1	102		10/27/2025

Matrix Spike Summary

Client:	Europastry	SDG No.:	Q3455
Project:	MCUA Permit No 14241 - 571 Jersey Ave NB NJ	Sample ID:	Q3455-01
Client ID:	Q3455 -01MS	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Oil and Grease	mg/L	78-114	287		267		20.0	1	101		10/27/2025

Matrix Spike Summary

Client:	Europastry	SDG No.:	Q3455
Project:	MCUA Permit No 14241 - 571 Jersey Ave NB NJ	Sample ID:	Q3455-01
Client ID:	Q3455 -01MSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Oil and Grease	mg/L	78-114	287		267		20.0	1	102		10/27/2025