

Report of Analysis

Client:	PSEG	Date Collected:	10/24/25
Project:	PSEG Hoboken	Date Received:	10/24/25
Client Sample ID:	WC-3	SDG No.:	Q3460
Lab Sample ID:	Q3460-09	Matrix:	Solid
Analytical Method:	NJEPH	% Solid:	94.3
Sample Wt/Vol:	30.06 g	Test:	EPH_NF
Prep Method :		Final Vol:	2000 uL
		Prep Date :	10/27/25

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Datafile	Date Ana.	Prep BatchID
TARGETS										
Aliphatic C28-C40	Aliphatic C28-C40	8.42		1	1.25	2.12	mg/kg	FE056522.D	10/27/25 19:19	PB170258
Aliphatic C9-C28	Aliphatic C9-C28	4.24	U	1	0.96	4.24	mg/kg	FE056522.D	10/27/25 19:19	PB170258
Total AliphaticEPH	Total AliphaticEPH	8.42			2.21	6.36	mg/kg			
Total EPH	Total EPH	8.42			2.21	6.36	mg/kg			

* As samples are not fractionated, all aliphatic and aromatic carbon compounds in the C9-C40 carbon range are calculated against the aliphatic calibration curve, and reported as Aliphatic EPH. Therefore, the aliphatic C9-C40 concentration for the sample is reported as the Total EPH.

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

Report of Analysis

Client:	PSEG	Date Collected:	10/24/25
Project:	PSEG Hoboken	Date Received:	10/24/25
Client Sample ID:	WC-3	SDG No.:	Q3460
Lab Sample ID:	Q3460-09	Matrix:	Solid
Analytical Method:	NJEPH	% Solid:	94.3
Sample Wt/Vol:	30.06 g	Final Vol:	2000 uL
Prep Method :		Prep Date	10/27/25
		Test:	EPH_NF

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	Prep BatchID
TARGETS									
Aliphatic C9-C28	Aliphatic C9-C28	0.000	U	1	0.96	4.24	mg/kg	10/27/25	PB170258
Aliphatic C28-C40	Aliphatic C28-C40	8.42		1	1.25	2.12	mg/kg	10/27/25	PB170258
SURROGATES									
3383-33-2	1-chlorooctadecane (SURR)	25.7			40 - 140	51%	SPK: 50		
84-15-1	ortho-Terphenyl (SURR)	25.0			40 - 140	50%	SPK: 50		

Quantitation Report For Aliphatic EPH Range.

Lab Sample ID:	Q3460-09	Acq On:	27 Oct 2025 19:19
Client Sample ID:	WC-3	Operator:	YP\AJ
Data file:	FE056522.D	Misc:	
Instrument:	FID_E	ALS Vial:	22
Dilution Factor:	1	Sample Multiplier:	1.00

Compound	R.T.		Response	Conc	highest_standard	Units
Aliphatic C9-C12	3.245	6.871	449131	2.721	300	ug/ml
Aliphatic C12-C16	6.872	10.320	555278	3.256	200	ug/ml
Aliphatic C16-C21	10.321	13.697	439107	2.458	300	ug/ml
Aliphatic C21-C28	13.698	17.368	832276	4.647	400	ug/ml
Aliphatic C28-C40	17.369	22.326	16568118	119.364	600	ug/ml
Aliphatic EPH	3.245	22.326	18843910	132.446		ug/ml
ortho-Terphenyl (SURR)	11.987	11.987	5110492	25.02		ug/ml
1-chlorooctadecane (SURR)	13.426	13.426	3927933	25.74		ug/ml
Aliphatic C9-C28	3.245	17.368	2275792	13.082	1200	ug/ml