

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: MATR02

Lab Code: ACE

SDG NO.: Q3461

Continuing Calib Date: 10/28/2025

Initial Calibration Date(s): 10/09/2025

10/09/2025

Continuing Calib Time: 10:28

Initial Calibration Time(s): 14:14

22:29

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1248-1 (1)	5.49	5.49	5.39	5.59	0.00
Aroclor-1248-2 (2)	5.76	5.76	5.66	5.86	0.00
Aroclor-1248-3 (3)	5.96	5.96	5.86	6.06	0.00
Aroclor-1248-4 (4)	6.36	6.36	6.26	6.46	0.00
Aroclor-1248-5 (5)	6.39	6.40	6.30	6.50	0.01
Tetrachloro-m-xylene	4.35	4.36	4.26	4.46	0.01
Decachlorobiphenyl	9.93	9.93	9.83	10.03	0.00

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: MATR02

Lab Code: ACE

SDG NO.: Q3461

Continuing Calib Date: 10/28/2025

Initial Calibration Date(s): 10/09/2025 10/09/2025

Continuing Calib Time: 10:28

Initial Calibration Time(s): 14:14 22:29

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1248-1 (1)	4.73	4.74	4.64	4.84	0.01
Aroclor-1248-2 (2)	4.97	4.98	4.88	5.08	0.01
Aroclor-1248-3 (3)	5.01	5.02	4.92	5.12	0.01
Aroclor-1248-4 (4)	5.18	5.19	5.09	5.29	0.01
Aroclor-1248-5 (5)	5.57	5.58	5.48	5.68	0.01
Tetrachloro-m-xylene	3.65	3.66	3.56	3.76	0.01
Decachlorobiphenyl	8.65	8.67	8.57	8.77	0.02