

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: MATR02

Lab Code: ACE

SDG NO.: Q3461

Continuing Calib Date: 10/28/2025

Initial Calibration Date(s): 10/09/2025

10/09/2025

Continuing Calib Time: 11:05

Initial Calibration Time(s): 14:14

22:29

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1268-1 (1)	8.51	8.51	8.41	8.61	0.00
Aroclor-1268-2 (2)	8.60	8.60	8.50	8.70	0.00
Aroclor-1268-3 (3)	8.82	8.82	8.72	8.92	0.00
Aroclor-1268-4 (4)	9.22	9.22	9.12	9.32	0.00
Aroclor-1268-5 (5)	9.61	9.61	9.51	9.71	0.00
Tetrachloro-m-xylene	4.35	4.36	4.26	4.46	0.01
Decachlorobiphenyl	9.93	9.93	9.83	10.03	0.00

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: MATR02

Lab Code: ACE

SDG NO.: Q3461

Continuing Calib Date: 10/28/2025

Initial Calibration Date(s): 10/09/2025 10/09/2025

Continuing Calib Time: 11:05

Initial Calibration Time(s): 14:14 22:29

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1268-1 (1)	7.55	7.56	7.46	7.66	0.01
Aroclor-1268-2 (2)	7.61	7.63	7.53	7.73	0.02
Aroclor-1268-3 (3)	7.82	7.84	7.74	7.94	0.02
Aroclor-1268-4 (4)	8.10	8.12	8.02	8.22	0.02
Aroclor-1268-5 (5)	8.39	8.41	8.31	8.51	0.02
Tetrachloro-m-xylene	3.65	3.66	3.56	3.76	0.01
Decachlorobiphenyl	8.65	8.67	8.57	8.77	0.02