

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** MATR02
Lab Code: ACE **SDG NO.:** Q3461
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/09/2025 10/09/2025

Client Sample No.: CCAL02 **Date Analyzed:** 10/28/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO114709.D **Time Analyzed:** 15:28

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.489	5.392	5.592	556.400	500.000	11.3
Aroclor-1016-2	5.511	5.414	5.614	519.190	500.000	3.8
Aroclor-1016-3	5.573	5.476	5.676	520.180	500.000	4.0
Aroclor-1016-4	5.668	5.571	5.771	551.110	500.000	10.2
Aroclor-1016-5	5.960	5.863	6.063	557.680	500.000	11.5
Aroclor-1260-1	7.072	6.975	7.175	540.750	500.000	8.2
Aroclor-1260-2	7.326	7.229	7.429	531.910	500.000	6.4
Aroclor-1260-3	7.682	7.585	7.785	506.100	500.000	1.2
Aroclor-1260-4	7.905	7.807	8.007	514.930	500.000	3.0
Aroclor-1260-5	8.211	8.113	8.313	490.340	500.000	-1.9
Decachlorobiphenyl	9.932	9.834	10.034	49.160	50.000	-1.7
Tetrachloro-m-xylene	4.352	4.255	4.455	54.010	50.000	8.0

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** MATR02
Lab Code: ACE **SDG NO.:** Q3461
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/09/2025 10/09/2025

Client Sample No.: CCAL02 **Date Analyzed:** 10/28/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO114709.D **Time Analyzed:** 15:28

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.734	4.643	4.843	473.670	500.000	-5.3
Aroclor-1016-2	4.753	4.662	4.862	471.720	500.000	-5.7
Aroclor-1016-3	4.927	4.836	5.036	494.870	500.000	-1.0
Aroclor-1016-4	4.969	4.879	5.079	502.040	500.000	0.4
Aroclor-1016-5	5.181	5.092	5.292	501.730	500.000	0.3
Aroclor-1260-1	6.212	6.125	6.325	527.860	500.000	5.6
Aroclor-1260-2	6.399	6.313	6.513	540.510	500.000	8.1
Aroclor-1260-3	6.552	6.466	6.666	509.140	500.000	1.8
Aroclor-1260-4	7.023	6.939	7.139	577.700	500.000	15.5
Aroclor-1260-5	7.263	7.179	7.379	555.390	500.000	11.1
Decachlorobiphenyl	8.647	8.571	8.771	48.520	50.000	-3.0
Tetrachloro-m-xylene	3.654	3.560	3.760	46.400	50.000	-7.2