

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** MATR02
Lab Code: ACE **SDG NO.:** Q3461
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/21/2025 10/21/2025

Client Sample No.: CCAL04 **Date Analyzed:** 10/28/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP076142.D **Time Analyzed:** 19:44

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.791	5.695	5.895	509.290	500.000	1.9
Aroclor-1016-2	5.812	5.717	5.917	519.210	500.000	3.8
Aroclor-1016-3	5.875	5.780	5.980	512.070	500.000	2.4
Aroclor-1016-4	5.972	5.877	6.077	516.790	500.000	3.4
Aroclor-1016-5	6.264	6.169	6.369	506.220	500.000	1.2
Aroclor-1260-1	7.381	7.287	7.487	537.190	500.000	7.4
Aroclor-1260-2	7.634	7.540	7.740	504.740	500.000	0.9
Aroclor-1260-3	7.992	7.899	8.099	528.380	500.000	5.7
Aroclor-1260-4	8.220	8.127	8.327	563.830	500.000	12.8
Aroclor-1260-5	8.545	8.452	8.652	521.070	500.000	4.2
Decachlorobiphenyl	10.407	10.319	10.519	51.440	50.000	2.9
Tetrachloro-m-xylene	4.639	4.542	4.742	52.070	50.000	4.1

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** MATR02
Lab Code: ACE **SDG NO.:** Q3461
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/21/2025 10/21/2025

Client Sample No.: CCAL04 **Date Analyzed:** 10/28/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP076142.D **Time Analyzed:** 19:44

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.874	4.778	4.978	557.140	500.000	11.4
Aroclor-1016-2	4.930	4.836	5.036	551.490	500.000	10.3
Aroclor-1016-3	5.051	4.954	5.154	511.650	500.000	2.3
Aroclor-1016-4	5.092	4.996	5.196	561.810	500.000	12.4
Aroclor-1016-5	5.306	5.210	5.410	532.360	500.000	6.5
Aroclor-1260-1	6.332	6.238	6.438	551.380	500.000	10.3
Aroclor-1260-2	6.519	6.425	6.625	539.130	500.000	7.8
Aroclor-1260-3	6.673	6.578	6.778	527.040	500.000	5.4
Aroclor-1260-4	7.141	7.047	7.247	563.200	500.000	12.6
Aroclor-1260-5	7.382	7.287	7.487	515.060	500.000	3.0
Decachlorobiphenyl	8.774	8.682	8.882	45.340	50.000	-9.3
Tetrachloro-m-xylene	3.779	3.681	3.881	55.170	50.000	10.3