

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** MATR02
Lab Code: ACE **SDG NO.:** Q3461
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/21/2025 10/21/2025

Client Sample No.: CCAL05 **Date Analyzed:** 10/29/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP076148.D **Time Analyzed:** 09:48

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.794	5.695	5.895	527.180	500.000	5.4
Aroclor-1016-2	5.815	5.717	5.917	536.660	500.000	7.3
Aroclor-1016-3	5.878	5.780	5.980	527.600	500.000	5.5
Aroclor-1016-4	5.975	5.877	6.077	534.540	500.000	6.9
Aroclor-1016-5	6.267	6.169	6.369	523.280	500.000	4.7
Aroclor-1260-1	7.385	7.287	7.487	556.850	500.000	11.4
Aroclor-1260-2	7.638	7.540	7.740	522.670	500.000	4.5
Aroclor-1260-3	7.996	7.899	8.099	536.890	500.000	7.4
Aroclor-1260-4	8.223	8.127	8.327	567.900	500.000	13.6
Aroclor-1260-5	8.549	8.452	8.652	521.700	500.000	4.3
Decachlorobiphenyl	10.412	10.319	10.519	51.700	50.000	3.4
Tetrachloro-m-xylene	4.642	4.542	4.742	53.500	50.000	7.0

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** MATR02
Lab Code: ACE **SDG NO.:** Q3461
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/21/2025 10/21/2025

Client Sample No.: CCAL05 **Date Analyzed:** 10/29/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP076148.D **Time Analyzed:** 09:48

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.876	4.778	4.978	524.950	500.000	5.0
Aroclor-1016-2	4.933	4.836	5.036	540.960	500.000	8.2
Aroclor-1016-3	5.053	4.954	5.154	506.660	500.000	1.3
Aroclor-1016-4	5.093	4.996	5.196	592.930	500.000	18.6
Aroclor-1016-5	5.308	5.210	5.410	524.670	500.000	4.9
Aroclor-1260-1	6.335	6.238	6.438	550.050	500.000	10.0
Aroclor-1260-2	6.521	6.425	6.625	564.370	500.000	12.9
Aroclor-1260-3	6.675	6.578	6.778	569.060	500.000	13.8
Aroclor-1260-4	7.143	7.047	7.247	584.200	500.000	16.8
Aroclor-1260-5	7.384	7.287	7.487	541.930	500.000	8.4
Decachlorobiphenyl	8.777	8.682	8.882	44.830	50.000	-10.3
Tetrachloro-m-xylene	3.782	3.681	3.881	55.500	50.000	11.0