

Surrogate Summary

SDG No.: **Q3463**

Client: **Chemtech Consulting Group**

Analytical Method: **SW8260-Low**

Lab Sample ID	Client ID	Parameter	Spike	Result	Recovery (%)	Qual	Limits (%)	
							Low	High
Q3463-01	STORAGE-BLANK-SOIL-REF-6	1,2-Dichloroethane-d4	50	60.0	120		74	125
		Dibromofluoromethane	50	52.7	105		75	124
		Toluene-d8	50	49.4	99		86	113
		4-Bromofluorobenzene	50	52.9	106		77	121
Q3463-02	STORAGE-BLANK-WATER-REF-5	1,2-Dichloroethane-d4	50	58.4	117		74	125
		Dibromofluoromethane	50	51.5	103		75	124
		Toluene-d8	50	49.3	99		86	113
		4-Bromofluorobenzene	50	53.1	106		77	121
Q3463-03	STORAGE-BLANK-WATER-REF-4	1,2-Dichloroethane-d4	50	62.6	125		74	125
		Dibromofluoromethane	50	52.0	104		75	124
		Toluene-d8	50	49.2	98		86	113
		4-Bromofluorobenzene	50	55.0	110		77	121
Q3463-04	STORAGE-BLANK-SAMPLE-MANAC	1,2-Dichloroethane-d4	50	59.6	119		74	125
		Dibromofluoromethane	50	53.0	106		75	124
		Toluene-d8	50	49.8	100		86	113
		4-Bromofluorobenzene	50	51.9	104		77	121
VN1024WBL01	VN1024WBL01	1,2-Dichloroethane-d4	50	59.6	119		74	125
		Dibromofluoromethane	50	51.7	103		75	124
		Toluene-d8	50	48.9	98		86	113
		4-Bromofluorobenzene	50	52.2	104		77	121
VN1024WBS01	VN1024WBS01	1,2-Dichloroethane-d4	50	48.5	97		74	125
		Dibromofluoromethane	50	46.9	94		75	124
		Toluene-d8	50	46.5	93		86	113
		4-Bromofluorobenzene	50	47.5	95		77	121
VN1024WBSD0	VN1024WBSD01	1,2-Dichloroethane-d4	50	50.2	100		74	125
		Dibromofluoromethane	50	47.2	94		75	124
		Toluene-d8	50	46.1	92		86	113
		4-Bromofluorobenzene	50	46.7	93		77	121

