

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: CHEM02
 Lab Code: ACE SDG NO.: Q3463
 Lab File ID: VN088087.D Date Analyzed: 10/24/2025
 Instrument ID: MSVOA_N Time Analyzed: 12:28
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	337154	8.21	612856	9.08	535872	11.84
UPPER LIMIT	674308	8.706	1225710	9.582	1071740	12.341
LOWER LIMIT	168577	7.706	306428	8.582	267936	11.341
EPA SAMPLE NO.						
STORAGE-BLANK-SOIL-REF-6	213039	8.21	478121	9.08	452361	11.84
STORAGE-BLANK-WATER-REF-5	215750	8.21	480262	9.08	450816	11.84
STORAGE-BLANK-WATER-REF-4	240302	8.21	541591	9.08	517069	11.84
STORAGE-BLANK-SAMPLE-MANAGEMENT	212273	8.21	470823	9.08	440380	11.84
VN1024WBL01	216989	8.21	492274	9.08	454718	11.85
VN1024WBS01	356256	8.21	651642	9.08	567094	11.84
VN1024WBSD01	335260	8.21	621300	9.08	541423	11.85

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.