

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QC Batch ID # VN102325

Review By	John Carlone	Review On	10/24/2025 8:16:57 AM
Supervise By	Maresh Dadoda	Supervise On	10/27/2025 2:53:14 PM
SubDirectory	VN102325	HP Acquire Method	HP Processing Method 82N102325W.M

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP135909 VP135921,VP135922,VP135923,VP135924,VP135925,VP135926 VP135927

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VN088077.D	23 Oct 2025 09:09		JC\MD	Ok
2	VSTDICC001	VSTDICC001	VN088078.D	23 Oct 2025 09:42	COM.#74 Peak not detected in 1PPB	JC\MD	Ok,M
3	VSTDICC005	VSTDICC005	VN088079.D	23 Oct 2025 10:38		JC\MD	Ok,M
4	VSTDICC020	VSTDICC020	VN088080.D	23 Oct 2025 10:59		JC\MD	Ok,M
5	VSTDICCC050	VSTDICCC050	VN088081.D	23 Oct 2025 11:20	Method failed for Comp. #74	JC\MD	Ok,M
6	VSTDICC100	VSTDICC100	VN088082.D	23 Oct 2025 11:41		JC\MD	Ok,M
7	VSTDICC150	VSTDICC150	VN088083.D	23 Oct 2025 12:02		JC\MD	Ok,M
8	IBLK	IBLK	VN088084.D	23 Oct 2025 12:23		JC\MD	Ok
9	VSTDICV050	ICVVN102325	VN088085.D	23 Oct 2025 14:54		JC\MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QC Batch ID # VN102725

Review By	John Carlone	Review On	10/28/2025 8:12:39 AM
Supervise By	Semsettin Yesilyurt	Supervise On	10/28/2025 1:20:10 PM
SubDirectory	VN102725	HP Acquire Method	HP Processing Method 82N102325W.M

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP135937
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP135938,VP135939

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VN088101.D	27 Oct 2025 09:58		JC\MD	Ok
2	VSTDCCC050	VSTDCCC050	VN088102.D	27 Oct 2025 10:40	pH#Lot#V12668	JC\MD	Ok,M
3	VN1027MBL01	VN1027MBL01	VN088103.D	27 Oct 2025 11:14		JC\MD	Ok
4	VN1027WBL01	VN1027WBL01	VN088104.D	27 Oct 2025 11:34		JC\MD	Ok
5	VN1027WBS01	VN1027WBS01	VN088105.D	27 Oct 2025 11:55	BS Failed High For Com.#30,47	JC\MD	Ok,M
6	VN1027WBSD01	VN1027WBSD01	VN088106.D	27 Oct 2025 12:29	BSD Failed High For Comp.#47	JC\MD	Ok,M
7	Q3454-01	101625	VN088107.D	27 Oct 2025 12:50	vial A pH<2 foamy sample	JC\MD	Ok
8	Q3464-01	MW2	VN088108.D	27 Oct 2025 13:10	vial B pH<2	JC\MD	Ok
9	Q3464-02	MW3	VN088109.D	27 Oct 2025 13:31	vial B pH<2	JC\MD	Ok
10	PB170240TB	PB170240TB	VN088110.D	27 Oct 2025 13:52		JC\MD	Ok
11	Q3449-06	CF-620-COMP-41	VN088111.D	27 Oct 2025 14:13	vial A pH#5.0	JC\MD	Ok
12	Q3439-04	TP-9	VN088112.D	27 Oct 2025 14:34	vial A pH#5.0	JC\MD	Ok,M
13	Q3439-08	TP-10	VN088113.D	27 Oct 2025 14:54	vial A pH#5.0	JC\MD	Ok
14	Q3440-04	JB-2	VN088114.D	27 Oct 2025 15:15	vial A pH#5.0	JC\MD	Ok,M
15	Q3451-04	TP-11	VN088115.D	27 Oct 2025 15:36	vial A pH#5.0	JC\MD	Ok
16	Q3451-08	TP-12	VN088116.D	27 Oct 2025 15:56	vial A pH#5.0	JC\MD	Ok
17	Q3452-04	JB-1	VN088117.D	27 Oct 2025 16:17	vial A pH#5.0	JC\MD	Ok
18	Q3426-02	TWP3	VN088118.D	27 Oct 2025 16:37	vial A pH<2	JC\MD	Ok

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QC Batch ID # VN102725

Review By	John Carlone	Review On	10/28/2025 8:12:39 AM
Supervise By	Semsettin Yesilyurt	Supervise On	10/28/2025 1:20:10 PM
SubDirectory	VN102725	HP Acquire Method	HP Processing Method 82N102325W.M
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP135937		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP135938,VP135939		

19	Q3426-03	TWP5	VN088119.D	27 Oct 2025 16:58	vial A pH<2	JC\MD	Ok,M
20	Q3426-04	MW2	VN088120.D	27 Oct 2025 17:19	vial A pH<2	JC\MD	Ok
21	Q3426-01	TWP2	VN088121.D	27 Oct 2025 17:39	vial A pH<2 Surrogate Fail;Need 5X Dilution	JC\MD	Dilution
22	Q3468-01	MW-06	VN088122.D	27 Oct 2025 18:00	vial A pH<2 E flag of previous sample	JC\MD	ReRun
23	Q3468-03	MW-10	VN088123.D	27 Oct 2025 18:21	vial A pH<2 Confirm Concentration	JC\MD	Not Ok
24	Q3468-04	MW-11	VN088124.D	27 Oct 2025 18:41	vial A pH<2	JC\MD	Ok
25	VSTDCCC050	VSTDCCC050EC	VN088125.D	27 Oct 2025 19:22		JC\MD	Ok,M

M : Manual Integration