

Surrogate Summary

SDG No.: **Q3464**

Client: **G Environmental**

Analytical Method: **SW8260-Low**

Lab Sample ID	Client ID	Parameter	Spike	Result	Recovery (%)	Qual	Limits (%)	
							Low	High
Q3464-01	MW2	1,2-Dichloroethane-d4	50	62.7	125		70 (74)	130 (125)
		Dibromofluoromethane	50	51.8	104		70 (75)	130 (124)
		Toluene-d8	50	47.6	95		70 (86)	130 (113)
		4-Bromofluorobenzene	50	53.6	107		70 (77)	130 (121)
Q3464-02	MW3	1,2-Dichloroethane-d4	50	59.0	118		70 (74)	130 (125)
		Dibromofluoromethane	50	51.0	102		70 (75)	130 (124)
		Toluene-d8	50	47.5	95		70 (86)	130 (113)
		4-Bromofluorobenzene	50	50.3	101		70 (77)	130 (121)
VN1027WBL01	VN1027WBL01	1,2-Dichloroethane-d4	50	61.1	122		70 (74)	130 (125)
		Dibromofluoromethane	50	52.8	106		70 (75)	130 (124)
		Toluene-d8	50	49.9	100		70 (86)	130 (113)
		4-Bromofluorobenzene	50	52.0	104		70 (77)	130 (121)
VN1027WBS01	VN1027WBS01	1,2-Dichloroethane-d4	50	56.2	112		70 (74)	130 (125)
		Dibromofluoromethane	50	53.2	106		70 (75)	130 (124)
		Toluene-d8	50	51.1	102		70 (86)	130 (113)
		4-Bromofluorobenzene	50	52.6	105		70 (77)	130 (121)
VN1027WBSD0	VN1027WBSD01	1,2-Dichloroethane-d4	50	56.4	113		70 (74)	130 (125)
		Dibromofluoromethane	50	53.0	106		70 (75)	130 (124)
		Toluene-d8	50	51.0	102		70 (86)	130 (113)
		4-Bromofluorobenzene	50	52.9	106		70 (77)	130 (121)

