

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: GENV01
 Lab Code: ACE SDG NO.: Q3464
 Lab File ID: VN088102.D Date Analyzed: 10/27/2025
 Instrument ID: MSVOA_N Time Analyzed: 10:40
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	321965	8.21	587119	9.08	522069	11.84
UPPER LIMIT	643930	8.706	1174240	9.582	1044140	12.341
LOWER LIMIT	160983	7.706	293560	8.582	261035	11.341
EPA SAMPLE NO.						
MW2	242209	8.21	543451	9.08	522532	11.84
MW3	223573	8.21	506770	9.08	486491	11.84
VN1027WBL01	198801	8.21	446179	9.08	422679	11.84
VN1027WBS01	300941	8.21	563874	9.08	505093	11.85
VN1027WBSD01	293635	8.21	537984	9.08	482887	11.85

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.