

Initial and Continuing Calibration Verification

Client: HDR, Inc.

SDG No.: Q3466

Project: PVWC Linear Construction

RunNo.: LB137657

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV Hexavalent Chromium	mg/L	0.497	0.5	99	90-110	10/27/2025
Sample ID: CCV1 Hexavalent Chromium	mg/L	0.494	0.5	99	90-110	10/27/2025
Sample ID: CCV2 Hexavalent Chromium	mg/L	0.489	0.5	98	90-110	10/27/2025
Sample ID: CCV3 Hexavalent Chromium	mg/L	0.493	0.5	99	90-110	10/27/2025
Sample ID: CCV4 Hexavalent Chromium	mg/L	0.489	0.5	98	90-110	10/27/2025

Initial and Continuing Calibration Verification

Client: HDR, Inc.

SDG No.: Q3466

Project: PVWC Linear Construction

RunNo.: LB137663

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: Corrosivity	ICV	pH	7.01	7	100	90-110	10/27/2025
Sample ID: Corrosivity	CCV1	pH	2.01	2.00	101	90-110	10/27/2025
Sample ID: Corrosivity	CCV2	pH	12.02	12.00	100	90-110	10/27/2025
Sample ID: Corrosivity	CCV3	pH	2.02	2.00	101	90-110	10/27/2025

Initial and Continuing Calibration Verification

Client: HDR, Inc.

SDG No.: Q3466

Project: PVWC Linear Construction

RunNo.: LB137669

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1 Reactive Cyanide	mg/L	0.093	0.099	94	85-115	10/28/2025
Sample ID: CCV1 Reactive Cyanide	mg/L	0.24	0.25	96	90-110	10/28/2025
Sample ID: CCV2 Reactive Cyanide	mg/L	0.24	0.25	96	90-110	10/28/2025
Sample ID: CCV3 Reactive Cyanide	mg/L	0.24	0.25	96	90-110	10/28/2025
Sample ID: CCV4 Reactive Cyanide	mg/L	0.24	0.25	96	90-110	10/28/2025

Initial and Continuing Calibration Verification

Client: HDR, Inc.

SDG No.: Q3466

Project: PVWC Linear Construction

RunNo.: LB137670

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID:	ICV1						
Cyanide		mg/L	0.095	0.099	96	90-110	10/28/2025
Sample ID:	CCV1						
Cyanide		mg/L	0.25	0.25	100	90-110	10/28/2025
Sample ID:	CCV2						
Cyanide		mg/L	0.25	0.25	100	90-110	10/28/2025