

Initial and Continuing Calibration Blank Summary

Client: HDR, Inc.

SDG No.: Q3466

Project: PVWC Linear Construction

RunNo.: LB137657

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB Hexavalent Chromium	mg/L	< 0.0050	0.0050	U	0.0029	0.01	10/27/2025
Sample ID: CCB1 Hexavalent Chromium	mg/L	< 0.0050	0.0050	U	0.0029	0.01	10/27/2025
Sample ID: CCB2 Hexavalent Chromium	mg/L	< 0.0050	0.0050	U	0.0029	0.01	10/27/2025
Sample ID: CCB3 Hexavalent Chromium	mg/L	< 0.0050	0.0050	U	0.0029	0.01	10/27/2025
Sample ID: CCB4 Hexavalent Chromium	mg/L	< 0.0050	0.0050	U	0.0029	0.01	10/27/2025

Initial and Continuing Calibration Blank Summary

Client: HDR, Inc.

SDG No.: Q3466

Project: PVWC Linear Construction

RunNo.: LB137669

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB1 Reactive Cyanide	mg/L	0.0012	0.0025	J	0.00096	0.005	10/28/2025
Sample ID: CCB1 Reactive Cyanide	mg/L	0.0012	0.0025	J	0.00096	0.005	10/28/2025
Sample ID: CCB2 Reactive Cyanide	mg/L	0.0015	0.0025	J	0.00096	0.005	10/28/2025
Sample ID: CCB3 Reactive Cyanide	mg/L	< 0.0025	0.0025	U	0.00096	0.005	10/28/2025
Sample ID: CCB4 Reactive Cyanide	mg/L	0.0014	0.0025	J	0.00096	0.005	10/28/2025

Initial and Continuing Calibration Blank Summary

Client: HDR, Inc.

SDG No.: Q3466

Project: PVWC Linear Construction

RunNo.: LB137670

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB1 Cyanide	mg/L	0.0014	0.0025	J	0.00096	0.005	10/28/2025
Sample ID: CCB1 Cyanide	mg/L	0.0024	0.0025	J	0.00096	0.005	10/28/2025
Sample ID: CCB2 Cyanide	mg/L	0.0012	0.0025	J	0.00096	0.005	10/28/2025