

Instrument ID: CV1

Daily Analysis Runlog For Sequence/QC Batch ID # LB137668

Review By	mohan	Review On	10/30/2025 3:55:07 PM
Supervise By	jaswal	Supervise On	10/30/2025 3:55:31 PM
STD. NAME	STD REF.#		
ICAL Standard	MP87754,MP87755,MP87756,MP87757,MP87758,MP87759		
ICV Standard	MP87760		
CCV Standard	MP87762		
ICSA Standard			
CRI Standard	MP87764		
LCS Standard			
Chk Standard	MP87761,MP87763,MP87765,MP87751		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	10/28/25 12:32		mohan	OK
2	S0.2	S0.2	CAL2	10/28/25 12:35		mohan	OK
3	S2.5	S2.5	CAL3	10/28/25 12:37		mohan	OK
4	S5	S5	CAL4	10/28/25 12:39		mohan	OK
5	S7.5	S7.5	CAL5	10/28/25 12:42		mohan	OK
6	S10	S10	CAL6	10/28/25 12:45		mohan	OK
7	ICV90	ICV90	ICV	10/28/25 12:51		mohan	OK
8	ICB90	ICB90	ICB	10/28/25 12:53		mohan	OK
9	CCV97	CCV97	CCV	10/28/25 12:56		mohan	OK
10	CCB97	CCB97	CCB	10/28/25 12:58		mohan	OK
11	CRA	CRA	CRDL	10/28/25 13:04		mohan	OK
12	HighStd	HighStd	HIGH STD	10/28/25 13:06		mohan	OK
13	ChkStd	ChkStd	SAM	10/28/25 13:08		mohan	OK
14	PB170271BL	PB170271BL	MB	10/28/25 13:11		mohan	OK
15	PB170271BS	PB170271BS	LCS	10/28/25 13:14		mohan	OK
16	Q3451-01	TP-11	SAM	10/28/25 13:16		mohan	OK
17	Q3451-01DUP	TP-11DUP	DUP	10/28/25 13:21		mohan	OK
18	Q3451-01MS	TP-11MS	MS	10/28/25 13:24		mohan	OK

Instrument ID: CV1

Daily Analysis Runlog For Sequence/QC Batch ID # LB137668

Review By	mohan	Review On	10/30/2025 3:55:07 PM
Supervise By	jaswal	Supervise On	10/30/2025 3:55:31 PM
STD. NAME	STD REF.#		
ICAL Standard	MP87754,MP87755,MP87756,MP87757,MP87758,MP87759		
ICV Standard	MP87760		
CCV Standard	MP87762		
ICSA Standard			
CRI Standard	MP87764		
LCS Standard			
Chk Standard	MP87761,MP87763,MP87765,MP87751		

19	Q3451-01MSD	TP-11MSD	MSD	10/28/25 13:26		mohan	OK
20	Q3451-05	TP-12	SAM	10/28/25 13:29		mohan	OK
21	CCV98	CCV98	CCV	10/28/25 13:31		mohan	OK
22	CCB98	CCB98	CCB	10/28/25 13:33		mohan	OK
23	Q3452-01	JB-1	SAM	10/28/25 13:36	Hg is High , Need dilution	mohan	Dilution
24	Q3457-01	TR-05-102425	SAM	10/28/25 13:38		mohan	OK
25	Q3460-01	WC-1	SAM	10/28/25 13:41		mohan	OK
26	Q3460-05	WC-2	SAM	10/28/25 13:44		mohan	OK
27	Q3460-09	WC-3	SAM	10/28/25 13:46		mohan	OK
28	Q3460-13	WC-4	SAM	10/28/25 13:49		mohan	OK
29	Q3460-17	WC-5	SAM	10/28/25 13:51		mohan	OK
30	Q3465-01	SU-03-10242025	SAM	10/28/25 13:54		mohan	OK
31	Q3466-02	B8-0.0-1.0-20251023	SAM	10/28/25 13:56		mohan	OK
32	Q3466-05	B9-0.0-1.0-20251023	SAM	10/28/25 13:58		mohan	OK
33	CCV99	CCV99	CCV	10/28/25 14:01		mohan	OK
34	CCB99	CCB99	CCB	10/28/25 14:03		mohan	OK
35	Q3466-07	B7-1.0-2.0-20251023	SAM	10/28/25 14:06		mohan	OK
36	Q3469-01	TP-13	SAM	10/28/25 14:08		mohan	OK
37	Q3451-01L	TP-11L	SD	10/28/25 14:11		mohan	OK
38	Q3451-01A	TP-11A	PS	10/28/25 14:13		mohan	OK

Instrument ID: CV1

Daily Analysis Runlog For Sequence/QC Batch ID # LB137668

Review By	mohan	Review On	10/30/2025 3:55:07 PM
Supervise By	jaswal	Supervise On	10/30/2025 3:55:31 PM
STD. NAME	STD REF.#		
ICAL Standard	MP87754,MP87755,MP87756,MP87757,MP87758,MP87759		
ICV Standard	MP87760		
CCV Standard	MP87762		
ICSA Standard			
CRI Standard	MP87764		
LCS Standard			
Chk Standard	MP87761,MP87763,MP87765,MP87751		

39	Q3452-01DL	JB-1DL	SAM	10/28/25 14:20	10X For Hg	mohan	Confirms
40	CCV100	CCV100	CCV	10/28/25 14:23		mohan	OK
41	CCB100	CCB100	CCB	10/28/25 14:25		mohan	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB137697

Review By	Janvi	Review On	10/31/2025 10:49:23 AM
Supervise By	mohan	Supervise On	11/3/2025 10:12:07 AM
STD. NAME	STD REF.#		
ICAL Standard	MP87637,MP87643,MP87641,MP87640,MP87639,MP87638		
ICV Standard	MP87644,MP87643		
CCV Standard	MP87647		
ICSA Standard	MP87645,MP87646		
CRI Standard	MP87643		
LCS Standard			
Chk Standard	MP87668,MP87648,MP87651,MP87649,MP87650		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	10/29/25 13:36		Jaswal	OK
2	S1	S1	CAL2	10/29/25 13:40		Jaswal	OK
3	S2	S2	CAL3	10/29/25 13:44		Jaswal	OK
4	S3	S3	CAL4	10/29/25 13:49		Jaswal	OK
5	S4	S4	CAL5	10/29/25 13:53		Jaswal	OK
6	S5	S5	CAL6	10/29/25 13:57		Jaswal	OK
7	ICV01	ICV01	ICV	10/29/25 14:01		Jaswal	OK
8	LLICV01	LLICV01	LLICV	10/29/25 14:07		Jaswal	OK
9	ICB01	ICB01	ICB	10/29/25 14:12		Jaswal	OK
10	CRI01	CRI01	CRDL	10/29/25 14:16		Jaswal	OK
11	ICSA01	ICSA01	ICSA	10/29/25 14:21		Jaswal	OK
12	ICSAB01	ICSAB01	ICSAB	10/29/25 14:25		Jaswal	OK
13	ICSADL	ICSADL	ICSA	10/29/25 14:29		Jaswal	OK
14	ICSABDL	ICSABDL	ICSAB	10/29/25 14:33		Jaswal	OK
15	CCV01	CCV01	CCV	10/29/25 14:37		Jaswal	OK
16	CCB01	CCB01	CCB	10/29/25 14:41		Jaswal	OK
17	LR-1	LR-1	HIGH STD	10/29/25 14:46		Jaswal	OK
18	LR-2	LR-2	HIGH STD	10/29/25 14:50		Jaswal	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB137697

Review By	Janvi	Review On	10/31/2025 10:49:23 AM
Supervise By	mohan	Supervise On	11/3/2025 10:12:07 AM
STD. NAME	STD REF.#		
ICAL Standard	MP87637,MP87643,MP87641,MP87640,MP87639,MP87638		
ICV Standard	MP87644,MP87643		
CCV Standard	MP87647		
ICSA Standard	MP87645,MP87646		
CRI Standard	MP87643		
LCS Standard			
Chk Standard	MP87668,MP87648,MP87651,MP87649,MP87650		

19	LR-3	LR-3	HIGH STD	10/29/25 14:55		Jaswal	OK
20	Q3465-01	SU-03-10242025	SAM	10/29/25 15:00		Jaswal	OK
21	PB170288BL	PB170288BL	MB	10/29/25 15:04		Jaswal	OK
22	PB170288BS	PB170288BS	LCS	10/29/25 15:08		Jaswal	OK
23	Q3447-05DL2	P001-TW3-251023-01	SAM	10/29/25 15:12	10X for Cu,Pb	Jaswal	Confirms
24	Q3466-02	B8-0.0-1.0-20251023	SAM	10/29/25 15:16		Jaswal	OK
25	Q3466-05	B9-0.0-1.0-20251023	SAM	10/29/25 15:20		Jaswal	OK
26	Q3466-07	B7-1.0-2.0-20251023	SAM	10/29/25 15:24		Jaswal	OK
27	CCV02	CCV02	CCV	10/29/25 15:43		Jaswal	OK
28	CCB02	CCB02	CCB	10/29/25 15:48		Jaswal	OK
29	Q3469-01	TP-13	SAM	10/29/25 15:52		Jaswal	OK
30	Q3469-01DUP	TP-13DUP	DUP	10/29/25 15:56		Jaswal	OK
31	Q3469-01L	TP-13L	SD	10/29/25 16:00		Jaswal	OK
32	Q3469-01MS	TP-13MS	MS	10/29/25 16:04		Jaswal	OK
33	Q3469-01MSD	TP-13MSD	MSD	10/29/25 16:08		Jaswal	OK
34	Q3469-01A	TP-13A	PS	10/29/25 16:12	0.1ml of each m6008 and m6017 in 10ml sample.	Jaswal	OK
35	PB170264BL	PB170264BL	MB	10/29/25 16:16		Jaswal	OK
36	PB170264BS	PB170264BS	LCS	10/29/25 16:24		Jaswal	OK
37	Q3442-01	SW-1	SAM	10/29/25 16:29		Jaswal	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB137697

Review By	Janvi	Review On	10/31/2025 10:49:23 AM
Supervise By	mohan	Supervise On	11/3/2025 10:12:07 AM
STD. NAME	STD REF.#		
ICAL Standard	MP87637,MP87643,MP87641,MP87640,MP87639,MP87638		
ICV Standard	MP87644,MP87643		
CCV Standard	MP87647		
ICSA Standard	MP87645,MP87646		
CRI Standard	MP87643		
LCS Standard			
Chk Standard	MP87668,MP87648,MP87651,MP87649,MP87650		

38	Q3464-02	MW3	SAM	10/29/25 16:33		Jaswal	OK
39	CCV03	CCV03	CCV	10/29/25 16:37		Jaswal	OK
40	CCB03	CCB03	CCB	10/29/25 16:41		Jaswal	OK
41	Q3464-01	MW2	SAM	10/29/25 16:45		Jaswal	OK
42	Q3464-01DUP	MW2DUP	DUP	10/29/25 16:50		Jaswal	OK
43	Q3464-01L	MW2L	SD	10/29/25 16:54		Jaswal	OK
44	Q3464-01MS	MW2MS	MS	10/29/25 16:58		Jaswal	OK
45	Q3464-01MSD	MW2MSD	MSD	10/29/25 17:02		Jaswal	OK
46	Q3464-01A	MW2A	PS	10/29/25 17:06	0.1ml of each m6008 and m6017 in 10ml sample.	Jaswal	OK
47	Q3476-01	NB-07-102825	SAM	10/29/25 17:19		Jaswal	OK
48	Q3477-01	SOUTH-MAHWAH-SO	SAM	10/29/25 17:23		Jaswal	OK
49	CCV04	CCV04	CCV	10/29/25 17:27		Jaswal	OK
50	CCB04	CCB04	CCB	10/29/25 17:31		Jaswal	OK
51	Q3480-01	RT3449	SAM	10/29/25 17:35	Ag is oversaturated, Need dilution	Jaswal	Dilution
52	Q3481-01	OR-02-102825	SAM	10/29/25 17:39		Jaswal	OK
53	Q3481-01DUP	OR-02-102825DUP	DUP	10/29/25 17:43		Jaswal	OK
54	Q3481-01L	OR-02-102825L	SD	10/29/25 17:48		Jaswal	OK
55	Q3481-01MS	OR-02-102825MS	MS	10/29/25 17:52		Jaswal	OK
56	Q3481-01MSD	OR-02-102825MSD	MSD	10/29/25 17:56		Jaswal	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB137697

Review By	Janvi	Review On	10/31/2025 10:49:23 AM
Supervise By	mohan	Supervise On	11/3/2025 10:12:07 AM

STD. NAME	STD REF.#
ICAL Standard	MP87637,MP87643,MP87641,MP87640,MP87639,MP87638
ICV Standard	MP87644,MP87643
CCV Standard	MP87647
ICSA Standard	MP87645,MP87646
CRI Standard	MP87643
LCS Standard	
Chk Standard	MP87668,MP87648,MP87651,MP87649,MP87650

57	Q3481-01A	OR-02-102825A	PS	10/29/25 17:59	0.1ml of each m6008 and m6017 in 10ml sample.	Jaswal	OK
58	CCV05	CCV05	CCV	10/29/25 18:03		Jaswal	OK
59	CCB05	CCB05	CCB	10/29/25 18:07		Jaswal	OK
60	Q3420-05	COMPOSITE	SAM	10/29/25 18:12	Na oversaturated	Jaswal	Dilution
61	Q3420-05DUP	COMPOSITEDUP	DUP	10/29/25 18:16	Na High	Jaswal	Dilution
62	Q3420-05L	COMPOSITEL	SD	10/29/25 18:20	Na High in original/dup/ms/msd	Jaswal	Dilution
63	Q3420-05MS	COMPOSITEMS	MS	10/29/25 18:25	Na High	Jaswal	Dilution
64	Q3420-05MSD	COMPOSITEMSD	MSD	10/29/25 18:29	Na High	Jaswal	Dilution
65	Q3420-05A	COMPOSITEA	PS	10/29/25 18:33	Na High, 0.1ml of each m6008 and m6017 in 10ml sample.	Jaswal	Dilution
66	CCV06	CCV06	CCV	10/29/25 18:55		Jaswal	OK
67	CCB06	CCB06	CCB	10/29/25 19:04		Jaswal	OK