



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: HDR, Inc.

SDG No.: Q3466

Contract: HDRI08

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB90	Mercury	0.076	+/-0.2	U	0.20	CV	10/28/2025	12:53	LB137668

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: HDR, Inc.

SDG No.: Q3466

Contract: HDRI08

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB97	Mercury	0.076	+/-0.2	U	0.20	CV	10/28/2025	12:58	LB137668
CCB98	Mercury	0.076	+/-0.2	U	0.20	CV	10/28/2025	13:33	LB137668
CCB99	Mercury	0.076	+/-0.2	U	0.20	CV	10/28/2025	14:03	LB137668
CCB100	Mercury	0.076	+/-0.2	U	0.20	CV	10/28/2025	14:25	LB137668

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: HDR, Inc.

SDG No.: Q3466

Contract: HDRI08

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Aluminum	11.3	+/-50	U	100	P	10/29/2025	14:12	LB137697
	Antimony	6.76	+/-25	U	50.0	P	10/29/2025	14:12	LB137697
	Arsenic	5.12	+/-10	U	20.0	P	10/29/2025	14:12	LB137697
	Barium	14.6	+/-50	U	100	P	10/29/2025	14:12	LB137697
	Beryllium	0.56	+/-3	U	6.00	P	10/29/2025	14:12	LB137697
	Cadmium	0.50	+/-3	U	6.00	P	10/29/2025	14:12	LB137697
	Calcium	234	+/-1000	U	2000	P	10/29/2025	14:12	LB137697
	Chromium	2.12	+/-5	U	10.0	P	10/29/2025	14:12	LB137697
	Cobalt	2.26	+/-15	U	30.0	P	10/29/2025	14:12	LB137697
	Copper	4.60	+/-10	U	20.0	P	10/29/2025	14:12	LB137697
	Iron	23.4	+/-50	U	100	P	10/29/2025	14:12	LB137697
	Lead	2.30	+/-6	U	12.0	P	10/29/2025	14:12	LB137697
	Magnesium	244	+/-1000	U	2000	P	10/29/2025	14:12	LB137697
	Manganese	5.94	+/-10	U	20.0	P	10/29/2025	14:12	LB137697
	Nickel	3.06	+/-20	U	40.0	P	10/29/2025	14:12	LB137697
	Potassium	918	+/-1000	U	2000	P	10/29/2025	14:12	LB137697
	Selenium	9.64	+/-10	U	20.0	P	10/29/2025	14:12	LB137697
	Silver	1.62	+/-5	U	10.0	P	10/29/2025	14:12	LB137697
	Sodium	868	+/-1000	U	2000	P	10/29/2025	14:12	LB137697
	Thallium	4.38	+/-20	U	40.0	P	10/29/2025	14:12	LB137697
	Vanadium	6.26	+/-20	U	40.0	P	10/29/2025	14:12	LB137697
	Zinc	16.7	+/-20	U	40.0	P	10/29/2025	14:12	LB137697

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: HDR, Inc.

SDG No.: Q3466

Contract: HDRI08

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Aluminum	11.3	+/-50	U	100	P	10/29/2025	14:41	LB137697
	Antimony	6.76	+/-25	U	50.0	P	10/29/2025	14:41	LB137697
	Arsenic	5.12	+/-10	U	20.0	P	10/29/2025	14:41	LB137697
	Barium	14.6	+/-50	U	100	P	10/29/2025	14:41	LB137697
	Beryllium	0.56	+/-3	U	6.00	P	10/29/2025	14:41	LB137697
	Cadmium	0.50	+/-3	U	6.00	P	10/29/2025	14:41	LB137697
	Calcium	234	+/-1000	U	2000	P	10/29/2025	14:41	LB137697
	Chromium	2.12	+/-5	U	10.0	P	10/29/2025	14:41	LB137697
	Cobalt	2.26	+/-15	U	30.0	P	10/29/2025	14:41	LB137697
	Copper	4.60	+/-10	U	20.0	P	10/29/2025	14:41	LB137697
	Iron	23.4	+/-50	U	100	P	10/29/2025	14:41	LB137697
	Lead	2.30	+/-6	U	12.0	P	10/29/2025	14:41	LB137697
	Magnesium	244	+/-1000	U	2000	P	10/29/2025	14:41	LB137697
	Manganese	5.94	+/-10	U	20.0	P	10/29/2025	14:41	LB137697
	Nickel	3.06	+/-20	U	40.0	P	10/29/2025	14:41	LB137697
	Potassium	918	+/-1000	U	2000	P	10/29/2025	14:41	LB137697
	Selenium	9.64	+/-10	U	20.0	P	10/29/2025	14:41	LB137697
	Silver	1.62	+/-5	U	10.0	P	10/29/2025	14:41	LB137697
	Sodium	868	+/-1000	U	2000	P	10/29/2025	14:41	LB137697
	Thallium	4.38	+/-20	U	40.0	P	10/29/2025	14:41	LB137697
	Vanadium	6.26	+/-20	U	40.0	P	10/29/2025	14:41	LB137697
	Zinc	16.7	+/-20	U	40.0	P	10/29/2025	14:41	LB137697
CCB02	Aluminum	11.3	+/-50	U	100	P	10/29/2025	15:48	LB137697
	Antimony	6.76	+/-25	U	50.0	P	10/29/2025	15:48	LB137697
	Arsenic	5.12	+/-10	U	20.0	P	10/29/2025	15:48	LB137697
	Barium	14.6	+/-50	U	100	P	10/29/2025	15:48	LB137697
	Beryllium	0.56	+/-3	U	6.00	P	10/29/2025	15:48	LB137697
	Cadmium	0.50	+/-3	U	6.00	P	10/29/2025	15:48	LB137697
	Calcium	234	+/-1000	U	2000	P	10/29/2025	15:48	LB137697
	Chromium	2.12	+/-5	U	10.0	P	10/29/2025	15:48	LB137697
	Cobalt	2.26	+/-15	U	30.0	P	10/29/2025	15:48	LB137697
	Copper	4.60	+/-10	U	20.0	P	10/29/2025	15:48	LB137697
	Iron	23.4	+/-50	U	100	P	10/29/2025	15:48	LB137697
	Lead	2.30	+/-6	U	12.0	P	10/29/2025	15:48	LB137697
	Magnesium	244	+/-1000	U	2000	P	10/29/2025	15:48	LB137697
	Manganese	5.94	+/-10	U	20.0	P	10/29/2025	15:48	LB137697
	Nickel	3.06	+/-20	U	40.0	P	10/29/2025	15:48	LB137697
	Potassium	918	+/-1000	U	2000	P	10/29/2025	15:48	LB137697
	Selenium	9.64	+/-10	U	20.0	P	10/29/2025	15:48	LB137697

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: HDR, Inc.

SDG No.: Q3466

Contract: HDRI08

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB02	Silver	1.62	+/-5	U	10.0	P	10/29/2025	15:48	LB137697
	Sodium	868	+/-1000	U	2000	P	10/29/2025	15:48	LB137697
	Thallium	4.38	+/-20	U	40.0	P	10/29/2025	15:48	LB137697
	Vanadium	6.26	+/-20	U	40.0	P	10/29/2025	15:48	LB137697
	Zinc	16.7	+/-20	U	40.0	P	10/29/2025	15:48	LB137697
CCB03	Aluminum	11.3	+/-50	U	100	P	10/29/2025	16:41	LB137697
	Antimony	6.76	+/-25	U	50.0	P	10/29/2025	16:41	LB137697
	Arsenic	5.12	+/-10	U	20.0	P	10/29/2025	16:41	LB137697
	Barium	14.6	+/-50	U	100	P	10/29/2025	16:41	LB137697
	Beryllium	0.56	+/-3	U	6.00	P	10/29/2025	16:41	LB137697
	Cadmium	0.50	+/-3	U	6.00	P	10/29/2025	16:41	LB137697
	Calcium	234	+/-1000	U	2000	P	10/29/2025	16:41	LB137697
	Chromium	2.12	+/-5	U	10.0	P	10/29/2025	16:41	LB137697
	Cobalt	2.26	+/-15	U	30.0	P	10/29/2025	16:41	LB137697
	Copper	4.60	+/-10	U	20.0	P	10/29/2025	16:41	LB137697
	Iron	29.0	+/-50	J	100	P	10/29/2025	16:41	LB137697
	Lead	2.30	+/-6	U	12.0	P	10/29/2025	16:41	LB137697
	Magnesium	244	+/-1000	U	2000	P	10/29/2025	16:41	LB137697
	Manganese	5.94	+/-10	U	20.0	P	10/29/2025	16:41	LB137697
	Nickel	3.06	+/-20	U	40.0	P	10/29/2025	16:41	LB137697
	Potassium	918	+/-1000	U	2000	P	10/29/2025	16:41	LB137697
	Selenium	9.64	+/-10	U	20.0	P	10/29/2025	16:41	LB137697
	Silver	1.62	+/-5	U	10.0	P	10/29/2025	16:41	LB137697
	Sodium	868	+/-1000	U	2000	P	10/29/2025	16:41	LB137697
	Thallium	4.38	+/-20	U	40.0	P	10/29/2025	16:41	LB137697
	Vanadium	6.26	+/-20	U	40.0	P	10/29/2025	16:41	LB137697
	Zinc	16.7	+/-20	U	40.0	P	10/29/2025	16:41	LB137697
CCB04	Aluminum	11.3	+/-50	U	100	P	10/29/2025	17:31	LB137697
	Antimony	6.76	+/-25	U	50.0	P	10/29/2025	17:31	LB137697
	Arsenic	5.12	+/-10	U	20.0	P	10/29/2025	17:31	LB137697
	Barium	14.6	+/-50	U	100	P	10/29/2025	17:31	LB137697
	Beryllium	0.56	+/-3	U	6.00	P	10/29/2025	17:31	LB137697
	Cadmium	0.50	+/-3	U	6.00	P	10/29/2025	17:31	LB137697
	Calcium	234	+/-1000	U	2000	P	10/29/2025	17:31	LB137697
	Chromium	2.12	+/-5	U	10.0	P	10/29/2025	17:31	LB137697
	Cobalt	2.26	+/-15	U	30.0	P	10/29/2025	17:31	LB137697
	Copper	4.60	+/-10	U	20.0	P	10/29/2025	17:31	LB137697
	Iron	23.4	+/-50	U	100	P	10/29/2025	17:31	LB137697
	Lead	2.30	+/-6	U	12.0	P	10/29/2025	17:31	LB137697

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: HDR, Inc.

SDG No.: Q3466

Contract: HDRI08

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB04	Magnesium	244	+/-1000	U	2000	P	10/29/2025	17:31	LB137697
	Manganese	5.94	+/-10	U	20.0	P	10/29/2025	17:31	LB137697
	Nickel	3.06	+/-20	U	40.0	P	10/29/2025	17:31	LB137697
	Potassium	918	+/-1000	U	2000	P	10/29/2025	17:31	LB137697
	Selenium	9.64	+/-10	U	20.0	P	10/29/2025	17:31	LB137697
	Silver	1.62	+/-5	U	10.0	P	10/29/2025	17:31	LB137697
	Sodium	868	+/-1000	U	2000	P	10/29/2025	17:31	LB137697
	Thallium	4.38	+/-20	U	40.0	P	10/29/2025	17:31	LB137697
	Vanadium	6.26	+/-20	U	40.0	P	10/29/2025	17:31	LB137697
	Zinc	16.7	+/-20	U	40.0	P	10/29/2025	17:31	LB137697
CCB05	Aluminum	12.2	+/-50	J	100	P	10/29/2025	18:07	LB137697
	Antimony	6.76	+/-25	U	50.0	P	10/29/2025	18:07	LB137697
	Arsenic	5.12	+/-10	U	20.0	P	10/29/2025	18:07	LB137697
	Barium	14.6	+/-50	U	100	P	10/29/2025	18:07	LB137697
	Beryllium	0.56	+/-3	U	6.00	P	10/29/2025	18:07	LB137697
	Cadmium	0.50	+/-3	U	6.00	P	10/29/2025	18:07	LB137697
	Calcium	234	+/-1000	U	2000	P	10/29/2025	18:07	LB137697
	Chromium	2.12	+/-5	U	10.0	P	10/29/2025	18:07	LB137697
	Cobalt	2.26	+/-15	U	30.0	P	10/29/2025	18:07	LB137697
	Copper	4.60	+/-10	U	20.0	P	10/29/2025	18:07	LB137697
	Iron	23.4	+/-50	U	100	P	10/29/2025	18:07	LB137697
	Lead	2.30	+/-6	U	12.0	P	10/29/2025	18:07	LB137697
	Magnesium	244	+/-1000	U	2000	P	10/29/2025	18:07	LB137697
	Manganese	5.94	+/-10	U	20.0	P	10/29/2025	18:07	LB137697
	Nickel	3.06	+/-20	U	40.0	P	10/29/2025	18:07	LB137697
	Potassium	918	+/-1000	U	2000	P	10/29/2025	18:07	LB137697
	Selenium	9.64	+/-10	U	20.0	P	10/29/2025	18:07	LB137697
	Silver	1.62	+/-5	U	10.0	P	10/29/2025	18:07	LB137697
	Sodium	868	+/-1000	U	2000	P	10/29/2025	18:07	LB137697
	Thallium	4.38	+/-20	U	40.0	P	10/29/2025	18:07	LB137697
	Vanadium	6.26	+/-20	U	40.0	P	10/29/2025	18:07	LB137697
	Zinc	16.7	+/-20	U	40.0	P	10/29/2025	18:07	LB137697
CCB06	Aluminum	11.3	+/-50	U	100	P	10/29/2025	19:04	LB137697
	Antimony	6.76	+/-25	U	50.0	P	10/29/2025	19:04	LB137697
	Arsenic	5.12	+/-10	U	20.0	P	10/29/2025	19:04	LB137697
	Barium	14.6	+/-50	U	100	P	10/29/2025	19:04	LB137697
	Beryllium	0.56	+/-3	U	6.00	P	10/29/2025	19:04	LB137697
	Cadmium	0.50	+/-3	U	6.00	P	10/29/2025	19:04	LB137697
	Calcium	234	+/-1000	U	2000	P	10/29/2025	19:04	LB137697

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: HDR, Inc.

SDG No.: Q3466

Contract: HDRI08

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB06	Chromium	2.12	+/-5	U	10.0	P	10/29/2025	19:04	LB137697
	Cobalt	2.26	+/-15	U	30.0	P	10/29/2025	19:04	LB137697
	Copper	4.60	+/-10	U	20.0	P	10/29/2025	19:04	LB137697
	Iron	23.4	+/-50	U	100	P	10/29/2025	19:04	LB137697
	Lead	2.30	+/-6	U	12.0	P	10/29/2025	19:04	LB137697
	Magnesium	244	+/-1000	U	2000	P	10/29/2025	19:04	LB137697
	Manganese	5.94	+/-10	U	20.0	P	10/29/2025	19:04	LB137697
	Nickel	3.06	+/-20	U	40.0	P	10/29/2025	19:04	LB137697
	Potassium	918	+/-1000	U	2000	P	10/29/2025	19:04	LB137697
	Selenium	9.64	+/-10	U	20.0	P	10/29/2025	19:04	LB137697
	Silver	1.62	+/-5	U	10.0	P	10/29/2025	19:04	LB137697
	Sodium	868	+/-1000	U	2000	P	10/29/2025	19:04	LB137697
	Thallium	4.38	+/-20	U	40.0	P	10/29/2025	19:04	LB137697
	Vanadium	6.26	+/-20	U	40.0	P	10/29/2025	19:04	LB137697
	Zinc	16.7	+/-20	U	40.0	P	10/29/2025	19:04	LB137697