

LB137673


7470A

INSTRUMENT ID: CV1

Linear

μ Abs.:

4638E



A= 0.0000e+000

B= 2.2105e-004

C= -5.5752e-002

Rho= 0.9986353

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.019	0.019	337	0.000	337	0			
0.2	0.200	0.210	0.010	1204	0.0 %	1204	0			
2.5	2.500	2.527	0.027	11682	0.0 %	11682	0			
5.0	5.000	5.147	0.147	23538	0.0 %	23538				
7.5	7.500	7.099	-0.401	32367	0.0 %	32367				
10.0	10.000	10.198	0.198	46386	0.0 %	46386				

LB137673 INSTRUMENT ID:CV1

Method: 7470A

Operator: Admin

Date of Analysis: 28 Oct 2025 14:55:36

Type	Sample ID	Extended ID	Conc.	µ Abs.	Units	Type	Std Conc	Method	Date
S	0.0 - 1	50	-	337	PPB	Std	0.0000	7470A	28 Oct 2025 14:55:54
S	0.2 - 1	502	-	1204	PPB	Std	0.2000	7470A	28 Oct 2025 14:58:10
S	2.5 - 1	52.5	-	11682	PPB	Std	2.5000	7470A	28 Oct 2025 15:00:28
S	5.0 - 1	50	-	23538	PPB	Std	5.0000	7470A	28 Oct 2025 15:02:45
S	7.5 - 1	52.5	-	32367	PPB	Std	7.5000	7470A	28 Oct 2025 15:05:04
S	10.0 - 1	506	-	46386	PPB	Std	10.0000	7470A	28 Oct 2025 15:10:12
U	ICV91 - 1	ICV91	3.6502	16765	PPB	SMPL	-7470A		28 Oct 2025 15:15:30
U	ICB91 - 1	ICB91	0.0068	283	PPB	SMPL	-7470A		28 Oct 2025 15:17:46
U	CCV01 - 1	CCV01	5.1447	23526	PPB	SMPL	-7470A		28 Oct 2025 15:20:04
U	CCB01 - 1	CCB01	-0.0102	206	PPB	SMPL	-7470A		28 Oct 2025 15:22:19
U	CRA - 1	CRA	0.2111	1207	PPB	SMPL	-7470A		28 Oct 2025 15:24:39
U	HighStd - 1	HighStd	10.1546	46190	PPB	SMPL	-7470A		28 Oct 2025 15:26:55
U	ChkStd - 1	ChkStd	6.4387	29380	PPB	SMPL	-7470A		28 Oct 2025 15:29:11
U	PB170292BL - 1	PBW	-0.0734	-80	PPB	SMPL	-7470A		28 Oct 2025 15:31:38
U	PB170292BS - 1	LCSW	3.5116	16138	PPB	SMPL	-7470A		28 Oct 2025 15:34:00
U	Q3460-04 - 1	WC-1	-0.0447	50	PPB	SMPL	-7470A		28 Oct 2025 15:36:17
U	Q3460-04DUP - 1	WC-1DUP	0.0090	293	PPB	SMPL	-7470A		28 Oct 2025 15:38:36
U	Q3460-04MS - 1	WC-1MS	3.6104	16585	PPB	SMPL	-7470A		28 Oct 2025 15:40:53
U	Q3460-04MSD - 1	WC-1MSD	3.4470	15846	PPB	SMPL	-7470A		28 Oct 2025 15:43:10
U	Q3460-08 - 1	WC-2	-0.0387	77	PPB	SMPL	-7470A		28 Oct 2025 15:45:28
U	CCV02 - 1	CCV02	5.1730	23654	PPB	SMPL	-7470A		28 Oct 2025 15:47:46
U	CCB02 - 1	CCB02	-0.0000	252	PPB	SMPL	-7470A		28 Oct 2025 15:50:01
U	Q3460-12 - 1	WC-3	0.0035	268	PPB	SMPL	-7470A		28 Oct 2025 15:52:24
U	Q3460-16 - 1	WC-4	0.0119	306	PPB	SMPL	-7470A		28 Oct 2025 15:57:04
U	Q3460-20 - 1	WC-5	0.0263	371	PPB	SMPL	-7470A		28 Oct 2025 16:00:54
U	Q3466-03 - 1	B8-0.0-1.0-20251023	0.0139	315	PPB	SMPL	-7470A		28 Oct 2025 16:03:10
U	Q3469-04 - 1	TP-13	0.0143	317	PPB	SMPL	-7470A		28 Oct 2025 16:05:26
U	Q3472-05 - 1	COMP-1	0.0035	268	PPB	SMPL	-7470A		28 Oct 2025 16:07:43
U	Q3472-07 - 1	COMP-2	0.0095	295	PPB	SMPL	-7470A		28 Oct 2025 16:09:59
U	Q3472-09 - 1	COMP-3	-0.0058	226	PPB	SMPL	-7470A		28 Oct 2025 16:12:16
U	Q3472-11 - 1	COMP-4	0.0017	260	PPB	SMPL	-7470A		28 Oct 2025 16:14:33
U	Q3472-13 - 1	COMP-5	0.0097	296	PPB	SMPL	-7470A		28 Oct 2025 16:16:49
U	CCV03 - 1	CCV03	5.0211	22967	PPB	SMPL	-7470A		28 Oct 2025 16:19:05
U	CCB03 - 1	CCB03	-0.0294	119	PPB	SMPL	-7470A		28 Oct 2025 16:21:20
U	Q3472-15 - 1	155	0.0685	562	PPB	SMPL	-7470A		28 Oct 2025 16:23:40
U	Q3472-17 - 1	142	0.1260	822	PPB	SMPL	-7470A		28 Oct 2025 16:25:56
U	PB170261TB - 1	PB170261TB	-0.0018	244	PPB	SMPL	-7470A		28 Oct 2025 16:28:11
U	Q3460-04LX5 - 1		-0.0518	18	PPB	SMPL	-7470A		28 Oct 2025 16:30:27
U	Q3460-04A - 1		3.4934	16056	PPB	SMPL	-7470A		28 Oct 2025 16:32:42
U	CCV04 - 1	CCV04	5.0580	23134	PPB	SMPL	-7470A		28 Oct 2025 16:34:59
U	CCB04 - 1	CCB04	-0.0874	-143	PPB	SMPL	-7470A		28 Oct 2025 16:37:18