

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: HDR, Inc.

SDG No.: Q3466

Contract: HDRI08

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV91	Mercury	3.65	4.0	91	90 - 110	CV	10/28/2025	15:15	LB137673

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: HDR, Inc.

SDG No.: Q3466

Contract: HDRI08

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Mercury	5.14	5.0	103	90 - 110	CV	10/28/2025	15:20	LB137673
CCV02	Mercury	5.17	5.0	104	90 - 110	CV	10/28/2025	15:47	LB137673
CCV03	Mercury	5.02	5.0	100	90 - 110	CV	10/28/2025	16:19	LB137673
CCV04	Mercury	5.06	5.0	101	90 - 110	CV	10/28/2025	16:34	LB137673

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: HDR, Inc.

SDG No.: Q3466

Contract: HDRI08

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Arsenic	3820	4000	95	90 - 110	P	10/28/2025	12:30	LB137677
	Barium	7600	8000	95	90 - 110	P	10/28/2025	12:30	LB137677
	Cadmium	1900	2000	95	90 - 110	P	10/28/2025	12:30	LB137677
	Chromium	783	800	98	90 - 110	P	10/28/2025	12:30	LB137677
	Lead	3730	4000	93	90 - 110	P	10/28/2025	12:30	LB137677
	Selenium	3820	4000	95	90 - 110	P	10/28/2025	12:30	LB137677
	Silver	963	1000	96	90 - 110	P	10/28/2025	12:30	LB137677

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: HDR, Inc.

SDG No.: Q3466

Contract: HDRI08

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Arsenic	20.3	20.0	102	80 - 120	P	10/28/2025	12:34	LB137677
	Barium	93.0	100	93	80 - 120	P	10/28/2025	12:34	LB137677
	Cadmium	5.86	6.0	98	80 - 120	P	10/28/2025	12:34	LB137677
	Chromium	10.6	10.0	106	80 - 120	P	10/28/2025	12:34	LB137677
	Lead	12.0	12.0	100	80 - 120	P	10/28/2025	12:34	LB137677
	Selenium	20.7	20.0	104	80 - 120	P	10/28/2025	12:34	LB137677
	Silver	12.0	10.0	120	80 - 120	P	10/28/2025	12:34	LB137677

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: HDR, Inc.

SDG No.: Q3466

Contract: HDRI08

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Arsenic	4950	5000	99	90 - 110	P	10/28/2025	13:03	LB137677
	Barium	10000	10000	100	90 - 110	P	10/28/2025	13:03	LB137677
	Cadmium	2480	2500	99	90 - 110	P	10/28/2025	13:03	LB137677
	Chromium	989	1000	99	90 - 110	P	10/28/2025	13:03	LB137677
	Lead	4910	5000	98	90 - 110	P	10/28/2025	13:03	LB137677
	Selenium	4960	5000	99	90 - 110	P	10/28/2025	13:03	LB137677
	Silver	1240	1250	99	90 - 110	P	10/28/2025	13:03	LB137677
CCV02	Arsenic	5220	5000	104	90 - 110	P	10/28/2025	14:01	LB137677
	Barium	10200	10000	102	90 - 110	P	10/28/2025	14:01	LB137677
	Cadmium	2610	2500	104	90 - 110	P	10/28/2025	14:01	LB137677
	Chromium	1060	1000	106	90 - 110	P	10/28/2025	14:01	LB137677
	Lead	5150	5000	103	90 - 110	P	10/28/2025	14:01	LB137677
	Selenium	5230	5000	105	90 - 110	P	10/28/2025	14:01	LB137677
	Silver	1290	1250	104	90 - 110	P	10/28/2025	14:01	LB137677
CCV03	Arsenic	4950	5000	99	90 - 110	P	10/28/2025	14:53	LB137677
	Barium	10500	10000	105	90 - 110	P	10/28/2025	14:53	LB137677
	Cadmium	2510	2500	101	90 - 110	P	10/28/2025	14:53	LB137677
	Chromium	1040	1000	104	90 - 110	P	10/28/2025	14:53	LB137677
	Lead	4950	5000	99	90 - 110	P	10/28/2025	14:53	LB137677
	Selenium	4940	5000	99	90 - 110	P	10/28/2025	14:53	LB137677
	Silver	1280	1250	102	90 - 110	P	10/28/2025	14:53	LB137677
CCV04	Arsenic	4930	5000	98	90 - 110	P	10/28/2025	15:45	LB137677
	Barium	9920	10000	99	90 - 110	P	10/28/2025	15:45	LB137677
	Cadmium	2500	2500	100	90 - 110	P	10/28/2025	15:45	LB137677
	Chromium	1000	1000	100	90 - 110	P	10/28/2025	15:45	LB137677
	Lead	4920	5000	98	90 - 110	P	10/28/2025	15:45	LB137677
	Selenium	4930	5000	98	90 - 110	P	10/28/2025	15:45	LB137677
	Silver	1240	1250	99	90 - 110	P	10/28/2025	15:45	LB137677
CCV05	Arsenic	5100	5000	102	90 - 110	P	10/28/2025	16:35	LB137677
	Barium	10300	10000	103	90 - 110	P	10/28/2025	16:35	LB137677
	Cadmium	2560	2500	102	90 - 110	P	10/28/2025	16:35	LB137677
	Chromium	1030	1000	103	90 - 110	P	10/28/2025	16:35	LB137677
	Lead	5060	5000	101	90 - 110	P	10/28/2025	16:35	LB137677
	Selenium	5090	5000	102	90 - 110	P	10/28/2025	16:35	LB137677

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: HDR, Inc.

SDG No.: Q3466

Contract: HDRI08

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV05	Silver	1280	1250	102	90 - 110	P	10/28/2025	16:35	LB137677
CCV06	Arsenic	4960	5000	99	90 - 110	P	10/28/2025	17:27	LB137677
	Barium	10000	10000	100	90 - 110	P	10/28/2025	17:27	LB137677
	Cadmium	2520	2500	101	90 - 110	P	10/28/2025	17:27	LB137677
	Chromium	1010	1000	101	90 - 110	P	10/28/2025	17:27	LB137677
	Lead	4960	5000	99	90 - 110	P	10/28/2025	17:27	LB137677
	Selenium	4970	5000	99	90 - 110	P	10/28/2025	17:27	LB137677
	Silver	1260	1250	101	90 - 110	P	10/28/2025	17:27	LB137677
CCV07	Arsenic	4970	5000	99	90 - 110	P	10/28/2025	18:17	LB137677
	Barium	10200	10000	102	90 - 110	P	10/28/2025	18:17	LB137677
	Cadmium	2520	2500	101	90 - 110	P	10/28/2025	18:17	LB137677
	Chromium	1010	1000	101	90 - 110	P	10/28/2025	18:17	LB137677
	Lead	4950	5000	99	90 - 110	P	10/28/2025	18:17	LB137677
	Selenium	4970	5000	99	90 - 110	P	10/28/2025	18:17	LB137677
	Silver	1260	1250	101	90 - 110	P	10/28/2025	18:17	LB137677
CCV08	Arsenic	5110	5000	102	90 - 110	P	10/28/2025	19:06	LB137677
	Barium	10500	10000	105	90 - 110	P	10/28/2025	19:06	LB137677
	Cadmium	2600	2500	104	90 - 110	P	10/28/2025	19:06	LB137677
	Chromium	1030	1000	103	90 - 110	P	10/28/2025	19:06	LB137677
	Lead	5090	5000	102	90 - 110	P	10/28/2025	19:06	LB137677
	Selenium	5120	5000	102	90 - 110	P	10/28/2025	19:06	LB137677
	Silver	1290	1250	104	90 - 110	P	10/28/2025	19:06	LB137677
CCV09	Arsenic	4990	5000	100	90 - 110	P	10/28/2025	19:58	LB137677
	Barium	10300	10000	104	90 - 110	P	10/28/2025	19:58	LB137677
	Cadmium	2530	2500	101	90 - 110	P	10/28/2025	19:58	LB137677
	Chromium	1010	1000	101	90 - 110	P	10/28/2025	19:58	LB137677
	Lead	4960	5000	99	90 - 110	P	10/28/2025	19:58	LB137677
	Selenium	5000	5000	100	90 - 110	P	10/28/2025	19:58	LB137677
	Silver	1280	1250	102	90 - 110	P	10/28/2025	19:58	LB137677
CCV10	Arsenic	4880	5000	98	90 - 110	P	10/28/2025	20:49	LB137677
	Barium	10100	10000	101	90 - 110	P	10/28/2025	20:49	LB137677
	Cadmium	2480	2500	99	90 - 110	P	10/28/2025	20:49	LB137677
	Chromium	987	1000	99	90 - 110	P	10/28/2025	20:49	LB137677
	Lead	4850	5000	97	90 - 110	P	10/28/2025	20:49	LB137677

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: HDR, Inc.

SDG No.: Q3466

Contract: HDRI08

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV10	Selenium	4880	5000	98	90 - 110	P	10/28/2025	20:49	LB137677
	Silver	1230	1250	99	90 - 110	P	10/28/2025	20:49	LB137677
