



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: HDR, Inc.

SDG No.: Q3466

Contract: HDRI08

Lab Code: ACE

Initial Calibration Source: _____

Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRI01	Arsenic	19.5	20.0	97	65 - 135	P	10/28/2025	12:42	LB137677
	Barium	103	100	103	65 - 135	P	10/28/2025	12:42	LB137677
	Cadmium	5.75	6.0	96	65 - 135	P	10/28/2025	12:42	LB137677
	Chromium	10.4	10.0	104	65 - 135	P	10/28/2025	12:42	LB137677
	Lead	11.5	12.0	96	65 - 135	P	10/28/2025	12:42	LB137677
	Selenium	20.5	20.0	102	65 - 135	P	10/28/2025	12:42	LB137677
	Silver	10.6	10.0	106	65 - 135	P	10/28/2025	12:42	LB137677
CRA	Mercury	0.21	0.2	106	70 - 130	CV	10/28/2025	15:24	LB137673