

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: HDR, Inc.

SDG No.: Q3466

Contract: HDRI08

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB170291BS							
Arsenic	ug/L	4000	3800		95	80 - 120	P
Barium	ug/L	1000	994		99	80 - 120	P
Cadmium	ug/L	1000	937		94	80 - 120	P
Chromium	ug/L	2000	1980		99	80 - 120	P
Lead	ug/L	5000	4620		92	80 - 120	P
Selenium	ug/L	10000	9530		95	80 - 120	P
Silver	ug/L	380	366		96	80 - 120	P

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: HDR, Inc.

SDG No.: Q3466

Contract: HDRI08

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB170292BS Mercury	ug/L	4.0	3.51		88	80 - 120	CV