

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: HDRI08

Lab Code: ACE

SDG NO.: Q3466

Continuing Calib Date: 10/28/2025

Initial Calibration Date(s): 10/09/2025

10/09/2025

Continuing Calib Time: 20:03

Initial Calibration Time(s): 14:14

22:29

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	5.51	5.49	5.39	5.59	-0.02
Aroclor-1016-2	(2)	5.53	5.51	5.41	5.61	-0.02
Aroclor-1016-3	(3)	5.59	5.58	5.48	5.68	-0.01
Aroclor-1016-4	(4)	5.69	5.67	5.57	5.77	-0.02
Aroclor-1016-5	(5)	5.98	5.96	5.86	6.06	-0.02
Aroclor-1260-1	(1)	7.09	7.08	6.98	7.18	-0.01
Aroclor-1260-2	(2)	7.35	7.33	7.23	7.43	-0.02
Aroclor-1260-3	(3)	7.71	7.69	7.59	7.79	-0.01
Aroclor-1260-4	(4)	7.93	7.91	7.81	8.01	-0.02
Aroclor-1260-5	(5)	8.23	8.21	8.11	8.31	-0.02
Tetrachloro-m-xylene		4.37	4.36	4.26	4.46	-0.01
Decachlorobiphenyl		9.96	9.93	9.83	10.03	-0.03

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: HDRI08

Lab Code: ACE

SDG NO.: Q3466

Continuing Calib Date: 10/28/2025

Initial Calibration Date(s): 10/09/2025

10/09/2025

Continuing Calib Time: 20:03

Initial Calibration Time(s): 14:14

22:29

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	4.73	4.74	4.64	4.84	0.01
Aroclor-1016-2	(2)	4.75	4.76	4.66	4.86	0.01
Aroclor-1016-3	(3)	4.93	4.94	4.84	5.04	0.01
Aroclor-1016-4	(4)	4.97	4.98	4.88	5.08	0.01
Aroclor-1016-5	(5)	5.18	5.19	5.09	5.29	0.01
Aroclor-1260-1	(1)	6.21	6.23	6.13	6.33	0.02
Aroclor-1260-2	(2)	6.40	6.41	6.31	6.51	0.01
Aroclor-1260-3	(3)	6.55	6.57	6.47	6.67	0.02
Aroclor-1260-4	(4)	7.02	7.04	6.94	7.14	0.02
Aroclor-1260-5	(5)	7.26	7.28	7.18	7.38	0.02
Tetrachloro-m-xylene		3.65	3.66	3.56	3.76	0.01
Decachlorobiphenyl		8.65	8.67	8.57	8.77	0.02