

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: HDRI08

Lab Code: ACE

SDG NO.: Q3466

Continuing Calib Date: 10/28/2025

Initial Calibration Date(s): 10/21/2025

10/21/2025

Continuing Calib Time: 15:40

Initial Calibration Time(s): 09:06

17:01

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1268-1 (1)	8.87	8.88	8.78	8.98	0.01
Aroclor-1268-2 (2)	8.96	8.97	8.87	9.07	0.01
Aroclor-1268-3 (3)	9.20	9.21	9.11	9.31	0.01
Aroclor-1268-4 (4)	9.63	9.64	9.54	9.74	0.01
Aroclor-1268-5 (5)	10.06	10.07	9.97	10.17	0.01
Tetrachloro-m-xylene	4.64	4.65	4.55	4.75	0.01
Decachlorobiphenyl	10.41	10.42	10.32	10.52	0.01

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: HDRI08

Lab Code: ACE

SDG NO.: Q3466

Continuing Calib Date: 10/28/2025

Initial Calibration Date(s): 10/21/2025

10/21/2025

Continuing Calib Time: 15:40

Initial Calibration Time(s): 09:06

17:01

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1268-1 (1)	7.67	7.67	7.57	7.77	0.00
Aroclor-1268-2 (2)	7.73	7.74	7.64	7.84	0.01
Aroclor-1268-3 (3)	7.94	7.94	7.84	8.04	0.00
Aroclor-1268-4 (4)	8.22	8.23	8.13	8.33	0.01
Aroclor-1268-5 (5)	8.52	8.53	8.43	8.63	0.01
Tetrachloro-m-xylene	3.78	3.79	3.69	3.89	0.01
Decachlorobiphenyl	8.78	8.79	8.69	8.89	0.01