

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** HDRI08
Lab Code: ACE **SDG NO.:** Q3466
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/09/2025 10/09/2025

Client Sample No.: CCAL01 **Date Analyzed:** 10/28/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO114694.D **Time Analyzed:** 09:33

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.489	5.392	5.592	591.510	500.000	18.3
Aroclor-1016-2	5.510	5.414	5.614	563.860	500.000	12.8
Aroclor-1016-3	5.573	5.476	5.676	562.350	500.000	12.5
Aroclor-1016-4	5.668	5.571	5.771	594.810	500.000	19.0
Aroclor-1016-5	5.960	5.863	6.063	598.130	500.000	19.6
Aroclor-1260-1	7.072	6.975	7.175	577.060	500.000	15.4
Aroclor-1260-2	7.326	7.229	7.429	571.520	500.000	14.3
Aroclor-1260-3	7.682	7.585	7.785	552.490	500.000	10.5
Aroclor-1260-4	7.904	7.807	8.007	557.750	500.000	11.6
Aroclor-1260-5	8.211	8.113	8.313	529.700	500.000	5.9
Decachlorobiphenyl	9.929	9.834	10.034	53.140	50.000	6.3
Tetrachloro-m-xylene	4.350	4.255	4.455	58.410	50.000	16.8

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** HDRI08
Lab Code: ACE **SDG NO.:** Q3466
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/09/2025 10/09/2025

Client Sample No.: CCAL01 **Date Analyzed:** 10/28/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO114694.D **Time Analyzed:** 09:33

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.733	4.643	4.843	465.960	500.000	-6.8
Aroclor-1016-2	4.752	4.662	4.862	503.000	500.000	0.6
Aroclor-1016-3	4.927	4.836	5.036	521.470	500.000	4.3
Aroclor-1016-4	4.968	4.879	5.079	530.080	500.000	6.0
Aroclor-1016-5	5.181	5.092	5.292	530.520	500.000	6.1
Aroclor-1260-1	6.212	6.125	6.325	556.560	500.000	11.3
Aroclor-1260-2	6.399	6.313	6.513	576.930	500.000	15.4
Aroclor-1260-3	6.552	6.466	6.666	534.140	500.000	6.8
Aroclor-1260-4	7.023	6.939	7.139	605.710	500.000	21.1
Aroclor-1260-5	7.263	7.179	7.379	583.620	500.000	16.7
Decachlorobiphenyl	8.648	8.571	8.771	50.690	50.000	1.4
Tetrachloro-m-xylene	3.654	3.560	3.760	50.130	50.000	0.3