

### CALIBRATION VERIFICATION SUMMARY

**Lab Name:** Alliance **Contract:** HDRI08  
**Lab Code:** ACE **SDG NO.:** Q3466  
**GC Column:** ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/21/2025 10/21/2025

**Client Sample No.:** CCAL03 **Date Analyzed:** 10/28/2025  
**Lab Sample No.:** AR1660CCC500 **Data File :** PP076126.D **Time Analyzed:** 14:35

| COMPOUND             | RT     | RT WINDOW |        | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D   |
|----------------------|--------|-----------|--------|--------------------|-------------------|------|
|                      |        | FROM      | TO     |                    |                   |      |
| Aroclor-1016-1       | 5.790  | 5.695     | 5.895  | 515.540            | 500.000           | 3.1  |
| Aroclor-1016-2       | 5.811  | 5.717     | 5.917  | 514.030            | 500.000           | 2.8  |
| Aroclor-1016-3       | 5.874  | 5.780     | 5.980  | 515.830            | 500.000           | 3.2  |
| Aroclor-1016-4       | 5.971  | 5.877     | 6.077  | 517.190            | 500.000           | 3.4  |
| Aroclor-1016-5       | 6.263  | 6.169     | 6.369  | 506.700            | 500.000           | 1.3  |
| Aroclor-1260-1       | 7.381  | 7.287     | 7.487  | 549.360            | 500.000           | 9.9  |
| Aroclor-1260-2       | 7.633  | 7.540     | 7.740  | 521.250            | 500.000           | 4.3  |
| Aroclor-1260-3       | 7.992  | 7.899     | 8.099  | 543.700            | 500.000           | 8.7  |
| Aroclor-1260-4       | 8.219  | 8.127     | 8.327  | 558.900            | 500.000           | 11.8 |
| Aroclor-1260-5       | 8.544  | 8.452     | 8.652  | 522.930            | 500.000           | 4.6  |
| Decachlorobiphenyl   | 10.407 | 10.319    | 10.519 | 51.990             | 50.000            | 4.0  |
| Tetrachloro-m-xylene | 4.639  | 4.542     | 4.742  | 51.490             | 50.000            | 3.0  |

### CALIBRATION VERIFICATION SUMMARY

**Lab Name:** Alliance **Contract:** HDRI08  
**Lab Code:** ACE **SDG NO.:** Q3466  
**GC Column:** ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/21/2025 10/21/2025

**Client Sample No.:** CCAL03 **Date Analyzed:** 10/28/2025  
**Lab Sample No.:** AR1660CCC500 **Data File :** PP076126.D **Time Analyzed:** 14:35

| COMPOUND             | RT    | RT WINDOW |       | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D   |
|----------------------|-------|-----------|-------|--------------------|-------------------|------|
|                      |       | FROM      | TO    |                    |                   |      |
| Aroclor-1016-1       | 4.871 | 4.778     | 4.978 | 546.360            | 500.000           | 9.3  |
| Aroclor-1016-2       | 4.931 | 4.836     | 5.036 | 529.640            | 500.000           | 5.9  |
| Aroclor-1016-3       | 5.050 | 4.954     | 5.154 | 509.890            | 500.000           | 2.0  |
| Aroclor-1016-4       | 5.091 | 4.996     | 5.196 | 562.330            | 500.000           | 12.5 |
| Aroclor-1016-5       | 5.305 | 5.210     | 5.410 | 515.240            | 500.000           | 3.0  |
| Aroclor-1260-1       | 6.332 | 6.238     | 6.438 | 546.780            | 500.000           | 9.4  |
| Aroclor-1260-2       | 6.519 | 6.425     | 6.625 | 556.220            | 500.000           | 11.2 |
| Aroclor-1260-3       | 6.672 | 6.578     | 6.778 | 542.250            | 500.000           | 8.5  |
| Aroclor-1260-4       | 7.140 | 7.047     | 7.247 | 548.730            | 500.000           | 9.7  |
| Aroclor-1260-5       | 7.381 | 7.287     | 7.487 | 540.840            | 500.000           | 8.2  |
| Decachlorobiphenyl   | 8.774 | 8.682     | 8.882 | 46.520             | 50.000            | -7.0 |
| Tetrachloro-m-xylene | 3.779 | 3.681     | 3.881 | 56.460             | 50.000            | 12.9 |