

Surrogate Summary

SDG No.: Q3466

Client: HDR, Inc.

Analytical Method: 8081B

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Recovery(%)	Qual	Limits(%)	
								Low	High
I.BLK-PD090647.D	PIBLK-PD090647.D	Decachlorobiphen	1	20	21.3	106		30 (57)	150 (171)
		Tetrachloro-m-xyl	1	20	19.4	97		30 (61)	150 (148)
		Decachlorobiphen	2	20	20.3	102		30 (57)	150 (171)
		Tetrachloro-m-xyl	2	20	20.0	100		30 (61)	150 (148)
I.BLK-PD090879.D	PIBLK-PD090879.D	Decachlorobiphen	1	20	23.1	115		30 (57)	150 (171)
		Tetrachloro-m-xyl	1	20	21.6	108		30 (61)	150 (148)
		Decachlorobiphen	2	20	24.9	125		30 (57)	150 (171)
		Tetrachloro-m-xyl	2	20	21.8	109		30 (61)	150 (148)
PB170283BL	PB170283BL	Decachlorobiphen	1	20	20.5	103		30 (20)	150 (144)
		Tetrachloro-m-xyl	1	20	19.2	96		30 (19)	150 (148)
		Decachlorobiphen	2	20	22.2	111		30 (20)	150 (144)
		Tetrachloro-m-xyl	2	20	19.2	96		30 (19)	150 (148)
PB170283BS	PB170283BS	Decachlorobiphen	1	20	19.5	98		30 (20)	150 (144)
		Tetrachloro-m-xyl	1	20	14.8	74		30 (19)	150 (148)
		Decachlorobiphen	2	20	21.8	109		30 (20)	150 (144)
		Tetrachloro-m-xyl	2	20	14.9	75		30 (19)	150 (148)
I.BLK-PD090891.D	PIBLK-PD090891.D	Decachlorobiphen	1	20	22.8	114		30 (57)	150 (171)
		Tetrachloro-m-xyl	1	20	21.5	108		30 (61)	150 (148)
		Decachlorobiphen	2	20	23.0	115		30 (57)	150 (171)
		Tetrachloro-m-xyl	2	20	21.6	108		30 (61)	150 (148)
Q3466-02	B8-0.0-1.0-20251023	Decachlorobiphen	1	20	15.8	79		30 (20)	150 (144)
		Tetrachloro-m-xyl	1	20	19.8	99		30 (19)	150 (148)
		Decachlorobiphen	2	20	16.4	82		30 (20)	150 (144)
		Tetrachloro-m-xyl	2	20	20.5	102		30 (19)	150 (148)
Q3466-02MS	B8-0.0-1.0-20251023MS	Decachlorobiphen	1	20	15.1	75		30 (20)	150 (144)
		Tetrachloro-m-xyl	1	20	18.1	90		30 (19)	150 (148)
		Decachlorobiphen	2	20	15.1	75		30 (20)	150 (144)
		Tetrachloro-m-xyl	2	20	19.1	96		30 (19)	150 (148)
Q3466-02MSD	B8-0.0-1.0-20251023MSD	Decachlorobiphen	1	20	14.8	74		30 (20)	150 (144)
		Tetrachloro-m-xyl	1	20	19.3	96		30 (19)	150 (148)
		Decachlorobiphen	2	20	15.2	76		30 (20)	150 (144)
		Tetrachloro-m-xyl	2	20	20.3	102		30 (19)	150 (148)
Q3466-05	B9-0.0-1.0-20251023	Decachlorobiphen	1	20	12.8	64		30 (20)	150 (144)
		Tetrachloro-m-xyl	1	20	18.6	93		30 (19)	150 (148)
		Decachlorobiphen	2	20	14.4	72		30 (20)	150 (144)
		Tetrachloro-m-xyl	2	20	18.6	93		30 (19)	150 (148)
Q3466-07	B7-1.0-2.0-20251023	Decachlorobiphen	1	20	11.7	59		30 (20)	150 (144)
		Tetrachloro-m-xyl	1	20	13.4	67		30 (19)	150 (148)
		Decachlorobiphen	2	20	9.81	49		30 (20)	150 (144)
		Tetrachloro-m-xyl	2	20	13.8	69		30 (19)	150 (148)
I.BLK-PD090900.D	PIBLK-PD090900.D	Decachlorobiphen	1	20	18.7	94		30 (57)	150 (171)

Surrogate Summary

SDG No.: Q3466

Client: HDR, Inc.

Analytical Method: 8081B

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Recovery(%)	Qual	Limits(%)	
								Low	High
I.BLK-PD090900.D	PIBLK-PD090900.D	Tetrachloro-m-xyl	1	20	19.4	97		30 (61)	150 (148)
		Decachlorobiphen	2	20	17.5	87		30 (57)	150 (171)
		Tetrachloro-m-xyl	2	20	21.3	107		30 (61)	150 (148)