

Instrument ID: **BNA_G**

Daily Analysis Runlog For Sequence/QC Batch ID # BG102925

Review By	rahul	Review On	10/30/2025 8:51:46 AM		
Supervise By	Jagrut	Supervise On	10/30/2025 4:39:08 PM		
SubDirectory	BG102925	HP Acquire Method	BNA_G	HP Processing Method	BG102425
STD. NAME		STD REF.#			
Tune/Reschk		SP6875			
Initial Calibration Stds		SP6857,SP6858,SP6859,SP6860,SP6861,SP6862,SP6863,SP6864			
CCC		SP6860			
Internal Standard/PEM		S13180,10ul/1000ul sample			
ICV/I.BLK		SP6872			
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	DFTPP	BG064617.D	29 Oct 2025 10:41	RC/JU	Ok
2	SSTDCCC040	BG064618.D	29 Oct 2025 11:22	RC/JU	Ok
3	PB170285BL	BG064619.D	29 Oct 2025 12:03	RC/JU	Ok
4	PB170285BS	BG064620.D	29 Oct 2025 12:44	RC/JU	Ok,M
5	PB170215BL	BG064621.D	29 Oct 2025 13:25	RC/JU	Ok
6	PB170215BS	BG064622.D	29 Oct 2025 14:06	RC/JU	Ok,M
7	Q3477-01	BG064623.D	29 Oct 2025 14:50	RC/JU	Ok
8	Q3477-01MS	BG064624.D	29 Oct 2025 15:31	RC/JU	Ok,M
9	Q3477-01MSD	BG064625.D	29 Oct 2025 16:12	RC/JU	Ok,M
10	Q3419-01	BG064626.D	29 Oct 2025 16:53	RC/JU	Ok
11	Q3419-02MS	BG064627.D	29 Oct 2025 17:34	RC/JU	Ok,M
12	Q3419-03MSD	BG064628.D	29 Oct 2025 18:15	RC/JU	Ok,M
13	Q3468-01	BG064629.D	29 Oct 2025 18:56	RC/JU	Ok,M
14	Q3468-02	BG064630.D	29 Oct 2025 19:37	RC/JU	Ok,M
15	Q3468-03	BG064631.D	29 Oct 2025 20:18	RC/JU	Ok
16	Q3468-04	BG064632.D	29 Oct 2025 20:59	RC/JU	Ok
17	Q3468-05	BG064633.D	29 Oct 2025 21:40	RC/JU	Ok
18	Q3468-06	BG064634.D	29 Oct 2025 22:21	RC/JU	Ok

M : Manual Integration