

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: Alliance

Contract: HDRI08

Lab Code: ACE

SDG NO.: Q3466

Lab File ID: BP026038.D

DFTPP Injection Date: 10/30/2025

Instrument ID: BNA_P

DFTPP Injection Time: 10:34

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
68	Less than 2.0% of mass 69	0.5 (1.7) 1
69	Mass 69 relative abundance	31
70	Less than 2.0% of mass 69	0.2 (0.6) 1
197	Less than 2.0% of mass 198	0.3
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.6
365	Greater than 1% of mass 198	3.7
441	Present, but less than mass 443	13.9
442	Greater than 50% of mass 198	88.8
443	15.0 - 24.0% of mass 442	17.2 (19.3) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC040	SSTDCCC040	BP026039.D	10/30/2025	11:15
B9-0.0-1.0-20251023	Q3466-05	BP026050.D	10/30/2025	20:30
B7-1.0-2.0-20251023	Q3466-07	BP026051.D	10/30/2025	21:11