

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: HDRI08
 Lab Code: ACE SDG NO.: Q3466
 Lab File ID: VY023585.D Date Analyzed: 10/27/2025
 Instrument ID: MSVOA_Y Time Analyzed: 09:12
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS4 AREA #	RT #				
12 HOUR STD	333946	13.346				
UPPER LIMIT	667892	13.846				
LOWER LIMIT	166973	12.846				
EPA SAMPLE NO.						
B8-0.5-1.0-20251023	392869	13.35				
B9-0.5-1.0-20251023	361763	13.35				
B7-1.5-2.0-20251023	398135	13.35				
VY1027SBL01	410194	13.35				
VY1027SBS01	345467	13.35				
VY1027SBSD01	325402	13.35				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.