

NEW JERSEY LAB ID#:20012 : NEW YORK LAB ID#: 11376

GC/MS SEMI-VOLATILE ANALYSIS CONFORMANCE/NON-CONFORMANCE SUMMARY

CHEMTECH PROJECT NUMBER: BG103025

SequenceID : BG103025

	NA	NO	YES
1. Chromatograms Labeled/Compounds Identified. (Field samples and Method Blanks)	_____	_____	_____✓_____
2. GC/MS Tuning Specifications. DFTPP Meet Criteria Criteria (NOTE THAT THERE ARE DIFFERENT CRITERIA FOR NY ASP CLP, CLP AND NJ)	_____	_____	_____✓_____
3. GC/MS Tuning Frequency - Performed every 24 hours for 600 series and 12 hours for 8000 series	_____	_____	_____✓_____
4. GC/MS Calibration - Initial Calibration performed within 30 days before sample analysis and continuing calibration performed within 24 hours of sample analysis for 600 series and 12 hours for 8000 series	_____	_____	_____✓_____
5. GC/MS Calibration Met:	_____	_____	_____✓_____
a. Initial calibration Meet Criteria	_____	_____	_____✓_____
If not met, list those compounds and their recoveries which fall outside the acceptable range.	_____	_____	_____✓_____
b. Continuous Calibration(CCC) Meet Criteria	_____	_____✓_____	_____✓_____
If not met, list those compounds and their recoveries which fall outside the acceptable range.	_____	_____✓_____	_____✓_____
Compound Nitrobenzene-d5			
Surrogate,Nitrobenzene,2-Nitrophenol,Hexachlorocyclopentadiene,2,4-Dinitrophenol,4,6-Dinitro-2-methylphenol failed from			
6. Blank Contamination - If yes, list compounds and concentrations in each blank:	_____	_____	_____✓_____
a. B/N Fraction			
d. Acid Fraction			

7. Surrogate Recoveries Meet Criteria			✓
If not met, list those compounds and their recoveries which fall outside the acceptable ranges.			
a. B/N Fraction			
d. Acid Fraction			
8. Matrix Spike/Matrix Spike Duplicate Recoveries Meet Criteria	✓		
If not met, list those compounds and their recoveries which fall outside the acceptable range.			
a. B/N Fraction			
d. Acid Fraction			
9. Internal Standard Area/Retention Time Shift Meet Criteria			✓
Comments:			
10. Extraction Holding Time Met			✓
If not met, list number of days exceeded for each sample:			
11. Analysis Holding Time Met			✓
If not met, list number of days exceeded for each sample:			

ADDITIONAL COMMENTS:

Recovery of only one compound bis(2-Chloroethyl)ether is marginally biased low in PB170286BS & PB170286BSD, and Nitrobenzene-d5 Surrogate in PB170306BL was marginally biased high. The data will be used for the hard copies.

dhruv

 Analyst

10/30/2025

 Date

NEW JERSEY LAB ID#:20012 : NEW YORK LAB ID#: 11376

GC/MS SEMI-VOLATILE ANALYSIS CONFORMANCE/NON-CONFORMANCE SUMMARY

CHEMTECH PROJECT NUMBER: BG103025

SequenceID : BG103025 NA NO YES