

Surrogate Summary

SW-846

SDG No.: Q3468

Client: LiRo Engineers, Inc.

Analytical Method: 8270E

Lab Sample ID	Client ID	Parameter	Spike (PPM)	Result (PPM)	Recovery (%)	Qual	Limits (%)	
							Low	High
PB170286BL	PB170286BL	Nitrobenzene-d5	100	102	102		67	132
		2-Fluorobiphenyl	100	87.4	87		52	132
		Terphenyl-d14	100	79.4	79		42	152
PB170286BS	PB170286BS	Nitrobenzene-d5	100	93.9	94		67	132
		2-Fluorobiphenyl	100	79.9	80		52	132
		Terphenyl-d14	100	79.5	79		42	152
PB170286BSD	PB170286BSD	Nitrobenzene-d5	100	97.0	97		67	132
		2-Fluorobiphenyl	100	79.8	80		52	132
		Terphenyl-d14	100	80.6	81		42	152
Q3468-01	MW-06	Nitrobenzene-d5	100	120	120		67	132
		2-Fluorobiphenyl	100	95.3	95		52	132
		Terphenyl-d14	100	96.2	96		42	152
Q3468-02	MW-08	Nitrobenzene-d5	100	118	118		67	132
		2-Fluorobiphenyl	100	85.9	86		52	132
		Terphenyl-d14	100	87.9	88		42	152
Q3468-03	MW-10	Nitrobenzene-d5	100	132	132		67	132
		2-Fluorobiphenyl	100	110	110		52	132
		Terphenyl-d14	100	104	104		42	152
Q3468-04	MW-11	Nitrobenzene-d5	100	123	123		67	132
		2-Fluorobiphenyl	100	101	101		52	132
		Terphenyl-d14	100	92.9	93		42	152
Q3468-05	MW-13	Nitrobenzene-d5	100	113	113		67	132
		2-Fluorobiphenyl	100	94.7	95		52	132
		Terphenyl-d14	100	90.7	91		42	152
Q3468-06	MW-12	Nitrobenzene-d5	100	117	117		67	132
		2-Fluorobiphenyl	100	86.3	86		52	132
		Terphenyl-d14	100	92.3	92		42	152