

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q3468

Analytical Method: 8270E

Client: LiRo Engineers, Inc.

DataFile: BG064639.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	RPD	Limits		RPD
								Qual	Low	High	
PB170286BSD	Acenaphthylene	50	44.5	ug/L	89	0			79	103	20
	Acenaphthene	50	45.3	ug/L	91	0			59	113	20
	Fluorene	50	40.0	ug/L	80	2			64	107	20
	Phenanthrene	50	40.3	ug/L	81	0			62	109	20
	Anthracene	50	40.1	ug/L	80	0			65	110	20
	Fluoranthene	50	41.3	ug/L	83	1			64	110	20
	Pyrene	50	37.9	ug/L	76	0			71	103	20
	Benzo(a)anthracene	50	40.7	ug/L	81	0			62	107	20
	Chrysene	50	40.5	ug/L	81	1			61	108	20
	Benzo(b)fluoranthene	50	43.9	ug/L	88	1			77	113	20
	Benzo(k)fluoranthene	50	43.0	ug/L	86	0			77	105	20
	Benzo(a)pyrene	50	44.8	ug/L	90	1			72	131	20
	Indeno(1,2,3-cd)pyrene	50	43.3	ug/L	87	0			72	105	20
	Dibenz(a,h)anthracene	50	44.4	ug/L	89	2			78	115	20
	Benzo(g,h,i)perylene	50	44.6	ug/L	89	1			75	118	20